

A novel high-throughput scanning microscope for label-free detection of protein and small-molecule chemical microarrays

Y. Y. Fei, J. P. Landry, Y. S. Sun, and X. D. Zhu^{a)}

Department of Physics, University of California at Davis, Davis, California 95616, USA

J. T. Luo, X. B. Wang, and K. S. Lam

Division of Hematology and Oncology, Department of Internal Medicine, UC Davis Cancer Center, University of California at Davis, Sacramento, California 95817, USA

(Received 31 August 2007; accepted 10 December 2007; published online 17 January 2008)

We describe a novel scanning optical microscope based on a polarization-modulated nulling ellipsometry. The new microscope employs a combination of scanning mirror and sample translation and thus enables high-throughput label-free detection of biomolecular microarrays with more than 10 000 protein or small-molecule targets. For illustration, we show the image of a 2760-spot protein microarray on a functionalized glass slide obtained with such a microscope. The new scanning microscope is also capable of determining, in parallel, the real-time binding kinetics of multiple molecular species under aqueous conditions. © 2008 American Institute of Physics.

[DOI: [10.1063/1.2830286](https://doi.org/10.1063/1.2830286)]

I. INTRODUCTION

Microarrays of surface-immobilized macromolecules or chemical ligands are powerful tools for *in vitro* analysis of multiple biochemical binding events in a single experiment.^{1,2} They are particularly useful for proteomics and glycomics research as the number and variety of structures and functions of proteins, glycans, and their ligands are large. The cooperative effect in a probe-target binding reaction due to the presence of other molecules further increases the complexity of proteomic and glycomic studies. Microarray technology in various forms enables parallel detection of hundreds or thousands of biochemical binding processes, thus promises to shorten the discovery and analysis time. In most cases the microarray technology is associated with some form of fluorescence detection such that molecular probes are labeled with fluorescent tags either extrinsically (e.g., with Alexa 480) or “intrinsically” through genetic engineering (e.g., incorporation of green fluorescent protein or phytochrome proteins).^{3–5} The fluorescence detection is widely used for its superior sensitivity and low background. The choices of fluorescent tags are many, some being nearly free of photobleaching.

However, fluorescent labeling has its shortcomings. High cost and variation in labeling efficiency are well known. For proteomics and glycomics studies, tagging molecules of interest with fluorescent agents inevitably changes the properties of host molecules. For example, the rate of incorporating deoxyribonucleotide triphosphates (dNTP such as deoxycytidine triphosphate or dCTP, deoxy-adenosine triphosphate or dATP, etc.) by high-fidelity polymerases is reduced by more than an order of magnitude when dNTP's are fluorescently labeled with Cy3 or Cy5 molecules. The binding affinity of a protein to a small molecule or peptide ligand may

also be significantly reduced when the protein is labeled by a fluorescent molecule such as Cy3. Often such an effect of fluorescent labeling is not known or characterized *a priori* due to lack of comparative label-free studies.^{6,7} It is thus important and necessary to develop label-free optical methods for high-throughput detection of microarrays.

II. NOVEL SCANNING OPTICAL MICROSCOPE BASED ON DETECTION OF OBLIQUE-INCIDENCE REFLECTIVITY DIFFERENCE

At oblique incidence a surface-bound change such as binding of molecules in liquid phase to surface-immobilized molecular targets alters the ratio of reflection coefficients, $r_p/r_s = \tan \psi \exp(i\delta)$, for *p*-polarized (transverse magnetic mode) and *s*-polarized (transverse electric mode) components of a monochromatic light beam. Optical ellipsometry in one form or another measures Ψ and δ , and thus can be used to study the surface-bound process. Based on the detection of oblique-incidence optical reflectivity difference (OI-RD), a special form of polarization-modulated nulling ellipsometry, we have developed an optical scanning microscope for label-free, high-throughput detection of microarrays over large areas (equivalent of the field of view). In this report, we describe the optical arrangement and performance of a scanning OI-RD microscope capable of *in situ* detecting protein, peptide, and small-molecule microarrays with thousands of target spots immobilized on conventional microscope glass slides. The microscope readily scans over 10 000 distinct molecular target spots and is capable of real-time kinetic as well as endpoint measurements of thousands of binding reactions.

Compared to imaging surface plasmon resonance (SPR) scanners,^{8–11} an OI-RD microscope does not depend on gold-coated substrates for detection and can have a “field of view” as large as $2 \times 5 \text{ cm}^2$. We should note that microarrays with

^{a)}Electronic mail: xdzhu@physics.ucdavis.edu.

up to 10 000–15 000 spots on a single solid support (e.g., a 1×3 in.² glass slide) are suitable high-throughput platforms for proteomic and glycomic studies or for drug lead optimization. This is because varieties of proteins and glycans and small-molecule ligand candidates are in thousands.

Compared to imaging ellipsometers based on the scheme polarizer-compensator-sample-analyzer (PCSA)^{12–14} the OI-RD microscope is inherently more sensitive to surface-bound changes (e.g., thickness, density, etc.) by at least one order of magnitude. In a typical PCSA imaging ellipsometer, the phase compensator (C) is fixed while the polarizer (P) and the analyzer (A) are adjusted. The reflected light beam from an illuminated area on the sample surface (S) passes through the analyzer (A) and forms an image on a charge coupled device (CCD) camera. To yield ψ_0 and δ_0 maps of the illuminated region before, for example, a molecular binding reaction takes place on the sample surface, P and A are adjusted until the photocurrents are minimized over the image region on the CCD. During the subsequent binding reaction, the thickness of a surface-bound layer changes by Δd . This causes ψ and δ to deviate from ψ_0 and δ_0 by $\Delta\psi \sim \Delta d/\lambda$ and $\Delta\delta \sim \Delta d/\lambda$. However, the corresponding change in photocurrent under this off-null condition is proportional to $(\Delta\psi)^2$ and/or $(\Delta\delta)^2$. The quadratic dependence of the off-null photocurrent on already small quantities $\Delta\psi$ and $\Delta\delta$ sets the detection limit of this type of imaging ellipsometers to roughly $\Delta\psi \sim 0.01^\circ$ and $\Delta\delta \sim 0.01^\circ$ (i.e., ~ 0.0002 rad).

The OI-RD technique is a more sensitive form of ellipsometry.^{15–17} It is a polarization-modulated nulling ellipsometry in which the measurable harmonics of modulated photocurrents are directly proportional $\Delta\psi$ and $\Delta\delta$. This makes an imaging ellipsometer based on OI-RD signals at least an order of magnitude more sensitive than a typical PCSA imaging ellipsometer, namely, with the detection limit to $\Delta\psi \sim 0.001^\circ$ and $\Delta\delta \sim 0.001^\circ$. Such sensitivity is needed for high-throughput affinity detection of low molecular weight analytes. The detection limit of $\Delta\delta \sim 0.001^\circ$ corresponds to 0.01 nm in detected protein thickness, similar to that of an imaging SPR microscope.^{8–11}

The optical arrangement of our new scanning OI-RD microscope is sketched in Fig. 1. It is a dual-axis mechanical scanning microscope. The key difference between this scanning microscope and our previous OI-RD microscopes is that the sample-holding stage in the present microscope only moves in one direction for x -scan and the y -scan is accomplished with a combination of a scan mirror and an f -theta lens. Rotating a low-mass scan mirror to achieve a y -scan of 20 mm on the focal plane of the f -theta lens takes much less time than moving the sample-holding stage. To illustrate the concept of such a hybrid scanning microscope and the performance in terms of field of view and image quality of protein microarrays, we mount a metal-coated mirror on a backlash-free rotation stage (from Physik Instruments, Germany) driven by a custom stepping motor.

Changes in surface mass density and/or thickness of a thin nonabsorbing layer of molecules on a transparent substrate only alter δ significantly. Our OI-RD microscope measures the change in δ as follows. For illumination, we use a 30 mW, linearly polarized, continuous-wave neodymium




FIG. 1. (a) Top view of a hybrid scanning OI-RD microscope with a combination of y -scan by a combination of a RM and a FTL and x -scan by translation of the sample-holding stage. The sample is a biomolecular microarray printed on a glass slide. The microarray-bearing surface is in contact with an aqueous solution as a part of a fluidic handling system. (b) Side view of the microscope that illustrates the y -scan. PEM: photoelastic modulator. PS: phase shifter. FM: fixed mirror. OBJ: objective lens. A: analyzer. PD: photodiode detector.

doped yttrium aluminum garnet laser operated at $\lambda = 532$ nm. The beam intensity I_{inc} is stable within $\pm 0.4\%$ over at least 2 h. The beam is initially s -polarized. As shown in Fig. 1(a) it passes through a photoelastic modulator (PEM) (Hinds Instruments, OR) so that the beam emerging after the modulator changes from being s -polarized to p -polarized at a frequency Ω . The PEM is made of a fused quartz slab driven by a transducer at $\Omega = 50$ kHz (a mechanical frequency of the slab). The resultant beam then passes through a phase shifter (PS) that alters the relative phase between the s - and p -polarized components by a variable amount Φ_{PS} for nulling ellipsometry. The phase shifter here is a rotatable multi-order crystalline quartz half-wave plate for $\lambda = 532$ nm. It is mounted on a precision rotation stage with the rotation axis coinciding with the fast axis of the wave plate. We align the fast axis of the wave plate parallel to the s -polarized component of the laser beam and the slow axis parallel to the p -polarized component of the beam. By rotating the wave plate about the fast axis, we variably change the path length of the laser beam and in turn the phase between the two polarization components. We then use an assembly of a rotating mirror (RM) and a f -theta lens (FTL) to focus the beam into a spot (~ 10 μm in diameter) on the microarray-bearing surface of a glass slide (with optical dielectric constant ϵ_0) at incident angle $\phi_{\text{inc}} = 34.7^\circ$. The microarray-bearing surface is in contact with an aqueous solution (characterized by optical dielectric constant ϵ_s). The reflected beam from the illuminated spot passes through an analyzer (A) with its transmission axis set at θ_A (45° in the present study) from p -polarization. The beam is then imaged with an objective lens onto a single-element Si p - i - n photodiode operated in photoconductive mode. The photodiode converts




FIG. 2. Image of a 2760-spot BSA microarray acquired with the hybrid scanning OI-RD microscope as illustrated in Fig. 1. The image area is $18.3 \times 6.4 \text{ mm}^2$ and the pixel dimension is $17.8 \times 18.7 \text{ } \mu\text{m}^2$. The image contrast has been converted from the optical signal to the surface mass density of surface-bound BSA molecules.

the light intensity incident onto its surface, $I(t)$, into a photoelectric current. Through a current-to-voltage converter with a gain resistor $R=5 \times 10^4 \text{ } \Omega$, the photocurrent is detected by an SR830 Digital lock-in amplifier (Stanford Research Systems, CA). The light intensity $I(t)$ consists of harmonics of polarization-modulation frequency Ω . In a typical OI-RD experiment, we measure the first and second harmonics. The amplitude of the first harmonic is given by $I(\Omega) = I_{\text{inc}} |r_p r_s| \sin 2\theta_A \sin(\eta_{\text{sys}} + \delta + \Phi_{\text{PS}})$.¹⁶ The second harmonic is given by $I(2\Omega) = I_{\text{inc}} (|r_p|^2 \cos^2 \theta_A - |r_s|^2 \sin^2 \theta_A)$.¹⁶ η_{sys} is the phase difference between the p -polarized and s -polarized components of the light beam contributed by optical components in the beam path other than the sample surface. δ is the phase difference due to the reflection from the microarray-bearing glass surface alone.

Let δ_0 be the value for the bare glass slide surface. We adjust Φ_{PS} until $\eta_{\text{sys}} + \delta_0 + \Phi_{\text{PS}} = 0$ on the bare surface so that $I(\Omega) = 0$. When a thin layer of molecules (characterized by

thickness d and optical dielectric constant ϵ_d) is added to the bare surface or when the region of the surface covered with microarray features (again characterized by a thickness d and an optical dielectric constant ϵ_d) is moved under the illumination optics, the first harmonic becomes finite, namely, $I(\Omega) = I_{\text{inc}} |r_p r_s| \sin 2\theta_A \sin(\delta - \delta_0)$. We determine $I_{\text{inc}} |r_p r_s| \sin 2\theta_A \equiv I_{\text{max}}(\Omega)$ separately by adjusting Φ_{PS} until $I(\Omega) = I_{\text{inc}} |r_p r_s| \sin 2\theta_A \sin(\eta_{\text{sys}} + \delta + \Phi_{\text{PS}})$ reaches its maximum. The ratio of $I_{\text{inc}} |r_p r_s| \sin 2\theta_A \sin(\delta - \delta_0)$ to $I_{\text{max}}(\Omega)$ yields $\sin(\delta - \delta_0) \approx \delta - \delta_0 \equiv \Delta\delta$.

When the thickness d of the interrogated microarray feature or the added molecular layer is much less than the wavelength λ , $\Delta\delta$ is related to the properties of the surface-bound layer by¹⁵⁻¹⁷

$$\Delta\delta \cong - \frac{4\pi\sqrt{\epsilon_s\epsilon_0}}{(\epsilon_s - \epsilon_0)(\epsilon_s/\epsilon_0 - \cot^2 \phi_{\text{inc}})} \frac{(\epsilon_d - \epsilon_0)(\epsilon_d - \epsilon_s)}{\epsilon_d} \left(\frac{d}{\lambda}\right) \Theta. \quad (1)$$

Here Θ is the fractional surface coverage of the molecules, defined as the area covered by the molecules divided by the total area available. Θ can be determined experimentally by exposing the microarray-covered glass surface to a solution of bovine serum albumin (BSA). The additional change in $\Delta\delta$ from the originally Θ -covered region is proportional to $1 - \Theta$ while the signal from the originally uncovered region is proportional to 1. From the two values of $\Delta\delta$, one extracts Θ . With our current scan microscope, we can detect a submonolayer of protein molecules such as bovine serum albumin at a level as small as $\Theta = 0.005$ or $1/200$ of a monolayer.

From Eq. (1) it is clear that the first-harmonic $I(\Omega)$, after the initial nulling step, increases linearly with d/λ . In contrast, when the ambient, the substrate, and the molecular layer are all transparent to the laser beam (as in our present study), $I(2\Omega)$ changes quadratically with d/λ and is thus negligible.¹⁷

An image in $\Delta\delta$ of a biomolecular microarray on a glass slide is obtained by scanning the illumination beam in y -direction with the rotating mirror as shown in Fig. 1(b) and moving the sample stage in x -direction. The incremental angle for the rotating mirror driven by a stepping motor is as small as $42.5 \text{ } \mu\text{rad}$ in our present microscope. Together with the f -theta lens, this gives a rise to a linear step of $4.68 \text{ } \mu\text{m}$ in y -direction on the sample surface (on the focal plane of the f -theta lens). In actual scan, we use a step size of $18.7 \text{ } \mu\text{m}$ in order to reduce the scan time and to match the beam diameter on the sample surface ($\sim 10 \text{ } \mu\text{m}$).

III. EXPERIMENTAL RESULTS

For target microarrays, we use rabbit immunoglobulin-G (IgG), human IgG, mouse IgG, BSA (all purchased from Jackson Immuno Research Laboratories), and biotin-BSA conjugates that we have synthesized in house.¹⁸ The choice of biotin-BSA conjugates as a target is motivated by our intent to apply the OI-RD microscope to high-throughput screening small-molecule libraries for protein ligands. Our strategy is to conjugate small molecules to carrier proteins such as BSA through a flexible linker and then print small-molecule-BSA conjugate microarrays on an epoxy-

functionalized glass surface. The conjugate binds to the epoxy groups on the glass surface through amine residues on the surface of BSA, leaving the small-molecule target accessible to protein probes in subsequent analysis. In our case more than one biotin molecule is attached to one BSA molecule. All the target molecules, in $1 \times \text{PBS}$ (phosphate buffered saline) at same concentration of $7.6 \mu\text{M}$, are printed on an epoxy-functionalized glass slide (CEL Associates, TX) with a contact-printing robot (OmniGrid 100, Genomic Solutions, MI). The center-to-center separation between neighboring spots is $200 \mu\text{m}$. The spots are more or less circular with an average diameter of $\sim 100 \mu\text{m}$. After printing, the microarray-bearing slide is washed with $1 \times \text{PBS}$ ($\text{pH}=7.4$) to remove unbound protein targets. The slide is then assembled into a fluid-handling system filled with $1 \times \text{PBS}$ so that the microarray is examined *in situ* in either the buffer solution or a probe solution during subsequent reactions with the scanning OI-RD microscope. After the excess printed

target materials are washed away, the microarray spots are optically flat on the glass substrate. There are no observable edge effects when these spots are imaged with our scanning microscope.

We first demonstrate that the new scanning microscope is capable of imaging large microarrays under aqueous conditions. Figure 2 shows the image of a 2760-spot BSA microarray immersed in a $1 \times \text{PBS}$ buffer solution after the washing step. The microarray covers an area of $1.83 \times 0.64 \text{ cm}^2$, total of $\sim 1 \text{ cm}^2$. The scan steps are $17.8 \mu\text{m}$ in x -direction and $18.7 \mu\text{m}$ in y -direction, the dimension of one image pixel is $17.8 \times 18.7 \mu\text{m}^2$. If we reduce the step size by a factor of 2, this OI-RD microscope can scan a protein microarray four times as dense over the same area, with over 11 000 spots. In Fig. 2 we have converted the optical signal $\Delta\delta$ to the surface mass density Γ using the following procedure. Using the fact that $\Gamma_{\text{BSA}}/\rho_{\text{BSA}} = \Theta_{\text{BSA}} d_{\text{BSA}}$, we rewrite Eq. (1) into

$$\begin{aligned} \Delta\delta &\cong - \frac{4\pi\sqrt{\varepsilon_s}\varepsilon_0}{(\varepsilon_s - \varepsilon_0)(\varepsilon_s/\varepsilon_0 - \cot^2\phi_{\text{inc}})} \frac{(\varepsilon_{\text{BSA}} - \varepsilon_s)(\varepsilon_{\text{BSA}} - \varepsilon_0)}{\varepsilon_{\text{BSA}}} \left(\frac{d_{\text{BSA}}\Theta_{\text{BSA}}}{\lambda} \right) \\ &= - \frac{4\pi\sqrt{\varepsilon_s}\varepsilon_0}{(\varepsilon_s - \varepsilon_0)(\varepsilon_s/\varepsilon_0 - \cot^2\phi_{\text{inc}})} \frac{(\varepsilon_{\text{BSA}} - \varepsilon_s)(\varepsilon_{\text{BSA}} - \varepsilon_0)}{\varepsilon_{\text{BSA}}} \left(\frac{\Gamma_{\text{BSA}}}{\rho_{\text{BSA}}\lambda} \right). \end{aligned} \quad (2)$$

By rearranging Eq. (2) and putting in the values of $\varepsilon_0 = 2.307$ for the glass slide, $\varepsilon_s = 1.788$ for the aqueous buffer solution, $\varepsilon_d = \varepsilon_{\text{BSA}} = 2.031$ for BSA in solution, and $\rho_{\text{BSA}} = 1.35 \text{ gm/cm}^3$, we arrive at

$$\Gamma_{\text{BSA}} = (4.98 \times 10^{-5} \text{ g/cm}^2) |\Delta\delta|. \quad (3)$$

From Eq. (3) we convert the experimental $\Delta\delta$ to the surface mass density of the BSA layer. If we model a bovine serum albumin molecule as an ellipsoid with major axes being $4 \times 4 \times 14 \text{ nm}^3$,¹⁹ a fully packed monolayer of BSA molecules with their long axes *parallel* to the surface (side-on geometry) will have a surface mass density of $1.9 \times 10^{-7} \text{ g/cm}^2$. A full monolayer of BSA molecules with the long axes *perpendicular* to the surface (end-on geometry) will have a surface mass density of $6.8 \times 10^{-7} \text{ g/cm}^2$. Figure 2 shows that after the excess protein targets are washed off, nearly all printed spots consist of a full monolayer of BSA in end-on geometry.

The image of a nearly 3000-spot BSA microarray takes 69 min to acquire. Out of the total image acquisition time, only 10 min are used for signal averaging at 2 ms per “pixel,” while the remaining 59 min are spent to rotate the scan mirror that currently is driven by a stepping motor. By using a galvanometer-based scan mirror, the time needed to rotate the mirror in order to acquire the same image can be reduced to less than 3 min. This means that the scan time for an image as shown in Fig. 2 can be acquired in less than 15 min.

We subsequently use the new hybrid scanning OI-RD microscope to detect two known protein-protein and protein-small-molecule binding reactions on a separately prepared microarray. The microarray consists of rabbit IgG (RB), human IgG (HM), and mouse IgG (MS) as antigen targets; biotin-BSA conjugate as a small-molecule target, and unmodified BSA as a control target. The microarray is printed with each target in $1 \times \text{PBS}$ solution at a concentration of $7.6 \mu\text{M}$. Before the optical scan, the printed microarray is washed to remove the excess targets. In Fig. 3(a) we show the image of the target microarray obtained with the microarray immersed in $1 \times \text{PBS}$. Each column consists of one target and has 90 spots (only 15 spots are shown). The sections in the middle and on the right are duplicates. The spot size and the surface mass density are uniform for each target; they do vary for different targets presumably due to the detailed makeup of surface residues of the printed proteins. The average surface mass density of the immobilized BSA and biotin-BSA targets is $\Gamma_{\text{BSA}} = 5.8 \times 10^{-7} \text{ g/cm}^2$ [calculated using Eq. (3) from the optical signals $\Delta\delta$]. The surface mass densities of the immobilized IgG targets [as shown in Fig. 3(a)] are also converted from the respective optical signals using Eq. (3) since the optical dielectric constant and volume mass density for IgG are roughly the same as BSA.²⁰ The average surface mass density of the IgG targets is $\Gamma_{\text{IgG}} = 11.5 \times 10^{-7} \text{ g/cm}^2$. Again if we model an IgG molecule as an ellipsoid with three major axes measuring $4.4 \times 4.4 \times 23.5 \text{ nm}^3$, a full monolayer of IgG molecules with their long axes perpendicular to the surface will a surface mass




FIG. 3. Images of a protein microarray after a sequence of postprinting treatment. The microarray consists of mouse IgG (MS), human IgG (HM), rabbit IgG (RB), biotin-BSA (BBBSA), and BSA as targets printed in columns. Dashed lines separate three microarray replicates. (a) Image in surface mass density after washing; (b) mass density change due to BSA-blocking treatment; (c) mass density after reaction with goat polyclonal antibody against the Fc domain of mouse IgG; (d) mass density change after reaction with streptavidin.

density of $13 \times 10^{-7} \text{ g/cm}^2$. This suggests that both IgG and BSA targets bind to epoxy-functionalized surface in end-on geometry. It is reasonable to expect that half of the immobilized IgG targets have the Fc fragments pointing away from the surface and thus accessible to antibodies against these fragments.

Before reactions, we flow 0.5 mg/ml BSA (i.e., $7.6 \mu\text{M}$) in $1 \times \text{PBS}$ into the flow cell so that free epoxy groups on the microarray-bearing surface are reacted with the BSA molecules. This “blocking” step prevents nonspecific binding during subsequent reactions. We take an image after the

blocking step (not shown here). The differential image [as shown in Fig. 3(b)] taken by subtracting the image after blocking from the image shown in Fig. 3(a) reveals that the coverage Θ of the printed targets is uniform and full. It means that the variation in $\Delta\delta$ or in surface mass density from target to target in Fig. 3(a) results from the variation in mean thickness of the targets. The mean thickness depends on the orientation of the targets.

We expose the microarray first with unlabeled goat antibody against the Fc fragments of the mouse IgG at a probe concentration of 0.01 mg/ml (i.e., $0.067 \mu\text{M}$.) In Fig. 3(c)

we show the difference between the image taken after the reaction and the image taken after the BSA blocking step. Only the mouse IgG targets show the strong evidence of binding with the goat antimouse IgG probe. The average change in surface mass density on the mouse IgG targets is $\Delta\Gamma_{\text{IgG}}=5.3\times 10^{-7}$ g/cm². This is one-half of a full monolayer of goat IgG probe in end-on geometry. Since roughly one-half of the mouse IgG targets in end-on geometry has the Fc fragments accessible to the probe, this means that all accessible mouse IgG targets have reacted with the probe. Figure 3(c) also reveals noticeable cross reactivity of goat antimouse IgG with surface-immobilized rabbit IgG and human IgG targets. The net change in surface mass density due to the cross-reacted goat IgG is five times smaller than on the mouse IgG targets. There is no evidence of binding for goat antimouse IgG probe on either biotin-BSA conjugate or BSA control targets.

We next subject the microarray to reaction with streptavidin at a probe concentration of 0.17 μM . Figure 3(d) shows the differential image by subtracting the image taken before the reaction from the image after the reaction. Only on biotin-BSA conjugates we observe uniform changes in $\Delta\delta$ or in surface mass density, while no changes at all on other targets. This indicates, not surprisingly, that the streptavidin molecules bind specifically to biotin targets that are conjugated to BSA through linker molecules. The surface mass density of captured streptavidin molecules by biotin-BSA conjugates is $\Delta\Gamma=6.9\times 10^{-7}$ g/cm². It is more than the surface mass density expected of a fully packed monolayer of streptavidin, which is $\sim\Gamma_{\text{streptavidin}}=4\times 10^{-7}$ g/cm². This can be understood as follows: the area number density of biotin molecules is higher than that of a full monolayer of streptavidin; the linker molecules make the “excess” biotin targets accessible to streptavidin molecules beyond one fully packed monolayer.

In addition to endpoint measurements, the present high-throughput OI-RD microscope is capable of monitoring reaction kinetics at multiple targets. In this mode of operation, we choose one pixel from each of the five targets and one pixel from the unprinted region as the reference “channel” and only measure $\Delta\delta$ from these pixels in sequence. By repeating the measurement sequence, we obtain the surface mass density change in real time on the selected targets. We minimize the effect of any drift in η_{sys} in the microscope by subtracting $\Delta\delta$ of the reference channel from $\Delta\delta$ of the targets. Figure 4 shows the surface mass density change versus time on five targets measured at 1/3 Hz (one time point every 3 seconds). The solution of goat antimouse IgG at a concentration of 0.067 μM is quickly introduced into the sample fluidic system at $t=0$. The signal on the mouse IgG target increases until it reaches the equilibrium value $\Delta\Gamma=5.3\times 10^{-7}$ g/cm² as also displayed in Fig. 3(c). The signal on biotin-BSA conjugate and BSA control targets remain unchanged. On HM and RB target the signals change noticeably but by a significantly smaller amount, $\Delta\Gamma=1.5\times 10^{-7}$ g/cm², as also shown in Fig. 3(c). Other than confirming the endpoint measurement shown in Fig. 3(c), the real-time data reveal the kinetics of the observed binding reactions.




FIG. 4. Simultaneous measurement of the binding reaction in real time of goat polyclonal antibody against mouse IgG (the Fc domain) with a protein microarray consisting of mouse IgG, human IgG, rabbit IgG, biotin-BSA conjugate, and BSA as surface-immobilized target using the hybrid OI-RD microscope as illustrated in Fig. 1.

IV. CONCLUSION

In summary, we have developed a new high-speed oblique-incidence reflectivity difference (OI-RD) microscope that can detect biomolecular microarrays with 2760 features in 1 cm² area. This scanning microscope uses a combination of a rotating scan mirror for y -scan and a single-axis translation for x -scan. By increasing the scan field of view or the density of microarray spots by a factor of 4, this microscope should be able to detect biomolecular microarrays with over 10 000 features. For microarray spots with 100 μm diameters, this new microscope can acquire an endpoint image of a 2760-spot microarray in less than 15 min with a pixel dimension of $20\times 20\ \mu\text{m}^2$. The new scanning OI-RD microscope is also capable of following the kinetics of molecular interactions on biomolecular microarrays in a high-throughput fashion.

ACKNOWLEDGMENTS

This work was supported by NIH under NIH-R01-HG003827-02.

- ¹T. Kodadek, *Chem. Biol.* **8**, 105 (2001).
- ²G. MacBeath, *Nat. Genet.* **32**, 526 (2002).
- ³M. Schena, *Microarray Analysis* (Wiley, Hoboken, 2003).
- ⁴H. Zhu, M. Bilgin, R. Bangham, D. Hall, A. Casamayor, P. Bertone, N. Lan, R. Jansen, S. Bidingmaier, T. Houfek, T. Mitchell, P. Miller, R. A. Dean, M. Gerstein, and M. Snyder, *Science* **293**, 2101 (2001).
- ⁵P. Mitchell, *Nat. Biotechnol.* **20**, 225 (2002).
- ⁶O. Shliom, M. Huang, B. Sachais, A. Kuo, J. W. Weisel, C. Nagaswami, T. Nassar, K. Bdeir, E. Hiss, S. Gawlak, S. Harris, A. Mazar, and A. A.

- Higazi, J. Biol. Chem. **275**, 24304 (2000).
- ⁷R. Karlsson, J. Mol. Recognit. **17**, 151 (2004).
- ⁸B. K. Singh and A. C. Hillier, Anal. Chem. **78**, 2009 (2006).
- ⁹C. Boozer, G. Kim, S. Cong, H. Guan, and T. Londergan, Curr. Opin. Biotechnol. **17**, 400 (2006).
- ¹⁰J. S. Shumaker-Parry and C. T. Campbell, Anal. Chem. **76**, 907 (2004).
- ¹¹K. Usui-Aoki, K. Shimada, M. Nagano, M. Kawai, and H. Koga, Proteomics **5**, 2396 (2005).
- ¹²R. M. A. Azzam and N. M. Bashara, *Ellipsometry and Polarized Light* (Elsevier Science, Science, 1987).
- ¹³G. Jin, R. Jansson, and H. Arwin, Rev. Sci. Instrum. **67**, 2930 (1996).
- ¹⁴Z. H. Wang and G. Jin, Anal. Chem. **75**, 6119 (2003).
- ¹⁵J. P. Landry, X. D. Zhu, and J. P. Gregg, Opt. Lett. **29**, 581 (2004).
- ¹⁶P. Thomas, E. Nabighian, M. C. Bartelt, C. Y. Fong, and X. D. Zhu, Appl. Phys. A: Mater. Sci. Process. **79**, 131 (2004).
- ¹⁷X. D. Zhu, J. P. Landry, Y. S. Shun, J. P. Gregg, K. S. Lam, and X. W. Guo, Appl. Opt. **46**, 1890 (2007).
- ¹⁸Q. Xu, S. Miyamoto, and K. S. Lam, Mol. Divers. **8**, 301 (2004).
- ¹⁹A. K. Wright and M. R. Thompson, Biophys. J. **15**, 137 (1975).
- ²⁰J. L. Oncley, G. Scatchard, and A. Brown, J. Phys. Colloid Chem. **51**, 184 (1947).