

SALTIN, B., GAGGE, A. P. and STOLWIJK, J. A. J. (1968) Muscle temperature during submaximal exercise in man. *J. Appl. Physiol.* **25**, 679-688.
 TOGAWA, T. (1973) Medical thermometer making use of zero heat-flow method (in Japanese). Report of the Institute for Medical and Dental Engineering **7**, 75-83.

TATSUO TOGAWA
 TETSU NEMOTO

*Institute for Medical & Dental Engineering
 Tokyo Medical & Dental University
 Tokyo, Japan*

*Dept. of Pediatrics
 Tokyo Women's University
 Tokyo, Japan*

TOYO YAMAZAKI

*National Food Research Institute
 Ministry of Agriculture & Forestry
 Tokyo, Japan*

TOSHIO KOBAYASHI

Errata

DERRETT, C. J., and TAVNER, P. J. (1975) Active filters: some simple design methods for biomedical workers. *Med. & Biol. Eng.*, **13**, (6), 883-888

The authors would like to thank Dr. H. Berlin of the University of Delaware for drawing their attention to the following errors in the printed version of their paper:

Eqn. 1 should read:

$$V' = \frac{-V_{out} Y_5}{Y_2}$$

Eqn. 2 should read:

$$\frac{V_{out}}{V_{in}} = \frac{-Y_1 Y_2}{Y_2 Y_4 + Y_5 (Y_1 + Y_2 + Y_3 + Y_4)}$$

The equation for Y_5 in Section 3.1 should read:

$$Y_5 = sC_5$$

The equation for R_5 in Section 3.2 should read:

$$R_5 = 4R_3 = \frac{\sqrt{2}}{\omega_n C}$$

Eqn. 7 should read:

$$\frac{V_{out}}{V_{in}} = \frac{-\sqrt{(2)} \omega_n s}{s^2 + \sqrt{(2)} \omega_n s + \omega_n^2}$$

Dr. Berlin has also correctly pointed out that it is possible to synthesise the filter transfer function in the paper using different circuit configurations from those given.