

Chip-based wide field-of-view nanoscopy

Robin Diekmann^{#1}, Øystein I. Helle^{#2}, Cristina I. Øie², Peter McCourt³, Thomas R. Huser^{1,4}, Mark Schüttpelz¹ and Balpreet S. Ahluwalia²

Link to the manuscript:

<https://www.nature.com/nphoton/journal/v11/n5/full/nphoton.2017.55.html>

doi:10.1038/nphoton.2017.55

Chip-based wide field-of-view nanoscopy

Robin Diekmann^{#1}, Øystein I. Helle^{#2}, Cristina I. Øie², Peter McCourt³, Thomas R. Huser^{1,4}, Mark Schüttpehl¹ and Balpreet S. Ahluwalia²

These authors contributed equally to this work.

1 Department of Physics, Bielefeld University, 33615 Bielefeld, Germany.

2 Department of Physics and Technology, UiT-The Arctic University of Norway, 9037 Tromsø, Norway.

3 Department of Medical Biology, UiT-The Arctic University of Norway, 9037 Tromsø, Norway.

4 Department of Internal Medicine and NSF Center for Biophotonics, University of California, Davis, Sacramento, CA 95817, USA

Correspondence should be addressed to M.S. (schuettp@physik.uni-bielefeld.de) or B.S.A. (balpreet.singh.ahluwalia@uit.no).

Present optical nanoscopy techniques use a complex microscope for imaging and a simple glass slide to hold the sample. Here, we demonstrate the inverse: the use of a complex, but mass-producible optical chip, which hosts the sample and provides a waveguide for the illumination source, and a standard low-cost microscope to acquire super-resolved images via two different approaches. Waveguides composed of a material with high refractive-index contrast provide a strong evanescent field that is used for single-molecule switching and fluorescence excitation, thus enabling chip-based single molecule localization microscopy. Additionally, multimode interference patterns induce spatial fluorescence intensity variations that enable fluctuation-based super-resolution imaging. As chip-based nanoscopy separates the illumination and detection light paths, total-internal-reflection fluorescence excitation is possible over a large field of view, with up to 0.5 mm x 0.5 mm being demonstrated. Using multi-colour chip-based nanoscopy, we visualize fenestrations in liver sinusoidal endothelial cells.

Various different methods of super-resolution microscopy have emerged within the last few years¹. Techniques include structured illumination microscopy (SIM)^{2,3}, stimulated emission depletion (STED)^{4,5} microscopy and temporal signal fluctuation-based techniques such as super-resolution optical fluctuation Imaging (SOFI)⁶ and entropy based super-resolution imaging (ESI)⁷. However, single molecule localization microscopy (SMLM) approaches such as (direct) stochastic optical reconstruction microscopy ((d)STORM)^{8,9} and (fluorescence) photo-activated localization microscopy (F)PALM^{10,11} are possibly the most widespread implementations currently used in applied nanoscopy. SMLM techniques rely on spatially and temporally separated signals from single emitters such as organic fluorophores⁹ or quantum dots^{12,13} to localize their coordinates with sub-pixel precision. The localizations coordinate table from up to multiple thousands of frames is used to render the super-resolved image,

after possible filtering steps based on the localizations list. In contrast, fluctuation based techniques do not require strictly separated emitters but make use of temporal fluctuations of the fluorescence intensity signal over several to hundreds of frames. The super-resolved image is generated by time-dependent statistical analysis of the pixel values in the fluorescence image stack, e.g. calculation of cumulants in SOFI⁶ or entropy values in ESI⁷.

Ongoing developments suggest that an impressive diversity of biological and medical questions will be answered using optical nanoscopy. However, this development is frequently accompanied by the need for bulky and expensive optical setups that are rather cumbersome to operate and to maintain, especially for untrained users. Typically, a standard microscope is adapted by the use of bulk optical components to generate, steer and deliver the required illumination pattern to the sample. Often, the same objective lens is used for both, fluorescence excitation and detection, and there are thus overlaid excitation and emission paths. Furthermore, specific objective lenses with a high numerical aperture (NA) are needed to support optical sectioning in the form of total internal reflection or highly inclined and laminated optical sheet (HILO)¹⁴ illumination. Another approach to optical sectioning is given by Bessel beam¹⁵ or lattice light sheet¹⁶ illumination, again at the cost of rather complex and bulky optical setups. Besides the technical demands, using a high-NA objective lens with a high magnification consequently limits the field-of-view (FOV). This makes simultaneous super-resolution imaging of several to many cells, difficult or impossible.

Optical microscopy based on waveguide chips significantly reduces the complexity of the entire optical setup, enabling miniaturization by completely removing the excitation light path from the microscope. Instead, waveguides which tightly confine the guided light by total internal reflection due to a high refractive index contrast (HIC) to the surrounding media such as water and cells are used to deliver the illumination light to the sample. The evanescent field on top of the waveguide can be utilized for total internal reflection fluorescence (TIRF) excitation over an almost arbitrarily wide FOV that is intrinsically independent of the detection objective lens and in principle only limited by the waveguide design. Evanescent field excitation using waveguides was first introduced by Grandin et al.¹⁷, where a slab waveguide was used to generate an evanescent field over the large stretch of the waveguide chip. Slab waveguides (**Fig. 1a**) have later been used for both fixed and live cell imaging¹⁸⁻²⁰ and recently for label free imaging²¹. In another work²², a coverslip was used to support guided modes of light, providing an evanescent field across the entire coverslip. Although these previous approaches to waveguide-based illumination have been used for fluorescence microscopy for some time, none of these techniques have yet demonstrated imaging with a resolution beyond the diffraction limit. Here, we present super-resolution optical microscopy

using chip-based waveguides made of HIC materials, opening research activities towards the field of nanoscopy based on photonic integrated circuits (PICs).

Waveguide chips potentially allow for easy integration of multiple optical functions on a common platform. Particularly, devices based on HIC materials allow tighter confinement of light inside the waveguide reducing the footprints and opening new possibilities of integrated chip-based optical systems such as Raman spectroscopy²³, bio-sensors²⁴, optical coherence tomography²⁵, and others.

Here, we demonstrate waveguide chip-based super-resolution fluorescence imaging by two complementary approaches using ESI and *d*STORM. The high intensity in the evanescent field generated by the HIC waveguide material is used for optical switching of fluorophores as required by *d*STORM. In addition, the intrinsically multi-mode interference pattern within the waveguide is used to generate fluctuating intensity patterns for ESI. To demonstrate the applicability of waveguide chip-based super-resolution microscopy we visualize the connection of the actin cytoskeleton and plasma membrane fenestrations in liver sinusoidal endothelial cells (LSECs).

RESULTS

Chip-based single molecule localization microscopy

The performance of chip-based *d*STORM is shown by imaging immunostained microtubules in rat LSECs²⁶ plated directly on the waveguide (**Fig. 2a**). Measuring the lateral profile along one straight microtubule filament reveals a hollow structure²⁷ which has been used earlier in localization microscopy as a benchmark sample²⁸⁻³⁰, discussed in detail in³¹. This shows a resolution of better than 50 nm (**Fig. 2b**), confirmed by full-width-at-half-maximum (FWHM) values, localization precision³², and Fourier ring correlation^{33,34} (FRC) calculations (**Supplementary Fig. 1**). The resolution capability was further investigated by using DNA origami nanorulers that provide markers at (50 ± 5) nm distance as references. These can be clearly resolved in chip-based *d*STORM (**Fig. 2c,d, Supplementary Fig. 2**) which shows a comparable performance to the widely used architecture of an inverted TIRF *d*STORM setup (**Fig. 2c,d, Supplementary Fig. 3**).

As an advantage over conventional setups, waveguide chip-based nanoscopy greatly benefits from the fact that the fluorescence excitation is independent of the detection objective lens. As fluorescence is excited by the evanescent field of the waveguide, the technique provides optical sectioning and excellent signal to background ratios at penetration depths below 200 nm (**Supplementary Fig. 4, Supplementary Fig. 5, Video 1**), similar to

objective-based TIRF. Beyond that, the planar extent of the evanescent field only depends on the waveguide geometry and can thus be adjusted by the waveguide layout. While for epi-illumination, *d*STORM imaging over a FOV of 100 μm side lengths has recently been demonstrated³⁵, we significantly extend this range by chip-based TIRF imaging: the use of a broad waveguide in combination with a low magnification/low NA objective lens enables simultaneous imaging of more than 50 cells (**Fig. 2e**) on a FOV with 0.5 millimetre width and height while details below 140 nm could still be resolved (**Fig. 2f,g**), again confirmed by multiple methods (**Supplementary Fig. 6**).

Samples can be prepared directly on the waveguide chip (**Fig. 1b**) following standard protocols used for cell attachment to common glass coverslips³⁶. Besides acting as a substrate for the sample, the chip additionally provides an evanescent field used for fluorescence excitation through the integrated waveguide (**Fig. 1c**) while the fluorescent signal is captured by a conventional upright microscope. Since light guidance and shaping is integrated in the waveguide chip, the entire optical setup features relatively low complexity and only requires efficient coupling of light into the waveguide. We meet this requirement by either focusing different laser lines through an objective lens to the input facet of the waveguide or directly coupling to it from an optical fibre (**Fig. 1d**). As different wavelengths can be combined simultaneously, dyes can be effectively photoswitched for *d*STORM imaging (**Supplementary Fig. 7**).

To develop suitable waveguides for chip-based nanoscopy, tantalum pentoxide (Ta_2O_5) and silicon nitride (Si_3N_4) materials were used because of their high refractive indices, low propagation losses, low absorption, and low auto-fluorescence at visible wavelengths³⁷. High intensities in the evanescent field induce switching of organic fluorophores. To achieve this, we use channel-like waveguides of rib or strip geometry (**Fig. 1a**) that transversely confine the guided light in contrast to the slab geometry used in other waveguide-imaging approaches^{17-19,21}. Channel breadths (i.e. the waveguide extend orthogonal to the wave propagation direction) range from 25 to 500 μm in order to span the appropriate length scale for imaging many samples simultaneously, e.g. several cells.

The waveguides used in this work are highly multi-moded. Back reflections and multi-mode interference inside the waveguide cause a spatially stable distribution of laterally non-uniform evanescent fields, which strongly depends on the coupling into the waveguide. Accordingly, fluorescence excitation is inhomogeneous and, hence, impairs the reconstructed *d*STORM image. To counteract this problem, a piezo stage is used to oscillate the coupling objective lens or fibre back and forth along the input facet of the waveguide during the measurement.

This maintains continuous coupling but shifts the mode pattern to obtain an average distribution which shows significantly less modulation over larger length scales as compared to conventional TIRF setups (**Supplementary Fig. 8, Supplementary Fig. 9, Video 2**). Consequently, the distinct modes are not visible in the reconstructed *d*STORM images.

Chip-based fluctuation imaging

In contrast, fluctuations induced by the shifted mode pattern are desired in the case of chip-based ESI. Although original implementations of the fluctuation-based approaches SOFI and ESI use intrinsic quantum dot or fluorophore temporal intensity fluctuations (e.g. due to blinking and bleaching), it has recently been shown that speckle pattern illumination³⁸ can also invoke temporal emission fluctuations allowing for super-resolved fluctuation imaging³⁹. Intrinsic intensity fluctuations originate from single emitters and therefore spatially tightly confined sources. On the contrary, the spatial frequencies of the waveguide illumination pattern define the length scales on which the fluctuations occur, and, hence, the obtainable resolution (**Supplementary Fig. 10, Supplementary Note 1**). Using *d*STORM, we have measured fringes in the multi-mode interference pattern of a waveguide. The FWHM of the structure sizes goes down to 140 nm for a vacuum laser excitation wavelength of 660 nm (**Supplementary Fig. 11**).

To utilize the illumination intensity fluctuations induced by the waveguide for ESI analysis, we oscillate the coupling along the input facet such that the mode distribution changes from image to image. As random intensity fluctuations are desired, there is no need to further control the illumination pattern besides changing it from frame to frame. Low input power is used, keeping the intensity under the threshold of undesired single molecule switching, contrary to the *d*STORM acquisition procedure where single molecule switching is required. While continuously changing the excitation pattern in this manner, about 200 frames are recorded. The acquired data is used as input for the ESI reconstruction algorithm to generate a super-resolved image (**Fig. 3a**) which we demonstrate via imaging tubulin in LSECs (**Fig. 3b,c**).

Following the ESI acquisition, the same sample is again imaged using *d*STORM by increasing the illumination intensity. This allows for a direct comparison of the ESI image with the *d*STORM image of higher resolution but requiring a much higher number of input frames (**Supplementary Table 1**). Hence, *d*STORM serves as a control reference for assessing the performance of chip-based ESI. This verifies that the resolution of the ESI image is on the order of 110 nm, as adjacent microtubules at 106 nm distance are still resolved and simultaneously observed in both super-resolved images (**Fig. 3c,d**). This resolution is also

confirmed by a FWHM of 104 nm of a single tubule in the ESI image (**Supplementary Fig. 12**). Thus, chip-based ESI utilizing spatial excitation pattern fluctuations readily achieves a resolution enhancement of about a factor of 2, using an NA 1.2 objective lens for fluorescence detection (**Supplementary Note 1**). Chip-based *d*STORM pushes the resolution even further, admittedly at the cost of longer acquisition times.

Scalable super-resolution imaging

As the evanescent field generation in waveguide chip-based nanoscopy does not depend on the objective lens used for fluorescence detection, the presented approaches can be applied for successive image acquisitions at different magnifications allowing for scalable FOV imaging: To obtain an overview image with a large FOV, low magnification/ low NA lenses can be used. If higher resolution is desired, specific regions of interest (ROI) can be imaged at superior resolution afterwards switching to a high magnification/ high NA objective lens.

Intensity fluctuations owing to the shifted mode pattern enabled us to record ESI images with TIRF excitation in a FOV of 0.46 mm width using a 20x magnification/ NA 0.4 objective lens (**Figure 4b,e**). While a resolution not better than 653 nm is expected for diffraction-limited imaging (**Fig. 4a,d**), actin filaments of 334 nm distance can still be resolved in the large FOV ESI image (**Fig. 4h**). A filament FWHM value of 333 nm confirms this finding (**Supplementary Fig. 13**). This demonstrates a resolution enhancement at these particular positions of approx. a factor of 2, which was achieved from 208 frames recorded in less than 25 seconds. For further resolution improvement, *d*STORM imaging was performed using the same low NA objective lens (**Fig. 4c,f**).

The use of a low NA 0.45 (**Fig. 2e,f**) or 0.4 (**Fig. 4c,f**) objective lens lacks resolution compared to the use of a higher NA 1.2 objective lens (**Fig. 2a, Fig 3b,c, Fig. 4g**) as the obtainable localization precision and accuracy strongly depends on the NA of the utilized objective lens (**Supplementary Note 2, Supplementary Note 3, Supplementary Fig. 14**). Nevertheless, it shows that TIRF-based microscopy is not limited to high NA and high magnification lenses and can be extended to wide fields-of-view. Using the low NA 0.4 objective lens, as few as 208 frames for ESI and 3,000 frames for *d*STORM were sufficient to capture the structure of interest at significantly improved resolution in comparison to the diffraction limit (**Supplementary Note 4, Supplementary Fig. 15, Supplementary Fig. 16**). Fast software tools for (almost) real-time *d*STORM reconstruction are readily available^{40,41}, such that *d*STORM is mainly limited by the acquisition time. Hence, either low NA ESI or *d*STORM can be used for identification of cells of interest or for finding rare biological events, while the choice of the proper method can be made by prioritizing either the acquisition time

(choosing ESI) or the resolution (choosing *d*STORM). Subsequently, an identified region of interest may be imaged using high NA *d*STORM to achieve uncompromised resolution (**Fig. 4g**). Additionally, the relatively short acquisition times for the low NA images help to prevent photobleaching of fluorophores, allowing for successive acquisition procedures of the same sample (**Supplementary Fig. 15**).

Imaging membrane/cytoskeleton interaction in LSECs

Besides the cytoskeleton, the plasma membrane of LSECs is of distinct importance for their function^{26,42}. LSECs are highly specialized endothelial cells forming the sinusoidal blood vessels in the liver. They contain multiple trans-cytoplasmic pores (fenestrations) of 50 to 200 nm in diameter, making optical nanoscopy a particularly well-suited tool for their investigation⁴³. Fenestrations are typically grouped together in structures called sieve plates, due to their function as a molecular sieve between the blood and the hepatocytes. Fenestrations allow free passage of plasma from the sinusoidal lumen through to the Space of Disse and beyond to the underlying hepatocytes (and in the reverse direction), while retaining the blood cells in the lumen. Constituents in the plasma (e.g. lipoproteins, pharmaceuticals) can thus be cleared and processed by hepatocytes, and products synthesized by hepatocytes can be released back into the circulation via this route. We applied waveguide imaging to visualize the interaction between the actin network and fenestrations (**Fig. 5a, Supplementary Fig. 17**). Multi-colour chip-based *d*STORM shows that small actin filaments colocalize with the membrane supporting the individual fenestrations (**Fig. 5b**), reinforcing earlier findings with different methods^{44,45}. Thicker actin bundles surround multiple fenestrations forming sieve plates. Recently, it has been demonstrated that sieve plates are also surrounded by tubulin⁴⁴. Taken together with our results, this suggests that sieve plates have a combined actin and tubulin outer framework. Super-resolution optical methods that allow multiplexing will contribute to a greater understanding of how sieve plates and fenestrations function and are regulated, especially if these methods can be extended to living cells.

DISCUSSION

Our approach to resolution-scalable nanoscopy allows for localization-based, as well as fluctuation-based super-resolution imaging on a simple setup that only demands coupling of the excitation light source into the waveguide. Using a 20x objective lens, a FOV of approx. 0.5 mm width can be imaged at a resolution of about 340 nm in less than 25 seconds as shown above, which will also enable live-cell applications. Accepting longer acquisition times of approx. 8 to 30 minutes, a resolution even below 140 nm can be achieved for this large FOV without image stitching (**Supplementary Note 5**). Thus, our waveguide chip-based

approach fills the missing gap of acquiring highly resolved images over an unprecedented large FOV in TIRF microscopy of which the maximum breadth is predetermined by the waveguide layout. Changing to a 60x objective lens with NA 1.2 results in an optical resolution of below 50 nm for *d*STORM or about 110 nm for ESI. Though the presented approach is more flexible in the choice of the imaging lens than other implementations of TIRF microscopy, it shares the feature of a shallow sample illumination. Thus, fluorescence excitation is limited to a few hundred nanometres above the substrate, which can be either advantageous or disadvantageous depending on the application. While the illumination over the entire stretch of the waveguide enables large FOV TIRF imaging, the illuminated area usually goes beyond the FOV. All samples located on top of the currently imaged waveguide are simultaneously exposed to fluorescence excitation which must be considered in the experimental design, e.g. concerning photobleaching.

Though both ESI and *d*STORM are based on fluorescence signal dynamics, the underlying processes significantly differ for their chip-based implementations. As in conventional realisations, chip-based *d*STORM relies on the ability of the dyes to exhibit photoswitching, but the demonstrated implementation of ESI induces spatial intensity fluctuations by the waveguide excitation. It is therefore not restricted to specific probes, which makes it applicable to an even wider range of scenarios. So far, we have focused on a comparison of ESI and *d*STORM, which on the one hand allowed us to validate chip-based ESI by *d*STORM. On the other hand, experiments were designed for *d*STORM compatible dyes and imaging was performed in a buffer that favours photoswitching beyond a certain intensity threshold. As photoswitching was not desired for ESI data acquisition, excitation intensities were kept relatively low, entailing the need for rather long camera exposure times. However, if chip-based ESI is performed independently of *d*STORM, much higher intensities accompanied with significantly shorter exposure times can be applied. This will enable the acquisition of a fully reconstructed ESI image within 1 second, allowing for live-cell imaging at high temporal resolution.

While *d*STORM and fluctuation-based approaches already belong to the most wide-spread techniques for super-resolution imaging, the waveguide chip-based implementation dramatically reduces the complexity of the setup and, thus, has the potential to make it accessible for a wider range of users. Incorporating it into PICs will make it compatible with fast and compact optical fibre components, e.g. fibre-based switches, modulators, multiplexer and lasers. A waveguide chip, either fibre pigtailed or with a FC connector, will simply allow adding or removing any optical fibre components, thus eliminating the need of system re-alignment which is often the case when using bulk optical components. By guiding

the illumination light through optical fibres and waveguides, many standard optical microscopes can be used to acquire super-resolved images. Future PIC developments towards on-chip laser generation, filtering, and steering of entire illumination systems will potentially further extend the capabilities. Additionally, the integrated platform makes combinations with different lab-on-a-chip methods, e.g. microfluidics⁴⁶, optical trapping⁴⁷ or other detection techniques⁴⁸ now straightforward to implement. Furthermore, other analysis modes based on SMLM can be utilized e.g. for highly sensitive quantitative binding assays⁴⁹.

METHODS

Methods and any associated references are available in the online version of the paper.

ACKNOWLEDGMENTS

The authors thank Idir Yahiatène and Marcel Müller for help with the ESI reconstruction algorithm and Viola Mönkemöller for help with sample preparation and membrane dye staining. This work was supported by the European Research Council (grant no. 336716 to B.S.A.), the Research Council of Norway (grant no. 244764/F11 to B.S.A.), and the German Academic Exchange Service (grant no. 57160327 to M.S.). R.D. acknowledges additional support by grant no. KF2140610NT4 of the German Federal Ministry for Economic Affairs and Energy.

AUTHOR CONTRIBUTIONS

B.S.A. and M.S. conceived the project. All authors designed the research. C.I.Ø. isolated the cells, and stained and prepared the biological samples. R.D. and Ø.I.H. built the setup, prepared the non-biological samples, performed the experiments, performed the simulations, reconstructed the images, analysed the data, and created the figures. R.D., Ø.I.H., M.S., and B.S.A. mainly wrote the paper. All authors reviewed the manuscript.

COMPETING FINANCIAL INTERESTS

M.S. and B.A.S. applied for patent GB1606268.9 for chip-based optical nanoscopy. The other authors declare no competing financial interest.

Figure 1 | The implementation of chip-based nanoscopy. (a) Channel-like waveguide geometries are realized by etching the slab waveguide either partially or completely down to the SiO₂ substrate. In either case, the light is mainly guided by the channels of 25 to 500 μm breadth. (b) Five channels can easily be seen in the photograph of a strip waveguide chip, marked by the arrows. (Scale bar, 1 cm.) (c) Light guided inside the waveguide is the source of the evanescent field illuminating samples on top of the surface. (d) The optical setup consists of a simple upright microscope for fluorescence detection and an illumination unit to provide coupling to the input facet of the waveguides either through an objective lens or via a lensed fibre.




Figure 2 | Demonstration of chip-based *d*STORM. (a) Diffraction limited and *d*STORM imaging of tubulin in liver sinusoidal endothelial cells (LSECs). (Scale bar, 2 μ m. Scale bar on inset, 1 μ m.) Measuring a lateral profile of 540 nm width along a straight microtubule (magenta marking in the inset of the *d*STORM image) reveals its hollow structure (b). (c) The resolution capability is further investigated by imaging DNA-origami nanorulers of (50 ± 5) nm specified length that can be clearly resolved with waveguide-based *d*STORM similar to objective-based TIRF *d*STORM. (Scale bars, 50 nm.) (d) Analysing their line profiles, a mean nanoruler length of 49 nm is found in both cases affirming that the chip-based implementation shows comparable performance to a conventional inverted *d*STORM setup. (e) Additionally, waveguide chip-based illumination allows for using a low magnification/ low NA (20x/ NA 0.45) objective lens for *d*STORM imaging over a field-of-view of 0.5 mm \times 0.5 mm. (f) shows a detail marked by the white box in (e), where the profile over adjacent tubulin filaments reveals their separation by 138 nm (g).




Figure 3 | Demonstration of chip-based ESI. (a) Spatial fluorescence intensity fluctuations are induced by changing the mode pattern of the waveguide during image acquisition. Accordingly, these measurements show a diffraction limited image of the labelled structure multiplied by the mode pattern. A stack of approx. 200 frames is used as input data for the fluctuation analysis reconstruction algorithm, resulting in one super-resolved image. (b) Imaging tubulin in an LSEC. The comparison of the diffraction limited image, the corresponding ESI reconstruction, and the *d*STORM image shows the gradually increasing resolution. (Scale bar, 5 μm .) (c) shows a magnification of the same region (indicated by the rectangle in (b)) for the three different imaging modalities. (Scale bar, 0.5 μm .) The *d*STORM image of the same structure verifies the applicability of chip-based ESI: the line profiles reveal a resolution of 106 nm for chip-based ESI (green line) by imaging two adjacent microtubules, simultaneously observed in the *d*STORM image (blue dashed line) (d).




Figure 4 | Imaging the same sample under varied acquisition conditions reveals the specific strengths of the different approaches. As TIRF excitation is provided over the entire width of the waveguide, arbitrary objective lenses can be used for detection. Using a 20x magnification objective lens allows for chip-based TIRF imaging over a field-of-view of 0.46 mm width (a,d). Both fluctuation-based ESI (b,e) and localization-based *d*STORM (c,f) result in an optical resolution enhancement obtained using the NA 0.4 lens, and, thus, resolve actin bundles at 334 nm distance (h). Hence, both techniques provide a tool for scanning large fields-of-view at high resolution to identify cells of interest that can be re-imaged using a high magnification/ high NA lens in a subsequent step (g), accomplishing even superior resolution. The choice between wide field-of-view ESI or *d*STORM can be made on either prioritizing short acquisition times (choosing ESI) or best resolution (choosing *d*STORM). (d-g) show a detail of the region marked by the box in (a). (Scale bar on (a), 20 μ m. Scale bar on (d), 2 μ m.)





Figure 5 | Multi-colour chip-based dSTORM reveals the interplay between actin (magenta) and the membrane (green) in LSECs. (a) Groups of fenestrations form sieve plate superstructures which are surrounded by thicker actin bundles. As can be seen in the inset, actin is present between neighbouring fenestrations where it colocalizes with the plasma membrane. (Scale bar, 5 μm . Scale bar on inset, 1 μm .) (b) The line profiles taken at different positions in the liver cell as shown in (a) reveal diameters of approx. 200 nm and smaller for the fenestrations chosen, that can be resolved by super-resolution microscopy only. The profiles underline the visual impressions of colocalization.



REFERENCES

- 1 Schermelleh, L., Heintzmann, R. & Leonhardt, H. A guide to super-resolution fluorescence microscopy. *Journal of Cell Biology* **190**, 165-175, doi:10.1083/jcb.201002018 (2010).
- 2 Gustafsson, M. G. Surpassing the lateral resolution limit by a factor of two using structured illumination microscopy. *J Microsc* **198**, 82-87 (2000).
- 3 Heintzmann, R. & Cremer, C. G. Laterally modulated excitation microscopy: improvement of resolution by using a diffraction grating. *Proc. SPIE 3568, Optical Biopsies and Microscopic Techniques III* **3568**, 185-196, doi:10.1117/12.336833 (1999).
- 4 Hell, S. W. & Wichmann, J. Breaking the diffraction resolution limit by stimulated emission: stimulated-emission-depletion fluorescence microscopy. *Opt Lett* **19**, 780-782 (1994).
- 5 Willig, K. I., Rizzoli, S. O., Westphal, V., Jahn, R. & Hell, S. W. STED microscopy reveals that synaptotagmin remains clustered after synaptic vesicle exocytosis. *Nature* **440**, 935-939, doi:10.1038/nature04592 (2006).
- 6 Dertinger, T., Colyer, R., Iyer, G., Weiss, S. & Enderlein, J. Fast, background-free, 3D super-resolution optical fluctuation imaging (SOFI). *Proc Natl Acad Sci U S A* **106**, 22287-22292, doi:10.1073/pnas.0907866106 (2009).
- 7 Yahiatene, I., Hennig, S., Muller, M. & Huser, T. Entropy-Based Super-Resolution Imaging (ESI): From Disorder to Fine Detail. *Acs Photonics* **2**, 1049-1056, doi:10.1021/acsphotonics.5b00307 (2015).
- 8 Rust, M. J., Bates, M. & Zhuang, X. Stochastic optical reconstruction microscopy (STORM) provides sub-diffraction-limit image resolution. *Nature Methods* **3**, 793-795 (2006).
- 9 Heilemann, M. *et al.* Subdiffraction-resolution fluorescence imaging with conventional fluorescent probes. *Angew Chem Int Edit* **47**, 6172-6176, doi:10.1002/anie.200802376 (2008).
- 10 Betzig, E. *et al.* Imaging intracellular fluorescent proteins at nanometer resolution. *Science* **313**, 1642-1645, doi:10.1126/science.1127344 (2006).
- 11 Hess, S. T., Girirajan, T. P. K. & Mason, M. D. Ultra-high resolution imaging by fluorescence photoactivation localization microscopy. *Biophys J* **91**, 4258-4272, doi:10.1529/biophysj.106.091116 (2006).
- 12 Hoyer, P., Staudt, T., Engelhardt, J. & Hell, S. W. Quantum dot blueing and blinking enables fluorescent microscopy. *Nano Lett* **11**, 245-250 (2010).
- 13 Xu, J. Q., Tehrani, K. F. & Kner, P. Multicolor 3D Super-resolution Imaging by Quantum Dot Stochastic Optical Reconstruction Microscopy. *Acs Nano* **9**, 2917-2925, doi:10.1021/nn506952g (2015).
- 14 Makio Tokunaga, N. I., Kumiko Sakata-Sogawa. Highly inclined thin illumination enables clear single-molecule imaging in cells. *Nature Methods* **5**, 159-161 (2007).
- 15 Planchon, T. A. *et al.* Rapid three-dimensional isotropic imaging of living cells using Bessel beam plane illumination. *Nature Methods* **8**, 417-U468, doi:10.1038/Nmeth.1586 (2011).
- 16 Chen, B. C. *et al.* Lattice light-sheet microscopy: Imaging molecules to embryos at high spatiotemporal resolution. *Science* **346**, 439-+, doi:10.1126/Science.1257998 (2014).
- 17 Grandin, H. M., Stadler, B., Textor, M. & Voros, J. Waveguide excitation fluorescence microscopy: A new tool for sensing and imaging the biointerface. *Biosens Bioelectron* **21**, 1476-1482, doi:10.1016/j.bios.2005.06.011 (2006).
- 18 Agnarsson, B., Ingthorsson, S., Gudjonsson, T. & Leosson, K. Evanescent-wave fluorescence microscopy using symmetric planar waveguides. *Opt Express* **17**, 5075-5082, doi:10.1364/Oe.17.005075 (2009).

- 19 Agnarsson, B., Jonsdottir, A. B., Arnfinnsdottir, N. B. & Leosson, K. On-chip modulation of evanescent illumination and live-cell imaging with polymer waveguides. *Opt Express* **19**, 22929-22935, doi:10.1364/OE.19.022929 (2011).
- 20 Shen, H. *et al.* TIRF microscopy with ultra-short penetration depth. *Opt Express* **22**, 10728-10734, doi:10.1364/OE.22.010728 (2014).
- 21 Agnarsson, B. *et al.* Evanescent Light-Scattering Microscopy for Label-Free Interfacial Imaging: From Single Sub-100 nm Vesicles to Live Cells. *ACS Nano* **9**, 11849-11862, doi:10.1021/acs.nano.5b04168 (2015).
- 22 Ramachandran, S., Cohen, D. A., Quist, A. P. & Lal, R. High performance, LED powered, waveguide based total internal reflection microscopy. *Scientific reports* **3** (2013).
- 23 Dhakal, A. *et al.* Evanescent excitation and collection of spontaneous Raman spectra using silicon nitride nanophotonic waveguides. *Optics Letters* **39**, 4025-4028, doi:10.1364/OI.39.004025 (2014).
- 24 Fedyanin, D. Y. & Stebunov, Y. V. All-nanophotonic NEMS biosensor on a chip. *Scientific Reports* **5**, doi:10.1038/srep10968 (2015).
- 25 Yurtsever, G. *et al.* Photonic integrated Mach-Zehnder interferometer with an on-chip reference arm for optical coherence tomography. *Biomed Opt Express* **5**, 1050-1061, doi:10.1364/BOE.5.001050 (2014).
- 26 Sørensen, K. K., Simon-Santamaria, J., McCuskey, R. S. & Smedsrød, B. Liver Sinusoidal Endothelial Cells. *Comprehensive Physiology* (2015).
- 27 Weber, K., Rathke, P. C. & Osborn, M. Cytoplasmic microtubular images in glutaraldehyde-fixed tissue culture cells by electron microscopy and by immunofluorescence microscopy. *Proc Natl Acad Sci U S A* **75**, 1820-1824 (1978).
- 28 Olivier, N., Keller, D., Gonczyk, P. & Manley, S. Resolution Doubling in 3D-STORM Imaging through Improved Buffers. *Plos One* **8**, doi:10.1371/journal.pone.0069004 (2013).
- 29 Vaughan, J. C., Jia, S. & Zhuang, X. Ultra-bright Photoactivatable Fluorophores Created by Reductive Caging. *Nat Methods* **9**, 1181-1184, doi:10.1038/nmeth.2214PMCID: PMC3561463 (2012).
- 30 Endesfelder, U. & Heilemann, M. Art and artifacts in single-molecule localization microscopy: beyond attractive images. *Nat Methods* **11**, 235-238, doi:10.1038/nmeth.2852 (2014).
- 31 Bourg, N. *et al.* Direct optical nanoscopy with axially localized detection. *Nature Photonics* **9**, 587-593, doi:10.1038/nphoton.2015.132 (2015).
- 32 Endesfelder, U., Malkusch, S., Fricke, F. & Heilemann, M. A simple method to estimate the average localization precision of a single-molecule localization microscopy experiment. *Histochemistry and cell biology* **141**, 629-638 (2014).
- 33 Nieuwenhuizen, R. P. *et al.* Measuring image resolution in optical nanoscopy. *Nat Methods* **10**, 557-562, doi:10.1038/nmeth.2448 (2013).
- 34 Banterle, N., Bui, K. H., Lemke, E. A. & Beck, M. Fourier Ring Correlation as a resolution criterion for super resolution microscopy. *Journal of Structural Biology* **183** (2013).
- 35 Douglass, K. M., Sieben, C., Archetti, A., Lambert, A. & Manley, S. Super-resolution imaging of multiple cells by optimised flat-field epi-illumination. *Nat Photonics* **10**, 705-708, doi:10.1038/nphoton.2016.200 (2016).
- 36 Smedsrød, B. & Pertoft, H. Preparation of pure hepatocytes and reticuloendothelial cells in high yield from a single rat liver by means of Percoll centrifugation and selective adherence. *Journal of leukocyte biology* **38**, 213-230 (1985).
- 37 Ahluwalia, B. S. *et al.* Fabrication of submicrometer high refractive index Tantalum Pentoxide waveguides for optical propulsion of microparticles. *Photonics Technology Letters, IEEE* **21**, 1408-1410 (2009).
- 38 Ventalon, C. & Mertz, J. Quasi-confocal fluorescence sectioning with dynamic speckle illumination. *Optics Letters* **30**, 3350-3352, doi:10.1364/OI.30.003350 (2005).

- 39 Kim, M., Park, C., Rodriguez, C., Park, Y. & Cho, Y. H. Superresolution imaging with optical fluctuation using speckle patterns illumination. *Sci Rep-Uk* **5**, 16525, doi:10.1038/srep16525 (2015).
- 40 Wolter, S. *et al.* rapidSTORM: accurate, fast open-source software for localization microscopy. *Nature methods* **9**, 1040-1041 (2012).
- 41 Smith, C. S., Joseph, N., Rieger, B. & Lidke, K. A. Fast, single-molecule localization that achieves theoretically minimum uncertainty. *Nature methods* **7**, 373-375 (2010).
- 42 Cogger, V. C., Roessner, U., Warren, A., Fraser, R. & Le Couteur, D. G. A Sieve-Raft Hypothesis for the regulation of endothelial fenestrations. *Computational and structural biotechnology journal* **8**, 1-9 (2013).
- 43 Mönkemöller, V. *et al.* Imaging fenestrations in liver sinusoidal endothelial cells by optical localization microscopy. *Phys Chem Chem Phys* **16**, 12576-12581, doi:10.1039/c4cp01574f (2014).
- 44 Mönkemöller, V., Øie, C., Hübner, W., Huser, T. & McCourt, P. Multimodal super-resolution optical microscopy visualizes the close connection between membrane and the cytoskeleton in liver sinusoidal endothelial cell fenestrations. *Scientific reports* **5** (2015).
- 45 Braet, F. & Wisse, E. Structural and functional aspects of liver sinusoidal endothelial cell fenestrae: a review. *Comp Hepatol* **1**, 1 (2002).
- 46 Wang, X. *et al.* Enhanced cell sorting and manipulation with combined optical tweezer and microfluidic chip technologies. *Lab on a Chip* **11**, 3656-3662 (2011).
- 47 Helle, Ø. I., Ahluwalia, B. S. & Hellesø, O. G. Optical transport, lifting and trapping of micro-particles by planar waveguides. *Opt Express* **23**, 6601-6612 (2015).
- 48 Dullo, F. T. & Hellesø, O. G. On-chip phase measurement for microparticles trapped on a waveguide. *Lab on a Chip* **15**, 3918-3924 (2015).
- 49 Jain, A. *et al.* Probing cellular protein complexes using single-molecule pull-down. *Nature* **473**, 484-U322, doi:10.1038/nature10016 (2011).

METHODS

Optical setups.

Waveguide-based imaging was performed on a conventional upright microscope (BXFM, Olympus), retrofitted with a waveguide chip module consisting of a piezo stage (17MAX604, Melles Griot) holding an objective lens (LMPlanFI 50x/0.5 BD, Olympus) or a lensed fibre (HP460sm, WT&T) used for coupling into the waveguide. The piezo allows fine tuning of the coupling, and was also programmed through a piezo controller (BPC303, Thorlabs) to oscillate during image acquisition. The chip was held by a vacuum chuck on a translation stage (MBT402D, Thorlabs) to prevent movement during coupling. During *d*STORM image acquisition, the vacuum pump was turned off to minimize sample drift but suction was retained. Imaging was performed using a 20x NA 0.4 (Plan N 20x/0.4, Olympus), 20x NA 0.45 (LUCPlanFL N 20x/0.45, Olympus), a 60x NA 1.2 water immersion (UplanSApo 60x/1.2 w, Olympus), or a 60x NA 1.42 oil immersion (PlanApo N 60x/1.42 oil, Olympus) objective lens. Different emission filter sets of both LP and BP filters were used (488LP (AHF) and 520/36BP (Edmund), or 664LP (AHF) and 692/40BP (AHF)), and the signal was captured by a sCMOS camera (Hamamatsu Orca Flash 4 v2), controlled by the manufacturer software (Hamamatsu HImage). The backprojected pixel width and height was 75.9 nm for Atto 488 and CellMask Deep Red measurements and 108.33 nm for Alexa 647 measurements.

Illumination was provided to the chip either via coupling a laser into the lensed fibre or by combining several laser lines through the coupling objective lens. Consequently, multi-colour imaging could be performed simultaneously if the emission path was split into two spectrally separated channels. The illumination unit consisted of two solid state lasers (Oxxius LBX-488-200-CSB 488 nm and Cobolt Flamenco 660 nm). Each beam profile width was adjusted by an individual Keplerian telescope (not shown in the sketch of Figure 1g) and spatial overlay was achieved by a dichroic mirror (DMPL505; Thorlabs).

Conventional *d*STORM imaging was performed on an inverted setup build around a microscope body (IX-71, Olympus) using a 60X NA 1.49 oil-immersion objective lens (Apo N, 60X NA 1.49; Olympus), described in detail elsewhere⁵⁰. Briefly, laser light was emitted from an argon-krypton ion laser (Innova 70C, Coherent) and the 647 nm laser line was selected by an acousto-optic tunable filter (AOTFnC-VIS-TN 1001; AA Opto Electronic). Fluorescence emission was captured by an EMCCD camera (iXon DV887DCS-BV; Andor) with backprojected pixel widths and heights of 112 nm through two emission filters (LP02-647RU-25; Semrock, ET700/75m; Chroma).

The TIRF excitation profile on the commercial OMX v4 Blaze system (GE Healthcare) was recorded using the ring TIRF mode and the field concentrator as the recommended illumination mode for *d*STORM imaging. The microscope was equipped with a 642 nm

excitation laser, an oil-immersion TIRF objective lens (60x NA1.49 ApoN, Olympus), and a sCMOS camera.

Waveguide chip fabrication.

The Ta₂O₅ waveguides were fabricated by sputtering a thin layer (150 to 220 nm) of Ta₂O₅ onto a silicon dioxide (SiO₂) substrate. Magnetron sputtering was employed and the deposition was optimized using an iterative process. The parameters that were optimized for the waveguide parameters were substrate temperature, magnetron power and the oxygen flow rate. The optimized parameters chosen were 200°C substrate temperature, magnetron power of 300 W and O₂ and Ar flow rates of 5 and 20 sccm, respectively. Post-deposition, standard photolithography was used to deposit photoresist for marking the waveguide geometry and ion-beam milling was used to fabricate the challenged waveguides (strip and rib). For rib waveguides, the slab was only partially etched down, while for strip the slab was etched completely down to the substrate. The sidewall roughness, which is the major source of propagation loss, was reduced by optimization of the ion-beam milling process. Finally, plasma etching was performed to remove the remaining photoresist from the photolithography step. Post-fabrication annealing for 5 hours at 600°C was employed to reduce the propagation losses.

The Si₃N₄ waveguides were fabricated first by thermally growing a silica layer with a thickness of 2 μm on silicon wafers. The thin Si₃N₄ layer (150 nm thick) was deposited using low-pressure chemical vapor deposition (LPCVD) at 800°C. Post-deposition, standard photolithography was used to deposit photoresist for marking the waveguide geometry, followed by reactive ion etching (RIE) to fabricate the challenged waveguides (strip and rib). The unwanted photoresist was then removed and a top cladding layer was deposited by plasma-enhanced chemical vapor deposition (PECVD) at 300°C of thickness 1.5 μm. At imaging areas, the top cladding was opened using RIE and wet etching to enable seeding of the cells. Further details of the fabrication process for Ta₂O₅ and Si₃N₄ waveguides can be found elsewhere^{37, 51}.

Before attaching cells onto the waveguides, the chip was thoroughly cleaned by submerging the chip in 5 % (v/v) Hellmanex (Sigma Aldrich) for 10 minutes at 70°C. Hellmanex was removed using distilled H₂O, and another cleaning step using isopropanol (Sigma Aldrich) was performed before rinsing once more with dH₂O.

Sample preparation.

Sprague Dawley male rats (Scanbur) were kept under standard conditions and fed standard chow ad libitum (Scanbur). The experimental protocols were approved by the Norwegian National Animal Research Authority (NARA) in accordance with the Norwegian Animal

Experimental and Scientific Purposes Act of 1986. The rats (body weight 150–300 g) were anesthetized with a mixture of medetomidin (Domitor vet, Orion) and ketamine (Ketalar, Pfizer) and LSECs were isolated and purified as described³⁶. Cleaned waveguide chips were coated with human fibronectin (50 µg/ml) for 10 min at room temperature. Isolated LSECs in RPMI 1640 were seeded on the fibronectin coated waveguide chips and incubated for 1 h at 37°C, followed by a washing step to remove non-attached cells and debris and allowed to spread their cytoplasm for another 2 h. The cells were then washed with phosphate buffered saline (PBS) and fixed with 4% paraformaldehyde (PFA) in PBS and 0.02 M sucrose, pH 7.2 for 30 minutes. Following fixation, the cell membranes were permeabilized with 0.1 % Triton X in PBS for 10 min at RT. Actin was stained with Atto 488-phalloidin (Sigma Aldrich), 1:400 dilution in PBS by 20 min incubation at RT. Membranes were stained with CellMask Deep Red (Life Technologies), 1:1000 dilution in PBS by 10 min incubation at RT. Tubulin was either stained with 1:400 alpha-tubulin antibody, Alexa Fluor 488 (ThermoFisher Scientific), at 1:400 dilution in PBS containing 5% bovine serum albumin, fraction V (AppliChem) by 20 min incubation at RT or with mouse anti beta-tubulin (Sigma Aldrich), at 1:400 dilution in PBS/BSA for 1 h at RT, followed by goat anti mouse Alexa Fluor 647 (ThermoFisher Scientific) at the same concentration as the primary antibody, for 1 h at RT.

Dye surfaces were prepared by submerging approx. 50 nM Alexa 647 in 0.05 % Poly-L-lysine (Sigma Aldrich) and incubating it for approx. 20 min on the waveguide chip or coverglass followed by an optional washing step. For measurements of the achievable localization precision using high and low NA objective lenses on the inverted conventional setup, 200 nm TetraSpeck Microspheres (Thermo Fisher Scientific) were diluted at approx. 1:50 in Vectashield H-1000 (Vector Laboratories).

Alexa 647-labelled R50 nanorulers (Gattaquant) were prepared on the waveguide by subsequently placing a Polydimethylsiloxane (PDMS) chamber on the surface, washing it thrice with PBS, incubating it with 0.5 mg/ml BSA-biotin (Sigma Aldrich) in PBS for 5 min, incubating it with 0.5 mg/ml Neutravidin (VWR) in PBS for 5 min, and incubating it with 0.5 µl R50 stock solution in 10 µl 60 mM Magnesium chloride in PBS for 5 min. Each step beginning from the BSA-biotin incubation was followed by two or three times washing with 60 mM Magnesium chloride (MgCl₂) in PBS. For imaging on the conventional inverted setup, the sample was prepared on a coverglass (#1.5) (Thermo Fisher Scientific Gerhard Menzel) by placing a PDMS chamber on the surface, washing it thrice with PBS, incubating it with 5 mg/ml BSA-biotin (Sigma Aldrich) in PBS for 5 min, washing it thrice with in PBS, incubating it with 0.5 mg/ml Neutravidin (Thermo Fisher Scientific) in PBS for 10 min, and incubating it with 2.5 µl R50 stock solution in 25 µl 10 mM Magnesium chloride in PBS for 1h. Each step

beginning from the Neutravidin incubation was followed by three times washing with 10 mM Magnesium chloride in PBS.

Imaging buffer was prepared by mixing H₂O-based 22.5 μ l oxygen scavenger system solution (based on glucose oxidase (Sigma-Aldrich) and catalase (Sigma-Aldrich))⁵², with 30 μ l PBS. The buffer was supplemented with 2 mM cyclooctatetraene (COT) (Sigma Aldrich) in case of Alexa 647 imaging and with different concentrations of Mercaptoethylamine (MEA) (Sigma Aldrich). An MEA concentration of 95 mM was used in case of CellMask Deep Red and Atto 488 imaging which we found to yield good blinking behavior on the waveguide for both CellMask Deep Red and Atto 488. In case of all Alexa 647 imaging experiments on the waveguide, we used an MEA concentration of 20 mM. The buffer for Alexa 647-labelled DNA nanorulers on the conventional inverted microscope was prepared from D₂O-based components, supplemented with 45 mM Tris, and contained MEA at a concentration of 10 mM. The same MEA concentration with additional 2 mM COT was used in H₂O-based buffer to record the example raw dSTORM data on the conventional inverted setup. Sub-diffraction sized, fluorescent 100 nm TetraSpeck Microspheres (Life Technologies) at a low concentration (approx. 1:5000 from stock) aid algorithmic drift correction of super-resolved images.

To prevent light from coupling into the coverslip, we mounted a custom made PDMS chamber of 120 μ m height on top of the chip to lift the coverglass (#1.5) (Thermo Fisher Scientific Gerhard Menzel) except for the experiment shown in **Fig. 5** and **Supplementary Fig. 17**. It was verified that only light guided by the waveguide excites fluorescence (**Supplementary Fig. 5, Video 1**). The chamber additionally served as a vessel for the imaging buffer. Alternatively, the buffer was held by capillary action between the chip and the coverslip directly deposited on the waveguide. In the latter case, a small piece of black PDMS of significantly greater height (approx. 2 mm) was applied close to the input facet to block any unguided light before the coverslip (experimental results shown in **Fig. 5** and **Supplementary Fig. 17**). This PDMS was manufactured such that a thin layer touching the waveguide was transparent, to prevent light absorption by the black PDMS.

Image acquisition and data analysis.

Several hundred frames were acquired at relatively low power at the input facet for waveguide-based imaging. During acquisition, the coupling objective lens was oscillated along the input facet. These image stacks were used as input for the ESI reconstruction. Alternatively, averaging over the acquired frames yields a diffraction-limited image of the sample. Exposure times for dSTORM ranged from 10 to 160.84 ms. For multi-colour

*d*STORM images (**Fig. 5, Supplementary Fig. 17**), different channels were recorded subsequently. A detailed register of imaging conditions is found in **Supplementary Table 1**. Image processing and analysis was carried out using Fiji⁵³ and Matlab (Mathworks). ESI reconstruction and postprocessing was performed using the Fiji plugin ESI⁷. Reconstructions were run in two iterations at second ESI order with an intermediate gamma correction step using a gamma factor of 0.5 to compensate for non-linear intensity amplification. For *d*STORM reconstruction and postprocessing the Fiji plugin ThunderSTORM⁵⁴ was used. Localizations were filtered for unphysical values of fitted FWHM width and localization precision. Non-linear drift correction was performed via cross-correlation of substacks. Rendered *d*STORM images were convolved with a 2D Gaussian kernel (see **Supplementary Table 1** for kernel width) in some cases. The look up tables and intensity values for shown figures were adjusted to achieve appropriate contrast including gamma correction, however all quantifications were made on linearly scaled images. For localization precision calculations³², we used custom written Matlab code. FRC analysis was carried out using the Fiji plugin FRC³³. Diffraction limited images were computed by averaging over all acquired frames recorded prior to *d*STORM imaging or the raw data also used as input for the ESI reconstruction.

Evanescent field intensities on the waveguide were measured using the fluorescent response of single sub-diffraction sized 100 μm TetraSpeck Microspheres (Life Technologies) as intensity sensors. A reference measurement was conducted by imaging the microspheres on the conventional inverted setup in epi-illumination mode. The photon response as a function of the illumination intensity was recorded and corrected for multiple factors including laser power, camera quantum efficiency, single-frame exposure time, the Gaussian-like beam profile in epi-illumination, and the waveguide breadth in waveguide illumination. Similar measurements were carried out on the waveguide to find the photon response as a function of the power at the input facet of the waveguide and estimate the corresponding evanescent field intensity. Though several factors have been corrected for, these numbers should only be taken as a rough estimation as the efficiency of coupling light into the waveguide differs between different chips.

ESI simulation.

Chip-based ESI was simulated by generating raw data of 200 frames in Matlab (Mathworks) as input for the ESI plugin. Multi-mode interference was modelled as sinusoidal patterns rendered to a fine grid with equally distributed random phases and equally distributed pattern wavelengths between 280 nm and 840 nm. The lower bound wavelength of 280 nm corresponds to a FWHM of 140 nm in the multi-mode interference pattern that we have measured for excitation with the laser of 660 nm vacuum wavelength (**Supplementary Fig.**

11). The fluorescent response of emitter pairs at different distances was modelled by assuming a linear response to the illumination, i.e. taking the respective illumination intensity at the emitters' positions as the fluorescent response in each frame. This signal was convolved with a Gaussian point-spread-function with a FWHM of $\lambda/2NA$ with the emission wavelength $\lambda = 678$ nm and the numerical aperture $NA = 1.2$, and afterwards scaled down to a grid with pixel widths and heights of 75.9 nm. Shot noise was modelled as Poisson noise and the signal was converted from counts to photons at 2.4 photons per count at a maximum of 500 photons per pixel. To consider additional camera noise, we used the sCMOS camera to record 200 frames with no illumination on the detector and added these to the modelled frames. The resulting image stack was imported to Fiji for ESI analysis.

Data availability statement

The data that support the images and plots within this paper and other findings of this study are available from the corresponding author upon reasonable request.

REFERENCES FOR METHODS

- 7 Yahiatene, I., Hennig, S., Muller, M. & Huser, T. Entropy-Based Super-Resolution Imaging (ESI): From Disorder to Fine Detail. *Acs Photonics* **2**, 1049-1056, doi:10.1021/acsp Photonics.5b00307 (2015).
- 32 Endesfelder, U., Malkusch, S., Fricke, F. & Heilemann, M. A simple method to estimate the average localization precision of a single-molecule localization microscopy experiment. *Histochemistry and cell biology* **141**, 629-638 (2014).
- 33 Nieuwenhuizen, R. P. *et al.* Measuring image resolution in optical nanoscopy. *Nat Methods* **10**, 557-562, doi:10.1038/nmeth.2448 (2013).
- 36 Smedsrød, B. & Pertoft, H. Preparation of pure hepatocytes and reticuloendothelial cells in high yield from a single rat liver by means of Percoll centrifugation and selective adherence. *Journal of leukocyte biology* **38**, 213-230 (1985).
- 37 Ahluwalia, B. S. *et al.* Fabrication of submicrometer high refractive index Tantalum Pentoxide waveguides for optical propulsion of microparticles. *Photonics Technology Letters, IEEE* **21**, 1408-1410 (2009).
- 50 Diekmann, R. *et al.* Nanoscopy of bacterial cells immobilized by holographic optical tweezers. *Nat Commun* **7**, 13711, doi:10.1038/ncomms13711 (2016).
- 51 Prieto, F. *et al.* An integrated optical interferometric nanodevice based on silicon technology for biosensor applications. *Nanotechnology* **14**, 907-912, doi:10.1088/0957-4484/14/8/312 (2003).
- 52 van de Linde, S. *et al.* Direct stochastic optical reconstruction microscopy with standard fluorescent probes. *Nature protocols* **6**, 991-1009 (2011).
- 53 Schindelin, J. *et al.* Fiji: an open-source platform for biological-image analysis. *Nature methods* **9**, 676-682 (2012).
- 54 Ovesny, M., Krizek, P., Borkovec, J., Svindrych, Z. & Hagen, G. M. ThunderSTORM: a comprehensive ImageJ plug-in for PALM and STORM data analysis and super-resolution imaging. *Bioinformatics* **30**, 2389-2390, doi:10.1093/bioinformatics/btu202 (2014).

Chip-based wide field-of-view nanoscopy

Robin Diekmann^{#1}, Øystein I. Helle^{#2}, Cristina I. Øie², Peter McCourt³,
Thomas R. Huser^{1,4}, Mark Schüttpelz¹ and Balpreet S. Ahluwalia²

These authors contributed equally to this work.

1 Department of Physics, Bielefeld University, 33615 Bielefeld, Germany.


2 Department of Physics and Technology, UiT-The Arctic University of Norway, 9037 Tromsø, Norway.

3 Department of Medical Biology, UiT-The Arctic University of Norway, 9037 Tromsø, Norway.


4 Department of Internal Medicine and NSF Center for Biophotonics, University of California, Davis, Sacramento, CA 95817, USA

Correspondence should be addressed to M.S. (schuettp@physik.uni-bielefeld.de) or B.S.A. (balpreet.singh.ahluwalia@uit.no).


Supplementary Figure 1 | The resolution of chip-based dSTORM acquired with a 60x NA 1.2 objective lens is investigated using different methods. The lateral profile along a microtubule (**Fig. 2a,b**) shows two distinct peaks corresponding to the separation distance of the shell-like structure of an antibody stained microtubule filament. Their distance of (47 ± 5) nm is measured by fitting the sum of two Gaussians (a). Their FWHM values of (49 ± 7) nm and (40 ± 5) nm affirm the resolution below 50 nm. (b) Consistently, we find a Fourier-ring-correlation resolution^{33,34} of (41.3 ± 2.3) nm for the dSTORM image of **Fig. 2a**. Localization precision values serve as lower bounds for the obtainable resolution. The localization precision distribution found from the photon statistics^{55,56} using the reconstruction software⁵⁴ shows a mean value of 8.5 nm (a), which is in good agreement with the alternative method of using the statistics of nearest neighbours in adjacent frames³² that shows a mean localization precision of (8.3 ± 0.2) nm. Note that the localizations were filtered for localization precisions less than 13 nm.




Supplementary Figure 2 | Mechanical drift in the setup for chip-based *d*STORM. Subdividing the localizations table into multiple stacks, calculating their cross-correlation, and interpolating between the discrete points allows for finding the drift that occurred during the acquisition of the raw data^{54,57}. The graph shows the mechanical drift during the chip-based measurement shown in **Fig. 2c**, which corresponds to about 500 nm after about 25,000 frames. Although this value lies significantly above the obtained resolution it can easily be corrected for in the localizations table before the image is rendered.



Supplementary Figure 3 | Examples of single frames from *d*STORM imaging of Alexa 647 on a conventional upright setup with objective-based TIRF excitation (a) and waveguide chip-based *d*STORM (b) show comparable quality of the raw data. (Scale bars, 1 μm .)




Supplementary Figure 4 | The waveguide allows for TIRF-like imaging, providing optical sectioning via evanescent excitation. We measured the penetration depth of the evanescent field of the waveguide to be (190 ± 59) nm when using illumination of 660 nm vacuum wavelength, and (172 ± 54) nm when using illumination of 488 nm vacuum wavelength. Measurements were conducted by using beads with a mean diameter of $8.18 \mu\text{m}$, whose surface was labelled with Atto 647N or Atto 488, respectively. Measuring the individual bead diameter and finding its centre allows for calculating the height of its surface over the waveguide. The spatially dependent strength of the evanescent field is then measured by the fluorescence intensity^{58,59}. The plot shows a typical result, which can be used to determine the penetration depth by fitting an exponentially decaying function to the data points of a height over the waveguide ranging from 0 to 300 nm.




Supplementary Figure 5 | Fluorescence is excited only by the evanescent field of the waveguide and not by scattered light. We verified this by laterally shifting the strip waveguide chip, which leads to a loss in coupling accompanied with a loss of most of the fluorescence signal from fluorescent beads (200 nm diameter). This is recovered by moving the waveguide back to its original location (**Video 1**).


This figure shows the first frame of the video. As a guide to the eye, the approximate position of the edges of the waveguide (grey line) and the coupling objective lens are shown.




Supplementary Figure 6 | Resolution estimates for chip-based *d*STORM using a 20x NA 0.45 objective lens. (a) The separation distance between adjacent tubulin filaments as shown in **Fig. 2f,g**. Fitting the sum of two Gaussians finds the distance of 138 nm for the distinct peaks. The resolution below 140 nm is confirmed by the FWHM values of (99 ± 14) nm and (105 ± 21) nm as well as a Fourier ring correlation^{33,34} resolution of (122 ± 10) nm (b). This resolution is further verified by the mean localization precision estimates from both the photon statistics^{55,56} using the reconstruction software⁵⁴ (c) and from nearest neighbour analysis³² (d). The nearest neighbour analysis was conducted on a ROI of $90 \mu\text{m} \times 72 \mu\text{m}$ size including the line profile of (a). Note that the localizations were filtered for localization precisions less than 60 nm.




Supplementary Figure 7 | Different wavelengths can be combined simultaneously in chip-based dSTORM by overlaying them before coupling into the waveguide. This allows for effective photoswitching, here demonstrated for CellMask Deep Red. At frame 7650, the 488 nm laser is switched on, providing power at the input facet of the waveguide in addition to the 660 nm laser. Activation of the 488 nm laser line helps in switching CellMask Deep Red back to the fluorescent state, indicated by an increase by 33% in the number of mean localizations per 30 frames.




Supplementary Figure 8 | Diffraction limited illumination pattern measurements. (a) Dense dye labelling of the waveguide surface and averaging over 200 frames allows measuring the homogeneity of the illumination (b). Ignoring the edges, we find a modulation depth of 16 % over the stretch of the waveguide (c) that is used for fluorescence excitation and single-molecule photoswitching. A low modulation depth does not significantly influence the quality of the *d*STORM image as discussed in ³⁵ and verified by our results as the structure of the illumination pattern is not visible in the reconstructed *d*STORM images (**Supplementary Fig. 9**). (d,e) To compare the performance to a commercial setup for *d*STORM imaging, we used ring-TIRF illumination on an OMX v4 Blaze system (GE Healthcare) equipped with a field concentrator. To limit the modulation depth of the illumination intensity to 16 % in this case restricts the field of view to about $18 \mu\text{m} \times 18 \mu\text{m}$ (f). Outside this region, the illumination intensity drops significantly which limits the applicable FOV for *d*STORM imaging. The illumination profile of waveguide TIRF excitation is less smooth, but its modulation does not become apparent in the *d*STORM images. However, in this case the illuminated area is only limited by the waveguide design and no significant drop in illumination intensity is observed over the stretch of the waveguide in both lateral directions. (Scale bars, $10 \mu\text{m}$.)




Supplementary Figure 9 | The use of broad waveguides (25 to 500 μm) results in multimode patterns that cause a laterally inhomogeneous evanescent field. Consequently, fluorescence excitation is non-uniform and the shape of the multimode pattern masks the imaged structures (a, b). Oscillating the coupling objective in ranges that still maintain coupling leads to a movement of the multimode pattern (**Video 2**), (a) and (b) show a ROI from different frames of the video, allowing for correcting for the non-uniform excitation by averaging over multiple frames (c). In this case, we have averaged over 61 frames. During dSTORM imaging, the coupling objective lens is also oscillated. Accordingly, the pattern is not visible in the super-resolved reconstruction of the LSEC membrane (d). (Scale bar, 2 μm .)




Supplementary Figure 10 | Simulation of chip-based ESI. We modelled a random phase sinusoidal illumination pattern with random fringe widths comparable to the measured values exciting emitter pairs at different distances (**Methods**). (a) The average signal over 200 frames shows that two emitters at the diffraction limit of 283 nm distance are just resolved (d), while emitters at 150 nm (c) and 125 nm (b) are not resolved. The ESI reconstruction results in a separation of the emitters at 150 nm distance (c), while the emitters at 125 nm distance are not separated (b), but the emitters at 283 nm distance become more clearly separated (d). (Scale bars, 1 μm .)




Supplementary Figure 11 | The multi-mode interference pattern of a strip waveguide imaged using *d*STORM. A surface of Alexa 647 dye molecules was prepared on top of the waveguide and a *d*STORM image was acquired giving the instant lateral field-distribution of the evanescent field. Interference fringes show structures down to 140 nm as seen in the line profiles. (Scale bar, 5 μm .)




Supplementary Figure 12 | (a) Chip-based ESI reconstruction of tubulin in the LSEC shown in **Fig 3b**. (Scale bar, 5 μm .) (b) and (c) show the region marked by a white box in (a) for diffraction limited imaging and chip-based ESI, respectively. (Scale bars, 1 μm .) (d) Fitting a Gaussian function to the line plots for a tubulin filament shows a FWHM of 231 nm for the diffraction limited image and a FWHM of 104 nm for the ESI image.




Supplementary Figure 13 | (a) Chip-based ESI reconstruction of actin in LSECs as shown in **Fig 4b**. (Scale bar, 20 μm .) (b) and (c) show the region marked with a white box in (a) for diffraction limited imaging and chip-based ESI, respectively. (Scale bars, 2 μm .) (d) Fitting a Gaussian function to the line plots for a single actin filament shows a FWHM of 691 nm for the diffraction limited image and a FWHM of 333 nm for the ESI image.




Supplementary Figure 14 | Comparison of the experimentally achieved localization precision and the theoretical predictions for Atto 488 on the waveguide (a) and for fluorescent beads on a conventional inverted microscope (b). (c)-(f) show example raw data taken under different conditions and (g)-(j) show the corresponding reconstructed images for the entire stack of raw data. See **Supplementary Note 2** for a discussion of the data. Within this figure, the following references are used: [a] = Thompson et al. 2002⁵⁵, [b] = Mortensen et al. 2010⁵⁶, [c] = Ovesny et al. 2014⁵⁴, [d] = Endesfelder et al. 2014³². (Scale bars, 1 μm .)



Supplementary Figure 15 | The comparison of the ROI shown in **Fig. 4f** reconstructed from 3,000 frames and 10,044 frames, respectively, shows that 3,000 frames were sufficient to reconstruct the image with comparable quality using low magnification *d*STORM (**Supplementary Note 4**). E.g. the two separated actin filaments at the edge of the cell are visible. Using a single frame exposure time of 160.84 ms and 3,000 frames leads to an overall acquisition time of 8 minutes, i.e. this approach enables a fast overview of a field of view of large width (**Figure 4c**) at a mean localization precision of 87.6 nm (**Supplementary Figure 16**). The short acquisition time helps to preserve fluorophores for subsequent re-imaging of cells of interest using a higher NA/ higher magnification objective lens. However, we have found that using 95 mM MEA in glucose oxidase and catalase based oxygen scavenger system buffer allows for at least 60 minutes *d*STORM imaging on the waveguide for both Atto 488 and CellMask Deep Red. (Scale bars, 1 μ m.)




Supplementary Figure 16 | The use of a 20x NA 0.4 objective lens allows for imaging a field-of-view of approx. 0.46 mm width (**Fig. 4c**). The measured distance between actin filaments shows a resolution of at least 334 nm in the case of the low NA objective lens used for chip-based *d*STORM. This result is confirmed by localization precision calculations³², estimating the average localization precision to be 87.6 nm as shown here.




Supplementary Figure 17 | Comparison of diffraction limited images and *d*STORM images of the cells shown in **Fig. 5a**. Diffraction limited images were recorded prior to the respective *d*STORM measurement. (Scale bars, 10 μ m.)


a Actin, Atto488-phalloidin
Widefield



b Membrane,
CellMask Deep Red
Widefield



c Actin, Atto488-phalloidin
*d*STORM
Membrane,
CellMask Deep Red
*d*STORM



Supplementary Note 1: Obtainable resolution for waveguide-based ESI

In former work on fluctuating random pattern illumination by Kim et al.³⁹, the authors point out that achievable resolution enhancement with this technique depends on the size of the structures in the speckle pattern. In their work, diffraction limited illumination at a laser vacuum wavelength of 532 nm through a high NA 1.4 objective lens was used, resulting in speckle sizes in the illumination pattern down to $\lambda/2NA = 190$ nm. The authors find a resolution of 144 nm when using SOFI with speckle pattern illumination, which gives a ratio between the theoretically smallest structures in the illumination pattern and the obtained resolution of 1.32. A similar value can be found for their results when using a low NA 0.4 objective lens. In this case, the speckle sizes go theoretically down to 665 nm, while a resolution of 500 nm was demonstrated, resulting in a ratio of 1.33.

In our case, the interference pattern is generated using a laser of 488 nm vacuum wavelength and a waveguide with a refractive index of about 1.7. This results in a lower bound for the interference fringe period in the illumination pattern of about $\lambda/2n_{\text{eff}} = 143$ nm. With this, a spatial resolution on the order of 110 nm (**Fig. 3d, Supplementary Fig. 12**) was achieved using a high NA 1.2 objective lens, which gives a ratio between the theoretically smallest illumination pattern sizes and the obtained resolution of 1.3, and matches the previously reported results in this field. However, we found a lower ratio of 0.43 when using a low NA 0.4 objective lens, though the illumination pattern is independent of the detection objective lens in chip-based ESI, which is in contrast to illumination being provided through the objective lens as performed by Kim et al. Presumably, the signal to noise ratio and other factors also influence the obtainable resolution. Hence, it will be helpful to further investigate these effects in chip-based ESI which might also favour the optimization of the imaging process.

By theory, the lowest bound for possible fringe size resulting from mode interference is found for counterpropagating coherent waves which result in an intensity pattern in the form of a standing wave with half of the wavelength of the light wave. If the fringe size is expressed by a FWHM value, it is half of the wavelength of the wave forming the intensity pattern, i.e. one quarter of the wavelength of the interfering light. Related to the waveguide, the smallest possible value for the structure FWHM of the interference fringes is given by $\text{FWHM}_{\text{min}} = \lambda/(4n_{\text{eff}})$, which results in 97 nm assuming a vacuum wavelength of $\lambda = 660$ nm and an effective refractive index of $n_{\text{eff}} = 1.7$ inside the waveguide.

To be able to compare to this theoretical value, we have measured the multi-mode interference pattern using dSTORM (**Supplementary Fig. 11**) and find fringe sizes that go down to 140 nm, which is about 44 % higher than the theoretically possible minimum fringe size. It is reasonable that the fringe size does not go down to the lowest possible value as

the pattern results from the interference between multiple modes which are propagating in the same direction inside the waveguide.

We use the measured fringe sizes to model the image formation in waveguide-based ESI for emitter pairs at different distances. We simulate their subsequent illumination by sinusoidal patterns with random phases and random FWHM fringe sizes equally distributed between the experimentally obtained lowest value of 140 nm and an upper value of 420 nm (**Methods**). Analysing the data from 200 frames with ESI shows that two emitters spaced at the diffraction limited are easily resolved (**Supplementary Fig. 10**). This also applies for emitters spaced by 150 nm which is well below the diffraction limit while closer emitters at 125 nm spacing are not resolved. Again, we find a ratio between the illumination pattern from 660 nm vacuum wavelength and the achieved resolution of about 1.3.

As in case of waveguide based ESI, the interference pattern is generated within the waveguide and not by diffraction limited optics, the obtainable resolution enhancement is bigger when comparing to the diffraction limited resolution. Hence, we found a resolution of about 110 nm for Alexa 488 emitting at approx. $\lambda = 523$ nm while the diffraction limited resolution limit when employing a NA 1.2 lens is 218 nm, and, thus, a resolution enhancement of about a factor of 2.

Supplementary Note 2: Obtainable resolution for objective lenses of different NA in localization microscopy

As it is rather uncommon to record data in SMLM with low NA objective lenses, we investigate the theoretical limit for the achievable localization precision and compare it to the experimentally obtained values in **Supplementary Fig. 14**.

Both the original model of Thompson et al.⁵⁵ (here referred to as [a]) and its updated version by Mortensen et al.⁵⁶ (referred to as [b]) for the localization precision depending on the statistics of the captured image of an emitter are widely used. A slightly adapted version of the latter is also implemented in ThunderSTORM⁵⁴ (here referred to as [c]), the software that we use for image reconstruction. **Supplementary Fig. 14a** shows the theoretical prediction for the localization precision using either a high NA 1.2 water immersion lens (solid line) or a low NA 0.4 air lens (dashed line) calculated from both models.

For calculations, we used the experimentally achieved parameters for Atto 488 of a mean count of 248 photons/localization and a mean standard deviation of the background of 3.9 photons for the NA 1.2 lens, and a mean count of 128 photons/localization and a mean standard deviation of the background of 4.7 photons for the NA 0.4 lens. Note that different photon count values might stem from different sources, e.g. a different single frame exposure time that was used for the measurements. The photon count values were extrapolated to other NA values by taking the light collection efficiency depending on the acceptance cone of the objective lens into account. Assuming an emission wavelength of $\lambda = 523$ nm for Atto 488, we calculate the standard deviation σ of the fitted Gaussian-shaped point spread function (PSF) by assuming a FWHM equal to the Abbe resolution limit. The backprojected pixel widths A were calculated by assuming a perfect match between the Nyquist sampling capacity of the camera and the resolution of the optical system, i.e. $A = \lambda / 2NA \cdot 1 / 2\sqrt{2}$. Note that while this match is highly appreciated in diffraction-limited imaging, it is often not fulfilled for setups that only serve for SMLM imaging as adjusted backprojected pixel sizes may lead to better fitting results of single molecule emissions. However, this match is almost given for our setup: utilizing the NA 1.2 objective lens for Atto 488 imaging, we use a backprojected pixel width of $A = 75.9$ nm, while the Nyquist-optimal value is $A = 77.0$ nm. In case of NA 0.4, the used backprojected pixel width is $A = 228.3$ nm, while the Nyquist-optimal value is $A = 231.1$ nm.

To compare to the theory, the experimentally achieved localization precisions are estimated in two different ways:

- Using the statistics of the captured image of each emitter in the raw data, the reconstruction software [c] assigns to all localizations an estimated localization precision. The mean value over all localizations for NA 1.2 and for NA 0.4 is drawn in grey in **Supplementary Fig. 14a**, considering the standard deviation as error intervals.
- Another approach to estimating the experimentally achieved average localization precision is proposed by Endesfelder et al.³² (referred to as [d]). This method relies on computations on the distance to nearest neighbours of localizations in adjacent frames of the raw data as these possibly stem from the same molecule. Because this assumption is not always fulfilled, correction terms are considered in the computation. Using the set of localizations from the entire image stack, the findings for the localization precision are drawn in black in **Supplementary Fig. 14a**. Please note that this method is based solely on the localization coordinates and that it ignores the statistics on the raw data, i.e. photon count, background etc.

Experimentally achieved localization precisions estimated by both methods and theoretical predictions nicely fit in case of the NA 1.2 lens. (For better clarity, the representation of the experimental results is slightly shifted in the direction of the abscissa, but both refer to the case of NA 1.2.)

A strong deviation becomes apparent in case of the NA 0.4 lens being used. It is self-evident that the statistics of the localization precision computed by the reconstruction software [c] match the theoretical predictions of the model of [b] as they are based on the same assumptions, while the broad variance is possibly due to high noise in the raw data, of which **Supplementary Fig. 14c** shows an example. For comparison, **Supplementary Fig. 14d** shows an example of the raw data for the NA 1.2 lens being used, and **Supplementary Fig. 14g** and **Supplementary Fig. 14h** show reconstruction of the same area for the entire image stack.

In contrast to the method of [c], the estimation by the method of [d] results in significantly higher values for the localization precision. To explain this, we performed experiments on a conventional inverted microscope configured for SMLM experiments (**Supplementary Fig. 14b**). As samples, we use 100 nm sized fluorescent beads in a low concentration. This approach allows for modelling of single molecule localization experiments with adjustable image statistics comparable to the experiments on the waveguide (e.g. photon count values) by appropriately choosing excitation power and exposure times. Examples from the raw data are shown in **Supplementary Fig. 14e** and **Supplementary Fig. 14f** for the used objective lenses of NA 0.35 and NA 1.49, respectively.

As the beads do not blink, it was possible to localize them over a large number of frames ($n = 2046$ for NA 1.49 and $n = 4153$ for NA 0.35), such that the localization precisions in x-direction (red circle in **Supplementary Fig. 14b**) and y-direction (red diamond) can be measured directly as the standard deviations of the localization distribution, shown in **Supplementary Fig. 14i** and **Supplementary Fig. 14j**. The comparison to the estimation by both methods shows good agreement with the directly measured values for both high and low NA lenses utilized. The higher value for the localization precision in x-direction for the NA 0.35 case is possibly due to drift on the order of 10 nm during the experiment. Again, the localization precision values computed by reconstruction software show a broad variance, which is consistent with the findings on the waveguide.

Inconsistently, we find that in case of the low NA for the bead experiment (**Supplementary Fig. 14b**), the deviation between the estimation of the methods of [c] and [d] is not given as in case of the waveguide experiment (**Supplementary Fig. 14a**). Taking the raw data in case of the low NA lens (**Supplementary Fig. 14c**) and the high NA lens (**Supplementary Fig. 14d**) on the waveguide into account shows that the emitter density looks comparable. However, the individual PSFs are much worse separated in case of the low NA simply due to the significantly larger PSF widths. As the locations of the blinking dyes in a *d*STORM experiment are distributed stochastically, wider PSFs increase the chances of overlapping PSFs (**Supplementary Note 3**) which can lead to (1) a wrong emitter position estimation as a localization will be found in between the real positions of the multiple emitters, and (2) a wrong estimation on the single emitter photon count as this will be based on the photon emission of the multiple emitters simultaneously. This can now explain the deviation of the localization precision estimations by the methods of [c] and [d] considering their underlying principles: The method of [c] performs an estimation of the precision for each localization on the statistics of the raw data independent of the complete reconstructed image. If an erroneous localization occurs due to overlapping PSFs of multiple emitters, this affects the quality of the reconstructed image but also the localization precision statistics. In contrast, [d] performs computations only on the localization coordinate list that is used to render the super-resolved image. Hence, deviations between the two methods can indicate the problem of too dense single molecule blinking in the raw data.

However, due the stochastic nature of the process, overlapping of the PSFs of multiple emitters cannot be fully avoided in *d*STORM, but the probability can be decreased by tuning the experimental conditions in terms of imaging buffer composition, excitation power, single frame exposure time, and not least choosing an appropriate dye. While Atto 488 performs well for multiple subsequent acquisitions of the same sample that we used to compare between different modalities (**Fig. 3**) due to its relative good photostability⁶⁰, Alexa 647 is

often considered as a better option if the focus is laid on *d*STORM imaging only. Consequently, we could achieve a higher resolution in *d*STORM imaging using the latter dye (**Fig. 2**). This is confirmed by the very good agreement for the localization precisions determined by both methods for low and high NA chip-based *d*STORM imaging (**Supplementary Fig. 1** and **Supplementary Fig. 6**). Anyway, it was possible to achieve a resolution beyond the diffraction limit for low NA (0.4 and 0.45) lenses on the waveguide for both Atto 488 and Alexa 647, showing that our approach allows for large FOV *d*STORM imaging in TIRF excitation.

Supplementary Note 3: Resolution estimation on emitter density

A condition for the lower bound on the achievable resolution referring to the density of simultaneously active emitters, i.e. the rate of blinking, was introduced by Xu et al.¹³ and given as $(\pi(\lambda/(2NA))^2 r_{\max})^{1/2}$. As the parameters of the equation, we assume the emission wavelength $\lambda = 523$ nm (estimated here as the emission spectrum peak of Atto 488), the numerical aperture of the setup $NA = 0.4$, and the ratio between the maximum number of detected localization in one frame divided by the overall number of localization $r_{\max} = 1,821/9,628,118$. This estimation gives a lower bound of 16 nm which is significantly lower than the actually achieved resolution, so it should not set a limit in this case.

However, this assumption takes dense packing of emitters without overlap of their PSFs into account. As the locations of active emitters are independent of each other and randomly distributed, dense packing without overlap is unlikely to occur in a *d*STORM experiment, but overlap of active emitters by their PSF will occur at a certain probability depending on the labelling density, photoswitching properties, and PSF size⁶¹. Assuming the size of a diffraction limited spot as $D = \pi(\lambda/(2NA))^2$ similar to the estimation by Xu et al.¹³ and the overall imaged area A , the probability for one specific emitter to overlap with a second emitter is given by D/A . Hence the probability for no overlap with a second emitter follows as $1-D/A$ and for no overlap with m other emitters as $(1-D/A)^m$. Consequently, the probability for one specific emitter to not overlap with any other emitter in case of $n = m+1$ simultaneously active emitters is given by $1-(1-D/A)^{n-1}$. For the *d*STORM image shown in **Fig. 4c**, we have $A = 3,456 \mu\text{m}^2$, $\lambda = 523$ nm, $NA = 0.4$. The maximum number of simultaneously active emitters is $n_{\max} = 1,821$ while the mean number of simultaneously active emitters is $n_{\text{mean}} = 959$ (**Supplementary Note 4**). It follows that for n_{\max} , each emitter has the probability to overlap with another emitter of 50,7%, while for n_{mean} , the probability is 31,1 %. Overlapping PSFs will lead to false localizations and can therefore affect the quality of the reconstructed image (**Supplementary Note 2**), also depending on the ability of the reconstruction algorithm to discern close emitters.

Supplementary Note 4: Resolution estimation on localization density

As pointed out in ^{62,63}, the Nyquist resolution limit can be defined as $2/(\text{localization density})^{1/2}$ in the case of two-dimensional localization microscopy, e.g. dSTORM. Accordingly, to allow for a Nyquist resolution on the order of 140 nm, the localization density has to exceed 205 localizations/ μm^2 . However, as has been discussed recently⁶⁴, oversampling by at least a factor of 5 is desired due to the stochastic nature of localization microscopy. Hence, the localization density has to be 1020 localizations/ μm^2 to support a resolution of 140 nm. The low NA dSTORM reconstruction of Atto 488-phalloidin stained LSECs shown in **Fig. 4c** comprises of 9,628,118 localizations in 10,044 frames over a field-of-view of $3,456 \mu\text{m}^2$. Within this field-of-view, about 2/3 of the area is actually covered by cells, densely packed with actin. From these numbers, we estimate the localization density in a reconstruction from 3,000 frames to be $9,628,118 \text{ localizations} \times (3,000 \text{ frames}) / (10,044 \text{ frames}) \times (2/3 \times 3,456 \mu\text{m}^2)^{-1} = 1,248 \text{ localizations}/\mu\text{m}^2$. Thus, the achieved localization density exceeds 1020 localizations/ μm^2 and gives a lower bound for the resolution on the order of 140 nm. In case of Atto 488 imaging, we achieved a localization precision of 88 nm (**Supplementary Fig. 16**) indicating an achievable resolution on the order of $2 \times (2 \times \ln(2))^{1/2} \times \text{localization precision} \approx 207 \text{ nm}$. So the localization precision determines the limit for the resolution, but not the labelling density, which is further confirmed by the reconstructed image (**Supplementary Fig. 15**).

Supplementary Note 5: Comparison to large FOV *d*STORM via mosaic stitching

Fig. 2e shows *d*STORM imaging at approx. 140 nm resolution over a FOV of 0.5 mm × 0.5 mm that has been reconstructed from 21,716 raw frames recorded in approximately 18 minutes (**Supplementary Table 1**). Circumventing the need for modifying the illumination scheme, the FOV for localization microscopy can also be extended by stitching of multiple areas as it has recently been demonstrated by Nahidiazar et al.⁶⁵. The authors show that the stitching of eight regions resulted in the *d*STORM image of an almost entire HUVEC cell. For each region, 20,000 raw frames at 100 Hz had been recorded on a commercial setup. Hence, 160,000 raw frames recorded in approx. 26 minutes and subsequent stitching were necessary for the *d*STORM reconstruction of the entire cell. Though the authors do not give an estimated resolution for this example, we assume that the obtained resolution is much better than 140 nm. However, using the waveguide-based approach and accepting a worse resolution (which is still well beyond the diffraction limit) more than 50 cells compared to 1 cell could be imaged in approx. 18 minutes compared to 26 minutes.

Supplementary Table 1: Experimental conditions

	Experiment	Waveguide material and coupling device	Waveguide breadth (μm)	Imaging objective lens	Laser vacuum wavelength (nm)	Dye	Approximate illumination intensity (kW/cm^2)	Exposure time (ms)	# frames	Gaussian smoothing kernel width (sigma) (nm)
Figure 2a	dSTORM	Si_3N_4 , fibre	25	60x NA 1.2	660	Alexa 647 anti tubulin (primary and secondary antibody staining)	0.9	50	35043	-
Figure 2a	Diffraction limited	Si_3N_4 , fibre	25	60x NA 1.2	660	Alexa 647 anti tubulin (primary and secondary antibody staining)	0.05	100	505	-
Figure 2c	dSTORM	Ta_2O_5 , objective	65	60x NA 1.2	660	Alexa 647	0.1	30	24665	10
Figure 2c	dSTORM	-	-	60x NA 1.49	647	Alexa 647	5	50	6026	10
Figure 2e,f	dSTORM	Si_3N_4 , objective	500	20x NA 0.45	660	Alexa 647 anti tubulin (primary and secondary antibody staining)	0.9	50	21716	-
Figure 2e,f	Diffraction limited	Si_3N_4 , objective	500	20x NA 0.45	660	Alexa 647 anti tubulin (primary and secondary antibody staining)	0.0004	100	525	-
Figure 3b,c, Supplementary Figure 12b	Diffraction limited	Ta_2O_5 , objective	50	60x NA 1.2	488	Alexa 488 anti alpha-tubulin	0.003	300	202	-
Figure 3b,c, Supplementary Figure 12a,c	ESI	Ta_2O_5 , objective	50	60x NA 1.2	488	Alexa 488 anti alpha-tubulin	0.003	300	202	-
Figure 3b,c	dSTORM	Ta_2O_5 , objective	50	60x NA 1.2	488	Alexa 488 anti alpha-tubulin	0.3	20	30041	16.3
Figure 4a,d Supplementary Figure 13b	Diffraction limited	Ta_2O_5 , objective	65	20x NA 0.4	488	Atto 488 -phalloidin	0.03	100	200	-
Figure 4b,e, Supplementary Figure 13a,c	ESI	Ta_2O_5 , objective	65	20x NA 0.4	488	Atto 488 -phalloidin	0.03	100	208	-
Figure 4c,f, Supplementary Figure 15	dSTORM	Ta_2O_5 , objective	65	20x NA 0.4	488	Atto 488 -phalloidin	0.4	160.84	10044	68.5
Figure 4g	dSTORM	Ta_2O_5 , objective	65	60x NA 1.2	488	Atto 488 -phalloidin	0.4	40	15048	22.8
Figure 5a, Supplementary Figure 17c	dSTORM	Ta_2O_5 , objective	65	60x NA 1.42	488	Atto 488 -phalloidin	0.4	50	75018	30.4
Figure 5a, Supplementary Figure 17c	dSTORM	Ta_2O_5 , objective	65	60x NA 1.42	660 and 488	CellMask Deep Red	all frames: $9 \text{ kW}/\text{cm}^2$ for 660 nm; frames 7650 to 125565: $0.4 \text{ kW}/\text{cm}^2$ for 488 nm	70	125565	30.4
Supplementary Figure 3a	Diffraction limited	-	-	60x NA 1.49	647	Alexa 647		30	1	-
Supplementary Figure 3b	Diffraction limited	Ta_2O_5 , objective	42	60x NA 1.2	660	Alexa 647	10	10	1	-
Supplementary Figure 5, Video 2	Diffraction limited	Ta_2O_5 , objective	75	60x NA 1.2	660	200 nm TetraSpeck Microspheres	6	100	229 (Video)	-
Supplementary figure 8a	Diffraction limited	Ta_2O_5 , objective	42	60x NA 1.2	660	Alexa 647	0.01	100	200	-
Supplementary figure 8d	Diffraction limited	-	-	60x NA 1.49	642	Alexa 647		100	200	-
Supplementary Figure 9a,b,c, Video 1	Diffraction limited	Ta_2O_5 , objective	42	60x NA 1.42	660	CellMask Deep Red	2	100	61 (Video)	-
Supplementary Figure 9d, Video 1	dSTORM	Ta_2O_5 , objective	42	60x NA 1.42	660	CellMask Deep Red	13	100	416770	30.4
Supplementary Figure 11	dSTORM	Ta_2O_5 , objective	42	60x NA 1.2	660	Alexa 647	11	10		
Supplementary Figure 14e	Emitter detection	-	-	20x, NA 0.35	647	200 nm TetraSpeck Microspheres	0.03	15.37	2052	-
Supplementary Figure 14f	Emitter detection	-	-	60x NA 1.49	647	200 nm TetraSpeck Microspheres	0.01	5	4153	-
Supplementary Figure 17a	Diffraction limited	Ta_2O_5 , objective	65	60x NA 1.42	488	Atto 488 -phalloidin	0.03	200	200	-
Supplementary Figure 17b	Diffraction limited	Ta_2O_5 , objective	65	60x NA 1.42	660	CellMask Deep Red	5	100	300	-

REFERENCES FOR SUPPLEMENTARY INFORMATION

- 13 Xu, J. Q., Tehrani, K. F. & Kner, P. Multicolor 3D Super-resolution Imaging by Quantum Dot Stochastic Optical Reconstruction Microscopy. *Acs Nano* **9**, 2917-2925, doi:10.1021/nn506952g (2015).
- 32 Endesfelder, U., Malkusch, S., Fricke, F. & Heilemann, M. A simple method to estimate the average localization precision of a single-molecule localization microscopy experiment. *Histochemistry and cell biology* **141**, 629-638 (2014).
- 33 Nieuwenhuizen, R. P. *et al.* Measuring image resolution in optical nanoscopy. *Nat Methods* **10**, 557-562, doi:10.1038/nmeth.2448 (2013).
- 34 Banterle, N., Bui, K. H., Lemke, E. A. & Beck, M. Fourier Ring Correlation as a resolution criterion for super resolution microscopy. *Journal of Structural Biology* **183** (2013).
- 35 Douglass, K. M., Sieben, C., Archetti, A., Lambert, A. & Manley, S. Super-resolution imaging of multiple cells by optimised flat-field epi-illumination. *Nat Photonics* **10**, 705-708, doi:10.1038/nphoton.2016.200 (2016).
- 39 Kim, M., Park, C., Rodriguez, C., Park, Y. & Cho, Y. H. Superresolution imaging with optical fluctuation using speckle patterns illumination. *Sci Rep-Uk* **5**, 16525, doi:10.1038/srep16525 (2015).
- 54 Ovesny, M., Krizek, P., Borkovec, J., Svindrych, Z. & Hagen, G. M. ThunderSTORM: a comprehensive ImageJ plug-in for PALM and STORM data analysis and super-resolution imaging. *Bioinformatics* **30**, 2389-2390, doi:10.1093/bioinformatics/btu202 (2014).
- 55 Thompson, R. E., Larson, D. R. & Webb, W. W. Precise nanometer localization analysis for individual fluorescent probes. *Biophys J* **82**, 2775-2783, doi:10.1016/S0006-3495(02)75618-X (2002).
- 56 Mortensen, K. I., Churchman, L. S., Spudich, J. A. & Flyvbjerg, H. Optimized localization analysis for single-molecule tracking and super-resolution microscopy. *Nat Methods* **7**, 377-381, doi:10.1038/nmeth.1447 (2010).
- 57 Mlodzianoski, M. J. *et al.* Sample drift correction in 3D fluorescence photoactivation localization microscopy. *Opt Express* **19**, 15009-15019, doi:10.1364/OE.19.015009 (2011).
- 58 Mattheyses, A. L. & Axelrod, D. Direct measurement of the evanescent field profile produced by objective-based total internal reflection fluorescence. *Journal of biomedical optics* **11**, 014006-014006-014007 (2006).
- 59 Stabley, D. R., Oh, T., Simon, S. M., Mattheyses, A. L. & Salaita, K. Real-time fluorescence imaging with 20 nm axial resolution. *Nature communications* **6** (2015).
- 60 Dempsey, G. T., Vaughan, J. C., Chen, K. H., Bates, M. & Zhuang, X. Evaluation of fluorophores for optimal performance in localization-based super-resolution imaging. *Nat Methods* **8**, 1027-1036, doi:10.1038/nmeth.1768 (2011).
- 61 van de Linde, S., Wolter, S., Heilemann, M. & Sauer, M. The effect of photoswitching kinetics and labeling densities on super-resolution fluorescence imaging. *J Biotechnol* **149**, 260-266, doi:10.1016/j.jbiotec.2010.02.010 (2010).
- 62 Shroff, H., Galbraith, C. G., Galbraith, J. A. & Betzig, E. Live-cell photoactivated localization microscopy of nanoscale adhesion dynamics. *Nature Methods* **5**, 417-423, doi:10.1038/Nmeth.1202 (2008).
- 63 Shim, S. H. *et al.* Super-resolution fluorescence imaging of organelles in live cells with photoswitchable membrane probes. *P Natl Acad Sci USA* **109**, 13978-13983, doi:10.1073/pnas.1201882109 (2012).
- 64 Legant, W. R. *et al.* High-density three-dimensional localization microscopy across large volumes. *Nat Methods*, doi:10.1038/nmeth.3797 (2016).
- 65 Nahidiazar, L., Agronskaia, A. V., Broertjes, J., van den Broek, B. & Jalink, K. Optimizing Imaging Conditions for Demanding Multi-Color Super Resolution Localization Microscopy. *Plos One* **11**, doi:10.1371/journal.pone.0158884 (2016).