

# Dental Pulp Neurophysiology: Part 2. Current Diagnostic Tests to Assess Pulp Vitality

Ashraf Abd-Elmeguid, BDS, MDSc; Donald C. Yu, DMD, CAGS, MScD, FRCD(C)

## Contact Author

Dr. Yu  
Email: [donaldyu@ualberta.ca](mailto:donaldyu@ualberta.ca)



## ABSTRACT

In this second part of our 2-part review, we discuss recent research about pulp tests that determine the vitality of the tooth and clinically accepted pulp testers. A pain response to hot, cold or an electric pulp tester indicates the vitality of only a tooth's pulpal sensory supply; the response does not give any idea about the state of the pulp. Although the sensitivity of these tests is high, when false-positive and false-negative results occur, they may affect the treatment of the tooth. A tooth falsely diagnosed as nonvital with an electric pulp tester may undergo an unnecessary root canal, whereas one falsely diagnosed as vital may be left untreated, causing the necrotic tissue to destroy the supporting tissues (resorption). The vascular supply is more important to the determination of the health of the pulp than the sensory supply. Pulp death is caused by cessation of blood flow and may result in a necrotic pulp, even though the pulpal sensory supply may still be viable. The pulp can be healed only if the circulating blood flow is healthy. Although still under investigation, diagnostic devices that examine pulpal blood flow, such as the pulse oximeter and laser Doppler flowmetry, show promising results for the assessment of pulp vitality.

For citation purposes, the electronic version is the definitive version of this article: [www.cda-adc.ca/jcda/vol-75/issue-2/139.html](http://www.cda-adc.ca/jcda/vol-75/issue-2/139.html)

In part 1 of this 2-part review, we discussed the importance of understanding the neurophysiology of the dental pulp.<sup>1</sup> In part 2, we examine the relation between the pulpal neural distribution and several pulp tests that determine the neural response rather than the state of the vascular supply which determines the vitality of any tissue. We also examine recent innovative tests that determine vascularity. We do not discuss the different techniques that help clinicians reach a definitive diagnosis, such as digital radiography and radiovisiography, because these are beyond the scope of this review. Neither do we discuss the cavity test as a mechanical way to test

whether the coronal tooth pulp is necrotic because this test relies mainly on a negative response (i.e., no response) to the electric pulp test, together with information obtained from a periapical radiograph. In this second part of our 2-part review, we discuss recent research about pulp tests that determine the vitality of the tooth and clinically accepted pulp testers.

## Diagnostic Tests Related to the Neurophysiology of the Dental Pulp

### Electric Pulp Tester

The electric pulp tester is widely used to differentiate between lesions of endodontic origin and those not seen on radiographs.<sup>2</sup>



Figure 1: Electric pulp tester.

This device (Fig. 1) is designed to deliver an electric current to stimulate the closest myelinated A-delta fibres; the device does not usually stimulate the unmyelinated C fibres because of their higher threshold.<sup>3</sup> The electric pulp tester indicates the neural transmission and presence of vital nerve fibres, but does not measure the health or integrity of the pulp. Recently traumatized teeth that may temporarily lose their sensory function have no response to the device, even though their vascularity is intact (false-negative),<sup>4,5</sup> whereas teeth that are partly necrotic may give a response, even though they lack a blood supply (false-positive).<sup>6</sup>

An interface medium is necessary to conduct electrical impulses to the tooth<sup>7</sup>; this medium should be non-liquid-based because a liquid-based medium may give false-positive results if contact with the gingival tissues<sup>8,9</sup> or saliva<sup>10</sup> occurs. In a recent study, Mickel and others<sup>8</sup> found that K-Y lubricating gel and Crest baking soda and peroxide whitening tartar provided maximum conduction of electrical current to the cathode. The authors concluded that a good conduction medium is important in cases in which a false-negative result is predicted, such as in pulp-canal obliteration and recently traumatized teeth.

Also, the location of the electrode on the buccal surface was tested in several studies.<sup>7,11</sup> Bender and others<sup>11</sup> found that the best placement of the electrode is the incisal third of anterior teeth where the least amount of electric current provoked a response. Others placed the electrode on the occlusal third of the buccal surface,<sup>12</sup> centred between the gingival margin and the occlusal edge of the buccal surface<sup>13-15</sup> or on the gingival third of the buccal surface.<sup>7,16,17</sup> A recent study<sup>18</sup> specifically designed to determine the best site for the placement of the electrode on the molar tooth surface found no significant

difference between maxillary and mandibular molars in male and female subjects. Placement of the electrode on the mesiobuccal cusp tip elicited the lowest response. Placing the electrode more apically and to the centre of the supporting cusps showed an increase in the threshold response level. These results were related to the presence of pulp horns, where there is a high concentration of neural elements.<sup>19,20</sup>

On the basis of our clinical experience, it is our opinion that the best placement of the electrode is at the gingival third of the buccal surface of the natural tooth structure and that an electrocardiogram gel should be used as the interface medium.

### Thermal Tests (Application of Cold and Heat)

The application of heated water (hot water bath) or softened, heated gutta-percha to the tooth is commonly used to deliver heat to the pulp. These methods may produce sufficient heat to stimulate the C fibres and produce pain that lingers, is prolonged and is usually delayed about 2 to 4 seconds. It is important to use heat cautiously to avoid any damage to the pulp tissue.<sup>21,22</sup>

Several methods are used to apply cold to the teeth, such as ice sticks (0°C), CO<sub>2</sub> sticks (-78°C), ethyl chloride (-5°C) and dichlorodifluoromethane ([DDM] -50°C).<sup>23</sup> Fuss and others,<sup>24</sup> who assessed the reliability of thermal tests and electrical tests for adults and young patients, found that CO<sub>2</sub> and DDM are more effective than ice and ethyl chloride. CO<sub>2</sub> produces a greater decrease in pulp temperature than DDM<sup>25</sup> and a quick response from the pulp. This greater decrease in temperature has no detrimental effect on the pulp tissue.<sup>24,26-29</sup> In another study<sup>30</sup> to determine the effect of different carriers on transferring the cold application to the tooth, the authors reported that a large cotton pellet is preferable to a small one, to cotton rolls or to a wooden-handle cotton tip. They also reported that sprayed DDM produced a more efficient, colder effect than the dipping technique. The manufacturer reformulated the DDM into 1,1,1,2-tetrafluoroethane that has a low liquid temperature and, presumably, is more environmentally safe. Jones and others<sup>31</sup> found that this refrigerant spray was more likely to produce a response than the CO<sub>2</sub> dry ice. The spray also cooled the tooth in significantly less time, regardless of whether the tooth was restored or covered, or neither.

### Importance of Assessing the Blood Supply as an Indication for Pulp Vitality

As Cohen and Burns<sup>10</sup> reported, response to current clinical tests indicates only that sensory fibres are vital. However, 10%–16% of the results of these tests are false.<sup>6</sup> The nervous system, which is highly resistant to inflammation, may remain reactive, even though all surrounding tissues have degenerated; therefore testing the sensory supply may give a positive response when the

pulp is damaged (i.e., a false-positive result).<sup>32</sup> This test may also leave the patient with an unpleasant sensation.<sup>33</sup> A false-negative result (i.e., no response) may be obtained in cases of calcific metamorphosis,<sup>34</sup> recently traumatized teeth<sup>35,36</sup> and incomplete root formation.<sup>34</sup>

The vitality of the pulp is determined according to the health of the vascular supply, not of the sensory fibres.<sup>4,34,37,38</sup> The pulp receives its blood supply through thin-walled arterioles entering through the apical and accessory foramina. These arterioles run longitudinally through the centre of the pulp, branching out to its periphery where they form a capillary network in the subodontoblastic area. These capillaries do not enter the dentin; they drain into the venules that run alongside the arterioles and pass out through the same apical foramen.<sup>39,40</sup>

Different methods may be used to assess the blood flow in the pulp: for example, isotope clearance,<sup>41</sup> local hydrogen-gas desaturation<sup>42</sup> and labelled microspheres.<sup>43</sup> Because of the limitations on the use of isotopes with humans, these methods remain experimental (in vitro). A study<sup>44</sup> to determine whether a change in tooth temperature can trigger pulpal blood flow concluded that this method of assessing blood flow in the pulp was not clinically reliable.

### **Experimental Diagnostic Methods to Assess the Blood Supply**

Although they have a few drawbacks, the techniques described in this section have the most potential for clinical application in the future.

#### **Dual-wavelength Spectrophotometry**

Dual-wavelength spectrophotometry (DWS), which is done with a noninvasive portable instrument, can be used to test pulpal blood flow. Oximetry by spectrophotometer determines the level of oxygen saturation in the pulpal blood supply with a dual-wavelength light source (760 and 850 nm).<sup>45,46</sup> This instrument may be useful for determining pulp necrosis and the inflammatory status of the pulp.<sup>45</sup>

Nissan and others<sup>46</sup> did an in vitro study to determine the feasibility of using DWS to identify teeth with pulp chambers that are either empty, filled with fixed pulp tissue or filled with oxygenated blood. Their findings indicated that continuous-wave spectrophotometry may be a useful method for testing pulp.

#### **Pulse Oximetry**

Since the study of Nissan and others,<sup>46</sup> further research has focused on pulse oximetry, which is based on DWS. Pulse oximetry is widely used in medical practice to measure levels of oxygen saturation during the administration of intravenous anesthesia,<sup>47</sup> and is routinely used in emergency rooms and in situations in which sedation

and analgesia are used.<sup>48</sup> Pulse oximetry is noninvasive and atraumatic, characteristics that make it valuable to dentistry. This technology is based on a modification of Beer's law: namely, the absorption of light by a solute is related to its concentration at a given wavelength.<sup>49</sup> Pulse oximetry also uses the characteristics of hemoglobin in the red and infrared range: oxyhemoglobin absorbs more light in the infrared range than deoxyhemoglobin, and vice versa in the red range. Schmitt and others<sup>50</sup> found that pulse oximetry effectively determined the oxygen saturation in a tooth model in vitro. Noblett and others<sup>33</sup> who used a rubber dam clamp as a base for the sensor design in an in vitro pulpal-circulation model found that this design accurately determined oxygen saturation in blood circulating through the pulp chamber of a tooth model. Kahan and others<sup>51</sup> examined a modified tooth probe and found no consistency between its results and those of a finger probe. Gopi Krishna and others<sup>52</sup> produced consistent readings when they compared a customized dental sensor with the finger sensor. They recommended that sensors should suit the anatomy of the tooth and that the light-emitting sensor and the photoreceptor should be parallel to each other. They also recommended that the probe firmly grip the tooth surface for accurate measurement. Gopikrishna and others<sup>53</sup> compared a custom-made pulse-oximeter dental probe with thermal and electrical tests for the assessment of pulp vitality. They found that the sensitivity of the pulse oximeter is 100%; the cold test, 81%; and the electrical pulp tester, 71%. (Sensitivity indicates the ability of a test to report disease in patients who have the disease.<sup>6</sup>) This same group carried out a study<sup>54</sup> comparing the ability of the new pulse-oximeter method with that of the electrical and thermal tests to measure pulp vitality in recently traumatized teeth, whose condition can be complicated by a delay in diagnosis. They used a modified pulse oximeter that had a multisize oxygen sensor of small dimensions suitable for placement on human teeth and a sensor holder that maintained the stability of the teeth and the sensor. They reported constant vitality readings throughout the study from 0 to 6 months with the pulse oximeter, and variable readings with the electrical and thermal tests (responses varied from none on day 0 to beginning to respond on day 28 to almost total response for the 3-month period).

#### **Laser Doppler Flowmetry**

Laser Doppler flowmetry (LDF) is an accurate, noninvasive, reproducible, reliable method<sup>55-58</sup> of assessing blood flow in microvascular systems with a diode that projects an infrared light beam through the crown and pulp chamber. This light beam is scattered through moving red cells and static tissues.<sup>59</sup> Its frequency shifts when the beam passes through moving red blood cells, but remains constant when the beam passes through

static tissue.<sup>59,60</sup> The LDF technique takes about an hour to produce recordings, making it impractical for dental practices unless its time frame can be shortened to a few minutes.

In dentistry, LDF was used to assess pulpal blood flow as an indication of the vitality of traumatized teeth.<sup>56,61–63</sup> LDF was also used to assess gingival blood flow in flaps after ridge augmentation<sup>64</sup> and during Le Fort I osteotomy,<sup>65</sup> and to assess blood flow in intact teeth in animals<sup>66</sup> and in man.<sup>55</sup> Gazelius and others<sup>55</sup> used LDF to study pulpal blood flow with He-Ne light, a general purpose for LDF, rather than one optimized for measuring pulpal blood flow. Pettersson and Öberg<sup>59</sup> designed an LDF instrument for measuring blood flow in human pulp and used it to assess the viability of pulp in intact and traumatized teeth. They used an infrared laser diode with a longer wavelength that gave better penetration than the He-Ne wavelength. Sasano and others<sup>67,68</sup> designed and developed a transmitted laser-light flow meter that used high-powered laser light to monitor the pulpal blood flow of teeth rather than the conventional light flow-meter apparatus. Konno and others<sup>69</sup> used the same apparatus to evaluate changes in pulpal blood flow in dogs' molar intrusion, using a skeletal anchorage system.

LDF is reported to be technique-sensitive<sup>61</sup>: its readings are affected by the movement of the patient, a non-fixed probe or a mobile tooth.<sup>70,71</sup> The technique yields false-positive results when used for endodontically treated teeth<sup>72</sup> and when the gingival blood flow is measured.<sup>70,71</sup> Moreover, intracoronary and extracoronary scattering of the laser beam calls for special precautions such as covering the gingiva and the crown of the tooth.<sup>73,74</sup>

## Conclusions

Information about the physiology of pulpal pain and the sensory fibres causing this pain, together with information gathered from the patient, and the use of appropriate devices to test pulp sensitivity and vitality are very critical to reaching an accurate diagnosis on which to base an appropriate treatment plan. Multiple devices that test pulp viability are available on the market, but they test the viability of nerve fibres as a measure of pulp vitality, resulting sometimes in false-positive or false-negative results. These can lead to unnecessary endodontic procedures if these tests are not substantiated with results of other diagnostic measures. Pulpal blood flow, which is at least as important as testing the neural supply of the pulp, must also be examined. Although still being studied, methods to test blood flow look very promising and should soon be in use in the dental clinic. ➤

## THE AUTHORS



**Dr. Abd-Elmeguid** is a PhD student, medical sciences and dentistry, department of dentistry, University of Alberta, Edmonton, Alberta.



**Dr. Yu** is a clinical professor and head of the endodontic division, department of dentistry, University of Alberta, Edmonton, Alberta.

**Correspondence to:** Dr. Donald Yu, Room 4021, Dentistry/Pharmacy Centre, Department of dentistry, University of Alberta, Edmonton, AB T6G 2N8.

The authors have no declared financial interests.

This article has been peer reviewed.

## References

1. Abd-Elmeguid A, Yu D. Dental pulp neurophysiology: Part 1. Clinical and diagnostic implications. *J Can Dent Assoc* 2009; 75(1):55–9.
2. Cooley RL, Stillely J, Lubow RM. Evaluation of a digital pulp tester. *Oral Surg Oral Med Oral Pathol* 1984; 58(4):437–42.
3. Narhi M, Virtanen A, Kuhta J, Huopaniemi T. Electrical stimulation of teeth with a pulp tester in the cat. *Scand J Dent Res* 1979; 87(1):32–8.
4. Bhaskar SN, Rappaport HM. Dental vitality tests and pulp status. *J Am Dent Assoc* 1973; 86(2):409–11.
5. Bernick S, Nedelman C. Effect of aging on the human pulp. *J Endod* 1975; 1(3):88–94.
6. Petersson K, Soderstrom C, Kiani-Anaraki M, Levy G. Evaluation of the ability of thermal and electrical tests to register pulp vitality. *Endod Dent Traumatol* 1999; 15(3):127–31.
7. Cooley RL, Robison SF. Variables associated with electric pulp testing. *Oral Surg Oral Med Oral Pathol* 1980; 50(1):66–73.
8. Mickel AK, Lindquist KA, Chogle S, Jones JJ, Curd F. Electric pulp tester conductance through various interface media. *J Endod* 2006; 32(12):1178–80.
9. Michaelson RE, Seidberg BH, Guttuso J. An in vivo evaluation of interface media used with the electric pulp tester. *J Am Dent Assoc* 1975; 91(1):118–21.
10. Cohen S, Burns RC, editors. Pathways of the pulp. 8th ed. St. Louis: Mosby; 2002.
11. Bender IB, Landau MA, Fonseca S, Trowbridge HO. The optimum placement-site of the electrode in electric pulp testing of the 12 anterior teeth. *J Am Dent Assoc* 1989; 118(3):305–10.
12. Anderson RW, Pantera EA Jr. Influence of a barrier technique on electric pulp testing. *J Endod* 1988; 14(4):179–80.
13. Carnes PL, Cook B, Eleazer PD, Scheetz JP. Change in pain threshold by meperidine, naproxen sodium, and acetaminophen as determined by electric pulp testing. *Anesth Prog* 199; 45(4):139–42.
14. Branco FP, Ranali J, Ambrosano GM, Volpato MC. A double-blind comparison of 0.5% bupivacaine with 1:200,000 epinephrine and 0.5% levobupivacaine with 1:200,000 epinephrine for the inferior alveolar nerve block. *Oral Surg Oral Med Oral Pathol Oral Radiol Endod* 2006; 101(4):442–7.
15. Goodman A, Reader A, Nusstein J, Beck M, Weaver J. Anesthetic efficacy of lidocaine/meperidine for inferior alveolar nerve blocks. *Anesth Prog* 2006; 53(4):131–9.
16. Certosimo AJ, Archer RD. A clinical evaluation of the electric pulp tester as an indicator of local anesthesia. *Oper Dent* 1996; 21(1):25–30.
17. Modaresi J, Dianat O, Mozayeni MA. The efficacy comparison of ibuprofen, acetaminophen-codeine, and placebo premedication therapy on the depth of anesthesia during treatment of inflamed teeth. *Oral Surg Oral Med Oral Pathol Oral Radiol Endod* 2006; 102(3):399–403.
18. Lin J, Chandler N, Purton D, Monteith B. Appropriate electrode placement site for electric pulp testing first molar teeth. *J Endod* 2007; 33(11):1296–8.
19. Lilja J. Sensory differences between crown and root dentin in human teeth. *Acta Odontol Scand* 1980; 38(5):285–91.

20. Byers MR, Dong WK. Autoradiographic location of sensory nerve endings in dentin of monkey teeth. *Anat Rec* 1983; 205(4):441–54.
21. Zach L, Cohen G. Pulp response to externally applied heat. *Oral Surg Oral Med Oral Pathol* 1965; 19:515–30.
22. Baldissara P, Catapano S, Scotti R. Clinical and histological evaluation of thermal injury thresholds in human teeth: a preliminary study. *J Oral Rehabil* 1997; 24(11):791–801.
23. Fleury A, Regan JD. Endodontic diagnosis: clinical aspects. *J Ir Dent Assoc* 2006; 52(1):28–38.
24. Fuss Z, Trowbridge H, Bender IB, Rickoff B, Sorin S. Assessment of reliability of electrical and thermal pulp testing agents. *J Endod* 1986; 12(7):301–5.
25. Augsburger RA, Peters DD. In vitro effects of ice, skin refrigerant, and CO<sub>2</sub> snow on intrapulpal temperature. *J Endod* 1981; 7(3):110–6.
26. Peters DD, Mader CL, Donnelly JC. Evaluation of the effects of carbon dioxide used as a pulpal test. 3. In vivo effect on human enamel. *J Endod* 1986; 12(1):13–20.
27. Rickoff B, Trowbridge H, Baker J, Fuss Z, Bender IB. Effects of thermal vitality tests on human dental pulp. *J Endod* 1988; 14(10):482–5.
28. Ingram TA, Peters DD. Evaluation of the effects of carbon dioxide used as a pulpal test. Part 2. In vivo effect on canine enamel and pulpal tissues. *J Endod* 1983; 9(7):296–303.
29. Peters DD, Lorton L, Mader CL, Augsburger RA, Ingram TA. Evaluation of the effects of carbon dioxide used as a pulpal test. 1. In vitro effect on human enamel. *J Endod* 1983; 9(6):219–27.
30. Jones DM. Effect of the type carrier used on the results of dichlorodifluoromethane application to teeth. *J Endod* 1999; 25(10):692–4.
31. Jones VR, Rivera EM, Walton RE. Comparison of carbon dioxide versus refrigerant spray to determine pulpal responsiveness. *J Endod* 2002; 28(7):531–3.
32. Radhakrishnan S, Munshi AK, Hegde AM. Pulse oximetry: a diagnostic instrument in pulpal vitality testing. *J Clin Pediatr Dent* 2002; 26(2):141–5.
33. Noblett WC, Wilcox LR, Scamman F, Johnson WT, Diaz-Arnold A. Detection of pulpal circulation in vitro by pulse oximetry. *J Endod* 1996; 22(1):1–5.
34. Klein H. Pulp responses to an electric pulp stimulator in the developing permanent anterior dentition. *ASDC J Dent Child* 1978; 45(3):199–202.
35. Andreasen FM, Zhijie Y, Thomsen BL, Andersen PK. Occurrence of pulp canal obliteration after luxation injuries in the permanent dentition. *Endod Dent Traumatol* 1987; 3(3):103–15.
36. Andreasen FM. Histological and bacteriological study of pulps extirpated after luxation injuries. *Endod Dent Traumatol* 1988; 4(4):170–81.
37. Chambers IG. The role and methods of pulp testing in oral diagnosis: a review. *Int Endod J* 1982; 15(1):1–15.
38. Sasano T, Onodera D, Hashimoto K, Iikubo M, Satoh-Kuriwada S, Shoji N, and other. Possible application of transmitted laser light for the assessment of human pulp vitality. Part 2. Increased laser power for enhanced detection of pulpal blood flow. *Dent Traumatol* 2005; 21(1):37–41.
39. Rowe AH, Pitt Ford TR. The assessment of pulpal vitality. *Int Endod J* 1990; 23(2):77–83.
40. Nanci A, Ten Cate AR. Ten Cate's oral histology: development, structure, and function. 6th ed. St. Louis: Mosby; 2003.
41. Edwall L, Kindlova M. The effect of sympathetic nerve stimulation on the rate of disappearance of tracers from various oral tissues. *Acta Odontol Scand* 1971; 29(4):387–400.
42. Tonder KH, Aukland K. Blood flow in the dental pulp in dogs measured by local H<sub>2</sub> gas desaturation technique. *Arch Oral Biol* 1975; 20(1):73–9.
43. Kim S. Regulation of pulpal blood flow. *J Dent Res* 1985; 64 Spec No: 590–6.
44. Smith E, Dickson M, Evans AL, Smith D, Murray CA. An evaluation of the use of tooth temperature to assess human pulp vitality. *Int Endod J* 2004; 37(6):374–80.
45. Sigurdsson A. Pulpal diagnosis. *Endodontic Topics* 2003; 5:12–25.
46. Nissan R, Trope M, Zhang CD, Chance B. Dual wavelength spectrophotometry as a diagnostic test of the pulp chamber contents. *Oral Surg Oral Med Oral Pathol* 1992; 74(4):508–14.
47. Samraj R, Indira R, Srinivasan MR, Kumar A. Recent advances in pulp vitality testing. *Endodontology* 2003; 15:14–7.
48. Severinghaus JW, Kelleher JF. Recent developments in pulse oximetry. *Anesthesiology* 1992; 76(6):1018–38.
49. Schnapp LM, Cohen NH. Pulse oximetry. Uses and abuses. *Chest* 1990; 98(5):1244–50.
50. Schmitt JM, Webber RL, Walker EC. Optical determination of dental pulp vitality. *IEEE Trans Biomed Eng* 1991; 38(4):346–52.
51. Kahan RS, Gulabivala K, Snook M, Setchell DJ. Evaluation of a pulse oximeter and customized probe for pulp vitality testing. *J Endod* 1996; 22(3):105–9.
52. Gopi Krishna V, Kandaswamy D, Gupta T. Assessment of the efficacy of an indigenously developed pulse oximeter dental sensor holder for pulp vitality testing. *Indian J Dent Res* 2006; 17(3):111–3.
53. Gopikrishna V, Tinagupta K, Kandaswamy D. Evaluation of efficacy of a new custom-made pulse oximeter dental probe in comparison with the electrical and thermal tests for assessing pulp vitality. *J Endod* 2007; 33(4):411–4.
54. Gopikrishna V, Tinagupta K, Kandaswamy D. Comparison of electrical, thermal, and pulse oximetry methods for assessing pulp vitality in recently traumatized teeth. *J Endod* 2007; 33(5):531–5.
55. Gazelius B, Olgart L, Edwall B, Edwall L. Non-invasive recording of blood flow in human dental pulp. *Endod Dent Traumatol* 1986; 2(5):219–21.
56. Olgart L, Gazelius B, Lindh-Stromberg U. Laser Doppler flowmetry in assessing vitality in luxated permanent teeth. *Int Endod J* 1988; 21(5):300–6.
57. Ingolfsson AE, Tronstad L, Riva CE. Reliability of laser Doppler flowmetry in testing vitality of human teeth. *Endod Dent Traumatol* 1994; 10(4):185–7.
58. Yanpiset K, Vongsavan N, Sigurdsson A, Trope M. Efficacy of laser Doppler flowmetry for the diagnosis of revascularization of reimplanted immature dog teeth. *Dent Traumatol* 2001; 17(2):63–70.
59. Pettersson H, Öberg PA. Pulp blood flow assessment in human teeth by laser Doppler flowmetry. *Lasers in Orthopedic, Dental, and Veterinary Medicine*. 1991; SPIE vol. 1424:116–9.
60. Wilder-Smith PE. A new method for the non-invasive measurement of pulpal blood flow. *Int Endod J* 1988; 21(5):307–12.
61. Evans D, Reid J, Strang R, Stirrups D. A comparison of laser Doppler flowmetry with other methods of assessing the vitality of traumatized anterior teeth. *Endod Dent Traumatol* 1999; 15(6):284–90.
62. Strobl H, Gojer G, Norer B, Emshoff R. Assessing revascularization of avulsed permanent maxillary incisors by laser Doppler flowmetry. *J Am Dent Assoc* 2003; 134(12):1597–603.
63. Strobl H, Haas M, Norer B, Gerhard S, Emshoff R. Evaluation of pulpal blood flow after tooth splinting of luxated permanent maxillary incisors. *Dent Traumatol* 2004; 20(1):36–41.
64. Zanetta-Barbosa D, Klinge B, Svensson H. Laser Doppler flowmetry of blood perfusion in mucoperiosteal flaps covering membranes in bone augmentation and implant procedures. A pilot study in dogs. *Clin Oral Implants Res* 1993; 4(1):35–8.
65. Dodson TB, Neuenschwander MC, Bays RA. Intraoperative assessment of maxillary perfusion during Le Fort I osteotomy. *J Oral Maxillofac Surg* 1994; 52(8):827–31.
66. Edwall B, Gazelius B, Berg JO, Edwall L, Hellander K, Olgart L. Blood flow changes in the dental pulp of the cat and rat measured simultaneously by laser Doppler flowmetry and local 125I clearance. *Acta Physiol Scand* 1987; 131(1):81–91.
67. Sasano T, Shoji N, Kuriwada S, Sanjo D, Izumi H, Karita K. Absence of parasympathetic vasodilatation in cat dental pulp. *J Dent Res* 1995; 74(10):1665–70.
68. Sasano T, Nakajima I, Shoji N, Kuriwada S, Sanjo D, Ogino H, and other. Possible application of transmitted laser light for the assessment of human pulp vitality. *Endod Dent Traumatol* 1997; 13(2):88–91.
69. Konno Y, Daimaruya T, Iikubo M, Kanzaki R, Takahashi I, Sugawara J, and other. Morphologic and hemodynamic analysis of dental pulp in dogs after molar intrusion with the skeletal anchorage system. *Am J Orthod Dentofacial Orthop* 2007; 132(2):199–207.
70. Musselwhite JM, Klitzman B, Maixner W, Burkes EJ Jr. Laser Doppler flowmetry: a clinical test of pulpal vitality. *Oral Surg Oral Med Oral Pathol Oral Radiol Endod* 1997; 84(4):411–9.
71. Akpinar KE, Er K, Polat S, Polat NT. Effect of gingiva on laser doppler pulpal blood flow measurements. *J Endod* 2004; 30(3):138–40.
72. Ingolfsson AR, Tronstad L, Hersh EV, Riva CE. Efficacy of laser Doppler flowmetry in determining pulp vitality of human teeth. *Endod Dent Traumatol* 1994; 10(2):83–7.
73. Polat S, Er K, Polat NT. Penetration depth of laser Doppler flowmetry beam in teeth. *Oral Surg Oral Med Oral Pathol Oral Radiol Endod* 2005; 100(1):125–9.
74. Polat S, Er K, Polat NT. The lamp effect of laser Doppler flowmetry on teeth. *J Oral Rehabil* 2005; 32(11):844–8.