

Dielectrophoresis for manipulation of micro/nano particles in microfluidic systems

C. Zhang · K. Khoshmanesh · A. Mitchell ·
K. Kalantar-zadeh

Published online: 3 September 2009
© Springer-Verlag 2009

Erratum to: Anal Bioanal Chem
DOI 10.1007/s00216-009-2922-6

Unfortunately, there was a mistake in Equations (12) and (13).
The corrected equations for (12) and (13) are:

$$f_{CM\text{-cylindrical}} = \frac{\varepsilon_p^* - \varepsilon_m^*}{\varepsilon_m^*}, \quad (12)$$

$$f_{CM\text{-spherical}} = \frac{\varepsilon_p^* - \varepsilon_m^*}{\varepsilon_p^* + 2\varepsilon_m^*}, \quad (13)$$

The online version of the original article can be found at <http://dx.doi.org/10.1007/s00216-009-2922-6>.

C. Zhang · A. Mitchell · K. Kalantar-zadeh (✉)
School of Electrical and Computer Engineering, RMIT University,
Melbourne, Australia
e-mail: kourosh.kalantar@rmit.edu.au

K. Khoshmanesh
School of Engineering, Deakin University,
Melbourne, Australia