

Effect of electrical conditions on an impedimetric immunosensor based on a modified conducting polypyrrole

Imen Hafaid^{a,b*}, Syrine Chebil^c, Hafsa Korri-Youssoufi^c, François Bessueille^a, Abdelhamid Errachid^a, Zina Sassi^b, Zulfiqur Ali^d, Adnane Abdelghani^e, Nicole Jaffrezic-Renault^a

^aUniversity of Lyon, Laboratory of Analytical Sciences, UMR-CNRS 5180, Claude Bernard University, Lyon I, 43 Boulevard du 11 Novembre 1918, 69622 Villeurbanne Cedex, France

^bAmpere Laboratory, Bât L. De Vinci - INSA -20, avenue Albert Einstein, F-69621, Villeurbanne Cedex, France

^cEquipe de Chimie Bioorganique et Bioinorganique, ICMMO, UMR-CNRS 8182, Paris-Sud University, Bât 420, 91405 Orsay France

^dUniversity of Teesside Middlesbrough Tees Valley TS1 3BA UK

^eUnité de Recherche de Physique des Semi-conducteurs et Capteurs, IPEST, La Marsa, 2070, Tunis, Tunisia

ABSTRACT

This work has explored the development of impedimetric immunosensors based on a conducting copolypyrrole film to which a histidine tag antibody has been bound using a complex NTA chelator coordinated with copper as metal. The electrical properties of the copolymer film have been characterized by Electrochemical Impedance Spectroscopy (EIS). The experimental study of EIS was performed over a range of dc voltages from 0 to -1.4 V (vs saturated calomel electrode SCE) where the polypyrrole layers transit from their conducting to insulating state and to demonstrate the contribution of the copper complex as redox probe. The electrical behavior of the biomaterial at these various potentials was studied according to an equivalent circuit fitting the experimental parameters. The results indicated that better immunosensor characteristics were obtained at an applied potential of -1400 mV where the reduction of copper occurs. The impedance obtained at this potential shows a linear variation of charge transfer resistance (R_{ct}) with the specific antigen concentration in the range of 0.1 ng mL⁻¹ to 10 ng mL⁻¹ with a detection limit of 0.1 ng mL⁻¹. Moreover, this study demonstrates that these immunosensors exhibit high sensitivity, stability and reproducibility.

Key words: Immunosensors, Conducting polymer, impedance spectroscopy, His-tag antibody

1. Introduction

Immunosensors have generated great interest with expectation of providing fast and highly sensitive immunological response. Immunosensors have widespread applications in clinical diagnostics, food safety and quality control, biological analysis and environmental monitoring [1-4]. In recent past years, research in this field has evolved quickly with the aim of improving the performance of the biosensors (specificity, stability, sensitivity, detection limit). Recent studies have focused on the use of conducting polymers as tools for amplification of the response signal from antibody - antigen interactions, which leads to the improvement of the performance characteristics of the biosensor [5,6].

Generally, in physics, conducting polymers are of great scientific and technological importance because of their electrical, electronic, magnetic, and optical properties [7-9]. These beneficial material properties mean that conducting polymers have wide utility in various applications such as anti-static and anti-corrosion coating, batteries and super capacitors, and more recently light emitting diodes (LEDs), electrochromic devices and transparent electrode materials [10-13]. Furthermore, the intrinsic characteristics of conducting polymers allow their exploitation for the preparation of nanocomposites with entrapped nano-scaled biomolecules, mainly proteins [14] and single stranded DNA oligomers [15]. Some conducting polymers doped and/or covalently or non covalently modified by bionanomaterials exhibit unique catalytic or affinity properties that can be easily applied in the design of bioanalytical sensors [6,16]. Polypyrrole (PPy) is one of the most extensively studied for the design of bioanalytical sensors. It is attractive because of its good biocompatibility as polymer *in vivo* [17,18], high electrical conductivity, good redox properties and long-term environmental stability [19, 20]. PPy is widely considered as an effective material for immobilization of biomaterials and for transducing /amplification of analytical signal within immunosensing devices [18,21].

The immobilization of bioactive macromolecules in or on electrically conductive polymers has been extensively explored in an effort to provide intimate contact between these two elements [22]. In fact, the immobilization of biologically active molecules on a conducting

polymer can be achieved by three methods [23,24]: (i) entrapping biomolecules in an electropolymerized film formed with a mixture of monomers and biomolecules; however, a possible drawback is the denaturation of the molecules (ii) electrostatic binding of biomolecules directly onto specific groups generated on the substrate surface, which constitutes a convenient method for immobilizing proteins; however, the major drawback of this method is the strength of interaction, which is dependent on the surrounding solution, and so changes in ionic strength and/or in pH can cause the release of the attached moiety; and (iii) covalent attachment of the biomolecules to an electropolymerized polymer film carrying pendant-reactive groups such as amine or carboxylic group ($-\text{NH}_2$, $-\text{COOH}$). This latter approach is the most promising, as growth of the polymer film and immobilization of the biomolecules can be performed under different experimental conditions, allowing for the denaturation problem to be overcome. Furthermore, there is minimum physisorption, therefore reducing the likelihood of leaching of the captured biomolecules [25]. Most importantly, it is critical that covalent immobilization methods maintain the activity of the molecules, increase stability, and ensure accessibility of the analyte to perform biological events such as hybridization of complementary oligonucleotides, antigen-antibody binding, or enzyme-catalyzed reactions [26].

In this paper, we report on the development of an electrochemical immunosensor based on electropolymerized functionalized co-polypyrrole film on which a fragment of an antibody marked by a biomolecule (Hist-tag Single Chain Antibody) is immobilized by the intermediary of a NTA ($\text{N}\alpha,\text{N}\alpha$ -Bis(Carboxymethyl)-L-lysine Hydrate) graft coordinated with a metal ion [27-28], for the antigen recognition which is an antigenic peptide coupled with BSA (Bovine Serum Albumin). This approach that has never been used until now for antibody immobilization, is interesting because it raises the possibility of reversibly binding the antibody by using a complexing agent of copper without any denaturation of the antibody.

Electrochemical impedance spectroscopy (EIS) has been chosen as the main characterization technique to assess the immunosensor performance, since it is known as an effective tool for the qualitative and quantitative characterization of electrochemical processes occurring in the conducting polymer films [29,30], and is able to probe the interfacial properties of the modified electrode. EIS has also often been used for understanding chemical

transformations and processes associated with the conductive supports as well as probing antibody-antigen interaction in order to study the immunosensor responses [31-35].

In a first step the EIS technique was applied to study the electrical properties of the copolymer at various dc potentials ranging from 0 to -1.4V/SCE. The contribution of the potential range and then of the electrical and redox properties of the layer was studied to determine the contribution of the conductivity of polypyrrole and of the NTA/copper complex as redox probe in the analytical features of this immunosensor. Next, EIS was applied to monitor the formation of the immunosensor biolayer and of the recognition of a specific antigen by the antibody bonded to the modified copolymer film. Finally, the immunosensor performance was evaluated over a range of potentials according to its sensitivity to the antibody-antigen recognition event.

2. Materials and methods

2.1. Reagents

Pyrrole monomer (Py) (Sigma) was distilled prior to use. N-hydroxyphthalimide pyrrole (Py-NHP) was synthesized as described previously [28]. Lithium Perchlorate (LiClO₄), Acetonitrile and N α ,N α -Bis(Carboxymethyl)-L-lysine Hydrate (NTA) were obtained from Fluka. Anhydrous 1,2-Propylene carbonate (C₄H₆O₃), Acetic Acid (C₂H₄O₂), Copper (II) Acetate and Sodium Acetate (CH₃COONa.3H₂O) were purchased from Aldrich. The 6-His tag reduced antibody (Fab fragment hyst-ScAc) and the antigen (peptide conjugated to BSA) were obtained from Wyeth Company.

Copper acetate solution was obtained by diluting 0.2 M copper (II) acetate in acetate buffer solution (pH 4.6). The buffer solution used for all experiments was Phosphate Buffer Saline (PBS) containing 137 mM NaCl, 2.7 mM KCl, 0.01 M KH₂PO₄ and 0.01 M K₂HPO₄, pH 7.4. All solutions were made in ultrapure water (resistance 18.2 M Ω cm⁻¹) produced by a Millipore Milli-Q system.

2.2. Instrumentation

2.2.1. Electrochemical deposition and analysis

Electrochemical deposition of copolymer and EIS measurements were performed using a Voltalab 80 impedance analyzer. A conventional three electrode cell was used, including a saturated calomel electrode (SCE) as reference electrode, a platinum plate as counter electrode (0.54 cm^2) and a modified gold working electrode (0.19 cm^2). The impedance spectroscopy measurements were carried out in the frequency range from 0.05 Hz to 100 kHz at different dc potentials 0, -200, -400, -600, -800 and -1400 mV versus SCE electrode, using a modulation voltage of 10 mV. Data simulation was made with the commercial software Zview (Scribner and associates, Charlottesville, VA). All electrochemical measurements were carried out at room temperature in PBS solution at pH 7.4.

2.3. *Electropolymerization of the PPy/PPy-NHP film on gold electrodes*

Gold electrodes with a defined surface of 0.19 cm^2 were provided by the Laboratoire d'Analyse et d'Architecture des Systèmes (LAAS), CNRS Toulouse. The electrodes were made by evaporating a gold layer of 300 nm thickness on a silicon substrate capped with a 30 nm titanium adhesion layer. Before electropolymerization, the gold electrode was cleaned in an ultrasonic bath for 10 min in acetone and dried under a dry nitrogen flow. Subsequently, the electrode was dipped for 2 min at room temperature in an aqua regia mixture ($\text{HCl}/\text{HNO}_3/\text{H}_2\text{O}$) and then in a basic mixture ($\text{H}_2\text{O}_2/\text{NH}_4\text{OH}/\text{H}_2\text{O}$) for 5 min at 80°C , 3:1:16 and 1:1:5 by volume concentration respectively. Between and after these treatments, the gold electrodes were rinsed thoroughly with ultrapure water.

After cleaning, the gold electrodes were immediately placed as working electrode on the three-electrode electrochemical cell. Electropolymerization was performed at 0.9V/SCE for 2 min in a propylene carbonate solution containing 0.04 M pyrrole (Py), 0.06 M 3-N-hydroxyphthalimide pyrrole (Py-NHP) and 0.5 M LiClO_4 to give the copolymer film P(Py-PyNHP). The solution was de-aerated through nitrogen bubbling for 20 min before use. The measured charge consumed during the electropolymerization process reached $2.54 \cdot 10^{-3} \text{ C}$ corresponding to a charge density of 13.3 mC cm^{-2} . The functionalized polypyrrole film thus obtained is in its electrically p-type conducting state, i.e partially oxidized by the ClO_4^- anion.

2.4. *Immobilization of the antibody*

After electropolymerization, the copolypyrrole modified electrode was rinsed respectively with propylene carbonate and PBS. Then, it was immersed in 7 mM NTA solution for 1 h at room temperature to functionalize the copolypyrrole surface with NTA groups via amide links. After thoroughly rinsing with PBS, the electrode was immediately rinsed with sodium acetate/acetic acid buffer solution and immersed in 0.01 M copper (II) acetate solution at pH 4.2 for 15 min for the coordination of Cu^{2+} ions with grafted NTA groups. A single Chain Antibody (scab) fragment with a Histidine tag (0.4 mg/ml) was subsequently immobilized on PPy-NHP/NTA/ Cu^{2+} films for 1h at room temperature, followed by thorough rinsing with PBS. Fully functionalized electrodes were then incubated at different concentrations of the antigen for 30 min at room temperature.

Figure 1 shows the schematic steps for the formation of the bilayer and the antigen-antibody interaction. After the electropolymerization reaction, the electrode surface became covered with polypyrroles bearing activated ester (Py-NHP), which can form covalent bonds with the NTA molecule by reaction with its terminal amine group and the activated ester of the polymer. The new membrane enriched with carboxyl groups stemming from the three terminal carboxyl regions of NTA, was allowed to form a complex with metal ions such as Cu^{2+} . Immediately, these borderline Cu (II) ions were chelated with the histidine ligand attached to the single-chain antibody fragment by completing coordination with the imidazole ring of the histidine tagged antibody [36].

3. Results and discussion

3.1. *Monitoring of the redox process by cyclic voltammetry*

Cyclic voltammetry of the copolymer film P(Py-PyNHP) was carried out to determine the potential range where the polypyrrole layer transits from a conducting to an insulating state. Figure 2.1a shows the oxidation wave peaks at -200 mV where the polymer is strongly conducting, whilst at less than -800 mV the polypyrrole layer becomes insulating.

After the immobilization of NTA and copper interaction the cyclic voltammetry (figure 2.1b) shows at very negative potential a broken reduction wave of copper corresponding to the

reduction of $\text{Cu}^{2+}/\text{Cu}^0$ and followed by two oxidation waves observed at 200 mV and 500 mV corresponding to the oxidation state of Cu/Cu^+ and $\text{Cu}^+/\text{Cu}^{2+}$. The differential pulse voltammetry was investigated to avoid capacitive current and in order to determine the potential peak of the reduction waves of copper. Figure 2.2 shows the DPV waves in reduction state of the copolymer after NTA/copper immobilization. Besides the reduction processes of the polypyrrole layer a reduction wave between -1.4 V to -1.1 V is observed corresponding to the reduction of Cu^{2+} to Cu^0 .

3.2. Copolymer behavior at various dc potentials

The Nyquist impedance plots of the copolymer films at different applied dc potentials are presented in Fig. 3, where $Re(z)$ is the real part and $Im(z)$ is the imaginary part of the complex impedance Z . The obtained plots were analyzed by using several models of equivalent circuits which are appropriate for a conducting polymer film as described in the literature [29,37-39], but the fittings were unsatisfactory (poor χ^2 correlation coefficients). In contrast, we obtained good fittings for frequencies ranging from 0.05 to 100000 Hz ($\chi^2 \approx 10^{-3}$) using the equivalent circuit depicted in Fig.4, where the various parameters are represented as: R_s solution resistance, C_f capacitance of the copolymer film, R_f polymer film resistance, CPE constant phase element, R_i resistance of ionic transfer and W Warburg Impedance. All of the values of the electrical parameters are presented in table 1 (first line). In fact, this circuit allowed distinguishing between the intrinsic electrical characteristics related to the copolymer film in C_f and R_f which is related to the mobility of ionic charge into the polymer backbone during the doping dedoping process and the R_{ct} charge transfer resistance at the polymer|electrolyte interface or charge transfer resistance at the Pt|polymer interface, according to [29].

Je garderais R_{ct} au lieu de R_i sur le schema 4.

The impedance spectra were measured at dc potentials ranging between -1.4 and 0 V, covering the range in which the polymer film transits from insulating to conducting states [40, 41]. Nonetheless, the spectra show considerable changes with potential throughout the entire frequency range due to the transformations occurring in the film during its insulator-conductor transition.

Figures 5 and 6 show the variation with applied dc potential of capacitance (C_f) and resistance (R_f) respectively, of the polymer film as deduced from fitting the high frequency impedance data to the equivalent circuit shown in Fig. 4. At large negative potentials between -800 mV and -1400 mV the purely resistive behavior of the film is observed (see Figures 5, 6) and no ionic transfer occurs, with the capacitance remaining low at about $3.5\mu\text{F cm}^2$. At less negative potential (-200 mV) the polymer film undergoes a transition from the undoped state to the doped state, which is accompanied by the ionic exchange of ions from solution to a polymer layer to assure the doping process. The capacitance of the film in this case increases to around 1.7mF cm^2 . As expected, the film resistance also drops as the film becomes conductive.

The expression of the capacitance of the polypyrrole film as follows:

$$C_f = \frac{\varepsilon_0 \varepsilon A}{d}$$

Where the ε_0 the vacuum dielectric constant and the ε is the double layer dielectric constant, A is the surface area and d the thickness of the film.

The increase of the capacitance of the film will be dominated by the increase in negative charge originating from the phosphate anions, i.e. the increasing of C_f with the potential is related to the doping level of the film, as already observed in [29, 42]. The doping process of the polypyrrole layer also leads to a slight increase of the thickness of the film by the swelling of the polymer layer due to incorporation of the counter ions, however, this will have a negligible effect on the change in the polymer capacitance.

3.3. *Electrical characterization of the biofilm at various dc potentials by impedance spectroscopy*

Figure 7 shows Nyquist plots, for the assembly of the biofilm on gold electrode for various applied potentials. All the electrical parameters values, calculated from fitting the impedance data to the equivalent circuit shown in Fig. 4, are presented in Table 1.

After the immobilization of NTA/ Cu^{2+} complex, the charge transfer resistance R_{ct} is due to the redox process of the copper. Then, the resistance variation indicate that the attachment of

NTA/Cu²⁺ complex facilitates charge transfer through the modified electrode, at potentials -800 mV and -1400 mV. Finally, the subsequent immobilization of scAb on the electrode was indicated by an increase of R_{ct}.

The impedance spectra for the different layers at the potentials from 0 to -0.8 V indicate a diffusion behavior, when Warburg impedance is higher than R_{ct}. This can be explained by the diffusion of ions into the polypyrrole layer to assure the electroneutrality in the dedoping process. Concerning the electrochemical behavior of the layer at -1400 mV, a low Warburg impedance was obtained which is attributed to the decrease in ionic movement from the film-electrolyte interface as the polypyrrole layer is in insulating state and no ionic exchange was obtained at this potential.

At -1400 mV a low charge transfer resistance is obtained for the P(Py-PyNHP)/NTA/Cu²⁺/His-Tag film where R_{ct}= 272.6 Ωcm² compared to the value of R_{ct}= 9943 Ωcm² obtained at 0 V. At -1400 mV the reduction of the NTA/Cu²⁺/ His-Tag-Ab complex results in an increase of Faradaic current. In comparison, it is known that NTA/Co²⁺ complex undergoes the reduction process Co²⁺/Co at low potential range as already demonstrated in the case of polypyrrole functionalized with NTA/Co²⁺ [27] and monolayers of NTA/Co²⁺ deposited on glassy carbon electrode [43] where the reduction potential is observed around -1.2V. Consequently, this redox process leads to a decrease of the charge transfer resistance of the immunosensor membrane at this low potential.

3.4. Antigen detection using EIS

Systematically, the electrode modified with histidine-tagged Antibody fragment was equilibrated with a range of concentrations of the specific antigen (from 0 to 100 ng mL⁻¹). The Bode impedance diagrams at 0 V, -800 mV and -1400 mV and Nyquist plot at -1400 mV obtained after incubation with different concentrations of antigen, are shown in Fig. 8 and Fig. 9. The obtained spectra clearly show that the impedance increases with the increase of the antigen concentration, which indicates that the antigen was linked to the specific sites of the antibody immobilized on P(Py-PyNHP)/NTA/Cu²⁺ film and disturbing the electrical properties of the layer. Furthermore, as previously for the dc potentials from 0 to -800 mV/SCE the impedance

spectra of antigen-antibody (Ag – Ab) interactions indicate a diffusion behavior, where as the impedance spectra obtained at -1400 mV show that the biolayer presents an electroactivity rather high by comparison to its behavior at the other potentials due to the reduction of copper ions in the NTA complex.

The electrical parameters obtained from the fitted data are summarized in table 2. From these results, the R_{ct} parameter appears as the important value in the case of antibody –antigen interaction. Variation of R_f and C_f parameters corresponding to the electrical characteristics of the polypyrrole layer are also observed specially at 0 V where the polypyrrole is in its conducting state. However these variations are low compared to those obtained for R_{ct} corresponding to the charge transfer occurring in the biolayer. The increasing of electronic conduction at -1400 mV caused by reduction of copper ion in the complex exhibits a much higher R_{ct} variation than for the other dc potentials.

In order to study the immunosensor response, at various dc potentials, the calibration curves corresponding to the variation of the charge transfer resistance R_{ct} versus specific antigen concentrations in PBS solution, from the fitted data (Table 2), is presented in Fig.10. The immunosensor calibration curve exhibits a linear relation between R_{ct} and the antigen concentration followed by a saturation process. The dynamic range of detection varies, depending on the applied dc potential during the measurements. The sensitivity, detection limit calculated from the slope of the calibration curve in the linear part and the correlation coefficient of the linear regression of these plots are summarized in Table 3. The sensitivity, detection limit and dynamic range of the immunosensor varies according the dc potential of impedance measurement.

The sensitivity and dynamic range obtained at applied dc potential of 0 and -200 mV, where the polypyrrole layer is still in conducting state were of same order of magnitude: $3 \times 10^5 \Omega \text{cm}^2 / \mu\text{g mL}^{-1}$, in concentration of antigen ranging from 0.1 to 0.5 ng mL^{-1} , the detection limit corresponding to the lower concentration value measured is 40 pg mL^{-1} . In the case of the measurements obtained in dc potential range from -400 mV to -800mV where the polypyrrole layer is in the insulating state, the dynamic range is very low and saturation is obtained at 0.1 ng mL^{-1} . Whereas, for measurement carried out at -1400mV the sensitivity of the immunosensor

reaches a maximal value of $4.33 \cdot 10^5 \text{ } \Omega\text{cm}^2/\mu\text{g mL}^{-1}$ and a dynamic range from 0.1ng mL^{-1} to 10 mg mL^{-1} was obtained; the detection limit corresponding to the lower concentration value measured is 21.4 pg mL^{-1} . Also a good reproducibility was obtained for the immunosensor working at -1400 mV : 9% standard deviation on measurements performed with four different immunosensors.

Conclusion

In this study, we have introduced a new method for the fabrication of an immunosensor based on immobilization of His-tag antibody to an electropolymerized copolypyrrole film bearing a NTA/copper complex. Firstly, we have demonstrated that by using electrochemical impedance spectroscopy combined with theoretical equivalent circuits, it is possible to determine the electrical properties and to model the behavior of the copolymer film according to dc applied potential. We have also shown that EIS can be used to monitor the engineering of the biolayer.

We demonstrate that, depending on the dc potential, the electrical behavior of the biolayer is modified depending on the redox activity process occurring. At a potential range from 0 to -400 mV where the polypyrrole layer is in a conducting state, diffusion behavior is observed corresponding to the exchange of ions from electrolyte to the layer. At a highly negative potential of -1400 mV , the polypyrrole layer is electrically insulating and at this potential the reduction of the NTA/copper complex takes place, leading to a small increase in electronic conductivity and correspondingly low charge transfer resistance of the biolayer. The detection of the specific antigen was monitored by EIS by varying the dc potential to determine the potential where the sensor is more sensitive. Results show that at a dc potential of -1400mV where the redox process of the complex in the biolayer (NTA/Cu²⁺/his-Tag) occurs leads to sensitive modification of the charge transfer resistance and the immunosensor operates over a larger dynamic range compared to those obtained at less negative potentials.

Acknowledgements

This work was supported by European Commission through DVT-IMP project (*FP6-IST- 34256*) and by NATO through CLG grant # *CLG 982802*.

References

- [1] I. S. Park, N. Kim, Development of a chemiluminescent immunosensor for chloramphenicol, *Analytica Chimica Acta*, 578 (2006) 19-24.
- [2] J. Ramón-Azcón, E. Valera, Á. Rodríguez, A. Barranco, B. Alfaro, F. Sanchez-Baeza, M.P. Marco, An impedimetric immunosensor based on interdigitated microelectrodes (ID±mE) for the determination of atrazine residues, *Biosensors and Bioelectronics*, 23 (2008) 1367- 1373.
- [3] I. M. Ciomasu, P. M. Krämer, C.M. Weber, G.Kolb, D. Tiemann, S. Windisch, I.S. Frese, A.A. Kettrup, A new, versatile field immunosensor for environmental pollutants, *Biosensors and Bioelectronics*, 21 (2005) 354- 364.
- [4] B. D. Malhotra, A. Chaubey , S.P. Singh, Prospects of conducting polymers in biosensors, *Analytica Chimica Acta*, 578 (2006) 59-74.
- [5] J. Travas-Sejdic, H. Peng, R.P. Cooney, G.A. Bowmaker, M.B. Cannell, C. Soeller, Amplification of a conducting polymer-based DNA sensor signal by CdS nanoparticles, *Current Applied Physics*, 6 (2006) 562- 566.
- [6] P. N. Bartlett, P. R. Birkin, The application of conducting polymers in biosensors, *Synthetic Metals*, 61 (1993) 15-21.
- [7] M. Omastová, I. Chodák , J. Pionteck, Electrical and mechanical properties of conducting polymer composites, *Synthetic Metals*, 102 (1999) 1251-1252.
- [8] T. Sugano and M. Kinoshita, Magnetic properties of conducting polymers: Sulfur-linked aromatic hydrocarbons with and without acceptor doping, *Synthetic Metals*, 17 (1987) 685-690.
- [9] H. Stubb, E. Punkka , J. Paloheimo, Electronic and optical properties of conducting polymer thin films, *Materials Science Reports* 10 (1993) 85-140.
- [10] R.C. Patil , S. Radhakrishnan, Conducting polymer based hybrid nano-composites for enhanced corrosion protective coatings, *Progress in Organic Coatings*, 57 (2006) 332-336.
- [11] K. Gurunathan, D. P. Amalnerkar D. C. Trivedi, Synthesis and characterization of conducting polymer composite (PAn/TiO₂) for cathode material in rechargeable battery, *Materials Letters* 57 (2003) 1642-1648.
- [12] D. Chinn, M. DeLong, A. Fujii, S. Frolov, K. Yoshino, Z.V. Vardeny, Synthesis and processing of conducting polymer microlasers and microsize light emitting diodes, *Synthetic Metals* 102 (1999) 930-932.
- [13] E. Sahin, E.Sahmetlioglu, I. M. Akhmedov, C.Tanyeli, L.Toppare, Synthesis and characterization of a new soluble conducting polymer and its electrochromic devices, *Organic Electronics* 7 (2006) 351- 362.
- [14] A. Azioune, F. Siroti, J. Tanguy, M. Jouini, M. M. Chehimi, B. Miksa, S. Slomkowski, Interactions and conformational changes of human serum albumin at the surface of electrochemically synthesized thin polypyrrole films, *Electrochimica Acta* 50 (2005) 1661-1667.

- [15] B. Saoudi, Ch. Despas, M. M. Chehimi, N. Jammul, M. Delamar, J. Bessière, A. Walcarius, Study of DNA adsorption on polypyrrole: interest of dielectric monitoring, *Sensors and Actuators B-chemical* 62 (2000) 35-42.
- [16] A. Ramanavicius, A. Ramanaviciene, A. Malinauskas, Electrochemical sensors based on conducting polymer-polypyrrole, *Electrochimica Acta* 51 (2006) 6025-6037.
- [17] A. Ramanaviciene, A. Kausaite, S. Tautkus and A. Ramanavicius, Biocompatibility of polypyrrole particles: an in vivo study in mice, *Journal of Pharmacy and Pharmacology* 59 (2007) 311- 315.
- [18] A. Sargent, O. A. Sadic, Monitoring antibody–antigen reactions at conducting polymer-based immunosensors using impedance spectroscopy, *Electrochimica Acta* 44 (1999) 4667-4675.
- [19] M. Ogasawara, K. Funahashi, T. Demura, T. Hagiwara, K. Iwata, Enhancement of electrical conductivity of polypyrrole by stretching, *Synthetic Metals* 14 (1986) 61-69.
- [20] K.K. Kanazawa, A.F. Diaz, M.T. Krounbi, G.B. Street, Electrical properties of pyrrole and its copolymers, *Synthetic Metals* 4 (1981) 119-130.
- [21] A. Ramanavicius, K. Habermüller, E. Csöregi, V. Laurinavičius, W. Schuhmann, Polypyrrole entrapped quinohemoprotein alcohol dehydrogenase. evidence for direct electron transfer via conducting polymer chains, *Analytical Chemistry* 71 (1999) 3581-3586.
- [22] M. Gerard, A. Chaubey and B.D. Malhotra, Application Of Conducting Polymers To Biosensors, *Biosensors Bioelectronics* 17 (2002) 345-359.
- [23] W. Khan, M. Kapoor and N. Kumar, Covalent attachment of proteins to functionalized polypyrrole-coated metallic surfaces for improved biocompatibility, *Acta Biomaterialia* 3 (2007) 541- 549.
- [24] U. Lange, N. V. Roznyatovskaya, V. M. Mirsky, Conducting polymers in chemical sensors and arrays, *Analytica Chimica Acta* 614 (2008) 1-26.
- [25] O. Ouerghi, A. Senillou, N. Jaffrezic-Renault, C. Martelet, H. Ben Ouada, S. Cosnier, Gold electrode functionalized by electropolymerization of a cyano N-substituted pyrrole: application to an impedimetric immunosensor, *Journal of Electroanalytical Chemistry* 501(2001) 62-69.
- [26] N. K. Guimarda, N. Gomez Ch. E. Schmidt, Conducting polymers in biomedical engineering, *Progress in Polymer Science* 32 (2007) 876–921.
- [27] N. Haddour, S. Cosnier, Ch. Gondran, Electrogeneration of a Poly(pyrrole)-NTA Chelator Film for a Reversible Oriented Immobilization of Histidine-Tagged Proteins, *Journal of American Chemical Society* 127 (2005) 5752- 5753.
- [28] S. Chebil, I. Hafaid, H. Sauriat-Dorizon, N. Jaffrezic-Renault, A. Errachid, V. Auger, Z. Ali, H. Korri-Youssoufi, Electrochemical real-Time detection of model in plasma based on functionalized polypyrroles-NTA conjugate to His-tag scAntibody, *Adv. Mater.*(submitted).
- [29] A. Hallik, A. Alumaa, J. Tamm, V. Sammelselg, M. Vaartnou, A. Janes, E. Lust, Analysis of electrochemical impedance of polypyrrole|sulfate and polypyrrole|perchlorate films, *Synthetic Metals* 156 (2006) 488- 494.

- [30] C. Ehrenbeck, K. Jüttner, S. Ludwig, G. Paasch, The electrochemical impedance of a free-standing polypyrrole membrane, *Electrochimica Acta* 43 (1998) 2781-2789.
- [31] D.D. Macdonald, Reflections on the history of electrochemical impedance spectroscopy, *Electrochimica Acta* 51 (2006) 1376-1388.
- [32] M. Wang, L. Wang, G. Wang, X. Ji, Y. Bai, T. Li, S. Gong, J. Li, Application of impedance spectroscopy for monitoring colloid Au-enhanced antibody immobilization and antibody-antigen reactions, *Biosensors and Bioelectronics* 19 (2004) 575-582.
- [33] R. Pei, Z. Cheng, E. Wang, X. Yang, Amplification of antigen-antibody interactions based on biotin labeled protein-streptavidin network complex using impedance spectroscopy, *Biosensors and Bioelectronics* 16 (2001) 355-361.
- [34] S. Helali, A. Abdelghani, I. Hafaiedh, C. Martelet, M.I. Prodromidis, T. Albanis and N. Jaffrezic-Renault, Functionalization of niobium electrodes for the construction of impedimetric biosensors, *Materials Science and Engineering C* 28 (2008) 826-830.
- [35] I. Hafaiedh, S. Helali, K. Cherif, A. Abdelghani and G. Tournier, Characterization of tin dioxide film for chemical vapors sensor, *Materials Science and Engineering C* 28 (2008) 584-587.
- [36] G. Bayramoglu, A. Ulku Senel, M. Yakup Arica, Effect of spacer-arm and Cu(II) ions on performance of l-histidine immobilized on poly(GMA/MMA) beads as an affinity ligand for separation and purification of IgG, *Separation and Purification Technology* 50 (2006) 229-239.
- [37] A. A. Diagne, M. Fall, M. Guene, M. M. Dieng, F. Deflorian, S. Rossi, P. Bonora, C. Della Volpe, Electrochemical impedance spectroscopy of poly(bithiophene) films in an aqueous LiClO₄ solution, *Comptes rendus Chimie* 10 (2007) 558-563.
- [38] S. A. Sezai, S. Sibel, A. Murat, T. C. Metehan, Electrochemical impedance spectroscopy and morphological analyses of pyrrole, phenylpyrrole and methoxyphenylpyrrole on carbon fiber microelectrodes, *Surface and Coatings Technology* 202 (2008) 3997-4005.
- [39] N.F. Atta, A. Glal, F. Khalifa, Electrodeposited metals at conducting polymer electrodes: I—Effect of particle size and film thickness on electrochemical response, *Applied surface Science* 253 (2007) 4273-4282.
- [40] A.F. Diaz, I.C. Catillo, J.A. Logan, L. Wen-Young, Electrochemistry of conducting polypyrrole films, *Journal of Electroanalytical Chemistry* 129 (1981) 115-132.
- [41] T. Shimidzu, A. Ohtani, T. Iyoda, K. Honda, Charge-controllable polypyrrole/polyelectrolyte composite membranes: Part II. Effect of incorporated anion size on the electrochemical oxidation-reduction process, *Journal of Electroanalytical Chemistry* 224 (1987) 123-135.
- [42] J. Heinze, R. Bilger, Ion movements during redox switching of polypyrrole. Experiment and simulation, *Berichte Der Bunsen-Gesellschaft-Physical Chemistry Chemical Physics* 97 (1993) 502-506.
- [43] R. Blankespoor, B. Limoges, B. Schollhorn, J.-L. Syssa-Magalé, D. Yazidi, Dense monolayers of metal-chelating ligands covalently attached to carbon electrodes electrochemically and their

useful applications in affinity binding of histidine-tagged proteins, *Langmuir*. 21(**2005**) 3362-3375.

Figure captions

Fig.1. Schematic representation of the fabrication process of the multilayer system of the immunosensor on the gold electrode.

Fig. 2.1. Cyclic voltammetry of copolypyrrole film analysed in PBS pH 7.0. after polymerization reaction , with scan rate of 50 mV/s (a) P(Py-PyNHP), (b) P(Py-PyNHP)/NTA-Cu²⁺

Fig. 2.2. DPV waves in reduction state of Py-PyNTA-Cu²⁺

Fig. 3 Nyquist plot obtained for copolymer film electrodeposited on the gold electrode, at different formal dc potentials vs. SCE in 0.01 mM PBS solutions pH 7.4. The dc potentials used are: 0, -200, -400, -600, -800 and -1400 mV. The spectra were obtained between 0.05 Hz – 100 KHz. Amplitude of alternative voltages: 10 mV.

Fig.4 Equivalent electrical circuit model for complex impedance plot.

Fig.5 Dependence of the polymer film resistance R_f , with dc potential, obtained from fit of equivalent circuit parameters to experimental data.

Fig.6 Dependence of the polymer film capacitance with dc potential, obtained from fit of equivalent circuit parameters to experimental data.

Fig. 7 Nyquist plot ($-\text{Im}(z)$ vs. $\text{Re}(z)$) for the faradic impedance measurements at (a) 0 mV, (b)-800mV and (c)-1400 mV) vs. SCE, in 0.01 mM PBS pH 7.4 solution, corresponding to the different layers grafted onto the gold electrode : P(Py-PyNHP), P(Py-PyNHP)/NTA-Cu²⁺ and P(Py-PPyNHP)/NTA-Cu²⁺/Ab. Spectra were obtained between 0.05 Hz – 100 KHz. Amplitude of Alternative voltage :10 mV.

Fig. 8 Bode diagrams ($|Z(f)|$ vs. $\log(f)$) at 0 mV, (b)-800 mV and (c)-1400 mV vs.SCE, in PBS solution pH 7.4, obtained for the faradic impedance measurement on modified gold electrode with histidine tagged reduced Antibody under various concentrations of

antigen. Spectra were obtained between 0.05 Hz – 100 KHz. Amplitude of alternative voltage: 10 mV.

Fig.9 Nyquist plots ($-\text{Im}(z)$ vs. $\text{Re}(z)$) at -1400 mV vs.SCE in PBS solution pH 7.4, obtained for the faradic impedance measurement on modified gold electrode with histidine tagged reduced Antibody under various concentrations of antigen. Spectra were obtained between 0.05 Hz – 100 kHz. Amplitude of alternative voltage: 10 mV.

Fig.10 Calibration curves describing the variation of charge transfer resistance R_{ct} against antigen concentration for various potential (0, -200, -400, -600,-800, -1400 mV).

Fig.11 Calibration curves describing the variation of charge transfer resistance R_{ct} against antigen concentration for potential -1400mV.


Table captions

Table 1 Electrochemical parameters obtained from fitting of the experimental results of P(Py/Py-NHP) layer at different potentials(0, -200, -400, -600,-800, -1400 mV).


Table 2 Electrochemical parameters obtained from fitting of the experimental results of antigen detection at different potentials (0, -200, -400, -600, -800, -1400 mV).

Table 3 Sensitivity, detection limit, correlation coefficient and the dynamic range of immunosensor responses at various dc potentials.


Figure(s)
[Click here to download high resolution image](#)




Figure(s)

[Click here to download high resolution image](#)


Figure(s)
[Click here to download high resolution image](#)





Figure(s)
[Click here to download high resolution image](#)




Figure(s)
[Click here to download high resolution image](#)








Figure(s)
[Click here to download high resolution image](#)




Figure(s)


[Click here to download high resolution image](#)



Figure(s)
[Click here to download high resolution image](#)




Figure(s)
[Click here to download high resolution image](#)




Figure(s)


[Click here to download high resolution image](#)




Figure(s)
[Click here to download high resolution image](#)




Figure(s)
[Click here to download high resolution image](#)



Figure(s)
[Click here to download high resolution image](#)





Potential (mV)	Layers	R_s (Ωcm^2)	R_f (Ωcm^2)	C_f (Fcm^2)	R_i/ct (Ωcm^2)	W (Ωcm^2)	CPE (Fcm^2)-	α_{CPE}	χ^2
0	PPY	38.89	11.003	1.60 E-3	7679	48789	4.10 E-4	0.81	0.0018
	Cu^{+2}	35.25	58.07	1.10 E-3	7135	7550	3.60 E-4	0.82	0.0018
	Ab	35.01	86.56	8.50 E-4	9943	10164	2.29 E-4	0.82	0.0018
-200	PPY	37.38	24	1.73 E-3	5071	7546	6.48 E-4	0.80	0.0012
	Cu^{+2}	29.09	73	4.07 E-3	4877	13456	3.40 E-4	0.81	0.0015
	Ab	26.92	23	1.38 E-3	7793	12250	1.61 E-4	0.86	0.0019
-400	PPY	37.24	78	1.01 E-3	5506	9205	2.04 E-4	0.87	0.0021
	Cu^{+2}	28.9	87	9.10 E-4	4225	10184	1.85 E-4	0.88	0.0021
	Ab	27.2	97	9.40 E-4	6749	9806	1.50 E-4	0.85	0.0022
-600	PPY	36.51	181	1.61 E-4	4150	7581	1.53 E-4	0.87	0.0021
	Cu^{+2}	28.78	168	1.50 E-4	4520	9121	1.51 E-4	0.79	0.0019
	Ab	26.91	160	2.10 E-4	4901	9357	1.00 E-4	0.78	0.0023
-800	PPY	35.31	224	1.26 E-4	3202	5683	3.04 E-5	0.87	0.0013
	Cu^{+2}	28.02	284	2.50 E-4	3102	12040	1.89 E-5	0.91	0.0014
	Ab	25.75	253	2.24 E-4	3402	14100	1.80 E-5	0.92	0.0012
-1400	PPY	27.53	355.8	3.60 E-5	136	64.07	6.77 E-5	0.90	0.0019
	Cu^{+2}	26.51	325.8	2.84 E-5	212.1	64.99	1.57 E-4	0.91	0.0020
	Ab	27.73	473.6	2.51 E-5	272.6	81.11	2.59 E-4	0.89	0.0016

Table 1

Potential (mV)	Ag Concentration	Rs (Ωcm^2)	Rf(Ωcm^2)	Cf(Fcm^2)	Rct(Ωcm^2)	W (Ωcm^2)	CPE (Fcm^2)	α_{CPE}	χ^2
0	0	28.01	12.30	1.57 E-4	6165	1960	5.40 E-4	0.87	0.0022
	100 pg/ml	29.68	12.50	7.81E-5	6229	2047	4.24 E-4	0.85	0.0015
	500 pg/ml	29.33	18.99	1.91 E-4	6313	6017	3.78 E-4	0.80	0.0021
	4.5 ng/ml	30.89	29.51	2.71 E-4	6365	16491	2.50 E-4	0.76	0.0014
	100.ng/ml	29.66	44.10	1.64 E-4	6447	19933	1.59 E-4	0.74	0.0016
-200	0	27.61	53.80	1.06 E-2	5025	19327	6.03 E-4	0.84	0.0022
	100 pg/ml	29.02	48.60	1.98 E-3	5050	12327	5.08 E-4	0.79	0.0019
	500 pg/ml	28.22	58.48	1.07 E-3	5198	13188	2.85 E-4	0.71	0.0015
	4.5 ng/ml	28.72	94.90	1.07 E-3	5317	20296	1.71 E-4	0.70	0.0024
	100.ng/ml	29.01	80.70	1.60 E-3	5367	23943	2.98 E-4	0.70	0.0023
-400	0	26.66	96.31	2.46E-5	4864	9771	4.60 E-4	0.87	0.0017
	100 pg/ml	27.23	102.36	4.27 E-5	4919	8923	2.91 E-4	0.82	0.0020
	500 pg/ml	26.11	83.10	1.16 E-4	4948	11103	1.74 E-4	0.69	0.0023
	4.5 ng/ml	28.19	82.00	5.65 E-5	4986	9709	8.14 E-5	0.77	0.0020
	100.ng/ml	27.53	97.30	3.67 E-5	5036	8943	6.27 E-5	0.80	0.0026
-600	0	27.42	108.72	3.40 E-5	4595	10469	2.10 E-4	0.79	0.0020
	100 pg/ml	26.98	145.90	7.48 E-5	4634	10100	1.00 E-4	0.78	0.0021
	500 pg/ml	26.45	134.50	3.34 E-5	4754	14008	4.84 E-5	0.83	0.0025
	4.5 ng/ml	28.55	125.70	4.74 E-5	4866	19203	3.20 E-5	0.87	0.0019
	100.ng/ml	27.00	123	3.56 E-5	4865	22277	2.27 E-5	0.88	0.0024
-800	0	26.02	150	2.17 E-5	3920	24100	6.04 E-5	0.82	0.0023
	100 pg/ml	26.87	198	2.18 E-5	3957	25588	6.08 E-5	0.83	0.0024
	500 pg/ml	26.82	134	2.19 E-5	3986	28869	5.83 E-5	0.82	0.0024
	4.5 ng/ml	28.18	130	2.34 E-5	4016	45289	4.90 E-5	0.83	0.0023
	100.ng/ml	29.45	189	1.45 E-5	4072	52910	4.93 E-5	0.79	0.0025
-1400	0	28.38	284.1	2.78 E-5	214.4	44.03	1.86 E-4	0.90	0.0017
	100 pg/ml	29.14	217.7	3.48 E-5	279.9	44.02	1.20 E-4	0.89	0.0018
	500 pg/ml	27.72	338.7	3.17 E-5	438.4	78.25	1.49 E-4	0.87	0.0017
	10 ng/ml	27.73	302.5	3.42 E-5	614.4	119.5	9.74E-5	0.86	0.0018
	100.ng/ml	26.58	308.6	3.84 E-5	891.4	195.4	7.71E-5	0.87	0.0017

Table 2

Potential (mV)	Sensitivity ($\Omega\text{cm}^2/\mu\text{g mL}^{-1}$)	Detection limit	Dynamic measurement	Coefficient regression R^2
0	$2.92 \cdot 10^5$	62.7 pg /mL	0 - 500 pg /mL	0.950
-200	$3.5 \cdot 10^5$	40.0 pg /mL	0 - 500 pg /mL	0.996
-400	$4.1 \cdot 10^5$	32.0 pg /mL	0 - 100 pg /mL	0.986
-600	$3.1 \cdot 10^5$	37.7 pg /mL	0 - 500 pg /mL	0.997
-800	$3.7 \cdot 10^5$	40.2 pg /mL	0 - 100 pg /mL	0.997
-1400	$4.3 \cdot 10^5$	21.4 pg /mL	0 - 10 ng /mL	0.990

Table 3

