

RESEARCH

Open Access



Effects of acquisition device, sampling rate, and record length on kinocardiography during position-induced haemodynamic changes

Amin Hossein^{1,3*} , Jérémy Rabineau¹, Damien Gorlier¹, Farhana Pinki¹, Philippe van de Borne², Antoine Nonclercq¹ and Pierre-François Migeotte¹

*Correspondence:

amin.hossein@ulb.ac.be

¹ LPHYS, Université Libre de Bruxelles, Brussels, Belgium

Full list of author information is available at the end of the article

Abstract

Background: Kinocardiography (KCG) is a promising new technique used to monitor cardiac mechanical function remotely. KCG is based on ballistocardiography (BCG) and seismocardiography (SCG), and measures 12 degrees-of-freedom (DOF) of body motion produced by myocardial contraction and blood flow through the cardiac chambers and major vessels.

Results: The integral of kinetic energy (iK) obtained from the linear and rotational SCG/BCG signals was computed over each dimension over the cardiac cycle, and used as a marker of cardiac mechanical function. We tested the hypotheses that KCG metrics can be acquired using different sensors, and at 50 Hz. We also tested the effect of record length on the ensemble average on which the metrics were computed. Twelve healthy males were tested in the supine, head-down tilt, and head-up tilt positions to expand the haemodynamic states on which the validation was performed.

Conclusions: KCG metrics computed on 50 Hz and 1 kHz SCG/BCG signals were very similar. Most of the metrics were highly similar when computed on different sensors, and with less than 5% of error when computed on record length longer than 60 s. These results suggest that KCG may be a robust and non-invasive method to monitor cardiac inotropic activity.

Trial registration Clinicaltrials.gov, NCT03107351. Registered 11 April 2017, <https://clinicaltrials.gov/ct2/show/NCT03107351?term=NCT03107351&draw=2&rank=1>.

Keywords: Cardiac contractility, Cardiac kinetic energy, Cardiac strength, Ballistocardiography, Seismocardiography, Kinocardiography, Wearable monitoring, Head down tilt, Head up tilt

Background

Cardiac contractions occur more than 2 billion times during an average lifespan. Early detection of any cardiac impairment would allow for intervention and the prevention of irreversible damage [1]. Therefore, the cardiac chronotropic and inotropic states of



© The Author(s) 2021. This article is licensed under a Creative Commons Attribution 4.0 International License, which permits use, sharing, adaptation, distribution and reproduction in any medium or format, as long as you give appropriate credit to the original author(s) and the source, provide a link to the Creative Commons licence, and indicate if changes were made. The images or other third party material in this article are included in the article's Creative Commons licence, unless indicated otherwise in a credit line to the material. If material is not included in the article's Creative Commons licence and your intended use is not permitted by statutory regulation or exceeds the permitted use, you will need to obtain permission directly from the copyright holder. To view a copy of this licence, visit <http://creativecommons.org/licenses/by/4.0/>. The Creative Commons Public Domain Dedication waiver (<http://creativecommons.org/publicdomain/zero/1.0/>) applies to the data made available in this article, unless otherwise stated in a credit line to the data.

patients at risk of a cardiovascular complication should be measured regularly. Currently, the main non-invasive modes of cardiac function assessment include electrocardiography (ECG), echocardiography (Echo), and cardiac magnetic resonance (cMR). The gold standard for cardiac contractility assessment is cMR. However, Echo is the most commonly used technique in daily clinical practice. Both cMR and Echo require trained operators, and are time-consuming. This makes them unsuitable for regular, weekly, or monthly assessments for all at-risk patients. Alternative technologies are therefore needed for more regular and autonomous measures of cardiac function.

Recently, there has been renewed interest in ballistocardiography (BCG) and seismocardiography (SCG) for automated assessment of the inotropic state of the heart [2]. These technologies are based on measurements of body movements induced by cardiac contraction, and blood flow in the cardiac chambers and major vessels of the heart [3]. While BCG primarily measures overall body recoil in response to blood ejection from the heart, and pulsatile blood flow in the aorta and its main branches [3], SCG primarily records precordial accelerations or vibrations of the chest that result from the complex series of movements caused by myocardial contraction [4]. The relationship between cardiac contractility and ballistic signals was identified, and extensively studied by Starr in the twentieth century [3]. Several studies have shown that the signals recorded with BCG and SCG are good indicators of myocardial function and dysfunction [5-7]. A review of recent advances in the field SCG was published by Taebi et al. recently [8].

We previously reported on kinocardiography (KCG) as a subject-specific calibrated combination of linear and rotational SCG and BCG techniques. KCG measures a total of 12 degrees-of-freedom (DOF) of body motion. This consists of 6-DOF from three-dimensional (3D) linear and 3D angular motion recorded from sensors attached to the sternum (SCG) and 6-DOF (3D linear and 3D angular) from whole-body motion recorded from the lower back area (BCG). We introduced the time integral of kinetic energy (iK) over the cardiac cycle as a measure of total cardiac mechanical effort [6]. We showed that, in healthy individuals, this metric could detect dobutamine-induced haemodynamic changes with a high degree of accuracy, sensitivity, and specificity ($\geq 94\%$). Furthermore, the metric iK , observed with increasing doses of dobutamine injection, was correlated with stroke volume (SV) and cardiac output (CO) [6]. The same iK metric showed a significant increase during a brief end expiratory breath hold, and an even larger increase during subsequent spontaneous inspiration, possibly due to its sensitivity to changes in the right heart volume [9]. Previous studies have also shown that cardiovascular deconditioning resulting from long-duration head-down tilt bed rest was visible on KCG metrics. In particular, the evolution of linear BCG iK (iK_{Lin}) followed changes in stroke volume [10, 11]. KCG metrics show promising results, but would benefit from further validations. Indeed, to date, there have been no studies on inter-sensor acquisition differences. Furthermore, in our previous study [6], KCG metrics were computed on SCG and BCG signals acquired at 50 Hz, while these signals are acquired at higher frequencies which range from 100 Hz to 1 kHz [2, 12].

The BCG and SCG signals are prone to artefacts from non-cardiac voluntary or involuntary movements of the body due to their sensitivity [2, 3]. The ensemble average (EA) technique minimises artefacts by pooling and temporally aligning the signals from several heart beats to construct a new signal that represents an average cardiac cycle [8]. In

a linear and a stationary system, the longer the record, the better the EA would represent this average heartbeat [13]. However, it is impractical to record signals during a very long period. To our knowledge, there are no studies on the quantification of the effects of record duration on the EA constructed on SCG and BCG signals, and consequently on the KCG metrics.

The acute head-down tilt (HDT) and head-up tilt (HUT) positions are known to increase [14] and decrease [15] the SV, respectively. While the effects of postural changes on BCG and SCG have been reported [16, 17], there are no studies on the effect of acute HDT and HUT on BCG and SCG. In this study, healthy volunteers were tested in a tilt test protocol at baseline, and after HDT and HUT, used to widen the haemodynamic states on which the validation was performed, and to observe the effects of these on the KCG metrics.

This study aimed to consolidate the technical aspects of KCG by: (1) comparing KCG metrics acquired using two different devices; (2) validating that a sampling rate of 50 Hz allows a precise measure of KCG-derived metrics; (3) analysing the effects of the EA window length on these KCG metrics; and (4) documenting the effects of acute changes in body position on the KCG parameters, and comparing them to existing literature.

Results

The cohort HR and KCG parameters in the supine, HDT, and HUT positions, are presented in Table 1.



Sampling rate validation

Bland–Altman plots were generated for each metric computed with Dev1 on the initial 1 kHz acquisition against its equivalent on the down-sampled version at 50 Hz. The difference between the two paired measurements was plotted against the mean of the two measurements. No trends were observed for any of the computed metrics (Fig. 1). The limits of agreement of the compared data were 97.92% for BCG iK_{Lin} , 93.06% for BCG iK_{Rot} , 95.83% for SCG iK_{Lin} , and 97.22% for SCG iK_{Rot} . The mean difference in percentage between the two methods for each metric was 0.015%, 0.01%, 0.1%, and 0.03%, respectively (Fig. 1).

Table 1 Physiological and KCG parameters presented as median [Q1; Q3] during supine (baseline) and for each position (HDT, and HUT)

Measuring device	Parameter	Supine	HDT	HUT	
ECG-Dev1	HR (bpm)	65.3 [56.4; 75.3]	64.7 [57.2; 72.9]	78.5 [71.7; 81.1] [†]	
Dev1	SCG	iK_{Lin} (μs)	66.5 [40.3; 99.9]	77.6 [50.2; 139]	33.6 [24.3; 46.4] [†]
		iK_{Rot} (μs)	331 [266; 389]	270 [189; 533]	236 [124; 355]*
	BCG	iK_{Lin} (μs)	1.3 [0.7; 1.9]	1.8 [1.1; 3.8] [†]	0.7 [0.6; 1.9]
		iK_{Rot} (μs)	2.7 [1.9; 4.3]	7.8 [3.7; 14.7]*	3.3 [1.5; 11.6]
Dev2	SCG	iK_{Lin} (μs)	61 [37.3; 91.2]	70 [44.2; 124]	27 [19.6; 42.4] [†]
		iK_{Rot} (μs)	335.6 [268; 410]	273.5 [202; 551]	257.4 [128; 372]*
	BCG	iK_{Lin} (μs)	1.4 [0.7; 2.0]	1.9 [1.1; 3.7] [†]	0.7 [0.6; 1.8]
		iK_{Rot} (μs)	4.7 [3.6; 7.3]	11.7 [8.0; 16.4]*	5.7 [3.4; 11.9]

Paired comparison for each parameter to compare HDT and HUT to Supine. * $p < 0.005$; [†] $p < 0.0001$



Sensors comparison

The Bland–Altman plots of the KCG metrics computed from the Dev1 versus the Dev2 acquisitions are shown in Fig. 2. No trends were seen for BCG iK_{Lin} , BCG iK_{Rot} , SCG


iK_{Lin} , and $SCG iK_{Rot}$. The limits of agreement of the compared data were 92.36% for $BCG iK_{Lin}$, 96.53% for $BCG iK_{Rot}$, 95.14% for $SCG iK_{Lin}$, and 95.14% for $SCG iK_{Rot}$. The mean difference in percentage between the two devices for each metric was 2%, 50%, 12%, and 3%, respectively (Fig. 2).

Ensemble average length effect

The KCG metrics $BCG iK_{Lin}$, $BCG iK_{Rot}$, $SCG iK_{Lin}$, and $SCG iK_{Rot}$ were computed with Dev2 based on an EA computed on increasing window lengths: the first 20 s, 30 s, 40 s, 50 s, 60 s, 70 s, 80 s, and 90 s of the record. Figure 3 shows the differences expressed as a percentage of the value with a 95% CI of each KCG metric computed on several window lengths, compared to the metric computed on a 90 s window. The $SCG iK_{Lin}$ showed the fastest stabilisation, as early as 40 s of window length, the EA has a mean error inferior to 5% when compared to a 90 s window length. The $SCG iK_{Rot}$ and $BCG iK_{Lin}$ showed the same trends and tend to have less than 5% of error from window lengths of 50 s and a 60 s, respectively. The $BCG iK_{Rot}$ showed a slower stabilisation, when computed with a 70 s window length, the metrics still had a mean error of 10% when compared to a 90 s window length. Of interest, the effect of position (supine, HDT, and HUT) on the stabilisation of the metrics showed the same trends (Fig. 6 in Appendix A).

Position-induced haemodynamic changes

The linear mixed-effects model showed significant differences between positions for all KCG parameters ($p < 0.0001$). Paired t tests were used to compare the HDT and HUT to the supine position. The $SCG iK_{Lin}$ and $SCG iK_{Rot}$ were stable from the supine to the HDT positions, and decreased to values below the baseline when participants were in the HUT position ($p < 0.0001$, and $p < 0.005$, respectively). The $BCG iK_{Lin}$ and $BCG iK_{Rot}$



increased from the supine to HDT positions ($p < 0.0001$ and $p < 0.005$, respectively), and returned to the baseline level in the HUT position (Fig. 4).


Discussion

Main findings

Kinocardiography is emerging as a promising non-invasive technique to monitor cardiac inotropic activity. The main findings of this study are that the KCG metrics, namely BCG iK_{Lin} , BCG iK_{Rot} , SCG iK_{Lin} , and SCG iK_{Rot} , can be computed based on BCG and SCG signals acquired using different sensors at either 1 kHz or 50 Hz indifferently. Moreover, a recording of 60 s is enough to acquire BCG iK_{Lin} , SCG iK_{Lin} , and SCG iK_{Rot} readings with less than 5% margin of error.

Sampling rate effects

It has previously been shown that BCG signals' content is expected to lay in the 0 Hz–15 Hz frequency range [12], while SCG content is expected to be in the 0 Hz–20 Hz [18], 25 Hz [19], or 30 Hz [20] frequency range. Others also argue that Phonocardiogram signals in the range superior to 25 Hz should also be considered as high-frequency seismocardiograms [21]. Both BCG and SCG have been recorded at high frequency, ranging from 200 Hz to 1 kHz [2, 12]. Acquiring these signals at a lower frequency would enable: (1) real-time transmission using low power wireless techniques, especially when multiple axes, such as the 12 used in this study, are acquired simultaneously, and (2) a reduction in the amount of data to be stored, which is a growing challenge that Internet of things (IoT) and medical devices have to tackle [22]. However, this needs to be done while maintaining the integrity and quality of the data. The scope of this work was not to evaluate if SCG and BCG were accurately acquired at 50 Hz. It was rather to evaluate if iK metrics computed through kinocardiography and derived from



SCG and BCG signals acquired at 50 Hz were similar to the same metrics acquired at 1 kHz. In other words, to which extent SCG and BCG signals acquired at 50 Hz contain all the information needed to compute the iK metrics. The Bland–Altman comparison between the 1 kHz and 50 Hz acquisition showed a satisfactory limit of agreement of more than 90% for all four KCG metrics. The plots show no trends and the differences between both measures represent less than 0.1% for all four metrics. This demonstrates that the SCG and BCG iK metrics computed from signals recorded at 50 Hz or 1 kHz can always be considered to have a precision greater than 1%.

Sensor comparison

The same 4 metrics computed on the signals acquired by Dev2 at 50 Hz was comparable to the metrics computed with Dev1 acquired at 1 kHz, and demonstrated a satisfactory limit of agreement above 90% for all 4 metrics. The plots showed no trends, and the differences in the mean value of both measures were 2% for BCG iK_{Lin} , 50% for BCG iK_{Rot} , 12% for SCG iK_{Lin} , and 3% for SCG iK_{Rot} . While the limits of agreement showed that the difference of the metrics computed with Dev1 and Dev2 were constant, there was a large difference compared to the mean value of BCG iK_{Rot} . This was likely due to the poor 100 mdps resolution of the gyroscopes used in Dev1, which was 25 times larger compared to the 4 mdps resolution of the sensors used in Dev2. This hypothesis is supported by the high-resolution gyroscopes of approximately 9 mdps used by Tadi et al. when performing gyrocardiography [23].

Ensemble average window length effect

BCG and SCG are prone to artefacts due to non-voluntary and voluntary movements, as well as respiratory variations. Generally, an average beat is constructed using EA to mitigate these undesired noisy activities [13, 20]. However, the length of the window or the number of beats on which the EA is computed is arbitrary. Studies have applied variable average recording lengths on which the EA is computed, and these have ranged from 5 to 15 s [20], 5 to 10 heartbeats [17], and 90 s [6] to 140 s [24]. For a modified weighing scale measuring BCG, Inan et al. computed the average beat on 30 heartbeats, which is close to 30 s [25]. The EA can be computed on all the beats indistinctly or depending on the respiratory phase [24, 26] to take into account the BCG-SCG respiration variability. In this study, we showed how KCG parameters derived from BCG and SCG are modified by the length on which the EA is computed on all the heartbeats indistinctly. Our hypothesis was that after 90 s, each metric reaches a reasonably stable value, and that a longer recording would be impractical in the clinical setting. We have shown that BCG iK_{Lin} , BCG iK_{Rot} , SCG iK_{Lin} , and SCG iK_{Rot} reach the threshold of below 10% of discrepancy after 50 s, 70 s, 30 s, and 30 s, respectively. Therefore, when using KCG metrics, depending on the amplitude of the event to be detected, if metrics are computed on different EA window lengths, attention should be given to the varying results obtained solely due to the EA window length differences. For a given metric, the IC of the mean is relatively stable, and has little dependence on the EA window length. Even with BCG iK_{Rot} , for which we observed a 58% mean difference between the 30 s and 90 s window lengths, the IC was consistently lower than 15% for all time frames. Hence, depending

on the precision needed on the metric, short window lengths with a rather acceptable error margin could be used to compare metrics computed on the same window length.

In a previous study [6], when a small dose of dobutamine ($5 \mu\text{g kg}^{-1} \text{min}^{-1}$) was administered to healthy participants, there was an increase in all the metrics; BCG iK_{Lin} (50%), BCG iK_{Rot} (150%), SCG iK_{Lin} (100%), and SCG iK_{Rot} (188%). These results help better understand the potential contribution of the EA window length in light of changes induced by haemodynamic change. Indeed, in the occurrence of dopamine-induced increases on each metric, a 30 s window length-based EA would have enabled a true positive detection of the influence of 5 gamma of dobutamine on all the KCG metrics.

During the change of position from supine to head-up tilt, HR significantly increases for all subjects. However, as one can see on Fig. 6, the dynamic of stabilisation of each metric does not change in function of the position. This result might surprise as a higher heart rate should allow more heartbeats to be included in the EA and therefore provoke a faster stabilisation of the metrics computed with the EA. However, as we speculate that EA stabilisation rate is mainly linked to respiration rate, a higher heart rate without change in respiration rate should indeed not change the stabilisation dynamic of the metrics extracted from the EA.

HDT and HUT effects

Long term HDT-6 bed rest is used as an Earth-based analogue to the effect of micro-gravity exposure on the cardiovascular system [27]. While long-term—above 2 weeks—HDT-6 is known to lead to orthostatic intolerance characterised by a decrease in SV, the first 30–60 min of HDT are characterised by a significant increase in SV and central venous pressure (CVP) [28, 29]. The head-up ($+80^\circ$) tilt is characterised by an increase in peripheral resistance, a decrease in SV, and an increase in HR [28]. The BCG iK_{Lin} and iK_{Rot} increased significantly from the supine to the HDT positions by 52% and 51%, respectively. These results are consistent with the findings of Butler et al. on SV variation due to HDT [28] and the relationship between BCG iK_{Lin} and SV [6]. However, BCG iK_{Lin} and iK_{Rot} did not change significantly from the supine to HUT positions, while SV is expected to decrease. In contrast, SCG iK_{Lin} and iK_{Rot} decreased significantly from the supine to the HUT position, but did not change from the supine to the HDT position. While these results are of interest, in this study, no conclusions on the KCG ability to follow cardiovascular changes during postural changes are drawn. Indeed, the aim of the postural changes was simply to widen the range of haemodynamic activity on which the validation was performed. For a comprehensive understanding of the effects of HDT and HUT on KCG parameters, a dedicated study protocol should be performed with the presence of alternative haemodynamic measures.

Strengths and limitations

Metrics computed with KCG are based on integration of ensemble-averaged kinetic energy signals. These metrics are automatically computed and do not necessitate the intervention of an operator. There is no need to manually select the signal window on which the computations are performed. Thus, they may be particularly suitable for real-time analysis and telemonitoring applications. Other research teams are also pursuing similar research aims. Inan et al. presented a method based on fusion of

three normalised linear SCG axes and use of a graph similarity score (GSS), and found significant differences in GSS between compensated and decompensated heart failure patients [5]. Mechanical parameters extracted from KCG and GSS-derived models are more robust against motion artefacts and other disturbances, as they do not require accurate event-based detection of peaks upon which most other BCG/SCG-based methods rely. Moreover, a known issue to EA synchronised with the R-wave is its tendency to blur events that varies within a record, such as cardiac mechanical events present in the SCG or BCG signal (e.g., aortic opening, peak systolic ejection, etc.) [30]. As the method presented here is based on an integrated energy on the whole cardiac cycle, it is more robust to this kind of distortions in the waveform too.

The study presents potential limitations that require consideration. Dev1 and Dev2 sensors for BCG and SCG recordings were placed on top of each other to obtain simultaneous signals. Therefore, the sensors were not acquiring signals from the exact same position, which may have induced a difference in the accelerations and angular rate acquired between Dev1 and Dev2. Furthermore, calibration of the sensors of each device according to the manufacturer's datasheet could have introduced a small error. We compare KCG parameters computed on signals acquired at 50 Hz and 1 kHz to highlight that the acquisition and storage of raw signals can be optimised by sampling at lower frequencies. Whereas this aim is reached here, a study searching for the optimal sampling rates for KCG parameters could be implemented in the future.

Changes in body position lead to both acute and long-term adaptations of the cardiovascular state [14]. For instance, an increase of stroke volume is reported after several minutes' exposure to head-down tilt bed rest, because of increased venous return; however, a decrease of stroke volume is observed after several days, due to depletion of plasma volume [10]. To acquire the steadiest cardiovascular state achievable, the starting point of the records (after 10 min of rest in each position) has been chosen in such a way that acute adaptations caused by changes in body position are already established [31], while the duration of these records (up to 90 s) is significantly smaller than the time constant of the long-term adaptations (e.g., several days for adaptation of blood volume).

The maximal length of the EA window was arbitrarily chosen as 90 s. A longer window could have been used, but we opted for a recording session more suitable in a clinical setting. Notably, this was longer than the 30 s used in other studies [20, 25]. The EA was computed on all heart beats without taking the respiration phase into account. This procedure is expected to cancel the effect of respiration as it is known to have a large influence on BCG and SCG [24]. If the averaged beat was computed only on beats within a single respiratory phase (i.e., inspiration or expiration), we would expect a faster stabilisation. Also, the sub-sections of the record compared to the maximal length were always the first seconds of the record. This makes the shorter EA window length metrics less stable as they come after a transition between different respiration cycle lengths. However, these changes were determined to be small compared to the changes observed between metrics computed on different EA window lengths. Additionally, participants were subjected to a fixed respiration cycle of 10 s. We expect that a faster respiration cycle would also lead to a faster stabilisation of the metrics as the EA window would include more breathing cycles.

Given that the order of positioning from supine to HDT and to HUT was not randomised, the effect of the HDT preceding the HUT position could have impacted the measures. However, the potential effect of positioning order may have been minimised as participants were placed in the supine position for 10 min before HDT or HUT. Furthermore, the physiological effects of the changes in posture were a supplementary result of the study. Participant numbers were small, and they were all male. However, assessing 12 or fewer individuals for a validation study is common [29, 32]. The absence of female participants should not have [33], or may only slightly have affected [31] the conclusions drawn from the postural changes. It has been shown that the BCG and SCG waveforms are affected by gender [34], and these differences are hypothesised to be mainly due to the inherent cardiac physiology distinction that can be found between men and women [35]. Therefore, the relatively wide haemodynamic range induced by head tilts in this study allows to partially mitigate the absence of females. However, a validation performed on a larger group including females would be interesting as a next step to confirm our findings. Also, these validations were performed on a relatively young participants group, as it is common to see on validation studies [29, 33]. Despite the points highlighted here, we believe that these potential limitations do not preclude the conclusions brought by this research.

Conclusions

In this study, time integrals of kinetic energy (iK), based on the ensemble average of Ballistocardiography and Seismocardiography signals, were computed on a relatively wide range of haemodynamic statuses induced by the acute head-up and head-down tilt positions. The iK were computed simultaneously using two different sensors, at high frequency (1 kHz) and lower frequency (50 Hz), and at an increasing duration of recording time. These metrics showed to be robust in light of sensor differences, sufficiently accurate when BCG and SCG were acquired at 50 Hz, and accurate with a precision of $\pm 5\%$ for most of the tested metrics when computed on a 60 s record length. These results support the use of kinocardiography to monitor cardiac inotropic activity non-invasively and robustly complement information obtained using standard techniques in clinical settings.

Methods

Protocols and participants

Twelve healthy male volunteers who had no history of cardiac disease were non-smokers, and had BMIs between 20 and 25 kg m⁻², were recruited. Participants had a mean age of 22.1 years (± 3.1), and a mean BMI of 23.2 kg m⁻² (± 4.2). None of the participants took drugs or any medication. The study protocol complied with the Declaration of Helsinki, was approved by the local Ethics Committee (Hôpital Erasme—CCB: B406201630013), and was registered on ClinicalTrials.org (April 11, 2017), identifier *NCT03107351*. The prototype of the KCG device used in this clinical trial was authorised by the Belgian Federal Agency for Medicine and Health Products (FAMHP). Written informed consent was obtained from each participant prior to the experimental testing procedure.

Sensors were placed on participants to connect to the ECG, BCG, and SCG recording systems. They had to lie in a supine position on an inversion table for 10 min for stabilisation. The recording was then started, and continuous data were acquired for the next 7 min. Participants were then put in HDT of 6 degrees for 10 min of stabilisation

followed by 7 min of recording. In the final position in HUT of 80 degrees, the participants underwent another 10 min of stabilisation, followed by 7 min of recording.

Participants were asked to remain as still as possible, and to not talk or fall asleep for the duration of the recording.


The influence of breathing on the BCG signal is well known [3, 24, 36]. Hence, we accounted for the influence of respiration using an Imposed and Controlled Breathing (ICB) protocol. The ICB consisted of breathing for a fixed duration during the inspiration and expiration phases matching the normal inspiratory over total breath ratio (35%). An audio system with inspiratory and expiratory sounds was used to guide the participants. Participants were briefed with breathing instructions prior to performing 10 repetitions of 4-, 6-, 8-, and 10-s breathing cycles in succession. Post-hoc analysis showed that they maintained these cycles accurately.

Data acquisition and devices

Two devices (Dev1 and Dev2) were used to simultaneously record the ECG, SCG, and BCG signals of the participants. Dev1 is the Cardiovector (Medical Computer Systems Ltd., Zelenograd, Russian Federation), a system used to monitor autonomic function on the International Space Station [37]. It allows the synchronous performance of: (a) electrocardiography (ECG), (b) impedance cardiography (ICG) in tetrapolar configuration, (c) plethysmography (PTG) using a nasal thermistor to evaluate breathing, (d) 6-DOF SCG (3-axis linear accelerations and 3-axis angular velocities, LIS344ALH and LPY403AL, STMicroelectronics), and (e) 6-DOF BCG (3-axis linear accelerations and 3-axis angular velocities, LIS344ALH and LPY403AL, STMicroelectronics). The acceleration was acquired at 0.5 mg and an RMS noise of $500 \mu\text{g}\cdot\text{Hz}^{-1/2}$, while the angular rate was acquired at a 100 mdps resolution and $10 \text{ mdps}\cdot\text{Hz}^{-1/2}$ RMS noise. An 80 Hz first-order analogue low-pass filter was applied to both linear accelerations and angular rates. All signals were sampled at 1 kHz and stored on a computer. The Dev1 had separate small plastic covers containing the accelerometer and gyroscope sensors that could be individually attached to the body. More details on the system can be found in [37].

Dev2 was the kinocardiograph (HeartKinetics SRL, Gosselies, Belgium), a portable device with two modules. Each module contained a 3-axis accelerometer and 3-axis gyroscope sensor (LSM6DSL, STMicroelectronics), and was attached to the body with standard sticky gel ECG electrodes (3 M). The acceleration and angular rate sensitivity range of the sensor were set to $\pm 2 \text{ g}$ and $\pm 250 \text{ dps}$, respectively, with a resolution of 0.061 mg/LSB and 4.375 mdps/LSB and an RMS noise of $80 \mu\text{g}\cdot\text{Hz}^{-1/2}$ and $4 \text{ mdps}\cdot\text{Hz}^{-1/2}$ with an output bandwidth of 416 Hz. The Dev2 was controlled with a smartphone or a tablet connected via Bluetooth. It collects a two-lead ECG at 200 Hz (ADS1292R, AD Instruments) together with 3-DOF linear accelerations and 3-DOF rotational velocities from each detector. A micro-controller (STM32F411, STMicroelectronics) was used to acquire synchronously data from each chip and operate the Bluetooth low energy (SPTLE-RF, STMicroelectronics). A total of 12-DOF linear acceleration and angular velocity signals were recorded at 50 Hz, and a 25 Hz first-order analogue low-pass filter was applied.

Dev2 consists of two parts which corresponds to the SCG and BCG sensors. For both BCG and SCG, the Dev1 sensors were adhesively attached on top of the Dev2 sensors, which were then positioned on the body with their ECG electrodes. The BCG sensors



were placed in the lumbar lordosis curvature, between the second and the third lumbar vertebrae and close to the participant’s centre of mass. The SCG sensors were placed on the manubrium of the sternum, below the clavicle, and in the superior mediastinum region where the great vessels emerge from the heart (Fig. 5).

Data acquisition from both the Dev1 and Dev2 devices was initiated at the same time, and continuous data were acquired for the duration of the recording.

Data analyses

All data analyses and statistics were performed on a custom-built KCG analysis toolbox written in Matlab 2018b (Mathworks, NL).

Signals acquired using Dev1 and Dev 2 were synchronised based on the ECG signal. The delay between the recordings of the two devices was determined by cross-correlating the ECG signals from both devices to find the point of correspondence. Signal synchronisation was validated visually for all records. The EA method, which is described in the next section, is based on the R waves detected on the ECG. However, there may have been slight differences in the ECG readings of the two devices and in the synchronisation of the ECG of both Dev1 and Dev2. Therefore, the ECG of Dev1 was used for the rest of the analyses for both devices.

The SCG and BCG signals of Dev2 were up-sampled to 1 kHz. Detection of the P, Q, R, S, and T waves was conducted using a modified Pan and Tompkins algorithm [38], and a pattern matching algorithm on the ECG channel. The detections were validated visually. The heart rate (HR) was calculated based on the RR intervals. Afterwards, the initial record was divided into 4 sub-records, one for each of the 4 phases of the ICB protocol (4-, 6-, 8-, and 10-s breathing cycles), and each sub-record was processed independently. In this study, only data from the 10-s breathing cycle protocol are presented. This was done to obtain the longest record possible within a constant breathing cycle.

Ensemble average method

The time reference for the EA is a fixed interval that took into account the cardiac activity preceding atrial depolarisation occurring before the R peak. Indeed, the beginning of

the n th interval started before the P wave, at time $k_{n,start} = R_n - \Delta$, where R_n is the time of the n -th R peak and Δ is 200 ms. The end of the cardiac cycle of interest was assumed to be $k_{n,end} = R_n + \max[RR_i]$ where RR_i represent all the RR intervals of the current record. The mean was taken on all the beats to obtain an average ECG signal. Based on this, an EA was calculated for each channel of the BCG and SCG recordings.

Kinocardiography data analysis

Participant height and weight were used to assess their inertial parameters [39]. Linear accelerations (\vec{a}) first undergo single-time integration to provide velocity (\vec{v}). From this, the linear and rotational kinetic energy transmitted by cardiac contraction to the body was computed using the following equations on the ensemble-averaged signals:

$$K_{Lin} = \frac{1}{2}m(v_x^2 + v_y^2 + v_z^2) \tag{1}$$

$$K_{Rot} = \frac{1}{2}(I_{xx}\omega_x^2 + I_{yy}\omega_y^2 + I_{zz}\omega_z^2), \tag{2}$$

where m is the body mass of the subject, K_{Lin} is the linear kinetic energy, v_x, v_y, v_z are components of the velocity vector \vec{v} , K_{Rot} is the rotational kinetic energy, $I_{xx}, I_{yy},$ and I_{zz} are the orthogonal components of the moment of inertia, and $\omega_x, \omega_y,$ and ω_z are components of the measured angular velocity $\vec{\omega}$ coming from the gyroscope.

These computations gave the following ensemble-averaged signals: BCG K_{Lin} , BCG K_{Rot} , SCG K_{Lin} , and SCG K_{Rot} . The time integral (iK) of K over the ensemble-averaged cardiac cycle (CC) was computed as follows:

$$SCG\ iK_{Lin} = \int_{CC} SCG\ K_{Lin}(t)dt \tag{3}$$

$$SCG\ iK_{Rot} = \int_{CC} SCG\ K_{Rot}(t) dt \tag{4}$$

$$BCG\ iK_{Lin} = \int_{CC} BCG\ K_{Lin}(t)dt \tag{5}$$

$$BCG\ iK_{Rot} = \int_{CC} BCG\ K_{Rot}(t)dt. \tag{6}$$

Where CC, which denotes the entire cardiac cycle, defined as starting with the P wave of the cycle i and ending with the P wave of the cycle $i + 1$, was delimited based on the ensemble-averaged ECG. The integral was indeed performed on this specific CC interval and not the whole ensemble-averaged signal. This leads to the scalar metrics BCG iK_{Lin} , BCG iK_{Rot} , SCG iK_{Lin} , and SCG iK_{Rot} .

Detailed information on the signal processing of multidimensional BCG and SCG records can be found in our previous study [6].

Statistical analyses

Influence of sample rate and sensor comparison

Based on the 1 kHz acquisitions of Dev1, the sample rates BCG and SCG data were decreased by a factor of 20 by keeping the first sample and then every 20th sample after the first. Signals at 50 Hz were therefore obtained. Bland-Altman plots [40] were generated for each KCG metric calculated on the initial 1 kHz acquisition versus the down-sampled version at 50 Hz acquisition. To compare KCG metrics computed on basis of SCG and BCG signals acquired by different sensors, Bland-Altman plots were also generated for each metric to compare Dev1 and Dev2 acquisitions. The measures were considered similar when Bland-Altman plots displayed (1) an absence of positive or negative trend and (2) the presence of more than 90% of the points between ± 1.96 standard deviation (STD).

Ensemble average window length effect

Based on the 100 s signal from the 10 repetitions of 10 s breathing cycle, each record was categorised into 8 sections which corresponded to the first 20 s, 30 s, 40 s, 50 s, 60 s, 70 s, 80 s, and 90 s of the signal, respectively, to account for the influence of duration on the EA metrics. Data were grouped according to the different positions (supine, HDT, and HUT). To quantify the difference in the EA metrics caused by the window lengths, a plot of the differences expressed as a percentage of the value [(Method A – Method B)/mean %] with 95% Confidence Interval (CI), as described by Giavarina [40], was generated.

Influence of position

The KCG parameters were entered into a linear mixed-effects model [41] together with position and time as fixed effects to assess their interactions. The random effects were the intercepts for the subjects. Indeed, position and time are considered as fixed effects, because their impact on haemodynamics is expected to be seen on the variables of interest by a change of the intercept. This fixed effect on the intercept is subject to some randomness caused by inter-individual differences, hence the choice to consider subjects as a random effect. Visual inspection of the residual plots was performed to detect any deviation from homoscedasticity or normality. p values were obtained using likelihood ratio tests of the full model with the effect in question, against the model without this effect. Statistical significance was set at $\alpha = 0.05$ (two-tailed hypothesis tests). KCG parameters were compared between positions by a paired t test for data with a normal distribution, or a Wilcoxon signed-rank test for skewed data. A Lilliefors test was used to test whether the difference between the compared sample populations was normally distributed. A Bonferroni correction was applied to account for multiple comparisons. Therefore, as 4 metrics were computed, a p value less than 0.0125 was considered to compute 95% confidence intervals.

Acknowledgements

The authors would like to acknowledge the contribution of the volunteers. We would also like to thank Dr. Martin Choumont for comments that greatly improved the manuscript.

Authors' contributions

PFM and AH conceived the idea and designed the study. AH had full access to all of the data in the study and takes responsibility for the integrity of the data and the accuracy of the data analysis. PFM conceived the KCG system which AH developed further. PFM, AH, JR, and DG developed the signal processing algorithms for the ECG and KCG signals. AH performed the statistical analysis. FP enrolled the participants and took outcome measurements. AH drafted the

manuscript. PFM, AN, JR, DG, and PVDB critically reviewed the manuscript. All the authors proofread and corrected the manuscript.

Funding

This study was supported by a grant from the European Space Agency and the Belgian Federal Scientific Policy Office (PRODEX PEA 4000110826). Jeremy Rabineau is supported by the Fonds de la Recherche Scientifique (Mandat Aspirant F.R.S.—FNRS FC 29801). Published with the assistance of the Fondation Universitaire de Belgique.

Availability of data and materials

The datasets used and/or analysed during the current study are available from the corresponding author on reasonable request.

Ethics approval and consent to participate

The study protocol complied with the Declaration of Helsinki, was approved by the local Ethics Committee (Hôpital Erasme—CCB: B406201630013), and was registered on ClinicalTrials.org (April 11, 2017), identifier NCT03107351. The prototype of the KCG device used in this clinical trial was authorised by the Belgian Federal Agency for Medicine and Health Products (FAMHP). Written informed consent was obtained from each participant prior to the experimental testing procedure.

Consent for publication

Not applicable.

Competing interests


A.H, P.F.M., D.G.: HeartKinetics minority shareholder. The other authors declare no conflict of interest.

Author details

¹ LPHYS, Université Libre de Bruxelles, Brussels, Belgium. ² Department of Cardiology, Erasme Hospital, Université Libre de Bruxelles, Brussels, Belgium. ³ BEAMS, Université Libre de Bruxelles, Brussels, Belgium.


Appendix A

The KCG metrics $BCG iK_{Lin}$, $BCG iK_{Rot}$, $SCG iK_{Lin}$, and $SCG iK_{Rot}$ were computed with Dev2 based on an EA computed for increasing windows lengths, using the first 20, 30, ... and 90 s of the record. Figure 6 shows differences expressed as a percentage of the value with 95% CI for each KCG metric computed on several window lengths compared to the metric computed on a 90 s window. The HDT position shows the fastest stabilisation, while HUT seems to delay the stabilisation of the metrics as a function of the window length.



Appendix B

Figure 7 shows linear and rotational kinetic energy for BCG and SCG signals for one subject in HDT position.



Received: 17 September 2020 Accepted: 10 December 2020

Published online: 06 January 2021

References

1. Hunt SA, et al. ACC/AHA guidelines for the evaluation and management of chronic heart failure in the adult: executive summary a report of the American college of cardiology/american heart association task force on practice guidelines (committee to revise the 1995 guidelines for the evaluation and management of heart failure): developed in collaboration with the international society for heart and lung transplantation; endorsed by the heart failure society of America. *Circulation*. 2001;104(24):2996–3007.
2. Inan OT, et al. Ballistocardiography and seismocardiography: a review of recent advances. *IEEE J Biomed Health Inform*. 2015;19(4):1414–27. <https://doi.org/10.1109/JBHI.2014.2361732>.
3. Starr I. The relation of the ballistocardiogram to cardiac function. *Am J Cardiol*. 1958;2(6):737–47. [https://doi.org/10.1016/0002-9149\(58\)90271-6](https://doi.org/10.1016/0002-9149(58)90271-6).
4. Salerno DM, Zanetti J. Seismocardiography for monitoring changes in left ventricular function during ischemia. *Chest*. 1991;100(4):991–3. <https://doi.org/10.1378/chest.100.4.991>.
5. Inan OT, et al. Novel wearable seismocardiography and machine learning algorithms can assess clinical status of heart failure patients. *Circ Heart Fail*. 2018;11(1):e004313. <https://doi.org/10.1161/CIRCHEARTFAILURE.117.004313>.
6. Hossein A, et al. Accurate detection of dobutamine-induced haemodynamic changes by kino-cardiography: a randomised double-blind placebo-controlled validation study. *Sci Rep*. 2019;9(1):1–11. <https://doi.org/10.1038/s41598-019-46823-3>.
7. Aydemir VB, Fan J, Dowling S, Inan OT, Rehg JM, Klein L. Ballistocardiography for ambulatory detection and prediction of heart failure decompensation. *J Card Fail*. 2018;24(8):S116. <https://doi.org/10.1016/j.cardfail.2018.07.425>.
8. Taebi A, Solar BE, Bomar AJ, Sandler RH, Mansy HA. Recent advances in seismocardiography. *Vibration*. 2019;2(1):64–86. <https://doi.org/10.3390/vibration2010005>.

9. Morra S, et al. Modification of the mechanical cardiac performance during end-expiratory voluntary apnea recorded with ballistocardiography and seismocardiography. *Physiol Meas*. 2019;40(10):105005. <https://doi.org/10.1088/1361-6579/ab4a6a>.
10. Rabineau J, et al. Cardiovascular adaptation and countermeasure efficacy assessed by ballistocardiography and seismocardiography during 60 days of simulated microgravity in head-down bed rest. *Sci Rep*. 2020;10:1–31.
11. Caiani EG et al. Effectiveness of high-intensity jump training countermeasure on mitral and aortic flow after 58-days head-down bed-rest assessed by phase-contrast MRI. In: Presented at the international astronautical congress, Bremen, Germany. 2018.
12. Burger HC, Noordergraaf A. Physical basis of ballistocardiography. III. *Am Heart J*. 1956;51(2):179–85.
13. Lejeune L, Caiani EG, Prisk GK, Migeotte PF. Evaluation of ensemble averaging methods in 3D ballistocardiography. In: *Conf. Proc. Annu. Int. Conf. IEEE Eng. Med. Biol. Soc. IEEE Eng. Med. Biol. Soc. Annu. Conf.*, vol. 2014; 2014. pp. 5176–5179. <https://doi.org/10.1109/EMBC.2014.6944791>.
14. Blomqvist CG, Nixon JV, Johnson RL, Mitchell JH. Early cardiovascular adaptation to zero gravity simulated by head-down tilt. *Acta Astronaut*. 1980;7(4–5):543–53. [https://doi.org/10.1016/0094-5765\(80\)90043-0](https://doi.org/10.1016/0094-5765(80)90043-0).
15. Bundgaard-Nielsen M, Sørensen H, Dalsgaard M, Rasmussen P, Secher NH. Relationship between stroke volume, cardiac output and filling of the heart during tilt. *Acta Anaesthesiol Scand*. 2009;53(10):1324–8. <https://doi.org/10.1111/j.1399-6576.2009.02062.x>.
16. Soames RW, Atha J. Three-dimensional ballistocardiographic responses to changes of posture. *Clin Phys Physiol Meas*. 1982;3(3):169–77. <https://doi.org/10.1088/0143-0815/3/3/001>.
17. Di Rienzo M, et al. Wearable seismocardiography: Towards a beat-by-beat assessment of cardiac mechanics in ambulant subjects. *Auton Neurosci*. 2013;178(1):50–9. <https://doi.org/10.1016/j.autneu.2013.04.005>.
18. Jerosch-Herold M, et al. The seismocardiogram as magnetic-field-compatible alternative to the electrocardiogram for cardiac stress monitoring. *Int J Card Imaging*. 1999;15(6):523–31. <https://doi.org/10.1023/A:1006364518204>.
19. Zanetti JM, Tavakolian K (2013) Seismocardiography: past, present and future. In: 2013 35th annual international conference of the IEEE engineering in medicine and biology society (EMBC); 2013. pp. 7004–7. <https://doi.org/10.1109/EMBC.2013.6611170>.
20. Marcus FI, et al. Accelerometer-derived time intervals during various pacing modes in patients with biventricular pacemakers: comparison with normals. *Pacing Clin Electrophysiol PACE*. 2007;30(12):1476–81. <https://doi.org/10.1111/j.1540-8159.2007.00894.x>.
21. Castiglioni P, Faini A, Parati G, Di Rienzo M. Wearable seismocardiography; 2007. pp. 3954–3957. <https://doi.org/10.1109/IEMBS.2007.4353199>.
22. Cai H, Xu B, Jiang L, Vasilakos AV. IoT-based big data storage systems in cloud computing: perspectives and challenges. *IEEE Internet Things J*. 2017;4(1):75–87. <https://doi.org/10.1109/JIOT.2016.2619369>.
23. JafariTadi M, et al. Gyrocardiography: a new non-invasive monitoring method for the assessment of cardiac mechanics and the estimation of hemodynamic variables. *Sci Rep*. 2017;7(1):6823. <https://doi.org/10.1038/s41598-017-07248-y>.
24. Prisk GK, Verhaeghe S, Padeken D, Hamacher H, Paiva M. Three-dimensional ballistocardiography and respiratory motion in sustained microgravity. *Aviat Sp Environ Med*. 2001;72(12):1067–74.
25. Inan OT, Etemadi M, Wiard RM, Giovangrandi L, Kovacs GTA. Robust ballistocardiogram acquisition for home monitoring. *Physiol Meas*. 2009;30(2):169–85. <https://doi.org/10.1088/0967-3334/30/2/005>.
26. Prisk GK, Migeotte P-F. Physiological insights from gravity-free ballistocardiography. In: *Conf. Proc. Annu. Int. Conf. IEEE Eng. Med. Biol. Soc. IEEE Eng. Med. Biol. Soc. Annu. Conf.*, vol. 2013; 2013. pp. 7282–85. <https://doi.org/10.1109/EMBC.2013.6611239>.
27. Fortney SM, Schneider VS, Greenleaf JE. The physiology of bed rest. In: *Comprehensive physiology*, American Cancer Society; 2011. pp. 889–939.
28. Butler GC, Xing HC, Hughson RL. Cardiovascular response to 4 hours of 6 degrees head-down tilt or of 30 degrees head-up tilt bed rest. *Aviat Sp Environ Med*. 1990;61(3):240–6.
29. Gaffney FA, Nixon JV, Karlsson ES, Campbell W, Dowdey ABC, Blomqvist CG. Cardiovascular deconditioning produced by 20 hours of bedrest with head-down tilt (–5°) in middle-aged healthy men. *Am J Cardiol*. 1985;56(10):634–8. [https://doi.org/10.1016/0002-9149\(85\)91025-2](https://doi.org/10.1016/0002-9149(85)91025-2).
30. Hurwitz BE, Shyu L, Reddy SP, Schneiderman N, Nagel JH. Coherent ensemble averaging techniques for impedance cardiography. In: [1990] Proceedings. Third annual IEEE symposium on computer-based medical systems; 1990. pp. 228–35. <https://doi.org/10.1109/CBMSYS.1990.109403>.
31. Frey MA, Tomaselli CM, Hoffler WG. Cardiovascular responses to postural changes: differences with age for women and men. *J Clin Pharmacol*. 1994;34(5):394–402. <https://doi.org/10.1002/j.1552-4604.1994.tb04979.x>.
32. Iwasaki K-I, Zhang R, Zuckerman JH, Pawelczyk JA, Levine BD. Effect of head-down-tilt bed rest and hypovolemia on dynamic regulation of heart rate and blood pressure. *Am J Physiol-Regul Integr Comp Physiol*. 2000;279(6):R2189–99. <https://doi.org/10.1152/ajpregu.2000.279.6.R2189>.
33. Patel K, et al. Effect of postural changes on cardiovascular parameters across gender. *Medicine (Baltimore)*. 2016;95(28):e4149. <https://doi.org/10.1097/MD.0000000000004149>.
34. Cao X, Wang J, Zhang X, Tang J. A ballistocardiogram measurement system for home monitoring: design, performance, and evaluation. *Chin Sci Bull*. 2014;59(23):2909–17. <https://doi.org/10.1007/s11434-014-0462-8>.
35. Harrison WK, Smith E. Sex differences in cardiac function of a group of young adults. *Cardiology*. 1980;66(2):74–84. <https://doi.org/10.1159/000170853>.
36. Tavakolian K, Vaseghi A, Kaminska B. Improvement of ballistocardiogram processing by inclusion of respiration information. *Physiol Meas*. 2008;29(7):771–81. <https://doi.org/10.1088/0967-3334/29/7/006>.
37. Baevsky RM, et al. Autonomic function testing on board the ISS—update on 'Pneumocard'. *Acta Astronaut*. 2007;61(7):672–5. <https://doi.org/10.1016/j.actaastro.2006.11.017>.
38. Pan J, Tompkins WJ. A real-time QRS detection algorithm. *IEEE Trans Biomed Eng*. 1985;32(3):230–6.
39. Matsuo A, Ozawa H, Goda K, Fukunaga T. Moment of inertia of whole body using an oscillating table in adolescent boys. *J Biomech*. 1995;28(2):219–23. [https://doi.org/10.1016/0021-9290\(94\)00055-9](https://doi.org/10.1016/0021-9290(94)00055-9).

40. Giavarina D. Understanding bland altman analysis. *Biochem Med.* 2015;25(2):141–51. <https://doi.org/10.11613/BM.2015.015>.
41. Wolfinger R, O'connell M. Generalized linear mixed models a pseudo-likelihood approach. *J Stat Comput Simul.* 1993;48(3–4):233–43. <https://doi.org/10.1080/00949659308811554>.

Publisher's Note

Springer Nature remains neutral with regard to jurisdictional claims in published maps and institutional affiliations.

Ready to submit your research? Choose BMC and benefit from:

- fast, convenient online submission
- thorough peer review by experienced researchers in your field
- rapid publication on acceptance
- support for research data, including large and complex data types
- gold Open Access which fosters wider collaboration and increased citations
- maximum visibility for your research: over 100M website views per year

At BMC, research is always in progress.

Learn more biomedcentral.com/submissions

