

Flow analysis: A novel approach for classification

Christina Vakh, Marina Falkova, Irina Timofeeva, Alexey Moskvina, Leonid Moskvina & Andrey Bulatov

To cite this article: Christina Vakh, Marina Falkova, Irina Timofeeva, Alexey Moskvina, Leonid Moskvina & Andrey Bulatov (2015): Flow analysis: A novel approach for classification, Critical Reviews in Analytical Chemistry, DOI: [10.1080/10408347.2015.1087301](https://doi.org/10.1080/10408347.2015.1087301)

To link to this article: <http://dx.doi.org/10.1080/10408347.2015.1087301>



Accepted online: 12 Sep 2015.



Submit your article to this journal [↗](#)



View related articles [↗](#)

Flow analysis: A novel approach for classification

Christina Vakh*, Marina Falkova, Irina Timofeeva, Alexey Moskvina, Leonid Moskvina,

Andrey Bulatov

Department of Analytical Chemistry, Institute of Chemistry, Saint-Petersburg University

St.Petersburg State University, SPbSU, SPbU, 7/9 Universitetskaya nab., St. Petersburg,

199034 Russia

Abstract

We have suggested a novel approach for classification of flow methods according to the conditions under which the mass transfer processes and chemical reactions take place in the flow mode: *dispersion-convection flow methods (1)* and *forced-convection flow methods (2)*. The first group includes continuous flow analysis, flow injection analysis, all injection analysis, sequential injection analysis, sequential injection chromatography, cross injection analysis, multicommutated flow analysis, multisyringe flow injection analysis, multi-pumping flow systems, loop flow analysis and simultaneous injection effective mixing flow analysis. The second group includes segmented flow analysis, zone fluidics, flow batch analysis, sequential injection analysis with a mixing chamber, stepwise injection analysis and multicommutated stepwise injection analysis. The offered classification allows to systematize a large number of the flow methods. Recent development and application of dispersion-convection flow methods and forced-convection flow methods are presented.

* Corresponding author. *E-mail address: kristina-fulmes@spbu.ru; kristina-fulmes@mail.ru; Tel./fax +79112613385*

Keywords

Flow analysis, classification of flow methods, automation, miniaturization

Abbreviation

AIA – All injection analysis

BI – Bead-injection

CFA – Continuous flow analysis

CIA – Cross injection analysis

DCFM – Dispersion-convection flow methods

FCFM – Forced-convection flow methods

FBA – Flow batch analysis

FIA – Flow injection analysis

LAV – Lab-at-valve

LFA – Loop flow analysis

LOV – Lab-on-valve

MC – Mixing coil

MCh – Mixing chamber

MCFA – Multicommutated flow analysis system

MCSWIA – Multicommutated stepwise injection analysis

MPFS – Multi-pumping flow system

MSFIA – Multisyringe flow injection analysis

SFA – Segmented flow analysis

SIA – Sequential injection analysis

SIA MCh – Sequential injection analysis with a mixing chamber

SIC – Sequential injection chromatography

SIEMA – Simultaneous injection effective mixing flow analysis

SWIA – Stepwise injection analysis

ZF – Zone fluidics

1. Introduction

There is a tendency of automation of chemical analysis due to the necessity to carry out a large number of analyses of environmental, food, pharmaceuticals and chemical industrial samples. Another current tendency is a miniaturization of analytical systems since it allows reducing the sample and reagent consumption and waste generation. In this matter the flow methods have been recognized as universal tool for automation and miniaturization of various analytical procedures.

Flow methods have been invented in the second half of the 20th century [1-6] and have become an attractive field for researchers in automation of chemical analysis. Initially flow methods were focused on automation of liquid samples analysis, but later it became possible to automate gaseous [7, 8] and solid samples [9]. Herewith, the flow methods allow to automate the main stages of chemical analysis: sampling, sample pre-treatment (separation, derivatization et al.), measurement of the analytical signal. To automate the chemical analysis the flow systems usually include pumps, valves, commutative tubes, mixing/reaction devices and detectors. A sequence of all analytical procedures is often controlled by a computer or a microprocessor.

The flow methods are well described in quite numerous monographs [10-13] and reviews [14-17] where their main fundamental principles and applications are presented. The evolution of the flow methods has been discussed from different points of view such as historical aspects [18], the commutation concept [19], the effect on analytical methodologies [16] and towards the development of Green analytical chemistry [17].

In the early 21st century, almost simultaneously several groups of scientists have paid attention to the possibility of carrying out the analytical procedures in the special mixing chambers. The aim of the researchers was to provide the automation of chemical analysis with high sensitivity and versatility of the flow manifold.

At the present time the numerous flow methods have been developed. It has been previously suggested to divide flow methods into two groups [20]: *flow analysis with continuous sampling* and *flow analysis with intermittent sampling*, where sample portions are injected into the system from a sampling loop. The main criterion for the classification of these methods is the type of sample injection into the flow system. Nevertheless, this classification does not consider the processes occurring in the flow manifold.

We have suggested the classification of the flow methods, which is based on the conditions under which the mass transfer processes and chemical reactions take place in the flow mode. The offered classification will allow to systematize a large number of the flow methods and to discuss their general advantages and disadvantages.

2. Classification

The conditions under which the mass transfer processes and chemical reactions take place in the flow mode may be used as the main criteria for dividing of all flow methods into two groups (Fig. 1): *dispersion-convection flow methods (1)* and *forced-convection flow methods (2)*.

The concept of the first group assumes the delivering of sample zone in the laminar flow of a carrier to a detector. The two mass transfer phenomena primarily responsible for the transportation of samples through dispersion-convection flow systems are convection and

diffusion. They both affect the broadening of the sample zone, which is referred to as sample dispersion in analogy with both chromatography and chemical reaction engineering [21]. Herewith, on the one hand, the diffusion provides mixing of the sample with reagents, but on the other hand, it leads to the dispersion of the sample in the flow of a carrier. It should be pointed out, that generally the equilibrium of the chemical reaction, which is usually used in the DCFM, is not achieved during the moving of the sample zone in the commutative tubes to the flow detector, what causes the reduction of the sensitivity of analysis. In this case, the analytical signal is formed by convection under a laminar flow regime and diffusion phenomena.

The concept of the second group assumes the mixing of the sample and reagents under forced convection, which provides high efficiency of the mixing and elimination of the dispersion. This flow-batch approach was first used in a flow technique [22] and it is frequently called as flow-batch mode. This group of methods is characterized by the involving special MCh into the flow manifold, where the solution of samples and reagents are delivered.

Moreover, it is possible to achieve the equilibrium of the chemical reaction proceeding in the MCh. Unlike DCFM the analytical signal in FCFM is formed by forced convection under a turbulent regime.

The forms of the analytical signals obtained by using the DCFM and FCFM are presented in Fig. 1. The analytical signal in case of FCFM is the difference between the detector signals corresponding to the sample solution and the background, like the signals measured in manual techniques [23]. The analytical signal measured by using DCFM is an asymmetric peak [12], which is less than the maximum achieved by FCFM due to the FCFM provide complete mixing of sample and reagent solutions in the MCh.

Moreover, in case of DCFM the dispersion of the sample depends on several parameters such as sample volume, flow rate, length and diameter of the commutative tubes, configuration of the mixing coils and detector design [12]. The dispersion of the sample in the DCFM leads to the decrease of sensitivity in comparing with manual procedures. However, such reduction of the sensitivity is not observed in case of FCFM. This possibility was demonstrated on the determination of epinephrine in pharmaceuticals [24].

3. Dispersion-convection flow methods

3.1. *The concept and capabilities*

The first invented flow method was the continuous flow analysis (Fig. 2 a) [25]. The principal concept of CFA assumes the continuous analysis of liquid samples. Mixing of the sample with reagents solutions in the CFA is carried out in the reaction/mixing coils under the convection and diffusion and the analytical signal of sample is continuously measured. The CFA has been widely used in analytical practice for on-line analysis [26, 27]. But its main disadvantage is the significant consumption of reagents and respectively large volume of waste generation.

Another flow method which is included to the DCFM is the flow injection analysis (Fig. 2 b) [28, 29]. The principal concept of FIA assumes the periodic injection of discrete portions of sample into a continuous laminar and non-segmented flow of carrier by using the valve. Mixing of sample and reagents' zones is occurred in the MC under the influence of diffusion and convection. To achieve high reproducibility strictly constant values of flow rate as well as the diameter of the coils and commutative tubes are used. Mostly the equilibrium of the reaction is

not achieved. It leads to the sensitivity reduction. The stopped-flow mode can be used to increase sensitivity especially for kinetic methods [30].

The FIA systems exploiting reagent injection in a sample stream allow to reduce the volume of reagent solution significantly. This strategy was demonstrated at the molybdenum blue method and reagent consumption was reduced by up to 240-fold in comparison to FIA with sample injection into a carrier [31].

Another approach for reducing volumes of reagents was reported in [32]. The authors have developed a novel flow injection technique, called as an all injection analysis, where all reagent solutions are injected into a reaction coil and all solutions are circulated for a fixed time. By this circulating process, the amount of the reagents' consumption is extremely eliminated.

Cyclic FIA allows to realize others approaches to minimize reagent consumption. This opportunity has been shown in cyclic flow-injection spectrophotometric determination of lead (II) based on its reaction with Arsenazo III [33]. A cation-exchange resin AmberliteJRA-120 was included after the detection cell for regeneration of Arsenazo III. After analyte determination, the lead (II) was retained in the column and the released reagent was directed back to its original reservoir. Similar approaches can be used to determine other anions.

Nevertheless, the common drawback of FIA like as CFA and AIA is the necessity to redesign manifold for each analyte analysis.

The first versatile flow manifold was realized in sequential injection analysis (Fig. 2 c) [34]. The SIA manifold includes a multi-way valve, a holding coil, a syringe or/and peristaltic pump, a reaction coil and a flow detector. The SIA concept assumes the sequential delivery of portions of a carrier, a sample and reagent solution into the holding coil. After switching of the

valve and reversing of the pump the sample and reagent solutions are moved through the reaction coil to the detector. In this case, a concentration gradient is formed, which leads to the partially overlapping of sample and reagent zones, forming an area where the reaction product is generated. Efficiency of the sample and reagent zones overlapping influences on the analytical signal and depends on the physical parameters of the system (the injected sample volume, the flow rate, the length and diameter of tubes in a manifold, the configuration and volume of the holding and reaction coils, the detector design) and solution properties (viscosity, molecular diffusion coefficients) [35]. SIA compared to CFA and FIA allows to reduce the reagent consumption and waste generation significantly. In this respect, the most progressive implementation of SIA has become the SIA «Lab-on-valve» (SIA LOV) [36-38], which assumes the performing analysis in the channels of multi-way valve. The SIA LOV is attractive from the viewpoint of minimizing the sample volumes, especially for the analysis of biological samples [39] and expensive reagents consuming [40]. Later a simpler approach, SIA with lab-at-valve (LAV) concept, has been proposed [41-48]. It is employed by attaching a device integrating sample processing and detection units on a port of a multiposition selection valve. This makes the SIA LAV simpler than the SIA LOV. The SIA LAV unit can be built using an ordinary and less precise machine tool, to have suitable functions for chemistries of interest and with a nut that can plug in a port of the valve in the usual way.

The sequential injection chromatography [49-52] should be also included to the DCFM. The SIC involves the combination of liquid chromatography and sequential injection analysis. Sample solution and eluent by means of syringe pump and a switching valve are sequentially aspirated through a chromatographic column included into the SIC manifold (Fig. 2 d). The SIC

can be realized in reversed-phase mode [53]. The monolithic chromatographic columns with a high porosity are used in SIC. They allow providing high efficiency of analytes separation at low back pressure (2.5 MPa) which is produced in flow systems. The monolithic columns consist of a single piece of high-purity polymeric silica gel rod with a bimodal pore structure: mesopores (average size 13 nm) used for separation and macropores (average size 2 μm) used for mobile phase flowing. In the above mentioned review [49] capabilities of SIC and high performance liquid chromatography were compared. The main advantages of the SIC are the significant reduction of the reagents' consumption and the equipment cost. Furthermore, it becomes possible to perform the derivatization in the automated mode.

The idea of DCFM is also implemented in multicommutated flow analysis system [54-58], simultaneous injection-effective mixing analysis [59, 60], multi-pumping flow system [61-64], multisyringe flow injection analysis [65-68], loop flow analysis [69] and cross injection analysis [70] methods. These methods are characterized by high reduction of sample and reagents consumption in compared with CFA, FIA, AIA and even SIA.

The multicommutated flow analysis system (Fig. 3 a) consists of a peristaltic pump and a set of solenoid valves, by means of which the required portions of reagents and sample solutions are injected into a carrier flow. The location of valves and configuration of all communications depend on an automated technique. Elimination of overheating of valves is an important aspect in the operation of MCFA. If the valve is switched ON for a long time, the heating takes place and deformation of polytetrafluoroethylene channels of the valves is observed. This problem is solved by the installation of special protective electronic systems [12].

The simultaneous injection-effective mixing analysis [59, 60] is a hybrid format of FIA, SIA and MCFA. Sample and reagent solutions are aspirated into the several holding coils through the solenoid valves by a syringe pump (Fig. 3 b), and then the zones are simultaneously transferred in a carrier flow into a MC by reversed flow toward a detector. It leads to effective mixing and rapid detection.

The multi-pumping flow system (Fig. 3 c) includes a solenoid piston pumps operating in a pulse mode. The sample and reagents are injected into the flow system by means of pumps and then are mixed in the mixing coils. The efficiency of the sample and reagents zones mixing increases due to the pulsation of the piston pumps in MPFS. However, dispersion is not excluded. The great advantages of MPFS are the high throughput of analysis, flexibility, easy configuration, and robustness.

The multisyringe flow injection analysis manifold (Fig. 4 a) includes special panel equipped with four syringes. Each syringe in the top has a three-way valve which directs the solution from syringe to the flow system or returns it back to the reagent reservoir to avoid the mixing of solutions from other syringe. The main advantages of MSFIA are the high robustness of the system due to the absence of pumping tubes; the possibility of using aggressive solvents and reagents due to inert materials of syringes; the possibility to commutate the flow system with sample pretreatment devices (e.g. filters, sorption columns) with the opportunity to use high pressure.

The loop flow analysis has been introduced for water analysis [69]. The main parts of LFA manifold (Fig. 4 b) are the multichannel peristaltic pumps, multi-way rotary valve and cross-shaped flow cell. The hermetic closed loop provides full protection against background

interference. Firstly, a sample is introduced into the sample loop using pump. Then, pump is kept working and pump is used to propel the reagents solution into the reagent loop and the spectrophotometer is set to zero. When the valve is switched, the sample and reagent are mixed and both pumps are stopped for the formation of the colored complex, which is monitored by using a detector. The LFA was used for shipboard applications in marine science and in on-line environmental monitoring applications.

The last suggested method of the first group is the cross injection analysis [70]. This method assumes that sample and reagent solutions are injected perpendicularly into a carrier flow in a CIA cell (platform with cylindrical channels) by a peristaltic pump (Fig. 4 c). The mixing of the sample and reagent zones is carried out by their movement in a carrier flow from the CIA cell to the detector. The use of the CIA cell eliminates the need for valves using. Nevertheless, it does not provide the efficient mixing of the sample and reagent zones and the elimination of the dispersion.

3.2. *Recent development and application*

Nowadays, the DCFM are focused on the development of new automated sample preparation and multi-component methods and coupling of flow and separation methods (Table 1).

The liquid-liquid microextraction based on the DCFM has found wide application for sample preparation [71-73]. Several approaches have been developed: microcolumn phase separation [74], in-syringe approach [75] coupling with sequential injection system as well as magnetic stirring. The membrane methods of separation and preconcentration on the principles of DCFM such as pervaporation [76, 77] and gas diffusion [78-81] are also actively developed.

Moreover, the possibility of automation of single-drop headspace microextraction based on the SIA concept has been presented [82].

The bead-injection (BI) technique based on the principles of DCFM has used for sample preparation [83]. BI is the combination of the use of beads with a flowing stream of solution in a FIA/SIA system. Beads are utilized as solid surfaces to pre-concentrate or extract the analyte or to accommodate a chemical reaction. The flowing stream of solution is used to carry beads through the system. There is no need to regenerate the bead surfaces because they are discarded after each use and are replaced by fresh ones. It helps to reduce the risk of contamination, denaturation, and system clogging, and also, makes it possible to operate BI in the continuous flow system.

To increase the efficiency of SIA it was coupled with FIA [84]. Such coupling was implemented for the determination of lead (II) in water. The automated technique included the pre-concentration of the analyte in ion-exchange column operating in a sequential injection mode. After that, the elution of lead (II) was performed in flow injection mode for its subsequent spectrophotometric determination. Using such coupling flow system, it is possible not only to increase the sensitivity of lead (II) determination, but also to increase the sample throughput.

The effective implementation in DCFM is the coupling of the MSFIA and MPFS, which was applied for the determination of ^{226}Ra in water samples [85]. Such flow system allows increasing the sample throughput and reducing the reagent consumption.

To realize multi-component DCFM several approaches have been proposed. The first one means the simultaneous determination of several analytes provided by using of manifolds with several pumps, valves or detectors [86]. Additionally, the chemometric [87] and differential-

kinetic approaches [88] have been proposed and used in the multi-component flow analysis. The sample throughput of such systems is several times higher than conventional DCFM.

Recently attention was focused on coupling of DCFM with mass-spectrometry [89, 90], chromatography [91] and capillary electrophoresis [92-94] where the DCFM were used for automation of sample pretreatment. It was presented in the overview [94]. The benefits of hyphenated methods are high sensitivity and selectivity.

4. Forced convection flow methods

4.1. The concept and capabilities

The mixing under forced convection prevents the dispersion, that is common phenomenon in the methods of the first proposed group.

Mixing of the sample and reagents solution under forced convection is observed in the segmented flow analysis [95]. In SFA (Fig. 5 a) a continuous flow of the sample generated by a peristaltic pump is segmented by a gas bubbles and then mixed with the reagent flow in the mixing/reaction coils. The mixed flow is then moved to a flow detector, where the gas bubbles were preliminary removed. Sample segmentation by gas bubbles generates a turbulent flow, which leads to the homogenization of the reaction mixture. Furthermore, segmentation by the gas bubbles partially eliminates dispersion of the sample. SFA can be recognized as an intermediate approach between the flow methods of the first and the second groups.

The idea of forced convection has been better implemented in other FCFM: zone fluidics [96], flow batch analysis [23, 97-100], sequential injection analysis with a mixing chamber [101], stepwise injection analysis [102] and multicommutated stepwise injection analysis [103].

The last mentioned FCFM assume the main unit in the manifolds (Fig. 5 b-d, 6) – mixing chamber, where the portions of the samples and reagent solutions are sequentially delivered, mixed, thermostated (if necessary) and stored for a certain time to reach equilibrium.

Zone fluidics (Fig. 5 b) can be considered as a return to a SFA concept, but using the experience obtained in the SIA. ZF is defined as the precisely controlled physical, chemical, and fluid-dynamic manipulation of zones of miscible and immiscible fluids and suspended solids in narrow bore conduits to accomplish sample conditioning and chemical analysis. Fluids are propelled and manipulated in the manifold by means of a precise bi-directional flow pump. A holding coil between the pump and valve performs a similar role as in SIA. The ports of the multi-position valve are coupled to various reservoirs, reactors, unit operators, manifold devices, and detectors as indicated [96].

The mixing chamber in flow batch analysis is usually combined with a cell of the appropriate type of detector (Fig. 5 c). Portions of the sample and reagent solutions are sequentially delivered to the MCh by several peristaltic or solenoid piston pumps. Mixing of the reaction solutions is carried out in the MCh with a magnetic stirrer or fishing line connected to an electromotor, then pause is kept to complete the reaction and finally the measurement of an analytical signal is performed [104]. Nevertheless, the FBA manifold has certain limitations. Thus, the involving of special devices for mixing of solutions in the MCh in FBA manifold complicates the design of the analyzer. The combination of the MCh with the detector cell limits the possibility of varying the sample volume and using of several types of detectors in one flow analyzer. The increasing of the optical path length for measuring the analytical signal in

spectrophotometric analysis is less possible. In FBA, the optical path length usually does not exceed 10 mm due to the limitations of inner volume of the MCh.

The manifold of sequential injection analysis with a mixing chamber (Fig. 5 d) differs from FBA in the conditions of the sample zone formation. Mixing of the sample with the reagent solutions is carried out in the MCh, and then the solution of the reaction product is injected into a carrier flow and delivered through the reaction coil to a flow detector. This manifold is most similar to SIA, but it eliminates the problem associated with an inefficient mixing of the sample zone and reagent solutions as it was in SIA.

The idea of forced convection mixing was implemented in the stepwise injection analysis. The SWIA manifold (Fig. 6 a) is similar to the FBA manifold. The SWIA manifold includes a multi-way valve, a reversible peristaltic pump, a flow detector and a thermostated MCh. But the SWIA manifold always includes gas delivering channel to mix the sample with the reagent solutions into the MCh by babbling, unlike the FBA manifold, where the solutions are mixed using the magnetic stirrer or fishing line connected to an electromotor. MCh can be implemented for the dissolution of solid-phase samples or solid-phase extraction of analytes from the sample [105-107] as well as for the absorption of gaseous analytes [108, 109]. The concept of SWIA assumes that all stages of routine analysis are strictly performed: sampling; sample preparation, including analyte pre-concentration (if necessary) or derivatization; analyte absorption into solution, when gases are analyzed; the dissolution, when solid samples are analyzed; the addition of reagent solutions to the sample solution; mixing solutions by a babbling; thermostating (if necessary); a pause for the formation of reaction product; and finally the measurement of the analytical signal.

The multicommutated stepwise injection analysis was proposed for automation of a multicomponent spectrophotometric analysis. The MCSWIA manifold includes two similar eight-way solenoid valves and two peristaltic single-channel pumps (Fig. 6 b). The first valve is used for sequential injection of samples, reagent solutions and a gas phase into the flow system. The gas phase is used for mixing of solutions in the mixing chambers, which are coupled with the second valve. The number of MCh is determined by the number of analytes and the corresponding number of colour-forming reactions, which are necessary for their determination. In turn, the number of MCh is limited by the number of ports of a valve.

4.2. *Recent development and application*

The FCFM have already found applications for the automation of analysis of aqueous samples [110-112], biological fluids [113], pharmaceuticals [114], biofuels [115, 116] and other samples (Table 2). It should be noted that the benefit of FCFM is the versatility of flow manifolds. The involving of the MCh into the flow manifold allows to automate various procedures of sample pretreatment (dilution, liquid-liquid extraction, gas absorption, dissolution of soluble solid-phase samples as well as the extraction of the analyte from the solid-phase samples et al.) rather easily.

Thus, the liquid-liquid extraction of analytes can be realized directly in the MCh for the pre-concentration [117]. In this case, the effective mixing of the aqueous and organic phases, as well as the phases separation, are carried out in the glass MCh. Sample, reagents and organic solvent are introduced into the MCh by a peristaltic pump using air as a carrier.

The ZF measurement of octanol-water partition coefficient of drugs [114] was developed. In this case, the system is consisted of a syringe pump with a selection valve, a holding column,

a silica capillary flow-cell and an in-line spectrophotometer. Exact microliter volumes of solvents (octanol and phosphate buffer saline) and a solution of the drug, sandwiched between air segments, were sequentially loaded into the vertically aligned holding column. The distribution of the drug between the aqueous and octanol phases was occurred by the oscillation movement of the syringe pump piston.

The SWIA has been implemented for the dispersive liquid-liquid microextraction. The dispersion of the extractant was also performed directly in the MCh. Such procedure was used for the fully automated preconcentration and spectrophotometric determination of antipyrine in saliva [118].

The idea of automation of headspace single-drop micro-extraction has been implemented based on SWIA [119]. The most important features of the SWIA with headspace single-drop micro-extraction are: automated determination of volatile compounds in complicated matrices including suspension; the successful coupling of the continuous operating process of headspace single-drop micro-extraction with the UV-VIS technique. The efficiency of the proposed system was successfully demonstrated in ammonia determination in concretes.

In case of gas analysis, the gaseous sample is delivered to the MCh, which is filled with the acceptor solution. During the absorption the gaseous analytes are converted into the detectable forms. It was implemented in the SWIA determination of H₂S [108] and mercaptans [120] in the natural gas; phenols [109] and nitrogen oxides [121] in the atmospheric air. The developed techniques do not require the use of standard gas mixtures for the calibration of the analyzer. Its calibration is carried out by the standard solutions in the acceptor stream, which are delivered to the flow detector.

To automate the soluble solid-phase samples analysis the ZF [96] and the SWIA were used and applied for the determination of biologically active substances in medicinal herbs [107]. The extraction of the biologically active substances from medicinal plants was carried out in the MCh under ultrasonication.

Moreover, FCFMs allow also carrying out the standard addition method [122-124]. Standard addition method was implemented in the flow-batch procedure for iron determination by atomic absorption spectroscopy in the hydrated ethanol fuel [125]. In the developed FB procedure the MCh was coupled with a nebulizer of the flame atomic absorption spectrometer by means of the valve. In this procedure the portions of a fuel sample, a standard solution of the iron (III) and deionized water were mixed in the MCh. Then, the mixed solution from the MCh by the valve was delivered into the nebulizer of the flame atomic absorption spectrometer. The injected amount of iron into the fuel sample was regulated by the ratio of standard solution and solvent.

5. Conclusion

The proposed overview has been presented a critical discussion of the possibility to classify the flow methods into two groups according to the conditions under which the mass transfer processes and chemical reactions take place in the flow mode: dispersion-convection flow methods and forced-convection flow methods.

All methods of the first group are characterized by a high throughput due to the reactions generally do not achieve the chemical equilibrium. The mass transfer processes and chemical reactions are carried out under the influence of the convection and diffusion. Nevertheless, the strict order of reagents and sample injection into the flow system allows to achieve the excellent repeatability. In general, the sensitivity in dispersion-convection flow methods is lower in

comparison with manual procedures. This decreasing of sensitivity is caused by two factors. In case of kinetically slow chemical reactions, continuous flow of carrier does not allow to optimize the condition of reaction products' formation (optimal time and the temperature of reaction media). Analysis in a stopped-flow mode only partially solves the first problem, since in this case the dispersion of the sample is increased.

The forced-convection flow methods provide highly sensitive measurements due to the physical and chemical equilibriums of the analytical process are achieved and the dispersion of the sample is excluded. Moreover, another benefit of forced-convection flow methods is the versatility of the flow manifolds. The involving of the mixing chambers in the flow manifolds makes it easy to automate such operations as dilution, standard additions injection, liquid-liquid extraction, gas absorption and dissolution of solid-phase samples. The main drawback of forced-convection flow methods is a low throughput due to time-consuming procedures of sequential aspiration of reagent and sample solution and their mixing in mixing chamber to achieve equilibrium.

Acknowledgement

This work was supported by the Russian Foundation for Basic Research (projects no. 13-03-00031-a and no. 15-33-20068).

References

1. V. Wallace, Rapid automated analysis involving liquid-liquid extraction: Atropine by a prudhomme method, *Anal. Biochem.* 20 (3) (1967) 411-418.
2. N.R. Kuzel, Automated analysis of tertiary amines, *J. Pharm. Sci.* 57 (1968) 852–855.
3. N.R. Kuzel, H.E. Roudebush, C.E. Stevenson, Automated techniques in pharmaceutical analysis, *J. Pharm. Sci.* 58 (4) (1969) 381-406.
4. D.J. Blackmore, A.S. Curry, S. Hayes, E.R. Rutter, Automated Analysis for Drugs in Urine, *Clin. Chem.* 17 (1971) 896-902.
5. B. Karlberg, S. Thelander, Extraction based on the flow-injection principle: Part I. Description of the Extraction System, *Anal. Chim. Acta* 98 (1978) 1-7.
6. H. Bergamin, F.J.X. Medeiros, B.F. Reis, E.A.G. Zagatto, Solvent extraction in continuous flow injection analysis : Determination of molybdenum in plant material, *Anal. Chim. Acta* 101 (1978) 9-16.
7. P. Sritharathikhun, M. Oshima, S. Motomizu, On-line collection/concentration of trace amounts of formaldehyde in air with chromatomembrane cell and its sensitive determination by flow injection technique coupled with spectrophotometric and fluorometric detection, *Talanta* 67 (2005) 1014-1022.
8. Y. Wei, M. Oshima, J. Simon, L.N. Moskvina, S. Motomizu, Absorption, concentration and determination of trace amounts of air pollutants by flow injection method coupled with a chromatomembrane cell system: application to nitrogen dioxide determination, *Talanta* 58 (2002) 1343-1355.

9. E. Lamparter, C. Lunkenheimer, The automation of dissolution testing of solid oral dosage forms, *J. Pharm. Biomed. Anal.* 10 (1992) 727-733.
10. M. Trojanowicz, *Advances in flow analysis*, Wiley, Weinheim, 2008.
11. W. Frenzel. *Flow injection analysis. Principles, techniques and applications*. Technical University of Berlin, 1993.
12. A. Cerda, V. Cerda, *An introduction to flow analysis*, Sciware, Palma de Mallorca, 2009.
13. Y.A. Zolotov, L.N. Moskvina, L.N. Moskvina, A.V. Bulatov, *Chemical Flow analysis*, Nauka, Moscow, 2014.
14. E.H. Hansen, 30 years of flow injection analysis – and passing the torch, *Anal. Chim. Acta* 600 (2007) 4-5.
15. A.M. Idris, Overview of generations and recent versions of flow injection techniques, *Crit. Revs. Anal. Chem.* 40 (2010) 150-158.
16. E.H. Hansen, M. Miro, How flow-injection analysis (FIA) over the past 25 years has changed our way of performing chemical analyses, *Trends Anal. Chem.* 26 (2007) 18-26.
17. W.R. Melchert, B.F. Reis, F.R.P. Rocha, Green chemistry and the evolution of flow analysis. A review, *Anal. Chim. Acta.* 714 (2012) 8-19.
18. J. Ruzicka, E.H. Hansen, From beaker to microfluidics, *Anal. Chem.* 72 (2000) 212A.
19. E.A.G. Zagatto, B.F. Reis, C.C. Oliveira, R.P. Sartini, M.A.Z. Arruda, Evolution of the commutation concept associated with the development of flow analysis, *Anal. Chim. Acta* 400 (1999) 249-256.
20. W.E. van der Linden, Classification and definition of analytical methods based on flowing media, *Pure Appl. Chem.* 66 (1994) 2493-2500.

21. S.D. Kolev, I.D. Mckelvie, *Advances in flow injection analysis and related techniques*, Elsevier, Hungary, 2008.
22. A.C. Javier, S.R. Crouch, H.V. Malmstadt, Automated fast reaction-rate system for phosphate determinations in the millisecond range, *Anal. Chem.* 41 (1969) 239-243.
23. P.H.G.D. Diniz, L.F. de Almeida, D.P. Harding, M.C.U. de Araujo, Flow-batch analysis, *Trends Anal. Chem.* 35 (2012) 39-49.
24. A.V. Bulatov, A.V. Petrova, A.B. Vishnikin, A.L. Moskvina, L.N. Moskvina, Stepwise injection spectrophotometric determination of epinephrine, *Talanta* 96 (2012) 62–67.
25. W.B. Furman, W.H.C. Walker, *Continuous Flow Analysis: Theory and Practice*, first ed., Dekker, New York, 1976.
26. P. Herzsprung, A. Duffek, K. Friese, M. de Rechter, M. Schultze, W.v. Tümpling Jr., Modification of a continuous flow method for analysis of trace amounts of nitrate in iron-rich sediment pore-waters of mine pit lakes, *Water Res.* 39 (2005) 1887-1895.
27. E. Muñoz, S. Palmero, M.A. García-García, A continuous flow system design for simultaneous determination of heavy metals in river water samples, *Talanta* 57 (2002) 985–992.
28. J. Ruzicka, E.H. Hansen, *Flow injection analysis*, *Anal. Chim. Acta* 78 (1975) 145-157.
29. J. Ruzicka, E.H. Hansen, *Flow injection analysis*, second ed., Wiley, New York, 1988.
30. J. Ruzicka, The second coming of flow-injection analysis, *Anal. Chim. Acta* 261 (1992) 3-157.
31. P.B. Martelli, F.R.P. Rocha, R.C.P. Gorga, B.F. Reis, A Flow System for Spectrophotometric Multidetermination in Water Exploiting Reagent Injection, *J. Braz. Chem. Soc.* 13 (2002) 642-646.

32. H. Itabashi, H. Kawamoto, T. Kawashima, A Novel Flow Injection Technique: All Injection Analysis, *Analytical Sciences* 17(1) (2001) 229-231.
33. M. Zenki, K. Minamisawa, T. Yokoyama, Clean analytical methodology for the determination of lead with Arsenazo III by cyclic flow-injection analysis, *Talanta* 68 (2005) 281-286.
34. J. Ruzicka, G.D. Marshall, Sequential injection: a new concept for chemical sensors, process analysis and laboratory assays, *Anal. Chim. Acta* 237 (1990) 329-343.
35. T. Gubeli, G. Christian, J. Ruzicka, Fundamentals of sinusoidal flow sequential injection spectrophotometry, *Anal. Chem.* 63 (1991) 2407-2413.
36. L.D. Scampavia, J. Ruzicka, Micro-sequential injection: a multipurpose lab-on-valve for advancement of bioanalytical assays, *Anal. Sci.* 17 (2001) 429-431.
37. M.S. Decuir, H.M. Boden, A.D. Carroll, J. Ruzicka, Principles of micro sequential injection analysis in the lab-on-valve format and its introduction into a teaching laboratory, *J. Flow Injection Anal.* 24 (2007) 103-108.
38. J. Ruzicka, Lab-on-Valve: universal microflow analyzer based on sequential and bead injection, *Analyst* 125 (2000) 1053-1060.
39. C.M. Schulz, J. Ruzicka, Real-time determination of glucose consumption by live cells using a lab-on-valve system with an integrated microreactor, *Analyst* 127 (2002) 1293-1298.
40. H. Erxleben, J. Ruzicka, Atomic absorption spectroscopy for mercury, automated by sequential injection and miniaturized in lab-on-valve system, *Anal. Chem.* 77 (2005) 5124-5128.
41. K. Grudpan, Some recent developments on cost-effective flow-based analysis, *Talanta* 64 (2004) 1084-1090.

42. A.N. Anthemidis, E.I. Daftsis, N.P. Kalogiouri, A sequential injection lab-at-valve (SI-LAV) platform for hydride generation atomic absorption spectrometry (HG-AAS): on-line determination of inorganic arsenic, *Anal. Methods* 6 (2014) 2745-2750.
43. C. Mitani, A.N. Anthemidis, On-line liquid phase micro-extraction based on drop-in-plug sequential injection lab-at-valve platform for metal determination, *Anal. Chim. Acta* 771 (2013) 50-55.
44. K. Ponghong, S.K. Hartwell, K. Grudpan, Sequential injection Lab-at-valve (SI-LAV) segmented flow system for kinetic study of an enzyme, *Talanta* 85 (2011) 804-808.
45. W. Thongchai, B. Liawruangrath, S. Liawruangrath, Sequential injection analysis with lab-at-valve (SI-LAV) for the determination of solasodine in *Solanum* species, *Talanta* 81 (2010) 565-571.
46. R. Burakham, R. Burakham, J. Jakmune, K. Grudpan, Development of sequential injection-lab-at-valve (SI-LAV) micro-extraction instrumentation for the spectrophotometric determination of an anionic surfactant, *Anal. Sci.* 22 (2006)137-140.
47. R. Burakham, S. Lapanantnoppakhun, J. Jakmune, K. Grudpan, Exploiting sequential injection analysis with lab-at-valve (LAV) approach for on-line liquid-liquid micro-extraction spectrophotometry, *Talanta* 68 (2005) 416-421.
48. J. Jakmune, L. Patimapornlert, S. Suteerapataranon, N. Lenghor, K. Grudpan, Sequential injection with lab-at-valve (LAV) approach for potentiometric determination of chloride, *Talanta* 65 (2005) 789-793.

49. J. Huclova, D.Šatínský, O. Pavlíček, L. Vedralová, R.Karlíček, Using on-line solid phase extraction for determination of amiloride in human urine by sequential injection technique, *Anal. Chim. Acta* 573-574 (2006) 376-382.
50. R.E.E. Elgorashe, A.M. Idris, M.A. Abdelrahman, A.E.M. Saeed, Facile Assay Method for Norfloxacin and Ciprofloxacin by Sequential Injection Chromatography, *Acta Chromatographica* 26 (2014) 321-334.
51. A.M. Idris, The Second Five Years of Sequential Injection Chromatography: Significant Developments in the Technology and Methodologies, *Crit. Rev. Anal. Chem.* 44 (2014) 220-232.
52. L.B.O. dos Santos, C.M.C. Infante, J.C. Masini, Development of a sequential injection chromatography (SIC) method for determination of simazine, atrazine, and propazine, *J. Sep. Sci.* 32 (2009) 494-500.
53. R. de Prá Urio, J.C. Masini, Evaluation of sequential injection chromatography for reversed phase separation of triazine herbicides exploiting monolithic and core-shell columns, *Talanta* 131 (2015) 528-534.
54. M.A. Feres, P.R. Fortes, E.A.G. Zagatto, J.L.M. Santos, J.L.F.C. Lima, Multicommution in flow analysis: concepts, applications and trends, *Anal. Chim. Acta* 618 (2008) 1-17.
55. F.R.P. Rocha, B.F. Reis, E.A.G. Zagatto, J.L.F.C. Lima, R.A.S. Lapa, J.L.M. Santos, Multicommution in flow analysis: concepts, applications and trends, *Anal. Chim. Acta* 468 (2002) 119-131.
56. M.C. Icardo, J.V.G. Mateo, J.M. Calatayud, Multicommution as a powerful new analytical tool, *Trends Anal. Chem.* 2 (2002) 366-378.

57. C. Pons, R. Forteza, A.O.S.S. Rangel, V. Cerda, The application of multicommutated flow techniques to the determination of iron, *Trends Anal. Chem.* 25 (2006) 583-588.
58. M.A. Segundo, A.O.S.S. Rangel, Flow analysis: a critical view of its evolution and perspective, *J. Flow Injection Anal.* 19 (2002) 3-8.
59. N. Teshima, D. Noguchi, Y. Joichi, N. Lenghor, N. Ohno, T. Sakai, S. Motomizu, Simultaneous injection-effective mixing analysis of Palladium, *Anal. Sciences* 26 (2010) 143-144.
60. N. Ratanawimarnwong, K. Ponghong, N. Teshima, D. Nacapricha, K. Grudpan, T. Sakai, S. Motomizu, Simultaneous injection effective mixing flow analysis of urinary albumin using dye-binding reaction, *Talanta* 96 (2012) 50-54.
61. R.A.S. Lapa, J.L.F.C. Lima, B.F. Reis, J.L.M. Santos, E.A.G. Zagatto, Multi-pumping in flow analysis: concepts, instrumentation, potentialities, *Anal. Chim. Acta* 466 (2002) 125-132.
62. J.L.M. Santos, M.F.T. Ribeiro, J.L.F.C. Lima, A.C.B. Dias, E.A.G. Zagatto, Multipumping flow systems: an alternative approach to sample handling in spectroscopy measurements, *Spectrosc. Lett.* 40 (2007) 41-50.
63. J.L.M. Santos, M.F.T. Ribeiro, A.C.B. Dias, J.L.F.C. Lima, E.E.A. Zagatto, Multipumping flow systems: the potential of simplicity, *Anal. Chim. Acta* 600 (2007) 21-28.
64. J.L.F.C. Lima, J.L.M. Santos, A.C.B. Dias, M.F.T. Ribeiro, E.A.G. Zagatto, Multi-pumping flow systems: an automation tool, *Talanta* 64 (2004) 1091-1098.
65. A.V. Cerda, R. Forteza, J.M. Estela, Potential of multisyringe flow-based multicommutated systems, *Anal. Chim. Acta* 600 (2007) 35-45.

66. M.A. Segundo, L.M. Magalhaes, Multisyringe flow injection analysis: state-of-the-art and perspectives, *Anal. Sci.* 22 (2006) 3-8.
67. M. Miro, V. Cerda, J.M. Estela, Multisyringe flow injection analysis: characterization and applications, *Trends Anal. Chem.* 21 (2002) 199-210.
68. B. Horstkotte, O. Elsholz, V. Cerda, Review on automation using multisyringe flow injection analysis, *J. Flow Injection Anal.* 22 (2005) 99-109.
69. J. Ma, Q. Li, D. Yuan, Loop flow analysis of dissolved reactive phosphorus in aqueous samples, *Talanta* 123 (2014) 218-223.
70. D. Nacapricha, P. Sastranurak, T. Mantim, N. Amornthammarong, K. Uraisin, C. Boonpanaid, C. Chuyprasartwattana, P. Wilairat, Cross injection analysis: Concept and operation for simultaneous injection of sample and reagents in flow analysis, *Talanta* 110 (2013) 89-95.
71. M. Fernando, B. Horstkotte, J.M. Estela, V. Cerdà, Automated in-syringe dispersive liquid-liquid microextraction, *Trends Anal. Chem.* 59 (2014) 1-8.
72. M. Miró, E.H. Hansen, On-line sample processing involving microextraction techniques as a front-end to atomic spectrometric detection for trace metal assays, *Anal. Chim. Acta.* 782 (2013) 1-11.
73. S.Clavijo, J. Avivar, R. Suárez, V. Cerdà, Analytical strategies for coupling separation and flow-injection techniques, *Trends Anal. Chem.* 67 (2015) 26-33.
74. A.N. Anthemidis, K.G. Ioannou, On-line sequential injection dispersive liquid-liquid microextraction system for flame atomic absorption spectrometric determination of copper and lead in water samples, *Talanta*. 79 (2009) 86-91.

75. B. Horstkotte, M. Alexovič, F. Maya, C.M. Duarte, V. Andruch, V.Cerdá, Automatic determination of copper by in-syringe dispersive liquid-liquid microextraction of its bathocuproine-complex using long path-length spectrophotometric detection, *Talanta* 99 (2012) 349-356.
76. E. Mataix, M.D. Luque de Castro, Determination of total and free sulfur dioxide in wine by pervaporation-flow injection analysis, *Analyst* 123 (1998) 1547-1553.
77. T. Rupasinghe, T.J. Cardwell, R.W. Cattrall, I.D. Potter, S.D. Kolev, Determination of arsenic by pervaporation-flow injection hydride generation and permanganate spectrophotometric detection, *Anal. Chim. Acta* 510 (2004) 225-230.
78. A. Dhaouadi, L. Monser, S. Sadok and N. Adhoum, Validation of a flow-injection-gas diffusion method for total volatile basic nitrogen determination in seafood products, *Food Chem.* 103 (2007) 1049-1053.
79. N. Amini, T.J. Cardwell, R.W. Cattrall, S. Kolev, Determination of mercury(II) at trace levels by gas-diffusion flow injection analysis with amperometric detection, *Anal. Chim. Acta* 539 (2005) 203-207.
80. S.D. Kolev, P. Fernandes, D. Satinsky, P. Solich, Highly sensitive gas-diffusion sequential injection analysis based on flow manipulation, *Talanta*. 79 (2009) 1021-1026
81. R.B. Mesquita, A.O. Rangel, Gas diffusion sequential injection system for the spectrophotometric determination of free chlorine with o- dianisidine, *Talanta* 68 (2005) 268-273.

82. A.N. Anthemidis, I.S.I. Adam, Development of on-line single-drop micro-extraction sequential injection system for electrothermal atomic absorption spectrometric determination of trace metals, *Anal. Chim. Acta* 632 (2009) 216-220.
83. S. Kradtap Hartwell, G.D. Christian, K. Grudpan, Bead injection with a simple flow-injection system: an economical alternative for trace analysis, *Trends Anal. Chem.* 23 (2004) 619-623.
84. R.B.R. Mesquita, S.M.V. Fernandes, A.O.S.S. Rangel, A flow system for the spectrophotometric determination of lead in different types of waters using ion-exchange for pre-concentration and elimination of interferences, *Talanta* 62 (2004) 395-401.
85. Y. Fajardo, E. G´omez, F. Garcias, V. Cerd`a, M. Casas, Development of an MSFIA-MPFS pre-treatment method for radium determination in water samples, *Talanta* 71 (2007) 1172-1179.
86. M. Dolores Luque de Castro, M. Valcarcel Cases, Simultaneous Determinations in Flow Injection Analysis. A Review, *Analyst* 109 (1984) 413-419.
87. J. Saurina, S. Hernandez-Cassou, Quantitative determinations in conventional flow injection analysis based on different chemometric calibration strategies: a review, *Anal. Chim. Acta* 438 (2001) 335-352.
88. L.K. Shpigun, Ya.V. Shushenachev, P.M. Kamilova, Kinetic separation in flow injection spectrophotometry: Simultaneous determination of copper and zinc in a single run, *Anal.Chim. Acta* 28 (2006) 360-365.
89. C.M.P.V. Lopes, A.A. Almeida, M.L.M.F.S. Saraiva, J.L.F.C. Lima, Determination of Rh, Pd and Pt in urine samples using a pre-concentration sequential injection analysis system

coupled to a quadrupole-inductively coupled plasma-mass spectrometer, *Anal. Chim. Acta* 600 (2007) 226-232.

90. A. Sabarudin, N. Lenghor, M. Oshima, L. Hakim, T. Takayanagi, Y.-H. Gao, S. Motomizu, Sequential-injection on-line preconcentration using chitosan resin functionalized with 2-amino-5-hydroxy benzoic acid for the determination of trace elements in environmental water samples by inductively coupled plasma-atomic emission spectrometry, *Talanta* 72 (2007) 1609-1617.

91. H.M. González-San Miguel, M. Fernández, J.M. Estela, V. Cerdà, Contribution of multi-commuted flow analysis combined with monolithic columns to low-pressure, high-performance chromatography, *Trends Anal. Chem.* 28 (2009) 336-346.

92. P. Kuban, B. Karlberg, Flow/sequential injection sample treatment coupled to capillary electrophoresis. A review, *Anal. Chim. Acta* 648 (2009) 129-145.

93. B. Horstkotte, O. Elsholz, V. Cerda, Development of a capillary electrophoresis system coupled to sequential injection analysis and evaluation by the analysis of nitrophenols, *Intern. J. Environ. Anal. Chem.* 87 (2007) 797-811.

94. B. Horstkotte, V. Cerdà, Coupling of flow techniques with capillary electrophoresis: review of operation principles, challenges, potentials, and applications, *J. Chromatogr. Sci.* 47 (2009) 636-647.

95. S. Coverly. Segmented flow analysis. *Encyclopedia Anal. Sci.* (Second Ed.) (2005) 210-216

96. G. Marshall, D. Wolcott, D. Olson, Zone fluidics in flow analysis: potentialities and applications, *Anal. Chim. Acta* 499 (2003) 29-40.

97. J. Skrlíkova, V. Andruch, H. Sklenářová, P. Chocholouš, P. Solich, I.S. Balogh, A novel dual-valve sequential injection manifold (DV-SIA) for automated liquid–liquid extraction. Application for the determination of picric acid, *Anal. Chim. Acta* 666 (2010) 55-61.
98. J. Skrlíkova, V. Andruch, H. Sklenářová, P. Chocholouš, P. Solich, I.S. Balogh, An air-assisted liquid–liquid extraction using a dual-valve sequential injection manifold (DV-SIA): Determination of copper, *Anal. Methods* 2 (2010) 1134-1139.
99. C.C. Acebal, H. Sklenářová, J. Škrliková, I. Šrámková, V. Andruch, I.S. Balogh, P. Solich, Application of DV-SIA manifold for determination of thiocyanate ions in human saliva samples, *Talanta* 96 (2012) 107-112.
100. M. Alexovic, V. Andruch, I.S. Balogh, J. Šandrejová, A single-valve sequential injection manifold (SV-SIA) for automation of air-assisted liquid-phase microextraction: stopped flow spectrophotometric determination of chromium(VI), *Anal. Methods* 5 (2013) 2497-2502.
101. T. McCormack, J.F van Staden, Use of a sequential injection technique to evaluate the effect of mixing chambers on zone penetration, *Anal. Chim. Acta* 367 (1998) 111-121.
102. A.V. Bulatov, A.L. Moskvin, L.N. Moskvin, A.V. Mozhuhin, The stepwise injection analysis as a new opportunity for automation of chemical analysis of liquid, gaseous and solid-phase samples, *J. Flow Injection Anal.* 27 (2010) 13-19.
103. Ch.S. Fulmes, A.V. Bulatov, O.G. Yasakov, E.A. Freze, A. N. Moskvin, Y. M. Dedkov, Leonid N. Moskvin, Multicommutated Stepwise injection analysis as new approach for simultaneous determination of nickel (II), copper (II) and zinc (II) in wet aerosols, *Microchem. J.* 110 (2013) 649-65.

104. S.S. Monte-Filho, M.B. Lima, S.I.E. Andrade, D.P. Harding, Y.N.M. Fagundes, S. R.B. Santos, Sherlan G. Lemos, M.C.U. Araújo, Flow-batch miniaturization, *Talanta* 86 (2011) 208-213.
105. M.T. Falkova, M.O. Pushina, A.V. Bulatov, G.M. Alekseeva, L.N. Moskvina, Stepwise injection spectrophotometric determination of flavonoids in medicinal plants, *Anal. Lett.* 47 (2014) 970-982.
106. M. Falkova, M. Alexovic, M. Pushina, A. Bulatov, L. Moskvina, V. Andrich, Fully automated on-line flow-batch based ultrasound-assisted surfactant-mediated extraction and determination of anthraquinones in medicinal plants, *Microchem. J.* 116 (2014) 98-106.
107. M.T. Falkova, A.V. Bulatov, M.O. Pushina, A.A. Ekimov, G. M. Alekseeva, L.N. Moskvina, Multicommutated stepwise injection determination of ascorbic acid in medicinal plants and food samples by capillary zone electrophoresis ultraviolet detection, *Talanta* 133 (2015) 82-87.
108. A.V. Bulatov, D.K. Goldvirt, L.N. Moskvina, A.L. Moskvina, E.A. Vaskova, Stepwise injection photometric determination of hydrogen sulfide in natural gas, *J. Flow Injection Anal.* 23 (2006) 102-106.
109. A. Bulatov, M. Soloviev, A. Petrova, A. Moskvina, L. Moskvina, Stepwise injection photometric determination of phenols in air, *J. Flow Injection Anal.* 27 (2010) 158-161.
110. S.I.E. Andrade, M.B. Lima, I.S. Barreto, W.S. Lyra, L.F. Almeida, M.C.U. Araújo, E.C. Silva, A digital image-based flow-batch analyzer for determining Al(III) and Cr(VI) in water, *Microchem. J.* 109 (2013) 106-111.

111. L.F. Almeida, M.G.R. Vale, M.B. Dessuy, M.M. Silva, R.S. Lima, V.B. Santos, P.H.D. Diniz, M.C.U. Araújo, A flow-batch analyzer with piston propulsion applied to automatic preparation of calibration solutions for Mn determination in mineral waters by ET AAS, *Talanta* 73 (2007) 906-912.
112. R.A.C Lima, S.R.B Santos, R.S. Costa, G.P.S Marcone, R.S. Honorato, V.B. Nascimento, M.C.U Araujo, Hardness screening of water using a flow-batch photometric system, *Anal. Chim. Acta* 518 (2004) 25-30.
113. M.J. Silva, E.V. Anjos, R.S. Honorato, M.F. Pimentel, A.P.S. Paim, Spectrophotometric cocaine determination in a biphasic medium employing flow-batch sequential injection analysis, *Anal. Chim. Acta* 629 (2008) 98-103.
114. P. Wattanasin, P. Saetear, P. Wilairat, D. Nacapricha, S. Teerasong, Zone fluidics for measurement of octanol–water partition coefficient of drugs, *Anal. Chim. Acta* 860 (2015) 1-7.
115. M.B. Lima, M. Insausti, C.E. Domini, M.F. Pistonesi, M.C.U.Araujo, B.S.F. Band, Automatized flow-batch method for fluorescent determination of free glycerol in biodiesel samples using on-line extraction, *Talanta* 89 (2012) 21-26.
116. Ch.S. Vakh, A.V. Bulatov, A.Y. Shishov, A.V. Zabrodin, L.N. Moskvina, Determination of silicon, phosphorus, iron and aluminum in biodiesel by multicommutated stepwise injection analysis with classical least squares method, *Fuel* 135 (2014) 198-204.
117. M.C.T. Diniza, O.F. Filhob, J.J.R. Rohwedder, An automated system for liquid–liquid extraction based on a new micro-batch extraction chamber with on-line detection. Preconcentration and determination of copper (II), *Anal. Chim. Acta* 525 (2004) 281-287.

118. A. Bulatov, K. Medinskaia, D. Aseeva, S. Garmonov, L. Moskvina, Determination of antipyrine in saliva using the dispersive liquid-liquid microextraction based on a stepwise injection system, *Talanta* 133 (2015) 66-70.
119. I. Timofeeva, I. Khubaibullin, M. Kamencev, A. Moskvina, A. Bulatov, Automated procedure for determination of ammonia in concrete with headspace single-drop microextraction by stepwise injection spectrophotometric analysis, *Talanta* 133 (2015) 34-37.
120. A.V. Bulatov, D.K. Goltvitr, L.N. Moskvina, Stepwise injection photometric determination of mercaptans in hydrocarbon gases, *J. Anal. Chem.* 3 (2008) 253-258.
121. A. Bulatov, K. Medinskaia, A. Ipatov, A. Moskvina, L. Moskvina, Stepwise injection photometric determination of nitrogen oxides in atmospheric air, *J. Flow Injection Anal.* 28 (2011) 146-149.
122. J.E. Silva, F.A. Silva, M.F. Pimentel, R.S. Honorato, V.L. Silva, M.C.B.S.M. Montenegro, A.N. Araújo, A flow-batch internal standard procedure for iron determination in hydrated ethanol fuel by flame atomic absorption spectrometry, *Talanta* 70 (2006) 522-526.
123. A.V. Bulatov, Y.K. Protzenko, K.A. Subbotina, L.N. Moskvina, A.L. Moskvina, Stepwise injection photometric determination of phosphates in urine, *J. Flow Injection Anal.* 25 (2008) 24-28.
124. I.P.A. Morais, M.R.S. Souto, A.O.S.S. Rangel, Sequential injection standard addition system with a mixing chamber: determination of orthophosphate in waters, *J. Flow Injection Anal.* 20 (2003) 187-192.

125. J.E. Silva, F.A. Silva, M. F. Pimentel, R. S. Honorato, V.L.Silva, M.C.B.S.M. Montenegro, A.N. Araujo, A flow-batch internal standard procedure for iron determination in hydrated ethanol fuel by flame atomic absorption spectrometry, *Talanta* 70 (2006) 522-526.
126. İ. Durukan, Ç.A. Şahin, S. Bektaş, Determination of copper traces in water samples by flow injection-flame atomic absorption spectrometry using a novel solidified floating organic drop microextraction method, *Microchem. J.* 98 (2011) 215-219.
127. S. Satienerakul, P. Phongdong, S. Liawruangrath, Pervaporation flow injection analysis for the determination of sulphite in food samples utilising potassium permanganate–rhodamine B chemiluminescence detection, *Food Chem.* 121 (2010) 893-898.
128. A.N. Ramdzan, P.J. Mornane, M.J. McCullough, W. Mazurek, S.D. Kolev, Determination of acetaldehyde in saliva by gas-diffusion flow injection analysis, *Anal. Chim. Acta* 786 (2013) 70-77.
129. J.L. Adcock, P.S. Francis, K.M. Agg, G.D. Marshall, N.W. Barnett, A hybrid FIA/HPLC system incorporating monolithic column chromatography, *Anal. Chim. Acta* 600 (2007) 136-141.
130. A. Sixto, M. Knochen, Multicommutated flow system for the determination of glucose in honey with immobilized glucose oxidase reactor and spectrophotometric detection, *Talanta* 77 (2009) 1534-1538.
131. R.N. Fernandes, B.F. Reis, Flow system exploiting multicommutation to increase sample residence time for improved sensitivity. Simultaneous determination of ammonium and ortho-phosphate in natural water, *Talanta* 58 (2002) 729-737.

132. S. Koronkiewicz, S. Kalinowski, Application of direct-injection detector integrated with the multi-pumping flow system to photometric stop-flow determination of total iron, *Talanta* 96 (2012) 68-74.
133. R. Suárez, B. Horstkotte, V. Cerdà, In-syringe magnetic stirring-assisted dispersive liquid-liquid microextraction for automation and downscaling of methylene blue active substances assay, *Talanta* 130 (2014) 555-560.
134. V. Andruch, C.C. Acebal, J.Škrliková, H. Sklenářová, P.Solich, I.S. Balogh, F. Billes, L. Kocúrová, Automated on-line dispersive liquid-liquid microextraction based on a sequential injection system, *Microchem. J.* 100 (2012) 77-82.
135. A.N. Anthemidis, K.I. Ioannou, Development of a sequential injection dispersive liquid-liquid microextraction system for electrothermal atomic absorption spectrometry by using a hydrophobic sorbent material: determination of lead and cadmium in natural waters, *Anal. Chim. Acta* 668 (2010) 35-40.
136. N. Chu, S. Fan, Sequential injection kinetic spectrophotometric determination of quaternary mixtures of carbamate pesticides in water and fruit samples using artificial neural networks for multivariate calibration, *Spectrochim. Acta Part A* 74 (2009) 1173-1181.
137. Th. Mai, P.C. Hauser, Simultaneous separations of cations and anions by capillary electrophoresis with contactless conductivity detection employing a sequential injection analysis manifold for flexible manipulation of sample plugs, *J. Chromatogr. A* 1267 (2012) 266-272.
138. C. Vakh, E. Freze, A. Pochivalov, E. Evdokimova, M. Kamencev, L. Moskvina, A. Bulatov, Simultaneous determination of iron (II) and ascorbic acid in pharmaceuticals based on flow sandwich technique, *J. Pharmacol. Toxicol. Methods* 73 (2015) 56-62.

139. A.D. Batista, C.F. Nascimento, W.R. Melchert, F.R.P. Rocha, Expanding the separation capability of sequential injection chromatography: Determination of melamine in milk exploiting micellar medium and on-line sample preparation, *Microchem. J.* 117 (2014) 106-110.
140. J.R.E. Thabano, C.Tjell Jens, G.M. Sawula, Determination of orthophosphates using a macro segmented flow analyzer (MSFA) based on colorimetric detection, *Talanta* 64 (2004) 60-68.
141. M.JR. Lima, S.M.V Fernandes, A.O.S.S Rangel, Sequential injection titration of chloride in milk with potentiometric detection, *Food Control* 15 (2004) 609-613.
142. P.C.A.G. Pinto, M. L. M.F.S. Saraiva, J.L.F.C. Lima, A flow sampling strategy for the analysis of oil samples without pre-treatment in a sequential injection analysis system, *Anal. Chim. Acta* 555 (2006) 377-383.

Tables caption

Table 1. Applications of the dispersion-convection flow methods.

Flow method	Detection technique	Matrix	Analyte	On-line pre-treatment of sample	LOD	Sample throughput, h ⁻¹	Ref.
CFA	ASPM	river water	Zn ²⁺ , Cd ²⁺ , Fe ²⁺ , Cu ²⁺ , Ni ²⁺ , Co ²⁺ , Cr(VI)	-	0.176- 4.01 nM	-	27
FIA	FAAS	water	Cu	solidified floating organic drop microextraction	0.58 ng L ⁻¹	3	126
FIA	CL	food	SO ₃ ²⁻	pervaporation	0.2 mg L ⁻¹	30	127
FIA	UV-Vis	saliva	acetaldehyde	gas-diffusion	12.3 μg L ⁻¹	9	128
FIA	HPLS- UV	urine	opiate alkaloids biogenic amines	derivatization	0.2-5 10 ⁻⁷ M	-	129
MCFA	UV-Vis	honey	glucose	derivatization (immobilized glucose oxidase reactor)	0.073 g L ⁻¹	20	130
MCFA	UV-Vis	water	NH ₄ ⁺ , PO ₄ ³⁻	-	7 μg L ⁻¹ NH ₄ ⁺ 17 μg L ⁻¹ PO ₄ ³⁻	56	131

MPFS	UV-Vis	water	Fe	oxidation	0.15 mg L ⁻¹	180	132
MSFIA	UV-Vis	water	anionic surfactants	magnetic stirring-assisted dispersive liquid-liquid microextraction	7 μg L ⁻¹	10	133
SIA	UV-Vis	saliva	SCN ⁻	dispersive liquid-liquid microextraction	0.017 mg L ⁻¹	-	134
SIA	ETAAS	water	Cu, Cd	dispersive liquid-liquid microextraction	10 ngFe L ⁻¹ 2 ngCd L ⁻¹	10	135
SIA	ETAAS	water	Cd	single-drop micro-extraction	0.01 μg L ⁻¹	6	82
SIA	UV-Vis and BP-ANN	water	carbamate pesticides	-	0.2-0.4 mg L ⁻¹	18	136
SIA	Q-ICP-MS	urine	Rh, Pd, Pt	pre-concentration (Metalfix TM Chelamine TM resin)	0.4-1.2 ng L ⁻¹	9	89
SIA	CE-CD	water	NH ₄ ⁺ , K ⁺ , Ca ²⁺ , Na ⁺ , Mg ²⁺ , Mn ²⁺ , Zn ²⁺ , Cd ²⁺ , Ba ²⁺ , Cl ⁻ , S ₂ O ₃ ²⁻ , NO ₃ ⁻ , SO ₄ ²⁻ , NO ₂ ⁻	-	0.3-2 μM	-	137
SIA	UV-Vis	drugs	Fe (II), ascorbic acid	-	0.2 mg L ⁻¹ Fe (II), 0.2 mg L ⁻¹ ascorbic acid	41	138
SIC	UV-Vis	milk	melamine	dilution of the	0.6 mg	9	139

				sample with sodiumdodecyl sulfate using a multiposition valve	L^{-1}		
--	--	--	--	---	----------	--	--

UV-Vis – spectrophotometry, ETAAS – electrothermal atomic absorption spectrometry, FAAS – Flame atomic absorption spectrometry, CL – chemiluminescence, BP-ANN – back-propagation-artificial neural network algorithms for multivariate quantitative analysis, Q-ICP-MS – quadrupole-inductively coupled plasma-mass spectrometry, CE-CD – capillary electrophoresis with contactless conductivity detection, HPLS-UV – High-performance liquid chromatography with UV-detection, ASPM – adsorptive stripping potentiometry.

Table 2. Applications of the forced-convection flow methods.

Flow method	Detection technique	Matrix	Analyte	On-line pre-treatment of sample	LOD	Sample throughput, h ⁻¹	Ref.
SFA	UV-Vis	water and wastewater	PO ₄ ³⁻	-	4 µg L ⁻¹	40	140
FB	FAAS	hydrated ethanol fuel	Fe	standard-addition method	0.04 mg L ⁻¹	10	122
FB	UV-Vis	water	Cu ²⁺	liquid-liquid extraction	5 µg L ⁻¹	14	117
FB	UV-Vis	biodiesel	glycerol	derivatization	0.036 mg L ⁻¹	14	115
MCSWIA	UV-Vis	biodiesel	Al ³⁺ , Fe ³⁺ , Si, P	-	0.3 mg kg ⁻¹ Al ³⁺ 0.6 mg kg ⁻¹ Fe ³⁺ , Si, P	6	116
SIA-MC	PM	milk	Cl ⁻	pseudo-titration	0.1 mM	17	141
SIA-MC	UV-Vis	water	PO ₄ ³⁻	standard-addition method	0.024 mgP L ⁻¹	324	124
SIA-MC	UV-Vis	oil	Fe ³⁺	dilution	0.31 mg	20	142

					L ⁻¹		
SWIA	UV-Vis	urine	PO ₄ ³⁻	standard-addition method	0.6 mg L ⁻¹	10	123
SWIA	UV-Vis	saliva	antipyrine	derivatization, dispersive liquid-liquid microextraction	1 μM	5	118
SWIA	UV-Vis	natural gas	hydrogen sulfide	absorption	20 μg m ⁻³	20	108
SWIA	UV-Vis	medicinal plants	anthraquinones	ultrasound-assisted surfactant-mediated extraction	4 mg L ⁻¹	6	106
SWIA	UV-Vis	concretes	NH ₄ ⁺	headspace single-drop micro-extraction	30 μg kg ⁻¹	4	119
ZF	UV-Vis	drugs	partition coefficient	liquid-liquid extraction	-	-	114

UV-Vis – spectrophotometry, FAAS – Flame atomic absorption spectrometry, PM – potentiometry.

Figures caption




Fig. 1. Classification of the flow analytical methods based on the conditions under which the mass transfer processes and chemical reactions take place in the flow mode.




Fig. 2. The dispersion-convection flow methods: (a) CFA, (b) FIA, (c) SIA, (d) SIC




Fig. 3. The dispersion-convection flow methods: (a) MCFIA, (b) SIEMA, (c) MPFS.




Fig. 4. The dispersion-convection flow methods: (a) MSFIA, (b) LFA, (c) CIA.




Fig. 5. The forced-convection flow methods: (a) SFA, (b) ZF, (c) FBA, (d) SIA MC.




Fig. 6. The forced-convection flow methods: (a) SWIA, (b) MCSWIA.

