

Fourier Transform Measurement of NO₂ Absorption Cross-sections in the Visible Range at Room Temperature.

A.C. Vandaele,, C. Hermans, P.C. Simon, M. Van Roozendael
J.M. Guilmot, M. Carleer, and R. Colin

Abstract

New laboratory measurement of NO₂ absorption cross-sections were performed using a Fourier transform spectrometer at 2 and 16 cm⁻¹ (0.03 and 0.26 nm at 400 nm) in the visible range (380-830 nm) and at room temperature. The use of a Fourier transform spectrometer leads to a very accurate wavenumber scale (0.005 cm⁻¹, 8x10⁻⁵ nm at 400 nm). The uncertainty on the new measurements is better than 4%. Absolute and differential cross-sections are compared with published data, giving an agreement ranging from 2% to 5% for the absolute values. The discrepancies in the differential cross-section can however reach 18%. The influence of the cross-sections on the ground-based measurement of the stratospheric NO₂ total amount is also investigated.

Key words

Fourier Transform Spectroscopy, NO₂, absorption cross-sections, differential absorption cross-sections, visible, stratospheric and tropospheric measurements

Introduction

Nitrogen dioxide plays an important role in the chemistry of the troposphere and the stratosphere. It is produced from the oxidation of NO in the troposphere, where it acts as the main source of tropospheric ozone, and is a precursor to species, such as nitric acid, which play a role in the acidification of the environment. Its role in stratospheric photochemistry has been pointed out by Crutzen (1970). Beside its catalytic interaction in the control of ozone, it regulates the amounts of ClO, which in turn controls the ozone loss due to the chlorine catalytic cycle, and of ClONO₂, which is an important stratospheric reservoir of chlorine. It plays thus an important role in the coupling of the NO_x and ClO_x families.

Molecular absorption in the UV-Visible region has been widely used to measure the concentrations of gases in the atmosphere, either in the troposphere or in the stratosphere. The instruments used range from ground-based spectrometers measuring tropospheric or stratospheric concentration by the Differential Optical Absorption Spectroscopy (DOAS) technique (see for example Platt and Perner, 1980, Solomon et al., 1987, Edner et al., 1993, Vandaele et al.,1992, Evangelisti et al., 1995, Camy-Peyret et al., 1996), to spectrometers on board satellites such as the GOME (Global Ozone Monitoring Experiment) launched in April 1995 on board ERS-2 satellite, and to SCIAMACHY and GOMOS instruments to be launched in 1999 on board ENVISAT-1.



All these instruments require absorption cross-sections of the observed molecules, measured at a resolution of 0.02 nm or better and with an accuracy better than 5% (Chance et al, 1990). Accurate cross-sections are also needed for the chemical-dynamical-radioactive modeling of the atmosphere.

The measurement of the NO_2 absorption cross-section is complicated by the presence of its dimer N_2O_4 . Several studies have attempted to measure NO_2 cross-sections. Hall and Blacet(1952) measured absorption spectra of NO_2 - N_2O_4 mixtures at 298K and deduced the contribution of N_2O_4 . Johnston and Graham (1974) measured NO_2 cross-sections in the 185-420 nm spectral region at room temperature. Bass et al.(1976) investigated the 185-410 nm range at 298 K. They corrected their data for the presence of N_2O_4 . Leroy et al.(1987) reported values from 427 to 450 at 298K. Schneider et al.(1987) obtained NO_2 absorption cross-sections between 200 nm and 700 nm, at 298 K and determined the absorption cross-sections of N_2O_4 between 200 and 255 nm. Koffend et al.(1987) used a pulsed dye laser to perform high resolution measurement of NO_2 absorption structures in the 392-395 nm and 411-414 nm. Davidson et al.(1988) investigated the dependence of the NO_2 cross-sections on temperature and the influence of this dependence on the determination of the photolysis rate of NO_2 in the atmosphere. Harwood and Jones (1994) studied the temperature dependence of the ultraviolet-visible absorption cross section of NO_2 . The cross sections of N_2O_4 . were also derived by the latter, as well as new values for the equilibrium constant. Mérienne et al.(1995) measured NO_2 absorption cross-sections in the 300-500 nm region at 293K. The use of an absorption path length of 61m allowed them to work at very low pressure (<0.04 Torr) and to minimize the influence of the dimer. Discrepancies of the order of 20% or sometimes more are found between all the measurements.

This work presents new absolute absorption cross-sections of NO_2 between 12000 and 26000 cm^{-1} (380-830 nm) at 294K. The absorption cross-sections have been measured at two resolutions : 2 and 16 cm^{-1} (0.03 and 0.26 nm at 400 nm). Cross-sections obtained in this work have been measured with a Fourier Transform Spectrometer (FTS), which combines the advantages of a great sensitivity and a built-in wavenumber calibration. This last advantage is achieved by the presence of a He-Ne laser, which allows the interferogram to be digitised at equal intervals. This leads to a highly accurate and reproducible wavenumber calibration. With the Fourier Transform Spectrometer described below, an accuracy of about 0.005 cm^{-1} (8×10^{-5} nm at 400 nm) on the wavenumbers is achieved.



These new cross-sections will then be compared with published data sets, measured at similar resolutions. A comparison of cross-sections smoothed to a resolution of 1.0 nm will also be shown. Finally, the importance to dispose of cross-sections of high quality will be stressed, by investigating their influence on ground-based stratospheric NO₂ measurements in the visible range.

Experimental details

The experimental set-up consists of a light source, the spectrometer, and the gas handling system. The choice of the light source, beam splitter and detector is determined by the spectral region to be investigated.

Combinations of a high pressure O₃ free Xenon source (USHIO, 450 W), a Tungsten filament lamp, a Quartz Suprasil beam splitter with a Si diode or a UV vacuum solar blind diode have been used in order to cover the entire spectral region from 12000 to 26000 cm⁻¹. A blue filter provided by BRUKER or a copper sulphate solution filter have been used. The experimental conditions are reported in Table 1. A photo feedback system allowed the lamp intensity to be stabilised within 1% throughout the experiments.

The absorption spectra were recorded using either a BRUKER IFS120HR or a BRUKER 120M Fourier Transform Spectrometer. The theoretical resolving power attainable with these spectrometers is 106. However, the resolution is essentially limited in the UV by the quality of the mirrors and beam splitter, leading to a maximum resolution of the order of 10-2 cm⁻¹ in this region.

Following some recommendations concerning the measurements of absorption cross-sections (Chance, 1990), spectra have been recorded with resolutions of 2 and 16 cm⁻¹, corresponding respectively to 0.03 and 0.26 nm at 400 nm.

NO₂ gas (UCAR, stated purity of 99.5%) used without further purification, was introduced in a 5.15 cm absorption cell located in the sample compartment of the Fourier Spectrometer. The partial pressure of the gas was then monitored with a 100 Torr full scale Baratron gauge. We waited until the stabilisation of the partial pressure (10 to 30 min, without any lamp illumination) before filling with oxygen in order to obtain a total pressure of 1 atm. The presence of oxygen is of great importance as it induces the reconversion of any NO into NO₂, maintaining the NO₂ concentration constant. NO can be present in the cell as an initial impurity or produced by the photo dissociation of NO₂ during the experiment. The cell has been described in an earlier publication (Hurtmans et al., 1993). Its characteristics will be briefly summarised here : the cell is made of anodised aluminum and has quartz windows; the temperature of the gas is monitored inside the cell with a temperature transducer characterised by an accuracy of 0.2K in the temperature range used in the present work.



The temperature in the cell was stabilised with a liquid circulating around the cell. Finally, the whole experimental set-up was located in a temperature stabilised room.

Absorption of NO_2 on the inner surface of the cell is inevitable. To limit this effect on the pressure measurement, we waited before and after filling the cell with oxygen. This is however not enough to guarantee a stable NO_2 partial pressure throughout the experiment. Each spectrum was therefore the average of a number of scans (see Table I) which have been recorded by sequential blocks of either 256 or 1024 scans. Only the blocks of scans whose absorption did not differ from the first block by more than 1% were retained. We thus checked that the gradual decrease of the NO_2 pressure inside the cell was small. Moreover it proved that photochemical degradation of the sample due to the irradiation by the lamp was not important during the experiment.

Spectra have been obtained using a double sided recording mode, during the forward movement only of the mobile mirror and no apodization function was used. Blank spectra, i.e. with an empty cell, were recorded before and after each measurement.

Determination of the absorption cross-sections

Absorption cross-sections $s(l)$ are derived from the experimental data using the Beer-Lambert law:

$$I(l) = I_0(l) \exp(-n l s(l))$$

where n is the gas concentration in the cell,

l the absorption path length,

$I_0(l)$ and $I(l)$ the intensities of the signal with an empty cell and a filled cell.

The $\text{NO}_2 \rightleftharpoons \text{N}_2\text{O}_4$ equilibrium implies that N_2O_4 is always present in the cell. Absorption of N_2O_4 occurs at wavenumbers greater than 25000 cm^{-1} (Hall and Blacet, 1952, Schneider et al., 1987), but not in the spectral region investigated in this work. N_2O_4 should not therefore interfere with the absorption structures of NO_2 studied here.



However the presence of N_2O_4 must be considered when determining the partial pressure of NO_2 . This was done by considering the following equations :

$$P_t = P_{N_2O_4} + P_{NO_2}$$

where P_t is the total partial pressure in the cell and the partial pressures $P_{N_2O_4}$ and P_{NO_2} are related through the equilibrium constant K_p :

$$K_p = \frac{P_{NO_2}^2}{P_{N_2O_4}}$$

where K_p is the equilibrium constant.

Partial pressures of NO_2 and N_2O_4 can thus be calculated. Hurtmans et al. (1993) reviewed the values of the equilibrium constant found in the literature and determined an empirical relation for its temperature dependence. For temperatures ranging from 233 K to 403 K, the dependence was expressed by a fourth degree polynomial expansion. At 294.15K, the value of the constant K_p is 105.72 hPa. Using the above equations the partial pressure of NO_2 can be calculated from the various total pressures used.

The NO_2 absorption cross-sections at the resolution of 16 cm^{-1} were obtained by taking the mean value of all the spectra i.e. all the spectra taken at 16 cm^{-1} and the spectrum at 2 cm^{-1} degraded to 16 cm^{-1} . All measurements were found to agree within 2% which is lower than the uncertainty stated in Table II. The results are plotted in Figure 1 for the entire spectral region. Figure 2 shows a detailed region from 22200 to 23200 cm^{-1} of the spectrum at the resolution of 2 cm^{-1} .

Discussion

Error evaluation

The error budget on the absorption cross-sections was carefully evaluated, taking into account the errors on the pressure and temperature measurements, on the reaction constant K_p , the uncertainty on the absorption path length, the presence of possible impurities in the samples, the adsorption of NO_2 taking place inside the cell and the absorbance accuracy.



According to the error propagation principle, the uncertainty on the cross-sections can be derived from the following relation :

$$\frac{\delta\sigma}{\sigma} = \sqrt{\left(\frac{\delta A}{A}\right)^2 + \left(\frac{\delta T}{T}\right)^2 + \left(\frac{\delta l}{l}\right)^2 + \left(\frac{\delta P_{\text{NO}_2}}{P_{\text{NO}_2}}\right)^2}$$

where A is the absorbance ($A = \ln \frac{I_0}{I}$). From the definition of the partial

pressure of NO₂, δP_{NO_2} can be written as :

$$\delta P_{\text{NO}_2} = \left(\frac{\partial P_{\text{NO}_2}}{\partial K_P}\right)_{P_t} \delta K_P + \left(\frac{\partial P_{\text{NO}_2}}{\partial P_t}\right)_{K_P} \delta P_t$$

Setting , $a = (1 + 4P_t / K_P)^{1/2}$ one obtains

$$\frac{\delta P_{\text{NO}_2}}{P_{\text{NO}_2}} = \left(\frac{a-1}{2a}\right) \frac{\delta K_P}{K_P} + \left(\frac{4P_t}{2a(a-1)K_P}\right) \frac{\delta P_t}{P_t}$$

Concerning dK_P , Roscoe and Hind (1993) reviewed different measurements of K_P and came to the conclusion that the error on this constant was of the order of 5% for temperatures ranging from 233 to 300 K. The error on the total pressure arises from the presence of impurities inside the NO₂ bottle, from the adsorption of NO₂ and from the measurement itself (0.1% with the Baratron gauge). The manufacturer states a 99.5% purity NO₂. An infrared study of the sample

(Hurtmans et al. ,1993; Vander Auwera, personal communication) showed that small amounts of N₂O and CO₂ were present in the sample, as well as negligible amounts of HNO₂ and HNO₃. These species do not interfere with the NO₂ absorption in the visible region but cause an error on the determination of its partial pressure. NO is also present as an impurity in the sample, but the presence of oxygen ensures its conversion into NO₂. An estimated value of 2% is believed to cover the uncertainties due to all impurities. As already explained, the adsorption of NO₂ on the inner walls of the cell causes a decrease of absorption estimated at 1% throughout an experiment. The error on the absorbance for each individual spectrum was less than 1%.

The cumulative error is given in Table II. According to these figures, the accuracy on the present cross-section measurement of NO₂ can be stated to be better than 4%.



Comparison with published measurements

NO₂ absorption cross-sections will be compared below with the data from Schneider et al.(1987), Johnston and Graham (1974), Harwood and Jones(1994), and Mérienne et al. (1995), taking into account the conversion of wavelengths to wavenumbers and the correction of wavelengths from air to vacuum when necessary.

Comparison of data at high resolution

Data from Schneider et al.(298 K, resolution of 0.04 nm), Johnston and Graham (294 K, resolution of 0.1 nm), and Mérienne et al. (293 K, resolution of 0.01 nm) have been convolved in order to obtain a resolution equivalent to 16 cm⁻¹ and have been compared to our measurement at this resolution. The cross-sections were smoothed using a Gaussian function, whose width was given by the desired resolution. Results are plotted in Figures 3 and 4. Figure 3 shows the absolute absorption cross-sections of NO₂ from the four groups of authors in the spectral region from 415 and 465 nm. Figure 4 presents the relative difference (expressed in %) between our data and those of Johnston and Graham and of Mérienne et al. The results appear to be « noisy ». It must be remembered however that the necessary conversion of the wavelengths to wavenumbers may introduce artefacts, which could explain some of the noise found in the residuals. Moreover, it is almost impossible to convolve a spectrum so that its new resolution perfectly matches the resolution of another spectrum. This produces yet additional noise in the residual. But the most important source of noise comes from errors in the calibration of the wavelength scale. The wavenumber scale produced by the FTS is linear and absolute, and that wavelength scales produced by grating spectrometers are not and may be shifted or distorted. Moreover, as grating spectrometers usually work on smaller spectral regions, large errors may be introduced when joining two scanned regions. All these errors on the wavelength scale account for most of the noise found in the residuals.

The difference between our data and those of Schneider et al. is considerable (10% to 25%) and greatly varies with wavelength. Data from Schneider et al. are systematically lower in the 380-426 nm region, as expected if photo dissociation of NO₂ was taking place in their experiment. Moreover, a large discrepancy in the wavelength scale is observed in some parts of the spectrum. It seems that errors were introduced in the data set of Schneider et al. when merging two contiguous spectral regions.



The comparison with the results of Johnston and Graham shows that values found in our work are systematically 5% lower between 415 and 500 nm. For wavelengths higher than 525 nm, data from Johnston and Graham are lower by 5%. Attention should be paid to the fact that a wavelength shift of the complete data set of Johnston and Graham of 0.25 nm has been applied in order to make the absorption structures coincide.

The comparison of our data with the data of Mérienne et al. shows a very good agreement (less than 2%). This result is noteworthy as it must be pointed out that the two sets of data were obtained with completely different spectrometers and instrumental set-ups. Moreover a comparison of the two sets at the higher resolution (2 cm^{-1}) has been made, showing (Figure 5) again a difference of only 2% between the two sets of data.

Comparison of data at low resolution

In order to compare the NO_2 absorption cross-sections obtained in this work with the data from Harwood and Jones (1994) (298 K, resolution of 0.54 nm), all data sets have been convolved to a chosen resolution of 1.0 nm using a Gaussian function (see 4.2.1). The resolution of most of DOAS grating instruments lies usually in the range between 0.5 and 1.2 nm; a resolution of 1.0 nm is therefore representative for that kind of instruments.

Convolved absolute cross-sections of Harwood and Jones are systematically lower by 5% than the convolved data measured in this work. The differences reach however 15% in the 430 to 440 nm region.

As the DOAS technique is based on the differential cross-sections of the absorbing molecules rather than their absolute cross-sections, differential cross-sections derived from the measurements of the various groups were also compared. Differential cross-sections are obtained by smoothing the absolute cross-sections to remove all the absorption structures from the spectrum, and by subtracting this smoothed curve from the absolute cross-section. The data have been smoothed with a running average over a number of points corresponding to a spectral interval of 25 nm.

The comparison has been limited to the 410-460 nm region, where the DOAS stratospheric NO_2 measurements are usually performed (Hofmann et al., 1995). Results are reported in Figure 6, where differential cross-sections and their differences relative to our data are presented. Shifts in the wavelength scales in the data of Schneider et al. and of Johnston and Graham have not been corrected. The good agreement between the data of this work and those of Mérienne et al. is again confirmed.



Once the shift in the data of Johnston and Graham is taken into account, their differential absorption structures appear to agree very well (less than 2%). A shift in the data of Schneider et al. clearly visible above 425 nm could not be corrected as it is not constant over the entire range; differences up to 70% are therefore present.

Table III shows more clearly the differences existing between the various data sets. Each differential cross-section has been compared to the differential cross-sections of this work. Shifts in the wavelength scale and correction factor of the amplitude were thus determined. The fitted region ranged from 410 to 460 nm when possible. For values of Schneider et al., two regions were chosen, as their shift is highly dependent on the wavelength region. In the 410-420 nm region they do not present any significant shift compared to the data of Mérienne et al. In the 450-460 nm region, a shift of 0.58 nm relative to our data is present. However, once this shift has been taken into account, differential cross-sections agree well. Altogether differential cross-sections agree within 1.5% in amplitude except for the data of Harwood and Jones.

Influence of the cross-sections on stratospheric NO₂ measurements

The measurements of the NO₂ concentration in the atmosphere, either the stratosphere or the troposphere, mostly rely today on the use of the spectroscopic signature of the molecule. The NO₂ amount is usually determined using the intense absorption structures located between 400 and 600 nm, but its spectrum must be taken into account on the whole UV/visible range in order to eliminate its interference with the absorption of other molecules.

The need for NO₂ cross-section values of high quality has recently been illustrated in a NDSC (Network for the Detection of Stratospheric Change) campaign, focused on the measurements of stratospheric NO₂ using various UV/Visible spectrometers (Hofmann et al., 1995). Seven instruments participated in the campaign. Each instrument worked with a different NO₂ cross-section more suited to its instrumental function. It was decided to calibrate all the instruments by the use of NO₂ cell measurements. The participants thus derived a correction factor lying between 0.96 to 1.08 to apply on their NO₂ data. This means that based on the choice of the NO₂ absorption cross-sections, a discrepancy of 12% on the NO₂ concentration was avoided.

The Institute for Space Aeronomy participated in this campaign with a ground based instrument developed for stratospheric studies. This instrument has already been described in a previous paper (Van Roozendaal et al., 1994) and will only be briefly summarised here.



The instrument records spectra of the zenith scattered sunlight and is optimised for the detection of NO_2 and O_3 in the visible. The detector used is an EG&G Reticon photodiode array (1024 pixels). The spectral range extends from 400 to 560 nm, with a resolution of 1.0 nm corresponding to a sampling ratio of 6.6 pixels/FWHM. The software permits a shift and a stretch of the wavelength scale of the cross-section with respect to the analysed spectrum.

Zenith sky spectra recorded at Harestua (Norway) on day number 218, 1995 have been analysed with the different absorption cross-sections mentioned above, convolved with the instrumental function of the instrument. The relative differences in NO_2 column abundance are plotted in Figure 7a. Column abundances calculated with our cross-sections have been chosen as references. The scatter of points at lower NO_2 slant amount is due to the increase of the noise relative to the NO_2 differential signature. Results calculated with the differential cross-sections obtained in this work, those of Johnston and Graham, and those of Mérienne et al. do agree well (within less than 1%). Results derived from Harwood and Jones are about 18% higher. These results mainly reproduce the differences observed between the differential cross-sections themselves. More interesting are the results plotted in Figure 7b, where the standard deviations on the NO_2 slant amount have been represented as a function of the NO_2 slant column. This is representative of how well the absorption cross-sections compare with the absorption signatures found in the atmospheric spectra. Except at high NO_2 slant amounts, corresponding to very high zenith angles and thus very low signal to noise ratio, the standard deviation should be independent of the NO_2 slant amount if no instrumental factor interferes. The fact that the evolution is linear reflects the crucial effect of the shape of the differential structures of the absorption cross-sections. Cross-sections of this work, Mérienne et al. and Johnston and Graham lead to similar standard deviations, although cross-sections of Harwood and Jones lead to a higher dependence in the NO_2 amount.

We have also used cross-sections measured at lower temperatures by Coquart et al.(1995) to determine the NO_2 amount. It should be pointed out that the spectrum of NO_2 is composed of a series of narrow structures, whose width are of the order of or lower than the resolutions used for the laboratory data. Moreover their shape varies with temperature. This effect is of great importance when monitoring the stratospheric NO_2 , as temperature in the stratosphere can be as low as 190 K. Residuals of the fitting procedure applied to derive the NO_2 amount were found to be correlated with the temperature dependence of the cross-sections (Hoffman et al., 1995).



The cross-sections of Coquart et al. (1995) measured at 220 K and 240 K have also been used to measure the NO₂ amount in the spectra recorded at Harestua, during day number 218. Results are plotted in Figure 7. The NO₂ slant columns are 14% and 20% lower when using absorption cross-sections at 240K and 220K respectively, but the standard deviation on these results are lower. This lower error level is consistent with the observation of an absorption arising from NO₂ situated in the stratosphere at altitudes where the temperature is of the order of 220-240K. The evolutions of the standard deviation with the NO₂ slant column at the two low temperatures are very similar. The residual error level attained is believed to result mainly from errors due to the convolution process or the variation of the resolution in the spectral range studied. These results confirm the necessity to take into account the temperature dependence of the absorption cross-sections. This dependence affects not only the amplitude of the differential structures of the cross-sections, but also their shape.

Conclusions

NO₂ absolute absorption cross-sections have been measured at room temperature at resolutions of 2 and 16 cm⁻¹ (0.03 nm and 0.26 nm at 400 nm respectively). The accuracy on these new measurements is believed to be better than 4%, with an accuracy on the wavenumber scale of 0.005 cm⁻¹ (8x10⁻⁵ nm at 400 nm). Absolute and differential cross-sections have been compared to other data sets generally used in atmospheric measurements. Our data show good agreement with data of Mérienne et al. (1995) (better than 2%). Data from Johnston and Graham (1974) present a difference of 5% in the absolute cross-sections, however the differential cross-sections are in better agreement. The wavelength scale of their data shows a systematic shift compared to the scale of the Fourier transform spectrometer. Data from Schneider et al. (1987) show large discrepancies in the wavelength scale. Absolute cross-sections of Harwood and Jones (1994) are systematically 5% lower, and the differential cross-sections are 18% higher.

These cross-sections and those measured at lower temperatures by Coquart et al. (1995) have been convolved at a resolution of 1.0 nm and used to derive NO₂ slant amounts measured with a DOAS stratospheric instrument. The results of this comparison show the good quality of the new measurements but demonstrates also the need to take into account the temperature dependence of the cross-sections. New measurements of this temperature dependence will be undertaken soon.

All the data obtained during this work are available on floppy disks, either on a wavenumber or a wavelength scale.



Acknowledgements

This project has been supported by the Belgian State - Prime Minister's Service - Science Policy Office (contracts GC/35/002, E7/35/11P, and WI/35/004) and the "Fonds National de la Recherche Scientifique".

References

Amoruso A., Crescentini, L., Fiocco, G., and M. Volpe, 1993, New measurements of the NO₂ absorption cross-section in the 440- to 460-nm region and estimates of the NO₂ -N₂O₄ equilibrium constant, J. Geophys. Res. 98, 16857-16863

Bass, A.M., Ledford, A.E. and Laufer, A.H., 1976 , Extinction coefficients of NO₂ and N₂O₄, J. Res. Nat. Bur. Stand. A 80, 143-166

Camy-Peyret, C., Bergquist, B., Galle, B., Carleer, M, Clerbaux, C., Colin, R., Fayt, C., Goutail, F., Nunes-Pinharanda, M., Pommereau, J.P., Hausmann, M., Platt, U., Pundt, I., Rudolph, T., Hermans, C., Simon, P.C., Vandaele, A.C., Plane, J., and Smith, N., 1996, Intercomparison of Instruments for Tropospheric Measurements using Differential Optical Absorption Spectroscopy, J. Atm. Chem. 23, 51-80

Chance, K.V., Burrows, J.P., Meller, R., Moortgat, G.K., Perner, D., and Schneider, W., 1990, Database needs for UV/Visible atmospheric spectroscopy, in : Atmospheric Spectroscopy Applications Workshop, ed. A. Barbe, A. Ponomarev, N. Yu, and R. Zander, 186-194, Moscow

Coquart B., Jenouvrier, A., and Mérienne, M.F., 1995, The NO₂ absorption spectrum. II Absorption cross-sections at low temperatures in the 400-500 nm region, J. Atmos. Chem. 21, 251-261

Crutzen, P., 1970, The influence of nitrogen oxide on the atmospheric ozone content, Q. J. R. Meteorol. Soc. 96, 320

Davidson, J.A., Cantrell, C.A., McDaniel, A.H., Shifter, R.E., Madronich, S., and Calvert, J.G., 1988, Visible-ultraviolet absorption cross-sections for NO₂ as a function of temperature, J. Geophys. Res. 93, 7105-7112

Edner, H., Ragnarson, P., Spännare, S., and Svanberg, S., 1993, Differential optical absorption spectroscopy (DOAS) system for urban atmospheric pollution monitoring, Appl. Optics 32, 327-332

Evangelisti, F., Baroncelli, A., Bonasoni, P., Giovanelli, G., and Ravegnani, F., 1995, Differential optical absorption spectrometer for measurement of tropospheric pollutants, Appl. Optics 34, 2737-274



Hall, Jr. T.C., and Blacet, F.E., 1952, Separation of the absorption spectra of NO₂ and N₂O₄ in the range of 2400-5000Å, J. Chem. Phys. 20, 1745-1749

Harwood, M.H. and Jones, R.L., 1994, Temperature dependent ultraviolet-visible absorption cross-sections of NO₂ and N₂O₄ : low-temperature measurements of the equilibrium constant for 2NO₂ « N₂O₄, J. Geophys. Res. 99, 22955-22964

Hofmann, D.J., Bonasoni, P., De Mazière, M., Evangelisti, F., Francois, P., Giovanelli, G., Goldman, A., Goutail, F., Harder, J., Jakoubek, R., Johnston, P., Kerr, J., McElroy, T., McKenzie, R., Mount, G., Pommereau, J.-P., Simon, P., Solomon, S., Stutz, J., Thomas, A., Van Roozendael, M., and Wu, E., 1995, Intercomparison of UV/Visible Spectrometers for Measurements of Stratospheric NO₂ for the Network for the Detection of Stratospheric Change, J. Geophys. Res. 100, 16765-16791

Hurtmans, D., Herman, M., and Vander Auwera, J., 1993, Integrated band intensities in N₂O₄ in the infrared range, J. Quant. Spectrosc. Transfer 50, 595-602

Johnston, H.S. and Graham, R., 1974, Photochemistry of NO_x and HNO_x compounds, Can. J. Chem. 52, 1415-1423

Koffend, J.B., Holloway, J.S, Kwok, M.A., and Heider, R.F., 1987, High resolution absorption spectroscopy of NO₂, J. Quant. Spectrosc. Transfer 37, 449-453

Leroy, B., Rigaud, P. and Hicks, E., 1987, Visible absorption cross-sections of NO₂ at 298K and 235K, Annales Geoph. 5A, 247-250

Mérienne, M.F., Jenouvrier, A. and Coquart, B., 1995, The NO₂ absorption spectrum. I: Absorption cross-sections at ambient temperature in the 300-500 nm region, J. Atmos. Chem. 20, 281-297

Platt, U. and Perner, D., 1980, Direct measurements of atmospheric CH₂O, HNO₂, O₃, NO₂, and SO₂ by differential optical absorption in the near UV, J. Geophys. Res. 85, 7453-7458

Roscoe, H.K., and Hind, A.K., 1993, The equilibrium constant of NO₂ with N₂O₄ and the temperature dependence of the visible spectrum of NO₂ : a critical review and the implications for measurements of NO₂ in the polar stratosphere, J. Atmos. Chem. 16, 257-276



Schneider, W., Moortgat, G.K., Tyndall, G.S. , and Burrows, J.P., 1987, Absorption cross-sections of NO₂ in the UV and Visible region (200 - 700 nm) at 298K, J. Photochem. and Photobiol., A : Chem. 40, 195-217

Solomon S., Schmeltekoff A.L., and Sanders R.W., On the interpretation of zenith sky absorption measurements, J. Geophys. Res. 92, 8311 (1987)

Vandaele, A.C., Carleer, M., Colin, R., and Simon, P.C., 1992, Detection of urban O₃, NO₂, H₂CO and SO₂ using Fourier Transform Spectroscopy, in :Optical Methods in Atmospheric Chemistry,ed. H. Schiff and U. Platt, SPIE Proc. vol 1715, The International Society for Optical Engineering, Bellingham, Washington, 288-292

Vandaele, A.C., Simon, P.C., Guilmot, J.M., Carleer, M., and Colin, R., 1994, SO₂ Absorption Cross-section Measurement in the UV using a Fourier Transform Spectrometer, J. Geophys. Res. 99, 25599-25605

Van Roozendaal, M., Fayt, C., Bolsée, D., Simon, P.C., Gil, M., Yela, M. , and Cacho, J., 1984, Ground-based stratospheric NO₂ monitoring at Keflavik (Iceland) during EASOE, Geophys. Res. Lett. 21, 1379-1382



Table I : Experimental conditions

Res	Pt	Nb of scans	Source1	Filter2	Detector	Spectral Region
cm-1	hPa					cm-1
16	10.18	8 x 256	W	b	Si-diode	9500-26000 (380-1050 nm)
16	29.50	14 x 256	W	b	Si-diode	9500-26000
16	29.90	4 x 1024	W	b	Si-diode	9500-26000
16	50.50	4 x 1024	W	b	Si-diode	9500-26000
16	67.70	3 x 1024	W	b	Si-diode	9500-26000
16	10.70	4 x 1024	W	Cu ⁺⁺	Si-diode	15000-30000 (330-670 nm)
16	15.18	4 x 1024	W	Cu ⁺⁺	Si-diode	15000-30000
16	20.02	4 x 1024	W	Cu ⁺⁺	Si-diode	15000-30000
16	30.27	4 x 1024	W	Cu ⁺⁺	Si-diode	15000-30000
16	39.30	4 x 1024	W	Cu ⁺⁺	Si-diode	15000-30000
16	48.65	4 x 1024	W	Cu ⁺⁺	Si-diode	15000-30000
2	13.85	4 x 1024	X	b	UV-diode	15000-30000
1W=Tungsten filament lamp, X=High Pressure Xenon lamp						
2b=Blue filter, Cu ⁺⁺ = Copper solution						

Table II : Estimate of the accuracy of cross-sections measurements

	Error
	(%)
Sample impurities	2
Adsorption inside the cell	1
Absorbance on individual spectra	1
Absorption path length	0.5
Pressure	0.1
Temperature	0.1
Cumulative error	< 4



Table III : Comparison of the differential cross-sections

	Fit Range (nm)	Shift (nm)	Correction Factor
This work		0	1
Mérienne <i>et al.</i>	410-460	0.03	1.003
Harwood and Jones	410-460	0.04	1.160
Johnston and Graham	415-460	0.24	1.002
Schneider <i>et al.</i>	410-420	0.03	1.017
Schneider <i>et al.</i>	450-460	0.58	1.018

Figure captions

Figure 1. Absolute absorption cross-sections of NO₂ at the resolution of 1 cm⁻¹ at room temperature.

Figure 2. Absolute absorption cross-sections of NO₂ at the resolution of 2 cm⁻¹ at room temperature from 22200 to 23200 cm⁻¹.

Figure 3. Comparison of absolute cross-sections of this work ($\frac{3}{4}$), Mérienne *et al.*(1995) (x x x x x), Johnston and Graham (1974) (- x -), and Schneider *et al.* (1987) (- -). All data from the literature have been convolved to correspond to the resolution of 16 cm⁻¹. The wavenumber scale of Johnston and Graham has been shifted by 5 cm⁻¹. No shift in the wave number scale of Schneider *et al.* has been applied as it is not constant on the entire spectral range.

Figure 4. Relative differences between the absolute cross-sections of Mérienne *et al.*(1995) ($\frac{3}{4}$), Johnston and Graham (1974) (x x x x x), convolved to 16 cm⁻¹, and the cross-sections of this work. The wavenumber scale of Johnston and Graham has been shifted by 5 cm⁻¹. The mean value and the 1s deviation for the data of Mérienne *et al.* are also presented.

Figure 5. Relative differences between the absolute cross-sections of Mérienne *et al.*(1995) convolved to 2 cm⁻¹, and the cross-sections of this work. The mean value and the 1s deviation are also presented.



Figure 6. (a) Comparison of the NO₂ differential cross-sections in the 410-460 nm spectral range : (¾) this work, (x x x x x) Mérienne et al(1995), (- -) Johnston and Graham (1974), (- x -) Schneider et al. (1987), and (- x x -) Harwood and Jones (1994). Shifts in the wavelength scales of Johnston and Graham and Schneider et al. have not been corrected. (b) Differences in the NO₂ differential cross-sections with respect to the data of this work.

Figure 7. Influence of the absorption cross-section on the determination of stratospheric NO₂ column abundances. Results shown correspond to zenith sky measurements made at Harestua (Norway) on julian day 218, 1995 using the cross-sections of this work (I), Mérienne et al(1995) (m), Johnston and Graham (1974) (Š), Harwood and Jones (1994, data at 298 K) (Ñ), and Coquart et al. (1995) at 240 K (à) and 220 K (D). The differences in NO₂ column amount relative to the NO₂ amounts found with the cross-sections of this work are plotted in Figure 7.a, and the standards deviations on the measurements in Figure 7.b

Figure 1.




Figure 2.




Figure 3.




Figure 4.




Figure 5.




Figure 6.




Figure 7.

