

ARTICLE

Open Access

Full noncontact laser ultrasound: first human data

Xiang Zhang^{1,2}, Jonathan R. Fincke^{1,2}, Charles M. Wynn³, Matt R. Johnson³, Robert W. Haupt³ and Brian W. Anthony^{1,2}

Abstract

Full noncontact laser ultrasound (LUS) imaging has several distinct advantages over current medical ultrasound (US) technologies: elimination of the coupling mediums (gel/water), operator-independent image quality, improved repeatability, and volumetric imaging. Current light-based ultrasound utilizing tissue-penetrating photoacoustics (PA) generally uses traditional piezoelectric transducers in contact with the imaged tissue or carries an optical fiber detector close to the imaging site. Unlike PA, the LUS design presented here minimizes the optical penetration and specifically restricts optical-to-acoustic energy transduction at the tissue surface, maximizing the generated acoustic source amplitude. With an appropriate optical design and interferometry, any exposed tissue surfaces can become viable acoustic sources and detectors. LUS operates analogously to conventional ultrasound but uses light instead of piezoelectric elements. Here, we present full noncontact LUS results, imaging targets at ~5 cm depths and at a meter-scale standoff from the target surface. Experimental results demonstrating volumetric imaging and the first LUS images on humans are presented, all at eye- and skin-safe optical exposure levels. The progression of LUS imaging from tissue-mimicking phantoms, to excised animal tissue, to humans in vivo is shown, with validation from conventional ultrasound images. The LUS system design insights and results presented here inspire further LUS development and are a significant step toward the clinical implementation of LUS.

Introduction

Modern ultrasonography is well established in diagnostic and interventional medical imaging and is the most commonly used imaging modality for soft tissue¹. Ultrasound has advantages compared to other imaging methods, including being nonionizing, relatively low cost, and portable. Current embodiments of ultrasound technology range from cart-based bedside systems to portable handheld devices¹. Conventional ultrasound imaging requires the placement of piezoelectric transducers in contact with the patient to transmit and then detect reflected and scattered acoustic waves at the body surface. Compared to other imaging modalities, patient contact is a source of

variability unique to ultrasound. The imaging clinician applies variable contact forces on the ultrasound probe to the tissue, and the resultant tissue compression causes contact-sensitive images; quantitative ultrasound imaging methods such as shear wave elastography have been shown to be directly sensitive to tissue compression^{2–4}. Other contact-sensitive applications, such as remote patient/neonatal monitoring, tracking of wound healing, and imaging of sensitive skin areas, could significantly benefit from a noncontact ultrasound system.

In addition, the freehand reference frame and the cross-sectional or slice nature of conventional ultrasound imaging generates orientation-sensitive images. These two sources of image variability generally complicate longitudinal tracking (monitoring over time) of the tissue morphology using ultrasound. In comparison, MRI and CT have gantry fixed reference frames and generate volumetric images without patient contact. However, frequent imaging for continuous patient monitoring using

Correspondence: Xiang Zhang (xzhang88@mit.edu) or Brian W. Anthony (banthony@mit.edu)

¹Department of Mechanical Engineering, Massachusetts Institute of Technology, 77 Massachusetts Ave., Cambridge, MA 02139, USA

²Institute for Medical Engineering & Science, Massachusetts Institute of Technology, 45 Carleton St., Cambridge, MA 02142, USA

Full list of author information is available at the end of the article.

These authors contributed equally: Robert W. Haupt, Brian W. Anthony

© The Author(s) 2019



Open Access This article is licensed under a Creative Commons Attribution 4.0 International License, which permits use, sharing, adaptation, distribution and reproduction in any medium or format, as long as you give appropriate credit to the original author(s) and the source, provide a link to the Creative Commons license, and indicate if changes were made. The images or other third party material in this article are included in the article's Creative Commons license, unless indicated otherwise in a credit line to the material. If material is not included in the article's Creative Commons license and your intended use is not permitted by statutory regulation or exceeds the permitted use, you will need to obtain permission directly from the copyright holder. To view a copy of this license, visit <http://creativecommons.org/licenses/by/4.0/>.

MRI or CT is prohibitively expensive or would expose patients to significant ionizing radiation in the case of CT.

A volumetric noncontact ultrasound imaging method could resolve many existing limitations and extend ultrasound imaging to broader applications. As presented in this article, full noncontact laser ultrasound (LUS) – employing skin surface photoacoustic sources in combination with laser interferometric detection – generates image features in human studies comparable to conventional ultrasound and could address conventional ultrasound limitations. A fully optical noncontact LUS system is broadly applicable to contact-sensitive imaging applications – elastography, musculoskeletal disease tracking, and imaging of sensitive/painful tissue regions. In addition, LUS may see high usage in critical care and surgical situations where large-area and high-temporal-resolution imaging is often necessary but forgone due to cost, radiation, or inability to safely relocate patients into MRI scanners.

Other approaches making advances toward noncontact and volumetric ultrasound include ultrasound tomography (UST) and photoacoustics (PA) imaging. UST methods typically surround an imaging target with ultrasonic elements within a water tank^{5–10}. UST systems can produce volumetric ultrasound images of human tissue comparable to those of MRI or CT^{5,11}. However, the geometric constraints of water tanks and the inflexibility of large ultrasonic arrays limit UST systems to highly specific applications, such as breast imaging. Unlike UST, PA approaches – utilizing the conversion of optical energy to acoustic energy via thermoelastic expansion of the tissue – offers a pathway toward compression and coupling-agent-free ultrasound imaging^{12,13}. Since the first report of the PA effect over a century ago, modern PA systems are multiwavelength, multipoint, and multi-contrast and have been utilized to image length scales spanning microns to centimeters^{12–21}. Generally, PA uses an optical source for excitation and traditional piezoelectric elements for detection. Pulsed lasers irradiating biological tissue generate acoustic impulses via optical absorption and induce thermoelastic stress and relaxation within the tissue^{22,23}. By tuning the optical wavelength, varying photosensitive absorbers in biological tissue can be selectively imaged²⁰. To increase image quality or imaging depth, optical contrast agents such as nanoparticles or dyes can also be injected^{18,21,24}. Since PA signal generation is target specific, the spatial location of the optical to acoustic conversion point within the tissue can be localized only *a posteriori*. Depending on the irradiating optical beam, time inversion of multiple recorded acoustic signals may be necessary to localize the source position for PA image reconstruction^{25,26}. The PA imaging depth and resolution are object and application specific – dictated by the interaction of the specific light source and the tissue of interest; thus, optical parameters such as

wavelength, power, geometric focus, and repetition rate are critical in PA system design²².

In contrast to most PA systems, LUS uses an optical detector and is fully noncontact. The optical detection of ultrasound offers increased sensitivity, broader bandwidth, more compact packaging, and true noncontact measurements^{27–36}. The LUS technique is broadly used in nondestructive testing (NDT) for remote thickness measurement, fault detection, and material characterization^{27,36–38}. More recently, LUS has been demonstrated on tissue-mimicking phantoms, excised tissue samples, and a chicken chorioallantoic membrane^{39–44}. However, a full optical ultrasound system for *in vivo* human imaging has not been previously demonstrated.

In this article, we report the design and evaluation of an eye- and skin-safe, full noncontact, all-optical LUS imaging system evaluated on humans *in vivo*. Image results on tissue-mimicking phantoms and *ex vivo* animal tissue are also presented. Unlike an optical source in a PA system – maximizing optical penetration into the tissue, the optical source for the reported LUS system minimizes tissue penetration, specifically to convert optical energy to acoustic energy at the tissue surface. Typical PA systems rely on optical windows, where tissue optical absorption is low, to penetrate deeper into tissue and selectively image optical absorbers such as hemoglobin or other injected optical contrast agents¹⁵. Since optical attenuation is two to three orders of magnitude higher than acoustic attenuation in tissue, restricting optical conversion to the tissue surface and interrogating with acoustic propagation is significantly more efficient than optical propagation for deep anatomical imaging. By selecting optical wavelengths with a high optical absorption coefficient, optical penetration is minimized, while heat generation for thermoacoustic conversion is maximized. Furthermore, the spatial location of the generated acoustic source can be localized *a priori*, removing the need for array detection and inversion algorithms for source localization. In combination with an optical detector, the LUS system is noncontact and analogous to conventional probe-based ultrasound imaging except that it uses light. Free-space positioning of the optical source and detector points enables volumetric LUS imaging without costly 2D piezoelectric arrays. Imaging results from the LUS system present the first instance of a laser-based ultrasound system tested on human subjects. Results demonstrating LUS system feasibility and system design insights are reported. The human LUS images are encouraging and will motivate further research toward the clinical implementation of noncontact LUS imaging technology.

Results

Volumetric 3D Phantom Imaging

The LUS system was initially evaluated on tissue-mimicking phantoms and *ex vivo* porcine tissue prior to

experimentation on human volunteers. A 1540 nm pulsed source laser delivers the optical pulses to excite acoustic waves on the tissue surface, and a 1550 nm continuous wave (CW) Mach-Zehnder Laser Doppler Vibrometer (LDV) measures returning acoustic vibrations on the tissue surface. Both lasers are manufacturer-rated to be eye and skin safe. The source laser and LDV were measured to have 2.3 mJ per pulse and 9.8 mW, respectively. The source laser has a 2 mm beam diameter on the subject surface, and the LDV spot was manually focused to maximize optical backscatter. The reported noise equivalent power for the LDV is $<0.5 \mu\text{m/s}/\sqrt{\text{Hz}}$. We specifically selected wavelengths near 1500 nm to limit optical penetration to the tissue surface and maximize the converted source amplitude. Wavelengths near 1500 nm can leverage the high optical absorption of tissue near 1500 nm for efficient source conversion while maintaining eye and skin safety by having the highest maximum permissible exposure (MPE) limits, namely, 1 J/cm^2 and 0.1 W/cm^2 for 1500–1800 nm pulsed and CW lasers, respectively^{45,46}. With the safety limits computed using the limiting aperture diameters from the ANSI standard, the corresponding source and detector irradiances are found to be 0.024 J/cm^2 and 0.1 W/cm^2 , both within the eye and skin MPE limits. No surface enhancements were used for any LUS imaging experiments.

High-water-content gelatin phantoms with various inclusions were constructed to replicate the optical absorption characteristics of biological tissue in the infrared (IR) spectrum. Metallic spheres, rods, disks, and square-shaped inclusions were embedded in the phantoms to evaluate the LUS system imaging capabilities. The optical source and detection spots were mechanically calibrated, controlled, and collocated using steering mirrors and linear stages. Free-space control of the source and detection spots enables both 2D and 3D imaging. The time-series output of the LDV was recorded through a digital oscilloscope connected to the host computer. A simplified LUS system architecture is shown in Fig. 1a.

As shown in Figs. 1b and 2c, the image reconstruction of LUS line scans and raster scans using a synthetic aperture focusing technique (SAFT) algorithm in conjunction with a coherence factor reconstructed both 2D and 3D images of the gelatin phantoms³⁷. The LUS images were compared against a clinical GE Logiq E9 ultrasonic imager with a 9 MHz center frequency linear probe. By inspection, the LUS images and conventional B-mode images show broad feature agreements (Fig. 1b/c). However, the LUS system was also able to generate 3D volumetric images by raster scanning the LUS source and detector spots on the phantom surface (Fig. 2c). Prior LUS research imaged phantoms with embedded objects/tissue but relied on added retroreflective material on the phantom surface to enhance the optical reflectivity for the optical detector^{39,42}.


Ex vivo Animal Imaging

Excised porcine abdominal tissue was imaged in the LUS system. Using porcine tissue as a human analog in biomedical research is well established, including for use in toxicology, immunology, wound healing, and radiation^{47,48}. Porcine skin is similar to human skin in both anatomical structure and optical composition⁴⁹. Porcine abdominal tissue was obtained from the local market without specialized preparation to include natural skin variations. For each sample, the epidermis, dermis, subcutaneous fat, and muscle layers were clearly visible. Similar to phantom imaging, LUS line scans were completed on each sample. Reconstructed LUS images with verification from conventional ultrasound are shown in Fig. 3c, d. No damage or marking was visible on any tissue surface after the LUS experiments.

The reconstructed LUS image confirms that LUS is sensitive to soft-tissue features at eye- and skin-safe optical exposure levels. In the LUS image (Fig. 3c), highly reflective air–tissue interfaces are visible at $\sim 4.5 \text{ cm}$, while weakly reflective soft-tissue boundaries such as skin-fat and fat-muscle interfaces are also present. The subcutaneous fat layer and multiple muscle-fat boundaries are clearly present in the LUS image. Conventional ultrasound imaging (Fig. 3d) and visual inspection (Fig. 3b) verify all boundaries and features detected in the LUS images. Dominant soft-tissue boundaries are present in both the LUS and conventional ultrasound images at 1 cm, 2 cm, and 3.5 cm. In particular, the contour of the first reflecting muscle-fat layer at 1 cm matches in both the LUS and conventional images.

In vivo human LUS imaging

The MIT Committee on the Use of Humans as Experimental Subjects (COUHES) approved the human LUS imaging protocol. Four volunteers' forearms were imaged using the LUS system. Consent from each volunteer was obtained prior to LUS and conventional ultrasound imaging. Both the source and detection lasers were verified for safety against the ANSI standard. Similar to previous experiments, conventional ultrasound imaging using the GE Logiq E9 system followed each LUS imaging session for feature verification. LUS and conventional scans were completed on the inside and backside of volunteers' forearms. No volunteer reported any sensation, discomfort, or tissue change during or following any LUS imaging session. Within the reconstructed LUS image shown in Fig. 4c, tissue features such as muscle fascia boundaries and the bone surface are clearly detected. In Fig. 4c, the muscle fascia boundary is present from 0.5 cm to 1 cm, and the surface of the ulna is present from 2 cm to 2.5 cm. Comparing LUS with conventional ultrasound, the same soft-tissue and bone features are present at the same depth, verifying that the LUS system



can detect features presently detected by conventional ultrasound.


Discussion

LUS Results and Human Imaging

The LUS imaging results for porcine and human tissue demonstrate the capability of LUS to remotely and deeply image biological tissue at safe optical exposure levels. The tissue boundaries detected by LUS are in agreement with the tissue boundaries detected by conventional ultrasound. Components of LUS technology have been explored in prior research, but a full demonstration of LUS on a human subject has not previously been presented. The human LUS imaging presented here (Fig. 4c) is the first instance of LUS imaging on a human subject in vivo, and volumetric LUS imaging (Fig. 2c) demonstrates LUS capabilities for 3D imaging. By restricting optical penetration and selecting optical wavelengths that maximize both the generation and detection of acoustic waves, we can demonstrate LUS imaging on tissue-mimicking phantoms, excised porcine tissue, and human subjects without any surface enhancements for the optical source or detector performance. Validation of the LUS results on human subjects is a significant step toward

proving the clinical viability of LUS and motivates further LUS research and development.


LUS imaging in porcine tissue (Fig. 3c) shows a clear and coherent boundary layer in the first 1.5 cm, but deeper layers present significant artifacts near the boundary. The same effect is visible in the human LUS image (Fig. 4c) for deeper layers. These depth-dependent artifacts are likely due to the lack of elevation focusing in LUS. While a conventional ultrasound probe is elevation focused to minimize out-of-plane reflections, optical sources and receivers are unfocused and become sensitive to more out-of-plane reflection at deeper imaging depths. More artifacts are visible in the human LUS image (Fig. 4c) than in the porcine LUS image (Fig. 3c) and can be attributed to patient motion and variations in optical backscatter from the skin. Overall, less acoustic speckle is present in the LUS images due to the lower imaging frequency, bandwidth, and lateral resolution of LUS than those of conventional ultrasound. Currently, LUS performance is limited by the acoustic frequency generated by the optical source and the detector sensitivity. The frequency of LUS sources is determined by the impinging optical wavelength and tissue absorption, with bandwidth limits due to the acoustic attenuation and tissue surface



roughness; the expected source bandwidth for LUS at 1540 nm is ~ 1.5 MHz⁴⁵. For optical detection, noncontact interferometric methods are limited by the optical backscatter from the tissue surface. Optical detectors have more than sufficient bandwidth for ultrasound imaging but are generally reliant on ideal stationary reflective surfaces for measurement, without consideration of human safety³⁵. Based on our human subject experiments, there is significant variation in the optical backscatter, as the detector was used to scan the skin tissue. Optical reflectance on human subjects has been reported for optical wavelengths between 250 and 2500 nm, but the measurement of backscatter specularly is lacking^{50,51}. Full skin reflection characteristics must be investigated to design specialized optical detectors for clinical LUS. For the LUS system reported here, the 1550 nm LDV was specifically selected to maximize permissible optical backscatter from the skin while maintaining safety. Adaptive focusing could further

reduce variability associated with skin variations and patient motion, but feedback control of the optical output based on the optical backscatter will be necessary.


At this point, the LUS images presented here are comparable to images presented at the incipient stages of medical ultrasound imaging decades ago¹. Quantitative image comparison of the current LUS technology against modern medical ultrasound is still premature. The conventional ultrasound images presented here use a 9 MHz center frequency transducer with multielement beamforming and are of expectedly higher image quality and resolution than those of the LUS images. Nevertheless, similar structures and dimensions are consistently observed in both modalities. While further work remains prior to commercialization and clinical use, the core enabling technologies of LUS are available. Discussed in later sections, recent research advances in laser technology, silicon photonics, and hydrogels could accelerate future LUS development.



Optical system design

From prior literature, LUS sources can be approximated by an equivalent disk transducer below the tissue surface, with the geometry determined by the parameters of the impinging light and absorption characteristics of the irradiated tissue^{22,23}. With proper thermal and mechanical stress confinement, the equivalent disk is prescribed, in diameter, by the impinging optical beam diameter, and in thickness, by the optical penetration depth (inverse of the tissue optical absorption coefficient at the impinging optical wavelength)²². Thus, the central frequency of the converted acoustic wave and the conversion efficiency are directly determined by the optical absorption coefficient²². For practical applications, the beam diameter should be larger than the penetration depth such that the dominant propagation direction of the converted acoustic wave is directed into the tissue. For maximum conversion efficiency, the optical absorption coefficient should equal the

acoustic wavenumber of the desired acoustic frequency to be generated²². For biological tissue, the stress confinement conditions for LUS are satisfied by nanosecond pulsed infrared sources, restricting optical conversion at the tissue surface using high optical absorption regions of water⁴⁵. To generate acoustic frequencies relevant for clinical ultrasound (1–10 MHz), ideal optical wavelengths for LUS source generation are ~1500 and 2000 nm, corresponding to peaks in the water absorption spectrum. At 1500 nm, the optical absorption water is $\sim 40 \text{ cm}^{-1}$, which corresponds to a converted acoustic source near 1 MHz. Prior work reports that both 1500 nm and 2000 nm pulsed sources generate acoustic waves with bandwidths up to 1.5 MHz⁴⁵. However, the MPE limit for nanosecond sources at 2000 nm is an order of magnitude lower than that at 1500 nm, 0.1 J/cm^2 vs. 1 J/cm^2 , respectively⁴⁶. Thus, the 1540 nm optical source was selected to maximize the converted acoustic source amplitude while remaining



within the safety limits. Following the disk transducer model, the optical beam diameter dictates the beam profile of the converted acoustic source. For the 2 mm diameter optical source on tissue with an expected frequency of ~ 1.5 MHz, the acoustic beam width is approximately 60 degrees. A narrower beam spread can be generated using larger optical beam diameters, but significantly higher optical energy is required to maintain the overall fluence.

For optical detection, optical backscatter from the tissue surface must be maximized. Prior measurements of optical reflectance from human skin report that wavelengths between 500 nm and 1200 nm offer the highest optical reflection factor^{46,50,51}. However, the safety MPE limits must be considered again. For a 100 μ s measurement duration (corresponding to an imaging depth of 7.5 cm in tissue), assuming similar photodetector responsivity across the spectrum, if a tissue surface is irradiated at the maximum MPE limits for every spectral

region, the maximum quantity of reflected light to a detector is still in the far infrared, between 1500 and 1800 nm. Thus, the selected 1550 nm LDV balances detection sensitivity with subject safety. Considering subject safety and acoustic performance, ~ 1500 nm is the optimal spectral region for both LUS transmission and detection and is able to leverage the commercial abundance of 1550 nm optical components from the telecommunication industry⁴⁵.

LUS performance is directly linked to the intrinsic tissue properties of human skin. The absorption and reflection parameters dictate LUS transmission and detection, respectively. The tissue surface roughness limits the LUS transmit bandwidth and may contribute to variation and specularly in optical reflection, which degrades the optical detection sensitivity. Skin tissue inhomogeneity from pigmentation and melanin may further affect the LUS performance and warrant further investigation.

Averaging multiple traces improves the overall signal-to-noise ratio, but the method is limited by MPE limits and increases the overall imaging time; poor or a lack of optical backscatter from the tissue surface cannot be resolved by averaging. Surface treatment, adaptive focusing, or multichannel optics may be necessary for clinical LUS systems to mitigate tissue variability. Furthermore, unlike fixed-geometry ultrasound arrays, the equivalent optical LUS array conforms to the local tissue geometry. The array geometry for the LUS images presented can be approximated as planar, but the imaging of large tissue regions with significant local curvature requires active localization of the sources and detectors to accurately reconstruct an image.

Enabling LUS technologies

Future LUS development should focus on component improvements as well as surface treatments to enhance multiple facets of LUS. An amplitude-modulated optical source can excite higher acoustic bands to improve the image resolution²². Implementing a fast amplitude-modulated optical source could improve the LUS imaging depth, bandwidth, and resolution; in addition, conventional ultrasound techniques used to amplify the acoustic signal-to-noise ratio, such as pulse compression, matched filtering, and source encoding, can be leveraged. Currently, both LUS and PA are limited by the existing laser technology^{16–18,41,52}. Beyond amplitude modulation, the parallelization of optical sources and receivers to enable optical transmit and receive beamforming techniques will be a critical turning point, similar to how piezoelectric arrays enabled and revolutionized clinical ultrasound imaging. Since the spatial locations of LUS sources are restricted to the tissue surface and can be known *a priori*, transmit beamforming is possible only in LUS and not in penetrating PA. The adaptation of computer vision techniques such as 3D imaging and tracking technologies could make large-volume LUS imaging and optical beamforming feasible. Optical spot tracking is also required if patient motion is significant during data acquisition. With sufficiently fast data acquisition and coverage, LUS systems could behave similar to a body volume camera, capable of simultaneous 3D imaging of both the external and internal tissue geometries. The silicon photonics industry, fueled by sensors required for autonomous car operations, has developed chip-scale, solid-state, steerable laser technology applicable for both PA and LUS imaging^{53–57}. Since the ideal operating range for LUS is near 1500 nm, communication and silicon photonic innovations may directly impact LUS system development.

As discussed previously, the LUS system performance is subject to the optical and acoustic properties of the tissue. A minor surface treatment layer could bypass the tissue

limitations. Radio frequency (RF) coils and contrast agents are commonly used for MRI and CT, respectively; designing a surface treatment layer for LUS is not inconceivable. Enhanced optical reflection is commonly achieved with retroreflective dust or tapes^{39,42}. However, retroreflective dust is a respiratory irritant, and tape or dust on the tissue surface impedes the LUS source. Gel or gel pads embedded with retroreflectors could be designed to enhance optical detection without interference to the optical source; the bulk water content of the gel can preserve the LUS source characteristics in the IR spectrum, while embedded reflectors can enhance the optical backscatter. Furthermore, source bandwidth limitations due to the tissue roughness can be bypassed by generating the LUS source in the gel layer. Gel or gel pads routinely used for ultrasound imaging can be augmented for LUS imaging. Flexible hydrogels could also be designed for LUS; the bulk water content of hydrogels mixed with retroreflective particles can conform to the tissue surface and enhance the LUS performance⁵⁸. The treatment layer could expand the LUS source bandwidth, improve the optical detection sensitivity, and permit higher optical exposure limits. However, irradiance exceeding the MPE limits on the treatment layer may mandate additional safety measures such as enclosure or eye protection. New optical designs should be evaluated to accommodate the use of surface treatments in LUS imaging.

Outlook

Based on these encouraging results, LUS inspires confidence for further research and development. Human LUS images verify the feasibility of LUS for in vivo anatomical imaging without compromising patient safety. LUS is sensitive to both hard and soft anatomical features, similar to conventional ultrasound, but is fully non-contact. The real-time remote sensing of biological tissue would find broad applicability in noninvasive patient monitoring, contact sensitive imaging (elastography and musculoskeletal disease tracking), and intraoperative applications. Current embodiments of LUS for research are generally single-point transmission and detection due to the high cost and complexity. Clinical iterations of LUS will require multipoint optical transmission and detection to amplify the acoustic source amplitudes and reduce the data acquisition time. Analogous challenges existed in the nascent stages of conventional ultrasound imaging. Scaling from mechanically scanned single-element systems to highly parallelized real-time clinical imagers took decades of research and development¹. A similar pathway is likely for LUS – parallelization of single- to multipoint laser technology. The initial human results are encouraging; rapid advances in related industrial sectors will establish the technologies necessary to enable clinical implementation of LUS.

Materials and methods

System configuration and imaging

A 1540 nm passively Q-switched pulsed laser (Optitask OT-37) was selected for LUS source generation. A commercially available class-2 1550 nm laser Doppler Vibrometer (Polytec RSV-150) with a green tracking laser (<1 mW) was selected for the optical detection of ultrasonic signals. The bandwidth of the LDV is 2.5 MHz, with 49 mm/s/V sensitivity. The minimum focused diameter for the LDV is 135 μm . Both lasers are within the respective pulsed and continuous MPE limits in the IR spectrum for eye and skin safety. A significant margin of safety is available to engineer custom LUS solutions with higher power and faster repetition rates but is outside the scope of the human LUS feasibility study presented here.

For each slice or volumetric image, the optical source and detection points were collocated and sequentially scanned on the sample surface linearly or rastered, respectively. Optical steering and collocation were completed using fast steering mirrors (Optics in Motion LLC, OIM102.3) and linear stages (Sigma Koki Co, SGSP26–200 stage, SHOT-702 motion controller). The spatial sampling frequency in each instance satisfies the Nyquist limits for the expected 1.5 MHz maximum acoustic frequency to prevent grating lobe interference. A scan length of ~ 5 cm at 0.5 mm pitch was completed for each imaging instance. The aperture length for LUS was the same as that for the conventional ultrasound probe used for verification. For volumetric LUS imaging, a phantom was scanned with a 4 cm by 8 cm mesh grid with a 0.5 mm pitch. For each sampling point, 50 time traces in total were recorded and compiled for postprocessing and image reconstruction. Data acquisition from the LDV was recorded on a digital oscilloscope (National Instruments, NI PXIe-5170R, 14-bits, 250 MHz max bandwidth) nested in an NI PXIe-1073 chassis. Post data acquisition, an SAFT algorithm with a coherence factor – to minimize the side-lobe artifacts – was used to reconstruct images for each LUS dataset³⁷.

Phantom composition

The LUS phantoms consist of porcine gelatin (6% by weight, Sigma-Aldrich, gel strength 300, Type A) dissolved in deionized water at 85 °C. To match typical tissue sound speeds, 1-propanol (2% by weight, Alfa Aesar, A19902) was dissolved in the gelatin mixture. The solution cooled to 45 °C prior to pouring into molds with the desired inclusions. The solution was degassed in the molds and left to solidify at room temperature. All phantoms were sealed and stored at 4 °C and used in the experiments within one week of construction to avoid sound speed changes due to water loss over time.

Animal sample preparation

The use of animal tissue was approved by the MIT Committee on Animal Care, protocol number E17–09–0320. Porcine abdominal tissue was obtained from the local market. The sample tissue strips include skin and multiple layers of visible muscle and fat without deep cartilage or bone inclusions. The skin surfaces were not specially selected or treated to include natural skin variations in the experiments. The strips were ~ 5 cm thick and cut to 25 cm lengths prior to LUS imaging.

Human LUS

All human LUS experiments were completed with approval from the MIT Committee on the Use of Humans as Experimental Subjects (protocol # 1702850719R001). As an extra precaution, all laser component optical outputs were measured using an optical power meter (Ophir Nova II). As dictated by the Institute Review Board (IRB), all optical outputs were independently verified to be eye and skin safe. Laser safety glasses were still available if requested by any volunteer. All personal identifiers were removed from the LUS and conventional US data to preserve volunteer anonymity. Each imaging instance covered an ~ 5 cm line across a volunteer's forearm, with conventional ultrasound imaging completed at the same location immediately following LUS imaging. The forearm was selected for ease of accessibility by the LUS system and significant collections of soft- and hard-tissue features.

Conventional ultrasound validation

A standard clinical system, GE Logiq E9, with a 9 MHz linear probe provided verification of the LUS images using conventional ultrasound methods. The acquisition process with the conventional system consisted of applying ultrasound gel (Aquasonics 100) to the probe and acquiring B-mode ultrasound images at the same location as that of the LUS scan. Although the conventional ultrasound system operates at significantly higher frequencies and leverages beamforming, gross comparisons of large soft-/hard-tissue features, such as fascia, arteries, muscles, tendons, and bones, are still possible between LUS and conventional ultrasound. For this article, B-mode ultrasound serves as the gold standard to verify the tissue features detected using LUS.

Acknowledgements

This work was funded by the biomedical line program at the MIT Lincoln Laboratory and by the U.S. Army Medical Research and Materiel Command's Military Operational Medicine Research Program. We thank Kai Thomenius for the technical discussion and Anthony Samir for discussions on clinical applications. We thank Robert McLaren for his technical assistance with the laser construction. We thank LCDR Steele from the U.S. Army Medical Research and Materiel Command's Military Operational Medicine Research Program for their continued funding and support of this work. This work is funded by the

MIT Lincoln Laboratory Biomedical Line Program for the United States Air Force and by the U.S. Army Medical Research and Materiel Command's Military Operational Medicine Research Program. The views expressed in this paper are those of the author(s) and do not reflect the official policy of the Department of Army, Department of Defense, or the U.S. Government.

Author details

¹Department of Mechanical Engineering, Massachusetts Institute of Technology, 77 Massachusetts Ave., Cambridge, MA 02139, USA. ²Institute for Medical Engineering & Science, Massachusetts Institute of Technology, 45 Carleton St., Cambridge, MA 02142, USA. ³Lincoln Laboratory, Massachusetts Institute of Technology, 244 Wood Street, Lexington, MA 02421, USA

Author contributions

X.Z., J.R.F., C.M.W., and R.W.H. performed the experiments and designed the system components. M.R.J. built the electrical components for the optical source. X.Z. and J.R.F. analyzed and reconstructed the data. R.W.H. and B.W.A. provided guidance and advised the study. R.W.H. is the program manager and technical lead with C.M.W. of the noncontact laser ultrasound program at the MIT Lincoln Laboratory. All authors reviewed and edited the manuscript.

Data availability

The datasets generated and analyzed in the article are available from the corresponding author upon reasonable request.

Conflict of interest

The authors declare that they have no conflict of interest.

Received: 20 June 2019 Revised: 25 November 2019 Accepted: 27

November 2019

Published online: 20 December 2019

References

- Szabo, T. L. *Diagnostic Ultrasound Imaging: Inside Out*. (Elsevier/Academic Press, 2014).
- Syversveen, T. et al. Tissue elasticity estimated by acoustic radiation force impulse quantification depends on the applied transducer force: an experimental study in kidney transplant patients. *Eur. Radiol.* **22**, 2130–2137 (2012).
- Lam, A. C. L. et al. The influence of precompression on elasticity of thyroid nodules estimated by ultrasound shear wave elastography. *Eur. Radiol.* **26**, 2845–2852 (2016).
- Porra, L. et al. The effect of applied transducer force on acoustic radiation force impulse quantification within the left lobe of the liver. *Australas. J. Ultrasound Med.* **18**, 100–106 (2015).
- Hopp, T. et al. Breast imaging with 3D ultrasound computer tomography: results of a first in-vivo study in comparison to MRI images. Proceedings of the 12th International Workshop on Digital Mammography. Gifu City: Springer, 2014. https://doi.org/10.1007/978-3-319-07887-8_11
- Duric, N. et al. Detection of breast cancer with ultrasound tomography: first results with the Computed Ultrasound Risk Evaluation (CURE) prototype. *Med. Phys.* **34**, 773–785 (2007).
- Hollenhorst, M. et al. Ultrasound computed tomography in breast imaging: first clinical results of a custom-made scanner. *Ultraschall der Med.* **31**, 604–609 (2010).
- Waag, R. C. & Fedewa, R. J. A ring transducer system for medical ultrasound research. *IEEE Trans. Ultrason., Ferroelectr., Frequency Control* **53**, 1707–1718 (2006).
- Zhang, X. et al. In *Proceedings of the 2015 37th Annual International Conference of the IEEE Engineering in Medicine and Biology Society*. Milan, Italy: IEEE, 2015, 5541–5544.
- Zhang, X. et al. In *Proceedings of SPIE 9790, Medical Imaging 2016: Ultrasonic Imaging and Tomography*. San Diego: SPIE, 2016.
- Ranger, B. et al. In *Proceedings of SPIE 7629, Medical Imaging 2010: Ultrasonic Imaging, Tomography, and Therapy*. San Diego: SPIE, 2010, 762906.
- Cox, B. T. et al. Quantitative spectroscopic photoacoustic imaging: a review. *J. Biomed. Opt.* **17**, 061202 (2012).
- Wang, L. V. *Photoacoustic Imaging and Spectroscopy*. (Boca Raton: CRC Press, 2009).
- Bell, A. G. On the production and reproduction of sound by light. *Am. J. Sci.* **20**, 305–324 (1880).
- Xu, M. H. & Wang, L. V. Photoacoustic imaging in biomedicine. *Rev. Sci. Instrum.* **77**, 041101 (2006).
- Wang, L. V. & Yao, J. J. A practical guide to photoacoustic tomography in the life sciences. *Nat. Methods* **13**, 627–638 (2016).
- Wang, L. V. & Hu, S. Photoacoustic tomography: in vivo imaging from organs to organs. *Science* **335**, 1458–1462 (2012).
- Taruttis, A. & Ntziachristos, V. Advances in real-time multispectral optoacoustic imaging and its applications. *Nat. Photonics* **9**, 219–227 (2015).
- Haupt, R. et al. Non-contact laser ultrasound concept for biomedical imaging. *Proceedings of 2017 IEEE International Ultrasonics Symposium*. Washington: IEEE, 2017. <https://doi.org/10.1109/ULTSYM.2017.8091941>.
- Beard, P. Biomedical photoacoustic imaging. *Interface Focus* **1**, 602–631 (2011).
- Ku, G. & Wang, L. V. Deeply penetrating photoacoustic tomography in biological tissues enhanced with an optical contrast agent. *Opt. Lett.* **30**, 507–509 (2005).
- Lyamshev, L. M. *Radiation Acoustics*. (Boca Raton: CRC Press, 2004).
- Gusev, V. E. & Karabutov, A. A. *Laser Optoacoustics*. (New York: American Institute of Physics, 1993).
- Jathoul, A. P. et al. Deep in vivo photoacoustic imaging of mammalian tissues using a tyrosinase-based genetic reporter. *Nat. Photonics* **9**, 239–246 (2015).
- Hristova, Y., Kuchment, P. & Nguyen, L. Reconstruction and time reversal in thermoacoustic tomography in acoustically homogeneous and inhomogeneous media. *Inverse Probl.* **24**, 055006 (2008).
- Johnson, J. L., Shragge, J. & van Wijk, K. Image reconstruction of multi-channel photoacoustic and laser-ultrasound data using reverse time migration. Proceedings of SPIE 9323, Photons Plus Ultrasound: Imaging and Sensing 2015. San Francisco: SPIE, 2015, 932314.
- Monchalain, J. P. Optical detection of ultrasound. *IEEE Trans. Ultrason., Ferroelectr., Frequency Control* **33**, 485–499 (1986).
- Zhang, E., Laufer, J. & Beard, P. Backward-mode multiwavelength photoacoustic scanner using a planar Fabry-Perot polymer film ultrasound sensor for high-resolution three-dimensional imaging of biological tissues. *Appl. Opt.* **47**, 561–577 (2008).
- Dong, B. Q. et al. Isometric multimodal photoacoustic microscopy based on optically transparent micro-ring ultrasonic detection. *Optica* **2**, 169–176 (2015).
- Park, S. J. et al. Noncontact photoacoustic imaging based on all-fiber heterodyne interferometer. *Opt. Lett.* **39**, 4903–4906 (2014).
- Rousseau, G., Blouin, A. & Monchalain, J. P. Non-contact photoacoustic tomography and ultrasonography for tissue imaging. *Biomed. Opt. Express* **3**, 16–25 (2012).
- Hochreiner, A. et al. Non-contact photoacoustic imaging using a fiber based interferometer with optical amplification. *Biomed. Opt. Express* **4**, 2322–2331 (2013).
- Wissmeyer, G. et al. All-optical optoacoustic microscope based on wideband pulse interferometry. *Opt. Lett.* **41**, 1953–1956 (2016).
- Ashkenazi, S. et al. Optoacoustic imaging using thin polymer etalon. *Appl. Phys. Lett.* **86**, 134102 (2005).
- Wissmeyer, G. et al. Looking at sound: optoacoustics with all-optical ultrasound detection. *Light: Sci. Appl.* **7**, 53 (2018).
- Rousseau, G. et al. Non-contact biomedical photoacoustic and ultrasound imaging. *J. Biomed. Opt.* **17**, 061217 (2012).
- Lévesque, D. et al. Performance of laser-ultrasonic F-SAFT imaging. *Ultrasonics* **40**, 1057–1063 (2002).
- Monchalain, J. P. Laser-ultrasonics: from the laboratory to industry. *AIP Conf. Proc.* **700**, 3–31 (2004).
- Johnson, J. L., Shragge, J. & van Wijk, K. Nonconfocal all-optical laser-ultrasound and photoacoustic imaging system for angle-dependent deep tissue imaging. *J. Biomed. Opt.* **22**, 041014 (2017).
- Johnson, J. L., van Wijk, K. & Sabick, M. Characterizing phantom arteries with multi-channel laser ultrasonics and photo-acoustics. *Ultrasound Med. Biol.* **40**, 513–520 (2014).
- Paltauf, G. et al. Photoacoustic tomography using a Mach-Zehnder interferometer as an acoustic line detector. *Appl. Opt.* **46**, 3352–3358 (2007).
- Johnson, J. L. et al. All-optical extravascular laser-ultrasound and photoacoustic imaging of calcified atherosclerotic plaque in excised carotid artery. *Photoacoustics* **9**, 62–72 (2018).

43. Stratoudaki, T., Clark, M. & Wilcox, P. D. Adapting the full matrix capture and the total focusing method to laser ultrasonics for remote non destructive testing. Proceedings of 2017 IEEE International Ultrasonics Symposium. Washington: IEEE, 2017. <https://doi.org/10.1109/ULTSYM.2017.8092864>.
44. Eom, J., Park, S. J. & Lee, B. H. Noncontact photoacoustic tomography of *in vivo* chicken chorioallantoic membrane based on all-fiber heterodyne interferometry. *J. Biomed. Opt.* **20**, 106007 (2015).
45. Fincke, J. R. et al. Characterization of laser ultrasound source signals in biological tissues for imaging applications. *J. Biomed. Opt.* **24**, 021206 (2018).
46. US-ANSI. ANSI Z136.1-2007 American National Standard for Safe use of lasers (Laser Institute of America, Orlando, 2007).
47. Brożyna, A. et al. Porcine skin as a model system for studies of adverse effects of narrow-band uvb pulses on human skin. *J. Toxicol. Environ. Health, Part A* **72**, 789–795 (2009).
48. Summerfield, A., Meurens, F. & Ricklin, M. E. The immunology of the porcine skin and its value as a model for human skin. *Mol. Immunol.* **66**, 14–21 (2015).
49. Kong, R. & Bhargava, R. Characterization of porcine skin as a model for human skin studies using infrared spectroscopic imaging. *Analyst* **136**, 2359–2366 (2011).
50. Cooksey, C. C., Tsai, B. K. & Allen, D. W. A collection and statistical analysis of skin reflectance signatures for inherent variability over the 250 nm to 2500 nm spectral range. Proceedings of SPIE 9082, Active and Passive Signatures V. Baltimore: SPIE, 2014, 908206.
51. Cooksey, C. C. & Allen, D. W. Reflectance measurements of human skin from the ultraviolet to the shortwave infrared (250 nm to 2500 nm). Proceedings of SPIE 8734, Active and Passive Signatures IV. Baltimore: SPIE, 2013, 87340N.
52. Taruttis, A., van Dam, G. M. & Ntziachristos, V. Mesoscopic and macroscopic optoacoustic imaging of cancer. *Cancer Res.* **75**, 1548–1559 (2015).
53. Sun, J. et al. Large-scale nanophotonic phased array. *Nature* **493**, 195–199 (2013).
54. Cole, D. B. et al. Integrated heterodyne interferometer with on-chip modulators and detectors. *Opt. Lett.* **40**, 3097–3100 (2015).
55. Bradley, J. D. B. et al. Monolithic erbium- and ytterbium-doped microring lasers on silicon chips. *Opt. Express* **22**, 12226–12237 (2014).
56. DeRose, C. T. et al. Electronically controlled optical beam-steering by an active phased array of metallic nanoantennas. *Opt. Express* **21**, 5198–5208 (2013).
57. Poulton, C. V. et al. Coherent solid-state LIDAR with silicon photonic optical phased arrays. *Opt. Lett.* **42**, 4091–4094 (2017).
58. Sun, J. Y. et al. Highly stretchable and tough hydrogels. *Nature* **489**, 133–136 (2012).