

Full-polarized Tomographic Diffraction Microscopy Achieves a Resolution about One-Fourth of the Wavelength

T. Zhang,¹ Y. Ruan,¹ G. Maire,¹ D. Sentenac,³ A. Talneau,² K. Belkebir,¹ P.C. Chaumet,^{1,*} and A. Sentenac¹

¹Aix Marseille Université, CNRS, Centrale Marseille, Institut Fresnel, UMR 7249, 13013 Marseille, France

²Laboratoire de Photonique et de Nanostructure, 91460 Marcoussis, France

³European Gravitational Observatory, 56021 Cascina (PI), Italy

(Received 14 March 2013; revised manuscript received 23 July 2013; published 12 December 2013)

We present a marker-free microscope that records the phase, amplitude, and polarization state of the field diffracted by the sample for different illumination directions. The data are processed with an appropriate inversion method to yield the sample permittivity map. We observe that the full-polarized information ameliorates significantly the three-dimensional image of weakly scattering subdiffraction objects. A resolution about one-fourth of the illumination wavelength is experimentally demonstrated on complex samples.

DOI: [10.1103/PhysRevLett.111.243904](https://doi.org/10.1103/PhysRevLett.111.243904)

PACS numbers: 42.30.Wb, 42.25.Fx

Optical microscopy is an old, popular tool for imaging microscopic samples. In the past decades, many different configurations such as confocal, dark-field, phase [1] or the more recent tomographic diffraction [2] microscopies have been developed to improve the resolution, the sensitivity, and/or the contrast of nonfluorescent specimen images. All these techniques rely on the same physical principle: the sample interacts with an incident light beam (coherent or incoherent) and the image is formed analogically or numerically from the scattered far field. In the single scattering regime, a far-field microscope is ideally able to capture the sample information up to frequency $2/\lambda$ where λ is the illumination wavelength [3]. Thus, the resolution, defined as the full width at half maximum of the image of a point object, could, theoretically, reach $\lambda/4$, i.e., half the conventional diffraction limit. In practice, the noise and the image formation process deteriorate the restitution of the sample high spatial frequencies. Therefore, the resolution of the best current microscopes is still about $\lambda/2$ [4].

A promising way for ameliorating the resolution is to take advantage of the vectorial nature of light. Thus, the polarization-dependent behavior of a tightly focused beam [5,6] has been used to improve the resolution of a confocal-like microscope along one specific direction [7]. However, the intrinsic difficulty to combine confocal images obtained under various incident polarization states [7] has hampered the use of this approach to reach a resolution significantly beyond the diffraction limit in all directions of the image plane.

Another interesting approach is tomographic diffraction microscopy (TDM) which numerically reconstructs the sample three-dimensional image from many holograms recorded under different incident angles. Remarkably, a transverse resolution about one-fourth of the wavelength has been observed on elongated objects [8] with TDM. Yet, this achievement has never been evidenced on a three-dimensional specimen. The most likely reason is that

TDM does not account for the light polarization effect in the image formation process [9]. The holograms are obtained for one given polarization state and the reconstruction procedures are based on scalar approximate models [4,10,11].

In this Letter, we sought to exploit the full potential of TDM by taking advantage of the light polarization. We implemented a tomographic diffraction microscope that records and processes the vectorial complex field scattered by a sample for any incident polarization state [12]. The setup, described in Fig. 1, is a modified reflection microscope in which an off-axis holography technique has been introduced to recover the phase and amplitude of the field at the microscope image plane. A collimated laser beam (He-Ne wavelength $\lambda = 633$ nm), controlled by a mirror mounted on step motors (Newport NSA12), illuminates the sample through an air objective satisfying Sine-Abbe condition with numerical aperture $NA = 0.95$ (Zeiss




FIG. 1 (color online). Schematic of the full-polarized TDM setup. M the rotative mirror. BE the beam expander. D the diaphragm. BS the beam splitter. OL the objective lens. P the pinhole. L_1 the tube lens. L_2 and L_3 are the relay lenses ($f' = 3.5$ and 20 cm, respectively). HW_1 and HW_2 the half-wave plate on the incident field and reference field, respectively.

Epiplan-Apochromat $\times 50$) under various angles of incidence. At the image plane of the microscope, the field scattered by the sample, \mathbf{E} , reads, overlooking the magnification factor,

$$\mathbf{E}(\mathbf{r}_{\parallel}) = \int_{k_{\parallel} < k_0 \text{NA}} \frac{1}{\gamma} (e_s \hat{\mathbf{s}} + e_p \hat{\mathbf{p}}) \exp(i\mathbf{k}_{\parallel} \cdot \mathbf{r}_{\parallel}) d\mathbf{k}_{\parallel}, \quad (1)$$

where $\hat{\mathbf{z}}$ denotes the optical axis of the microscope, $\gamma = \sqrt{(2\pi/\lambda)^2 - k_{\parallel}^2}$, $\hat{\mathbf{s}} = \hat{\mathbf{z}} \times \hat{\mathbf{k}}$, $\hat{\mathbf{p}} = \hat{\mathbf{s}} \times \hat{\mathbf{k}}$ with $\mathbf{k} = \mathbf{k}_{\parallel} + \gamma \hat{\mathbf{z}}$ and (e_s, e_p) are the s and p complex amplitudes of the sample scattered plane wave propagating along \mathbf{k} . For each illumination, one records with a CCD camera (Kappa DX4-1020FW), the interference patterns between \mathbf{E} and an off-axis reference wave which is successively polarized along two different directions $\hat{\mathbf{d}}_1$ and $\hat{\mathbf{d}}_2$ [13]. Processing the holograms yields $\mathbf{E} \cdot \hat{\mathbf{d}}_1$ and $\mathbf{E} \cdot \hat{\mathbf{d}}_2$ from which (e_s, e_p) for all \mathbf{k}_{\parallel} such that $k_{\parallel} < k_0 \text{NA}$ are easily extracted. To obtain the vectorial complex field scattered by the sample for any incident polarization state, one needs to record (e_s, e_p) for two independent incident polarization states. Hence, the full-polarized TDM mode requires four intensity measurements for each incident angle (and possibly only two using [12]).

The different incident and reference polarization states are formed thanks to the half-wave plates that are placed on the illumination and reference wave paths, see Fig. 1. To avoid turning off the specular reflection (which is used for normalizing the data [14]), the incident polarization directions are set to $\hat{\mathbf{x}}$ and $\hat{\mathbf{y}}$, while the reference ones are $\hat{\mathbf{d}}_1 = (\hat{\mathbf{x}} + \hat{\mathbf{y}})/\sqrt{2}$ and $\hat{\mathbf{d}}_2 = (\hat{\mathbf{x}} - \hat{\mathbf{y}})/\sqrt{2}$. For comparison purposes, we also considered the classical implementation of TDM in which the incident and reference polarization directions are the same, either $\hat{\mathbf{x}}$ or $\hat{\mathbf{y}}$. In these configurations (hereafter indicated by $\hat{\mathbf{x}} \hat{\mathbf{x}}$ or $\hat{\mathbf{y}} \hat{\mathbf{y}}$), the data consist of the scalar projection of the scattered field on the incident polarization state. Thus, there are four times less data in the $\hat{\mathbf{x}} \hat{\mathbf{x}}$ or $\hat{\mathbf{y}} \hat{\mathbf{y}}$ modes than in the full-polarization mode.

In this Letter, the samples to be imaged are deposited on a silicon substrate and are described by a relative permittivity contrast $\chi(\mathbf{r})$ with respect to the planar geometry [9,15]. The unknown sample permittivity contrast is reconstructed from the TDM data using an iterative inversion method based on a rigorous vectorial electromagnetic model of the scattered far field (e_s, e_p) , which, in particular, accounts for the substrate [16]. The algorithm consists in estimating, iteratively, the relative permittivity contrast χ of a bounded domain W (which is known to contain the sample), such that it minimizes a cost function representing the distance between the measurements and the simulated scattered far field obtained with the permittivity estimation [9,17]. The minimization of the cost function is performed using a conjugate gradient algorithm and, to improve the reconstruction, the unknown relative permittivity in W is assumed to be real and positive. Note that the method is

adapted to any TDM data. In the full-polarization mode, the cost function involves the full vectorial field scattered by the sample, whereas in the $\hat{\mathbf{x}} \hat{\mathbf{x}}$ mode, it involves only the projection of the scattered field.

For investigating the performance of our setup, we consider a sample made of $h = 150$ nm high resin cylinders with diameter $D = 150$ nm and relative permittivity 2 deposited on a planar Si substrate of relative permittivity $15 + i0.15$, and centered at the summits of a square with sides 300 nm along $\hat{\mathbf{x}}$ and $\hat{\mathbf{y}}$ as depicted in Fig. 2(a). The scattered far field is measured in both the full-polarized and scalar ($\hat{\mathbf{x}} \hat{\mathbf{x}}$, $\hat{\mathbf{y}} \hat{\mathbf{y}}$) modes for 10 incident directions in the (x, z) plane and 10 in the (y, z) plane with azimuthal angles varying from -50° to 50° . With this angular configuration, the diffraction limit of the microscope is about $0.5\lambda/\text{NA} \approx 340$ nm, i.e., above the cylinders center interdistance. It is observed in Fig. 2(b) that, unsurprisingly, the four cylinders cannot be distinguished on a classical wide-field microscope image (which was synthesized by summing the image field intensities for all the illuminations).

We first analyze the achievement of a classical TDM implementation by processing the scalar $\hat{\mathbf{y}} \hat{\mathbf{y}}$ data with the usual inverse Fourier transform algorithm [2]. The latter relies on a scalar, free-space (no substrate), single scattering model which states that the far field scattered in the \mathbf{k} direction by a sample illuminated by a plane wave with wave vector \mathbf{k}_i is proportional to the three-dimensional Fourier transform of $\chi(\mathbf{r})$ taken at $(\mathbf{k} - \mathbf{k}_i)$. Theoretically, in a noise-free configuration, the transverse image resolution should be about $0.25\lambda/\text{NA} \approx 170$ nm with our angle configuration, i.e., below the cylinders center interdistance [3]. Yet, it is seen in Fig. 3(a) that the simple inversion scheme fails in retrieving the sample. Its noise sensitivity (due to the weak number of incidences) and its model errors are prohibitive for imaging such small objects.

Applying the iterative inversion method to the same scalar data provides a much better sample estimate. The cylinders height is now accurately retrieved. However, their shapes are strongly elongated along the $\hat{\mathbf{x}}$ axis, Fig. 3(b). This resolution anisotropy is a direct consequence of the incident and reference polarization choice, as was observed in Ref. [7], and will be explained below. We now turn to the reconstructions obtained either with the




FIG. 2 (color online). (a) Sample geometry. $D = 150$ nm and $h = 150$ nm. (b) Wide-field image of the sample obtained using spatially incoherent illumination: details of the object are not resolved.




FIG. 3 (color online). Sample reconstructions obtained from the scalar $\hat{y}\hat{y}$ data. (a) Longitudinal and transverse cuts of the three-dimensional image given by a direct inversion method based on a free-space scalar model. (b) Longitudinal and transverse cuts of the relative permittivity estimation ($\varepsilon = \chi + 1$) given by an iterative inversion method based on a rigorous vectorial model of the data accounting for the substrate and for the field projection onto the \hat{y} direction.

combined scalar $\hat{x}\hat{x}$ and $\hat{y}\hat{y}$ data, Figs. 4(a)–4(c), or the full-polarized vectorial mode, Figs. 4(d)–4(f). In both cases, the four cylinders are accurately retrieved with isotropic resolution, the full-polarized vectorial data yielding




FIG. 4 (color online). Relative permittivity reconstructions, $\varepsilon = \chi + 1$, obtained by the iterative inversion method based on a rigorous vectorial model of the field from the combined $\hat{x}\hat{x}$ and $\hat{y}\hat{y}$ data (a)–(c) and from the full-polarized vectorial data (d)–(f). (a),(d) xy section, (b),(e) yz section taken along the dashed line in (a),(d), (c),(f) reconstructed permittivity along the dashed line in (b),(e). The full width at half maximum of the reconstructed permittivity is about one-fourth of the wavelength.

a better quantitative estimation of the permittivity level. The resolution, defined as the full width at half maximum of the reconstructed permittivity peaks, Figs. 4(c) and 4(f), is about one-fourth of the wavelength.

The role of the incident and reference polarization states on the image resolution requires a careful analysis. Indeed, when single scattering is dominant (which is the case for this sample), the scalar $\hat{x}\hat{x}$ or $\hat{y}\hat{y}$ data give access to the sample permittivity Fourier coefficients within the same Fourier domain as the vectorial full-polarized data [15]. Hence, the resolution of the image depends only on the ability to extract, from noisy data, the permittivity Fourier coefficients at the highest accessible spatial frequencies. The latter are conveyed by the far field scattered at high angles, namely for large $\mathbf{k} - \mathbf{k}_i$. Now, in presence of a substrate, the far-field behavior in these directions significantly differs depending on the incident polarization state. In Fig. 5, we plot the scattered far-field modulus in the \mathbf{k}_{\parallel} plane obtained experimentally and theoretically when the sample is illuminated under 50° in the (x, z) plane with a p -polarized (corresponding to the \hat{x} direction for the incident half-wave plate) and s -polarized (corresponding to the \hat{y} direction) plane waves. We observe that the back scattered field is about twice bigger with the p -polarized incident wave than with the s -polarized incident wave. The better signal to noise ratio at high angles obtained with p -polarized incident waves explains the better resolution along the \hat{y} direction that is observed in Fig. 3(b). To confirm this interpretation, we plot, in Fig. 6, the relative permittivity map reconstructed by the iterative inversion




FIG. 5 (color online). Experimental (a),(c) and theoretical (b),(d) modulus of the far field scattered by the sample illuminated in the (x, z) plane under 50° of incidence. The scattered field is normalized by the incident magnitude times 10^8 . The black disk on the left indicates the nonexploitable angular domain about the specular reflection. (a),(b) the incident wave is s polarized. (c),(d) the incident wave is p polarized. One observes on the right of the plots that the back scattered field amplitudes, which convey the highest spatial frequency sample information, are twice as big in p polarization as in s polarization.




FIG. 6 (color online). xy (a),(c) and yz (b),(d) sections of the reconstructed permittivity given by the iterative “rigorous” inversion method from the measured far field. (a) and (b) with s -polarized incident waves; (c) and (d) with p -polarized incident waves. The superiority of p illumination for imaging the sample is evidenced.

method from the scattered vectorial field obtained using only s -polarized incident waves, Figs. 6(a) and 6(b), or only p -polarized incident waves, Figs. 6(c) and 6(d). As expected, the p -polarized mode yields an accurate sample reconstruction, similar to that of the full-polarized mode, while the s -polarized mode provides a totally blurred image.

To document further the potential of the full-polarized TDM, we image a complex sample made of twelve resin rods of width 100 nm, length 300 nm, and height 140 nm radially placed at the summit of a dodecagon [Fig. 7(a)]. The sample is illuminated by eight directions of incidence, defined by a fixed polar angle $\theta = 60^\circ$ and an azimuthal angle regularly spaced within 2π . The dark-field microscope image, simulated by summing the scattered




FIG. 7 (color online). (a) Scanning electron microscope image. (b) Dark-field optical microscope image. (c) Reconstructed permittivity averaged over the sample’s height using full-polarized TDM data. (d) Permittivity along the dashed circle in (c). Plain line: full-polarized data; dashed line: the combined $\hat{x}\hat{x}$ and $\hat{y}\hat{y}$ data.

intensities recorded at the image plane for all the illuminations, displays a blurred ring with no indication of individual rods [Fig. 7(b)]. In contrast, the full polarized TDM recovers the twelve rods, Fig. 7(c). The reconstructed permittivity taken along a circle passing through the middle of the rods demonstrates that the technique is able to distinguish the rods whatever their orientation even though their interdistance is about one-fourth of the wavelength. On this complex sample, the full-polarized permittivity reconstruction is significantly better than the $\hat{x}\hat{x}$ and $\hat{y}\hat{y}$ one, Fig. 7(d), especially for the rods on either side of the image diagonals.

In conclusion, we have developed the first optical digital microscope that exploits all the information accessible via the diffraction process (intensity, phase, and polarization state of the scattered field for any possible illumination within the numerical aperture of the objective). In the single scattering regime, this ultimate microscope is able to reconstruct permittivity maps with a resolution about one-fourth of the wavelength. This experimental achievement, which outperforms that of all existing far-field microscopes, points out the importance of accounting for light polarization when tackling super resolution and sets a landmark in optical far-field imaging.

This work has been financed by the ANR SURMITO.

*patrick.chaumet@fresnel.fr

- [1] D. B. Murphy and M. W. Davidson, *Fundamentals of Light Microscopy and Electronic Imaging* (Wiley, New York, 2012).
- [2] V. Lauer, *J. Microsc.* **205**, 165 (2002).
- [3] O. Haeberlé, K. Belkebir, H. Giovannini, and A. Sentenac, *J. Mod. Opt.* **57**, 686 (2010).
- [4] Y. Cotte, F. Toy, P. Jourdain, N. Pavillon, D. Boss, P. Magistretti, P. Marquet, and C. Depeursinge, *Nat. Photonics* **7**, 113 (2013).
- [5] R. Dorn, S. Quabis, and G. Leuchs, *Phys. Rev. Lett.* **91**, 233901 (2003).
- [6] D. Lara and C. Dainty, *Appl. Opt.* **45**, 1917 (2006).
- [7] K. A. Serrels, E. Ramsay, R. J. Warburton, and D. T. Reid, *Nat. Photonics* **2**, 311 (2008).
- [8] G. Maire, F. Drsek, J. Girard, H. Giovannini, A. Talneau, D. Konan, K. Belkebir, P. C. Chaumet, and A. Sentenac, *Phys. Rev. Lett.* **102**, 213905 (2009).
- [9] Y. Ruan, P. Bon, E. Mudry, G. Maire, P. C. Chaumet, H. Giovannini, K. Belkebir, A. Talneau, B. Wattellier, S. Monneret *et al.*, *Opt. Lett.* **37**, 1631 (2012).
- [10] Y. Sung, W. Choi, C. Fang-Yen, K. B. Zadegan, R. R. Dasari, and M. S. Feld, *Opt. Express* **17**, 266 (2009).
- [11] W. Choi, C. Fang-Yen, K. Badizadegan, S. Oh, N. Lue, R. R. Dasari, and M. S. Feld, *Nat. Methods* **4**, 717 (2007).
- [12] T. Colomb, F. Dürr, E. Cuche, P. Marquet, H. G. Limberger, R.-P. Salathé, and C. Depeursinge, *Appl. Opt.* **44**, 4461 (2005).
- [13] The experiment was automatized using the free software OPTICSBENCHUI, <http://www.opticsbenchui.com/>.

-
- [14] G. Maire, J. Girard, F. Drsek, H. Giovannini, A. Talneau, K. Belkebir, P. C. Chaumet, and A. Sentenac, *J. Mod. Opt.* **57**, 746 (2010).
- [15] E. Mudry, P. C. Chaumet, K. Belkebir, G. Maire, and A. Sentenac, *Opt. Lett.* **35**, 1857 (2010).
- [16] K. Belkebir, P. C. Chaumet, and A. Sentenac, *J. Opt. Soc. Am. A* **22**, 1889 (2005).
- [17] E. Mudry, P. C. Chaumet, K. Belkebir, and A. Sentenac, *Inverse Probl.* **28**, 065007 (2012).