

High-throughput microfluidic single-cell RT-qPCR

Adam K. White^{a,1}, Michael VanInsberghe^{a,1}, Oleh I. Petriv^{a,b}, Mani Hamidi^a, Darek Sikorski^{a,c,d}, Marco A. Marra^e, James Piret^{c,d}, Samuel Aparicio^{f,g}, and Carl L. Hansen^{a,b,2}

^aCentre for High-Throughput Biology, University of British Columbia, Vancouver, BC, Canada V6T 1Z4; ^bDepartment of Physics and Astronomy, University of British Columbia, Vancouver, BC, Canada V6T 1Z4; ^cDepartment of Chemical and Biological Engineering, University of British Columbia, Vancouver, BC, Canada V6T 1Z4; ^dMichael Smith Genome Sciences Centre, BC Cancer Agency, Vancouver, BC, Canada V5Z 4S6; ^eMichael Smith Laboratories, University of British Columbia, Vancouver, BC, Canada V6T 1Z4; ^fCentre for Translational and Applied Genomics, Vancouver, BC, Canada V5Z 4E6; and ^gDepartment of Pathology, University of British Columbia, Vancouver, BC, Canada V6T 2B5

Edited by Robert H. Austin, Princeton University, Princeton, NJ, and approved June 17, 2011 (received for review December 28, 2010)

A long-sought milestone in microfluidics research has been the development of integrated technology for scalable analysis of transcription in single cells. Here we present a fully integrated microfluidic device capable of performing high-precision RT-qPCR measurements of gene expression from hundreds of single cells per run. Our device executes all steps of single-cell processing, including cell capture, cell lysis, reverse transcription, and quantitative PCR. In addition to higher throughput and reduced cost, we show that nanoliter volume processing reduced measurement noise, increased sensitivity, and provided single nucleotide specificity. We apply this technology to 3,300 single-cell measurements of (i) miRNA expression in K562 cells, (ii) coregulation of a miRNA and one of its target transcripts during differentiation in embryonic stem cells, and (iii) single nucleotide variant detection in primary lobular breast cancer cells. The core functionality established here provides the foundation from which a variety of on-chip single-cell transcription analyses will be developed.

real-time PCR | lab on a chip | transcriptional noise | microRNA | single-cell diagnostics

Single cells represent the fundamental unit of biology; however, the vast majority of biological knowledge has emerged as a consequence of studying cell populations and not individual cells. Inevitably, there are fundamental and applied questions, such as those relating to transcriptional control of stem cell differentiation, intrinsic noise in gene expression, and the origins of disease, that can be addressed only at the single-cell level. For example, single-cell analysis allows for the direct measurement of gene expression kinetics, or for the unambiguous identification of coregulated genes, even in the presence of desynchronization and heterogeneity that could obscure population-averaged measurements. Similarly, single-cell methods are vital in stem cell research and cancer biology, where isolated populations of primary cells are heterogeneous due to limitations in purification protocols, and it is often a minority cell population that is the most relevant. High-throughput single-cell measurement technologies are therefore of intense interest and have broad application in clinical and research settings.

Existing methods for measuring transcript levels in single cells include RT-qPCR (1), single molecule counting using digital PCR (2) or hybridization probes (3, 4), and next generation sequencing (5). Of these, single-cell RT-qPCR provides combined advantages of sensitivity, specificity, and dynamic range, but is limited by low throughput, high reagent cost, and difficulties in accurately measuring low abundance transcripts (6).

Microfluidic systems provide numerous advantages for single-cell analysis: economies of scale, parallelization and automation, and increased sensitivity and precision that comes from small volume reactions. Considerable effort over the last decade has been directed toward developing integrated and scalable single-cell genetic analysis on chip (7, 8). Thus, many of the basic functionalities for microfluidic single-cell gene expression analysis have been demonstrated in isolation, including cell manipulation and trapping (9, 10), RNA purification and cDNA synthesis

(11–13), and microfluidic qPCR (14) following off-chip cell isolation, cDNA synthesis, and preamplification. In particular, microfluidic qPCR devices (Biomark Dynamic Array, Fluidigm) have recently been applied to single-cell studies (15, 16). Although these systems provide a high-throughput qPCR read-out, they do not address the front end sample preparation and require single-cell isolation by FACS or micropipette followed by off-chip processing and preamplification of starting template prior to analysis. The critical step of integrating all steps of single-cell analysis into a robust system capable of performing measurements on large numbers of cells has yet to be reported. A single demonstration of an integrated device for directly measuring gene expression in single cells was described by Toriello et al., combining all steps of RNA capture, PCR amplification, and end-point detection of amplicons using integrated capillary electrophoresis (17). Despite the engineering complexity of this system, throughput was limited to four cells per run, cell capture required metabolic labeling of the cells, and the analysis was not quantitative. Thus, there remains an unmet need for microfluidic technologies capable of scalable and quantitative single-cell genetic analysis.

Here we describe an integrated microfluidic device for high-throughput RT-qPCR analysis of mRNA and miRNA expression at a throughput of hundreds of single cells per experiment. We show that this technology provides a powerful tool for scalable single-cell gene expression measurements with improved performance, reduced cost, and higher sensitivity as compared to analysis in microliter volumes. This technology represents the implementation of robust and high-throughput single-cell processing and amplification of nucleic acids on a chip, thereby achieving a major milestone in microfluidic biological analysis.

Results and Discussion

Device Design. An integrated microfluidic device that performs 300 parallel RT-qPCR assays and executes all steps of single-cell capture, lysis, reverse transcription, and qPCR is shown in Fig. 1A. To facilitate the precise comparison of different samples and cell types, our prototype consists of six independent sample-loading lanes, each containing 50 cell-processing units. We resolved previously limiting technical pitfalls by the inclusion of design elements to (i) allow for efficient distribution of single cells without mechanical damage, (ii) minimize background signal arising from free RNA or cell debris in the medium, and (iii) avoid reaction inhibition by cell lysates in nanoliter volumes.

Author contributions: M.A.M., J.P., S.A., and C.L.H. designed research; A.K.W., M.V., O.I.P., M.H., and D.S. performed research; A.K.W., M.V., O.I.P., M.H., and D.S. analyzed data; and A.K.W., M.V., and C.L.H. wrote the paper.

Conflict of interest statement: C.L.H. has a financial interest in Fluidigm Corporation, which has products related to the subject matter of this study.

This article is a PNAS Direct Submission.

¹A.K.W. and M.V. contributed equally to this work.

²To whom correspondence should be addressed. E-mail: chansen@phas.ubc.ca.

This article contains supporting information online at www.pnas.org/lookup/suppl/doi:10.1073/pnas.1019446108/-DCSupplemental.




Fig. 1. Design and operation of the microfluidic device for single-cell gene expression analysis. (A) Schematic of microfluidic device. Scale bar: 4 mm. The device features 6 sample input channels, each divided into 50 compound reaction chambers for a total of 300 RT-qPCR reactions using approximately 20 μ L of reagents. The rectangular box indicates the region depicted in B. (B) Optical micrograph of array unit. For visualization, the fluid paths and control channels have been loaded with blue and red dyes, respectively. Each unit consists of (i) a reagent injection line, (ii) a 0.6-nL cell capture chamber with integrated cell traps, (iii) a 10-nL reverse transcription (RT) chamber, and (iv) a 50-nL PCR chamber. Scale bar: 400 μ m. (C) Optical micrograph of two cell capture chambers with trapped single cells indicated by black arrows. Each trap includes upstream deflectors to direct cells into the capture region. Scale bar: 400 μ m. (D–I) Device operation. (D) A single-cell suspension is injected into the device. (E) Cell traps isolate single cells from the fluid stream and permit washing of cells to remove extracellular RNA. (F) Actuation of pneumatic valves results in single-cell isolation prior to heat lysis. (G) Injection of reagent (green) for RT reaction (10 nL). (H) Reagent injection line is flushed with subsequent reagent (blue) for PCR. (I) Reagent for qPCR (blue) is combined with RT product in 50 nL qPCR chamber. Scale bar for D–I: 400 μ m.

In order to reduce device complexity and obviate the need for RNA purification, we optimized our device to be compatible with commercially available assays that use “one-pot” RT-qPCR protocols requiring only the sequential addition of reagents into a single reaction vessel. Each cell-processing unit consists of a compound chamber, formed by a cell capture chamber connected sequentially to two larger chambers for RT and qPCR (Fig. 1B). This simple fluidic architecture allows the implementation of either heat lysis followed by two-step RT-qPCR (Fig. 1D–I), or chemical lysis followed by one-step RT-qPCR. A detailed description of device operation for each of these protocols is provided in *SI Materials and Methods*. All lanes are connected to a common feed channel that, following the completion of each reaction step, is used to inject the next reaction master mix through the upstream chambers, thereby diluting the intermediate product (cell lysate or cDNA) and assembling the next reaction mixture. This parallelization of reaction assembly in a microfluidic format ensures equal timing of all reaction steps and greatly reduces technical variability associated with pipetting and mixing steps in microliter volumes. Fluorescence measurements were performed to ensure the efficient and reproducible transfer of reactants at each step, showing that losses in sample transfer are below 5%. To minimize device expense and complexity, temperature control and fluorescence detection were performed using peripheral hardware including a CCD detector mounted above a flatbed thermocycler plate.

We designed our chamber volumes to ensure sufficient dilution between each processing step to avoid reaction inhibition while at the same time maintaining high template concentrations and assay sensitivity. Initial attempts to perform RT-qPCR in low nanoliter volumes were found to produce highly variable results, including nonspecific amplification and inconsistent detection of abundant transcripts (18). Cell lysate dilutions showed that reaction inhibition becomes significant at concentrations in excess of 0.2 cells/nL, or 10 cells/50 nL-reaction (Fig. 2D). On the other hand, RT-qPCR measurement noise has been shown

to become the dominant source of variability when starting at concentrations below one copy per 100 nL (6), illustrating that minimizing reaction volumes is critical for precise measurements on limited template. Finally, experiments in tubes were performed to determine that a dilution ratio of at least 5:1 (PCR mix:RT product) is optimum for PCR efficiency. We therefore designed our combined reactors to have an aggregate total volume of 60.6 nL, consisting of a 0.6-nL cell capture chamber, a 10-nL RT chamber, and a 50-nL qPCR chamber. These volumes allow for the reliable amplification of single molecules (Fig. 2A) and result in a final template concentration of 330 ng/mL when starting from a single-cell equivalent of RNA (20 pg). The use of larger volume RT and PCR chambers has the added advantage of reducing their surface-to-volume ratio, thereby minimizing reagent evaporation through the gas permeable device material (polydimethylsiloxane).

Another critical step toward integration was to efficiently distribute single cells toward each location on the array without mechanical damage. To achieve reproducible and deterministic loading of single cells into each array element, we engineered a hydrodynamic single-cell trap within each capture chamber. Cell traps consisting of a single cup structure (19) were found to be highly inefficient, capturing less than 0.1% of cells passing in close proximity to the center of the channel structure. To improve capture efficiency, we incorporated upstream deflectors, located 22.5 μ m from the trap, to focus cells into the central streamlines where capture is most efficient (Fig. 1C). Using these structures we were able to achieve high single-cell occupancy of array locations (Fig. 3A and B). Over eight separate experiments, a loading protocol of approximately 60 s (10^6 cells/mL, 20 nL/s per lane) resulted in the successful isolation of single cells in 1,518/1,700 chambers (89.3%), with a cell capture efficiency of $5.0 \pm 0.5\%$. Staining with trypan blue was used to assess the viability of cells after loading and was determined to be equivalent to the viability of the input sample (97.4% viability vs. input 96.8%). Finally, measurements of the distribution of cell diameters prior




Fig. 2. Precision and sensitivity of microfluidic RT-qPCR. (A) Fluorescence image of entire device showing 300 reactions in 6 lanes. Image is taken after 40 cycles of PCR from dilution series of purified total RNA from K562 cells. (Left to Right) The samples are 40 pg/chamber, 5 pg/chamber, 625 fg/chamber, 78 fg/chamber, 10 fg/chamber, and no-template control (NTC). Single molecule amplification at limiting dilution results in a digital amplification pattern for 10- and 78-fg lanes. No amplification is observed in NTC lane ($N = 50$). (B) Three hundred real-time amplification curves generated from processing sequences of images similar to A. The threshold for determining CT values is indicated by the dashed line. (C) On-chip (black) and off-chip (blue) RT-qPCR for GAPDH from a $8\times$ serial dilution of purified total RNA shows improved sensitivity in nanoliter volume reactions. In the microfluidic system, CT values for the 10-fg sample correspond to single molecule amplifications detected in 19 of 50 chambers. The mean and standard deviation from single-cell measurements is shown in green for both on- and off-chip analysis. CT values obtained on chip correspond to a mean of 20 pg of RNA per cell. Off-chip measurements of single K562 cells washed twice in PBS and isolated by glass capillary exhibit artificially increased levels due to residual signal from debris and free RNA in the supernatant (red). Cells were transferred in approximately 2 μL of supernatant, which was measured to contain approximately 20 pg of extracellular RNA. Error bars represent standard deviation of measured CT values for all amplified reactions. (D) Real-time amplification curves of GAPDH in K562 cell lysate dilutions. Inhibition of RT-PCR occurs at cell lysate concentrations beyond 10 cell equivalents per 50 nL reaction. (E) Measured CT values for GAPDH in dilution series of cell lysate. No inhibition occurs for single-cell lysates.

to and after loading indicated that cell trapping did not introduce significant bias ($p = 0.67$, two-sample t test) in selecting cells of different sizes (Fig. S1). This cell trap geometry and loading protocol were used in all subsequent qPCR measurements presented below. Further improvement of trap and deflector geometries were found to achieve fill factors of $>99\%$ (100 single cells captured out of 100 traps analyzed) and cell capture efficiencies of $87.0 \pm 4.5\%$, with cell viability again matching the input sample ($>98\%$) and not significantly biasing cell sizes ($p = 0.35$, two-sample t test), making this method applicable to the analysis




Fig. 3. Single-cell loading and transcript measurements. (A) The locations of cells in each chamber along all six lanes of a device, as determined by bright-field microscopy, are represented as white circles and overlaid on a heat map of CT values obtained from RT-qPCR measurements of GAPDH in K562 cells. Red circles indicate NTC. (B) Scatter plot showing CT measurements for experiment shown in A. Histogram (*Inset*) shows 93.2% single-cell occupancy. (C) Distribution of the number of GAPDH transcripts measured in single K562 cells ($N = 233$).

of limited quantity samples such as rare stem cells or clinical samples.

The immobilization of cells in traps was also used for on-chip washing of cells prior to lysis to remove free RNA, cellular debris, and untrapped cells that would otherwise give rise to background signal or result in low single-cell occupancy (Fig. S2A and B). The efficiency of chamber washing, determined by loading purified RNA template (36.5 ng/ μ L), followed by washing and RT-qPCR analysis, was >99.99% (1.1×10^4 copies measured without wash, 0 copies detected after washing) (Fig. S2C). In addition, RT-qPCR measurements testing different cell loading and washing protocols demonstrated that on-chip washing allows for loading directly from culture medium with low background as compared to off-chip wash steps followed by analysis in microliter volumes (Fig. 2C). Importantly, on-chip washing allows for lysis within seconds of washing, thereby minimizing spurious transcriptional responses that may arise from sequential medium exchange and spin steps.

Validation of Integrated Single-Cell RT-qPCR. We first tested the sensitivity and precision of RT-qPCR in our device by performing measurements of GAPDH expression over an 8-fold dilution series of total RNA, ranging from 40 pg (approximately 2 cell equivalents) to 10 fg (approximately 1/2,000 cell equivalents). RNA was purified from K562 cells, a BCR-ABL positive human cell line derived from a patient with chronic myeloid leukemia (20) (Fig. 2A–C). The efficiency of amplification was determined over the four highest template concentrations (40 pg, 5 pg, 625 fg, and 78.125 fg) as the slope from a linear least squares fit of $\log_2(C)$ vs. cycle threshold (CT) and was found to be 0.988 ± 0.055 . The standard deviation of CT values was less than 0.15 at the three highest concentrations (SD = 0.08, 0.10, and 0.14 for the 40 pg, 5 pg, and 625 fg samples, respectively), indicating uniform amplification across the array and technical error of less than 10% in absolute concentration, near the limit of qPCR precision. The highest measurement variability was observed in the 78-fg sample, where shot noise (Poisson sampling noise) is most pronounced and accounts for approximately 50% of the measurement variance. Template amounts below 625 fg resulted in a digital pattern characteristic of single

molecule amplification (49/50 for 78 fg and 19/50 for 10 fg) and consistent with the expected occupancy of chambers as determined by a binomial distribution (2). Based on the frequency of single molecule detection in the 10-fg sample, we measured the average copy number of GAPDH to be 979 ± 240 transcript copies per single-cell equivalent (20 pg) (Fig. 2). This measurement is comparable to previous reports (11) and is in close agreement with an independent estimate based on normalizing the dilution series to CT values obtained for single molecules $(\text{copies}/20 \text{ pg} = 1/2 \times \text{copies}/40 \text{ pg} = 1/2 \times (1 + \text{efficiency})^{(\text{CT}(40 \text{ pg}) - \text{CT}(\text{single molecule}))}) = 1,407 \pm 153 \text{ copies}/20 \text{ pg})$. It should be noted that these estimates represent a lower bound because they do not account for RT efficiency; the RT efficiency of GAPDH has been previously estimated to be approximately 50% (13) but is dependent on transcript secondary structure and assay design. A comparison of CT values obtained from on-chip qPCR from cDNA synthesized off-chip demonstrated that on-chip RT efficiency is equal to that obtained off-chip when working from the same RNA concentrations (Fig. S3). Finally, comparison of the same dilution series of RNA, assayed for GAPDH both on-chip and in tubes (20- μL volume) (Fig. 2C), showed that on-chip analysis provides improved sensitivity.


We next evaluated the efficiency and reliability of on-chip cell processing by comparing our GAPDH measurements of purified RNA to measurements performed directly from single K562 cells (Figs. 2C and 3C). K562 cells were loaded directly from culture medium followed by washing and analysis using a chemical lysis and one-step RT-qPCR protocol (Cells DirectTM, Invitrogen). Using a CT threshold of 31.5, corresponding to the mean CT of a single molecule of GAPDH (CT = 30.5) plus two standard deviations (SD = 0.5), we observed successful amplification in 100% of single cells ($N = 233$) (Fig. 3A and B). Adjacent chambers that did not contain a cell were clearly separated from single-cell measurements with an average Δ CT value of 5.7 (five empty chambers, three of which amplified) (Fig. 3A and B and Fig. S4). Consistent with previous reports (21), we observed a log-normal distribution of GAPDH in single cells with mean CT values of 20.3 (SD = 0.8) and an average of 1,761 (SD = 648) copies per cell (Fig. 3C). These expression levels are consistent with previous estimates in single cells (11). Additionally, the mean CT of 20.3 observed for single cells matches measurements of single-cell equivalent lysate (CT = 20.2, Fig. 2D). Using digital PCR on cDNA prepared from K562 cell lysate, we measured an average of $1,229 \pm 72$ GAPDH molecules per single-cell equivalent. We conclude that the relative efficiency of on-chip single-cell lysis and mRNA extraction/accessibility is equal to that achieved when working from RNA purified from large numbers of cells. Finally, as expected, RT-qPCR measurements from chambers loaded with more than one cell show reduced variability and lower CT values (Figs. S24 and S5). Taken together, these results establish the precise measurement of mRNA abundance with single molecule sensitivity and the dynamic range needed for single-cell analysis.

Application to Measurement of Single-Cell miRNA Expression. We next applied our technology to the study of single-cell miRNA expression. miRNAs are thought to provide a unique signature of cellular state and are central players in orchestrating development and oncogenesis, making them a promising class of biomarker for single-cell analysis (15, 22, 23). Importantly, the short length of miRNAs (approximately 22 nucleotides) makes them difficult to detect by hybridization approaches, so that RT-qPCR is the dominant quantification strategy. To demonstrate the robustness and throughput of our technology, we performed a total of 1,672 single-cell measurements to examine single-cell variability in the expression of nine miRNAs spanning a wide range of abundance (>16,000 copies per cell to <0.2 average copies per cell). K562 cells were again chosen as a heterogeneous population for

this study because they are known to exhibit mixed characteristics of erythrocytes, granulocytes, and monocytes (20, 24). We first measured the expression of miR-16, a highly expressed microRNA that is found in many tissue types (25) and has been suggested as a suitable internal standard for normalization (26). We found that miR-16 was log-normally distributed across K562 cells, but with slightly lower expression and notably tighter regulation than GAPDH, having an average of 804 (SD = 261) copies per cell and a standard deviation of 30% (mean CT = 21.4, SD = 0.4). This strikingly low variability is within our estimates of cell volume differences (Fig. S1). Matched experiments on single cells, isolated by micropipette into 20- μL volume tubes displayed an increase in measurement variability to approximately 90% (mean CT = 29.5, SD = 0.9), demonstrating the improved precision of parallel microfluidic cell processing in nanoliter volumes (Fig. 4A). Microliter volume experiments also showed a pronounced increase in measured CT values that results from the low concentration of template and the large number of required PCR cycles.

To demonstrate the utility of our device for measuring differential expression in single cells, we next measured the expression of miR-223, a miRNA implicated in myeloid differentiation (24, 27). In contrast to miR-16, K562 cell miR-223 expression was found to be highly variable (mean CT = 22.2, SD = 1.6, copy number = 513, SD = 406) and was not log-normally distributed (Fig. 4B), consistent with the known functional heterogeneity of K562 cells. These measurements highlight the utility of single-cell miRNA expression analysis for assessing the heterogeneity of cell populations and for identifying miRNAs that are useful biomarkers of cellular state. To further explore this possibility, we measured the expression of an additional seven miRNAs (nine total) and plotted the patterns of single-cell expression in K562 populations (Fig. 4C). Following the procedure described above, we used single molecule CT values, obtained by digital PCR, to translate measured CT values to absolute copy number. Assuming 100% efficient amplification, we observed that the copy number, calculated as $2^{(\text{CT}(\text{single cell}) - \text{CT}(\text{single molecule}))}$, was well correlated (coefficient of 0.9932) with the average copy number obtained by digital PCR of cDNA prepared from bulk lysates (Fig. 4D). Single-cell measurements revealed distinct patterns of miRNA expression, with miR-16, miR-92, and miR-17-5p each exhibiting unimodal and tightly regulated distributions, whereas miR-223, miR-196a, and miR-145 showed multimodal distributions and a high level of cellular heterogeneity. Notably, for the lowest abundance miRNA, miR-200a, we detected expression in only a small fraction of cells and at levels below approximately five copies per cell. The average miR-200a copy number over all cells was within a factor of two of that obtained by digital PCR (0.2 copies per cell). In contrast, miR-92 was found to be the most abundant miRNA and was present at approximately 16,000 copies per cell. These measurements established miRNA quantification in single cells with a dynamic range of greater than 10^4 and at single molecule sensitivity.

Finally, to illustrate the utility of single-cell measurements in precisely assessing differences in both the average expression and the heterogeneity between two different cell populations, the expression levels of miR-16 and miR-223 in K562 cells were compared to those in CA1S cells (28, 29), a human embryonic stem cell line (hESC). Although miR-16 was found to be expressed in hESC at similar levels to K562 ($\Delta\text{CT} = 0.6$), we observed approximately a twofold greater variability in expression (mean CT = 22.0, SD = 0.7) (Fig. 4A). In contrast, when compared to K562, single CA1S cell measurements of miR-223 showed strong down-regulation, with miR-223 detected in only 3.6% of cells. The absence of significant miR-223 expression in hESC is expected due to the role of miR-223 as a differentiation-specific miRNA (24, 27).



from the primary sample, we found the ratio of mutant to wild-type SP1 alleles to be $18.7 \pm 2.3\%$, in agreement with the previously reported ratio of 21.9%, obtained by deep sequencing (31). The lower frequency of cells expressing the mutant SP1 allele may be due to allelic expression bias or an amplification of the SP1 mutant allele, both of which are supported by Shah et al. (31). Regardless, given that the frequency of tumor cells within the original sample was approximately 89% (31), both DNA molecule counting and single-cell RNA expression measurements show that the metastasis of this tumor is derived from multiple cancer cell lineages.

A

log₁₀ Oct4 Copy Number

ND

• Day 0
• Day 4
• Day 6
• Day 8

B

log₁₀ miRNA145 Copy Number

C

log₁₀ miRNA145 Copy Number

D

Wildtype Expression (C_t)

Mutant Expression (C_t)

• K562 Cells
• Primary Cells

94.6% 0.0%
72.6% 18.8%

5.4% 0.0%
7.7% 0.9%

ND 30 28 26 24 22

PNAS Early Edition | 5 of 6

Conclusion

Here we have demonstrated the implementation of scalable and quantitative single-cell gene expression measurements on an integrated microfluidic system. The presented device performs 300 high-precision single-cell RT-qPCR measurements per run, surpassing previous microfluidic systems by a factor of approximately 100 in throughput. Further scaling the throughput to over 1,000 measurements on a device with an area of one square inch is straightforward as each array element occupies an area of 0.6 mm². In terms of performance, we have established a dynamic range of at least 10⁴, measurement precision of better than 10%, single molecule sensitivity, and specificity capable of discriminating the relative abundance of alleles differing by a single nucleotide. Compared to tube-based single-cell RT-qPCR, microfluidic processing provides improved reproducibility, precision, and sensitivity, all of which may be critical in identifying subtle differences in cell populations. Nanoliter volume also results in a 1,000-fold reduction in reagent consumption, thereby enabling cost-effective analysis of large numbers of single cells.

In over 3,300 single-cell experiments, using adherent and suspension cell lines as well as clinical samples, we have shown that microfluidic RT-qPCR is well-suited to the quantitative analysis of miRNA expression and SNV detection, both of which are difficult or inaccessible by alternative hybridization methods. Notably, our device allowed for precise comparison of the distributions of GAPDH and miR-16 expression. miR-16 was found to be exquisitely regulated in K562 cells, a finding that is striking given the known functional heterogeneity of this population and the high variability in the expression of other measured miRNAs. We postulate that higher variability of GAPDH expression reflects the fundamentally stochastic process of transcriptional bursts followed by mRNA degradation. Incorporation of miRNA into the RNA-induced silencing complex (RISC) is known to provide enhanced stability so that miRNA are inherently less subject

to temporal fluctuations; miRNA are thus particularly suited as biomarkers for assessing single-cell state and population heterogeneity. We anticipate that scalable and precise single-cell miRNA analysis will become an invaluable tool in stratifying populations of mixed differentiation state (15).

Here we have established the critical element of combining all single-cell-processing steps into an integrated platform. This functionality provides a solid foundation upon which increasingly advanced microfluidic single-cell transcription analysis may be built. We anticipate that more complex fluid routing (32), to distribute cell contents across multiple chambers, will allow for the multiplexed measurements of tens of targets across hundreds of cells, and for combining this technology with single molecule detection by digital PCR. Alternatively, the microfluidic system described here could be used for single-cell processing and preamplification, with recovered reaction products analyzed by high-throughput microfluidic qPCR or sequencing. We contend that the simplicity of device operation will soon allow for the robust and automated implementation of single-cell RT-qPCR, leading to its widespread adoption in research applications and opening the prospect of diagnostic tests based on single-cell analysis.

Materials and Methods

Experimental details and protocols for device fabrication, device operation, on-chip and off-chip RT-qPCR, cell culture, hESC differentiation, digital PCR, mRNA-FISH, image analysis, transfer efficiency measurements, and cell trapping measurements are provided in [SI Materials and Methods](#).

ACKNOWLEDGMENTS. This research was funded by Genome BC, Genome Canada, Western Diversification, Canadian Institutes of Health Research (CIHR), Terry Fox Foundation, and the Natural Sciences and Engineering Research Council (NSERC). Salary support was provided by the Michael Smith Foundation for Health Research (to A.K.W. and C.L.H.), NSERC (to A.K.W., M.V., M.H., and D.S.), and CIHR (to C.L.H.).

1. Taniguchi K, Kajiya T, Kambara H (2009) Quantitative analysis of gene expression in a single cell by qPCR. *Nat Methods* 6:503–506.
2. Warren L, Bryder D, Weissman IL, Quake SR (2006) Transcription factor profiling in individual hematopoietic progenitors by digital RT-PCR. *Proc Natl Acad Sci USA* 103:17807–17812.
3. Raj A, van den Bogaard P, Rifkin SA, van Oudenaarden A, Tyagi S (2008) Imaging individual mRNA molecules using multiple singly labeled probes. *Nat Methods* 5:877–879.
4. Larsson C, Grundberg I, Soderberg O, Nilsson M (2010) In situ detection and genotyping of individual mRNA molecules. *Nat Methods* 7:395–397.
5. Tang F, et al. (2009) mRNA-Seq whole-transcriptome analysis of a single cell. *Nat Methods* 6:377–382.
6. Bengtsson M, Hemberg M, Rorsman P, Stahlberg A (2008) Quantification of mRNA in single cells and modelling of RT-qPCR induced noise. *BMC Mol Biol* 9:63–74.
7. Zare RN, Kim S (2010) Microfluidic platforms for single-cell analysis. *Annu Rev Biomed Eng* 12:187–201.
8. Sims CE, Allbritton NL (2007) Analysis of single mammalian cells on-chip. *Lab Chip* 7:423–440.
9. Wheeler AR, et al. (2003) Microfluidic device for single-cell analysis. *Anal Chem* 75:3581–3586.
10. Skelley AM, Kirak O, Suh H, Jaenisch R, Voldman J (2009) Microfluidic control of cell pairing and fusion. *Nat Methods* 6:147–152.
11. Marcus JS, Anderson WF, Quake SR (2006) Microfluidic single-cell mRNA isolation and analysis. *Anal Chem* 78:3084–3089.
12. Bontoux N, et al. (2008) Integrating whole transcriptome assays on a lab-on-a-chip for single cell gene profiling. *Lab Chip* 8:443–450.
13. Zhong JF, et al. (2008) A microfluidic processor for gene expression profiling of single human embryonic stem cells. *Lab Chip* 8:68–74.
14. Guo GJ, et al. (2010) Resolution of cell fate decisions revealed by single-cell gene expression analysis from zygote to blastocyst. *Dev Cell* 18:675–685.
15. Petriv OI, et al. (2010) Comprehensive microRNA expression profiling of the hematopoietic hierarchy. *Proc Natl Acad Sci USA* 107:15443–15448.
16. Diehn M, et al. (2009) Association of reactive oxygen species levels and radioresistance in cancer stem cells. *Nature* 458:780–783.
17. Toriello NM, et al. (2008) Integrated microfluidic bioprocessor for single-cell gene expression analysis. *Proc Natl Acad Sci USA* 105:20173–20178.
18. Gong Y, Ogunniyi AO, Love JC (2010) Massively parallel detection of gene expression in single cells using subnanolitre wells. *Lab Chip* 10:2334–2337.
19. Di Carlo D, Aghdam N, Lee LP (2006) Single-cell enzyme concentrations, kinetics, and inhibition analysis using high-density hydrodynamic cell isolation arrays. *Anal Chem* 78:4925–4930.
20. Lozzio BB, Lozzio CB, Bamberger EG, Feliu AS (1981) A multipotential leukemia cell line (K-562) of human origin. *Proc Soc Exp Biol Med* 166:546–550.
21. Bengtsson M, Stahlberg A, Rorsman P, Kubista M (2005) Gene expression profiling in single cells from the pancreatic islets of Langerhans reveals lognormal distribution of mRNA levels. *Genome Res* 15:1388–1392.
22. Lu J, et al. (2005) MicroRNA expression profiles classify human cancers. *Nature* 435:834–838.
23. Bartel DP (2004) MicroRNAs: Genomics, biogenesis, mechanism, and function. *Cell* 116(2):281–297.
24. Yuan JY, et al. (2009) MicroRNA-223 reversibly regulates erythroid and megakaryocytic differentiation of K562 cells. *J Cell Mol Med* 13:4551–4559.
25. Landgraf P, et al. (2007) A mammalian microRNA expression atlas based on small RNA library sequencing. *Cell* 129:1401–1414.
26. Tang F, Hajkova P, Barton SC, Lao K, Surani MA (2006) MicroRNA expression profiling of single whole embryonic stem cells. *Nucleic Acids Res* 34:e9.
27. Chen CZ, Li L, Lodish HF, Bartel DP (2004) MicroRNAs modulate hematopoietic lineage differentiation. *Science* 303:83–86.
28. Schulze HG, et al. (2010) Assessing differentiation status of human embryonic stem cells noninvasively using Raman microspectroscopy. *Anal Chem* 82:5020–5027.
29. Adewumi O, et al. (2007) Characterization of human embryonic stem cell lines by the International Stem Cell Initiative. *Nat Biotechnol* 25:803–816.
30. Xu N, Papagiannakopoulos T, Pan G, Thomson JA, Kosik KS (2009) MicroRNA-145 regulates OCT4, SOX2, and KLF4 and represses pluripotency in human embryonic stem cells. *Cell* 137:647–658.
31. Shah SP, et al. (2009) Mutational evolution in a lobular breast tumour profiled at single nucleotide resolution. *Nature* 461:809–813.
32. Huft J, Da Costa D, Walker D, Hansen CL (2010) Three-dimensional large-scale microfluidic integration by laser ablation of interlayer connections. *Lab Chip* 10:2358–2365.

Supporting Information

White et al. 10.1073/pnas.1019446108

SI Materials and Methods

Device Fabrication and Operation. Microfluidic devices were fabricated by multilayer soft lithography (1, 2). Planar silicon molds were defined by photolithography, using photomasks designed with CAD software (AutoCAD, Autodesk Inc.) and printed on transparency films at a resolution of 20,000 dots per inch (CAD/Art services). The “control” mold was fabricated using SU8-2025 photoresist (Microchem) to deposit valve features 24 μm in height. The “flow” mold was fabricated with three lithographic steps. First, the channels for reagent injection, and connections between chambers were fabricated using 13- μm high SPR220-7 photoresist (Shipley). The SPR220-7 channels were rounded to facilitate valve closure by incubation at 115 $^{\circ}\text{C}$ for 15 min. A hard bake at 190 $^{\circ}\text{C}$ for 2 h was used to prevent SPR photoresist erosion during addition of subsequent layers. Second, the cell trap features were defined in 14- μm SU8-2010 photoresist (Microchem). Finally, the large chambers and fluidic bus lines were constructed using 150- μm high SU8-100 photoresist. All photoresist processing was performed according to manufacturer specifications. All molds were fabricated on 4-inch silicon wafers (Silicon Quest International).

Microfluidic devices were cast from these molds in polydimethylsiloxane (PDMS, RTV615, General Electric). Each device consists of a three-layer elastomeric structure with a blank bottom layer, a middle control layer with channels that act as valves by pushing up and pinching closed channels in the above flow layer. The molds were first treated with chlorotrimethylsilane (TMCS, Aldrich) vapor for 2 min to prevent PDMS from bonding to the photoresist structures. The flow layer was made by pouring a mixture of PDMS (5 parts RTV615A:1 part RTV615B) onto the flow mold, degassing, and then baking for 60 min at 80 $^{\circ}\text{C}$. A thin control layer was made by spin coating the control mold with PDMS (20 parts RTV615A:1 part RTV615B) at 1,800 rpm and baking for 45 min at 80 $^{\circ}\text{C}$. After baking, the PDMS of the flow layer was peeled from the flow mold and aligned to the control layer. Following a 60-min bake at 80 $^{\circ}\text{C}$, the bonded two-layer structure was separated from the control mold, and channel access holes were punched. A blank layer (without channels) was prepared by spinning PDMS (20 parts RTV615A:1 part RTV615B) on a blank wafer (2,000 rpm) and baking 45 min at 80 $^{\circ}\text{C}$. The bonded flow and control structure was mounted onto the blank layer and baked for 3 h at 80 $^{\circ}\text{C}$. Finally, the three-layer bonded structure was removed from the blank mold, diced into individual devices, and these were each bonded to clean glass slides by baking overnight at 80 $^{\circ}\text{C}$.

The device operation requires control of nine pneumatic valves and may be operated using a simple manifold of manual valves. For the current study a semiautomated implementation was used in which microfluidic valves were controlled by solenoid actuators (Fluidigm Corp.) controlled through a digital input-output card (NI-DAQ, DIO-32H, National Instruments) operated using LabView drivers (National Instruments). Tygon tubing connected the solenoids to the microfluidic device by 20-gauge stainless steel pins (Small Parts Inc.) fitted into the control line ports. Krytox (DuPont) oil was used as the fluid in the control lines, and the valves were actuated with 30-psi pressure.

Microfluidic Single-Cell RT-qPCR. The device was designed to be compatible with commercially available RT-qPCR products. A protocol for heat lysis, followed by a two-step RT-qPCR was used with miRNA and OCT4 mRNA assays. Alternatively, a chemical

lysis, followed by one-step RT-qPCR, was used for mRNA measurements of single nucleotide variants (SNVs) and GAPDH.

Single-Cell Transcript Measurements by Heat Lysis and Two-Step RT-qPCR. The device was primed by flowing PBS containing 0.5 mg/mL BSA and 0.5 U/ μL RNase inhibitor through all channels, while keeping the reverse transcription (RT) and PCR chambers empty and isolated by valves. The BSA helped prevent cells from adhering to channel walls. After priming, but prior to cell loading, all valves were closed. A single-cell suspension was injected into the device by applying pressure (approximately 2–3 psi) to microcapillary pipette tips plugged into the sample inlets. The sample inlets were first dead-end filled against an inlet valve to prevent air bubbles from entering the device. The sample inlet valves, cell chamber valves, and outlet valve were opened to allow the cell suspension to flow through the sample channels. Cells were loaded into the device suspended in culture media (directly from culture). Cell loading concentrations were kept between 5×10^5 and 1×10^6 cells/mL, resulting in over 80% occupancy of cell traps with single cells in 1–2 min at a flow rate of approximately 20 nL/s. Lower concentrations were found to require proportionately longer times to achieve high occupancy of trapped single cells. Concentrations greater than 2×10^6 cells/mL were found to occasionally clog the inlet port or the channel at trap locations. A peristaltic pump was integrated into the device for controlling the flow rate; however, pressure-driven flow was used for the current study.

After injecting the cell suspension and trapping single cells the cell sample inlet valve was closed, and the cells were washed by flushing the line with the PBS solution. This removed untrapped single cells, extracellular RNA, and debris. Following on-chip washing, the cell chamber valves were closed to partition the cell loading channel and isolate individual cell reactors. Visual inspection of the cell-capture chambers under a microscope was used to confirm and count the number of cells in each chamber. The cells were lysed by placing the microfluidic device onto a flatbed thermocycler and heating it to 85 $^{\circ}\text{C}$ for 7 min (and then cooled to 4 $^{\circ}\text{C}$).

RT was performed in the device by using the ABI High Capacity Reverse Transcription kit (3), with the addition of a surfactant to prevent adsorption of nucleic acids and proteins to PDMS surfaces (2 μL 10 \times Reverse Transcription Buffer, 4 μL 5 \times RT stem-loop miRNA primer from ABI, 1 μL 100 mM dNTPs, 1.34 μL of 50 U/ μL Multiscribe Reverse Transcriptase, 0.26 μL of 20 U/ μL RNase Inhibitor, 2 μL 1% Tween 20, 9.4 μL PCR grade water). The RT mix was loaded into the device and flushed through the reagent injection channels. RT reagent was injected into the reaction by opening the valve connecting the cell chamber to the RT chamber, and the valve connecting the cell chamber to the reagent injection line. The RT chamber was dead-end filled before closing the connection to the reagent injection line. A pulsed temperature RT protocol was carried out by placing the microfluidic device on a flatbed thermocycler (2 min at 16 $^{\circ}\text{C}$, followed by 60 cycles of 30 s at 20 $^{\circ}\text{C}$, 30 s at 42 $^{\circ}\text{C}$, and 1 s at 50 $^{\circ}\text{C}$). RT enzyme was inactivated at 85 $^{\circ}\text{C}$ (5 min), and then the device was cooled to 4 $^{\circ}\text{C}$.

The PCR reagent was prepared with 25 μL of 2 \times TaqMan Universal Master Mix (ABI), 2.5 μL 20 \times Real-Time miRNA assays (primers and probe, ABI), 5 μL of 1% Tween 20, and 7.5 μL of PCR grade water. The PCR reagent was flowed through the reagent injection channels to flush away the RT reagent. Valves were opened and the PCR reagent was injected to dilute the

RT product into the PCR reaction chamber. After completely filling the PCR reaction chamber, the valves closing the PCR chambers were actuated, and the device was transferred to an enclosure for real-time PCR (Prototype version of Biomark™ Instrument, Fluidigm). The real-time PCR enclosure consists of a custom flatbed thermocycler, a xenon arc lamp and filter set,

and a CCD imager with optics for fluorescent imaging of the entire device periodically during PCR thermocycling (see description of real-time PCR instrumentation below). PCRs were thermocycled with the following conditions: 10 min at 95 °C, followed by 50 cycles of 15 s at 95 °C and 1 min at 60 °C. Images were acquired at 60 °C.

Protocol timing for performing heat lysis and two step RT-qPCR in the microfluidic system

Step	Description	Time
1	Prime device with PBS 0.5 mg/mL BSA and 0.5 U/μL RNase inhibitor	1 min
2	Inject cell suspension (passive cell trapping)	1 min
3	On-chip cell washing with PBS containing 0.5 mg/mL BSA and 0.5 U/μL RNase inhibitor	1 min
4	Close valves partitioning cell loading channel and isolating single cells	30 s
5	Count cells by visual inspection with microscope	7 min
6	Heat lysis by placing device on flatbed thermocycler and heating to 85 °C	7 min
7	Flush fluidic bus and reagent injection lines with reagent for RT	2 min
8	Inject RT reagent through the cell-capture chamber, dead-end filling the 10 nL RT chamber	1 min
9	Close reagent injection valve, creating isolated reactors combining the cell-capture chamber and RT chamber	30 s
10	Perform reverse transcription (pulsed temperature protocol) by placing device on flatbed thermocycler	2.5 h
11	Flush fluidic bus and reagent injection lines with reagent for PCR	2 min
12	Inject PCR reagent through combined cell-capture/RT chamber into 50 nL PCR chamber.	5 min
13	Close valve to PCR chamber. Allow for mixing by diffusion	40 min
14	Load device into BioMark real-time PCR system and focus camera	5 min
15	Run qPCR protocol	

Single-Cell Transcript Measurements by Chemical Lysis and One-Step RT-qPCR. Measurements of mRNA transcripts (SP1, GAPDH) were performed using the Cells Direct kit (Invitrogen). Operation of the microfluidic device for chemical lysis and one-step RT-qPCR was similar to the methods described for heat lysis and two-step RT-qPCR with several distinctions. The device was primed and cells were washed with PBS containing 0.5 mg/mL BSA. Additional RNase inhibitor was omitted as the chemical lysis buffer (10 μL lysis resuspension buffer, 1 μL lysis enhancer solution, Invitrogen, USA) contained RNA stabilizing agents. Cell loading was the same as in the heat lysis and two-step RT-qPCR scenario. Single cells were lysed by injecting a chemical lysis buffer through the cell-capture chamber and filling

the 10-nL chamber (used for RT reagent injection in the two-step protocol). The lysis reaction was incubated at room temperature for 10 min, followed by heat inactivation of the lysis reagent by placing the device on a flatbed thermocycler and incubating at 70 °C for 10 min. The one-step RT-qPCR mix [1 μL of SuperScript III RT/Platinum Taq Mix, 25 μL of 2× Reaction Mix (with ROX reference dye), 2.5 μL of 20× TaqMan Assay (primers and probes, ABI), 1 μL of 50 mM MgSO₄, 5.5 μL of H₂O, and 5 μL of 1% Tween 20] was then combined with the cell lysate into the final 50-nL reaction chamber. The device was transferred to the real-time PCR enclosure for temperature control and imaging of the one-step RT-qPCR (20 min at 50 °C for RT, followed by a hot-start at 95 °C for 2 min, and 50 cycles of 15 s at 95 °C and 30 s at 60 °C).

Protocol timing for performing chemical lysis and one step RT-qPCR in the microfluidic system

Step	Description	Time
1	Prime device with PBS 0.5 mg/mL BSA and 0.5 U/μL RNase inhibitor	1 min
2	Inject cell suspension (passive cell trapping)	1 min
3	On-chip cell washing with PBS containing 0.5 mg/mL BSA and 0.5 U/μL RNase inhibitor	1 min
4	Close valves partitioning cell loading channel and isolating single cells	30 s
5	Count cells by visual inspection with microscope	7 min
6	Flush fluidic bus and reagent injection lines with reagent for lysis	2 min
7	Inject lysis reagent through the cell-capture chamber, dead-end filling the 10-nL chamber	1 min
8	Close reagent injection valve, creating isolated reactors combining the cell-capture chamber and lysis reservoir chamber	30 s
9	Perform lysis at room temperature and heat inactivation of the lysis reagent at 75 °C by placing device on flatbed thermocycler	25 min
10	Flush fluidic bus and reagent injection lines with reagent for RT-qPCR	2 min
11	Inject RT-qPCR reagent through combined cell-capture/lysis chamber into 50 nL RT-qPCR chamber.	5 min
12	Close valve to RT-qPCR chamber. Allow for mixing by diffusion	40 min
13	Load device into BioMark real-time PCR system and focus camera	5 min
14	Run RT-qPCR protocol	

Digital PCR Experiments. For mRNA digital PCR analysis cells were washed with PBS containing 0.5 mg/mL BSA, lysed in chemical lysis buffer, reverse transcription was performed in tubes according to the protocol described above, and the resulting cDNA product was loaded into digital PCR arrays. For miRNA studies, cells were lysed in PBS containing 0.5 mg/mL BSA and 0.5 U/μL RNase inhibitor. Reverse transcription was performed using miRNA stem-loop primers (Applied Biosystems) and the High Capacity cDNA Reverse Transcription kit (Applied Biosystems) in 10-μL volumes. Prior to injection into microfluidic digital PCR arrays, RT product was added to the PCR reagent as in the on-chip two-step RT-qPCR protocol described above. Thermal cycling of digital PCR arrays was also performed using the same protocols as described above. PDMS digital PCR arrays consist-

ing of 765 2-nL individual PCR chambers, of similar design to those described in Warren et al. (4), were fabricated by multilayer soft lithography. After thermal cycling, positive chambers were counted and actual molecule numbers were derived based on the binomial distribution.

System for Real-Time PCR. The BioMark™ Reader is a commercially available real-time PCR instrument developed by Fluidigm and designed to run Fluidigm Integrated Fluidic Circuits (IFCs). The prototype version of this system allowed access to the flatbed thermocycler inside the enclosure, permitting the use of custom microfluidic devices in addition to the intended commercial IFCs. Fundamental specifications for data collection include the following:

Specifications for collecting qPCR images

Image resolution and bit depth:	4 Megapixel, 16 bit
Filters:	FAM: Ex 485 ± 20, Em 525 ± 25; CAL: Ex 530 ± 20, Em 570 ± 30; ROX: Ex 580 ± 25, Em 610 ± 15; QUASAR: Ex 580 ± 25, Em 680 ± 25
Light source:	175-W xenon arc bulb

RT-qPCR Assays. Measuring mRNA in the presence of genomic DNA requires primers designed to specifically target mature mRNA sequences. In many cases, this can be accomplished by designing intron-spanning primers. A specially designed stem-loop RT primer system (Applied Biosystems) is used for the specific targeting of mature miRNAs.

TaqMan assays for GAPDH (Applied Biosystems, Assay ID Hs99999905_m1) and miRNAs were obtained from Applied Biosystems. For GAPDH, a control experiment omitting the reverse transcriptase was performed off chip, in microliter volumes with bulk cell lysate (at equivalent concentration of a single cell on chip, 10^5 cells/mL), and showed no amplification after 40 cycles of PCR.

OCT4 (POU5F1) primer sequences were obtained from RTPrimerDB* and synthesized by Biosearch Technologies Inc; Forward primer: ACC CAC ACT GCA GCA GAT CA, Reverse primer: CAC ACT CGG ACC ACA TCC TTC T, Probe: Quasar⁶⁷⁰-CCA CAT CGC CCA GCA GCT TGG-BHQ-2, RT primer: TTG TGC ATA GTC GCT GCT TGA T. Measurement of OCT4 in single hESCs by microfluidic RT-qPCR without reverse transcriptase showed no amplification after 40 cycles of PCR.

BHQ-Plus probes with enhanced duplex stabilization (Biosearch Technologies Inc.) were used for SNV detection to allow for shorter sequence lengths and increased specificity. The SNV location for the SP1 locus was selected from table 2 in Shah et al. (5). Two hundred base pairs flanking this location on the hg18 sequence were used for assay design using Primer3. The resulting primer and probe sequences were as follows (the SNV is underlined):

SP1 Mutant Probe: FAM-AGGCCAGCAAAAACAAGG-BHQ-1

5' Modification: FAM, 3' Modification: BHQ-1 Plus. Tm = 62.7 °C

SP1 WT probe: Cal Fluor-CAGGCCAGCAAAAAGAA-BHQ-1

5' Modification: CAL Fluor Orange 560, 3' Modification: BHQ-1 plus. Tm = 62.1 °C

SP1 Forward Primer: CCAGACATCTGGAGGCTCATTTG Tm = 65.8 °C

SP1 Reverse Primer: TGAAGTAGCTGAGGCTGGATA Tm = 66.0 °C

Control experiments without reverse transcriptase showed positive amplification. Therefore the measurement of SP1 mutant and wild-type abundance in single cells by RT-qPCR does not discriminate between mature mRNA transcripts and genomic DNA.

Image Analysis. Fluorescence images of the entire device were taken in at least two different colors (one passive reference dye and one or more reporter dyes) after each PCR cycle and were analyzed using custom scripts written in MATLAB (MathWorks) to generate real-time amplification curves. Reaction chambers were segmented from the rest of the image using the first image of the passive reference dye. The image was manually rotated so that all of the reaction chambers were square with the edges of the image. Next, the average image intensities

across each row and column were calculated and a threshold was manually set to differentiate bright areas from background. Regions containing both bright rows and bright columns were assigned to the reaction chambers.

All subsequent images were automatically aligned to this initial image by minimizing the absolute distance between the average row and column intensities of the initial image, and the one being analyzed. For each image, the intensities of the reporter and passive dyes were recorded for each reaction chamber. Real-time amplification curves were generated by normalizing the intensity of each reporter dye to that of the passive dye. Linear components were removed from these curves by fitting the equation of a line to the preexponential region and extrapolating and subtracting the result from the entire curve. The threshold for determining CT values was automatically determined as the median normalized fluorescence value at the maximum second derivative of all amplification curves.

mRNA-FISH. Cells grown on LABTEK chambered cover glass were washed with PBS, fixed in 4% formaldehyde for 10 min at room temperature and permeabilized in 70% EtOH at 40 °C overnight. The next day cells were rinsed with wash buffer (15% formamide in 2× SSC) and then hybridized with the appropriate dilution of mRNA-FISH probes specific to OCT4 (Table S1) in hybridization solution (dextran sulfate, yeast tRNA, vanadyl ribonucleoside complex (New England Biolabs), BSA, 15% formamide in 2× SSC) overnight at 30 °C. The next morning the OCT4 hybridization solution was aspirated and cells were sequentially rinsed and incubated with wash buffer at 30 °C for 30 min, then washed with 2× SSC. One drop (25 µL) of Slowfade GOLD antifade reagent with DAPI was then added to the cells, covered immediately with a coverslip, and imaged. Stacks of 32–64 mRNA hybridization images (spaced by 0.5 µm) were acquired for each cell using a Leica DMI 6000B inverted microscope with a 100× objective (N.A. 1.3) in DAPI and Texas-red filter spectra.

Fluorescent spots corresponding to individual mRNA molecules in each image stack were evaluated manually because automatic thresholding using previously reported algorithms were found to be unreliable. Difficulty in automating this process was attributed to inconsistent signal to noise using reported protocols and may be related to the thickness of hESC cells (approximately 15 µm). In addition, manual intervention was needed to ascertain the boundaries of adjacent cells. To optimize the signal to noise we systematically varied the probe concentration, incubation time, incubation temperature, as well as the formamide concentration in the hybridization buffer solution.

Cell Culture. K562 cells were cultured in DMEM (Gibco) supplemented with 10% FBS (Gibco). Purified RNA was extracted from K562 cells using RNA MiniPrep (Qiagen).

CA1S hESCs (6, 7) were propagated in mTeSR (8) basal medium (STEMCELL Technologies, Inc.), additionally supplemented with antibiotic-antimycotic mix (100 U/mL penicillin, 100 mg/mL streptomycin, and 0.25 mg/mL amphotericin B) (Invitrogen). Upon passaging, hESCs were washed with PBS prior to incubating with TrypLE Express (Invitrogen) at 37 °C for 10 min to detach single hESCs from 4- to 8-d-old cultures depending on confluency. TrypLE Express was neutralized with mTeSR supplemented with antibiotic-antimycotic mix and suspensions were then transferred into new tissue culture dishes

*<http://www.rtpimerdb.org/>.

containing a precoated layer of 1:30 diluted Matrigel (Becton Dickinson) and mTeSR supplemented with antibiotic-antimycotic mix. For differentiation, mTeSR was replaced with DMEM with 10% FBS 1 d after plating cells.

When harvesting hESCs for qRT-PCR, cells were incubated with TrypLE Express (Invitrogen) at 37°C for 20 min in order to produce a more uniform single-cell suspension from 4- to 8-d-old cultures.

Cryovials of primary cells isolated from a lobular breast cancer metastasis were provided by the BC Cancer Agency in accordance with ethical guidelines of the University of British Columbia. To increase viability, cells were transferred to fresh culture medium and incubated for 2 d before analyzing in the microfluidic device.

Transfer Efficiency Measurements. A solution containing 10 μ M FAM-labeled 40-mer poly-A oligonucleotides (IDT), 0.1% Tween 20, and ROX passive reference dye (from CellsDirect kit, Invitrogen, P/N 54880) diluted 100 \times was loaded into the cell-capture chambers and sequentially pushed into the 10-nL and 50-nL chambers with water containing 0.1% Tween 20, and ROX reference dye diluted 100 \times . Fluorescence images acquired of FAM and ROX were used to measure the transfer of oligonucleotides from one chamber to the next. The transfer efficiency for each chamber was calculated as (Initial Signal – Final Signal)/(Initial Signal), where Signal = (FAM Intensity – FAM Background)/(ROX Intensity – ROX Background). A conservative estimate of the lower bound of transfer efficiency was taken to be one standard deviation from the mean measurement of transfer efficiency.

Cell-Capture Measurements. A custom microfluidic device with a linear array of cell trap geometries was fabricated using protocols described above. The device was mounted on an inverted microscope (Leica DM IRE2) and imaged in bright field using a CCD camera (Hamamatsu ORCA-ER). The device was primed with 0.05% BSA (Gibco) in PBS (Gibco). Prior to loading in the device, cells were washed twice in fresh culture media [DMEM (Gibco) supplemented with 10% FBS (Gibco)]. After the final wash cells were resuspended to be at a concentration of 1 million per mL. Input sample viability was measured with the Cedex Automated Cell Counter (Roche Innovatis AG).

To measure the capture efficiency, cells were pumped through the array using a downstream microfluidic peristaltic pump at a rate of approximately 1 nL/s, and the number of cells that bypassed each trap before a successful trapping event was recorded. These counts were fit using a maximum-likelihood estimator for a geometric distribution with the `fitdistr` function (MASS package version 7.3-6) in R (version 2.11.1). Efficiencies are reported as the probability of a successful capture for each cell.

To measure cell viability after loading, cells were loaded into the array using pressure driven flow as described above until high

trap occupancy was observed. 0.2% trypan blue (Gibco) in PBS was then flowed over the trapped cells. Viability was calculated as the number of unstained cells divided by the total number of cells.

Cell diameter was measured from Cedex images and images of cells trapped in the microfluidic device using ImageJ (version 1.43u). A two-sample *t* test was used to test the hypothesis that the resulting size distributions were significantly different. The assumption of equal variance was tested using an *F* test. For optimized cell trap geometries the cell trapping efficiency was improved to 87% by bringing the cup within one cell diameter of the focuser and by including a small bypass shunt through the cup, similar to the cup geometry presented in Skelley et al. (9).

Mixing by Diffusion. Mixing of solutions by diffusion was characterized in the microfluidic device by loading fluorescently labeled 40 base poly-A oligonucleotides into the 10-nL chambers, and pushing the contents of the chamber into the adjacent 50-nL chambers. Time-lapse imaging was used to measure the evolution of the distribution of fluorescently labeled oligonucleotides in the PCR chambers over time (Fig. S7). The standard deviation of the pixel intensities in each chamber through time was used as a metric of mixing. The resulting curves of all analyzed chambers (*N* = 200) were each fit to a decaying exponential using least squares regression to determine the characteristic mixing time constant. This resulted in a mean mixing time of 15.2 ± 1 min.

Using the Stokes–Einstein relation and assuming a random coil we estimate the diffusion constant of a 40 base oligonucleotide to be

$$D = \frac{K_B T}{6\pi\eta R_h}, \quad [S1]$$

where $K_B T$ is the thermal energy (4.1 pN·nm), η is the fluid viscosity (approximately 0.001 kg/m·s), and R_h is the coil hydrodynamic radius (10). The hydrodynamic radius is proportional to the radius of gyration R_g , and is given by

$$R_h \approx 0.5R_g \approx 0.5(Lp/3)^{1/2}, \quad [S2]$$

where L is the contour length of single stranded DNA (40 bases \times 4.3 Å/base) and p is the persistence length (approximately 40 Å) (10). This yields a diffusion value of approximately 9.0×10^{-11} m² s⁻¹, which is comparable to the diffusion constant of polymerase, the largest molecule in the PCR mix. Since the template solution constitutes only 1/5 of the final PCR reaction it must diffuse the longest distance to equilibrate across the chamber. Therefore, the measured diffusion time of 15.2 min represents an upper bound to the time constant for complete mixing of all components.

1. Unger MA, Chou HP, Thorsen T, Scherer A, Quake SR (2000) Monolithic microfabricated valves and pumps by multilayer soft lithography. *Science* 288:113–116.
2. Thorsen T, Maerkl SJ, Quake SR (2002) Microfluidic large-scale integration. *Science* 298:580–584.
3. Chen C, et al. (2005) Real-time quantification of microRNAs by stem-loop RT-PCR. *Nucleic Acids Res* 33:e179.
4. Warren L, Bryder D, Weissman IL, Quake SR (2006) Transcription factor profiling in individual hematopoietic progenitors by digital RT-PCR. *Proc Natl Acad Sci USA* 103:17807–17812.
5. Shah SP, et al. (2009) Mutational evolution in a lobular breast tumour profiled at single nucleotide resolution. *Nature* 461:809–813.
6. Schulze HG, et al. (2010) Assessing differentiation status of human embryonic stem cells noninvasively using Raman microspectroscopy. *Anal Chem* 82:5020–5027.
7. Adewumi O, et al. (2007) Characterization of human embryonic stem cell lines by the International Stem Cell Initiative. *Nat Biotechnol* 25:803–816.
8. Ludwig TE, et al. (2006) Feeder-independent culture of human embryonic stem cells. *Nat Methods* 3:637–646.
9. Skelley AM, Kirak O, Suh H, Jaenisch R, Voldman J (2009) Microfluidic control of cell pairing and fusion. *Nat Methods* 6:147–152.
10. Tinland B, Pluen A, Sturm J, Weill G (1997) Persistence length of single-stranded DNA. *Macromolecules* 30:5763–5765.




Fig. S1. Histograms showing the size distribution of cells in original sample as measured by Cedex (A) are consistent with the size distribution of cells isolated by microfluidic traps (B). Under the assumption of spherical cell shape the distribution of diameters of trapped cells corresponds to a mean volume of 4.2 pL with a standard deviation of 2.0 pL





Fig. S2. On-chip cell washing. (A) Measurements of GAPDH in cells washed in PBS off-chip prior to injection into microfluidic device, without an on-chip wash contain background signal from template in supernatant. Without on-chip washing, untrapped cells remain in the capture chambers, resulting in fewer single-cell measurements (histogram inlayed). Detection of residual RNA after washing is dramatically reduced by comparison to off-chip results (Fig. S4) due to small volume processing. (B) On-chip washing was found to reduce the background signal from free RNA in the supernatant and dramatically increased the number of single cells analyzed. (C) Comparison of GAPDH measurements from loading purified RNA and washing, or not washing, the cell-capture chambers.

Fig. S3. Comparison of GAPDH measurements from K562 cell lysate with RT performed in the microfluidic device (on-chip) or RT performed in tubes (off-chip) prior to qPCR in the device. Obtained CT values (*Inset*) show no significant difference in efficiency.






Fig. 55. Measurement of miR-16 in hESC cell aggregates demonstrates that the number of cells is reflected in corresponding CT values.




Fig. 56. mRNA-FISH of OCT4 (red) counterstained with DAPI (blue) in CA15 cells. (A) Representative image of mRNA-FISH of OCT4 in a CA15 cell after 7 d of FBS differentiation. Estimate of average copy number of OCT4 mRNA as determined by manual inspection of image stacks is 42 (SD = 41, $N = 6$). (B) Representative image of undifferentiated CA15 cells. Estimate of average copy number as determined by manual inspection of image stacks is 988 (SD = 368, $N = 6$). Scale bar, 10 μ m.




Fig. S7. Mixing by diffusion. Plot shows the standard deviation of pixel intensity values for a chamber as a function of time following the transfer of a solution of fluorescently labeled 40 base poly-A oligonucleotide from the RT chamber (10 nL) to the PCR chamber (50 nL) by flushing with buffer. An exponential fit to the data to each of 200 chambers yields a mean mixing time constant of 15.2 ± 1.0 min. A representative time-lapse series of images from one chamber is shown (*Right*).

Table S1. Multiple OCT4 probe sequences for mRNA FISH

Probe (5' → 3')	Name	Probe #	Position	% GC
tgaatgagggtctgcgaag	OCT4_1	1	2	50
aaatccgaagccagggtgcc	OCT4_2	2	61	55
atcacctccaccactctggag	OCT4_3	3	95	60
agggtccgaggatcaaccag	OCT4_4	4	138	60
aggagggtccttggaaactta	OCT4_5	5	161	55
aatccccacacctcgagcg	OCT4_6	6	215	60
atccccccacagaaactcata	OCT4_7	7	253	50
actagccccactccaactctg	OCT4_8	8	289	60
tcaggctgagaggtctccaa	OCT4_9	9	322	55
agttgtctctccaccgact	OCT4_10	10	354	60
ttctcttctccagcttcac	OCT4_11	11	418	50
ctctctcgggttttgctcca	OCT4_12	12	440	60
ttctgcagagctttgatgtc	OCT4_13	13	466	45
cttgccaaattgctcgagtt	OCT4_14	14	488	45
tgatctcttctgcttcagg	OCT4_15	15	510	50
atcggtctgtgtatatccca	OCT4_16	16	533	50
aaatagaacccccagggtga	OCT4_17	17	560	50
tgtttgggtcgaataccttc	OCT4_18	18	582	45
taagctgcagagcctcaaaag	OCT4_19	19	612	50
gcagcttacacatgttcttg	OCT4_20	20	636	45
tccaccactttctgcagcaa	OCT4_21	21	661	55
gattttcattgttgctagct	OCT4_22	22	684	35
tctgtttgcatactctctg	OCT4_23	23	706	45
actggttcgtttctctttc	OCT4_24	24	743	45
ttgcctctcactcggttctc	OCT4_25	25	766	55
ctgcaggacaacaaattctcca	OCT4_26	26	788	45
actctgcagctgtgggttt	OCT4_27	27	814	50
atccttctcgagcccaagct	OCT4_28	28	851	55
ttacagaaccacactcggac	OCT4_29	29	874	50
ttagctcgttctgatcgctt	OCT4_30	30	910	50
ctcaaaactctctggtgtg	OCT4_31	31	932	45
ctgagaaaggagaccaggca	OCT4_32	32	954	55
agaggaaaaggacactggtcc	OCT4_33	33	976	55
atagcttggggtaccaaaat	OCT4_34	34	1,010	45
agtcacgtgcagtgaaagtga	OCT4_35	35	1,038	45
ttccccctcagggaaggga	OCT4_36	36	1,064	60
tgacggagacagggggaaaag	OCT4_37	37	1,086	60
agtttgaatgcattgggagag	OCT4_38	38	1,116	45
attctagaagggtcaggcac	OCT4_39	39	1,139	55
ttttcttccctagctctc	OCT4_40	40	1,176	45
aaaaacacctgacacaaatc	OCT4_41	41	1,200	45
ccttagtgaatgaagaactt	OCT4_42	42	1,226	35
agctttgtgttcccaattc	OCT4_43	43	1,249	45
aaccagttgccccaaactcc	OCT4_44	44	1,278	55
cattgaactctcaccttccct	OCT4_45	45	1,300	45
gtgggattaaaaatcaagagc	OCT4_46	46	1,322	40
ccaggctctctatttaaga	OCT4_47	47	1,359	35
aagtgtgtctatctactgtg	OCT4_48	48	1,381	40