



## **University of Dundee**

# Holographic Optical Tweezers at the Tip of a Needle

Leite, Ivo T.; Turtaev, Sergey; Jiang, Xin; Siler, Martin; Cuschieri, Alfred; Russell, Philip St J.

Published in:

IEEE Photonics Society Summer Topicals Meeting Series, SUM 2018

DOI:

10.1109/PHOSST.2018.8456739

Publication date: 2018

Document Version Peer reviewed version

Link to publication in Discovery Research Portal

Citation for published version (APA):
Leite, I. T., Turtaev, S., Jiang, X., Siler, M., Cuschieri, A., Russell, P. S. J., & Cizmar, T. (2018). Holographic Optical Tweezers at the Tip of a Needle. In *IEEE Photonics Society Summer Topicals Meeting Series, SUM 2018* (pp. 133-134). [8456739] Institute of Electrical and Electronics Engineers Inc.. https://doi.org/10.1109/PHOSST.2018.8456739

**General rights** 

Copyright and moral rights for the publications made accessible in Discovery Research Portal are retained by the authors and/or other copyright owners and it is a condition of accessing publications that users recognise and abide by the legal requirements associated with

- Users may download and print one copy of any publication from Discovery Research Portal for the purpose of private study or research.
- You may not further distribute the material or use it for any profit-making activity or commercial gain.
  You may freely distribute the URL identifying the publication in the public portal.

If you believe that this document breaches copyright please contact us providing details, and we will remove access to the work immediately and investigate your claim.

Download date: 20. Aug. 2022

Author Accepted Manuscript version of Leite, IT, Turtaev, S, Jiang, X, Siler, M, Cuschieri, A, Russell, PSJ & Cizmar, T 2018, Holographic Optical Tweezers at the Tip of a Needle. in IEEE Photonics Society Summer Topicals Meeting Series, SUM 2018., 8456739, Institute of Electrical and Electronics Engineers Inc., pp. 133-134, 2018 IEEE Photonics Society Summer Topicals Meeting Series, SUM 2018, Waikoloa, United States, 9/07/18. DOI: 10.1109/PHOSST.2018.8456739

# Holographic Optical Tweezers at the Tip of a Needle

Ivo T. Leite Leibniz Institute of Photonic Technology Jena, Germany ivo.leite@leibniz-ipht.de

Alfred Cuschieri
Institute of Medical Science and
Technology, Univ. Dundee
Dundee, United Kingdom
a.cuschieri@dundee.ac.uk

Sergey Turtaev
Leibniz Institute of Photonic
Technology
Jena, Germany
sergey.turtaev@leibniz-ipht.de

Philip St. J. Russell Max Planck Institute for the Science of Light Erlangen, Germany philip.russell@mpl.mpg.de Xin Jiang
Max Planck Institute for the
Science of Light
Erlangen, Germany
xin.jiang@mpl.mpg.de


Tomáš Čižmár Leibniz Institute of Photonic Technology Jena, Germany tomas.cizmar@leibniz-ipht.de Martin Šiler Institute of Scientific Instruments of the Czech Academy of Science Brno, Czech Republic siler@isi.brno.cz

Abstract—Fibre-based optical tweezers typically rely on engineered fibre terminations yielding limited flexibility in number and positioning of trap sites. Here, we demonstrate holographic optical tweezers delivered by a lensless, high-NA multimode fibre with full positioning control of multiple tweezers independently and in all directions.

Keywords—optical tweezers, multimode fibres, digital holography, wavefront shaping

Optical tweezers allow exerting and measuring minuscule forces in a contactless manner and, particularly in the domain of Biophotonics, have become ubiquitous tools for studying the mechanical properties of cells, molecular motors and biomolecules [1]. Conventional systems are usually based on bulk optical components such as high-NA microscope objective lenses to create sharp and aberrations-free foci that produce gradient forces overcoming the radiation pressure, thus achieving stable 3D trapping using a single beam of light. Miniaturised designs using optical fibres have been the object of intense research in the past two decades, aiming at applications spanning from optofluidics chips to minimallyinvasive probes capable of taking the study of the mechanics of biological processes deep within living tissue. Such designs rely mostly on single-mode as well as multicore fibres with elaborate fibre terminations and the optical traps are either static or have limited flexibility in terms of positioning the trap sites [2].

We have developed an all-solid, step-index multimode fibre based on compound 'soft-glasses' yielding an unprecedented NA exceeding 1 at wavelengths smaller than 600 nm, and 0.96 at 1064 nm. By utilising and further extending the methods of holographic wavefront shaping for controlling light propagation in multimode fibres [3,4], we demonstrate three-dimensional optical trapping and manipulation via a bare, single fibre [5]. Being inherently holographic, our approach allows generating an arbitrary number of optical tweezers, as well as precisely repositioning them independently in all directions (Fig. 1). Surprisingly, the high-NA fibre probe shows a remarkable macro-bending resilience even for radii of curvature up to 5 cm (Fig. 2), which should be sufficient for a






Fig. 1. Dynamic manipulation of holographic optical tweezers delivered through a lensless multimode fibre. Multiple holographic tweezers trapping of (a) eight particles comprising a rotating 3D cube and (b) six particles with different size illustrating Bohr's model of the He atom. The particles appear to have different brightness according to their axial position with respect to the focal plane of the imaging microscope objective (particles appearing 'brighter' are at a greater distance from the fibre endface). Incoherent light for illumination is also delivered via the fibre core, which is visible as the blurred disk in the background.




Fig. 2. Resilience of the holographic optical tweezers to bending of the fibre. (a) A static optical trap confining a silica microsphere (1.5  $\mu m$  diameter) is delivered through a multimode fibre with length 65 mm. The unheld portion of the fibre (15 mm long) is initially straight and is then progressively bent due to the force applied by a motorised stage moving upwards. (b) Side-view image of the trapped particle in the initial, straight conformation of the fibre. (c) Side-view image of the trapped particle in the final, bent conformation of the fibre corresponding to a vertical displacement of 2 mm and a radius of curvature of 5 cm.

broad spectrum of real-world applications. Moreover, the possibility of generating aberration-free foci with NA exceeding 0.8 across the fibre core opens new perspectives for high-resolution holographic micro-endoscopy, paving the way for the delivery of advanced microscopy techniques through air-thin fibre-optic probes.

### ACKNOWLEDGMENT

I.T.L. and S.T. acknowledge funding from the People Programme (Marie Curie Actions) of the European Union's Seventh Framework Programme (FP7/2007–2013) under REA grant agreement no. 608144. I.T.L., S.T. and T.Č. also acknowledge financial support from the Thüringer Ministerium für Wirtschaft, Wissenschaft und Digitale Gesellschaft, the Thüringer Aufbaubank and the Federal Ministry of Education and Research, Germany (BMBF). M.Š. and T.Č. acknowledge support from the European Regional Development Fund

through project no. CZ.02.1.01/0.0/0.0/15003/0000476. T.Č. also acknowledges the University of Dundee and SUPA (Scottish Universities Physics Alliance; PaLS initiative) for financial support.

### REFERENCES

- [1] F. M. Fazal and S. M. Block, "Optical tweezers study life under tension," Nat. Photonics, vol. 5, pp. 318–321, 2011.
- [2] R. S. R. Ribeiro, O. Soppera, A. G. Oliva, A. Guerreiro, and P. A. S. Jorge, "New trends on optical fiber tweezers," J. Ligthw. Technol., vol. 33, pp. 3394–3405, 2015.
- [3] T. Čižmár and K. Dholakia, "Exploiting multimode waveguides for pure fibre-based imaging," Nat. Commun., vol. 3, pp. 1027, 2012.
- [4] S. Turtaev, I. T. Leite, and T. Čižmár, "Multimode fibres for microendoscopy," Optofluid. Microfluid. Nanofluid., vol. 2, pp. 31–35, 2015.
- [5] I. T. Leite et al., "Three-dimensional holographic optical manipulation through a high-numerical-aperture soft-glass multimode fibre," Nat. Photonics, vol. 12, pp. 33–39, 2018.