

Inelastic X-ray scattering with 0.75 meV resolution at 25.7 keV using a temperature-gradient analyzer

Daisuke Ishikawa,^{a,b,*} David S. Ellis,^{a,‡} Hiroshi Uchiyama^{a,b} and Alfred Q. R. Baron^{a,b}

^aMaterials Dynamics Laboratory, RIKEN SPring-8 Center, 1-1-1 Kouto, Sayo, Hyogo 679-5148, Japan, and ^bResearch and Utilization Division, Japan Synchrotron Radiation Research Institute (JASRI), 1-1-1 Kouto, Sayo, Hyogo 679-5198, Japan. *E-mail: disikawa@spring8.or.jp

The use of temperature-gradient analyzers for non-resonant high-resolution inelastic X-ray scattering is investigated. The gradient compensates for geometrical broadening of the energy resolution by adjusting the lattice spacing of the analyzer crystal. Applying a ~ 12 mK temperature gradient across a 9.5 cm analyzer, resolutions of 0.75 (2) meV FWHM at 25.7 keV for Si(13 13 13) and 1.25 (2) meV at 21.7 keV for Si(11 11 11) were measured, while retaining large (250 mm) clearance between the sample position and detector, and reasonable (9.3 mrad \times 8.8 mrad) analyzer acceptance. The temperature control and stability are discussed.

Keywords: inelastic X-ray scattering; high-resolution analyzer; X-ray optics.

1. Introduction

Non-resonant high-resolution inelastic X-ray scattering (IXS) with ~ 1 meV resolution is a spectroscopic technique that has been widely used to study atomic dynamics, including phonons in crystals and collective dynamics in disordered materials. With the advent of intense and highly collimated X-ray beams from synchrotron light sources, much effort has been made to develop instrumentation for IXS spectrometers (Dorner & Peisl, 1983; Burkel, 1991; Sette *et al.*, 1998; Baron *et al.*, 2000; Sinn *et al.*, 2001).

A key component in presently operating IXS spectrometers is the spherical analyzer (Fujii *et al.*, 1982; Masciovecchio *et al.*, 1996*a,b*; Verbeni *et al.*, 2005; Said *et al.*, 2011), and it is, in fact, usually the analyzer performance that limits the spectrometer energy resolution. Therefore, methods of improving energy resolution of the analyzers are of interest, especially if this is possible without reducing analyzer efficiency (*e.g.* Huotari *et al.*, 2005, 2006; Ishikawa & Baron, 2010). In this paper we show that the temperature-gradient analyzer proposed by Ishikawa & Baron (2010) can be used to provide 0.75 ± 0.02 meV resolution at the Si(13 13 13) reflection, whereas an analyzer without such a gradient is estimated to provide at best 0.81 meV in theory, and observed to provide not better than 0.9 meV resolution in practice. This validates the temperature-gradient concept for high resolution. In addition, we explore using such a gradient at the Si(11 11 11) reflection and show that it can gain practical improvement in resolution to 1.25 meV, although, for the (11 11 11) reflection, this level of

resolution is also theoretically possible with a uniform-temperature crystal.

2. Basic concepts for temperature-gradient analyzers

The essential idea for a temperature-gradient analyzer is to use a thermal gradient to tailor the d -spacing of a crystal to compensate for the ‘demagnification’ aberration. This can be useful when the experimental geometry forces the spectrometer to operate without the detector in a conventional Roland geometry, something that becomes increasingly necessary as resolution is improved and spherical analyzers must operate close to backscattering. Ishikawa & Baron (2010) provide an in-depth discussion of how temperature-gradient analyzers operate. Here we briefly discuss the concept again, but focus on the high (\sim meV) resolution case.

2.1. Spectrometer geometry

The high-resolution spectrometer at BL43LXU is designed for \sim meV energy resolution using photon energies $E = 21.7, 23.7, 25.7$ keV [for Si(nmn) reflections, $n = 11, 12, 13$]. The monochromator uses a single backscattering reflection with $\theta_B = 89.98^\circ$, while the analyzers are ‘diced’ spheres with a radius of curvature $R = 9.8$ m. Of the two geometries discussed by Ishikawa & Baron (2010), here we use the ‘off-Rowland’ geometry because it minimizes the required size of the detector, reducing noise, and allowing operation closer to backscattering which gives better resolution.

For a spherical analyzer, focusing in the detector requires $2/R = 1/L_1 + 1/L_2$ where L_1 is the sample–analyzer distance, L_2

[‡] Present address: Technion, Israel Institute of Technology, Technion City, Haifa 3200003, Israel.




Figure 1 Schematic of the IXS sample-analyzer-detector geometry (side view). Here, $R, L_1, L_2 \gg l \gg d$. The analyzer crystals are positioned to focus in the detector. The energy-position (detector vertical position) correlation ($E - E_0$ versus d_y) of crystallites (1)–(3) and all crystallites effects (total) are also illustrated.

the analyzer-detector distance, $l = L_1 - L_2$ is the sample-detector horizontal offset (here, $R \simeq L_1 \simeq L_2$, $\delta_0 \ll 1$, $\delta_0 \equiv \pi/2 - \theta_B$) (see Fig. 1). Taking the analyzer radius of curvature $R = 9800$ mm and clearance at sample $l = 250$ mm, then $L_1 = 9927$ and $L_2 = 9679$ mm.¹ Taking the analyzer (vertical) angular acceptance to be $\Omega = D/L_1 = 8.8$ mrad, the minimum detector offset in this geometry is given as $d_{\min} = \Omega l/2 + c' = 3$ mm. Here, $c' = c(1 + M)/2 \simeq 0.86$, where the analyzer pixel size $c \simeq 0.87$ mm and magnification $M = L_2/L_1 = 0.975$. In order to accommodate four rows of analyzers [as is useful for measuring transverse dispersion; see Baron *et al.* (2008)], two rows of detectors are used above and below the scattered beam, as shown in Fig. 2(a). This means that the spectrometer has two different offsets, d , for the detectors, with $d = 4$ mm for the first and fourth rows of analyzers and $d = 10$ mm for the second and third rows of analyzers. Detector elements have a size of $p = 2$ mm that is sufficient to accept the focused beam size of $2c' \simeq 1.7$ mm, and are arranged on a 3.03 mm pitch. Note that this off-Rowland geometry, as discussed by Ishikawa & Baron (2010), generally requires a non-linear (quadratic) temperature-gradient correction; however, a linear gradient is sufficient in this case due to large, 10 m, arm length and small solid angle $\Omega \simeq 10$ mrad at near-backscattering $\delta_0 < 0.5$ mrad.

2.2. Analyzer crystals

Analyzers were fabricated on rectangular substrates [100 (H) mm \times 95 (V) mm \times 30 (t) mm] to facilitate creating a one-directional temperature gradient, as discussed in detail

¹ The present spectrometer has flexibility to operate between on-Rowland and off-Rowland geometry. The sample-analyzer distance L_1 is adjustable from 9800 to 10100 mm.




Figure 2 Schematic view of the multi-element analyzer/detector focusing geometry at BL43LXU (not to scale). (a) Side view, (b) view from above. The signals from each analyzer are collected by individual detectors.

below. Single-crystal silicon was used as a substrate material due to its high thermal conductivity, which both promotes uniform temperature (Masciovecchio *et al.*, 1996b) and makes it relatively easy to control the small gradients (~ 0.01 K/10 cm) needed for operation. Wafers were cut from a high-purity single-crystalline silicon ingot (resistivity 2–4 k Ω cm) and then diced leaving a 0.2 mm backwall. The wafers were etched in KOH solution, and then the diced side was bonded onto the spherical substrate ($R = 9800 \pm 15$ mm) by gold diffusion, before removing the backwall by polishing and etching in acid. The accuracy of the bonded crystallite alignment is typically < 17 μ rad (r.m.s.) error from ideal curvature on the spherical substrate (Miwa, 2002). The final crystallites have typical dimensions of 0.87 mm \times 0.87 mm \times 4.9 mm on a 1.0 mm pitch on the 92 (H) mm \times 87 (V) mm active analyzer surface. 5 mm-thick crystallites were used to improve the tails of the resolution function (Said *et al.*, 2011). To reduce multi-beam effects at the Si(11 11 11) reflection, the (111)-normal wafers were cut with (1 $\bar{1}$ 0) 45° inclined from the (111) scattering plane.

2.3. Creating the temperature gradient

The gradient is created by using the analyzer as one component in a thermal circuit. The required heat flow through the circuit, for a desired temperature gradient $\Delta T/h$, may be estimated as $Q = \lambda S \Delta T/h$, where λ is the thermal conductivity, ΔT the temperature difference of the two surfaces, S the heat transfer area and h the heat transfer distance. Taking $\lambda = 156$ W m⁻¹ K⁻¹ [single-crystalline silicon at $T = 300$ K (Glassbrenner & Slack, 1964)], $S = 30$ mm \times 100 mm, $h = 95$ mm and $\Delta T/h = 11.9$ mK/95 mm (= 10 mK/80 mm), the required power is $Q = 58.5$ mW. To accomplish




Figure 3 Rectangular analyzer and its holder for a one-directional temperature gradient. T_1 – T_4 indicate positions of thermistors used for feedback. The arrows indicate thermal flow under steady-state conditions.

this, the analyzer is mounted between upper and lower L-shaped angle brackets made of Ni-coated oxygen-free high-conductivity (OFHC) copper. One of the brackets is heated and insulated from the other, as seen in Fig. 3. A 1 inch \times 1 inch film heater (78.4 Ω , Minco HK5163) is used as a main heater for base temperature control, and two power resistors (1.5 W maximum, each 100 Ω , in parallel, Alpha Electronics PDY100R00A) are used to generate the gradient. The assembly is mounted on a water-cooled plate. Heating points are located some distance away from the silicon to improve thermal homogeneity perpendicular to the gradient direction. The assembly is mounted in-vacuum (pressure < 5 Pa) to thermally isolate the system.

The contact between the crystal and the two L-brackets requires care. To improve thermal contact, the surfaces were mirror finished and parallel to <0.1 mrad. To promote temperature uniformity and reduce thermal contact resistance, several materials were tried, including indium foil, silver paste, thermal grease and InGa eutectic. The best results were obtained with a small amount of InGa eutectic.

Temperatures were measured using glass-encapsulated thermistors with a \sim 2 mm-diameter bead, 10 k Ω resistance (at 298 K) and four-wire readout,

for the thermistors. The voltage drop across each thermistor is then compared with that across a precision resistor using a switching digital multimeter (Keithley 3706A) with plug-in multiplexers cards (model 3720, 7.5 digit). Typically 24 thermistors are used in parallel with a given reference resistor. The

with four temperature sensors per gradient analyzer (Fig. 3). The sensors T_1 and T_2 were mounted near the center of the substrate at each side, and sensors T_3 and T_4 were attached 80 mm apart each from other. Sensors T_1 and T_2 were for monitoring the temperature of the center of the analyzer, $T_0 [\equiv (T_1 + T_2)/2]$, while the sensors T_3 and T_4 were for monitoring the temperature gradient in the vertical direction, $\Delta T_g [\equiv T_3 - T_4]$. T_0 and ΔT_g were used as the feedback parameters for control of the main heater and the offset heater, respectively.

2.4. Temperature control and stability

The temperature control system was built in-house in order to keep it both compact (see Fig. 4) and precise over many channels operated in parallel. Standard 1.5 V batteries (AA-cell) are used as ultralow-noise power supplies




Figure 4 Schematic of the compact multi-channel temperature control system.

Table 1

Resolution for the BL43LXU spectrometer as calculated *via* ray-tracing.

Resolutions given are FWHM. E : photon energy; d : sample–detector vertical offset; δ_0 : deviation angle from exact backscattering at the center of the analyzer; $\Delta\delta$ ($= \delta_{\max} - \delta_{\min}$): Bragg angle distribution over the analyzer; $(\Delta E/E)_{\text{geom}2} \equiv \Delta\delta \tan \delta_0$: fractional energy resolution by demagnification contribution; $(\Delta E/E)_{\text{geom}1} \equiv (c/L_1) \tan \delta_0$: fractional energy resolution by crystallite size contribution; ΔE_{geom} : geometrical broadening; ΔE_{tot} : theoretical total energy resolution; $\Delta E'_{\text{geom}}$, $\Delta E'_{\text{tot}}$, ΔE_{geom} and ΔE_{tot} with temperature gradient; ΔT_g : temperature gradient to reduce demagnification contribution. Here, radius of curvature: $R = 9800$ mm; sample–analyzer distance: $L_1 = 9926.6$ mm; analyzer–detector distance: $L_2 = 9676.6$ mm; sample–detector horizontal offset: $l = 250$ mm.

Case	E (keV)	d (mm)	δ_0 (mrad)	$\Delta\delta$ (mrad)	Uniform temperature			Temperature gradient			
					$(\Delta E/E)_{\text{geom}2}$ ($\times 10^{-8}$)	ΔE_{geom} (meV)	ΔE_{tot} (meV)	$(\Delta E/E)_{\text{geom}1}$ ($\times 10^{-8}$)	$\Delta E'_{\text{geom}}$ (meV)	$\Delta E'_{\text{tot}}$ (meV)	ΔT_g (mK/80 mm)
1	25.702	4	0.204	0.112	2.29	0.57	0.81	1.79	0.43	0.65	8.08
2	25.702	10	0.510	0.112	5.71	1.46	1.64	4.47	1.06	1.21	20.2
3	21.747	4	0.204	0.112	2.29	0.48	1.27	1.79	0.37	1.21	8.08
4	21.747	10	0.510	0.112	5.71	1.22	1.74	4.47	0.96	1.42	20.2

heater power was controlled using a multi-channel low-voltage power supply (W-IE-NE-R MPOD crate) with low-voltage modules type MPV-8060 (16-bit resolution, maximum 50 V). The temperature parameters T_0 and ΔT_g were controlled by a PID feedback program based on Bechhoefer (2005), which has online monitoring and logging. The feedback loop takes about 9 s per iteration to scan 72 temperature channels.

The temperatures must be stable to better than ± 0.3 mK² for this system to work properly over the \sim day, or longer, time scale of IXS measurements. In order to determine the PID feedback parameters, the open-loop thermal frequency response of the analyzer over long time scales was measured. Bode plots were made from 0.2 to 6 mHz, and the amplitude and phase open-loop response of the analyzer were fit to polynomials, which were used to construct a model of the analyzer’s thermal system in frequency space. Fig. 5 shows the temperature stability over a period of seven days, with temperature gradient $\Delta T_g = 10$ mK/80 mm. The stability was within ± 0.3 mK (≤ 0.08 mK r.m.s.) for all temperature sensors. During the initial adjustment phase, it took a few hours for the system to reach steady-state condition (*e.g.* from $\Delta T_g = 0$ to 10 mK; see inset in Fig. 5) after a change. Measurement of the energy resolution multiple times over a 30 day period confirmed the stability of the setup. The analyzer array is now operating with 72 channels of readout and 40 channels for control.

3. Results and discussion

Tests were performed at BL43LXU at SPring-8 in Japan (Baron, 2010). Two different geometries, $d = 4$ and 10 mm (Fig. 2a), and two different photon energies, 21.7 and 25.7 keV, were tested. The bandwidth of the X-rays from a segmented undulator source was reduced by a liquid-nitrogen-cooled high-heat-load double-crystal Si(111) monochromator followed by a pair of Si(400) offset crystals. A high-resolution 1.25° grazing-incidence backscattering monochromator using the Si(13 13 13) or Si(11 11 11) reflection was then used to

² ± 0.3 mK corresponds to ± 0.017 meV for 21.7 keV and ± 0.02 meV for 25.7 keV.




Figure 5

Temperature stability of the analyzer over one week. The temperature gradient is set to $\Delta T_g = 10$ mK/80 mm. T_1 – T_4 are sensors positions as indicated in Fig. 3. The inset shows the transient when establishing the temperature gradient ($\Delta T_g = 0$ to 10 mK/80 mm).

reduce the bandwidth to below 1 meV. The focal spot size at the sample was $40 \mu\text{m}$ (H) \times $45 \mu\text{m}$ (V) after focusing by an elliptically bent cylindrical mirror. The resolution was measured using a 2 mm-thick polymethyl-methacrylate (PMMA) sample with the analyzer placed at the structure-factor maximum. The full analyzer surface was illuminated, corresponding to momentum resolutions of $\Delta Q \approx 1.25 \text{ nm}^{-1}$ (full-width) at 25.7 keV and $\Delta Q \approx 1.05 \text{ nm}^{-1}$ (full-width) at 21.7 keV X-rays. The incident energy was scanned by changing the temperature of the backscattering monochromator with slope of 5 mK min^{-1} for 25.7 keV and 10 mK min^{-1} for 21.7 keV X-rays, while the analyzer temperature was kept constant. The energy scale was calibrated as described by Fukui *et al.* (2008).

3.1. Si(13 13 13), $d = 4$ and 10 mm

The best resolution was obtained using a temperature gradient at the Si(13 13 13) reflection, $E = 25.7$ keV and $d = 4$ mm. Fig. 6 shows the calculated energy–position correlation in the detector for a uniform temperature (top) and temperature gradient (bottom). The geometric contribution (Table 1, case 1) is calculated to be $\Delta E_{\text{geom}} = 0.57$ meV (Fig. 6c) for uniform temperature and 0.43 meV (Fig. 6d) for




Figure 6
 Simulated results for case 1 ($E = 25.7$ keV, $d = 4$ mm) with and without the temperature gradient. (a, b) Energy transfer versus detector vertical position. (c, d) Geometric resolution function. (e, f) Total resolution function convolved with the intrinsic response of the of Si(13 13 13) reflections.




Figure 7
 Resolution with the temperature gradient for $E = 25.7$ keV, $d = 4$ mm (case 1). (a) Energy resolution as a function of temperature gradient. The curve is to guide the eye. (b) Experimental resolution with best-case gradient ($\Delta T_g = 10$ mK/80 mm) and uniform temperature. The solid lines are fitted curves. The solid angle is 9.3 (H) mrad \times 8.8 (V) mrad.




Figure 8
 Resolution with the temperature gradient for $E = 25.7$ keV and $d = 10$ mm (case 2). See caption for Fig. 7. The best-case gradient is $\Delta T_g = 25$ mK/80 mm.

linear temperature gradient. The measured resolution as a function of the applied gradient is shown in Fig. 7, with the resolution improving from 0.96 meV without a gradient to 0.75 meV at the optimum gradient $\Delta T_g = 10$ mK/80 mm (58.7 mW) and then becoming worse as the optimum gradient is exceeded. The values of 0.75 and 0.96 meV are both slightly larger than our ray-tracing results of 0.65 and 0.81 meV, respectively, possibly resulting from residual non-uniformity in the lattice spacing over the analyzer surface. However, the value of 0.75 meV is, to the best of our knowledge, the narrowest resolution achieved with a spherical analyzer with ~ 10 mrad acceptance. The gradient also allows us to improve the resolution beyond the best value calculated without the gradient. It is also notable that the width of the tails on the resolution function are reduced (Table 3) by the gradient and that there are no changes in the integrated intensities to a level of 2%.

For the $d = 10$ mm case the geometrical contribution to the resolution is about 2.5 times larger than for $d = 4$ mm due to the larger deviation from backscattering, $(\Delta E/E)_{\text{geom}2} \equiv \tan \delta_0 \Delta \delta$ (Table 1, case 2). A larger gradient is then expected. Figs. 8(a) and 8(b) show the temperature gradient and resolution as a function of temperature gradient in case 2 ($d = 10$ mm). The energy resolution has a minimum at around $\Delta T_g = 25$ mK/80 mm (155.9 mW), which is roughly consistent with estimated value of $\Delta T_g = 20$ mK/80 mm (118 mW) as noted in Table 2. The experimental energy resolution was improved 32% from $\Delta E'_{\text{tot}} = 1.92$ to $\Delta E'_{\text{tot}} = 1.31$ meV (FWHM). The observed minimum resolution $\Delta E'_{\text{tot}} = 1.31$ meV is in reasonable agreement with our calculated limit $\Delta E'_{\text{tot}} = 1.21$ meV (FWHM). Note that, again, the tails on the resolution function are reduced compared with operation without the gradient (Table 3, cases 1 and 2) and no changes are observed in integrated intensities.

3.2. Si(11 11 11), $d = 4$ and 10 mm

Operation at the Si(11 11 11) reflection is interesting because the relaxed resolution, compared with the Si(13 13 13) reflection, allows an increase in count rates, although it is still desirable to obtain the best possible resolution. For the $d = 4$ mm geometry, the calculated contribution of the demagnification aberration is small enough that the gradient does not make that big a difference: 1.27 meV without the gradient becomes 1.21 meV at the

Table 2

Offset heater output for two focusing geometries at BL43LXU.

Parentheses indicate calculated minimum values. P : applied power for offset heaters; V_f : measured applied voltage for offset heaters.

Case	Measurement				Calculation	
	d (mm)	ΔT_g (mK/80 mm)	P (mW)	V_f (V)	ΔT_g (mK/80 mm)	P (mW)
1	4	10	58.7	1.6–1.7	8.1	47
2	10	25	155.9	2.6–2.8	20.2	118

optimal gradient. However, we find that the observed resolution improved from 1.48 meV to 1.25 meV by adding the gradient (Fig. 9*b*). While this suggests the presence of some residual gradient even when the gradient heater output was zero, it also suggests that practically some improvements can be made. For the case of $d = 10$ mm, when the analyzers are further from backscattering, the calculated improvement (1.72 to 1.45 meV) is more substantial, and we find that the best-case measured resolution is 1.55 meV (Fig. 10).




Figure 9

Resolution with the temperature gradient for $E = 21.7$ keV and $d = 4$ mm (case 3). See caption for Fig. 7. The best-case gradient is $\Delta T_g = 8.8$ mK/80 mm.




Figure 10

Resolution with the temperature gradient for $E = 21.7$ keV and $d = 10$ mm (case 4). See caption for Fig. 7. The best-case gradient is $\Delta T_g = 25$ mK/80 mm.

Table 3

Measured total energy resolution width with temperature gradient at BL43LXU.

Parentheses indicate uniform temperature values. The theoretical limits with the temperature gradient are also listed. FWHM/10: full-width at tenth of maximum; FWHM/100: full width at one-hundredth of maximum.

Case	E (keV)	d (mm)	Calculation		Measurement	
			FWHM (meV)	FWHM/10 (meV)	FWHM/100 (meV)	FWHM/100 (meV)
1	25.702	4	0.65	0.75 (0.96)	1.9 (2.2)	7.4 (8.0)
2	25.702	10	1.21	1.31 (1.92)	2.8 (3.7)	9.9 (11.9)
3	21.747	4	1.21	1.25 (1.50)	3.1 (3.6)	10.9 (12.2)
4	21.747	10	1.42	1.55 (1.97)	4.0 (4.5)	12.6 (14.2)

4. Conclusion

We have shown that careful application of a temperature gradient allows us to improve the resolution of \sim meV-scale-resolution spherical analyzers. We show a significant, \sim 22%, improvement in resolution, with a best case of 0.75 meV (FWHM) using 25.7 keV X-rays, without loss of intensity. The gradient also improves the tail of the response function. This shows that the gradient design, which was originally discussed primarily in the context of 10 meV-resolution analyzers with large solid angle (\sim 50 mrad \times 50 mrad), may also be used for high-resolution analyzers. Work is continuing on the 10 meV setup.

We acknowledge JASRI proposals Nos. 2010B2056 and 2010A1981 for optimization of analyzer fabrications. The authors thank S. Shibutani for technical assistance.

References

Baron, A. Q. R. (2010). *SPring-8 Inf. Newsl.* **15**, 14.
 Baron, A. Q. R., Sutter, J. P., Tsutsui, S., Uchiyama, H., Masui, T., Tajima, S., Heid, R. & Bohnen, K.-P. (2008). *J. Phys. Chem. Solids*, **69**, 3100–3102.
 Baron, A. Q. R., Tanaka, Y., Goto, S., Takeshita, K., Matsushita, T. & Ishikawa, T. (2000). *J. Phys. Chem. Solids*, **61**, 461–465.
 Bechhoefer, J. (2005). *Rev. Mod. Phys.* **77**, 783–836.
 Burkel, E. (1991). *Inelastic Scattering of X-rays with High Energy Resolution, Springer Tracts in Modern Physics*, Vol. 125. Berlin: Springer.
 Dorner, D. & Peisl, J. (1983). *Nucl. Instrum. Methods Phys. Res.*, **208**, 587–592.
 Fujii, Y., Hastings, J., Ulc, S. & Moncton, D. (1982). Technical Report VIII-95. SSRL, Stanford, CA, USA.

- Fukui, H., Katsura, T., Kuribayashi, T., Matsuzaki, T., Yoneda, A., Ito, E., Kudoh, Y., Tsutsui, S. & Baron, A. Q. R. (2008). *J. Synchrotron Rad.* **15**, 618–623.
- Glassbrenner, C. & Slack, G. (1964). *Phys. Rev.* **134**, A1058.
- Huotari, S., Albergamo, F., Vankó, G., Verbeni, R. & Monaco, G. J. (2006). *Rev. Sci. Instrum.* **77**, 053102.
- Huotari, S., Vankó, Gy., Albergamo, F., Ponchut, C., Graafsma, H., Henriquet, C., Verbeni, R. & Monaco, G. (2005). *J. Synchrotron Rad.* **12**, 467–472.
- Ishikawa, D. & Baron, A. Q. R. (2010). *J. Synchrotron Rad.* **17**, 12–24.
- Masciovecchio, C., Bergmann, U., Krisch, M., Ruocco, G., Sette, F. & Verbeni, R. (1996a). *Nucl. Instrum. Methods Phys. Res. B*, **111**, 181–186.
- Masciovecchio, C., Bergmann, U., Krisch, M., Ruocco, G., Sette, F. & Verbeni, R. (1996b). *Nucl. Instrum. Methods Phys. Res. B*, **111**, 339–340.
- Miwa, D. (2002). Master's thesis, Himeji Institute of Technology, Japan.
- Said, A. H., Sinn, H. & Divan, R. (2011). *J. Synchrotron Rad.* **18**, 492–496.
- Sette, F., Krisch, M. H., Masciovecchio, C., Ruocco, G. & Monaco, G. (1998). *Science*, **280**, 1550–1555.
- Sinn, H., Alp, E. E., Alatas, A., Barraza, J., Bortel, G., Burkel, E., Shu, D., Sturhahn, W., Sutter, J. P., Toellner, T. S. & Zhao, J. (2001). *Nucl. Instrum. Methods Phys. Res. A*, **467–468**, 1545–1548.
- Verbeni, R., Kocsis, M., Huotari, S., Krisch, M., Monaco, G., Sette, F. & Vanko, G. (2005). *J. Phys. Chem. Solids*, **66**, 2299–2305.