

**Intercomparison of
air ion spectrometers**

S. Gagné et al.

This discussion paper is/has been under review for the journal Atmospheric Measurement Techniques (AMT). Please refer to the corresponding final paper in AMT if available.

Intercomparison of air ion spectrometers: a basis for data interpretation

S. Gagné¹, K. Lehtipalo¹, H. E. Manninen¹, T. Nieminen¹, S. Schobesberger¹,
A. Franchin¹, T. Yli-Juuti¹, J. Boulon², A. Sonntag³, S. Mirme⁴, A. Mirme⁴,
U. Hörrak⁴, T. Petäjä¹, E. Asmi⁵, and M. Kulmala¹

¹Department of Physics, University of Helsinki, P.O. Box 64, 00014 Helsinki, Finland

²Laboratoire de Météorologie Physique, Blaise Pascal Univ., 63000 Clermont-Ferrand, France

³Leibniz Institute for Tropospheric Research, Permoserstrasse 15, 04303 Leipzig, Germany

⁴Institute of Physics, University of Tartu, Ulilikooli 18, 50090 Tartu, Estonia

⁵Finnish Meteorological Institute, P.O. Box 503, 00101 Helsinki, Finland

Received: 16 November 2010 – Accepted: 17 February 2011 – Published: 22 February 2011

Correspondence to: S. Gagné (stephanie.gagne@helsinki.fi)

Published by Copernicus Publications on behalf of the European Geosciences Union.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

◀

▶

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



Abstract

We evaluated 11 air ion spectrometers from Airel Ltd. after they had spent one year in field measurements as a part of the EUCAARI project: 5 Air Ion Spectrometers (AIS), 5 Neutral cluster and Air Ion Spectrometers (NAIS) and one Airborne NAIS (ANAIS). This is the first time that an ANAIS is evaluated and compared so extensively. The ion spectrometers' mobility and concentration accuracy was evaluated. Their measurements of ambient air were compared between themselves and to reference instruments: a DMPS, a BSMA, and an Ion-DMPS. We report on the simultaneous measurement of a new particle formation (NPF) event by all 11 instruments and the 3 reference instruments. To our knowledge, it is the first time that the size distribution of ions and particles is measured by so many ion spectrometers during a NPF event. The new particle formation rates ($\sim 0.2 \text{ cm}^{-3} \text{ s}^{-1}$ for ions and $\sim 2 \text{ cm}^{-3} \text{ s}^{-1}$ for particles) and growth rates ($\sim 25 \text{ nm h}^{-1}$ in the 3–7 nm size range) were calculated for all the instruments. The NAISs and the ANAIS gave higher concentrations and formation rates than the AISs. The latter agreed better with reference instruments. Finally, based on the results presented here, we give guidelines for data interpretation, when data from different ion spectrometers are compared.

1 Introduction

Ions in the atmosphere influence aerosol particles through their formation and growth mechanisms (Laakso et al., 2002; Lovejoy et al., 2004; Kulmala et al., 2004), through cloud processes (Harrison and Carslaw, 2003), and scavenging of particles (Andronache et al., 2006). In turn, aerosols affect the Earth's climate and the health of its inhabitants. Aerosol particles affect the climate through direct and indirect effects (e.g. Myhre et al., 2009). The direct effect consists of scattering and absorption of solar radiation by the aerosol particles present in the atmosphere. The indirect effects consist of the change of cloud albedo due to a change in aerosol particle concentrations,

Intercomparison of air ion spectrometers

S. Gagné et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

◀

▶

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



Intercomparison of air ion spectrometers

S. Gagné et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

◀

▶

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



as well as changes in the cloud lifetime and precipitation patterns (e.g., Twomey, 1991; Lohmann and Feichter, 2005; Stevens and Feingold, 2009). Aerosol particles larger than ~ 100 nm can act as cloud condensation nuclei (CCN), the size depending on particle and cloud properties.

5 Measurements of charged atmospheric particles (air ions) are the basis of several aerosol measurement techniques (e.g. Gerdien counters, ion spectrometers, air conductivity measurement techniques; see Hirsikko et al. (2011)). The charge of the particles can be used for their detection, and their electrical mobility for size classification. There are techniques that measure only the charged particles (e.g. ion spectrometers).
 10 Other techniques bring the particles to charge equilibrium, followed by mobility classification, allowing for the retrieval of the total particle concentration by data inversion. An example is the well-known and extensively used Differential/Scanning Mobility Particle Sizer (DMPS/SMPS, e.g. Wang and Flagan, 1990). These instruments have allowed the observation of formation and growth nanometer-sized particles above their detection limit, at 3 nm. However, the first steps of new particle formation happen below this limit.

New particle formation (NPF) and growth take place frequently in different environments around the world, including urban, rural and marine environments (see e.g. Kulmala et al., 2004; Laakso et al., 2008; Suni et al., 2008; Venzac et al., 2007, 2008; Vana et al, 2008; Lehtipalo et al., 2010; Crumeyrolles et al., 2010). The nucleation mechanisms are, as yet, not completely understood. Because traditional particle number concentration measurement instrumentation only covers a size range above ~ 3 nm,
 20 new instruments such as the Air Ion Spectrometers (AIS) were developed to allow for the detection of naturally charged particles below this limit. This gives the possibility to observe the behaviour of charged particles at the very first steps of nucleation (Kulmala et al., 2007).
 25

Intercomparison of air ion spectrometers

S. Gagné et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

◀

▶

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



The ion spectrometers evaluated in this work have been developed by Airel Ltd. (Tartu, Estonia) based on long-term measurements and instrumentation development made at the University of Tartu (Matisen et al., 1992). The Air Ion Spectrometer was first released in 2003 and measures the mobility distributions of small atmospheric ions and charged particles (0.8–40 nm in mobility diameter at NTP, Mirme et al., 2007). Later, a second generation instrument was released by the same company: the Neutral cluster and Air Ion Spectrometer (NAIS, Kulmala et al., 2007 and Manninen et al., 2009a). The NAIS can, in addition to the AIS operation modes (negative and positive ions), measure the total (neutral + charged) particle size distribution by using corona chargers for particle charging. An airtight and improved version of the NAIS, the Airborne NAIS (ANAIS, 2nd generation of NAIS, Mirme et al., 2010), was developed in 2007/2008 to allow measurements at different altitudes, including measurements on board an airplane. In this paper, results from all three different instrument types are presented.

The above mentioned ion spectrometers have been used in both field and laboratory measurements. They have been measuring, for example, at forest sites (in Finland: Kulmala et al., 2007; in Australia: Suni et al., 2008), at high altitudes (Venzac et al., 2007), in the South-African savannah (Laakso et al., 2008), in a marine environment in Ireland (Vana et al., 2008; Lehtipalo et al., 2010), on a trans-Siberian train (Vartiainen et al., 2007), indoors (Hirsikko et al., 2007), at a chamber experiment at CERN (Duplissy et al., 2010), in a hot-air balloon (Laakso et al., 2007a), on board an airplane (Mirme et al., 2010), and in many other places and environments (see Hirsikko et al., 2011 and references therein). Between spring 2008 and spring 2009, the instruments were measuring at different EUCAARI stations (European Integrated Project on Aerosol Cloud Climate Air Quality Interactions; Kulmala et al., 2009; Manninen et al., 2010, Kerminen et al., 2010).

The first air ion spectrometer calibration workshop (Asmi et al., 2009) took place during January and February 2008 in Helsinki, Finland, where 10 instruments (5 AISs and 5 NAISs) were calibrated. The purpose of the first calibration workshop was to verify

Intercomparison of air ion spectrometers

S. Gagné et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

◀

▶

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



the accuracy of the instruments so that they can be compared after being deployed in different sites around the world. Another purpose was to help in the development of such instruments by thoroughly characterizing them. Asmi et al. (2009) performed mobility and concentration comparisons to reference instruments and compared the ion spectrometers to each other. They concluded that the 10 instruments were comparing well with each other and were accurate. They found that the theoretical transfer function used for data inversion was comparable to the one they measured for AIS 5, and NAIS 3, 4, 5. They also found that the ion spectrometers were slightly overestimating the mobility. Although they found that the NAISs showed larger concentrations than the AISs in ambient measurements, they did not find such a difference during calibrations.

One of the two main aims of the second air ion spectrometer calibration workshop (this work), was to investigate the repeatability of the measurements, after the instruments had spent about one year in varying weather conditions in different environments. The second aim was to compare the instruments to each other and to reference instruments more extensively using ion and particle concentrations that reflect the ones observed in the field, also during new particle formation events. The latter aims at producing guidelines for data analysis, especially when comparing ion spectrometers with each other.

In this paper, we present the calibration results for mobility and concentration, and discuss the differences and similarities to the first calibration workshop. We also discuss the results of the intercomparison as well as the comparison with other instruments: a Balanced Scanning Mobility Analyzer (BSMA, Tammet, 2006), a Differential Mobility Particle Sizer (DMPS, Aalto et al., 2001) and an Ion-DMPS (Laakso et al., 2007b). We present the values obtained from NPF events and discuss the charged fraction measurements made with the different (A)NAISs. Thus, we evaluate the performance of different instruments and provide guidelines for data analysis and interpretation of field measurements.

2 Ion spectrometers

All the ion spectrometers (AIS, NAIS, ANAIS) are based on the same principle and share the same mobility analyzer structure. However, the models vary in their inlet part (including the chargers) and the air flow system (sampling and sheath air). Each ion spectrometer is identified with an individual name which is built from the instrument type and its serial number (e.g. NAIS3, see Table 1 for a list). These names were used during the EUCAARI campaign as well as in this paper.

2.1 Air Ion Spectrometer (AIS)

The Air Ion Spectrometer (AIS, Mirme et al., 2007) measures the size distribution of charged particles with a high time resolution. The measured mobilities range between 3.2 and $0.0013 \text{ cm}^2 \text{ V}^{-1} \text{ s}^{-1}$, corresponding to Stokes-Millikan mobility diameters (Mäkelä et al., 1996) between 0.8 and 42 nm in NTP conditions. The AIS has two identical differential mobility analyzers (DMA) functioning in parallel: one for negatively charged particles and the other for positively charged particles. Each analyzer has a flow rate of 90 lpm : 30 lpm of sample flow, and 60 lpm of closed loop sheath air flow. The sheath air is filtered for re-use by using a corona charger and an electrical filter. The sample air comes through a single inlet with a 60 lpm flow which is then divided into two (30 lpm for each polarity). The high flow rate allows for smaller diffusional losses, so that low concentrations of smaller ions can be detected with a reasonable signal to noise ratio. The high time resolution of the instrument is due to the design of its analyzers, which detect all mobility classes simultaneously, rather than by scanning the mobility distribution like in the SMPS or DMPS systems. The analyzers consist of an inner cylinder with 4 isolated sections to which different voltages are applied. The outer cylinder has 21 isolated cylindrical electrometers piled up vertically. The charged particles pass between the cylinders, perpendicular to an electric field, and are thus directed to an electrometer according to their electrical mobility. Each electrometer corresponds to a mobility channel and measures the current transmitted by

Intercomparison of air ion spectrometers

S. Gagné et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

◀

▶

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



these charged particles, from which their concentration is derived. The time resolution is user-adjustable but a minimum of one minute has been found to work with good accuracy, depending on noise levels (Asmi et al., 2009).

An offset measurement is made between each ion concentration measurement. A unipolar corona charger charges the particles with ions of the opposite polarity to that measured in the analyzer, and the charged particles are removed with an electric filter. This procedure allows for the measurement of particle-free air and thus the zero drift of the electrometers can be assessed as well as the noise due to parasitic currents. The offset is subtracted from the signal in the data inversion process. The data is inverted using the instrument equation of the ion spectrometer and the transfer function of the channels (based on geometry, flows, voltages and losses), and converted to a mobility distribution over 28 mobility bins. The inverting software is provided by Airel Ltd. It is also possible to record the raw electrometer electrical signal.

2.2 Neutral cluster and Air Ion Spectrometer (NAIS)

The Neutral cluster and Air Ion Spectrometer is an improved version of the AIS (Manninen et al., 2009a). An additional charging-filtering section was added in order to measure the total number distribution of particles. The sample (assumed to be at or close to the bipolar charge equilibrium) goes through a unipolar corona charger. The charged fraction of particles induced to the sampled air is known for all sizes (estimated from Fuchs theory, Fuchs and Sutugin, 1971). The corona ions (generally < 2 nm depending on concentration, air composition, polarity, etc.) are removed by the electrical filters, leaving a confidence size range between 2 and 42 nm (Asmi et al., 2009). The NAIS measures, in turn, the mobility distribution of particles (from negative and positive DMAs, particle measurements), and of naturally negatively and positively charged particles and ions (ion measurements), and the background (offset measurements). The NAIS is also capable of measuring in the so-called alternative measurement mode, during which extra charging units, with the opposite polarity to the main charger, are turned on. This mode of operation should allow the retrieval of the total number concentration

Intercomparison of air ion spectrometers

S. Gagné et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

◀

▶

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



of particles even if the charge distribution of the particles in the atmosphere was not at bipolar equilibrium. The alternative measurement operation mode was not used during the workshop.

2.3 Airborne Neutral cluster and Air Ion Spectrometer (Airborne NAIS)

5 The Airborne Neutral cluster and Air Ion Spectrometer (ANAIS, Mirme et al., 2009) is a second generation NAIS. It was developed to measure at varying altitudes (and thus pressures and temperatures). This was done by implementing an automatic sheath airflow adjustment system that compensates for the change in particle electrical mobility due to the change in pressure and temperature. The ANAIS has 3 or 4 blowers, 10 depending on the model, to control the flows separately (one for the sheath air of each mobility analyzer, one or two connected in series for the sampling line), while first generation NAISs had only one central blower. The data acquisition system was also upgraded to adjust to the conditions so that the data is always adjusted to NTP conditions. The charger current in the corona charger is controlled to keep the efficiency of diffusion charging stable. All the changes were made to insure that the DMA parameters of the ion spectrometer remain constant and the same data inversion procedure can be used. A new, 2nd generation inversion program was also developed along with the instrument. This new version of the NAIS can be easily used in changing atmospheric conditions, and requires little maintenance.

20 3 Measurement setups and procedure

During the calibration period (25 May 2009–26 June 2009), we used mobility standards, silver ions and silver particles to investigate the measurements of mobility and concentration. The flows were also adjusted prior to the calibrations, to insure the accuracy of the comparison. In this section, we will describe the different experimental setups (or stations) used to calibrate the ion spectrometers. Four main stations 25

Intercomparison of air ion spectrometers

S. Gagné et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

◀

▶

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



Intercomparison of air ion spectrometers

S. Gagné et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

◀

▶

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



can be distinguished: the flow adjustment station, the Hauke-type DMA station (4–40 nm particles), the high resolution DMA station (HDMA, < 5 nm particles), and the intercomparison station. The high resolution DMA setup allows for two different sub-
 5 setups: using mobility selected silver ions and mobility standards (Ude and Fernández de la Mora, 2005).

The Airborne NAIS as well as the AIS 2, which had not been calibrated in the first calibration workshop but participated in the EUCAARI measurement campaign, were calibrated for the first time. The AIS 5, that was present in the first calibration workshop, was not available this time. The same mobility and concentration calibration were per-
 10 formed on 5 AISs, 5 NAISs and the Airborne NAIS (ANAIS). The calibration procedures that were performed are summarized in Table 1.

3.1 Flow adjustment setup

When the instruments arrived in Helsinki, they were thoroughly cleaned and sent to the flow adjustment station to ensure the best mobility and concentration measurements. The different parts of the flow system were adjusted while keeping the sampling flow at ~ 60 lpm. In all ion spectrometers, except the ANAIS, five flows (sheath and sample
 15 flows for each analyzer ($2 \times (1 + 1) = 4$ flows) and the inlet/outlet flow, common to both analyzers) share one central blower. Each of those five flows were measured through the pressure drop in Venturi tubes, and adjusted if required. The flow balance was
 20 verified for leaks in the same fashion as described by Asmi et al. (2009), before sending the ion spectrometer further to mobility and concentration calibration. All the pressure drops had been stable during the field campaign, provided that the Venturi tubes were unobstructed, and only small adjustments were made. This means that, if maintenance cleaning is done regularly, the instruments can perform well for long periods in diverse
 25 field conditions.

3.2 HAUKE DMA

One of the purposes of this station was to compare the mobility diameter measured with the ion spectrometers with the mobility diameter selected with a Hauke-type DMA (10.9 cm in length). Also, the concentrations given by the ion spectrometers was compared with a Condensation Particle Counter (CPC, TSI 3025, Stoltzenburg and McMurry, 1991) and an aerosol electrometer (TSI 3068A).

Polydisperse silver particles were produced with a tube furnace (Carbolite Furnaces MFT 12/388), then sent into a bipolar charger (^{241}Am) to get charged and then size selected with a DMA. The particle size varied between 4 and 40 nm. The sample flow in the DMA was 4 lpm and the sheath air flow was 20 lpm. To make up for the total sampling flow of the detection instrument (ion spectrometers 60 lpm, CPC: 1.5 lpm, and aerosol electrometer: 3 lpm), 63 lpm of diluting air was introduced after the DMA. The 50% cut-off size of the CPC was 3 nm, while the electrometer, in principle, detects all ions and charged particles as long as the concentration is above $\sim 300 \text{ cm}^{-3}$.

The ion spectrometers were operated in two or three of their operation modes: ions, particles (in the case of NAISs only) and offset mode. Concentration and mobility measurements were made where the concentration or the mobility was kept stable over several measurement cycles of the ion spectrometers. We also performed measurements of the ion spectrometer DMA transfer function in which the mobility was scanned stepwise, each step spanning over one complete measurement cycle. The mobility range varied between 0.0014 and $0.13 \text{ cm}^2 \text{ V}^{-1} \text{ s}^{-1}$ ($\sim 4\text{--}40$ nm) and was divided over 100 size bins.

3.3 High resolution DMA (HDMA)

The high resolution DMA, or Herrmann DMA (HDMA, Eichler, 1997; de Juan and Fernández de la Mora, 1998; Ude and Fernández de la Mora, 2005; Herrmann et al., 2000) was used in mobility standard and silver ion calibrations as well as for transfer function measurements. The very high size/mobility resolution of the HDMA, due to

Intercomparison of air ion spectrometers

S. Gagné et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

◀

▶

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



its high sheath flow rate (more than 1000 lpm), allows for the very precise selection of the mobility and the use of small particle sizes (< 5 nm). The width of the DMA transfer function is much smaller than the resolution of the ion spectrometers. The HDMA was calibrated every day before starting the measurements, for establishing an accurate voltage to mobility conversion. The concentration was measured with an electrometer, in parallel with the ion spectrometers.

3.3.1 Calibration with mobility standards

Mobility standards were used for mobility calibration at particle sizes smaller than 3 nm in diameter. For producing positive standards, we used THAB (tetra-heptyl ammonium bromide) and TMAI (tetra-methyl ammonium iodide), and for negative mobilities only THAB was used. Those compounds were electrosprayed, producing singly charged ions with known mass and mobility (Ude and Fernández de la Mora, 2005). Four such positive ions and four negative ions were selected using the HDMA. Each peak was measured at different concentrations.

The mobilities of the positively charged standards were 2.18 (TMAI derived TMA⁺ ions), 0.97, 0.65 and 0.53 cm² V⁻¹ s⁻¹ (THAB derived THA⁺, (THAB)THA⁺ and (THAB)₂THA⁺, resp.) at NTP. The standards used here are the same that were used in the first calibration workshop and their values are well established.

Four mobility peaks were selected from the negative THAB spectrum as well, even though their composition has not been verified by mass spectrometry and their stability is not known. The mobilities of the peaks that were used are: 2.45 (most probably Br⁻ ions), 1.48, 0.87 and 0.64 cm² V⁻¹ s⁻¹.

3.3.2 Calibrations with silver ions

The transfer function of the ion spectrometers for small particles was also measured using silver ions, mobility-selected with a HDMA. The silver particles were produced using a tube furnace and charged downstream with a bipolar charger (²⁴¹Am) before

Intercomparison of air ion spectrometers

S. Gagné et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

◀

▶

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



**Intercomparison of
air ion spectrometers**S. Gagné et al.

[Title Page](#)[Abstract](#)[Introduction](#)[Conclusions](#)[References](#)[Tables](#)[Figures](#)[◀](#)[▶](#)[◀](#)[▶](#)[Back](#)[Close](#)[Full Screen / Esc](#)[Printer-friendly Version](#)[Interactive Discussion](#)

entering the HDMA. Below about 2 nm in mobility diameter, the particles were probably a mixture of ions produced in the charger and silver particles (Asmi et al., 2009). For positively charged particles, we selected 81 logarithmically spaced mobilities between 0.09 and 1.10 cm² V⁻¹ s⁻¹ (~ 1.4–4.8 nm). For negatively charged particles, the mobilities ranged between 0.38 and 1.48 cm² V⁻¹ s⁻¹ (~ 1.2–2.3 nm) separated over 51 mobility bins.

3.4 Intercomparison

The intercomparison took place in a classroom at the University of Helsinki Kumpula campus. The room was situated on the 4th floor of the physics building “Physicum”. The room had a door leading to a large balcony and another one leading to a corridor with offices. The room’s surface was about 42 m² with a height of about 4 m. When the door to the balcony was opened, the particle concentration rose rapidly. The ion spectrometers were measuring in this room whenever they were not being calibrated, or in cleaning or repair. The ion spectrometers measured indoor air, or mixed indoor and outdoor air, and during new particle formation and growth. The NPF event was provoked by peeling citrus fruits in the middle of the room at approximately the same distance from each instrument. There was no additional fan system insuring that the air was well mixed, however all ion spectrometers were able to detect the new particle formation almost instantly and simultaneously. After each original event, secondary events always took place a few hours later in the same closed room.

3.4.1 Accompanying instruments

Three other instruments were measuring in the same room along with the AIS and NAIS ion spectrometers: a BSMA, a DMPS and an Ion-DMPS. In this section, we describe those three instruments.

**Intercomparison of
air ion spectrometers**

S. Gagné et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

◀

▶

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



The Balanced Scanning Mobility Analyzer (BSMA, Tammet, 2006) measures the number size distribution of cluster ions and naturally charged particles in the size range 0.7–7 nm (Stokes-Millikan mobility diameters). The BSMA consists of two parallel plane-type DMAs for negative and positive ion classification and one common electrical amplifier as a detector. Here, the detector measures the electrical currents of air ions. The BSMA measures negative and positive ion spectra one after the other – not simultaneously. Due to high electrometer sensitivity, high flow rates and small wall losses, the BSMA is typically used as a reference for small ion concentrations. The BSMA agrees well with other instruments (see e.g. Hirsikko et al., 2005; Kulmala et al., 2007; Manninen et al., 2009a and Ehn et al., 2010).

The Differential Mobility Particle Sizer (DMPS, Aalto et al., 2001) measured the particle number size distribution between 10 and 300 nm in diameter. The DMPS consisted of a Hauke-type DMA (length 28.0 cm) in closed loop sheath flow arrangement (Jokinen and Mäkelä, 1996), a CPC (TSI 3025, Stolzenburg and McMurry, 1991) as a particle detector and a radioactive C-14 alpha neutralizer (370 MBq). Sampled particles were charged in an alpha-active bipolar charger and classified according to their electrical mobility in the DMA. Subsequently, the classified particles were counted by a particle detector (CPC), and the total concentration was retrieved after standard DMPS inversion. The DMPS was not calibrated, but the transfer function of the DMA and CPC were known from earlier calibration measurements.

The Ion-DMPS (Laakso et al., 2007b) is identical to a DMPS, with the exceptions that its bipolar charger can be switched on or off and the voltage applied to the DMA can be either positive or negative. The Ion-DMPS thus measured in four modes: positively charged particles neutralized (1. +neu.) or ambient (2. +amb.); negatively charged particles neutralized (3. –neu.) or ambient (4. –amb.). During the intercomparison, the Ion-DMPS was measuring particles between 2.2 and 11.5 nm in mobility diameter. The calibration curves of the instrument can be found in Laakso et al. (2007b). This instrument allowed for the calculation of the charge ratio of the particle size distribution by comparing the concentration of charged particles in the neutralized sample (bipolar

steady-state) and in the ambient sample. The charged fraction of particles can be calculated based on the bipolar steady-state charging probabilities (Wiedensohler, 1988).

4 Results and discussion

The performance of the 11 ion spectrometers (5 AISs, 5 NAISs, one Airborne NAIS) was tested against reference devices and compared to each other (Table 1), after a one-year field measurement campaign. Also, all the instruments measured a NPF event simultaneously for the first time. In this section, we discuss the mobility and concentration calibrations as well as the intercomparison and the new particle formation event.

4.1 Mobility, mobility standards and concentration calibrations

All 11 ion spectrometers measured electrospray-generated mobility standards according to the method described in Sect. 3.3.1. Results from three examples of mobility standards of each polarity are presented in Fig. 1. The AISs and NAISs alike detected the mobility of the concentration peak almost accurately and compare well with each other. A shift can be observed with the TMA⁺ ion (the smallest positive standard) and with negative standards. Part of the inaccuracies could result from clustering or fragmentation of the ions after mobility selection. All inaccuracies for negative standards point to an underestimation of the mobility, thus an overestimation of the particle's size. Another observation that can be made on Fig. 1 is that the AISs display a sharper peak than the NAISs. This was also seen in the previous workshop, as shown by Asmi et al. (2009). The fact that normalized concentration peaks are broader for NAISs than for AISs could suggest that the NAISs have a small background present in other size bins, in the ion measurement mode, that contribute to the total concentration. It could also be due to more turbulent flows in the NAISs than in the AISs, making particles land on the neighbouring electrometers.

Intercomparison of air ion spectrometers

S. Gagné et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

◀

▶

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



**Intercomparison of
air ion spectrometers**

S. Gagné et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

◀

▶

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



Figure 2 shows the results from the calibrations done with silver particles with the HDMA setup (< 5 nm) and the Hauke-DMA setup (4–40 nm). In the first column, we compare the mobilities detected by the ion spectrometers (the peak of the mobility distribution) and the mobilities selected by the DMAs. In the second column, the concentrations seen by the ion spectrometers and the reference instruments are compared. Finally, in the third column, the ratio of the concentrations detected by the ion spectrometers and the reference instruments is shown as a function of the mobility. The first and second rows show the negative and positive ions, respectively. Background measurements are presented in Fig. 3. We used the DMA as a filter by applying a zero voltage so that the ion spectrometers were measuring particle-free air. In the following analysis, we will first focus on the ion data (first two rows in Fig. 2), then we will discuss the results from the particle mode (third row in Fig. 2).

In ion measurement mode of the ion spectrometers, the mobilities were detected very accurately. In the positive mode at small diameters, however, the mobility was slightly overestimated. This was also observed by Asmi et al. (2009). In the second column, the concentrations are compared to those measured with an electrometer. The concentration detection is rather good, but the NAISs overestimate the concentrations, and this is especially noticeable for the positive polarity. The concentration for the ion spectrometers in Fig. 2 is the total concentration integrating the whole size range of the instrument. In the third column, the ratio of the concentrations detected with the ion spectrometers to the electrometer concentration is shown as a function of mobility. The AISs and NAISs behaved similarly at larger diameters (smaller mobilities) but the NAISs overestimated the concentration at smaller diameters. The Airborne NAIS generally followed the behaviour observed in first generation NAISs. One should note that the two setups cover different size ranges (Hauke DMA: $\sim 4\text{--}40$ nm = $0.0014\text{--}0.13$ cm² V⁻¹ s⁻¹; HDMA: $< \sim 5$ nm = > 0.083 cm² V⁻¹ s⁻¹). Hence, we have a gap for negative ions (the voltage supply only allowed for a smaller mobility range) and double lines for positive ions (the methods overlap on a small mobility range). The two methods are not in perfect agreement. This may be explained by that, at small sizes, the

**Intercomparison of
air ion spectrometers**

S. Gagné et al.

transfer function of the Hauke DMA is wider and the concentration of silver particles available is also smaller and the losses (and corrections) are more important. One can see in the third column that the concentrations are overestimated more as the diameter decreases for both setups. Part of this overestimation can be due to the setup, since the ratio falls back to one when the setup changes from the Hauke DMA to the Herrmann DMA. However, despite corrections for losses in the experimental setups, the trend remains. In the left column of Fig. 3, one can see the background of the ion spectrometers in ion measurement mode as a function of the diameter. The median total background concentration was 13.6 (23.5) cm^{-3} for AISs and 71.7 (53.2) cm^{-3} for NAISs for negative (and positive) ions.

In the particle measurement mode of the ion spectrometers, the mobility was detected accurately. In the case of the Airborne NAIS, the positive polarity DMA was underestimating the mobility while its negative polarity DMA was accurate. This can either be due to a temporary malfunction of the instrument (e.g. changes in flows) or to a difference between the positive and negative DMA data inversion. The large variation between the instruments for smaller particles is most probably due to corona ions being detected up to about 5 nm (0.08 cm^2 V^{-1} s^{-1}), making the maximum of the measured mobility distribution difficult to find.

In the case of the particle measurement mode, the total concentration was calculated from 3.4 nm (0.18 cm^2 V^{-1} s^{-1}) instead of calculating the total concentration between 0.8 and 42 nm. This was done in an attempt to avoid including most of the background corona ions (see Fig. 3). The smallest size selected with the Hauke DMA was 4 nm, the DMA had a transmission width of 0.3 nm. To avoid the corona ions completely, we changed the minimum size integrated in the total concentration from 3.4 to 5.6 nm, but the overestimation of the concentration by the ion spectrometers remained. The concentration ratio with the reference instrument was still between 2 and 3 for first generation NAISs, although it decreased by about 0.2 compared to the ratio presented in Fig. 2f. This means that, even when avoiding corona ions, the NAISs seem to overestimate the concentration, especially in particle measurement mode. The concentration

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

◀

▶

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



Intercomparison of air ion spectrometers

S. Gagné et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

◀

▶

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



ratio does not vary much as a function of particle size for the NAISs. However, the Airborne NAIS overestimated the concentrations more significantly at bigger sizes than at smaller sizes. This is most likely due to a different inversion program provided by the manufacturer with the second generation models. The background concentration in particle mode is presented on the right side of Fig. 3. Most of the background is probably corona ions from the chargers. The median total background was $1.6 \times 10^5 \text{ cm}^{-3}$ for the negative DMA and $2.5 \times 10^6 \text{ cm}^{-3}$ for the positive DMA. When we applied the same minimum size of 3.4 nm to avoid corona ions, the background became much smaller with medians of 71.9 cm^{-3} and 86.5 cm^{-3} for the negative and positive DMA respectively. It is important to note that the background varies with the particle concentration. When there are particles in the sample air, the corona ions charge the particles and do not contribute to the background anymore. It is thus difficult to evaluate what fraction of the concentration is imputable to the background.

Concentration calibrations have shown that the efficiency of the ion spectrometers does not depend on the concentration, so that the ratio of the ion spectrometer concentration to a reference instrument concentration remains the same as a function of the concentration. This was observed both by detecting 15 nm particles at varying concentrations and by plotting the concentration ratios presented in Fig. 2 as a function of concentration.

4.2 Intercomparison

The intercomparison period lasted roughly 20 days between 31 May 2009 and 22 June 2009. During that time, the instruments were measuring in the intercomparison room as described in Sect. 3.4, unless they were being calibrated, cleaned, fixed, or malfunctioning. The days on which most of instruments were measuring uninterruptedly in the room were on 6–7 and 13–14 June 2009.

4.2.1 Indoor and outdoor air measurements

On 6 June 2009, the instruments were measuring indoor air (Fig. 4a–d) and on 7 June 2009, the instruments were measuring outdoor air through an opened door giving on a balcony on the highest floor of the building (Fig. 4e–h). Here, we present the median concentration and current for each size channel. In both cases, if the median was a negative value, the value was replaced by $1\text{e-}2\text{ cm}^{-3}$ for the concentrations and by $1\text{e-}3\text{ fA}$ for the currents in order to facilitate visualization in logarithmic scale.

In the case of indoor air (Fig. 4a–d), the median concentration varied between about 0.01 and 10 000 particles per cm^3 per size channel (logarithmically spaced). The concentrations agreed fairly well from one instrument to another, especially for negatively charged particles. Once again, the NAISs display bigger concentrations than the AISs. This tendency is also observed when looking at the raw electrometer signal on the right panels. In the case of outdoor air (Fig. 4e–h), the median concentration varied between about 1 and 2000 particles per cm^3 per size channel. The instruments agree well with each other, again, better for negatively charged particles than for positively charged particles. One can notice that the concentration of small particles is bigger in the indoor air, while the concentration of large particles ($> 10\text{ nm}$) is bigger in outdoor air. This is probably due to coagulation of smaller particles onto the numerous larger particles found in outdoor air.

As can be seen from Fig. 4, the current of the AIS7 behaves like all other AISs, however, its concentrations are smaller than the other AISs in certain size ranges. The difference between the AIS7 and the other was caused by that the AIS7 was using different inverters, and had a much shorter cycle. The AIS7 cycle was about 10 times shorter than the other AISs. This made noise levels more important in respect to the concentration, leading to the smaller concentrations. We re-inverted the currents using an inversion matrix that takes into account the shorter measurement cycle time. The new concentrations are presented in Fig. 4 as dashed lines. One can see that the re-inverted output is much closer to the other AISs. Hence, we would like to point out

Intercomparison of air ion spectrometers

S. Gagné et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

◀

▶

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



the importance of using the appropriate inverters and upgrade the software in order to be comparable with other (N)AIS. This also shows that longer cycles are recommendable to allow better noise control leading to a better signal to noise ratio, and a better inversion of the currents.

4.2.2 Mean diameter and concentration comparison

The mean concentrations and diameters of ions and particles detected during the intercomparison workshop were calculated and are presented in Fig. 5. In general, all instruments showed similar concentrations and diameters, for both polarities, and also for neutral particles. However, in the case of negatively charged particles, measured with the ion measurement mode of the spectrometers, the AISs showed slightly smaller concentrations than the NAISs (Fig. 5a), consistent with the calibration results and the measurements of indoor and outdoor air. The BSMA, the reference instrument in this figure, agrees better with the AISs. For positively charged particle, the median concentrations varied more than for negatively charged particles. It is thus impossible to say whether the NAISs are overestimating the concentration for positively charged particles in this figure.

The concentrations measured with the NAISs in the particle measurement mode can be compared to the DMPS as a reference instrument (Fig. 5b). The concentrations measured by the ion spectrometers were all in the same range. Once again, the negative analyzers agreed better with each other than the positive analyzers. For both polarities, the DMPS yielded smaller concentrations, suggesting that the NAISs may be overestimating the concentrations. The only differences between the AISs and NAISs are the extra charging modules found in the NAISs (causing a more turbulent flow) and a slightly different inversion process that takes into account the diffusional losses happening in these extra modules. The cause for the different concentrations, in this case, is most likely due to the differences in the flow turbulence and the data inversion process.

Intercomparison of air ion spectrometers

S. Gagné et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

◀

▶

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



**Intercomparison of
air ion spectrometers**

S. Gagné et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

◀

▶

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



The size of small ions, smaller than ~ 3 nm, is influenced by air composition, temperature, pressure and sink due to aerosol particles (see e.g. Luts and Parts, 2002; Parts and Luts, 2004). The mean mobility or size of small ions has thus been reported in the literature (see Hirsikko et al., 2011 and references therein). During the intercomparison period, the diameter distribution of < 3 nm particles varied widely from one instrument to another (Fig. 5c). No systematic difference between the AISs and the NAISs was observed, however the BSMA measured smaller diameters than the ion spectrometers. The median diameter of small ions varied between 1.25 and 1.45 nm with the exception of the AIS 7's positive mode and AIS 1's negative mode. The median size of small ions can give indications about their chemical composition.

As can be seen from Fig. 5d, the NAISs tend to measure higher concentrations than the AISs, at least for negatively charged particles. In general, the BSMA and the ion spectrometers agree well with each other.

4.2.3 Measurements during a NPF event

Three NPF events were provoked in the room during the intercomparison: on the 10, 12 and 14 June 2009. The NPF events were provoked by peeling citrus fruits in the middle of the room. The clearest event, with the most instruments monitoring it, was on 12 June 2009 and a sample of the results is presented in Fig. 6. What may look like a burst at ca. 15:30 is only due to briefly opening the door giving on the outside air. The fruits were peeled only at the start of the first burst at ca. 17:20. Secondary, weaker events started after the main one at 18:50 and 20:00. The core of the analysis will focus on the first secondary event, starting at 18:50.

The first event was very strong and fast, so that it was impossible to calculate the growth and formation rates accurately. After this event, the population of small ions was almost completely depleted. This is probably partly due to the high coagulation sink of the newly formed particles and partly to the activation of these small ions into new particles. The small ion population slowly rebuilt allowing the formation and growth of

new particles (2nd and 3rd events). The concentration of small ions returned to normal levels once new particles had ceased to form, indicating the activation of small air ions during the event.

The charged fraction, which is the fraction of particles that are electrically charged, can be calculated for the NAIS data. It is presented for 12 June 2009 in Fig. 7. The charged fraction is the ratio of the concentration of charged particles (negatively + positively charged particles in the ion measurement mode) divided by the concentration of particles (NAIS particle measurement mode) for corresponding size ranges. Figure 7 shows the charged fraction of negatively (Fig. 7a–f) and positively (Fig. 7g–l) charged particles for 6 diameters corresponding to 6 channels of the Ion-DMPS. Since the NAISs provide two different measurements of the particle measurement mode (one for each DMA), the charged fraction was calculated for each DMA i.e. negative (positive) charged fraction = negative (positive) ion concentration ÷ particle concentration from the negative (positive) DMA. The charged fractions calculated from malfunctioning channel/polarity were removed from the figure.

The charged fractions measured with the NAISs were compared to the ones measured with the Ion-DMPS. The Ion-DMPS charged ratio was calculated as the ratio of the ambient mode concentration to the neutralized mode concentration (at steady-state charged fraction). The charged ratio was then multiplied by the steady state charging probability for the appropriate diameter (Wiedensohler, 1988) to get the charged fraction of the particles. The charged fraction derived from the Ion-DMPS data agrees with the range of charged fractions calculated with the different NAIS data. It is possible that the NAISs overestimate the total concentration in total operation mode when the charged fraction of the aerosol sample is much bigger than the equilibrium of the unipolar charger (overcharged situations). In most field situations this phenomena seems to be rare (see Manninen et al., 2010 and references therein), however the subject requires further investigation.

Particle formation rates at 2 nm (Manninen et al., 2009b) were calculated from each instrument during the second event on 12 June. The formation rate is the rate at which

Intercomparison of air ion spectrometers

S. Gagné et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

◀

▶

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



**Intercomparison of
air ion spectrometers**

S. Gagné et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

◀

▶

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



new particles are formed and appear in a given size range, in our case, at 2 nm (Kulmala et al., 2007). Values are shown in Table 2 for all modes and instruments for which it was possible to get a value. One has to bear in mind that the formation and growth rates calculated with the method of using the methods of Manninen et al. (2009) can be trusted within a factor 2, and that using different methods may cause an even bigger difference. The new particle formation events were mostly driven by negatively charged particles as is often observed in other environments, in laboratory experiments and in quantum chemistry calculations (see e.g. Vana et al., 2006; Gagné 2010; Winkler et al., 2008; Kurtén et al., 2009). For that reason, it was impossible to calculate the formation rate in the positive mode for all AISs and some NAISs. The formation rates calculated from the AISs' negative mode were smaller than those calculated from the NAISs' same mode. This is consistent with the finding that NAISs seem to overestimate the ion concentration, especially at small diameters where the formation rates were calculated. The formation rate of ions calculated based on the BSMA was higher than those of the other ion spectrometers, closer to the NAISs than the AISs. During the inter-comparison, the BSMA agreed better with the AISs because small ions (< 2 nm) were dominating the size distribution. However, during the NPF event, the concentration of small ions decreased and the concentration of larger particles increased. The BSMA agreed better with the NAISs in the larger size range.

The formation rate in the particle mode was calculated in the case of NAISs using the same method. The formation rate of all particles at 2 nm was also calculated from the DMPS. Due to the measurement range of the DMPS, the formation rate was first calculated at 10 nm and then scaled back to 2 nm using the formula described by Kerminen and Kulmala (2002) to be comparable to those calculated with the ion spectrometers. The formation rate J_2 of ions was found to be around 0.1–0.3 ($\text{cm}^{-3} \text{s}^{-1}$). The BSMA gave the highest rate ($0.27 \text{ cm}^{-3} \text{ s}^{-1}$) while the AISs and the NAISs gave an average of $0.15 \text{ cm}^{-3} \text{ s}^{-1}$ and $0.22 \text{ cm}^{-3} \text{ s}^{-1}$ respectively. The difference between the AISs and the NAISs is not a surprise since the NAISs seem to have a tendency to overestimate the concentrations. In particle mode, the DMPS gave a formation rate of $1.1 \text{ cm}^{-3} \text{ s}^{-1}$

while the NAIS average was $2.7 \text{ cm}^{-3} \text{ s}^{-1}$, again this is not too surprising given the results given in the 3rd row of Fig. 2.

Also, the growth rates of particles were calculated for the AISs, NAISs DMPS and BSMA. The growth rate is the rate at which particles in a given size range grow. The results are presented in Table 2. Due to the small concentrations in the positive mode, it was impossible to calculate them for all instruments. The growth rates were calculated in 3 different size ranges: 2–3 nm, 3–7 nm and 7–20 nm. As observed before, the growth is slower at small diameters and faster at bigger diameters (Hirsikko et al., 2007; Yli-Juuti et al., 2009). The AISs gave an average growth rate of 13.8, 27.0 and 35.5 nm h^{-1} in each size range, respectively. The NAISs, on the other hand showed growth rates of 13.6, 24.5, 38.1 nm h^{-1} for the negative polarity and slightly smaller values for the positive polarity. The BSMA yielded a growth rate of 11.4 nm h^{-1} in the 2–3 nm size range and 23.0 in the 3–7 nm range. The DMPS gave a growth rate of 37 nm h^{-1} in the 10–20 nm size range. No clear difference between the AISs and the NAISs was observed in the determination of the growth rate. This result can be explained by an accurate detection of the mobility for both types of instruments as shown in Figs. 1 and 2.

5 Conclusions

In this work, we inspected 11 ion spectrometers that had been deployed in field measurements during the EUCAARI project in 2007–2009. Three types of ion spectrometers were characterized and compared: 5 AISs, 5 NAISs and one second generation NAIS that is called an Airborne NAIS (ANAIS). We evaluated the response of the instruments regarding mobility (particle size) and concentration using mobility standards and silver particles in two experimental setups covering a size range between roughly 1 and 40 nm (the measurements performed are resumed in Table 1). We also had all the ion spectrometers and three reference instruments (BSMA, DMPS, ion-DMPS) measuring ambient indoor and outdoor air during the workshop, and compared their

Intercomparison of air ion spectrometers

S. Gagné et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

◀

▶

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



response. In addition, we provoked a new particle formation event in the room air that was monitored by all ion spectrometers side by side with the reference instruments. The formation and growth rates as well as the charged fraction were calculated and compared for all ion spectrometers and relevant reference instruments.

Based on the results presented in this paper, we present a number of results to keep in mind while performing analysis of air ion spectrometer data:

1. The mobility detection can be trusted for AISs and NAISs, provided that the instrument is clean and the flows are not obstructed.
2. The growth rates calculated from the ion spectrometer data are reliable (as a consequence of conclusion 1 and the analysis of a NPF event).
3. The concentration can vary from one individual instrument to the other by up to “10%” within the same instrument type.
4. In ion measurement mode, the NAISs give higher concentrations than the AISs, the AISs agreed better with the BSMA.
5. The NAISs can overestimate the concentration by a factor of 2–3 in particle measurement mode (based on calibration results and comparison with a DMPS).
6. The formation rates vary from one individual instrument to the other (based on conclusion 3; it is also important to note that formation rates can be trusted within a factor 2 with the method described by Manninen et al. (2009)).
7. The formation rates calculated from NAISs are higher than those calculated from AISs in ion measurement mode. The BSMA agreed better with the NAISs.
8. The formation rates of particles calculated from NAISs (in particle measurement mode) are higher than the reference instrument (based on conclusion 5 and on the analysis of a NPF event where it was compared to a DMPS).

Intercomparison of air ion spectrometers

S. Gagné et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

◀

▶

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



**Intercomparison of
air ion spectrometers**

S. Gagné et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

◀

▶

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



9. The charged fraction calculated from NAISs give a good estimate of the charged fraction and its behavior in time, at least when the sample is not too overcharged (based on the analysis of a NPF event where it was compared to an Ion-DMPS).

The ion spectrometers proved to be good mobility detectors in ion measurement mode (Fig. 1). In particle measurement mode, some NAIS models had better accuracy than others (Fig. 2). The concentration measurements were in good agreement with the reference instruments for AIS models (Figs. 2 and 5). The ion spectrometers seemed to overestimate the concentrations as the particle size decreases. This can be partly explained by uncertainties with the different experimental setups, but a general tendency remains (Fig. 2). The NAIS models also detected the concentration well, but slightly overestimated the concentration compared to the AISs and the reference instruments. This was observed both in calibration measurements (Fig. 2) and in ambient measurements (Figs. 4 and 5a). Moreover, the NAISs had bigger background concentrations than the AISs (Fig. 3). In particle measurement mode, the NAISs overestimated the concentration by a factor 2 to 3, again both in calibration (Fig. 2) and in ambient measurements (Fig. 5b). The overestimation in the particle measurement mode seemed to be independent of the mobility for all NAISs except for the Airborne NAIS (Fig. 2). This was attributed to the different inversion process that the ANAIS and second generation NAISs use. At smaller particle sizes, the concentration of the ANAIS data seemed to be closer to the reference values, whereas the concentration was overestimated at bigger particle sizes.

A new particle formation event was detected by all the ion spectrometers as well as a BSMA, DMPS and an Ion-DMPS. The formation rate J_2 of ions was found to be around 0.1–0.3 ($\text{cm}^{-3} \text{s}^{-1}$). The BSMA gave a rate of $0.27 \text{ cm}^{-3} \text{ s}^{-1}$, the highest rate of all instruments. The AISs gave an average of $0.15 \text{ cm}^{-3} \text{ s}^{-1}$ and the NAISs (including the ANAIS) an average of $0.22 \text{ cm}^{-3} \text{ s}^{-1}$. It is not surprising to get higher formation rates from the NAISs than from the AISs given that the NAISs show higher concentrations than the AISs. In particle measurement mode, the formation rate at 2 nm scaled back from the DMPS measurements was $1.1 \text{ cm}^{-3} \text{ s}^{-1}$ while the average

for NAISs was $2.7 \text{ cm}^{-3} \text{ s}^{-1}$. We thus recommend that this result be kept in mind when calculating formation rates from an NAIS, especially if compared with an AIS or with a DMPS/SMPS system.

The differences between the calculated formation rates may also have an effect on the ion-induced fraction, which has been previously calculated as the ratio of the ion formation rate to the particle formation rate (see e.g. Manninen et al., 2009b, 2010). Our results in Fig. 2 suggest that the overestimation is more important in particle measurement mode than in ion measurement mode. Thus the charged ratio, representing a part of the ion-induced fraction calculated from NAIS data, may be slightly underestimated (this does not affect the recombination part of the calculations). However, the charged fraction itself does not seem to be affected too greatly by this difference in formation rates (Fig. 7), probably because the charged fraction was not too high. The charged fraction calculated with the Ion-DMPS data was generally slightly higher than the one calculated with the NAISs data. The Ion-DMPS and the ion spectrometers yielded a similar behavior for the charged fraction, especially at larger diameters, where the Ion-DMPS is more reliable, suggesting that the charged fraction obtained from the NAISs is a good estimation, at least at larger sizes ($> 9 \text{ nm}$).

The growth rates, presented in Table 2, were similar for all the ion spectrometers, regardless of whether they were AISs or NAISs, and agreed well with the reference instruments. This is a direct result of the high performance of the instruments regarding mobility detection. The growth rates were about $14, 25, 35 \text{ nm h}^{-1}$ in the 2–3, 3–7, 7–20 nm size range, respectively.

The ion spectrometers evaluated in this paper performed well, despite having spent a year in varying weather conditions. They proved to be reliable and enduring instruments, although they need to be cleaned regularly to insure the quality of the data. Some systematic differences between the AISs and NAISs were observed, as well as minor differences between the first and second generation of NAISs. The reasons behind the overestimation of the concentration by NAISs did not become clear during the calibration workshop. Additionally, the inversion processes may need some

**Intercomparison of
air ion spectrometers**

S. Gagné et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

◀

▶

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



improvements, especially in particle measurement mode. A measurement-based data inversion could be a replacement solution to the present theory-based data inversion of the air ion spectrometers.

Acknowledgements. Jani Hakala and Jyri Mikkilä are acknowledged for their help in various tasks throughout the workshop. Ella-Maria Kyrö and Anna Frey are acknowledged for their help in preparing the instruments for calibration. This research was supported by the Academy of Finland Center of Excellence program (project number 1118615), the European Commission 6th Framework program project EUCAARI, contract no 036833-2 (EUCAARI) and the Estonian Science Foundation, grant number 8342. The work of A. Franchin and S. Schobesberger was supported by the European Commission under the 7th Framework Programme (grant agreement number 215072: Marie Curie Initial Training Network, CLOUD-ITN).

References

- Aalto, P., Hämeri, K., Becker, E., Weber, R., Salm, J., Mäkelä, J. M., Hoell, C., O'Dowd, C. D., Karlsson, H., Hansson, H.-C., Väkevää, M., Koponen, I. K., Buzorius, G., and Kulmala, M.: Physical characterization of aerosol particles during nucleation events, *Tellus* 53B, 344–358, 2004.
- Andronache, C., Grönholm, T., Laakso, L., Phillips, V., and Venäläinen, A.: Scavenging of ultrafine particles by rainfall at a boreal site: observations and model estimations, *Atmos. Chem. Phys.*, 6, 4739–4754, doi:10.5194/acp-6-4739-2006, 2006.
- Asmi, E., Sipilä, M., Manninen, H. E., Vanhanen, J., Lehtipalo, K., Gagné, S., Neitola, K., Mirme, A., Mirme, S., Tamm, E., Uin, J., Komsaare, K., Attoui, M., and Kulmala, M.: Results of the first air ion spectrometer calibration and intercomparison workshop, *Atmos. Chem. Phys.*, 9, 141–154, doi:10.5194/acp-9-141-2009, 2009.
- Crumeyrolle, S., Manninen, H. E., Sellegri, K., Roberts, G., Gomes, L., Kulmala, M., Weigel, R., Laj, P., and Schwarzenboeck, A.: New particle formation events measured on board the ATR-42 aircraft during the EUCAARI campaign, *Atmos. Chem. Phys.*, 10, 6721–6735, doi:10.5194/acp-10-6721-2010, 2010.

Intercomparison of air ion spectrometers

S. Gagné et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

◀

▶

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



**Intercomparison of
air ion spectrometers**

S. Gagné et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

◀

▶

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



- Duplissy, J., Enghoff, M. B., Aplin, K. L., Arnold, F., Aufmhoff, H., Avngaard, M., Baltensperger, U., Bondo, T., Bingham, R., Carslaw, K., Curtius, J., David, A., Fastrup, B., Gagné, S., Hahn, F., Harrison, R. G., Kellett, B., Kirkby, J., Kulmala, M., Laakso, L., Laaksonen, A., Lillestol, E., Lockwood, M., Mäkelä, J., Makhmutov, V., Marsh, N. D., Nieminen, T., Onnela, A., Pedersen, E., Pedersen, J. O. P., Polny, J., Reichl, U., Seinfeld, J. H., Sipilä, M., Stozhkov, Y., Stratmann, F., Svensmark, H., Svensmark, J., Veenhof, R., Verheggen, B., Viisanen, Y., Wagner, P. E., Wehrle, G., Weingartner, E., Wex, H., Wilhelmsson, M., and Winkler, P. M.: Results from the CERN pilot CLOUD experiment, *Atmos. Chem. Phys.*, 10, 1635–1647, doi:10.5194/acp-10-1635-2010, 2010.
- Eichler, T.: A Differential Mobility analyzer for ions and nanoparticles: Laminar flow at high Reynolds numbers, Senior Graduation Thesis presented to Fachhochschule Offenburg, Germany, May 1997.
- Ehn, M., Junninen, H., Schobesberger, S., Manninen, H., Franchin A., Sipilä, M., Petäjä, T., Kerminen, V.-M., Tammet, H., Mirme, A., Mirme, S., Hörrak, U., Kulmala, M., and Worsnop, D. R.: An instrumental comparison of mobility and mass measurements of atmospheric small ions, *Aerosol Sci. Tech.*, 45(4), 522–532, 2011.
- Fuchs, N. A. and Sutugin, A. G.: Highly dispersed aerosols, Vol H of *In Topics in Current Aerosol Research*, Pergamon Press, New York, 1971.
- Gagné, S., Nieminen, T., Kurtén, T., Manninen, H. E., Petäjä, Laakso, L., Kerminen, V.-M., Boy, M., and Kulmala, M.: Factors influencing the contribution of ion-induced nucleation in a boreal forest, Finland, *Atmos. Chem. Phys.*, 10, 3743–3757, doi:10.5194/acp-10-3743-2010, 2010.
- Harrison, R. G. and Carslaw, K. S.: Ion-aerosol-cloud processes in the lower atmosphere, *Rev. Geophys.*, 41, 1012, doi:10.1029/2002RG000114, 2003.
- Herrmann, W., Eichler, T., Bernardo, N., and Fernández de la Mora, J.: Turbulent transition arises at Re 35 000 in a short Viennatype DMA with a large laminarizing inlet, Abstract to the annual conference of the AAAR, St. Louis, MO, 6–10 October, 2000.
- Hirsikko, A., Laakso, L., Hörrak, U., Aalto, P. P., Kerminen, V.-M., and Kulmala, M.: Annual and size dependent variation of growth rates and ion concentrations in boreal forest, *Boreal Environ. Res.*, 10, 357–369, 2005.
- Hirsikko, A., Yli-Juuti, T., Nieminen, T., Vartiainen, E., Laakso, L., Hussein, T., and Kulmala, M.: Indoor and outdoor air ion and aerosol particles in the urban atmosphere of Helsinki: characteristics, sources and formation, *Boreal Environ. Res.*, 12, 295–310, 2007.

**Intercomparison of
air ion spectrometers**

S. Gagné et al.

[Title Page](#)[Abstract](#)[Introduction](#)[Conclusions](#)[References](#)[Tables](#)[Figures](#)[◀](#)[▶](#)[◀](#)[▶](#)[Back](#)[Close](#)[Full Screen / Esc](#)[Printer-friendly Version](#)[Interactive Discussion](#)

- Hirsikko, A., Nieminen, T., Gagné, S., Lehtipalo, K., Manninen, H. E., Ehn, M., Hörrak, U., Kerminen, V.-M., Laakso, L., McMurry, P. H., Mirme, A., Mirme, S., Petäjä, T., Tammet, H., Vakkari, V., Vana, M., and Kulmala, M.: Atmospheric ions and nucleation: a review of observations, *Atmos. Chem. Phys.*, 11, 767–798, doi:10.5194/acp-11-767-2011, 2011.
- 5 Jokinen, V. and Mäkelä, J.: Closed loop arrangement with critical orifice for DMA sheath/excess flow system, *J. Aerosol Sci.*, 28, 643–648, 1996.
- de Juan, L. and Fernández de la Mora, J.: Size analysis of nanoparticles and ions: Running a Vienna DMA of near optimal length at Reynolds numbers up to 5000, *J. Aerosol Sci.*, 29, 617–626, 1998.
- 10 Kerminen, V.-M. and Kulmala, M.: Analytical formulae connecting the “real” and the “apparent” nucleation rate and the nuclei number concentration for atmospheric nucleation events, *J. Aerosol Sci.*, 33, 609–622, 2002.
- Kerminen, V.-M., Petäjä, T., Manninen, H. E., Paasonen, P., Nieminen, T., Sipilä, M., Junninen, H., Ehn, M., Gagné, S., Laakso, L., Riipinen, I., Vehkamäki, H., Kurten, T., Ortega, I. K., Dal Maso, M., Brus, D., Hyvärinen, A., Lihavainen, H., Leppä, J., Lehtinen, K. E. J., Mirme, A., Mirme, S., Hörrak, U., Berndt, T., Stratmann, F., Birmili, W., Wiedensohler, A., Metzger, A., Dommen, J., Baltensperger, U., Kiendler-Scharr, A., Mentel, T. F., Wildt, J., Winkler, P. M., Wagner, P. E., Petzold, A., Minikin, A., Plass-Dülmer, C., Pöschl, U., Laaksonen, A., and Kulmala, M.: Atmospheric nucleation: highlights of the EUCAARI project and future directions, *Atmos. Chem. Phys.*, 10, 10829–10848, doi:10.5194/acp-10-10829-2010, 2010.
- 20 Kulmala, M., Laakso, L., Lehtinen, K. E. J., Riipinen, I., Dal Maso, M., Anttila, T., Kerminen, V.-M., Hörrak, U., Vana, M., and Tammet, H.: Initial steps of aerosol growth, *Atmos. Chem. Phys.*, 4, 2553–2560, doi:10.5194/acp-4-2553-2004, 2004a.
- Kulmala, M., Vehkamäki, H., Petäjä, T., Dal Maso, M., Lauri, A., Kerminen, V.-M., Birmili, W., and McMurry, P. H.: Formation and growth of ultrafine atmospheric particles: a review of observations, *J. Aerosol Sci.*, 35, 143–176, 2004b.
- 25 Kulmala, M., Riipinen, I., Sipilä, M., Manninen, H., Petäjä, T., Junninen, H., Dal Maso, M., Mordas, G., Mirme, A., Vana, M., Hirsikko, A., Laakso, L., Harrison, R. M., Hanson, I., Leung, C., Lehtinen, K. E. J., and Kerminen, V.-M.: Toward Direct Measurement of Atmospheric Nucleation, *Science*, 318, 89–92, doi:10.1126/science.1144124, 2007.
- 30 Kulmala, M., Asmi, A., Lappalainen, H. K., Carslaw, K. S., Pöschl, U., Baltensperger, U., Hov, Ø., Brenquier, J.-L., Pandis, S. N., Facchini, M. C., Hansson, H.-C., Wiedensohler, A., and O’Dowd, C. D.: Introduction: European Integrated Project on Aerosol Cloud Climate and Air

**Intercomparison of
air ion spectrometers**

S. Gagné et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

◀

▶

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



Quality interactions (EUCAARI) - integrating aerosol research from nano to global scales, Atmos. Chem. Phys., 9, 2825–2841, doi:10.5194/acp-9-2825-2009, 2009.

Kurtén, T., Ortega, I. K., and Vehkamäki, H.: The sign preference in sulfuric acid nucleation, J. Mol. Struct.-Theochem., 901, 169–173, 2009.

5 Laakso, L., Mäkelä, J. M., Pirjola, L., and Kulmala, M.: Model studies on ion-induced nucleation in the atmosphere, J. Geophys. Res-Atmos., 107, 4427, doi:10.1029/2002JD002140, 2002.

Laakso, L., Grönholm, T., Kulmala, L., Haapanala, S., Hirsikko, A., Lovejoy, E. R., Kazil, J., Kurtén, T., Boy, M., Nilsson, E. D., Sogachev, A., Riipinen, I., Stratmann, F., and Kulmala, M.: Hotair balloon as a platform for boundary layer profile measurements during particle
10 formation, Boreal Environ. Res., 12, 279–294, 2007a.

Laakso, L., Gagné, S., Petäjä, T., Hirsikko, A., Aalto, P. P., Kulmala, M., and Kerminen, V.-M.: Detecting charging state of ultra-fine particles: instrumental development and ambient measurements, Atmos. Chem. Phys., 7, 1333–1345, doi:10.5194/acp-7-1333-2007, 2007b.

15 Laakso, L., Laakso, H., Aalto, P. P., Keronen, P., Petäjä, T., Nieminen, T., Pohja, T., Siivola, E., Kulmala, M., Kgabi, N., Molefe, M., Mabaso, D., Phalatse, D., Pienaar, K., and Kerminen, V.-M.: Basic characteristics of atmospheric particles, trace gases and meteorology in a relatively clean Southern African Savannah environment, Atmos. Chem. Phys., 8, 4823–4839, doi:10.5194/acp-8-4823-2008, 2008.

20 Lehtipalo, K., Kulmala, M., Sipilä, M., Petäjä, T., Vana, M., Ceburnis, D., Dupuy, R., and O'Dowd, C.: Nanoparticles in boreal forest and coastal environment: a comparison of observations and implications of the nucleation mechanism, Atmos. Chem. Phys., 10, 7009–7016, doi:10.5194/acp-10-7009-2010, 2010.

Lohmann, U. and Feichter, J.: Global indirect aerosol effects: a review, Atmos. Chem. Phys., 5, 715–737, doi:10.5194/acp-5-715-2005, 2005.

25 Lovejoy, E. R., Curtius, J., and Froyd, K. D.: Atmospheric ion-induced nucleation of sulfuric acid and water, J. Geophys. Res., 109, D08204, doi:10.1029/2003JD004460, 2004.

Luts, A. and Parts, T.-E.: Evolution of negative small air ions at two different temperatures, J. Atmos. Sol.-Terr. Phys., 64, 763–774, 2002.

30 Manninen, H. E., Petäjä, T., Asmi, E., Riipinen, I., Nieminen, T., Mikkilä, J., Hörrak, U., Mirme, A., Mirme, S., Laakso, L., Kerminen, V.-M., and Kulmala, M.: Long-term field measurements of charged and neutral clusters using Neutral cluster and Air Ion Spectrometer (NAIS), Boreal Environ. Res., 14, 591–605, 2009a.

Manninen, H. E., Nieminen, T., Riipinen, I., Yli-Juuti, T., Gagné, S., Asmi, E., Aalto, P. P.,

**Intercomparison of
air ion spectrometers**

S. Gagné et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

◀

▶

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



Petäjä, T., Kerminen, V.-M., and Kulmala, M.: Charged and total particle formation and growth rates during EUCAARI 2007 campaign in Hyytiälä, *Atmos. Chem. Phys.*, 9, 4077–4089, doi:10.5194/acp-9-4077-2009, 2009b.

5 Manninen, H. E., Nieminen, T., Asmi, E., Gagné, S., Häkkinen, S., Lehtipalo, K., Aalto, P., Vana, M., Mirme, A., Mirme, S., Hörrak, U., Plass-Dülmer, C., Stange, G., Kiss, G., Hoffer, A., Töro, N., Moerman, M., Henzing, B., de Leeuw, G., Brinkenberg, M., Kouvarakis, G. N., Bougiatioti, A., Mihalopoulos, N., O'Dowd, C., Ceburnis, D., Arneth, A., Svenningsson, B., Swietlicki, E., Tarozzi, L., Decesari, S., Facchini, M. C., Birmili, W., Sonntag, A., Wiedensohler, A., Boulon, J., Sellegri, K., Laj, P., Gysel, M., Bukowiecki, N., Weingartner, E., Wehrle, G., Laaksonen, A., Hamed, A., Joutsensaari, J., Petäjä, T., Kerminen, V.-M., and Kulmala, M.: EUCAARI
10 ion spectrometer measurements at 12 European sites - analysis of new particle formation events, *Atmos. Chem. Phys.*, 10, 7907–7927, doi:10.5194/acp-10-7907-2010, 2010.

Matisen, R., Miller, F., Tamm, H. and Salm, J.: Air ion counters and spectrometers designed in Tartu University, *Acta et Comm. Univ. Tartuensis*, 947, 60–67, 1992.

15 Mirme, A., Tamm, E., Mordas, G., Vana, M., Uin, J., Mirme, S., Bernotas, T., Laakso, L., Hirsikko, A., and Kulmala, M.: A widerange multi-channel Air Ion Spectrometer, *Boreal Environ. Res.*, 12, 247–264, 2007.

Mirme, S., Mirme, A., Minikin, A., Petzold, A., Hörrak, U., Kerminen, V.-M., and Kulmala, M.: Atmospheric sub-3 nm particles at high altitudes, *Atmos. Chem. Phys.*, 10, 437–451, doi:10.5194/acp-10-437-2010, 2010.

20 Myhre, G., Berglen, T. F., Johnsrud, M., Hoyle, C. R., Berntsen, T. K., Christopher, S. A., Fahey, D. W., Isaksen, I. S. A., Jones, T. A., Kahn, R. A., Loeb, N., Quinn, P., Remer, L., Schwarz, J. P., and Yttri, K. E.: Modelled radiative forcing of the direct aerosol effect with multi-observation evaluation, *Atmos. Chem. Phys.*, 9, 1365–1392, doi:10.5194/acp-9-1365-2009, 2009.

25 Mäkelä, J. M., Riihelä, M., Ukkonen, A., Jokinen, V., and Keskinen, J.: Comparison of mobility equivalent diameter with Kelvin-Thomson diameter using ion mobility data, *J. Chem. Phys.*, 105, 1562–1571, 1996.


Parts, T.-E. and Luts, A.: Observed and simulated effects of certain pollutants on small air ion spectra: I. Positive ions, *Atmos. Environ.*, 38(9), 1283–1289, 2004.

30 Stevens, B. and Feingold, G.: Untangling aerosol effects on clouds and precipitation in a buffered system, *Nature*, 461, 607–613, 2009.

Stolzenburg, M. R. and McMurry, P. H.: An Ultrafine Aerosol Condensation Nucleus Counter,

**Intercomparison of
air ion spectrometers**

S. Gagné et al.

[Title Page](#)[Abstract](#)[Introduction](#)[Conclusions](#)[References](#)[Tables](#)[Figures](#)[◀](#)[▶](#)[◀](#)[▶](#)[Back](#)[Close](#)[Full Screen / Esc](#)[Printer-friendly Version](#)[Interactive Discussion](#)

Aerosol Sci. Tech., 14, 48–65, 1991.

Suni, T., Kulmala, M., Hirsikko, A., Bergman, T., Laakso, L., Aalto, P. P., Leuning, R., Cleugh, H., Zegelin, S., Hughes, D., van Gorsel, E., Kitchen, M., Vana, M., Hörrak, U., Mirme, S., Mirme, A., Sevanto, S., Twining, J., and Tardos, C.: Formation and characteristics of ions and charged aerosol particles in a native Australian Eucalypt forest, *Atmos. Chem. Phys.*, 8, 129–139, doi:10.5194/acp-8-129-2008, 2008.

Tammet, H.: Continuous scanning of the mobility and size distribution of charged clusters and nanometer particles in atmospheric air and the Balanced Scanning Mobility Analyzer BSMA, *Atmos. Res.*, 82, 523–535, 2006.

Twomey, S.: Aerosols, clouds and radiation, *Atmos. Environ.*, 25A, 2435–2442, 1991.

Ude, S. and Fernández de la Mora, J.: Molecular monodisperse mobility and mass standards from electrosprays of tetra-alkyl ammonium halides, *J. Aerosol Sci.*, 36, 1224–1237, 2005.

Vana, M., Tamm, E., Hörrak, U., Mirme, A., Tammet, H., Laakso, L., Aalto, P., and Kulmala, M.: Charging state of atmospheric nanoparticles during the nucleation burst events, *Atmos. Res.*, 82, 536–546, 2006.

Vana, M., Ehn, M., Petäjä, T., Vuollekoski, H., Aalto, P., de Leeuw, G., Ceburnis, D., O'Dowd, C. D., and Kulmala, M.: Characteristic features of air ions at Mace Head on the west coast of Ireland, *Atmos. Res.*, 90, 278–286, 2008

Vartiainen, E., Kulmala, M., Ehn, M., Hirsikko, A., Junninen, H., Petäjä, T., Sogacheva, L., Kuokka, S., Hillamo, R., Skorokhod, A., Belikov, I., Elansky, N., and Kerminen, V.-M.: Ion and particle number concentrations and size distributions along the Trans-Siberian railroad, *Boreal Environ. Res.*, 12, 375–396, 2007.

Venzac, H., Sellegri, K., and Laj, P.: Nucleation events detected at the high altitude site of the Puy de Dôme Research Station, France, *Boreal Environ. Res.*, 12, 397–408, 2007.

Venzac, H., Sellegri, K., Laj, P., Villani, P., Bonasoni, P., Marinoni, A., Cristofanelli, P., Calzolari, F., Fuzzi, S., Decesari, S., Facchini, M.-C., Vuillermoz, E., and Verza, G. P.: High frequency new particle formation in the Himalayas, *P. Natl. Acad. Sci. USA.*, 105(41), 15666–15671, 2008.

Wang, S. C. and Flagan, R. C.: Scanning electrical mobility spectrometer, *Aerosol Sci. Tech.*, 13, 230–240, 1990.

Wiedensohler, A.: An approximation of the bipolar charge distribution for particles in the sub-micron size range, *J. Aerosol Sci.*, 19, 387–389, 1988.

Winkler, P. M., Steiner, G., Virtala, A., Vehkamäki, H., Noppel, M., Lehtinen, K. E. J.,

Reischl, G. P., Wagner, P. E., and Kulmala, M.: Heterogeneous Nucleation Experiments bridging the scale from molecular ion clusters to nanoparticles, *Science*, 319, 1374, doi:10.1126/science.1149034, 2008.

- 5 Yli-Juuti, T., Riipinen, I., Aalto, P. P., Nieminen, T., Maenhaut, W., Janssens, I. A., Claeys, M., Salma, I., Ocskay, R., Hoffer, A., Imre, K., and Kulmala, M.: Characteristics of new particle formation events and cluster ions at K-puszta, Hungary, *Boreal Environ. Res.*, 14, 683–698, 2009.

AMTD

4, 1139–1180, 2011

**Intercomparison of
air ion spectrometers**

S. Gagné et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

I◀

▶I

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



Intercomparison of air ion spectrometers

S. Gagné et al.

Table 1. Tasks performed for each ion spectrometer. The signs + and – in the task description represent positive and negative polarities, “–” means not done or not possible, and the “X” sign means that the task was performed. In the first two rows (standards and mobilities) the number of mobility standards used is shown on the left, and the number on the right expresses the total number of different mobility and concentration combinations.

	AIS1	AIS2	AIS3	AIS6	AIS7	NAIS1	NAIS2	NAIS3	NAIS4	NAIS5	A-NAIS
+ standards (nb. mob./conc.)	4/12	4/12	4/15	4/14	4/23	4/13	–/–	4/15	4/15	4/15	4/10
– mobilities (nb. mob./conc.)	–/–	4/10	4/8	4/10	4/9	4/8	4/10	4/9	4/8	4/10	4/8
+ HDMA silver	–	X	X	–	X	–	–	X	X	X	X
– HDMA silver	–	X	X	–	X	–	–	X	X	X	X
+ HAUKE silver	X	X	X	X	X	X	–	X	X	X	X
– HAUKE silver	X	X	X	X	X	X	X	X	X	X	X
Neutral HAUKE silver	–	–	–	–	–	X	X	X	X	X	X

[Title Page](#)[Abstract](#)[Introduction](#)[Conclusions](#)[References](#)[Tables](#)[Figures](#)[I◀](#)[▶I](#)[◀](#)[▶](#)[Back](#)[Close](#)[Full Screen / Esc](#)[Printer-friendly Version](#)[Interactive Discussion](#)

Intercomparison of
air ion spectrometers

S. Gagné et al.

Table 2. Analysis of the 2nd new particle formation event on 12 June 2009 (18:50) by different instruments and ions spectrometers (see Fig. 6 for details). The mean and standard deviation for the AISs and NAISs are shown in the last two rows.

Instrument	J_2 ($\text{cm}^{-3} \text{s}^{-1}$)			GR (nm h^{-1})					
	negative	positive	particle	2–3 nm (-/+)		3–7 nm (-/+)		7–20 nm (-/+)	
AIS 1	–	–	–	–	–	–	–	–	–
AIS 2	0.13	–	–	12.3	–	24.5	–	34.9	–
AIS 3	0.18	–	–	12.8	–	22.6	–	33.8	–
AIS 6	–	–	–	–	–	–	–	–	–
AIS 7	0.15	–	–	16.3	–	33.8	–	37.7	–
NAIS 1	0.21	0.08	2.2 (weak)	15.2	13.3	21.4	23.4	39.6	39.3
NAIS 2	0.23	–	–	9.6	–	27.5	–	37.1	–
NAIS 3	0.20	–	3.9	11.2	–	27.1	–	34.6	–
NAIS 4	0.21	0.07	1.6 (weak)	13.7	10.8	22.7	19.1	37.3	36.2
NAIS 5	0.24	0.09	1.3 (weak)	15.8	12.2	22.9	25.1	34.9	38.9
A-NAIS	0.20	0.08	4.7	16.2	13.2	25.3	22.0	38.1	–
DMPS	–	–	1.1	–	–	–	–	37 (10–20 nm)	
BSMA	0.27	–	–	11.4	–	23.0	–	–	–
AIS (mean and st. dev.)	0.15±0.03	–	–	13.8±2.2	–	27.0±7.0	–	35.5±2.0	
NAIS (mean and st. dev.)	0.22±0.02	0.08±0.01	2.7±1.5	13.6±2.7	12.4±1.2	24.5±2.5	22.4±2.5	38.1±1.7	

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

◀

▶

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



Intercomparison of
air ion spectrometers

S. Gagné et al.




Fig. 1. Examples of mobility standards measurements. Subfigures **(a)**, **(b)** and **(c)** show the response of the ion spectrometers (normalized mobility distribution) for 3 positively charged mobility standards, and subfigures **(d)**, **(e)** and **(f)** for 3 negative mobilities. The black vertical lines represent the mobility that was selected through the HDMA. The mobility and size of the standards are above each plot. TMAI is tetra-methyl ammonium iodide and THAB is tetra-heptyl ammonium bromide.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

◀

▶

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



Intercomparison of
air ion spectrometers

S. Gagné et al.




Fig. 2. Comparison of the ion spectrometers to reference instruments. **(a), (b), (c)**: negative polarity; **(d), (e), (f)**: positive polarity; **(g), (h), (i)**: total particles (charged + neutral, NAISs only). Subfigures **(a), (d)** and **(e)** display the mobility measured by the ion spectrometers as a function of the mobility selected by the DMA. Subfigures **(b), (e)** and **(h)** display the total concentration measured by the ion spectrometers as a function of the concentration measured with reference instruments. Subfigures **(c), (f)** and **(i)** display the ratio of the ion spectrometer concentration to the concentration of the reference instrument as a function of the particle mobility. In subfigures **(a)** to **(f)**, the AISs are represented by magenta lines and red filling (standard deviation between instruments), the NAISs by cyan lines and blue filling, and the ANAIS by a green line. The electrometer is the reference instrument in these two lines. In subfigures **(g), (h)** and **(i)**, the total particles, cyan lines and blue fillings correspond to the negative DMAs of the ion spectrometers; magenta lines and red fillings correspond to the positive DMAs of the ion spectrometers. The green and red lines represent the negative and positive DMAs of the ANAIS, respectively. The reference instrument is the CPC in this line. The black dashed lines are the ideal values.

Intercomparison of
air ion spectrometers

S. Gagné et al.




Fig. 3. The background concentrations are presented as a function of particle diameter. The total background concentration in ion measurement mode (a and c) was generally below 30 cm^{-3} for AISs and in the $30\text{--}50 \text{ cm}^{-3}$ concentration range for NAISs. In particle measurement mode (b and d), the median total background concentration was $1.6 \times 10^5 \text{ cm}^{-3}$ and the median total background concentration above 3.4 nm was 86.5 cm^{-3} .

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

◀

▶

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



Intercomparison of
air ion spectrometers

S. Gagné et al.




Fig. 4. Median size distribution and electrometer current distribution for indoor air (negative: (a), (b); positive: (c), (d)), measured on 6 June 2009 and outdoor air (negative: (e), (f); positive: (g), (h)), measured on 7 June 2009. The AIS7 re-inverted concentrations are represented by the dashed line of the same color.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

◀

▶

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



Intercomparison of
air ion spectrometers

S. Gagné et al.




Fig. 5. Performance during the intercomparison period (7 June–25 June 2009). **(a)** Concentration of ions for AISs, NAISs and the BSMA (size range: 3–7 nm). **(b)** Concentration of particles for NAISs and the DMPS (size range: 10–40 nm). **(c)** Mean diameter for AISs, NAISs and the BSMA (0.8–3 nm). The red lines are the median value, the blue boxes are the 25th and 75th percentile, whiskers are 10th and 90th percentiles and red crosses are outliers. For readability purposes, the AISs are code “A” and the NAISs, “N”. The ANAIS is coded AN. **(d)** The total ion concentration of the ion spectrometers in the 3–7 nm size range is compared to the BSMA concentration. The AISs are represented by black circles and the NAISs by blue diamonds.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

◀

▶

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



Intercomparison of
air ion spectrometers

S. Gagné et al.




Fig. 6. Size distribution during a NPF event (12 June 2009) for a range of instruments. **(a)** DMPS, total concentration of aerosol particles, **(b)** AIS, positive air ions, **(c)** NAIS concentration of particles in particle operation mode (from the negative analyzer), **(d)** AIS negative air ions, **(e)** NAIS negative air ions, **(f)** BSMA negative air ions.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

◀

▶

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



Intercomparison of
air ion spectrometers

S. Gagné et al.




Fig. 7. Negative (a to f) and positive (g to l) charged fraction (fraction of charged particles in percents) at 6 different diameters as a function of time for all NAISs on 12 June 2009. The charged fraction retrieved with the Ion-DMPS is shown as black circles and the bipolar steady-state charged fraction is marked with a black dashed line.

Copyright of Atmospheric Measurement Techniques Discussions is the property of Copernicus Gesellschaft mbH and its content may not be copied or emailed to multiple sites or posted to a listserv without the copyright holder's express written permission. However, users may print, download, or email articles for individual use.