

Investigating the Use of Wearables for Monitoring Circadian Rhythms in Chemotherapy: Preliminary Results from a Pilot Study

R. Castaldo^{1,2}, M. Prati³, L. Montesinos¹, V. Kulkarni¹, M. Chappell¹, H. Byrne⁴, P. Innominato⁵, S. Hughes⁶ and L. Pecchia¹

¹ School of Engineering, University of Warwick, Coventry, UK

² Institute of Advanced Studies, University of Warwick, Coventry, UK

³ Ingegneria dell'informazione, informatica e statistica, Università degli Studi di Roma La Sapienza, Rome, Italy

⁴ Mathematical Institute, University of Oxford, Oxford, UK

⁵ Warwick Medical School, University of Warwick, Coventry, UK

⁶ North Wales Clinical Research Centre, Wrexham, UK

Abstract— Due to circadian rhythms, the efficacy and side effects of chemotherapy change throughout the day. Chemotherapy in turn alters circadian rhythm. This creates a closed-loop requiring control. While circadian rhythms can be monitored using blood, salivary and urine hormone tests, such tests are not practical at home and do not provide continuous real-time monitoring. Combining signal processing and artificial intelligence and signal processing with commercial sensors embedded in smartwatches or clothes that measure physiological and behavioural attributes offers unprecedented and as yet unexplored opportunities to monitor circadian rhythms in real time. This paper presents the initial steps towards the development of a model for real-time monitoring of the circadian rhythms. This model will allow to revolutionize the administration of chemotherapy, improve patient responses, and reduce side effects and costs. Preliminary analysis shows promising results to automatically classify cortisol levels as high or low, based on behavioral and physiological signals monitored by non-invasive wearable sensors.

Keywords— Circadian rhythms, biomedical signals, wearable sensors, machine learning, chemotherapy.

I. INTRODUCTION

“Inner clock adapts our physiology to the dramatically different phases of the day, [...] regulating critical functions such as behavior, hormone levels, sleep, body temperature and metabolism”. This phenomenon is known as the biological clock or circadian rhythm [1].

Cancer therapy has significant side effects. Among other, it increases mental stress, altering internal clock (i.e., circadian rhythm) and sleep cycles, further disrupting quality of life. Recent researches [2] proved that patient with major alterations in circadian cycles are significantly less likely to survive to cancer treatments.

Currently, there are no instrumental methods to measure accurately how cancer therapy alters circadian cycles in individual patients over time. Consequently, it is not possible to optimize therapy or design new interventions to minimize this effect.

In the future, cancer treatments will be dynamically personalized considering patient characteristics and their individual response during therapeutic administration, based on closed-loop control. Unfortunately, we are far from this vision, due to lack of knowledge, the absence of reliable models and deficiency of cost-effective tools.

One reason is that methods for circadian measurements are not suitable for continuous and simultaneous monitoring at home. In fact, circadian cycles are measured via laboratory tests (i.e., hormones measured via blood, urine or saliva specimens), which are expensive and not easy to be performed at home. Most recently, actigraphy has been explored for circadian rhythm estimations at home [3]. Nonetheless, benchmark methods are not yet available and non-invasive behavioral (i.e., actigraphy) and physiological monitoring has not been combined yet, especially in oncology.

Therefore, there is the need to combine wearable sensors, biomedical signal analysis and machine learning to develop methods and tools to quantify alterations in internal clock, helping oncologists to optimize cancer therapy.

Several cortisol indices are commonly used in the literature to determine circadian alteration such as amplitude, frequency and phase [4]. In particular, peak-to-trough difference is one of the most used index to assess rhythm alteration in cancer patients [5].

This paper presents a preliminary result from a pilot study conducted on healthy subjects to identify a model to monitor circadian rhythms (peaks and trough) in real-time using artificial intelligence and unobtrusive wearable behavioral and physiological monitors.

II. METHODS AND MATERIALS

A. Study Participants

8 healthy participants (4 men and 4 women, mean age (SD): 26.2 (3.3) years) in whom no abnormalities were detected by the medical history, were recruited in the study. Baseline characteristics, such as age, height, weight, general

health status and use of medications, were collected during a baseline assessment and briefing session. The participants did not report history of heart disease, diabetes, systemic hypertension or hypotension, or sleep-disorders, or consumption of any medication throughout the course of the study, which could alter physiological signals being acquired. They had healthy body mass index (BMI), i.e. between 18.5 and 24.9.

The Biomedical and Scientific Research Ethics Committee of the University of Warwick approved this study (ref. REGO-2018-2205), assuring anonymity and no side effects or possible disadvantages for the participants. All participants were carefully instructed, and informed consent was acquired prior to the experiment. Participants were compensated with a fixed fee.

B. Protocol

Participants were asked to wear two different wearable devices for three nights and two consecutive days. The first wearable device, the Zephyr BioPatch, recorded ECG, breathing rate and raw 3-axis accelerations, with a sampling rate of 250 Hz, 18 Hz and 100Hz respectively. The second one was a wireless data logger, the iButtons, which can be used to obtain a valid measurement of human skin temperature. One iButton was attached to each ankle, and another iButton was attached on each side of the chest, one or two inches below the clavicle in the mid-clavicular line in order to measure distal and proximal body temperature respectively [6]. The temperature sensors took a measurement every 10 minutes.

In this study, cortisol was used as a marker of circadian rhythm. Participants were instructed on how to take and store saliva samples, so as they could be sent to a specialized laboratory and analyzed for levels of salivary cortisol. Participants were instructed to take a sample immediately upon waking, and then to take further samples every two hours for the rest of the day until they went to bed. Saliva samples were acquired for two consecutive days via Salimetrics® Cortisol Enzyme Immunoassay Kit, which is an immunoassay specifically designed and validated for the quantitative measurement of salivary cortisol. The saliva was collected by the passive drool technique.

For each subject, behavioral and physiological signals were acquired for three nights and two days by wearable devices, and two days' worth of salivary samples were taken.

Participants were also asked to report physical activity and food intake [7]. Participants were also asked to complete the Pittsburgh Sleep Quality Index (PSQI) instrument [8] and a consensus sleep diary [9]. The PSQI and sleep diary results were used to compare reported sleep disturbances with alteration in circadian cycles.

All of the participants were asked to maintain ordinary daily schedules during the experiments.

C. Data Analysis

The maximum and minimum cortisol levels were obtained for each subject for the two days, and these were labelled respectively as "peak" and "trough". In the cases where the minimum or maximum cortisol level for a period appeared in more than one measurement, then each measurement was also labelled.

Since physical exercise can greatly affect cortisol levels [10], only periods of time during which there were similar levels of activity were considered. For each peak and trough, a window of two hours around the time of the saliva measurement was taken. Within each window of time, activity level and posture were evaluated. The activity as reported from the Zephyr BioPatch represents a measure of second-to-second activity and is sensitive to small movements. In order to reduce this sensitivity, the signal was smoothed using a moving average, with a sample window of 60. This smoothed activity, in essence, represents minute-to-minute activity.

In order to control for activity and posture in the two hours window around each peak or trough, only times when activity was less than 0.2 (which corresponds to a level of activity less intense than walking) and the posture was between -20° and 20° (which corresponds to times when the chest was roughly upright) were considered.

For each selected window of time, distal and proximal body temperature were also considered. Distal body temperature was calculated as the mean of all measurements from ankle temperature sensors during the selected window of time. Proximal body temperature was calculated in the same way as distal body temperature.

For each selected window of time, the RR interval time-series was extracted from ECG records using an automatic QRS detector, WQRS, available in the PhysioNet's toolkit [11]. QRS review and correction was performed using PhysioNet's WAVE. The fraction of total RR intervals labelled as normal-to-normal (NN) intervals was computed as NN/RR ratio. NN/RR ratio was then used to measure the reliability of the data. Records with NN/RR ratio less than 90% threshold were excluded from the analysis. HRV analysis was performed on 5 min excerpts using Kubios (version premium) [12]. Time and frequency-domain features were analyzed according to international guidelines [13], while non-linear measures were analyzed as described in [14]. Frequency domain features were extracted from power spectrum estimated with autoregressive (AR) model methods [12]. Finally, 20 HRV features were examined.

D. Statistical Analysis and Classification

Given that HRV features were found non-normally distributed, Median (MD), Median Absolute Deviation (MAD) and interquartile range (IQR) (i.e., non-parametric

descriptors) were computed for each repetition. The non-parametric Wilcoxon Signed-Rank Test was used to appraise statistical differences of HRV features and temperature variation between the “peak” and “trough” of cortisol measures.

In order to optimize the performance of the machine learning models, the number of features should be limited by the number of instances of the event to detect (in this instance, a peak or trough in cortisol levels). Furthermore, a reduction in the number of features greatly simplifies the medical interpretation of any results achieved. Therefore, the features selection was performed using relevance and redundancy analysis as described in [15]. Training of the machine-learning models (including the algorithm parameter tuning) was performed using a leave-one-out-cross-validation approach on 6 participants. Binary classification performance measures were adopted according to the standards reported in [14]. Five different machine-learning methods were used to train, validate and test the classifiers (SVM, MLP, IBK, RF and LDA); the model was chosen as the classifier achieving the highest Area under the Curve (AUC), which is a reliable estimator of both sensitivity and specificity rates. The model was then tested on the remaining 2 participants.

III. RESULTS AND CONCLUSION

Preliminary analysis shows promising results to automatically detect cortisol levels as high or low (peaks or troughs), based on HRV data extracted during periods where activity and posture are controlled for. Some moderately successful classifiers were produced. Random Forest outperformed the other classifiers achieving 78% AUC and 73% overall accuracy. These results provide encouragement that such a protocol may be successful with further refinement, and wearable devices (through the measurement of HRV) may indeed be useful in the real-time monitoring of circadian rhythm.

ACKNOWLEDGMENT

We acknowledge the EPSRC-funded grant under the Cyclops Healthcare Network. RC thanks the Institute of Advanced studies at the University of Warwick.

CONFLICT OF INTEREST

The authors declare that they have no conflict of interest.

REFERENCES

- [1] L. W. Van Laake, T. F. Lüscher, and M. E. Young, "The circadian clock in cardiovascular regulation and disease: Lessons from the Nobel Prize in Physiology or Medicine 2017," *European heart journal*, 2017.
- [2] F. Lévi, A. Okyar, S. Dulong, P. F. Innominato, and J. Clairambault, "Circadian timing in cancer treatments," *Annual review of pharmacology and toxicology*, vol. 50, pp. 377-421, 2010.
- [3] M. T. Smith *et al.*, "Use of actigraphy for the evaluation of sleep disorders and circadian rhythm sleep-wake disorders: an American Academy of Sleep Medicine clinical practice guideline," *Journal of Clinical Sleep Medicine*, vol. 14, no. 07, pp. 1231-1237, 2018.
- [4] L. D. Dorn, J. F. Lucke, T. L. Loucks, and S. L. Berga, "Salivary cortisol reflects serum cortisol: analysis of circadian profiles," *Annals of clinical biochemistry*, vol. 44, no. 3, pp. 281-284, 2007.
- [5] M. Mormont and F. Levi, "Circadian-system alterations during cancer processes: A review," *International Journal of Cancer*, vol. 70, no. 2, pp. 241-247, 1997.
- [6] M. J. Hasselberg, J. McMahon, and K. Parker, "The validity, reliability, and utility of the iButton® for measurement of body temperature circadian rhythms in sleep/wake research," *Sleep medicine*, vol. 14, no. 1, pp. 5-11, 2013.
- [7] M. A. A. de Assis, E. Kupek, M. V. c. Nahas, and F. Bellisle, "Food intake and circadian rhythms in shift workers with a high workload," *Appetite*, vol. 40, no. 2, pp. 175-183, 2003.
- [8] D. J. Buysse, C. F. Reynolds III, T. H. Monk, C. C. Hoch, A. L. Yeager, and D. J. Kupfer, "Quantification of subjective sleep quality in healthy elderly men and women using the Pittsburgh Sleep Quality Index (PSQI)," *Sleep*, vol. 14, no. 4, pp. 331-338, 1991.
- [9] C. E. Carney *et al.*, "The consensus sleep diary: standardizing prospective sleep self-monitoring," *Sleep*, vol. 35, no. 2, pp. 287-302, 2012.
- [10] H. Budde, S. Machado, P. Ribeiro, and M. Wegner, "The cortisol response to exercise in young adults," *Frontiers in behavioral neuroscience*, vol. 9, p. 13, 2015.
- [11] A. L. Goldberger *et al.*, "Physiobank, physiotoolkit, and physionet components of a new research resource for complex physiologic signals," *Circulation*, vol. 101, no. 23, pp. e215-e220, 2000.
- [12] M. P. Tarvainen and J.-P. Niskanen, "Kubios HRV User's Guide," *Biosignal Analysis and Medical Imaging Group (BSAMIG), Department of Physics University of Kuopio*, 2013.
- [13] T. Force, "Heart rate variability guidelines: Standards of measurement, physiological interpretation, and clinical use," *European Heart Journal*, vol. 17, pp. 354-381, 1996.
- [14] P. Melillo, M. Bracale, and L. Pecchia, "Nonlinear Heart Rate Variability features for real-life stress detection. Case study: students under stress due to university examination," (in English), *BioMedical Engineering OnLine*, vol. 10, no. 1, pp. 1-13, 2011/11/07 2011, Art. no. 96.
- [15] R. Castaldo, P. Melillo, R. Izzo, N. D. Luca, and L. Pecchia, "Fall Prediction in Hypertensive Patients via Short-Term HRV Analysis," *IEEE Journal of Biomedical and Health Informatics*, vol. 21, no. 2, pp. 399-406, 2017.

Author: Rossana Castaldo
 Institute: University of Warwick
 Street: Library Road
 City: Coventry
 Country: UK
 Email: r.castaldo.1@warwick.ac.uk