



Published in final edited form as:

J Quant Spectrosc Radiat Transf. 2016 May ; 175: 90–99. doi:10.1016/j.jqsrt.2016.02.002.

Line intensities and temperature-dependent line broadening coefficients of Q-branch transitions in the ν_2 band of ammonia near 10.4 μm

Ritabrata Sur, R. Mitchell Spearrin, Wen Y. Peng, Christopher L. Strand, Jay B. Jeffries, Gregory M. Enns, and Ronald K. Hanson

High Temperature Gasdynamics Laboratory, Thermosciences division, Stanford University, 452 Escondido Mall, Bldg 520, CA 94305, USA

Abstract

We report measured line intensities and temperature-dependent broadening coefficients of NH_3 with Ar, N_2 , O_2 , CO_2 , H_2O , and NH_3 for nine $sQ(J,K)$ transitions in the ν_2 fundamental band in the frequency range 961.5–967.5 cm^{-1} . This spectral region was chosen due to the strong NH_3 absorption strength and lack of spectral interference from H_2O and CO_2 for laser-based sensing applications. Spectroscopic parameters were determined by multi-line fitting using Voigt lineshapes of absorption spectra measured with two quantum cascade lasers in thermodynamically-controlled optical cells. The temperature dependence of broadening was measured over a range of temperatures between 300 and 600 K. These measurements aid the development of mid-infrared NH_3 sensors for a broad range of gas mixtures and at elevated temperatures.

Keywords

Ammonia; Spectroscopy; Mid-infrared; Quantum cascade laser; Line intensity; Broadening

1. Introduction

Spectroscopic sensors for gaseous ammonia are important for a number of applications, ranging from environmental monitoring [1,2], industrial process control [2], clinical analysis of human breath [3–6], and the recycling of bio-waste water [7]. Trace detection at ppb to ppm levels is often required [7–10]. Laser absorption techniques have been implemented in various forms for trace NH_3 detection, e.g., rapid-scan direct absorption [11,12], wavelength-modulation spectroscopy (WMS) [4,13], frequency-modulation spectroscopy (FMS) [14], cavity ring-down spectroscopy [6,15], and photoacoustic spectroscopy (PAS) [3,9,16], all using different types of near-infrared or mid-infrared distributed-feedback (DFB) lasers [9,11,13,14,17], vertical cavity surface-emitting lasers (VCSEL) [12], or quantum cascade lasers (QCL) [4,6,15]. The appropriate wavelength for spectroscopic

detection generally depends on relative spectral strength and isolation from other components in the gas mixture.

Fig. 1(a) shows the simulated rovibrational absorption spectrum of 10 ppm NH₃ in air at 308 K (typical human breath temperature, 35 °C), 1 atm for an optical path length of 1 m. Ammonia (NH₃) has six vibrational modes – symmetric stretch (ν_1), symmetric bend (ν_2), doubly-degenerate asymmetric stretch (ν_3), and doubly-degenerate asymmetric bend (ν_4). Moreover, ammonia is a symmetric top molecule with two rotational quantum numbers, J and K , where J represents the total angular momentum and K represents the angular momentum about the primary axis of symmetry. The near-infrared wavelength range of 1–2.5 μm , composed of combination bands and overtones, is the most widely used spectral region for detection of ammonia, primarily for atmospheric sensing and combustion slip measurements [11,14]. However, as can be seen from Fig. 1(a), the NH₃ ν_2 fundamental absorption band is the strongest of all the IR bands with nearly 50–100X larger absorption strength than the NIR bands, and provides a possibility for more sensitive detection. The recent availability of compact and coherent light sources in the mid-infrared has led to the development of highly sensitive and low-interference diagnostics for ammonia utilizing NH₃ transitions near 965.35 cm^{-1} [1], 967.35 cm^{-1} [5,6,18,19], 1046.4 cm^{-1} [3,16] and 1103.4 cm^{-1} [4,20], with applications in the atmosphere and exhaled human breath. For laser absorption measurements to be accurate and calibration-free, a comprehensive characterization of the fundamental spectroscopy is required in the wavelength domain of interest.

Several studies on ammonia spectroscopy have been reported in the literature. The HITRAN 2000 database lists several NH₃ lines as reported by Kleiner et al. [21]. In the ν_2 rovibrational band of ammonia, line intensities and broadening parameters due to various different collision partners including NH₃, N₂, O₂, H₂, CO₂ and H₂O were measured with FTIR spectrometers [22–26] and QCL-based spectroscopic techniques [27]. All these studies are restricted to the sP(J,K) and sR(J,K) manifolds. In addition, line intensities and broadening parameters due to multiple collision partners (e.g. self, N₂, O₂, CO₂, H₂, Ar and air) have been measured and calculated in the ν_4 [28–30] and ν_1 [31–33] NH₃ rovibrational bands. The effect of water vapor broadening on ammonia transitions near 1.51 μm has also been studied by Schilt [34]. However, the sQ(J,K) manifold of the ν_2 band including the lines near 962.17 cm^{-1} , 965.35 cm^{-1} and 967.35 cm^{-1} are not well-characterized despite the common selection of these wavelengths for gas sensing [1,5,6,18,19] and a high demand for mid-infrared experimental NH₃ data noted in the literature [35,36]. This motivates the current high-resolution spectroscopic study of the Q-branch of the ν_2 fundamental rovibrational band of NH₃.

As shown by the shaded box in Fig. 1(b), the Q-branch lies in a region with minimal interference from H₂O and CO₂, the two common interfering species in atmospheric, combustion and clinical breath-sensing applications. Also, as mentioned in the HITRAN 2012 article [35], the spectral parameters were found to deviate very significantly from the values calculated using measured P-branch parameters to predict values in the R manifold. The sQ(J,K) manifold is particularly difficult to study with FTIR spectrometers because of the close proximity of the lines with associated blending effects. The current study uses two

narrow-linewidth quantum cascade lasers and an approach similar to Owen et al. [27] at low pressure to obtain accurate values of line intensities and broadening parameters in the $sQ(J,K)$ manifold.

1.1. Theory

Although the fundamental basis of absorption spectroscopy is well known, a short review provides definition of symbols and units used here. The transmission coefficient (τ) of a monochromatic light beam at frequency ν is governed by the Beer–Lambert relation:

$$\tau(\nu) = \frac{I}{I_0} = e^{-\alpha(\nu)} \quad (1)$$

where I_0 is the incident beam intensity, I is the transmitted beam intensity and α is the spectral absorbance for a pressure P , path length L , mole fraction of the absorbing species χ , transition linestrength S and lineshape function $\phi(\nu)$, as defined by the expression

$$\alpha\nu = PL\chi S\phi(\nu) \quad (2)$$

The above expression assumes uniform gas composition and temperature along the laser line of sight. For a single isolated transition, the total area under the absorbance curve is given by:

$$A = \int_{-\infty}^{\infty} \alpha(\nu) d\nu = SP\chi L$$

The lineshape function $\phi(\nu)$ is approximated by a Voigt function ($\int_{-\infty}^{\infty} \phi(\nu) d\nu = 1$) characterized by the collision-broadened full-width at half-maximum (FWHM), $\Delta\nu_C$ [cm^{-1}] and the Doppler FWHM, $\Delta\nu_D$ [cm^{-1}]. $\Delta\nu_C$ relates to the transition-specific broadening coefficients (γ_{NH_3-Y}) of the absorbing species i.e. NH_3 in this case, unique to each collision-partner (Y) as

$$\Delta\nu_C = P \sum_Y \chi_Y 2\gamma_{\text{NH}_3-Y} \quad (3)$$

The dependence of the collision-broadening coefficient on temperature can be modeled as a power law expression:

$$\gamma_{(\text{NH}_3-Y)} = \gamma_{0,(\text{NH}_3-Y)} \left(\frac{T_0}{T} \right)^{n_{(\text{NH}_3-Y)}} \quad (4)$$

where $\gamma_{0,(\text{NH}_3-Y)}$ is the broadening coefficient at the reference temperature, T_0 (=296 K) and $n_{(\text{NH}_3-Y)}$ is the temperature-dependence exponent of NH_3 with the collision partner Y .

In this article, we report the measured values of $S(296\text{ K})$, $\gamma_{0,(NH_3-Y)}$ and $n_{(NH_3-Y)}$ for nine transitions in the $sQ(J,K)$ manifold of the ν_2 rovibrational band of NH_3 .

2. Experimental details

The schematic shown in Fig. 2 illustrates the experimental apparatus used for these experiments. It consisted of a gas cell with two wedged BaF_2 windows, inlet gas manifold for mixing the gases, outlet line to the vacuum pump to facilitate a flowing gas configuration, a distributed feedback QCL (swapped with a second QCL for a different wavelength range) and a detector. To enable accurate measurements of the spectral parameters, use of two slightly different setups with different path lengths was optimal as discussed below:

i. Line intensity measurements—The experiments for determination of line intensities of NH_3 necessitate the analysis of laser absorption through a gas sample with a known concentration of NH_3 . This was found to be difficult due to the following reasons:

1. *Sticky nature of NH_3 .* Ammonia mixture composition in the cell was found to be inconsistent with the certified bottle concentration due to the surface adsorption of ammonia to the walls. Even with a flow arrangement, the steady-state concentration did not accurately reflect the bottle composition and hence could not be used for line strength measurements.
2. *Possible stratification of NH_3 in the mixture.* Resultant absorbance was found to be dependent on the flow rate suggesting possible stratification in the mixture composition further magnified due to the adsorption effect.

Therefore, line intensity measurements were performed at 296 K in a short (3.36 mm) cell with flowing 99.99% pure anhydrous NH_3 (Praxair) at sub-atmospheric pressure (<35 Torr). Due to the highly toxic nature of pure NH_3 , these studies were performed inside a fume hood.

ii. Broadening parameter measurements—Measurements of broadening parameters were performed in a heated optical cell of 101 cm path. The cell temperature non-uniformity at the hottest setting of 600 K was found to be less than ± 4 K, and progressively less at lower temperatures. In this setup, a differential flow configuration of 1.3% NH_3 in Ar (Praxair certified gas mixture) and the pure bath gas (denoted generically as Y) ensured a highly dilute mixture of NH_3 with about 96% bath gas Y and 4% Ar. This enabled the creation of desirable final mixture composition with about 4:100 ratio of flow rates, easily accomplished by the flow controllers FC1 and FC2. The sample pressure was controlled using micrometer valves V1 and V2 and measured using MKS 627D capacitance manometers with 100 and 1000 Torr full-scale pressure ranges (uncertainty: 0.12% of reading). Experiments were performed over typical pressure ranges of 0–30 Torr and 0–300 Torr for line intensity and collision broadening parameter measurements, respectively. The dilution was taken into account while processing the broadening data. The flow configuration ensured a constant ammonia concentration in the cell despite the significant

adsorption of NH_3 on the walls of the optical cell and the connected Teflon/stainless steel tubing.

For water broadening measurements, the gas Y line was replaced by a degassed water flask in a water bath (which maintained the flask temperature to counter the heat loss due to evaporation). At 296 K the vapor pressure of water is 21 Torr, leading to a slow and ambiguous water flow rate. In response, the flow rate of the NH_3 – Ar mixture had to be kept to a very low value using a micrometer valve, and the water concentration independently verified. In this configuration, the mixture composition was inferred by simultaneously measuring the mixture water concentration with a near IR TDLAS water sensor utilizing a water transition centered at 7185.6 cm^{-1} . The mixture water mole fraction was found to be about 70% for most cases, and the rest was assumed to be Ar and NH_3 for broadening calculation purposes.

The current study of highly resolved spectroscopy of NH_3 in the Q-branch was made possible by the availability of narrow linewidth (typically <10 MHz) quantum cascade lasers (QCLs) from Alpes Lasers SA (Neuchâtel, Switzerland; www.alpeslasers.ch/) and Adtech Optics (www.atoptics.com/) at the specified wavelength range of interest. Both QCLs have distributed feedback (DFB) architecture and were collimated in high-heat-load (HHL) packages and mounted on water-cooled aluminum plates for effective temperature control. The water temperature was maintained at approximately $18 \text{ }^\circ\text{C}$ by a Thermocube water chiller. The two QCLs (Alpes QCL and Adtech QCL) were tunable over the ranges 960.8 – 965.8 cm^{-1} and 963.4 – 968.9 cm^{-1} , respectively, by tuning the laser temperature and injection current. The temperature was regulated using Alpes Lasers TC-3 and Arroyo 5305 temperature controllers, while the current was controlled using a ILX Lightwave LDX 3232 high compliance laser diode driver and an Arroyo 4300 laser source. The laser was scanned over the necessary wavelength range by providing a sawtooth voltage input signal to the current controller by the data acquisition (DAQ) system. The DAQ system consisted of a National Instruments PXIe-6124 card (16-bit analog I/O, up to 2.5 Ms/s) mounted on an external chassis NI PXIe-1073. The output laser beam is collected on a Vigo PVM-2TE-10.6 thermoelectrically cooled multijunction, optically immersed photovoltaic detector. The generated signal voltage from the detector was recorded by the DAQ. The relative laser frequency during the laser scan was determined by recording the scan with a germanium etalon of free spectral range (FSR) of 0.0163 cm^{-1} placed along the path of the laser.

3. Results and discussions

The line intensities and temperature-dependent broadening coefficients of NH_3 with Ar, O_2 , N_2 , CO_2 , H_2O and NH_3 of nine spectral transitions were studied between 961.5 and 967.5 cm^{-1} . These transitions in the Q-branch of the ν_2 fundamental band of NH_3 near $10.4 \text{ }\mu\text{m}$ are designated by numbers 1–9 in Fig. 3 and listed in Table 1. Some of these lines (5 and 7) have previously been used for room temperature sensing [1,5,18,19] as discussed earlier, and line 3 is ideally suited for sensing at elevated temperatures. The remaining lines (1,2,4,6,8,9) can influence the spectral characteristics of the three primary lines (3,5,7) at elevated pressures, and require characterization for quantitative sensing over a broad range of conditions. In contrast to their utility, none of these lines have been studied experimentally

with regards to their fundamental spectroscopic parameters, and previous sensing applications have relied on calibration methods. The HITRAN database provides computed spectroscopic parameters based on models anchored to data in the R and P branches of the ν_2 band [25,36]. The results of this study aim to provide improved certainty in parameters found in the HITRAN database (S , γ_{self} , γ_{air}) and to provide additional parameters for broadening and its temperature dependence for other collision partners (CO_2 , H_2O , Ar) common to many important gas mixtures (e.g. atmosphere, breath, combustion exhaust). An attempt was made to characterize the pressure shift, however, due to the slow drift in the laser output frequency, the absolute frequency of the spectral features could not be resolved with a high degree of confidence and hence these values are not reported here.

3.1. Line intensities

The absorbance was determined from the experimentally obtained I and I_0 (Eq. (1)) after correcting for the detector offset (including emission and dark current noise), which was collected for every scan by tuning the laser current below the lasing threshold. Empirical background measurements enabled the extraction of non-absorbed laser intensity (I_0) with high fidelity. The absorbance profile was then fit with multiple Voigt profiles positioned at the line centers of each transition with the integrated area ($SP\chi L$) and collision broadening FWHM as the free parameters in the measured frequency range (e.g. Fig. 4(a), (b), and (c)) using the Levenberg–Marquardt optimization algorithm. Sample experimental Voigt lineshape fits for the lines 2–9 are shown in Fig. 4(a), (b) and (c). The sample fit for line 1 was not shown in the figure for better clarity of the other fits. In all the reported cases, the errors in the fits were better than 2% of the absorbance at line center. Due to the small residual error of the Voigt fit (typically $\sim 1\%$ max), other lineshape functions were not considered. The current study was conducted below ambient pressure and therefore the absorbance lineshapes at higher pressures must be experimentally validated to check for line-mixing phenomena as observed in the same band [37,38]. Similarly, discrepancies due to collision narrowing effects for the temperature range investigated in this study (< 600 K) were found to be negligible. Collision narrowing effects should be reconsidered at higher temperatures, however, if a greater deviation is observed from the current spectral models.

The integrated area under each lineshape, A can be related to the line intensity, S as $A = SP\chi L$, where χ is the mole fraction of NH_3 . For pure ammonia measurements, $\chi=1$. Corresponding to every pressure, the fitted Voigt lineshape resulted in different values of the integrated area, A . These values are plotted versus pressure along with the best-fitted straight line in Fig. 5. The slopes of the best fitted straight lines after being divided by the absorption path length yielded the line intensities of each transition at the measurement temperature, 296 K. The values of line intensities and comparison with the values from the literature are listed in Table 2. It can be seen that all the reported values of line intensities lie within the uncertainty bounds of those in the HITRAN'12 database ($< 20\%$ for all lines). The fitted Voigt lineshapes also provided the NH_3 self-broadening parameters at 296 K as listed in Table 3.

It can be seen from Fig. 6, that for a particular quantum number, J , the measured line intensities (denoted by markers) of the transitions follow the trends predicted by

HITRAN'12 (dashed lines for a particular K). Further, due to nuclear spin effects of C_{3v} group symmetry ($\sigma=3$), the degeneracy is doubled if the quantum number K is divisible by 3. Though our data set is relatively sparse, this effect of nuclear spin can be observed in Fig. 6, where the line intensities of transitions with K divisible by 3 are stronger due to larger statistical weight [39].

3.2. Collision broadening

Ammonia absorption features consisting of blended Q-branch transitions can be drastically altered by pressure-dependent collisional line broadening. Moreover, the Lorentzian broadening effect has a strong dependence on collision partner (i.e. bath gas composition). Fig. 7 shows absorption measurements for 500 ppm NH_3 near 962.17 cm^{-1} at 150 Torr in the presence of 4% Ar (see experimental details) and different balance gases, Y (Ar, O_2 , N_2 or CO_2). The dramatic change in lineshape motivated the accurate characterization of the collision broadening coefficient, γ_{NH_3-Y} of these lines with various broadening partners common to gas mixtures in sensing applications. The collision broadening coefficient, γ_{A-B} relates to the optical collision diameter, d_{AB} and the reduced mass, μ as $\gamma_{A-B} \sim (d_{AB}^4/\mu)^{1/2}$ [39]. Ar has a lower optical diameter than N_2 and O_2 and is significantly heavier than both, leading to a lower collisional broadening with NH_3 . Both NH_3 and H_2O are light molecules with large optical collision diameters, resulting in large collisional broadening, whereas CO_2 , being a heavier molecule contributes to lesser extent to broadening but is significantly higher than N_2 , O_2 and Ar. These trends can be noted in Fig. 7 and Fig. 8 (a) and also found to be consistent with Owen et al. [27]. The collision broadening coefficient also has a strong dependence on temperature. Very few studies, with the exception of Nemtchinov et al. [25], have been devoted to the study of the temperature dependence of the broadening parameters in this band.

To characterize the broadening coefficients, absorption spectra were collected at several pressures, while switching the bath gas between Ar, O_2 , N_2 , CO_2 and H_2O . The absorption spectra when fit with Voigt lineshapes yield the collision broadening halfwidth (FWHM), $\Delta\nu_c$.

Fig. 8 illustrates the dependence of $\Delta\nu_c$ on pressure for (a) different collision partners at a fixed temperature (435 K) and (b) temperatures for CO_2 as a collision partner for the $sQ(9,9)$ transition. From Eq. (3), it can be stated that the slopes of the fitted straight lines are equal to twice the sum of the mole-fraction-weighted collision broadening coefficients γ_{NH_3-Y} at the corresponding temperatures. The linear fit eliminates the effect of any small offset that the pressure measurements may have. The broadening experiments done with various bath gases contained about 4% additional Ar. Hence, the fits containing additional Ar dilution was corrected for the contribution of γ_{NH_3-Ar} . Moreover, it was ensured that any error propagated towards γ_{NH_3-Y} due to the uncertainty in γ_{NH_3-Ar} was kept to a minimum by creating a highly diluted sample composed primarily of the bath gas (~96%).

The broadening coefficient obtained for each collision partner Y was then obtained for a range of temperatures. The broadening coefficients obeyed the usual power law form (Eq. (4)). Therefore the plot of $\ln(\gamma_{(NH_3-Y)})$ vs $\ln(T)$ had a linear slope that provided the value of $n_{(NH_3-Y)}$. Fig. 9 shows a plot of the broadening coefficient versus temperature for the

sQ(9,9) transition, confirming the power-law behavior (linear in ln-ln scale) of these fits. In addition, it was found that the broadening coefficients and their temperature dependence exponents vary significantly with collision partner.

The values of the measured broadening coefficients (HWHM/atm) at 296 K (γ_0) and the temperature-dependence exponent (n) of broadening are tabulated in Table 3. The parameters of transitions 1–4 were studied in the temperature range of 296–600 K, while the rest were studied from 296–425 K. The lines 1–4 were examined more thoroughly at higher temperatures, because of their greater applicability in high temperature environments, e.g. in combustion exhaust, whereas the other lines generally find greater relevance in lower temperature applications such as atmospheric monitoring and breath sensing. The water broadening temperature-dependence exponents were not measured for lines 5–9 for the same reason. The temperature exponent of NH_3 -self broadening could not be measured because the 3.36 mm cell was an unheated optical cell.

The measured values of the broadening parameters and linestrengths generally agreed within 10% of those in HITRAN'12, and the ratio of the measured values (Z_{meas}) to those in the HITRAN'12 [35] database ($Z_{HT'12}$) are shown in Fig. 10. The filled bars represent the measured values (ranging in the vertical direction by their respective measurement uncertainties) divided by the corresponding magnitudes reported in [35]. The unfilled bars denote the range of uncertainty specified in [35]. It can be noted that γ_{self} and some values of γ_{air} do not have corresponding unfilled bars. This is because [35] only specifies an “average or estimate” uncertainty estimate instead of a numerical one. Some general trends could be observed in the measured values of γ_0 with (J,K) quantum numbers. But since only a few transitions with sequential (J,K) values were studied and differences between broadening coefficients of adjacent J values were mostly comparable to the associated errors, no clear trend could be established. But in cases where the errors were small compared to the differences between the broadening values, trends consistent with Owen et al. [27] were observed. For example, for $J=3$, γ_{0,H_2O} decreases with $K=1-3$, while for $J=3$, $\gamma_{0,self}$ and γ_{0,CO_2} increase with $K=1-3$.

3.3. Uncertainty analysis

For a given line parameter, U , which is a function of measured variables, e_j , the uncertainty in U , (σ_U) is expressed by the following equation:

$$\sigma_U = \sqrt{\sum_i \left(\frac{\partial U}{\partial \varepsilon_i} \sigma_{\varepsilon_i} \right)^2}$$

where σ_{ε_j} is the uncertainty associated with the variable ε_j . Fig. 11 shows the factors considered for the uncertainty analysis and their impact on the measurement uncertainties for the sQ(10,7) transition as an illustrative example. The different factors that influenced the measurement of the measured quantities are explained in details as follows:

3.3.1. Line intensity—The most significant sources of error in the measurement of the line intensities are:

- I. Path length, L : The measurements were performed in the short path length cell. Therefore it was crucial to determine the path length precisely, which was done by measuring the integrated area of the absorption profile of pure CO_2 in the cell using a pre-characterized CO_2 transition centered at 2384.186 cm^{-1} with an uncertainty of less than 1%.
- II. Pressure, P : The manufacturer (MKS instruments) has specified the pressure measurement uncertainty to be 0.12% of the reading and this translates to 0.12% in line intensity as they are inversely proportional.
- III. Integrated area, A : The residual of each fit Voigt profile introduces uncertainty in the inferred integrated area of about 1% due to 2% RMS error at the peak of the profile.
- IV. Slope of the integrated area vs P , $\delta A/\delta P$: The standard error in the slope of the fitted line on the A vs P plot is taken into account by this factor.
- V. Mole fraction of the gas, χ : Since 99.99% pure anhydrous NH_3 was used to measure the line intensity, the mole fraction only leads to a 0.01% uncertainty.
- VI. Temperature of the sample, T : The thermocouple temperature had 0.2% uncertainty in the recorded values. The line intensities of the presented lines have a dependence on T that is primarily a function of the lower state energy of the transition, E'' .

3.3.2. Broadening coefficient—Key factors influencing the uncertainty in the estimation of the broadening coefficients are:

- i. Pressure, P : The broadening halfwidths are directly proportional to pressure. Hence, the error has an 1:1 correlation to uncertainty in P .
- ii. Collision FWHM, $\Delta \nu_C$: The residual of each fitted Voigt profile from the absorbance data can be attributed to some uncertainty in the collision FWHM. This correlates to about 1% uncertainty due to 1% RMS error at the peak.
- iii. Slope of the FWHM vs P , $\delta \Delta \nu_C/\delta P$: The standard error in the slope of the fitted line on the $\Delta \nu_C$ vs P plot is taken into account by this factor.
- iv. Slope of $\ln(\gamma)$ vs $\ln(T)$, $\delta \ln(\gamma)/\delta \ln(T)$: The linearity of this plot was essential to obtaining a power-law behavior expression in γ . This factor captures the effect of the inaccuracy of the power-law model.
- v. Mole fraction of the gas, χ : Presence of 4% Ar in the mixture with bath gas resulted in an associated uncertainty in the corrected broadening coefficient for a specific bath gas Y . This amounts to less than 0.2% uncertainty in γ_{0,NH_3-Y} for $Y \neq \text{Ar}$.
- vi. Temperature of the sample, T : The temperature non-uniformity in the cell (± 4 K), especially at higher temperatures led to an additional uncertainty of 2%.

The errors in the temperature dependence exponents (n) were estimated in a method similar to Nemtchinov et al. [25]. Lines 1–4 show a better uncertainty in n since the value was obtained over a broader temperature range with additional temperature samples.

4. Conclusions

Line intensities and temperature-dependent broadening coefficients of select lines in the ammonia ν_2 fundamental $sQ(J,K)$ manifold are reported. A Voigt profile fitting algorithm was used to fit empirical absorption lineshapes and determine collision broadening FWHMs and integrated areas. The line intensity and self-broadening measurements were performed in a 3.36 mm cell with 99.99% pure NH_3 at 296 K. The reported values were found to be mostly within the uncertainty bounds of the HITRAN'12 database. The broadening coefficients were measured with Ar, O_2 , N_2 , CO_2 and H_2O as collision partners in dilute mixtures within the temperature range of 296–600 K in a 101 cm cell. The measurements indicate a strong dependence on the collision partner. The power-law temperature-dependence exponent was also reported for each species, with the exception of NH_3 self-broadening. The uncertainties in the experimental results represent an improvement over computed or estimated parameters in the current literature, where comparable values are available. More notably, the comprehensive set of spectroscopic parameters determined in this study provides a basis for sensitive and quantitative ammonia sensing over a broad range of gas conditions and mixture compositions for relevant applications.

Acknowledgments

This work was conducted with support from the Spectrum-SPADA innovation grant of Stanford University (NIH UL1 TR 000093), and NIH NICHD STTR Award (1R41HD084055-01).

The content is solely the responsibility of the authors and does not necessarily represent the official views of the National Institutes of Health.

References

1. Lewicki, R., Kosterev, A., Thomazy, DM., Gong, L., Griffin, R., Day, T., et al. Ammonia sensor for environmental monitoring based on a 10.4 μm external-cavity quantum cascade laser. Proceedings of lasers, sources related photonic devices; Washington, D.C.: OSA. 2010; p. LTuD2
2. Schilt S, Thévenaz L, Niklès M, Emmenegger L, Hügli C. Ammonia monitoring at trace level using photoacoustic spectroscopy in industrial and environmental applications. Spectrochim Acta Part A Mol Biomol Spectrosc. 2004; 60:3259–68.
3. Bakhrkin, YA., Kosterev, AA., Wysocki, G., Tittel, FK., Risby, TH., Bruno, JD. quantum cascade laser-based sensor platform for ammonia detection in exhaled human breath. Proceedings of laser applications to chemical, security, and environmental analysis; Washington, D.C.: OSA. 2008; p. LMB4
4. Owen K, Farooq A. A calibration-free ammonia breath sensor using a quantum cascade laser with WMS 2f/1f. Appl Phys B. 2013; 116:371–83.
5. Moskalenko KL, Nadezhdinskii AI, Adamovskaya IA. Human breath trace gas content study by tunable diode laser spectroscopy technique. Infrared Phys Technol. 1996; 37:181–92.
6. Manne J, Sukhorukov O, Jäger W, Tulip J. Pulsed quantum cascade laser-based cavity ring-down spectroscopy for ammonia detection in breath. Appl Opt. 2006; 45:9230. [PubMed: 17151764]
7. Webber ME, Claps R, Englich FV, Tittel FK, Jeffries JB, Hanson RK. Measurements of NH_3 and CO_2 with distributed-feedback diode lasers near 2.0 μm in bioreactor vent gases. Appl Opt. 2001; 40:4395. [PubMed: 18360480]

8. Schilt, S., Besson, J-P., Thévenaz, L., Gyger, M. Continuous and Simultaneous multigas monitoring using a highly sensitive and selective photoacoustic sensor. Conference on lasers and electro-optics, optical society of america; 2005; p. CTuGG4
9. Webber ME, MacDonald T, Pushkarsky MB, Patel CKN, Zhao Y, Marcillac N, et al. Agricultural ammonia sensor using diode lasers and photoacoustic spectroscopy. *Meas Sci Technol*. 2005; 16:1547–53.
10. Pushkarsky MB, Webber ME, Baghdassarian O, Narasimhan LR, Patel CKN. Laser-based photoacoustic ammonia sensors for industrial applications. *Appl Phys B Lasers Opt*. 2002; 75:391–6.
11. Claps R, Englich FV, Leleux DP, Richter D, Tittel FK, Curl RF. Ammonia detection by use of near-infrared diode-laser-based overtone spectroscopy. *Appl Opt*. 2001; 40:4387. [PubMed: 18360479]
12. Totschnig G, Lackner M, Shau R, Ortsiefer M, Roskopf J, Amann MC, et al. High-speed vertical-cavity surface-emitting laser (VCSEL) absorption spectroscopy of ammonia (NH₃) near 1.54 μm. *Appl Phys B*. 2003; 76:603–8.
13. De Tommasi E, Casa G, Gianfrani L. High precision determinations of NH₃ concentration by means of diode laser spectrometry at 2.005 μm. *Appl Phys B*. 2006; 85:257–63.
14. Modugno G, Corsi C. Water vapour and carbon dioxide interference in the high sensitivity detection of NH₃ with semiconductor diode lasers at 1.5 μm. *Infrared Phys Technol*. 1999; 40:93–9.
15. Paldus BA, Harb CC, Spence TG, Zare RN, Gmachl C, Capasso F, et al. Cavity ringdown spectroscopy using mid-infrared quantum-cascade lasers. *Opt Lett*. 2000; 25:666. [PubMed: 18064145]
16. Filho MB, da Silva MG, Sthel MS, Schramm DU, Vargas H, Miklós A, et al. Ammonia detection by using quantum-cascade laser photoacoustic spectroscopy. *Appl Opt*. 2006; 45:4966. [PubMed: 16807606]
17. Kosterev AA, Tittel FK. Ammonia detection by use of quartz-enhanced photoacoustic spectroscopy with a near-ir telecommunication diode laser. *Appl Opt*. 2004; 43:6213. [PubMed: 15605564]
18. Manne J, Jäger W, Tulip J. Sensitive detection of ammonia and ethylene with a pulsed quantum cascade laser using intra and interpulse spectroscopic techniques. *Appl Phys B*. 2008; 94:337–44.
19. Lewicki, R., Kosterev, AA., Thomazy, DM., Risby, TH., Solga, S., Schwartz, TB., et al. Real time ammonia detection in exhaled human breath using a distributed feedback quantum cascade laser based sensor. Proceedings of SPIE OPTO international society for optics and photonics; 2011; p. 79450K–79450K-7.
20. Miller, DJ., Zondlo, MA. Open-path high sensitivity atmospheric ammonia sensing with a 9 μm quantum cascade laser. Proceedings of Conference on lasers electro-optics; Washington, D.C.: OSA. 2010; p. JThJ4
21. Kleiner I, Tarrago G, Cottaz C, Sagui L, Brown LR, Poynter RL, et al. NH₃ and PH₃ line parameters: the 2000 HITRAN update and new results. *J Quant Spectrosc Radiat Transf*. 2003; 82:293–312.
22. Fabian M, Ito F, Yamada KMT. N₂, O₂, and air broadening of NH₃ in ν₂ band measured by FTIR spectroscopy. *J Mol Spectrosc*. 1995; 173:591–602.
23. Aroui H, Nouri S, Bouanich J-P. NH₃ self-broadening coefficients in the ν₂ and ν₄ bands and line intensities in the ν₂ band. *J Mol Spectrosc*. 2003; 220:248–58.
24. Guinet M, Jeseck P, Mondelain D, Pepin I, Janssen C, Camy-Peyret C, et al. Absolute measurements of intensities, positions and self-broadening coefficients of R branch transitions in the ν₂ band of ammonia. *J Quant Spectrosc Radiat Transf*. 2011; 112:1950–60.
25. Nemtchinov V, Sung K, Varanasi P. Measurements of line intensities and half-widths in the 10-μm bands of. *J Quant Spectrosc Radiat Transf*. 2004; 83:243–65.
26. Fabian M, Yamada KMT. Absolute Intensity of the NH₃ ν₂ Band. *J Mol Spectrosc*. 1999; 198:102–9. [PubMed: 10527784]
27. Owen K, Es-sebbar E, Farooq A. Measurements of NH₃ linestrengths and collisional broadening coefficients in N₂, O₂, CO₂, and H₂O near 1103.46 cm⁻¹. *J Quant Spectrosc Radiat Transf*. 2013; 121:56–68.

28. Aroui H, Chevalier M, Broquier M, Picardbersellini A, Gherissi S, Legaysommaire N. Line mixing parameters in the ν_4 rovibrational band of NH_3 perturbed by N_2 . *J Mol Spectrosc*. 1995; 169:502–10.
29. Aroui H, Broquier M, Picard-Bersellini A, Bouanich JP, Chevalier M, Gherissi S. Absorption intensities, pressure-broadening and line mixing parameters of some lines of NH_3 in the ν_4 band. *J Quant Spectrosc Radiat Transf*. 1998; 60:1011–23.
30. Dhib M, Bouanich J-P, Aroui H, Picard-Bersellini A. Analysis of N_2 , O_2 , CO_2 , and air broadening of infrared spectral lines in the ν_4 band of NH_3 . *J Quant Spectrosc Radiat Transf*. 2001; 68:163–78.
31. Markov VN, Pine AS, Buffa G, Tarrini O. Self broadening in the ν_1 band of NH_3 . *J Quant Spectrosc Radiat Transf*. 1993; 50:167–78.
32. Pine AS, Markov VN, Buffa G, Tarrini O. N_2 , O_2 , H_2 , Ar and He broadening in the ν_1 band of NH_3 . *J Quant Spectrosc Radiat Transf*. 1993; 50:337–48.
33. Pine AS, Markov VN. Self- and foreign-gas-broadened lineshapes in the ν_1 band of NH_3 . *J Mol Spectrosc*. 2004; 228:121–42.
34. Schilt S. Impact of water vapor on 1.51 μm ammonia absorption features used in trace gas sensing applications. *Appl Phys B*. 2010; 100:349–59.
35. Rothman LS, Gordon IE, Babikov Y, Barbe A, Chris Benner D, Bernath PF, et al. The HITRAN 2012 molecular spectroscopic database. *J Quant Spectrosc Radiat Transf*. 2013; 130:4–50.
36. Down MJ, Hill C, Yurchenko SN, Tennyson J, Brown LR, Kleiner I. Reanalysis of ammonia spectra: updating the HITRAN 14 NH_3 database. *J Quant Spectrosc Radiat Transf*. 2013; 130:260–72.
37. Hadded S, Thibault F, Flaud P-M, Aroui H, Hartmann J-M. Experimental and theoretical study of line mixing in NH_3 spectra. I. Scaling analysis of parallel bands perturbed by He. *J Chem Phys*. 2002; 116:7544.
38. Aroui H, Ben Mabrouk K, Boussassi R. Line-mixing effect on NH_3 line intensities. *J Quant Spectrosc Radiat Transf*. 2013; 130:273–83.
39. Hanson, RK., Spearrin, RM., Goldenstein, CS. *Spectroscopy and optical diagnostics for gases*. 1. Springer; 2016.




Fig. 1. (a) Simulation of NH₃ bands in the infrared at 308 K (35 °C) from HITRAN' 12, (b) Close-up of the region studied in relation to the interfering species in this region.




Fig. 2. Experimental setup for performing spectroscopic measurements of NH_3 .




Fig. 3. Simulated absorption spectrum of 10 ppm NH₃ in air using spectral parameters from HITRAN'12 at 296 K and 50 Torr over a 100 cm absorption path length indicating the studied transitions.




Fig. 4. Sample spectral fits and residuals for measured absorbances of pure anhydrous ammonia at 296 K in a 3.36 mm optical cell near (a) 962.17 cm⁻¹, (b) 965.35 cm⁻¹ and (c) 967.35 cm⁻¹.




Fig. 5.
Measured integrated area vs pressure at 296 K for the NH_3 ν_2 transitions sQ(3,3), sQ(6,6), sQ(9,9) and sQ(9,8).




Fig. 6. Variation of NH₃ line intensities with J and K quantum numbers for NH₃ ν_2 transitions, $sQ(J,K)$. The dashed lines indicate the values listed in HITRAN'12 database, while the markers indicate the measured values of line intensities.




Fig. 7. Dependence of absorption lineshape on collision partner Y as revealed by the measured spectrum of 500 ppm NH_3 , 4% Ar and Y (Ar, O_2 , N_2 or CO_2) gas balance.




Fig. 8. Sample plots of collision broadening FWHM vs pressure for sQ(9,9) for various (a) collision partners at constant temperature and (b) temperatures with CO₂ as a collision partner.




Fig. 9. Temperature-dependent broadening coefficients ($\gamma(T)$) for the NH₃ ν_2 sQ(9,9) transition for various collision partners, Y displaying power-law behavior in the form of linearity in the ln-ln scale.




Fig. 10. Comparison of measured line intensities, self- and air-broadening coefficients with HITRAN'12 database. The unfilled bars indicate the uncertainties specified in HITRAN'12 database (H12). The transitions with missing H12 bars indicate “average or estimate” uncertainty estimate in HITRAN'12.




Fig. 11. Example uncertainty analysis for sQ(10,7): line intensity, S(296 K) and N₂ broadening coefficient, $\gamma_{0, \text{NH}_3-\text{N}_2}$.

Table 1

Studied transitions and their numerical designation in this article.

Line	Transition	Center frequency [cm^{-1}]
1	sQ(10,7)	961.7762
2	sQ(10,6)	962.1443
3	sQ(9,9)	962.1714
4	sQ(9,8)	962.3884
5	sQ(6,6)	965.3539
6	sQ(6,5)	965.4994
7	sQ(3,3)	967.3463
8	sQ(3,2)	967.4068
9	sQ(3,1)	967.449

Author Manuscript

Author Manuscript

Author Manuscript

Author Manuscript

Measured line intensities of Q-branch NH₃ transitions and comparison with HITRAN'12 database [35]. The superscripts denote the uncertainties in the measurements with respect to the last decimal digit (e.g. for line 1: S(296 K)=0.2195±0.0055). The values in parenthesis for the column for HITRAN'12 represent the uncertainty reported (±20%) in the database.

Table 2

Line	Transition	Center frequency, H12 [cm ⁻¹]	E ^u , H12 [cm ⁻¹]	S(296 K), H12 [cm ⁻² /atm]	S(296 K) Current study [cm ⁻² /atm]	% Difference w.r.t. H12
1	sQ(10,7)	961.7762	907.9651	0.212 ⁽⁴³⁾	0.2195 ⁽⁵⁵⁾	3.4
2	sQ(10,6)	962.1443	955.1064	0.251 ⁽⁵⁰⁾	0.292 ⁽¹⁵⁾	14.3
3	sQ(9,9)	962.1714	592.5867	3.769 ⁽⁷³⁴⁾	3.672 ⁽⁹²⁾	-2.6
4	sQ(9,8)	962.3884	655.6464	1.081 ⁽²¹⁶⁾	1.019 ⁽²⁶⁾	-6.1
5	sQ(6,6)	965.3539	283.574	10.61 ⁽²¹²⁾	10.39 ⁽²⁶⁾	-2.1
6	sQ(6,5)	965.4994	324.369	3.008 ⁽⁶⁰²⁾	3.096 ⁽⁷⁷⁾	2.9
7	sQ(3,3)	967.3463	85.861	12.52 ⁽²⁵¹⁾	13.26 ⁽³³⁾	5.9
8	sQ(3,2)	967.4068	104.422	2.58 ⁽³²⁾	2.621 ⁽⁶⁶⁾	1.6
9	sQ(3,1)	967.449	115.536	0.611 ⁽¹²²⁾	0.669 ⁽⁶⁷⁾	9.6

Table 3

Measured NH₃-Y broadening parameters, γ_0 , and temperature-dependence exponent, n , with associated uncertainties. For γ_0 and n values, the numbers in parentheses are the estimated uncertainties of the measurements in terms of the last digit (e.g. for line 1: $\gamma_{0,N_2}=0.087\pm 0.004$).

Line	Transition	Line center	Quantity	N ₂	O ₂	CO ₂	Ar	NH ₃	H ₂ O
1	sQ(10,7)	961.7762	γ_0	0.087 ⁽⁴⁾	0.044 ⁽²⁾	0.202 ⁽⁸⁾	0.033 ⁽¹⁾	0.443 ⁽¹⁴⁾	0.263 ⁽¹¹⁾
			n	0.66 ⁽⁵⁾	0.69 ⁽⁵⁾	1.14 ⁽⁵⁾	0.59 ⁽⁵⁾		0.86 ⁽¹⁰⁾
2	sQ(10,6)	962.1443	γ_0	0.104 ⁽⁵⁾	0.034 ⁽²⁾	0.216 ⁽²⁰⁾	0.033 ⁽¹⁾	0.423 ⁽²⁶⁾	0.257 ⁽⁸⁾
			n	0.98 ⁽¹⁰⁾	0.41 ⁽¹⁰⁾	1.09 ⁽¹⁰⁾	0.48 ⁽¹⁰⁾		0.86 ⁽²⁰⁾
3	sQ(9,9)	962.1714	γ_0	0.109 ⁽⁴⁾	0.051 ⁽²⁾	0.214 ⁽⁷⁾	0.042 ⁽²⁾	0.57 ⁽¹⁷⁾	0.307 ⁽⁹⁾
			n	0.81 ⁽⁵⁾	0.76 ⁽⁵⁾	0.74 ⁽⁵⁾	0.69 ⁽⁵⁾		0.98 ⁽¹⁰⁾
4	sQ(9,8)	962.3884	γ_0	0.098 ⁽³⁾	0.048 ⁽²⁾	0.187 ⁽¹⁰⁾	0.039 ⁽¹⁾	0.511 ⁽¹⁶⁾	0.316 ⁽¹³⁾
			n	0.70 ⁽⁵⁾	0.67 ⁽⁵⁾	0.69 ⁽⁵⁾	0.64 ⁽⁵⁾		0.90 ⁽¹⁰⁾
5	sQ(6,6)	965.3539	γ_0	0.103 ⁽³⁾	0.052 ⁽²⁾	0.18 ⁽⁶⁾	0.045 ⁽¹⁾	0.553 ⁽²⁸⁾	0.321 ⁽³²⁾
			n	0.76 ⁽¹⁰⁾	0.73 ⁽¹⁰⁾	0.77 ⁽¹⁰⁾	0.79 ⁽¹⁰⁾		
6	sQ(6,5)	965.4994	γ_0	0.102 ⁽³⁾	0.053 ⁽²⁾	0.171 ⁽⁵⁾	0.041 ⁽¹⁾	0.523 ⁽²⁶⁾	0.341 ⁽³⁴⁾
			n	0.63 ⁽¹⁰⁾	0.78 ⁽¹⁰⁾	0.71 ⁽¹⁰⁾	0.64 ⁽¹⁰⁾		
7	sQ(3,3)	967.3463	γ_0	0.109 ⁽³⁾	0.058 ⁽²⁾	0.178 ⁽⁹⁾	0.047 ⁽²⁾	0.557 ⁽²⁸⁾	0.298 ⁽³⁰⁾
			n	0.86 ⁽¹⁰⁾	0.78 ⁽¹⁰⁾	0.76 ⁽¹⁰⁾	0.76 ⁽¹⁰⁾		
8	sQ(3,2)	967.4068	γ_0	0.112 ⁽⁴⁾	0.062 ⁽²⁾	0.156 ⁽⁸⁾	0.049 ⁽²⁾	0.43 ⁽²²⁾	0.383 ⁽³⁹⁾
			n	0.81 ⁽¹⁰⁾	0.84 ⁽¹⁰⁾	0.60 ⁽¹⁰⁾	0.66 ⁽¹⁰⁾		
9	sQ(3,1)	967.449	γ_0	0.104 ⁽⁵⁾	0.053 ⁽³⁾	0.121 ⁽⁶⁾	0.044 ⁽²⁾	0.351 ⁽³⁵⁾	0.486 ⁽⁹⁷⁾
			n	0.83 ⁽¹⁰⁾	0.60 ⁽¹⁰⁾	0.61 ⁽¹⁰⁾	0.90 ⁽¹⁰⁾		