



Low-Cost High-Performance MRI

Citation

Sarracanie, Mathieu, Cristen D. LaPierre, Najat Salameh, David E. J. Waddington, Thomas Witzel, and Matthew S. Rosen. 2015. "Low-Cost High-Performance MRI." *Scientific Reports* 5 (1): 15177. doi:10.1038/srep15177. <http://dx.doi.org/10.1038/srep15177>.

Published Version

doi:10.1038/srep15177

Permanent link

<http://nrs.harvard.edu/urn-3:HUL.InstRepos:23474073>

Terms of Use

This article was downloaded from Harvard University's DASH repository, and is made available under the terms and conditions applicable to Other Posted Material, as set forth at <http://nrs.harvard.edu/urn-3:HUL.InstRepos:dash.current.terms-of-use#LAA>

Share Your Story

The Harvard community has made this article openly available.
Please share how this access benefits you. [Submit a story](#).

[Accessibility](#)

SCIENTIFIC REPORTS



OPEN

Low-Cost High-Performance MRI

Mathieu Sarraclanie^{1,2}, Cristen D. LaPierre^{1,2}, Najat Salameh^{1,2,3}, David E. J. Waddington^{1,2,4}, Thomas Witzel¹ & Matthew S. Rosen^{1,2,5}

Received: 05 May 2015

Accepted: 18 September 2015

Published: 15 October 2015

Magnetic Resonance Imaging (MRI) is unparalleled in its ability to visualize anatomical structure and function non-invasively with high spatial and temporal resolution. Yet to overcome the low sensitivity inherent in inductive detection of weakly polarized nuclear spins, the vast majority of clinical MRI scanners employ superconducting magnets producing very high magnetic fields. Commonly found at 1.5–3 tesla (T), these powerful magnets are massive and have very strict infrastructure demands that preclude operation in many environments. MRI scanners are costly to purchase, site, and maintain, with the purchase price approaching \$1 M per tesla (T) of magnetic field. We present here a remarkably simple, non-cryogenic approach to high-performance human MRI at ultra-low magnetic field, whereby modern under-sampling strategies are combined with fully-refocused dynamic spin control using steady-state free precession techniques. At 6.5 mT (more than 450 times lower than clinical MRI scanners) we demonstrate $(2.5 \times 3.5 \times 8.5)$ mm³ imaging resolution in the living human brain using a simple, open-geometry electromagnet, with 3D image acquisition over the entire brain in 6 minutes. We contend that these practical ultra-low magnetic field implementations of MRI (<10 mT) will complement traditional MRI, providing clinically relevant images and setting new standards for affordable (<\$50,000) and robust portable devices.

Magnetic Resonance Imaging (MRI) is a powerful, non-invasive technique for revealing the internal structure and function of the human body with a rich range of biological contrasts. Despite considerable improvements in imaging quality and speed, the underlying technology remains remarkably unchanged compared to the first generation scanners that emerged on the market 30 years ago. The fact that very strong magnetic fields are needed to overcome the intrinsic lack of sensitivity of NMR-based methods continues to dominate scanner construction, and drives both pricing and scanner siting requirements. MRI scanners are built around massive superconducting magnets with a nominal cost of \$1 M per tesla of magnetic field. With 1.5 tesla (T) and 3T scanners in common use, and increasing demand for 7T, the extreme cost of these devices limits the number of scanners on site and requires hospitals to carefully prioritize patients. Additionally, these massive scanners are strictly confined to the MRI suite within a hospital thus precluding mobile operation in many environments including surgical intervention, triage and primary care suites.

Undeniably, one of the next revolutions in health care will center on cost-effectiveness. Thus the prospect of low-cost (<\$50,000) but high-performance MRI systems to complement traditional MRI scanners is compelling. A promising solution is MRI at very low magnetic field where scalable electro-magnets become practical. Operation at low magnetic field enables imaging in environments where high magnetic fields would be contraindicated (such as in the presence of nearby ferrous materials), and raises the potential for scanners to be built at significantly reduced total cost, and with open geometry designs that ease patient handling and positioning.

The unique role that very low magnetic field MRI scanners can play in neurocritical care was recognized 30 years ago in the pioneering work of Sepponen, *et al.*¹, who explored the clinical validity of brain MRI acquired in a 20 mT scanner located in a hospital emergency department. These early images

¹MGH/A.A. Martinos Center for Biomedical Imaging, 149 13th St, Suite 2301, Charlestown MA 02129, USA.

²Department of Physics, Harvard University, 17 Oxford St, Cambridge, MA 02138, USA. ³Institute of Physics of Biological Systems, Ecole Polytechnique Fédérale de Lausanne, CH-1015 Lausanne, Switzerland. ⁴School of Physics, University of Sydney, Physics Rd, Sydney NSW 2006, Australia. ⁵Harvard Medical School, 25 Shattuck St, Boston, MA 02115, USA. Correspondence and requests for materials should be addressed to M.S.R. (email: mrosen@cfa.harvard.edu)

were acquired at the lowest field strength reported in clinical MRI at that time, and although limited to a single 15 mm slice, were obtained with good contrast in a reasonable four minute acquisition.

In an effort to improve the performance of very low field MRI systems, Macovski and Conolly introduced the concept of pre-polarized MRI (also known as PMRI) in 1993², which employs a strong, inhomogeneous pulsed magnet field to generate increased nuclear polarization, and a second much weaker homogeneous magnetic field for signal detection. This PMRI strategy has been the acquisition strategy for nearly all very low field MRI systems since its introduction. In 2006, PMRI in human subjects with metal implants was reported *in vivo* in human wrists³, where a 0.4 T field was used for pre-polarization, and a 54 mT field used for signal detection.

The ultra-low field (ULF) MRI regime is defined⁴ when the magnetic field used for signal detection is below 10 mT. In 2007, PMRI was demonstrated with detection in the ULF regime, orders of magnitude lower than reported in Venook *et al.*³, using arrays of very sensitive superconducting quantum interference devices (SQUIDs)⁵ as magnetometers to measure the spatially encoded nuclear spin precession⁶. Pre-polarized cryogenic SQUID-detected ULF MRI has been demonstrated in the human brain as well as in the human hand and wrist by several groups^{7–15}. Results from late 2013 demonstrate *in vivo* 2D images of the human brain (pre-polarized to 80 mT) with (2.5×1.9) mm² in-plane resolution over a (10×10) cm² region of interest and a 100 mm thick slice, acquired in ~26 minutes¹⁶. Very recent results from the Los Alamos ULF effort demonstrate 3D images of the human brain (pre-polarized to 100 mT) with $(2.1 \times 2.4 \times 15)$ mm³ resolution (5 slices) in 67 minutes¹⁵.

Although SQUIDs are the most technically mature of the non-inductive magnetometers used at ULF, several alternative detection technologies have been explored. Optical measurement of nitrogen-vacancy (NV) color centers in diamond^{17–19} form the basis of robust solid-state magnetometers with unmatched magnetic field sensitivity at nanoscale resolutions. As of yet, however, NV-diamond magnetometers do not provide obvious benefits for human scale MRI. Atomic magnetometers (AM) have also been applied to pre-polarized NMR²⁰ and MRI^{21,22}, and improvements in these devices have resulted in a magnetic field sensitivity approaching SQUID performance²³ without the need for cryogenics. The first attempt at imaging the living human brain with an atomic magnetometer was reported in 2013²⁴. In this work, nuclear spins are pre-polarized at 80 mT and detection is performed at 4 mT. Despite the ultra-high sensitivity and dynamic range of the AM magnetometer, the setup as described provides limited 3D coverage and significant improvement in resolution and SNR (Signal-to-Noise Ratio) is needed in order to clearly discern anatomical features, which will inevitably increase the acquisition time.

Independent of which detection technology is used, all pre-polarized ULF MRI suffers from intrinsically long acquisition times, most of which is incompressible, that result from the time needed to generate nuclear polarization. In the present work, we demonstrate fast and efficient brain ULF MRI at 6.5 mT with no pre-polarization nor cryogenics, combining under-sampling strategies with a high performance fully refocused steady-state-based acquisition in a simple, inexpensive system. With a novel inductive single channel detector, we report the fastest 3D MRI of the living human brain in the ULF regime compared to the state-of-the-art as reported in the literature^{11,12,15,16,24}.

Results

Ultra-low field acquisition strategy. High performance imaging at ultra-low magnetic field focuses on significantly reducing acquisition time using fast imaging techniques. Here, fast imaging was enabled using 3D balanced steady state free precession sequences (b-SSFP)²⁵. Originally described by Carr in 1958 as a technique for improving the signal-to-noise ratio (SNR) in NMR experiments²⁶, b-SSFP was implemented as an efficient acquisition strategy for MRI in 1986 by Oppelt *et al.*²⁷, and extensively investigated in the early 2000s^{25,28–32}. Unlike traditional gradient- and spin-echo techniques, b-SSFP sequences dynamically refocus spin magnetization following measurement, eliminating the extra delays typically used for T_2 decay and T_1 recovery. This considerably reduces acquisition times and provides the highest SNR per unit time of all imaging sequences^{25,27}. These sequences are very sensitive to the amount of spin dephasing that occurs between consecutive RF pulses (the pulse repetition time, TR), and typical banding artifacts are expected to appear within a range of $\pm 1/(2 \cdot TR)$ Hz that result from inhomogeneity in the static magnetic field²⁵. This sets a strict requirement on the absolute field homogeneity over the field-of-view (FOV), which for operation at 3 T is typically at the sub-PPM level.

In the millitesla regime, however, the fractional homogeneity requirement is three orders of magnitude lower, significantly easing the engineering burden for low-field magnet design. With a current TR = 22.5 ms, our b-SSFP sequence is completely immune to banding artifacts for up to 160 ppm inhomogeneity at 6.5 mT. Furthermore, magnetic susceptibility differences are significantly reduced at ULF, preventing off-resonance b-SSFP artifacts. As a result, provided reasonable magnetic field homogeneity, b-SSFP at very low magnetic field alleviates the necessity of ultra-short TRs and provides good image quality over a large FOV without the need for sophisticated ultrafast gradient power amplifiers. Our 6.5 mT MRI scanner³³ (Fig. 1) was upgraded for improved B_0 stability³⁴, and was used for all the low-field b-SSFP experiments described here.

RF Coil design. The design of inductive detection coils for use in ULF MRI presents a different set of challenges to those present in conventional high-field MRI. In particular, issues of coil resistance and




Figure 1. Ultra-low field MRI system. Custom built biplanar 6.5 mT electromagnet with biplanar gradients (G_x , G_y , and G_z). The diameter of the outermost B_0 coil is 220 cm. The subject lays supine in the scanner and a custom built single channel transmit/receive spiral head coil wound with litz wire for operation at 276 kHz is placed to cradle the head.




Figure 2. 3D renderings of the single channel form-fitting head coil. (A) isometric, (B) back, and (C) side views are shown. The final design was 3D printed on a Fortus 360 mc printer (Stratasys, Eden Prairie, MN, USA) in polycarbonate using fused deposition modeling technology. The 30-turn spiral was wound with Type 1 40/38 Litz wire, parallel resonated to 276 kHz, and capacitively matched to 50 ohms.

probe bandwidth manifest differently. In conventional MRI, the dominant source of noise is the presence of small currents in the lossy sample (the so-called “body noise” regime) to which a characteristic sample resistance R_s is attributed. Both the sample and the coil contribute to Johnson noise but in practice R_s is much larger than the coil resistance R_c (i.e. $R_s \gg R_c$), and thus R_c can be neglected in SNR calculations. However, at low field, R_s becomes much smaller and R_c becomes the dominant noise contribution (i.e., the so-called Johnson noise dominated regime). To minimize the coil resistance in a simple design, larger diameter wire or stranded litz wire can be used, but one needs to consider the impact this has on coil bandwidth. Given the maximum imaging gradient strength of ~ 1 mT/m attainable in our 6.5 mT electromagnet Low Field Imager (LFI), a 20 cm (head-sized) FOV will span a frequency encode bandwidth of ~ 10 kHz. This sets the minimum bandwidth needed for the detection circuit so as to not significantly convolve the coil response function with the object being imaged. At our Larmor frequency of 276 kHz, this corresponds to a maximum coil Q of ~ 30 . A single channel inductive coil for operation at 276 kHz (Fig. 2) was designed and built using 3D printing fused deposition modeling technology and multi-strand litz wire³⁵. A 30-turn 3D Archimedean spiral with an aligned turn-to-turn distance of 5.6 mm guided wire placement, thus ensuring that \vec{B}_1 produced by the spiral pattern is everywhere orthogonal to the main magnetic field \vec{B}_0 . The hemispheric spiral design results in a very homogeneous magnetic field^{36,37}




Figure 3. 3D images of the living brain acquired in 6 minutes at 6.5 mT in (a) axial, (b) coronal, and (c) sagittal orientation. The corresponding maximum SNRs are a. 15, b. 21, and c. 16. Acquisition matrix: $64 \times 75 \times 15$, voxel size: a. $(2.5 \times 3.5 \times 8.5) \text{ mm}^3$, b. $(2.5 \times 3.5 \times 11.5) \text{ mm}^3$, and c. $(2.5 \times 3.5 \times 14.4) \text{ mm}^3$.

over the volume of interest, making it suitable for both RF transmit and receive. The number of turns in the coil was chosen to obtain the inductance needed to achieve the desired Q . Litz wire was preferred in this low frequency application due to its lower AC resistance compared to solid copper wire of the same physical size.

Image reconstruction and processing. MRI images are reconstructed from frequency- and phase-encoded information in the k -space formalism^{38,39}. Previously, we described our use of under-sampling strategies to accelerate low-field imaging³⁴. We make use of this here by randomly sampling 50% of k -space using a variable density Gaussian pattern. The variable density Gaussian sampling pattern emphasizes sampling in the center of k -space, where most of the information is located, and randomly skips lines near the edges. The resulting images do not exhibit coherent artifacts, such as typical wrap-around ghosts due to FOV contraction. Missing values in the acquired k -space were set to zero. The standard deviation of the sampling pattern as a fraction of the FOV was optimized to preserve adequate high-frequency information. Once reconstructed, the images were apodized and processed using Perona and Malik anisotropic diffusion filtering^{40,41} (ADF). ADF is a powerful denoising filter that convolves images of interest with adaptive Gaussian kernels. The Perona and Malik approach works as an iterative multi-scale smoothing and edge detection process that removes noise but prevents image blurring by adjusting filter sharpness as a function of signal intensity gradients.

In vivo brain MRI at 6.5 mT. Three-dimensional under-sampled images acquired at 6.5 mT in 6 minutes are shown in Fig. 3 for each of the three spatial orientations (axial, coronal, and sagittal). The maximum image SNR was computed from the ratio of maximum signal amplitude to the standard deviation over a user defined noise region; SNR of 15, 21, and 16 were measured in axial, coronal, and sagittal orientations respectively. With a maximum gradient strength of $\sim 1 \text{ mT} \cdot \text{m}^{-1}$ and maximal slew rate of $0.7 \text{ mT} \cdot \text{m}^{-1} \cdot \text{ms}^{-1}$, no artifacts from concomitant field effect are seen over the 20 cm field of view. The sinuses are easily recognizable in black on the images, as well as the skull. Surrounding the brain, we can identify the dura in bright grey on the coronal and sagittal images (Fig. 3b,c). In the brain, the two hemispheres and the cerebellum are distinct, and cortical tissue can be distinguished from white matter. Liquid compartments, here CSF, appear in bright grey and white. Images acquired in the axial orientation are compared to images acquired in the same subject at high magnetic field (3 T) using traditional T_1 , T_2 , and proton density (PD) weighted sequences (Fig. 4). In b-SSFP, contrast is related to the T_2/T_1 ratio of the imaged sample²⁵. At high field, liquids and tissue typically have rather different relaxation times but at 6.5 mT their ratio (T_2/T_1) is of order unity resulting in the distinct PD-weighted contrast of Fig. 4b. Most of the anatomic features seen at 3 T can be identified in the ultra-low-field scans. Figure 5 compares b-SSFP at 3 T to 0.0065 T.




Figure 4. Comparison of single channel ULF MRI to 32-channel high magnetic field scans. (a) b-SSFP at 6.5 mT. (b–d), PD , T_2 , and T_1 weighted contrast at 3 T, respectively. Most of the anatomic features seen at higher magnetic field can be identified on the ultra-low field scans. At low field, T_2 approaches T_1 , and the resulting image contrast in (a) is very similar to PD -weighting (b).

The maximum SNR in the high field (HF) b-SSFP image is 317. In order to interpret the difference in SNR between the HF and the ULF scans, we scale the HF image SNR by a factor corresponding to the difference in the ULF spatial resolution ($2.5 \times 3.5 \times 8.5$ in the axial orientation), and by a second factor to account for signal averaging as done at ULF (Fig. 5b, $NA = 160$). The SNR in the downscaled HF scan is equivalent to $317 \times (2.5 \times 3.5 \times 8.5) \times \sqrt{160} \approx 300000$, which gives the ratio $\frac{SNR_{3T}}{SNR_{0.0065}} \approx 8300$ in the axial orientation. If we assume similar coil performance, and neglect the difference in magnetization at steady state for the two magnetic fields, the resulting 8300-fold difference in SNR agrees reasonably well with the simplified approximation that SNR increases with magnetic field to the $3/2$ power⁴², here $460^{3/2} = 9866$. Strong banding artifacts appear at high field (Fig. 5a, yellow arrows), mainly due to magnetic susceptibility differences at the air-tissue interface. At 276 kHz, on the other hand, no imaging artifact is seen over a 20 cm FOV despite a $3 \times$ longer TR (Fig. 5b). Our results demonstrate excellent immunity to magnetic field inhomogeneity of the order of ± 22 Hz, i.e., 160 ppm at 6.5 mT.

Discussion

The work presented here demonstrates the shortest acquisition times and highest SNR per unit time in ULF MRI to date owing to our use of modern sparse sampling strategies and a fully refocused sequence in an optimized electromagnet scanner. These images were acquired without pre-polarization techniques, at a fixed magnetic field and with a simple single channel inductive detector. With an eye towards optimization, we note that for a given spatial resolution, the minimum TR—and consequently the total scan time—is limited by the maximum attainable time-integrated gradient strength. The maximum gradient strength in the LFI is currently $\sim 1 \text{ mT} \cdot \text{m}^{-1}$, resulting in a minimum TR of ~ 23 ms. Weak gradients especially impact phase encoding in balanced sequences like b-SSFP, as every phase-encode pulse is paired with an opposite polarity rewinding pulse. An increase in gradient strength would allow shorter




Figure 5. Comparison of b-SFFP images at (a) 3T and (b) 6.5 mT. Strong banding artifacts appear at high magnetic field (yellow arrows) in all orientations (coronal, sagittal, and axial) whereas no artifact is seen in the images acquired at ultra-low field.

phase encode pulses, thus decreasing total imaging time while maintaining SNR, provided that image distortion from non-linear magnetic fields that accompany the desired encoding gradient (the so called “concomitant field” artifacts⁴³) can be mitigated. At 6.5 mT, an increase in gradient strength in the range of $2\text{--}5\times$, combined with efficient strategies to eliminate concomitant field artifacts^{44–46}, can reasonably be envisioned. Additionally, improvements in the electronic noise floor can go a long way to improving scanner sensitivity. In our system, the scanner noise floor is dominated by poor filtering of the high current lines from the gradient power amplifiers into the RF shielded enclosure of the LFI. More effective filtering of electronic noise coming from our gradient power amplifiers would reduce our system noise floor by a measured factor of 3, thus decreasing the total acquisition time by another factor of $3^2 = 9$. As SNR increases with magnetic field to the $3/2$ power⁴², a simple doubling of magnetic field would result in a sequence about 8 times faster with similar SNR. In the case of human brain imaging, images with similar resolution and SNR as presented here could be then acquired in less than 3 seconds in such an optimized scanner.

A key challenge in obtaining clinically relevant MRI images at ULF is the ability to acquire T_1 and/or T_2 relaxation-weighted images, and thereby provide contrast to different types of tissue. Typically, magnetization prepared gradient-echo, and spin-echo sequences are used to obtain relaxation-weighted images, but these types of imaging experiments become prohibitively time consuming at ultra-low magnetic fields where signal averaging and recovery of the longitudinal magnetization are required. We have investigated a new strategy to provide contrast based on b-SSFP called “magnetic resonance fingerprinting” (MRF)⁴⁷, and have successfully started its implementation at 6.5 mT.⁴⁸

Finally, theoretical frameworks exist that allow image reconstruction of highly undersampled datasets with multiple channel acquisition^{49–56}. Acquisition schemes combining high undersampling rates with parallel imaging techniques such as SENSE⁵⁷ or GRAPPA⁵⁸ could reduce the total acquisition time even further, to less than a second. Recent work from Murphy *et al.*⁵⁹ successfully mitigates the computational expense by exploiting massively parallelized computing.

We contend that ULF MRI scanners operating at this expected level of performance could complement traditional MRI by relieving hospital congestion and shortening triage delays. Outside of the radiology suite, mobile ULF scanners might be deployable during military conflicts or during sport events and enable the acquisition of immediate after-trauma knowledge, typically in the case of traumatic brain

injuries. Finally, ULF MRI technology may allow resource-poor environments access to MRI systems, without the strict siting requirements and high costs of conventional scanners.

Methods

Consent and IRB. Informed consent was obtained from each healthy human volunteer prior to the experiment in accordance with the Human Research Committee of the Massachusetts General Hospital (MGH). All MRI imaging was performed in accordance with approved guidelines and regulations, using experimental protocols that were approved by the MGH Human Research Committee.

Ultra-low field MRI. All ULF MR images were acquired at 6.5 mT in a healthy human volunteer with the 30-turn single channel well-fitting spiral head coil described above (Fig. 2). The subject was placed supine, head first into the electromagnet. The b-SSFP 3D sequence with under-sampling rate of 50%, acquisition matrix of $64 \times 75 \times 15$, at a spatial resolution of $(2.5 \times 3.5 \times 8.5) \text{ mm}^3$, $(2.5 \times 3.5 \times 14.4) \text{ mm}^3$, and $(2.5 \times 3.5 \times 11.5) \text{ mm}^3$, was used in axial, sagittal and coronal orientations, respectively. TR = 22.5 ms, TE = 11 ms. Total acquisition time was 6 minutes for images with number of averages (NA) = 30 (Fig. 3), and 32 minutes with NA = 160 (Figs 4 and 5). All acquired data were processed using anisotropic diffusion filtering and interpolated in Fourier domain in two dimensions to a 96×96 matrix.

High field MRI. Reference high magnetic field images of the head were acquired in the same subject at 3 T on a standard whole-body scanner (Skyra, Siemens Healthcare) using a 32-channel head receiver coil with the subject in a supine position. All high field sequences were acquired with an acceleration factor of 2. Proton density, T_2 , and T_1 weighted sequences were acquired with matrix = $256 \times 256 \times 176$ at a spatial resolution of $(1 \times 1 \times 1) \text{ mm}^3$, with total acquisition times of 10 minutes, 5 minutes and 6 minutes respectively. A b-SSFP sequence was acquired with matrix = $256 \times 256 \times 192$ at a spatial resolution of $(1 \times 1 \times 1) \text{ mm}^3$, and NA = 1. The total acquisition time was ~3 minutes.

References

- Sepponen, R. E., Sipponen, J. T. & Sivula, A. Low Field (0.02 T) Nuclear Magnetic Resonance Imaging of the Brain. *Journal of Computer Assisted Tomography* **9**, 237 (1985).
- Macovski, A. & Conolly, S. Novel approaches to low-cost MRI. *Magnetic Resonance in Medicine* **30**, 221–230 (1993).
- Venook, R. D. *et al.* Prepolarized magnetic resonance imaging around metal orthopedic implants. *Magnetic Resonance in Medicine* **56**, 177–186 (2006).
- Robert Kraus, J., Espy, M., Magnelind, P. & Volegov, P. *Ultra-Low Field Nuclear Magnetic Resonance*. (Oxford University Press, 2014). doi: 10.1093/med/9780199796434.001.0001/med-9780199796434.
- Clarke, J. & Braginski, A. I. *The SQUID Handbook: Fundamentals and Technology of SQUIDS and SQUID Systems, Volume I*. (Wiley-VCH Verlag GmbH & Co. KGaA, 2004).
- Clarke, J., Hatridge, M. & Mößle, M. SQUID-Detected Magnetic Resonance Imaging in Microtesla Fields. *Annu. Rev. Biomed. Eng.* **9**, 389–413 (2007).
- Zotev, V. S. *et al.* Parallel MRI at microtesla fields. *Journal of Magnetic Resonance* **192**, 197–208 (2008).
- Zotev, V. S. *et al.* Microtesla MRI of the human brain combined with MEG. *Journal of Magnetic Resonance* **194**, 115–120 (2008).
- Zotev, V. S. *et al.* SQUID-based microtesla MRI for *in vivo* relaxometry of the human brain. *Applied Superconductivity, IEEE Transactions on* **19**, 823–826 (2009).
- Savukov, I. *et al.* Non-cryogenic anatomical imaging in ultra-low field regime: Hand MRI demonstration. *Journal of Magnetic Resonance* **211**, 101–108 (2011).
- Magnelind, P. E. *et al.* Co-Registration of Interleaved MEG and ULF MRI Using a 7 Channel Low-Tc SQUID System. *Applied Superconductivity, IEEE Transactions on* **21**, 456–460 (2011).
- Vesonen, P. T. *et al.* Hybrid ultra-low-field MRI and magnetoencephalography system based on a commercial whole-head neuromagnetometer. *Magnetic Resonance in Medicine* **69**, 1795–1804 (2012).
- Espy, M., Matlashov, A. & Volegov, P. SQUID-detected ultra-low field MRI. *Journal of Magnetic Resonance* **228**, 1–15 (2013).
- Savukov, I., Karaulanov, T., Wurden, C. J. V. & Schultz, L. Non-cryogenic ultra-low field MRI of wrist-forearm area. *Journal of Magnetic Resonance* **233**, 103–106 (2013).
- Espy, M. A. *et al.* Progress Toward a Deployable SQUID-Based Ultra-Low Field MRI System for Anatomical Imaging. *Applied Superconductivity, IEEE Transactions on* **25**, 1–5 (2015).
- Inglis, B. *et al.* MRI of the human brain at 130 microtesla. *Proceedings of the National Academy of Sciences of the United States of America* **110**, 19194–19201 (2013).
- Maze, J. R. *et al.* Nanoscale magnetic sensing with an individual electronic spin in diamond. *Nature* **455**, 644–647 (2008).
- Taylor, J. M. *et al.* High-sensitivity diamond magnetometer with nanoscale resolution. *Nat Phys* **4**, 810–816 (2008).
- Devience, S. J. *et al.* Nanoscale NMR spectroscopy and imaging of multiple nuclear species. *Nat Nano* **10**, 129–134 (2015).
- Savukov, I. M. & Romalis, M. V. NMR detection with an atomic magnetometer. *Phys. Rev. Lett.* **94**, 123001 (2005).
- Xu, S., Rochester, S. M., Yashchuk, V. V., Donaldson, M. H. & Budker, D. Construction and applications of an atomic magnetic gradiometer based on nonlinear magneto-optical rotation. *Review of Scientific Instruments* **77**, 083106 (2006).
- Xu, S. *et al.* Magnetic resonance imaging with an optical atomic magnetometer. *Proceedings of the National Academy of Sciences* **103**, 12668–12671 (2006).
- Kominis, I. K., Kornack, T. W., Allred, J. C. & Romalis, M. V. A subfemtotesla multichannel atomic magnetometer. *Nature* **422**, 596–599 (2003).
- Savukov, I. & Karaulanov, T. Magnetic-resonance imaging of the human brain with an atomic magnetometer. *Applied Physics Letters* **103**, 043703 (2013).
- Scheffler, K. & Lehnhardt, S. Principles and applications of balanced SSFP techniques. *European Radiology* **13**, 2409–2418 (2003).
- Carr, H. Steady-State Free Precession in Nuclear Magnetic Resonance. *Physical Review* **112**, 1693–1701 (1958).
- Oppelt, A. *et al.* FISP—a new fast MRI sequence. *Electromedica* **54**, 15–18 (1986).
- Plein, S. *et al.* Steady-state free precession magnetic resonance imaging of the heart: Comparison with segmented k-space gradient-echo imaging. *Journal of Magnetic Resonance Imaging* **14**, 230–236 (2001).

29. Scheffler, K., Heid, O. & Hennig, J. Magnetization preparation during the steady state: Fat-saturated 3D TrueFISP. *Magnetic Resonance in Medicine* **45**, 1075–1080 (2001).
30. Deshpande, V. S. *et al.* 3D magnetization-prepared true-FISP: A new technique for imaging coronary arteries. *Magnetic Resonance in Medicine* **46**, 494–502 (2001).
31. Miller, K. L. *et al.* Functional brain imaging using a blood oxygenation sensitive steady state. *Magnetic Resonance in Medicine* **50**, 675–683 (2003).
32. Schmitt, P. *et al.* Inversion recovery TrueFISP: Quantification of T1, T2, and spin density. *Magnetic Resonance in Medicine* **51**, 661–667 (2004).
33. Tsai, L. L., Mair, R. W., Rosen, M. S., Patz, S. & Walsworth, R. L. An open-access, very-low-field MRI system for posture-dependent ³He human lung imaging. *Journal of Magnetic Resonance* **193**, 274–285 (2008).
34. Sarracanie, M., Armstrong, B. D., Stockmann, J. & Rosen, M. S. High speed 3D overhauser-enhanced MRI using combined b-SSFP and compressed sensing. *Magnetic Resonance in Medicine* **71**, 735–745 (2014).
35. LaPierre, C. D., Sarracanie, M., Waddington, D. E. J. & Rosen, M. S. A single channel spiral volume coil for *in vivo* imaging of the whole human brain at 6.5mT, Presented at the International Society of Magnetic Resonance in Medicine, 5902, Toronto (2015).
36. Harpen, M. D. The spherical birdcage resonator. *Journal of Magnetic Resonance* (1969) **94**, 550–556 (1991).
37. Everett, J. E. & Osmeikhian, J. E. Spherical coils for uniform magnetic fields. *J. Sci. Instrum.* **43**, 470 (1966).
38. Ljunggren, S. A simple graphical representation of fourier-based imaging methods. *Journal of Magnetic Resonance* (1969) **54**, 338–343 (1983).
39. Twieg, D. B. The k-trajectory formulation of the NMR imaging process with applications in analysis and synthesis of imaging methods. *Med Phys* **10**, 610–621 (1983).
40. Perona, P. & Malik, J. Scale-space and edge detection using anisotropic diffusion. *IEEE Trans. Pattern Anal. Machine Intell.* **12**, 629–639 (1990).
41. Weickert, J. In *Lecture Notes in Computer Science* (eds Haar Romeny, ter, B., Florack, L., Koenderink, J. & Viergever, M.) **1252**, 1–28–28 (Springer Berlin Heidelberg, 1997).
42. Hoult, D. I. & Richards, R. E. The signal-to-noise ratio of the nuclear magnetic resonance experiment. *Journal of Magnetic Resonance* **24**, 71–85 (1976).
43. Sica, C. T. & Meyer, C. H. Concomitant gradient field effects in balanced steady-state free precession. *Magnetic Resonance in Medicine* **57**, 721–730 (2007).
44. Myers, W. R., Mößle, M. & Clarke, J. Correction of concomitant gradient artifacts in experimental microtesla MRI. *Journal of Magnetic Resonance* **177**, 274–284 (2005).
45. Volegov, P. L., Mosher, J. C., Espy, M. A. & Kraus, R. H., Jr. On concomitant gradients in low-field MRI. *Journal of Magnetic Resonance* **175**, 103–113 (2005).
46. Nieminen, J. O. & Ilmoniemi, R. J. Solving the problem of concomitant gradients in ultra-low-field MRI. *Journal of Magnetic Resonance* **207**, 213–219 (2010).
47. Ma, D. *et al.* Magnetic resonance fingerprinting. *Nature* **495**, 187–192 (2013).
48. Sarracanie, M., Cohen, O. & Rosen, M. S. 3D Balanced-EPI Magnetic Resonance Fingerprinting at 6.5 mT, Presented at the International Society of Magnetic Resonance in Medicine, 6266, Toronto (2015).
49. Block, K. T., Uecker, M. & Frahm, J. Undersampled radial MRI with multiple coils. Iterative image reconstruction using a total variation constraint. *Magnetic Resonance in Medicine* **57**, 1086–1098 (2007).
50. Liang, D., Liu, B., Wang, J. & Ying, L. Accelerating SENSE using compressed sensing. *Magnetic Resonance in Medicine* **62**, 1574–1584 (2009).
51. Otazo, R., Kim, D., Axel, L. & Sodickson, D. K. Combination of compressed sensing and parallel imaging for highly accelerated first-pass cardiac perfusion MRI. *Magnetic Resonance in Medicine* **64**, 767–776 (2010).
52. Trzasko, J. D. *et al.* Sparse-CAPR: Highly accelerated 4D CE-MRA with parallel imaging and nonconvex compressive sensing. *Magnetic Resonance in Medicine* **66**, 1019–1032 (2011).
53. Ying, L. & Sheng, J. Joint image reconstruction and sensitivity estimation in SENSE (JSENSE). *Magnetic Resonance in Medicine* **57**, 1196–1202 (2007).
54. Knoll, F., Clason, C., Bredies, K., Uecker, M. & Stollberger, R. Parallel imaging with nonlinear reconstruction using variational penalties. *Magnetic Resonance in Medicine* **67**, 34–41 (2012).
55. Huang, F. *et al.* A rapid and robust numerical algorithm for sensitivity encoding with sparsity constraints: Self-feeding sparse SENSE. *Magnetic Resonance in Medicine* **64**, 1078–1088 (2010).
56. Lustig, M. & Pauly, J. M. SPIRiT: Iterative self-consistent parallel imaging reconstruction from arbitrary k-space. *Magnetic Resonance in Medicine* **64**, 457–471 (2010).
57. Pruessmann, K. P., Weiger, M., Scheidegger, M. B. & Boesiger, P. SENSE: sensitivity encoding for fast MRI. *Magnetic Resonance in Medicine* **42**, 952–962 (1999).
58. Griswold, M. A. *et al.* Generalized autocalibrating partially parallel acquisitions (GRAPPA). *Magnetic Resonance in Medicine* **47**, 1202–1210 (2002).
59. Murphy, M. *et al.* Fast L1-SPIRiT Compressed Sensing Parallel Imaging MRI: Scalable Parallel Implementation and Clinically Feasible Runtime. *Medical Imaging, IEEE Transactions on* **31**, 1250–1262 (2012).

Acknowledgements

This work was supported by the U.S. Army Medical Research and Materiel Command (USAMRMC), Defense Medical Research and Development Program (DMRDP) award W81XWH-11-2-0076 (DM09094). N.S. was supported by the Swiss National Science Foundation (P300P2_147768). D.E.J.W. was supported by ANSTO and the Australian-American Fulbright Commission. This research was carried out at the Athinoula A. Martinos Center for Biomedical Imaging at the Massachusetts General Hospital, using resources provided by the Center for Functional Neuroimaging Technologies, P41EB015896, a P41 Biotechnology Resource Grant supported by the National Institute of Biomedical Imaging and Bioengineering (NIBIB), National Institutes of Health. The authors thank Bruce Rosen and Ronald Walsworth for their comments on this manuscript.

Author Contributions

M.S., C.D.L., N.S., D.E.J.W., T.W. and M.S.R. all contributed significantly to the work presented in this manuscript.

Additional Information

Competing financial interests: The authors declare no competing financial interests.

How to cite this article: Sarracanie, M. *et al.* Low-Cost High-Performance MRI. *Sci. Rep.* **5**, 15177; doi: 10.1038/srep15177 (2015).



This work is licensed under a Creative Commons Attribution 4.0 International License. The images or other third party material in this article are included in the article's Creative Commons license, unless indicated otherwise in the credit line; if the material is not included under the Creative Commons license, users will need to obtain permission from the license holder to reproduce the material. To view a copy of this license, visit <http://creativecommons.org/licenses/by/4.0/>