

RESEARCH ARTICLE

Open Access



Measurement of scoliosis Cobb angle by end vertebra tilt angle method

Jing Wang^{1†}, Jin Zhang^{1†}, Rui Xu², Tie Ge Chen¹, Kai Sheng Zhou¹ and Hai Hong Zhang^{1*} 

Abstract

Background: Scoliosis is a common deformity, and its severity is usually assessed by measuring the Cobb angle on the spinal X-ray film. The measurement of the Cobb angle is an important basis for selecting therapeutic methods and evaluating therapeutic effects. To measure and calculate the scoliosis Cobb angle by end vertebra tilt angle method (tilt angle method) and assess its accuracy and usability.

Methods: It is deduced that the Cobb angle is the sum of upper and lower end vertebra tilt angles through the law of plane geometry. The project included 32 patients with scoliosis who have received treatment in our hospital from June 2011 to July 2016, whose Cobb angles were measured at various segments (total 50). The measuring results of the tilt angle method and the classical method were compared, and the time spent for the measurement of the two groups was respectively recorded with an electronic stopwatch for comparison. The interference of line marking in imaging data pixel in the two groups was compared using Beyond Compare software.

Results: The measuring results through PACS (picture archiving and communication systems) were regarded as the reference standard. There was no statistical difference for measuring the Cobb angle between the PACS method, end vertebra tilt angle method, and classical method. The end vertebra tilt angle method takes less measuring time than the classical method. The measuring error between the classical method and the tilt angle method showed no statistical significance for the difference.

Conclusion: The scoliosis Cobb angle can be measured accurately and rapidly using the principle of the Cobb angle being equal to the sum of tilt angles of the upper and lower end vertebra, where in the film data of imaging will not be easily contaminated. Under special conditions, the average measuring error is $\pm 3^\circ$.

Keywords: End vertebra, Tilt angle, Scoliosis, Cobb angle

Background

The scoliosis Cobb angle is an important index of disease assessment. The classical method is used to determine the upper/lower end vertebrae (UEV/LEV) on the whole spine anteroposterior X-ray film; then, draw a vertical line respectively at the upper/lower end vertebra endplate lines (UEVEL/LEVEL), and the included angle of the two vertical lines is the Cobb angle [1] (Fig. 1). Manually drawing a line on the image film for measurement is needed in this method and hence is slightly cumbersome, and the line markings can easily contaminate the image data. This study

deduced the geometry law of the classical measurement method to calculate the Cobb angle by the end vertebra tilt angle measurement method (tilt angle method) and assess its accuracy, quickness, and contaminated interference in the image data.

Methods

General data

This group of patients included 10 males and 22 females, aged 11~24 with an average age of 16. Among which, 22 patients suffered from idiopathic scoliosis, 8 patients suffered from congenital scoliosis, and 2 patients suffered from neuromuscular scoliosis. There were 50 scoliosis segments totally including 24 main thoracic scoliosis segments, 18

* Correspondence: m13519698516_1@163.com

[†]Jing Wang and Jin Zhang contributed equally to this work.

¹Department Of Orthopedics, Orthopedics Key Laboratory of Gansu Province, Lanzhou University Second Hospital, Cuiyingmen, Lanzhou 730030, Gansu province, China

Full list of author information is available at the end of the article






Fig. 1 Measurement of Cobb angle by classical method

thoracolumbar scoliosis segments, and 8 lumbar scoliosis segments.

Measurement of Cobb angle by PACS

Anteroposterior X-ray of the whole spine was performed on patients to determine the UEV and LEV of scoliosis segments; then, UEVEL and LEVEL were marked on the film using the PACS (picture archiving and communication systems) built-in measuring procedure, and the Cobb angle was automatically calculated. The measurement was made by the same physician. To reduce the self-measuring deviation of the observer, the angle data of each patient was measured three times every 2 weeks and then the average value was taken as the final result.

End vertebra tilt angle measurement method

The classical method for measuring the Cobb angle is to draw a vertical line respectively at UEVEL and LEVEL, and the included angle between the two vertical lines is the Cobb angle. The following auxiliary lines are drawn: the horizontal lines AB and CD. The angle of AB with UEVEL is α , and the angle of CD with LEVEL is β , which are respectively the upper/lower end vertebra tilt angles (UEVTA/LEVTA) (Fig. 2a). The parallel line AB is drawn through the vertex of Cobb angle O (Fig. 2b). The sideline of the Cobb angle that extended to line AB intersects at E, the parallel line CD is drawn through the vertex of Cobb angle O, and the sideline of the Cobb angle that extended to line CD intersects at F (Fig. 2c). It can be deduced in accordance with parallelogram law and supplementary angle law that:

$$\begin{aligned}\angle \text{Cobb} &= 180^\circ - \angle \text{AEO} - \angle \text{CFO} \\ &= 180^\circ - (90^\circ - \alpha) - (90^\circ - \beta) = \alpha + \beta\end{aligned}$$

The Cobb angle is the sum of upper and lower end vertebra tilt angles. The included angle of the upper vertebra endplate line with the horizontal line is measured on the imaging data (Fig. 2d), and the included angle of the lower vertebra endplate line with the horizontal line is measured on the imaging data (Fig. 2e). And then, the

sum of two measured included angles is the Cobb angle (Fig. 2f).

Time spent on measurement by end vertebra tilt angle

PACS spinal imaging pictures of this group of patients were exported and printed in A4 paper. After the upper and lower end vertebrae and top vertebra were determined, three spine surgeons directly marked and measured the Cobb angle of the same scoliosis segment of the same patient using the classical method and tilt angle method, respectively, and then recorded the time spent on each measurement method using an electronic stopwatch and calculated the average value.

Occupied pixel space marked by the end vertebra tilt angle measurement method

The patient's whole spine anteroposterior X-ray film was scanned and made into a figure with a resolution of 654×1024 pixels. The drawing tool in the same computer was used to simulate the line markings of the classical method and the tilt angle method to measure the Cobb angle, and the stroke parameters of all drawings were kept the same. The line markings for the same Cobb angle should meet the following requirements to facilitate quantitative comparison: (1) The EVEL line ends with both lateral margins of the vertebra endplate (Fig. 3a). (2) Two EVEL vertical lines of the classical method are intersected into an angle, and one fifth of the EVEL length is extended from the intersection point (Fig. 3b). (3) The starting point of the horizontal line of the tilt angle method is intersected with the starting point of the EVEL line, and the vertical line drawn at ending point could be exactly intersected with the ending point of the EVEL line (Fig. 3c). Figures marked with a line were imported, and the line markings of the two measurement methods were compared with Beyond Compare for Mac software. The occupied pixel space value and the difference value of the two groups were automatically calculated. Then, the visual interference degree of line markings of the two groups in the whole imaging data was compared (Fig. 3d–g).

Statistical processing

Analysis was performed using SPSS 23.0 for Mac statistical software; data were expressed in mean value \pm standard deviation. Complete random one-way ANOVA is used to test the results of the angle of the three methods whether there is any error. We select *t* test to compare whether there is a difference between the tilt angle method and the classical method in measuring time. To compare the measuring error between the tilt angle method and the classical method, whether there is statistical difference, we chose to use the Mann-Whitney




Fig. 2 a, b, c Measuring process of the Cobb angle by the end vertebra tilt angle method. **d** Measurement of the upper end vertebra tilt angle. **e** Measurement of the lower end vertebra tilt angle. **f** The upper end vertebra tilt angle $\alpha = 41^\circ$, the lower end vertebra tilt angle $\beta = 30^\circ$, and the Cobb angle $= \alpha + \beta = 71^\circ$


U test. $p < 0.05$ means there is statistical significance for the difference.

Results

The PACS measuring result regarded as the reference standard [2] was rounded to an integer to conform to an actual clinical application. To compare the results of the angle of the three methods (PACS method, end vertebra tilt angle method, classical method), whether there is any error, a complete random one-way ANOVA is used to test.

In the test, we also need pairwise comparison; we choose the Bonferroni method and homogeneity test of variance. The following results were obtained:

1. In the homogeneity test of variance, significant $p > 0.1$ was obtained, which indicated that the same variance could be used for single-factor ANOVA.
2. Adjust $\alpha' = \alpha/m = 2 \times 0.05/3 \times (3 - 1) = 0.167$
3. Result of the PACS method: The range of Cobb angles for 50 cases was 25~125°, the median was 60°, the average value was $60.96 \pm 21.08^\circ$, and 95% confidence interval was 54.9°, 66.7°. The tilt angle method result average was $61.34 \pm 21.24^\circ$, and 95% confidence interval was 55.4°, 67.5°. The classical method result average was $61.90 \pm 21.34^\circ$ and 95% confidence interval was 55.9°, 67.9°. The results of single-factor analysis of variance showed $F = 0.033$,



$p = 0.967$, indicating that there was no statistical difference between the three methods.

- The Bonferroni method was used for pairwise comparison of the three methods ($p > 0.05$), and there was no statistical difference between the three methods, indicating that there was no difference in measuring the angle.

To compare whether there is a difference between the tilt angle method and the classical method in measuring time, we select t test and obtained the following results:

- Correlation coefficient was 0.284 ($p < 0.05$).
- We got $p < 0.05$ through t test. The average time spent for the tilt angle method was 12.98 ± 2.14 s with 95% confidence intervals of 12.37 s, 13.59 s.

The average time spent for the classical method was 18.96 ± 2.65 s with 95% confidence intervals of 18.20s, 19.71 s, which indicated that the measuring time of the two methods was different.

- We got the mean difference of the measuring angle time between the tilt angle method and the classical method which is -5.98 , indicating that the time required to measure the angle by the tilt angle method is faster.

To compare the measuring error between the tilt angle method and classical method, whether there is statistical difference, we chose to use the Mann-Whitney U test. The measuring error range was -15° to approximately $+6^\circ$ through the classical method with an average error of $\pm 3.67^\circ$ and -9° to approximately $+5^\circ$ through the tilt

angle method with an average error of $\pm 3.19^\circ$, which showed no statistical significance for the difference ($Z = -0.430$; $p = 0.667$) (Fig. 4).


The software analysis of the pixel difference performed on picture results showed that the average pixel marked by the line drawing of the classical method was 3680 ± 533 pixels, accounting for 0.46~1.13% of total pixels, while the average pixel marked by the line drawing of the tilt angle method was 1539 ± 320 pixels, accounting for 0.12~0.32% of total pixels. Therefore, we can conclude that the pixel marked by the line drawing of the tilt angle method is less than that of the classical method, so the imaging data is less polluted.

Discussion

Scoliosis is a three-dimensional deformity of the spine. No matter how complicated the scoliosis is, the measurement of the Cobb angle is based on the coronal or sagittal plane of imaging [3]. The Cobb angle is closely related to the spinous process angle of the coronal plane and rotation of the apical vertebra [4]. As for the bigger bending deformity of the spine coronal plane, the Cobb angle is the included angle of the upper end vertebra endplate line directly intersected with the lower end vertebra endplate line. For the smaller deformity of the spine, the intersected point of two endplate lines is outside of the X-ray film, so the vertical line of the upper end vertebra endplate line and that of the lower end vertebra endplate line shall be drawn to perform the measurement. In recent years, new measurement methods have been reported, such as the smartphone software [5, 6], PACS, and other computer software, and those methods are reliable and convenient and can replace

the classical method to measure the Cobb angle [7–9]. In modern medical healthcare systems with digital radiographs and analyses, the idea of reducing drawing artifacts on an X-ray film is somewhat redundant. In developing countries, such as China, that still analyze radiographs on conventional X-ray films, steps for classical Cobb angle measurement are as follows: (1) Draw an endplate line between the two intersections of the end vertebra endplate and lateral margins on the film or a straight line drawn between the upper tangent of pedicles' eyes in the same vertebra. (2) Measure the rectangle angle of the upper endplate line to draw the vertical line, and measure the rectangle angle of the lower endplate line to draw the vertical line. (3) Measure the included angle between two vertical lines (Cobb angle). The classical Cobb method needs a line drawn in a large range, and this will easily contaminate imaging data. In addition, limited by conditions of radiology departments of different hospitals and imaging film size, it is hard to include the whole spine segments into one film, and films shall be taken segment by segment. Thus, the measurement of the Cobb angle shall be performed by manually splicing the films into one figure, so there are inconveniences and figure angle deviations.

According to the geometry law, it can be deduced that the Cobb angle is the sum of upper and lower end vertebra tilt angles, so the Cobb angle can be calculated by measuring end vertebra tilt angles. No matter how serious the curvature of scoliosis is, and whether the scoliosis segments are in one imaging film, the Cobb angle can be calculated accurately and rapidly just by determining the two end vertebrae and measuring the tilt angles. The measuring steps of the tilt angle method are as



follows: (1) Draw the upper and lower end vertebra endplate connecting line on the film. (2) Measure the tilt angles of the upper and lower endplates. (3) Add the two measured results to get the Cobb angle. Obviously, the tilt method reduces one measurement step, so it can reduce the measuring time. In this study, the average time spent when using tilt angle to measure an angle is about 6 s less than that of the classical method. If you are skilled in the method, you can utilize the rectangular structure of a measuring ruler to fast determine the horizontal line and measure the end vertebra tilt angle in combination with the straight edge of the figure on the imaging film, which is more fast and convenient than the classical method in which the two vertical lines shall be additionally drawn for measurement. If the endplate connecting line develops clearly, the marker line of a measuring ruler can be directly utilized to perform overlap measurement, which can be free from the line drawing step.

When judging the interference and containment degree of line marking in imaging data, the figure treatment and analysis can be used to compare the difference of pixels marked by lines, which is more precise than visual observation and judgment [10]; the interference in pixels marked by the tilt angle is only 23.9~28.3% of that by the classical method, greatly lowering the contamination of line on imaging data.

A previous study has suggested that the Cobb measurement method has several sources of errors [11]: non-standard position of patients or/and devices in imaging examination. To confirm the correct marker lines in the scoliosis segments which have anatomical variation of the vertebrae, different observers identify the different upper and lower end vertebrae. For those reasons, the measuring error range for the classical Cobb method was 6~9° [11, 12]. The tilt method is a methodological improvement based on the Cobb method, which has the same measuring errors as the former one. The efficacy and effectiveness of the tilt

method were observed and compared by the same observers via using the same medical images. Therefore, the most common errors are intrinsic to the measurement method. The tilt method needs to draw two horizontal lines on the X-ray film. It is hard to make an accurate judgment about the reference points associated with the horizontal line of the actual torso. In addition, when the film is placed on the table or on the radiographic view box, the horizontal plane judgment will deviate from the real plane when the film is tilted, and the horizontal line is not the same as the horizontal of the actual torso. It is easy to make measuring error. But we found through computer simulation measurements that even though the film was tilted or the real horizontal plane was difficult to determine, there was no obvious measuring error. As shown, it is exactly the same case of scoliosis X-ray image data (Fig. 5a, b). We tilted the film to simulate the actual film placed on the table or on the film viewing illuminator, so the drawing line may deviate from the horizontal line. The green line is a horizontal line based on the entire imaging data, and it was given by the computer automatically. The red line was respectively at the upper/lower end vertebra endplate line. The angle between the red and green lines is an end vertebra tilt angle. The angle is completely consistent according to the geometry law and the actual observation (Fig. 5a, b).

Even though there is no statistical difference between the measurement error of the tilt method and the classical method, there are still some procedures to avoid the measurement error as far as possible. For instance, the shape of the imaging film is a rectangle, and the ruler line on the film is standard vertical or horizontal; it was given by the computer. We can use it as a reference point (Fig. 5c). On the other hand, we use the ruler as a measuring tool. The shape of the ruler is a rectangular structure, so we can make full use of the rectangular structure of the ruler and the rectangular outline of the imaging film as the reference point of the horizontal




Fig. 5 **a** Normally placed film. **b** Tilted placed film. **c** Red dotted line: the rectangular structure of the imaging data itself, and the ruler line (arrow) is vertical to the real horizontal plane

line. For example, the wide edge of the ruler overlaps with the edge of the film, and the line drawn on the long edge of the ruler must be the true horizontal line of the imaging data.

Cobb angle $>10^\circ$ means that scoliosis exists, $10\sim 25^\circ$ means regular recheck shall be performed, and $25\sim 45^\circ$ means orthosis shall be needed. Cobb angle $>45^\circ$ means surgical interference is needed. Cobb angle $>5^\circ$ in two X-ray examinations indicates the scoliosis deformity progress [13]. Therefore, the measuring error for Cobb angle $>5^\circ$ will possibly interfere with the diagnosis and treatment results. There is always a difference existing in the measurement of the Cobb angle of the same patient, and it is related to the patient position and photography angle. The manual line drawing and artificial observation are still the main reasons for measuring error. This study took the Cobb angle measured by PACS as the reference standard; the measuring error range for the classical method was $-15\sim 6^\circ$ with an average error of $\pm 3.67^\circ$, and the measuring error range for the tilt angle method was $-9\sim 5^\circ$ with the average error of $\pm 3.19^\circ$. The first reason for error is that the cases with complicated scoliosis we included were less; the second reason is that the imaging data included in the study were made into pictures of the same resolution and size in advance and printed and then UEV and LEV were determined uniformly for measurement and comparison [14], which could statistically reduce the measurer's judgment bias and inter-group error.

Conclusions

The scoliosis Cobb angle can be measured accurately and rapidly by the principle of the Cobb angle being equal to the sum of tilt angles of upper and lower end vertebra, wherein the film data of imaging will not be easily contaminated. Under special conditions, the average measuring error is $\pm 3^\circ$.

Abbreviations

EVEL: End vertebra endplate lines; EVTA: End vertebra tilt angle method; LEV: Lower end vertebra; LEVEL: Lower end vertebra endplate lines; LEVTA: Lower end vertebra tilt angles; PACS: Picture archiving and communication systems; UEV: Upper end vertebra; UEVEL: Upper end vertebra endplate lines; UEVTA: Upper end vertebra tilt angles

Availability of data and materials

All data were true and effective, and all patients were hospitalized in our hospital. Supported by the Technology Innovation Project of Cuiying, Lanzhou University Second Hospital (CY2017-MS06).

Authors' contributions

JW and JZ contributed in the conception and design, analysis and interpretation, and writing of the manuscript. RX, TC, and KZ contributed to the data collection and material support. HZ contributed to the critical revision of the manuscript and supervision. All authors read and approved the final manuscript.

Ethics approval and consent to participate

"Ethical approval was obtained by Lanzhou University Second Hospital Medical Ethics Committee Review Form. Approval reference if relevant." We had all necessary consent from any patients involved in the study, including consent to participate in the study where appropriate.

Consent for publication

Not applicable

Competing interests

The authors declare that they have no competing interests.

Publisher's Note

Springer Nature remains neutral with regard to jurisdictional claims in published maps and institutional affiliations.

Author details

¹Department Of Orthopedics, Orthopedics Key Laboratory of Gansu Province, Lanzhou University Second Hospital, Cuiyingmen, Lanzhou 730030, Gansu province, China. ²Radiology Department, Lanzhou University Second Hospital, Cuiyingmen, Lanzhou 730030, Gansu province, China.

Received: 29 November 2017 Accepted: 22 August 2018

Published online: 04 September 2018

References

- Cobb J. Outline for the study of scoliosis. *Instr Course Lect.* 1947;5:261–75.
- He J, Yan Z, Zhi Y, et al. The study of the reliability in Cobb measurement on PACS workstation. *Zhongguo ji zhu ji sui za zhi.* 2006;16(10):732–4.
- Sangole AP, Aubin CE, Labelle H, et al. Three-dimensional classification of thoracic scoliotic curves. *Spine.* 2009;34(1):91–9.
- Morrison DG, Chan A, Hill D, et al. Correlation between Cobb angle, spinous process angle (SPA) and apical vertebrae rotation (AVR) on posteroanterior radiographs in adolescent idiopathic scoliosis (AIS). *Euro Spine J.* 2014;24(2):306–12.
- Qiao J, Liu Z, Xu L, et al. Reliability analysis of a smartphone-aided measurement method for the Cobb angle of scoliosis. *J Spinal Disord Tech.* 2012;25(4):88–92.
- Shaw M, Adam CJ, Izatt MT, et al. Use of the iPhone for Cobb angle measurement in scoliosis. *Euro Spine J.* 2012;21(6):1062–8.
- Wang C, Lao L, Niu W, et al. Computer-assisted versus traditional Cobb angle measurement on digital radiograph in scoliosis. *Zhonghua Yi Xue Za Zhi.* 2010;90(19):1300–3.
- Aubin CE, Bellefleu C, Joncas J, et al. Reliability and accuracy analysis of a new semiautomatic radiographic measurement software in adult scoliosis. *Spine.* 2011;36(12):780–90.
- Zhang J, Lou E, Le LH, et al. Automatic Cobb measurement of scoliosis based on fuzzy Hough transform with vertebral shape prior. *J Digit Imaging.* 2009;22(5):463–72.
- Thakkar SC, Mears SC. Visibility of surgical site marking: a prospective randomized trial of two skin preparation solutions. *J Bone Joint Surg Am.* 2012;94(2):97–102.
- Capasso G, Maffulli N, Testa V. The validity and reliability of measurements in spinal deformities: a critical appraisal. *Acta Orthop Belg.* 1992;58(2):126–35.
- Dickson RA, Lawton JO, Archer IA, et al. The pathogenesis of idiopathic scoliosis. Biplanar spinal asymmetry. *J Bone Joint Surg Br.* 1984;66(1):8–15.
- Lonstein JE. Adolescent idiopathic scoliosis. *Lancet.* 1994;344(8934):1407–12.
- Morrissy RT, Goldsmith GS, Hall EC, et al. Measurement of the Cobb angle on radiographs of patients who have scoliosis. Evaluation of intrinsic error. *J Bone Joint Surg.* 1990;72(3):320–7.

Ready to submit your research? Choose BMC and benefit from:

- fast, convenient online submission
- thorough peer review by experienced researchers in your field
- rapid publication on acceptance
- support for research data, including large and complex data types
- gold Open Access which fosters wider collaboration and increased citations
- maximum visibility for your research: over 100M website views per year

At BMC, research is always in progress.

Learn more biomedcentral.com/submissions

