

Mechanical properties of polypropylene mesh used in pelvic floor repair

J. S. Afonso · P. A. L. S. Martins · M. J. B. C. Girao ·
R. M. Natal Jorge · A. J. M. Ferreira · T. Mascarenhas ·
A. A. Fernandes · J. Bernardes · E. C. Baracat ·
G. Rodrigues de Lima · B. Patricio

Published online: 4 December 2007
© International Urogynecology Journal 2007

Erratum to: Int Urogynecol J
DOI 10.1007/s00192-007-0446-1

In the section “Materials and methods” (second paragraph, fifth sentence), the displacement rate was mistakenly given in millimeters per second instead of millimeters per minute. The correct sentence is:

“All the samples were tested with a constant displacement rate of 5 mm/min in a mechanical testing machine specially designed for low-load soft-tissue testing [12].”

The online version of the original article can be found at <http://dx.doi.org/10.1007/s00192-007-0446-1>.

J. S. Afonso (✉) · M. J. B. C. Girao · E. C. Baracat ·
G. Rodrigues de Lima
Department of Gynecology, Federal University of São Paulo,
São Paulo, Brazil
e-mail: afonso@histeroscopia.med.br

P. A. L. S. Martins · R. M. N. Jorge · A. A. Fernandes
Institute of Mechanical Engineering (IDMEC),
University of Porto,
Porto, Portugal

A. J. M. Ferreira
Department of Mechanical Engineering and
Industrial Management (DEMEGI),
University of Porto,
Porto, Portugal

T. Mascarenhas · J. Bernardes · B. Patricio
Department of Gynecology, Faculty of Medicine,
University of Porto,
Porto, Portugal