

RESEARCH

Open Access



Patient-specific neurosurgical phantom: assessment of visual quality, accuracy, and scaling effects

Felipe Wilker Grillo^{1*†}, Victor Hugo Souza^{1†}, Renan Hiroshi Matsuda¹, Carlo Rondinoni¹, Theo Zeferino Pavan¹, Oswaldo Baffa¹, Helio Rubens Machado² and Antonio Adilton Oliveira Carneiro¹

Abstract

Background: Training in medical education depends on the availability of standardized materials that can reliably mimic the human anatomy and physiology. One alternative to using cadavers or animal bodies is to employ phantoms or mimicking devices. Styrene-ethylene/butylene-styrene (SEBS) gels are biologically inert and present tunable properties, including mechanical properties that resemble the soft tissue. Therefore, SEBS is an alternative to develop a patient-specific phantom, that provides real visual and morphological experience during simulation-based neurosurgical training.

Results: A 3D model was reconstructed and printed based on patient-specific magnetic resonance images. The fused deposition of polyactic acid (PLA) filament and selective laser sintering of polyamid were used for 3D printing. Silicone and SEBS materials were employed to mimic soft tissues. A neuronavigation protocol was performed on the 3D-printed models scaled to three different sizes, 100%, 50%, and 25% of the original dimensions. A neurosurgery team (17 individuals) evaluated the phantom realism as “very good” and “perfect” in 49% and 31% of the cases, respectively, and rated phantom utility as “very good” and “perfect” in 61% and 32% of the cases, respectively. Models in original size (100%) and scaled to 50% provided a quantitative and realistic visual analysis of the patient’s cortical anatomy without distortion. However, reduction to one quarter of the original size (25%) hindered visualization of surface details and identification of anatomical landmarks.

Conclusions: A patient-specific phantom was developed with anatomically and spatially accurate shapes, that can be used as an alternative for surgical planning. Printed models scaled to sizes that avoided quality loss might save time and reduce medical training costs.

Keywords: Neurosurgery, Education, Medical training, Patient-specific, Neuronavigation, Phantom, 3D printing, Simulator

Background

Medical error can be defined as acts of commission or omission that have the potential to harm or that effectively harm patients [1]. A study has revealed that technical errors, such as problems in equipment use or in the performance of a procedure, cause 27.8% of these events [2, 3]. The image-guided navigation (IGN) systems have been employed to assist surgeons during

complex surgical procedures [4] that require extreme manual and visual abilities. IGN allows real-time surgical tool localization through co-registration of the patient’s body and tomographic images, such as computed tomography (CT), magnetic resonance imaging (MRI), and ultrasound imaging (US) [5]. Neuronavigation systems are IGN systems dedicated to help surgeons during neurosurgeries to locate a tool in a three dimensional (3D) space while moving it around or inside the patient’s head [6]. Mastering the use of navigation systems requires extensive training in a controlled environment to allow accurate and precise measurements.

* Correspondence: fvgrillo@usp.br

†Equal contributors

¹Department of Physics, Faculty of Philosophy, Science and Letters at Ribeirao Preto, University of Sao Paulo, Av. Bandeirantes, 3900, Monte Alegre, Ribeirão Preto, SP CEP 14040-901, Brazil

Full list of author information is available at the end of the article

Medical training is another method to minimize medical errors and to make procedures more accurate. Physicians train to develop technical skills and to achieve expertise [7, 8]; that is, physicians practice in order to acquire the ability to reproduce a given task with superior performance [9]. The classical approach to anatomy and surgery teaching and learning is based on the use of human cadavers [10] and animals. However, this strategy is becoming less common due to the high cost and ethical and/or logistic issues involved in collecting cadavers and in maintaining live animals prior to sacrifice. The number of medical training tools has increased significantly in recent years. Learning centers are constantly improving simulation environments to help students and professionals to reach expertise [11–13]. Computational simulation [11, 13–15] and phantoms [16] are alternative tools for training purposes.

Advances in material technology and computer sciences have provided inexpensive alternatives for medical training. Polymers that mimic the elastic and haptic properties of human tissues are examples of such alternatives. The tissue-mimicking materials (TMM) employed in phantoms can reproduce specific physical, chemical, and/or biological properties of the human anatomy. Indeed, TMM can mimic structures that range from homogeneous tissues to complex structures [17, 18]. Several TMM have been reported in the literature [17, 19–22]. Agar/gelatin is a water-based material that is easy to prepare, but it must be protected from bacterial attacks and dehydration. On the other hand, oil-based materials, like styrene-ethylene/butylene-styrene (SEBS) gels, are biologically inert and present tunable properties [20]. The mechanical properties and touch feeling of oil-based materials and biological soft tissues are similar, so the former are adequate for ultrasound applications [21, 23].

The 3D printing in neurosurgery training allows reproduction of the morphology and structural features of specific patient cases by means of segmented CT or MRI images. Even though this technique is versatile, it is not free of limitations, which include the physical properties of the printing material, the cost, and the relation between the manufacturing time and the quality of the printed structure [24]. Involvement of professionals of different medical, technological, and artistic areas is essential to integrate the different devices, software, and protocols that simulate the neurosurgical environment and lead to viable outcomes [25, 26]. Together, these techniques should improve surgical planning, minimize errors, and increase the surgeon's confidence. Exhaustive training is the best method to achieve expertise and to acquire appropriate skills [7, 13].

The aim of this study was to develop a patient-specific phantom for neurosurgery to reproduce brain cortical

morphology, cerebrospinal fluid, meninges and scalp. The model was obtained by 3D printing and SEBS modeling. The quality assessment of the hybrid model was made by inspecting fused images (real and phantom) and by applying a questionnaire to health specialists. Additionally, we assessed the effect of reducing the size of a segmented model in the measurements performed with a neuronavigation system.

Methods


Patient-specific phantom

The MRI scan of a three-year-old boy diagnosed with Sturge-Weber Syndrome [27] was acquired using a 3DT1 gradient echo sequence (TR = 4.8 ms; TE = 3.4 ms; acquisition matrix = $129 \times 164 \times 169$ mm; slice thickness = 1 mm; pixel size = 1×1 mm) in a scanner Achieva 3 T (Philips, The Netherlands). The MRI scan was segmented to reveal the cortical surface and a representative portion of the patient's face. A 3D model was reconstructed and exported to stereolithographic file (.STL) with the aid of the InVesalius software (Centro de Tecnologia da Informação Renato Archer, Campinas, Brazil) [28]. By using Blender (Blender Institute, Amsterdam, the Netherlands), the STL model was separated into three pieces: skull with facial information and left and right brain hemispheres [29]. The local ethics committee approved the experimental procedure and the patient's guardian provided a signed informed consent (CAAE 36460914.4.0000.5440).

The left and right brain hemispheres were printed (PLA fused filament) separately by using a Zmorph 2.0 S (Zmorph LLC, Wroclaw, Poland) printer with a z-layer: 0.1 mm, path width 0.4 mm. The 3D prints were used to make negative molds with a commercial white silicone rubber (Polglass, Ribeirão Preto, Brazil). The molds were filled with molten SEBS gel (Kraton Polymers, Houston, EUA), which was manufactured by mixing the SEBS copolymer with mineral oil at a concentration of 10% w/w; as described by Cabrelli et al. [23]. Silicone 3% (w/w) (Silaex, São Paulo, Brazil) and aniline were added to improve the mechanical properties and coloring, respectively.

To evaluate the Young's modulus of the gel, a TA.XT plus Texture Analyzer (Stable Micro Systems, Surrey, UK) equipped with a plate with diameter of 40 mm was used to conduct mechanical tests on a cylindrical sample with diameter of 2.5 cm and thickness of 2 cm.

The brain phantom was covered with a latex balloon filled with 10% w/w gelatin solution 250 Bloom (Gelita, Eberbach, Germany) to mimic the *dura mater* and the cerebrospinal fluid. This structure was inserted into the PLA 3D-printed patient's skull and skin. To verify the locations of each structure inside the final model, a CT scan was recorded on a Brilliance Big Bore scanner (Philips Medical Systems, Cleveland, USA) operating at



120 kVp; the slice thickness was 1 mm. The patient-specific phantom CT images were registered and fused with native patient's T1-weighted anatomical MRI by using fiducial registration in a 3D Slicer (Slicer, Cambridge, USA) [30]. Figure 1 shows the flowchart of all the steps used to build the patient-specific model.

The quality of the resulting anatomy was assessed by visually inspecting the fused images and by comparing the patient-specific phantom images with the real patient's images. An experienced neurosurgeon (HRM) traced the path for both the skin cut and the craniotomy on the right side of the skull. Next, a window was opened by drilling the indicated path to expose the insula region. Finally, a board of 17 individuals (including 2 neurosurgeons, 6 neurologists, 2 radiologists, 1 pediatric and 6 health professionals) evaluated the quality of the resulting phantom by answering a standard questionnaire. All individuals were instructed to compare with their previous experiences how much the phantom simulate each item, from "nothing" to "perfect".

Multiscale models

Three different sizes of 3D-printed models based on a real surgical case were produced to compare the performance during neuronavigation and the cost-benefit ratio regarding time and printing quality. The first

model (real-size model, designated B100) was printed in a Paragon device (Rapid Technologies, Middleton St George, UK) by using selective laser sintering; the total printing time was 24 h. Then, a model with half the original size (designated B50) was printed by PLA extrusion in a ZMorph 2.0 S printer (z-layer: 0.1 mm, path width 0.4 mm); the total printing time was approximately 10 h. Finally, a model with one quarter of the original size (designated B25) was printed by PLA extrusion in a Prusa i3 printer (Rep Rap, China) with same presets.

Neuronavigation protocol

The printed models and neuronavigation were used to assess the effect of scale reduction on the neuroanatomical localization. The navigation was performed by using the open source software InVesalius Navigator [31] installed on a personal computer and connected to the optical tracking device MicronTracker Sx60 (ClaroNav, Toronto, Canada). The images were registered by point-based registration and linearly adjusted to the corresponding scale factor. The fiducial registration error (FRE) was calculated for each procedure to control the navigation quality. The coordinates of nine anatomical landmarks of clinical relevance were digitized during neuronavigation of each model: left and right ear (LE and RE, respectively), LE and nasion (N), RE and N, left

and right post-central gyrus (LPG and RPG, respectively), Brodmann Area 44 (BA44), supramarginal gyrus (SMG), and left and right lateral-occipital gyrus (LOG and ROG, respectively). Measurements were performed by three independent raters in three distinct acquisition series each. Raters were physicists with at least four years of experience in neuronavigation and visual analysis of anatomical characteristics of MRI slices.

Statistical analysis

The Euclidian distances between the six pairs of digitized landmarks were calculated (LE-N, LE-RE, RE-N, LPG-RPG, SMG-BA44, LOG-ROG). Two-way analysis of variance (ANOVA) was used to test whether the measured distance depended on both the scale factor and the target pair of anatomical landmarks. Post hoc Tukey's test was employed to compare all the groups. Statistical significance was set to 5%.

Results

Patient-specific phantom

The molten SEBS gel used to obtain the morphological brain phantom has Young's modulus 32 ± 1 kPa. Figure 2a shows an intermediate layer of the printing plan, Fig. 2b depicts the final printing plan and Fig. 2c illustrates the printed head. We took the CT images of the final model and registered them with the patient's MRI. Figure 3 illustrates the fiducial registration of

the patient-specific phantom CT images and the patient's MRI.

Figure 4a shows the model after the craniotomy performed according to the neurosurgeon's indication. In addition, we can see the structure mimicking the *dura mater* (Fig. 4b), the insula exposure (Fig. 4c) and the patient-specific phantom after the craniotomy simulation (Fig. 4d).

Table 1 lists the questionnaire results. The neurosurgeons team evaluated the phantom realism as "very good" and "perfect" in 49% and 31% of the cases, respectively. The neurosurgeons rated the phantom as "very good" and "perfect" in 61% and 32% of the cases, respectively. It is relevant to observe that, the grade "perfect" is an individual perception considering the purpose of the models. Therefore, it might not reflect that the model has the same properties as cadaveric models or real human tissues.

Multiscale models

Figure 5 shows all the models we used during the neuro-navigation protocol. Inspection of the model B50, printed with PLA extrusion, and of the model B100, manufactured by selective laser sintering, provided a quantitative and realistic visual analysis of the patient's cortical anatomy without any distortion due to scale reduction. However, the model B25, which corresponded to one quarter of the original size, had reduced surface




Fig. 2 a 3D digital planning for printing showing an intermediate layer and b the final shape based on patient's data. c Head phantom printed with PLA extrusion




Fig. 3 a Patient's MRI. b Phantom CT image. c Fused images with highlighted phantom CT image contour

details, which hindered identification of the anatomical landmarks. Reduced model size drastically decreased the printing time, which was 3, 10, and 22 h for B25, B50, and B100, respectively.

The estimated FRE for neuronavigation (mean \pm standard deviation) was 2.5 ± 0.3 , 1.7 ± 0.2 , and 1.5 ± 0.3 mm for B25, B50, and B100, respectively. Two-way ANOVA revealed that the calculated distance depended on both the scale factor ($F_{2,144} = 40.56$; $p < 0.001$) and the pair of anatomical landmarks ($F_{5,144} = 55,951.70$; $p < 0.001$). Also, we found a significant interaction between the pair of fiducial coordinates and the scale factor ($F_{10,144} = 11.78$; $p < 0.001$). Multiple comparison tests revealed that the LE-RE distance was greater in B25 as compared to B50 and B100 (95% confidence interval 125.8–129.6, $p < 0.001$). The difference between the LE-RE distances in

B25 and B50 and in B25 and B100 was 2.6 and 2.7 mm, respectively. The RE-N distance was greater for B25 than for B50 and B100 (95% confidence interval 124.0–127.8, $p < 0.001$); the mean difference between the RE-N distances in B25 and B50 and in B25 and B100 was 5.8 and 5.1 mm, respectively. B25, B50, and B100 did not differ significantly in terms of the remaining pairs of anatomical landmarks. In general, the measured distances tended to display higher deviations in the smallest model as compared to B50 and B100.

Discussion

In an attempt to minimize medical errors, the simulation of medical procedures has become a common approach in medical training. In this scenario, simulating the specific characteristics of a patient constitutes a challenge





Fig. 4 Visual aspect of the mounted patient-specific model. a Lateral view after craniotomy; b initial cut of the rubber-mimicked meninges; c exposed surface of the brain tissue and cerebrospinal fluid mimicking; d overview of the resulting patient-specific realistic phantom

Table 1 Results of patient specific phantom assessment

Assessment of phantom realism					
	Nothing	Not Good	Good	Very Good	Perfect
Phantom general aspect (anatomical structures: proportions and locations)	0%	0%	24%	47%	29%
Haptic response	0%	0%	18%	59%	24%
Size of the brain internal structures	0%	0%	18%	53%	29%
Brain tissue appearance	0%	0%	18%	53%	29%
Relationship between internal anatomical structures used as reference for access	0%	0%	24%	35%	41%
Realism average	0%	0%	20%	49%	31%
Assessment of phantom educational potential (utility)					
	Nothing	Not Good	Good	Very Good	Perfect
To acquire basic skills necessary for surgery	0%	0%	6%	65%	29%
To acquire depth sensation (insertion / extraction) through the bony window	0%	6%	6%	59%	29%
To learn how to orient yourself during the surgical procedure	0%	0%	6%	59%	35%
To learn procedures	0%	0%	6%	65%	29%
To learn how to position drains and other equipment	0%	0%	6%	59%	35%
Utility Average	0%	1%	6%	61%	32%
Overall Average	0%	1%	13%	55%	31%



[9]. The use of negative molds is a well-known method to reproduce morphology [32, 33]. However, it is impossible to copy internal structures or organs for in-vivo studies of human cases. The use of 3D printers enables the development of realistic phantoms with anatomically and spatially accurate shapes [34]. Previous studies have reported the use of 3D printing to reproduce the patient’s anatomy for surgical and treatment planning and to facilitate understanding of the normal and pathologic anatomy of individual patients in different situations, such as aneurysm [35, 36]. Weinstock et al. used 3D printing and silicone molding and relied on the aid of Hollywood special effects technicians, who finished the model with makeup, to improve realism [37]. Therefore, the combination of 3D printing and molding technique helps to accelerate object manufacture [37]. Compared to 3D printing, in this study the silicone molds enabled faster production of multiple copies with the same details and even allowed the preparation of materials that are not yet available for 3D printing. We also verified that the soft tissue-mimicking material SEBS bore a striking resemblance to the patient’s brain cortex anatomy and enabled haptic feedback for surgical procedures. Further development should allow mimicking of internal brain structures and dynamic functions, such as blood circulation.

External details of the PLA printed face provided anatomical landmarks that may help surgeons to plan the surgery, thereby improving realistic training. We planned the craniotomy on the basis of the traditional pterional

craniotomy approach. Figures 4b and c show that the virtual planning and the final phantom were very similar. To improve the total time, we used a printing layer of 0.1 mm. If necessary, layers of 0.03 mm might be used to achieve a better level of details. Virtual printing planning is not always identical to the final printed piece [38]. Printing errors may cause deformations, overlaps, or missing areas as compared to the planning file [24]. Due to these errors, the layer height and other parameters, such as infill ratio, printing speed, and support settings should be considered for each case, and the printed object must be analyzed before it is used. In this study, another important observation concerned the dimensions of the object to be printed; the patient-specific model presented herein almost reached the limit of the 3D printer (Zmorph - printing area 300 × 235 × 165 mm). Therefore, manufacturing models of older or bigger patients would demand that pieces subdivided for printing and then used to mount the final phantom.

Fusion of the phantom CT images and of the native T1-weighted MRI scans revealed a brain shift in the patient-specific phantom as compared to the real patient's data. Figure 3b highlights the difference between the patient's MRI scan (Fig. 3a) and the phantom CT image (Fig. 3c). Future development of the procedure presented here should improve the methods for brain positioning inside the skull. Fixed guiding points could improve brain mounting by diminishing the shift in relation to the original anatomy. Another possibility would be to use neuronavigation to refine orientation during model construction.

Patient-specific phantom assessment revealed a didactic potential for medical training, mainly for unusual diagnosis. The insertion of individual characteristics during training for a specific surgery might minimize serious errors [1] like operating on the wrong side [39] or on the wrong patient. Moreover, phantoms allow for exhaustive training by repetition of a specific procedure step anywhere and anytime. Not only surgeons but the entire medical team can practice with patient-specific phantoms, thereby increasing the total training time and respective individual learning curves [40].

Multiscale models have been developed and tested for neurosurgery training with neuronavigation. The InVesalius Navigator software enables real-time localization and digitization of anatomical structures with fiducial registration errors within the limits recommended by the specialized literature for clinical practice, i.e. below 3 mm [41]. Neuronavigation on B50 and B100 provided more accurate and precise distance measurements as compared to neuronavigation on B25. The smallest model showed higher deviations for the measurements, specifically for the LE-RE and the RE-N distances. Both pairs of anatomical landmarks had the right ear as one

of the references, which suggested that a possible systematic deviation was associated with the acquisition of the coordinates in B25. Indeed, we expected that B25 would provide lower-quality results as compared to B50 and B100 because the small scale approximated all the measurements to the uncertainty levels. Furthermore, none of the remaining references showed any difference associated with distance measurements across scale factors. A possible explanation is that visually localizing the same reference points on the smallest model is more difficult compared to the larger ones. We should emphasize that B50 had enough accuracy for simulation purposes. Reducing the scale of printed models without quality loss might help to reduce the protocol duration and the costs in future iterations. Thus, neuronavigation with B50 might be a suitable alternative to the real size model because it provided similar visual detailing.

It is important to note that, direct comparison between production time and measurement might be limited because the three scaled models were produced using different devices and materials. Nevertheless, it is expected that production time is shorter for smaller models, and different materials might have a small impact in measurements during neuronavigation as the resolution from the printing devices are significantly greater than from the measurement device.

Conclusion

A patient-specific (Sturge Weber case) phantom was successfully created by using 3D printing techniques and molds that included a soft tissue mimicking material for the brain structure, meninx and skull. Our study revealed that 3D printing and SEBS are promising tools to develop a patient-specific phantom as a teaching tool, providing tissue mimicking material and reliable anatomical morphology. Moreover, the molding technique enabled the performance of fast copies for a single case, which confirmed the didactic potential of the presented models. The use of multiscale models is a successful alternative to improve time and to simulate the general steps of surgical procedures.

Abbreviations

3D: Three dimensional; BA44: Brodmann Area 44; CT: Computed tomography; FRE: Fiducial registration error; IGN: Image-guided navigation; LE: Left ear; LOG: Left lateral-occipital gyrus; LPG: Left post-central gyrus; MRI: Magnetic resonance imaging; N: Nasion; PLA: Polyactic acid; RE: Right ear; ROG: Right lateral-occipital gyrus; RPG: Right post-central gyrus; SEBS: Styrene-ethylene/butylene-styrene; SMG: Supramarginal gyrus; STL: Stereolithographic file; US: Ultrasound imaging

Acknowledgements

We thank DSc. Alexandre Colello Bruno for intellectual contribution, Agnelo Bastos, Clésio Fernandes, Carlos Renato, Lourenço Rocha and Sérgio Bueno for technical support.

Funding

This study was supported by CAPES, FAEPA, FAPESP (2014/50414–9 and 13/18854–6) and CNPq (448806/2014–2 and 140787/2014–3) and FINEP (2613/09).

Availability of data and materials

Data will be made available on request. The authors have the right to share the data used in this research and will provide it upon request.

Authors' contributions

FWG, VHS, RHM, CR and TZP, participated in the conception and design of the study, data analysis and interpretation. OB, HRM and AAOC supervised experimental execution and helped with data interpretation. All authors drafted the manuscript and approved the final version. FWG and VHS contributed equally to this work.

Authors' information

Not applicable

Competing interests

FWG is shareholder of Gphantom. All the others disclose no conflict of interest.

Publisher's Note

Springer Nature remains neutral with regard to jurisdictional claims in published maps and institutional affiliations.

Author details

¹Department of Physics, Faculty of Philosophy, Science and Letters at Ribeirão Preto, University of Sao Paulo, Av. Bandeirantes, 3900, Monte Alegre, Ribeirão Preto, SP CEP 14040-901, Brazil. ²Department of Surgery and Anatomy, Faculty of Medicine at Ribeirão Preto, University of Sao Paulo, Ribeirão Preto, Brazil.

Received: 7 December 2017 Accepted: 26 February 2018

Published online: 13 March 2018

References

- Rolston JD, Bernstein M. Errors in Neurosurgery. *Neurosurg Clin N Am*. 2015; 26(2):149–55. <https://doi.org/10.1016/j.nec.2014.11.011>.
- Oremakinde AA, Bernstein M. A reduction in errors is associated with prospectively recording them. *J Neurosurg*. 2014;121(2):297–304. <https://doi.org/10.3171/2014.5.JNS.132341>.
- Stal D, Hollier L. Prospective Error Recording in Surgery: An Analysis of 1108 Elective Neurosurgical Cases. *J Craniofac Surg*. 2007;18(5):1244. <https://doi.org/10.1097/scs.0b013e31814b2b7c>.
- Mezger U, Jendrewski C, Bartels M. Navigation in surgery. *Langenbeck's Arch Surg*. 2013;398(4):501–14. <https://doi.org/10.1007/s00423-013-1059-4>.
- Doi K. Computer-aided diagnosis in medical imaging: historical review, current status and future potential. *Comput Med Imaging Graph Off J Comput Med Imaging Soc*. 2007;31(4–5):198–211. <https://doi.org/10.1016/j.compmedimag.2007.02.002>.
- Grunert P, Darabi K, Espinosa J, Filippi R. Computer-aided navigation in neurosurgery. *Neurosurg Rev*. 2003;26(2):73–99. discussion 100–101
- Gasco J, Holbrook TJ, Patel A, et al. Neurosurgery simulation in residency training: feasibility, cost, and educational benefit. *Neurosurgery*. 2013;73(Suppl 1):39–45. <https://doi.org/10.1227/NEU.0000000000000102>.
- T. Forcht Dagi. Commentary: The Roles and Future of Simulation in Neurosurgery. *Neurosurgery*. 2013;73: S4–S5. https://doi.org/10.1093/neurosurgery/73.suppl_1.S4.
- Gélinas-Phaneuf N, Del Maestro RF. Surgical expertise in neurosurgery: integrating theory into practice. *Neurosurgery*. 2013;73(Suppl 1):30–8. <https://doi.org/10.1227/NEU.0000000000000115>.
- McLachlan JC, Blich J, Bradley P, Searle J. Teaching anatomy without cadavers. *Med Educ*. 2004;38(4):418–24. <https://doi.org/10.1046/j.1365-2923.2004.01795.x>.
- Bradley P. The history of simulation in medical education and possible future directions. *Med Educ*. 2006;40(3):254–62. <https://doi.org/10.1111/j.1365-2929.2006.02394.x>.
- McGaghie WC, Issenberg SB, Petrusa ER, Scalese RJ. A critical review of simulation-based medical education research: 2003–2009. *Med Educ*. 2010;44(1):50–63. <https://doi.org/10.1111/j.1365-2923.2009.03547.x>.
- Ziv A, Wolpe PR, Small SD, Glick S. Simulation-based medical education: an ethical imperative. *Acad Med*. 2003;78(8):783–8.
- Kunkler K. The role of medical simulation: an overview. *Int J Med Robot*. 2006;2(3):203–10. <https://doi.org/10.1002/rcs.101>.
- Schirmer CM, Elder JB, Roitberg B, Lobel DA. Virtual reality-based simulation training for ventriculostomy: an evidence-based approach. *Neurosurgery*. 2013;73(Suppl 1):66–73. <https://doi.org/10.1227/NEU.0000000000000074>.
- Müns A, Mühl C, Haase R, et al. A neurosurgical phantom-based training system with ultrasound simulation. *Acta Neurochir*. 2014;156(6):1237–43. <https://doi.org/10.1007/s00701-013-1918-3>.
- Madsen EL, Hobson MA, Shi H, Varghese T, Frank GR. Tissue-mimicking agar/gelatin materials for use in heterogeneous elastography phantoms. *Phys Med Biol*. 2005;50(23):5597–618. <https://doi.org/10.1088/0031-9155/50/23/013>.
- Cook JR, Bouchard RR, Emelianov SY. Tissue-mimicking phantoms for photoacoustic and ultrasonic imaging. *Biomed Opt Express*. 2011;2(11): 3193–206. <https://doi.org/10.1364/BOE.2.003193>.
- Madsen EL, Zagzebski JA, Frank GR. Oil-in-gelatin dispersions for use as ultrasonically tissue-mimicking materials. *Ultrasound Med Biol*. 1982;8(3): 277–87. [https://doi.org/10.1016/0301-5629\(82\)90034-5](https://doi.org/10.1016/0301-5629(82)90034-5).
- Cabrelli LC, Grillo FW, Carneiro AAO, Pavan TZ. Copolymer-in-oil tissue-mimicking material with tunable acoustic properties. *IEEE International Ultrasonics Symposium (IUS)*. 2016:1–4. <https://doi.org/10.1109/ULTSYM.2016.7728859>.
- Oudry J, Bastard C, Miette V, Willinger R, Sandrin L. Copolymer-in-oil phantom materials for elastography. *Ultrasound Med Biol*. 2009;35(7):1185–97. <https://doi.org/10.1016/j.ultrasmedbio.2009.01.012>.
- D'Souza WD, Madsen EL, Unal O, Vigen KK, Frank GR, Thomadsen BR. Tissue mimicking materials for a multi-imaging modality prostate phantom. *Med Phys*. 2001;28(4):688–700. <https://doi.org/10.1118/1.1354998>.
- Cabrelli LC, Pelissari PIBGB, Deana AM, Carneiro AAO, Pavan TZ. Stable phantom materials for ultrasound and optical imaging. *Phys Med Biol*. 2017;62(2):432–47. <https://doi.org/10.1088/1361-6560/62/2/432>.
- Choi J-Y, Choi J-H, Kim N-K, et al. Analysis of errors in medical rapid prototyping models. *Int J Oral Maxillofac Surg*. 2002;31(1):23–32. <https://doi.org/10.1054/ijom.2000.0135>.
- Cooper J, Taqueti V. A Brief history of the development of mannequin simulators for clinical education and training. *Qual Saf Health Care*. 2004; 13(Suppl 1):i11–8. <https://doi.org/10.1136/qshc.2004.009886>.
- Filho FVG, Coelho G, Cavalheiro S, Lyra M, Zymberg ST. Quality assessment of a new surgical simulator for neuroendoscopic training. *Neurosurg Focus*. 2011;30(4):E17. <https://doi.org/10.3171/2011.2.FOCUS10321>.
- Rocco CD, Tamburrini G. Sturge–Weber syndrome. *Childs Nerv Syst*. 2006; 22(8):909. <https://doi.org/10.1007/s00381-006-0143-2>.
- Amorim P, Moraes T, Silva J, InVesalius PH. An interactive rendering framework for health care support. In: *Advances in Visual Computing*. Cham: Springer; 2015. p. 45–54. https://doi.org/10.1007/978-3-319-27857-5_5.
- Tsai A, Barnewolt CE, Prahbu SP, et al. Creation and validation of a simulator for neonatal brain ultrasonography. *A Pilot Study Acad Radiol*. 2017;24(1): 76–83. <https://doi.org/10.1016/j.jacr.2016.09.007>.
- Pieper S, Halle M, Kikinis R. 3D Slicer. In: 2004 2nd IEEE International Symposium on Biomedical Imaging: Nano to Macro (IEEE Cat No. 04EX821), vol. 1; 2004. p. 632–5. <https://doi.org/10.1109/ISBI.2004.1398617>.
- Amorim P, Moraes T, Silva J, Pedrini H. InVesalius: An interactive rendering framework for health care support. In: *Bebis G, Boyle R, Parvin B, Koracin D, Pavlidis I, Ferris R, et al., editors. Adv. Vis. Comput. 11th Int. Symp. ISVC 2015, Las Vegas, NV, USA, December 14–16, 2015, Proceedings, Part I*, Cham, Switzerland: Springer International Publishing; 2015. p. 45–54. https://doi.org/10.1007/978-3-319-27857-5_5.
- Seitz H, Tille C, Irsen S, Bermes G, Sader R, Zeilhofer H-F. Rapid prototyping models for surgical planning with hard and soft tissue representation. *Int Congr Ser*. 2004;1268:567–72. <https://doi.org/10.1016/j.jics.2004.03.139>.
- Tsuji M, Noguchi N, Ihara K, Yamashita Y, Shikimori M, Goto M. Fabrication of a maxillofacial prosthesis using a computer-aided design and manufacturing system. *J Prosthodont*. 2004;13(3):179–83. <https://doi.org/10.1111/j.1532-849X.2004.04029.x>.
- Randazzo M, Pisapia JM, Singh N, Thawani JP. 3D printing in neurosurgery: a systematic review. *Surg Neurol Int*. 2016;7(Suppl 33):S801–9. <https://doi.org/10.4103/2152-7806.194059>.
- Khan IS, Kelly PD, Singer RJ. Prototyping of cerebral vasculature physical models. *Surgical Neurology International*. 2014;5:11. <https://doi.org/10.4103/2152-7806.125858>.

36. KONDO K, NEMOTO M, MASUDA H, et al. Anatomical reproducibility of a head model molded by a three-dimensional printer. *Neurol Med Chir (Tokyo)*. 2015;55(7):592–8. <https://doi.org/10.2176/nmc.oa.2014-0436>.
37. Weinstock P, Rehder R, Prabhu SP, Forbes PW, Roussin CJ, Cohen AR. Creation of a novel simulator for minimally invasive neurosurgery: fusion of 3D printing and special effects. *J Neurosurg Pediatr*. 2017:1–9. <https://doi.org/10.3171/2017.1.PEDS16568>.
38. Salmi M, Paloheimo K-S, Tuomi J, Wolff J, Mäkitie A. Accuracy of medical models made by additive manufacturing (rapid manufacturing). *J Cranio-Maxillofac Surg*. 2013;41(7):603–9. <https://doi.org/10.1016/j.jcms.2012.11.041>.
39. Cohen FL, Mendelsohn D, Bernstein M. Wrong-site craniotomy: analysis of 35 cases and systems for prevention. *J Neurosurg*. 2010;113(3):461–73. <https://doi.org/10.3171/2009.10.JNS091282>.
40. Ramsay CR, Grant AM, Wallace SA, Garthwaite PH, Monk AF, Russell IT. Assessment of the learning curve in health technologies. A systematic review. *Int J Technol Assess Health Care*. 2000;16(4):1095–108.
41. Kuehn B, Mularski S, Schoenherr S, et al. Sensor-based neuronavigation: evaluation of a large continuous patient population. *Clin Neurol Neurosurg*. 2008;110(10):1012–9. <https://doi.org/10.1016/j.clineuro.2008.06.017>.

Submit your manuscript to a SpringerOpen[®] journal and benefit from:

- ▶ Convenient online submission
- ▶ Rigorous peer review
- ▶ Open access: articles freely available online
- ▶ High visibility within the field
- ▶ Retaining the copyright to your article

Submit your next manuscript at ▶ springeropen.com
