

Sensitive surface plasmon resonance performance of cadmium sulfide quantum dots-amine functionalized graphene oxide based thin film towards dengue virus E-protein — [Source link](#)

Nur Alia Sheh Omar, Yap Wing Fen, Jaafar Abdullah, Mohd Hazani Mat Zaid ...+2 more authors





Institutions: Universiti Putra Malaysia

Published on: 01 Jun 2019 - Optics and Laser Technology (Elsevier)

Topics: Cadmium sulfide and Surface plasmon resonance

Related papers:

- [Optical and surface plasmon resonance sensing properties for chitosan/carboxyl-functionalized graphene quantum dots thin film](#)
- [Label-free optical spectroscopy for characterizing binding properties of highly sensitive nanocrystalline cellulose-graphene oxide based nanocomposite towards nickel ion.](#)
- [Exploration of surface plasmon resonance for sensing copper ion based on nanocrystalline cellulose-modified thin film.](#)
- [Incorporation of surface plasmon resonance with novel valinomycin doped chitosan-graphene oxide thin film for sensing potassium ion](#)
- [Development of an optical sensor based on surface plasmon resonance phenomenon for diagnosis of dengue virus E-protein](#)

Share this paper:    

View more about this paper here: <https://typeset.io/papers/sensitive-surface-plasmon-resonance-performance-of-cadmium-394ni84cfk>

Sensitive surface plasmon resonance performance of cadmium sulfide quantum dots-amine functionalized graphene oxide based thin film towards dengue virus E-protein

ABSTRACT

An optical sensor for the dengue virus (DENV) E-protein based on cadmium sulfide quantum dots composited with amine functionalized graphene oxide (CdS-NH₂GO) thin film was successfully developed. A specific monoclonal antibodies (IgM) were covalently attached to CdS-NH₂GO via EDC/NHS coupling to sense targeted E-proteins. The SPR sensor exhibited an excellent detection limit (0.001 nM/1 pM) with sensitivity of 5.49° nM⁻¹ for the detection of DENV E-protein. The binding affinity, as well as the performance of the Au/CdS-NH₂GO/EDC-NHS/IgM film, was successfully obtained at 486.54 nM⁻¹ in detecting DENV E-proteins. These results indicated that the Au/CdS-NH₂GO/EDC-NHS/IgM film shows high potential sensitive and stronger binding towards DENV E-protein.

Keyword: Surface plasmon resonance; Binding affinity; DENV E-protein