

ARTEFACT REPORT

Open Access



Significant artefactual noise in ^{90}Y TOF-PET imaging of low specific activity phantoms arises despite increased acquisition time

Michel Hesse and Stephan Walrand*

* Correspondence: stephan.walrand@uclouvain.be
Universite Catholique de Louvain,
Brussels, Brabant, Belgium

Abstract

Volumes of usual PET phantoms are about four to sixfold that of a human liver. In order to avoid count rate saturation and handling of very high ^{90}Y activity, reported TOF-PET phantom studies are performed using specific activities lower than those observed in liver radioembolization.

However, due to the constant random coincidence rate induced by the natural crystal radioactivity, reduction of ^{90}Y specific activity in TOF-PET imaging cannot be counterbalanced by increasing the acquisition time. As a result, most ^{90}Y phantom studies reported images noisier than those obtained in whole-body ^{18}F -FDG, and thus advised to use dedicated noise control in TOF-PET imaging post ^{90}Y liver radioembolization.

We performed acquisitions of the Jaszczak Deluxe phantom in which the hot rod insert was only partially filled with 2.6 GBq of ^{90}Y . Standard reconstruction parameters recommended by the manufacturer for whole-body ^{18}F -FDG PET were used.

Low specific activity setups, although exactly compensated by increasing the acquisition time in order to get the same number of detected true coincidences per millilitre, were impacted by significant noise. On the other hand, specific activity and acquisition time setup similar to that used in post ^{90}Y liver radioembolization provided image quality very close to that of whole-body ^{18}F -FDG.

This result clearly discards the use of low specific activity phantoms intended to TOF-PET reconstruction parameter optimization. Volume reduction of large phantoms can be achieved by vertically setting the phantoms or by adding Styrofoam inserts.

Background

The unusual 32 positron emissions per million decays and the speckled pattern observed in ^{90}Y TOF-PET liver images initiated the widespread paradigm that this modality requires dedicated noise control. This belief was consolidated by many studies using large volume phantoms involving low specific activities to avoid PET saturation resulting in doses as low as 5 Gy for some of them. In addition, the acquisition time increase was often too limited to get the same number of detected true coincidences per milliliter as in post ^{90}Y liver radioembolization.

However, due to the constant random rate originating from the L[Y]SO crystal radioactivity jointly with the low ^{90}Y positron yield, lower specific ^{90}Y activities cannot be counterbalanced by increased acquisition times [1]. Moreover, Monte Carlo

simulations [2], ex vivo autoradiography [3] and ^{166}Ho MRI [4] proved the speckled nature of the sphere distribution in the liver.

Phantom presentation

A Jaszczak Deluxe phantom was set vertically (Fig. 1). The hot rod insert (inner rods diameters 4.8, 6.4, 7.9, 9.5, 11.1, 12.7 mm) was partially filled with different solutions: first, with a ^{18}F -FDG solution corresponding to a mean liver SUV = 2 in a 300-MBq 75-kg patient study; secondly, with a 2.65-GBq ^{90}Y —300 ml DTPA water solution, giving a 18-mm height filling of the rods, which resulted in $a \approx 100$ Gy mean absorbed dose in each filled sector slices. A 20-cm-thick attenuating medium, set under the phantom, modelled the transverse attenuation observed for the fully filled phantom set in conventional horizontal position.

Two bed positions were acquired, and the times per bed position were 1.5 min for ^{18}F and 20 min for ^{90}Y , respectively. The ^{90}Y phantom was further imaged for 200 min per bed position after a 10- and 20-fold factor decaying; for this last setup, two slices were summed together to exactly compensate the activity reduction. The Philips Gemini TF64 system had a TOF-fwhm of 550 ps. Standard FDG reconstruction parameters were used for all the acquisitions, i.e. 3 iterations \times 33 subsets.




Fig. 1 Jaszczak deluxe phantom set in vertical position on a 20-cm-thick paper bloc modelling the conventional attenuation as only a part of the hot rod insert was filled with active solution in order to reach a typical clinical liver absorbed dose in selective liver ^{90}Y radioembolization

Hot rods were individually visualized in five sectors for the ^{18}F setup and for the 100 Gy \times 20 min/bed- ^{90}Y setup as well; this later setup was affected only by a little higher noise level (Fig. 2). A clear degradation of the hot rods intensities was observed for the 6.4, 7.9, and 9.5 mm rods in the 10 Gy \times 200 min/bed setup while all sectors were dramatically impacted in the 2 \times 5 Gy \times 200 min/bed despite the exact acquisition time compensation in both setups.

Conclusion

Despite a slightly higher noise level, the spatial resolution of the ^{90}Y 100 Gy setup image using standard reconstruction parameters was very close to that of the WB ^{18}F -FDG setup, explaining the recent success of TOF-PET-based EUD in dose-response prediction [5].



Although time acquisition compensated, the 10 Gy and 5 Gy setup images were impacted by significant artefactual noise which clearly rules out the use of low specific activity phantom setups intended to reconstruction parameter optimization.

Volume reduction of large phantoms can also be achieved by vertically setting the phantoms or by adding Styrofoam inserts [6]. The phantom mean absorbed dose should always be reported.

Acknowledgements

Nobody or organization to acknowledge.

Authors' contributions

Both authors equally contributed to the study. Both authors read and approved the final manuscript.

Funding

This study was not funded.

Availability of data and materials

There are available.

Ethics approval and consent to participate

N/A (no human, no animal in the study)

Consent for publication

We agree.

Competing interests

The authors declare that they have no competing interests.

Received: 23 July 2019 Accepted: 31 October 2019

Published online: 28 November 2019

References

1. Conti M. Effect of randoms on signal-to-noise ratio in TOF PET IEEE Trans. Nucl. Sci. 2006;53:1188–93.
2. Walrand S, Hesse M, Chiesa C, Lhommel R, Jamar F. The low hepatic toxicity per Gray of 90Y glass microspheres is linked to their transport in the arterial tree favoring a nonuniform trapping as observed in posttherapy PET imaging. *J Nucl Med.* 2014;55:135–40.
3. Högberg J, Rizell M, Hultborn R, et al. Heterogeneity of microsphere distribution in resected liver and tumour tissue following selective intrahepatic radiotherapy. *EJNMMI research.* 2014;4:48.
4. van de Maat GH, Seevinck PR, Elschoot M, et al. MRI-based biodistribution assessment of holmium-166 poly(L-lactic acid) microspheres after radioembolisation. *Eur Radiol.* 2013;23:827–35.
5. d'Abadie P, Hesse M, Jamar F, Lhommel R, Walrand S. 90Y TOF-PET based EUD reunifies patient survival prediction in resin and glass microspheres radioembolization of HCC tumours. *Phys Med Biol.* 2018;63:245010.
6. Van Elmbt L, Vandenberghe S, Walrand S, Pauwels S, Jamar F. Comparison of yttrium-90 quantitative imaging by TOF and non-TOF PET in a phantom of liver selective internal radiotherapy. *Phys Med Biol.* 2011;56:6759.

Publisher's Note

Springer Nature remains neutral with regard to jurisdictional claims in published maps and institutional affiliations.

Submit your manuscript to a SpringerOpen[®] journal and benefit from:

- Convenient online submission
- Rigorous peer review
- Open access: articles freely available online
- High visibility within the field
- Retaining the copyright to your article

Submit your next manuscript at ► [springeropen.com](https://www.springeropen.com)