

# The DLR MIRO: a versatile lightweight robot for surgical applications

U. Hagn, M. Nickl, S. Jörg, G. Passig, T. Bahls, A. Nothhelfer, F. Hacker, L. Le-Tien, A. Albu-Schäffer, R. Konietschke, M. Grebenstein, R. Warpuß, R. Haslinger, M. Frommberger and G. Hirzinger  
Institute of Robotics and Mechatronics, German Aerospace Center (DLR), Wessling, Germany

### Abstract

**Purpose** – Surgical robotics can be divided into two groups: specialized and versatile systems. Versatile systems can be used in different surgical applications, control architectures and operating room set-ups, but often still based on the adaptation of industrial robots. Space consumption, safety and adequacy of industrial robots in the unstructured and crowded environment of an operating room and in close human robot interaction are at least questionable. The purpose of this paper is to describe the DLR MIRO, a new versatile lightweight robot for surgical applications.

**Design/methodology/approach** – The design approach of the DLR MIRO robot focuses on compact, slim and lightweight design to assist the surgeon directly at the operating table without interference. Significantly reduced accelerated masses (total weight 10 kg) enhance the safety of the system during close interaction with patient and user. Additionally, MIRO integrates torque-sensing capabilities to enable close interaction with human beings in unstructured environments.

**Findings** – A payload of 30 N, optimized kinematics and workspace for surgery enable a broad range of possible applications. Offering position, torque and impedance control on Cartesian and joint level, the robot can be integrated easily into telepresence (e.g. endoscopic surgery), autonomous or soft robotics applications, with one or multiple arms.

**Originality/value** – This paper considers lightweight and compact design as important design issues in robotic assistance systems for surgery.

**Keywords** Robotics, Surgery, Kinematics, Product design

**Paper type** Research paper

## 1. Introduction

Surgical robotic systems can be divided into two major groups: specialized and versatile systems. Specialized systems focus either on a dedicated surgical technique, like endoscopic surgery with the da Vinci surgical system by Intuitive Surgical (Green *et al.*, 1995) and Artemis (Schurr *et al.*, 1996) or on the treatment of a specific medical disease (e.g. cancer in Phee *et al.*, 2005). These systems can fulfil the dedicated task very well, but link the financial amortization in the clinic to single medical procedures. With ongoing research in medical treatment, many of these specialized robotic systems are likely to lose their niche. On the other hand, today's versatile systems often still base on the adaptation of industrial robots (e.g. Caspar in Albers *et al.*, 2007). Industrial robots are targeted on high-absolute accuracy which is achieved by stiff structures and thus relatively high mass. Safety and adequacy in the unstructured and crowded environment of an operating room combined with close human robot interaction are therefore at least questionable.

In contrast, the design approach of the DLR KINEMEDIC and the new generation MIRO aims at a compact, slim and lightweight robot (LWR) arm as a versatile core component for various existing and future medical robotic procedures. With its low weight of 10 kg and dimensions similar to those of the human arm, the MIRO robot can assist the surgeon directly at the operating table where space is sparse. Like the DLR LWR (Albu-Schäffer *et al.*, 2007), MIRO integrates torque sensing capabilities to enable close interaction with humans in unstructured environments. Accelerated masses are reduced by consequently employing the lightweight approach for the MIRO development. The result is an increased safety during close interaction with patient and user (Haddadin *et al.*, 2007). The reflected motor inertia is reduced by torque control up to a factor of 5 (Albu-Schäffer *et al.*, 2007), while the link inertia is low due to the lightweight design. Compact and slim design simplifies the integration of one or multiple MIRO robots in the crowded operating room.

By adding specialized instruments and modifying the application workflows within the robot control, the MIRO robot can be adapted to many different surgical procedures. This versatility has been achieved by the design of the robotic arm itself and by the flexibility of the robot control architecture. To equalize the deficits of the lightweight

---

The current issue and full text archive of this journal is available at [www.emeraldinsight.com/0143-991X.htm](http://www.emeraldinsight.com/0143-991X.htm)



Industrial Robot: An International Journal  
35/4 (2008) 324–336  
© Emerald Group Publishing Limited [ISSN 0143-991X]  
[DOI 10.1108/01439910810876427]

---

The DLR MIRO robot is funded by the Bayerische Forschungstiftung.

approach in absolute accuracy, the MIRO robot control offers the possibility to integrate external position sensors (PSs) to close the position control loop. In almost every surgical procedure where high-absolute accuracy is needed, a procedure or/and an imaging systems for registration of the patient on pre-recorded diagnostic data are state-of-the-art. These systems (e.g. navigation systems) can be used as external PSs for the MIRO robot control to achieve a high-absolute accuracy with the robotic system, whereas latency and low-sample rate of available medical tracking systems limit the dynamics of the task. In the understanding of the MIRO developers, the four basic parts of a surgical robotic system can be described as:

- 1 *surgeon* – medical know how, surveillance, flexibility, responsibility;
- 2 *robot* – exact relative positioning, endurance;
- 3 *instruments* – adaptation to a specific surgical task, dexterity; and
- 4 *navigation* – exact acquisition and planning of absolute positions.

Beside classic position control, the robot can be guided by the user through exerting forces by hand on the robot's structure. This is implemented by using integrated torque sensors (TQs) in all joints and torque control methods. The control method is described in Albu-Schäffer *et al.* (2007) as soft robotics. While haptic constraints in robotics are usually implemented on a position/velocity interface, the soft robotics concepts address robots with torque interface, where compliance is implemented at joint level, but also in Cartesian coordinates. With the different control modes the following robotic control architectures are possible:

- One or more MIRO robots are tele-manipulated by the surgeon from a distant control interface for example in endoscopic surgery.
- The robot is equipped with an instrument (e.g. drilling machine) and is guided by the surgeon manually, whereas the robot limits the free motion based on pre-planned virtual spaces and directions in the form of haptic paths, targets, virtual walls (Ortmaier *et al.*, 2006).
- The robot is equipped with an instrument (e.g. tip of a passive mirror arm for laser osteotomy) and performs a planned trajectory autonomously and is registered by additional sensors/procedures.

Two exemplary soft robotics applications have been implemented already on the first generation, the DLR KINEMEDIC, which is in the commercialization phase at BrainLAB/Germany. Integrated into BrainLAB's surgical navigation system, the KINEMEDIC prototype performs biopsies in cranial neurosurgery and drills holes for pedicle screws in spine stabilization. In both applications, the robot guides the surgeon to the point of interest (tumour or bore axis) by haptic means, whereas in the final step the robot simply holds the correct position and the surgeon completes the task manually. Instead of performing the robotic task completely autonomous, this interactive approach can enhance the acceptance of the robotic system as described in Jakopcic *et al.* (2002).

Additionally, different setups in the operating room can be combined with these control architectures. Owing to the lightweight approach, programmable gravity compensation and the corresponding dimensioning of the robot's joints, the


MIRO robot can be mounted hanging from a ceiling stand (Frumento *et al.*, 2006), on a mobile cart or at the side rails of the operating table, as shown in Figure 1. The assignment of control modes, number of robots and the different setups is interchangeable.

## 2. MIRO system design

Beside the lightweight approach, one essential design criterion of the MIRO robot is the versatility of the system. It is designed to fit seamlessly into existing surgical procedures and clinical environments, but is also configurable and extendable to comply with rapidly changing development in medical treatment and safety. To achieve this challenging demand, the MIRO robot has been conceived by clearly distinguishing between platform and application according to the principles of platform-based design (Sangiovanni-Vincentelli *et al.*, 2004). In this context, the application is designated to configure and parameterize the platform for a specific task. To achieve the demanded versatility of the system, it is necessary to cover as many system constraints as possible by the configuration, due to the faster optimization cycles in application design. Only absolute necessary, static system constraints which are inherent to the field of use (surgery) or unlikely to change have been assigned to the platform. All remaining constraints are classified as open and can be realized within the application. In order to avoid limitations in the number of valid applications, the MIRO platform design encapsulates all fundamental functions of the system into small, decoupled and tested blocks. The communication infrastructure of the MIRO platform grants low-level, direct and independent but yet convenient access to these blocks and is open for future extensions of the platform. Hence, application design can start directly on the hardware level and is only limited by the set of basic functions of the platform. The extent and impact of this design approach in the MIRO robot can be illustrated by some of the assignments of open and static constraints to application and platform, as shown in Table I.

The MIRO platform enables motion of the TCP in six degrees of freedom (DoF), restrictions deriving from the

**Figure 1** DLR MIRO robots equipped with endoscopic instruments in a tele-manipulation scenario



**Note:** Heart surgery

**Table I** Assignment of constraints to platform and application in the MIRO robot

Constraint	Description	Platform	Application
Cartesian kinematics	Unlimited motion of the TCP in six DoF	○	
Application kinematics	Motion restrictions specific to a certain surgical procedure (e.g. motion about the invariant trocar in endoscopy)		○
Relative accuracy	Minimal size of anatomic structures (e.g. in neurosurgery)	○	
Absolute accuracy	Combination of relative accuracy with registration, additional sensors and model calculation		○
Sensory measurement of joint torques and positions	Integrated joint position and TSs are capable of oversampling	○	
Sensor sample rate and resolution	Task specific adjustment of the ratio between sensor sample rate and resolution		○
Space limitations in the operating room	Slim and compact design of the robot arm	○	
Diversity of operating rooms	Omnidirectional placement of the robots base to enable different setups		○
Basic safety functions	Independent basic safety functions (e.g. redundancy)	○	
Safety strategies	Combination of basic safety functions to comply with different standards and applications		○

surgical procedure can be considered as being part of the application. In endoscopic surgery, instruments are inserted into the patient's body through small incisions or orifices, which limit the motion of the instruments to four DoF. Different platform-based solutions have been introduced to restrict the instrument motions according to this limitation either by specialized mechanisms (e.g. daVinci system) or by additional passive joints (e.g. Aesop, Zeus). In contrast, the MIRO robot offers the possibility to solve this problem within the application. One way is Cartesian position or impedance control about a virtual point of entry, which is assumed stationary or detected by sensors. A second approach is to control two joints of the robot to output zero torque.

The last two entities in Table I partitions the safety of the system into basic safety functions of the platform (sensory and communication redundancy, safe stop function, etc.) and the combination of these functions by the application. This approach is necessary due to the different existing and oncoming safety standards in the field of medical robotics.

By methods of simulation (Konietschke *et al.*, 2004), cadaver testing (Ortmaier *et al.*, 2006) and interrogation of clinicians, sets of system requirements for a variety of challenging medical applications were identified. The range of applications includes endoscopic heart surgery, drilling for pedicle screws in spine stabilization, biopsy of tumours in neurosurgery and osteotomy with CO<sub>2</sub> lasers in maxillofacial surgery. After partitioning these constraints either as static for the platform design or as open realized by the application, a set of design criteria have been formulated for the MIRO robot:

- slim, compact design to reduce the fear of close contact and to simplify the integration into the operating room;
- lightweight to reduce the impact of collisions and to facilitate the setup in the operating room;
- redundant number of joints to enhance flexibility and to assist collision avoidance of the robot's links with other objects;
- integrated sensors to measure all relevant physical properties of the system;
- different control modes where the user can always be in charge;
- motor, sensor and communication electronics integrated into the robot arm to reduce the internal cable harness and therewith the robot's dimensions;
- pc-based external robot control for scalability reasons;
- re-configurable electronics and robot control for rapid application prototyping; and
- open communication concept to integrate external sensors and actuators.


### 3. Hardware design

#### 3.1 Kinematic design

The issue of acceptance of new technologies is important for an improved learning curve of clinicians and technicians. Close interaction with technical systems demand understanding of the system, thus a central design issue of the MIRO robot is an inherent predictability of the system's actions for the user. To achieve this, a serial kinematics with seven DoF which resembles those of the human arm has been developed and optimized for medical procedures. The joint morphology groups the MIRO arm in a dedicated shoulder (roll-pitch-yaw), upper arm, elbow (pitch-roll), forearm and wrist (pitch-roll), each group with intersecting axes, as shown in Figure 2. Consequently, the motions of the robot joints in relation to the Cartesian movement of the TCP are intuitive for the user. Only the wrist pitch-roll kinematics differs from the human pitch-yaw configuration. The rotational DoF about the shaft axis of endoscopic instruments has been identified as very frequently used during endoscopic surgery. Thus, it is and is realized by a dedicated joint of the robot in order to reduce expansive robot motions. The singularity imposed by axes 5 and 7 is avoided by methods of null-space optimization in the inverse kinematics. The chosen kinematics offers six DoF at



Figure 2 MIRO kinematics (left); overlay of extremal pitch joint configurations with plot of maximum reach (right)



the TCP, additional restrictions deriving from certain medical procedures (e.g. in endoscopic surgery) will be solved by the application. This almost anthropomorphic kinematics with dedicated groups of joints with intersecting axes and the slim design are the main differences in appearance between the MIRO robot and the DLR LWR.

Based on the different selected medical procedures the kinematics of the robot has been optimized by the means of gradient methods and genetic algorithms (Konietschke *et al.*, 2004). Beside conventional surgical procedures, the optimization process paid special attention to endoscopic procedures, due to the limitations in endoscopic surgery described in Section 2. After defining quality criteria such as position accuracy, manipulability, minimum dynamics, maximum force, etc. the joint angle ranges and the ratio between the length of the upper arm ( $a_3$ ) and the forearm ( $d_5$ ) have been optimized with the goal of compact dimensions of the robot. The Denavit-Hartenberg parameters of the optimized kinematics are shown in Table II according to the notation of Yoshikawa (1990) and Craig (1986).

Table II DH parameters and joint ranges of the MIRO robot

$i$	$a_{i-1}$ (mm)	$\alpha_{i-1}$ (°)	$d_i$ (mm)	$\Theta_0$ (°)	Joint	Range (°)	Software limit (°)
1	0	0	240	0	1	$\pm 172.5$	$\pm 162.5$
2	0	-90	0	-90	2	$\pm 51$	$\pm 41$
3	310	90	0	0	3	$\pm 51$	$\pm 41$
4	0	-90	0	90	4	-57/+127	-47/+117
5	0	90	385	0	5	$\pm 172.5$	$\pm 162.5$
6	0	-90	0	0	6	$\pm 172.5$	$\pm 162.5$
7	0	90	200	0	7	$\pm 172.5$	$\pm 162.5$

### 3.2 Mechanical design

The integration of all hardware features established by the DLR LWR (torque and PSs, safety brakes, integrated electronics, etc.) in a yet more slim design with dedicated groups of joints with intersecting axes is the most challenging aspect in the mechanical design of the MIRO robot. The design approach of the LWR, with a sequential chain of revolving single joints connected by shell structures (Hirzinger *et al.*, 2002) turned out to be not feasible for the MIRO robot. In respect of good accessibility to the integrated electronics and the upper arm and forearm being the only places to integrate the electronics, this approach is not scalable to the desired slim dimensions of the MIRO robot. Therefore, the mechanical design is based on a skeletal structure and unencumbered housings. Similar to the bones in the human anatomy, these structural parts have been designed as tubes whereas the joint electronics are fitted to the outside. Decoupling of structural parts and housings gives the opportunity to design compliant polymer housings in order to reduce the severity of collisions, a method which can also be found in automotive cockpit design. Additionally, the robot can be operated without housings as shown in Figure 1 during testing and maintenance.

To form compact groups of joints with intersecting axes, couple joints with two DoF have been designed in analogy to the joints in the DLR Hand II (Butterfaß *et al.*, 2001). The couple joint design offers various advantages:

- both motors/brakes on the same side of the double joint (bundling of electronic components);
- compact joint dimensions;
- intersecting axes; and
- efficiency.

The advancement in efficiency needs further annotations. With the MIRO two DoF couple joint design the torques of

the motors add for one DoF under certain circumstances. If one joint torque is zero, the second joint torque is formed by the sum of both motor torques. As with the human arm, maximum payload needs not to be applicable in every kinematic posture. It is reasonable to use adequate kinematic postures of the robot arm in high-payload procedures. In a robot with a redundant number of joints, Cartesian tasks of the TCP can be realized by various kinematic postures in the null-space (Grunwald *et al.*, 2004). The setup and posture of the robot can therefore be optimized to achieve certain advantages. In the case of the MIRO robot, a posture that creates mostly single axis loads on the couple joints offers the opportunity that the motor torques add as described above.

The robot's shoulder is formed by a single revolving joint and a two DoF couple joint with intersecting axes (Figure 3). Joint 1 is designed similarly to the joints in the DLR LWR; motor and brake (M/B) are connected to a reduction gear (RG), a TS and a PS on the joint side and PSs on motor (not depicted) side are integrated. All joints in the MIRO robot feature a hollow shaft, which enables internal cabling of the robots electronics. The pitch and yaw DoF (axes 2 and 3) are realized by said couple joint design. Two motors equipped with safety brakes (M/B) are connected to RGs by tooth belts (TB). The two RGs are coupled by a Cardanic differential bevel gear (BG), whereas two bevels are connected to the drive side of the RGs, one bevel is running free (optional) and the driven bevel is connected to link 3. When both motors rotate in the same direction, the couple joint rotates about axis 2, whereas opposite rotations of the motors result in a joint rotation about axis 3 (yaw). Because of the limited space in the joint due to the integration of the differential BG, the TS cannot be integrated in the same way as in joint 1. Both torques of the couple joint are measured by a two axis TS which is integrated into link 3. The joint PSs of all joints are


integrated in a way which is not affected by the coupling mechanism, thus acquire the joint positions directly.


To form a pitch roll configuration for the elbow of the MIRO robot, a variation of the couple joint is used. The index position of the cardanic differential BG is rotated by 90° about axis 4 and link 5 is connected to the driven bevel as shown in Figure 3. The TS measures the torque of axis 5 and two additional orthogonal torques. Based on the known orientation of axis 5, the torque about axis 4 can be calculated out of these two additional signals.

In order to achieve a very compact pitch-roll wrist a different coupling mechanism has been designed for joints 6 and 7, which does not offer the above-mentioned efficiency advantage. Two modules containing motor, brake (M/B) and RG are connected to the joint by TBs, as shown in Figure 3. One TB directly drives link 6, whereas the second belt is connected to the driving bevel of a differential gear assembly (BG). The driven bevel is connected to link 7 and the TS, whereas the measurement of the torques is realized in the same way as in joints 4 and 5. Additionally, this joint design offers large joint ranges for axes 6 and 7 which is necessary for the use with endoscopic instruments.

Endoscopic instruments can be portioned into the functional tip (gripper, scissors, optics, etc.), a long thin shaft and an extracorporeal supply unit (Figure 4). For safety reasons it is important that the instrument can be removed from the patients body along the axis of the shaft. Whereas conventional instruments (e.g. drill machines, saws, lasers), can be mounted to the MIRO as shown in Figure 4 left, the hollow shaft of axis 7 and the large range of axis 6 ( $\pm 162.5^\circ$ ) allows endoscopic instruments to be mounted as shown in Figure 4 (right). An endoscopic instrument (e.g. DLR's actuated and sensorized surgical instruments in Seibold *et al.*, 2005) can be inserted through the hollow shaft of joint 7 and

**Figure 3** Shoulder joints 1, 2, 3 (left), elbow joints 4, 5 (middle) and wrist joints 6, 7 (right) of the MIRO robot



**Figure 4** MIRO robot equipped with passive linear guide and a drill machine (left); with endoscopic instrument in two extremal positions (right)

points with its functional tip in the opposite direction of the  $\cos_7$  coordinate system's  $z$ -axis.

### 3.3 Electronic design

In contrast to the more physical-oriented view of mechanical designers, electronic designers consider robots as signal-oriented devices. For electronic designers, robots are a composition of interfaces to the physical world, i.e. sensor and actuator modules (AMs), which are connected to computing resources by a heterogeneous communication infrastructure with strict performance constraints.

Signal noise, cabling effort, and EMC influences are reduced radically by digitalizing signals near by the physical interface. Therefore, the electronic modules have to be tightly integrated with the mechanical structure. The design of sensor and actuator components is not isolated from the mechanical design, since the outline of the modules has to fit into the morphology of the robot without constraining it.

Moreover, adequate modularisation and the standardization of communication interfaces increase reusability and reduce design costs. Related functionality should be implemented as components using a common hardware platform. As a consequence, the art of sensor and actuator design for highly integrated robots is to find the best trade-off between high integration and efficient implementation, to meet the strict constraints, and determine a suitable modularization.

All physical interface modules implement BiSS (IC-Haus, 2003), a unified sensor/actuator communication bus with very small footprint. The BiSS protocol allows cascading: different BiSS-components can be connected to a chain and only the first component of the chain is connected to the BiSS-Master. Thus, multiple sensor modules can be combined to a stack with a single BiSS-interface, i.e. BiSS allows a high degree of integration and the distribution of the modules according to the different joint designs.

A MIRO robot joint consists of three types of common components that realize the interfaces to the physical world: torque sensing module (TSM), position sensing module (PSM) and actuation module (AM). Accordingly, a MIRO

joint is a dedicated composition of sensing and AM (Table III).

#### 3.3.1 Sensing modules

The implementation of sensing modules is state-of-the-art. The sensor modules (TSM and PSM) include signal amplifier with anti-aliasing filter for two input signals, two  $\Sigma\Delta$ -ADCs and a single BiSS interface implementation. Thus, the modules are self-contained units which are connected by power supply and BiSS communication. They can be combined in different arrangements as needed in the different joints of the robot (Figure 3). All TS modules and PS modules arranged in the same location can share the same power supply and the same BiSS interface. Hence, the expense of cabling is minimized.

The  $\Sigma\Delta$ -ADC devices offer programmable decimation filters to configure the resulting sample rate and resolution of the sensor. Thus, the sampling rate can be configured by software without adapting the hardware implementation.

Point of measurement of a PSM is the conductive plastic potentiometer, which measures the joint angle. The linearity error of the potentiometers is trimmed to less than 0.1 percent. The PS electronics acquires the resistive ratios of the potentiometers and converts them to digital values for further processing.

The TSMs are based on aluminium parts with applied strain gages. The aluminium TS bodies are integral part of the robot's structure as described in Section 3.2. The joint side torques and forces cause an elastic deformation of the sensor bodies. This deformation leads to a change in resistance of the strain gages. The strain gages are distributed over the sensor body in a way that the effects of unwanted forces and torques

**Table III** Distribution of sensor and actuator modules in the MIRO robot

	Actuation	Single sensor modules	Stacked sensor modules
<b>Joint1</b>	1 AM		1 PSM, 1 TSM
<b>Joint23</b>	2 AM	2 PSM	1 TSM
<b>Joint45</b>	2 AM	1 PSM	1 PSM, 2 TSM
<b>Joint67</b>	2 AM	1 PSM	1 PSM, 2 TSM



are compensated in a bridge while the torque parallel to the joint axis causes a detuning of the bridge.

### 3.3.2 Actuation module

The AM is a current controlled power converter which is directly integrated with the motors. The AM compound reduces cabling effort and EMC problems caused by the high-switching frequencies of the power converters. Assuming motor current is proportional to the motor torque, the AM can be considered as is a high-integrated implementation of a linear torque source with a digital interface.

The field-oriented current controller for the DC motors is realized as a state-of-the-art PI-controller with space-vector modulation (Quang and Dittrich, 1999). A simple classic CORDIC operator implements the field vector rotations. A magneto resistive PS allows current control with little phase error. The motor current is measured by three shunt resistors at the source side of the low-side MOSFETs of the power converter. The maximum sample rate of current and position measurement is higher than a suitable switching frequency for the power converter. So, the sample rate of the current control can be configured within a wide range.

Additionally, the module implements safety functions (e.g. emergency stop, safety brake), and housekeeping (e.g. temperature measurement).

## 4. Infrastructure design: hardware integration and software drivers

Infrastructure of a robotic system enables the integration of the distributed electronic hardware, such as sensors and actuators, with control applications. Therefore, it has to provide convenient interfaces for application developers and constitute efficient but transparent glue between the distributed components of the robotic system. Thus, infrastructure involves the operation environment for sensors and actuators, communication protocols and software drivers.

Infrastructure design for the MIRO robot is determined by two opposite prerequisites: on the one hand, infrastructure should be an open platform that can be configured for versatile applications. This asks for a flexible and configurable architecture that is rich in functionality. On the other hand the electronic components are to be highly integrated with the mechanical structure. Thus, infrastructure needs to be lean and efficient to honor the sparse resources.

To balance these opposing requirements, a hierarchical approach is used: close to the mechanical structure, where available space is sparse, components are most dedicated. Higher layers successively add flexibility to the system leading to the use of general-purpose architectures on the top layer (Figure 5).

The challenge for designers is the decision which functionality to implement on which layer. The dedicated components on the lower levels require high-design effort and long development cycles. Hence, only functionality related to the platform should be implemented on lower levels. It is desirable to put most of the functionality on higher levels, but dedicated constraints such as high bandwidth or safety draw functionality to bottom layers.

While the classical system view is rather software-centric, the MIRO-architecture is radically different: the entire system

is synchronized by the sensors. The high-level software implementations of control algorithms are triggered by the joint hardware and the software is executed on demand, i.e. when new data are available.

In contrast to the classical hierarchy of hardware/firmware/device-drivers/software that separates hardware from software, the MIRO-architecture is built on a component model. A component model is functionality driven and not implementation driven like the device driver view. Components can be either hardware or software implementations. They are the design entities on all levels of the system, e.g. electronic circuits as well as high-level software implementations.

The synchronous signal-oriented design model introduced by Benveniste and Berry (2001) is used to define the interfaces and the interaction of the distributed components. This formal definition allows the seamless composition of components on all levels of the system hierarchy, which in turn enables the flexible movement of functionality between layers, regardless of hardware or software implementation.

### 4.1 The platform architecture

To balance the opposing requirements of flexibility and efficient implementation the MIRO platform is designed as a layered architecture, starting at dedicated highly integrated components growing towards general-purpose commercial off-the-shelf (COTS) platforms (Figure 5).

The applied communication protocols follow the same hierarchy as the modules: close to the hardware, low-bandwidth protocols featuring small implementation footprint are used, while the general-purpose platforms are connected by high-bandwidth standard communication protocols.


The modular layout on each level together with the aggregation of components on successive levels by the means of suitable communication creates the desired platform flexibility. The four layers sensor/AMs, joint modules, real-time hosts, and auxiliary hosts have proven to be a suitable hierarchy with sufficient flexibility at affordable design costs for robotic systems.

The sensor-AMs constitute the most dedicated level of the architecture as they represent the interface to the physical world (Section 3). The next level is formed by the MIRO joint modules, which are the main building blocks of the robotic system. They provide the infrastructure for the integration of the various sensor and AMs. Hence, the key for an open platform is the flexibility of the joint modules and the communication connecting them.


At this granularity, Field Programmable Gate Array (FPGA) technology is a suitable means for providing the desired flexibility at reasonable design costs. For example, the great number of available I/O pins of the FPGA allows the parallel connection of all sensor/AMs. Thus, low-bandwidth communication protocols featuring small implementation footprints can be used here. The sensor-AMs are connected to the joint modules with the industry standard sensor bus BiSS in C-Mode version.

The MIRO joint module is an instantiation of the flexible joint module concept described in Jörg *et al.* (2006). It uses a XILINX Spartan 3E FPGA for the implementation of the connection to the local modules (Figure 6).

**Figure 5** The hardware architecture features three layers of communication: BiSS, SpaceWire and Ethernet



**Figure 6** MIRO joint module: an open platform that allows the easy adaptation to changing application and safety requirements



Joint housekeeping and state monitoring are also implemented on the FPGA.

The joint modules are connected to real-time computing hosts by SpaceWire featuring 1 GB/s bandwidth and a latency lower than 20  $\mu$ s (ECSS, 2003). The packet-based SpaceWire offers a flexible and ample communication. Its protocol is entirely implemented on the joint FPGAs.

Real-time computing hosts provide the necessary computing power for the control implementation. The top level is the interconnection of real-time computing hosts to auxiliary hosts via standard Gigabit Ethernet. For example, the hosts of a multi-arm robotic setup as shown in Figure 1 can be connected to one system in this manner.

#### 4.2 The system architecture

Following the strategy to implement as much functionality as possible on general-purpose architectures all control algorithms are implemented on the external controller host, which is a standard PC platform running the real-time OS QNX (QNX Software Systems, www.qnx.com). This approach allows the use of conventional development tools such as Mathwork’s Matlab Simulink and enables efficient algorithm and application development. Moreover, one directly benefits from the constantly rising computing power of COTS platforms.

To achieve good performance in terms of control, a tight integration of the joint modules with the host by the



infrastructure is required. In conformance to the signal model the distributed joint module functionality is transparently integrated by a signal-flow-oriented middleware specifically developed for this purpose. Utilizing a static system specification approach and compile time optimization the typical overhead of common middleware implementations is avoided and a small run-time footprint is achieved (Jörg *et al.*, 2006). The middleware not only integrates the hardware with the software but enables hardware-software co-design. It covers one important common task: the communication.

Figure 7 shows the system architecture of the MIRO system: the medical application configures the joint control and the Cartesian control laws of the Control Layer according to its requirements. Section 5 explains this procedure in more detail.

The hardware abstraction layer (HAL) provides a functional separation between robot hardware and robot control. For convenience, the HAL implementation provides a complete current controlled robot to the control designers and presents all sensor values as floating point SI-values. Since the Cartesian control and the joint control are implemented with Matlab Simulink the HAL is represented as a Simulink block. The HAL not only integrates the remote joint functionality but also implements all features that are not provided by the hardware but are necessary for a suitable view of a current controlled robot, such as sensor value calibration (Figure 7).

Only functionality with high bandwidth or safety requirements, like current control and dedicated platform specific functions, is implemented on the joint modules.

Efficiency in terms of control design can be expressed by bandwidth and latency. The 1GB/s bandwidth of the SpaceWire compared to typical data of 250 byte/cycle allow theoretical control cycles of more than 500 kHz, which is

several times more than the bandwidth of typical physical processes. Figure 8 shows the overall latency from sensor to actuator that was measured while the robot was performing a typical position controlled trajectory. One can see that the latency always stays well below the deadline of one joint control cycle (i.e. 333  $\mu$ s).


## 5. Robot control

As mentioned before, the MIRO robot has two main control modes. Mode selection is accomplished in the application (Figure 9). The first one is the classical position control mode, in which the robot follows the commanded trajectory in Cartesian or joint coordinates as accurately as possible and is controlled to overcome external disturbances. This mode is required for exact positioning applications, such as laser cutting or bone drilling. The second operation mode is the compliant one, i.e. the so-called soft robotics approach, briefly described in the following section.

### 5.1 Torque and impedance control


In this mode, the user can guide the robot to a desired position or on a desired trajectory by hand, in a highly intuitive manner (Albu-Schäffer *et al.*, 2007). This is done in the so-called “gravity compensation mode”, in which a torque controller on joint level reduces the effects of friction and provides the torques needed to sustain the own weight of the robot and the tool. In this free-floating mode, several virtual springs based on virtual potential fields are used to impose constraint forces which prevent the robot entering restricted areas. This helps avoiding collisions with the patient, the surgeon and medical staff or with other instruments. Critical areas defined by the application (e.g. blood vessels or the spinal cord) are also avoided by the robot using this method.

**Figure 7** The layered MIRO system architecture: only the current control is implemented on the joint modules



**Note:** The HAL provides the interface to control and servicing

**Figure 8** From the signal path to the middleware view: the signal path mapped to the robot hardware yields signal-oriented components (i.e. the joint modules)



**Note:** The plot depicts the typical latency of the entire path

Moreover, by defining different compliances for the various Cartesian displacements of the robot, the system can accommodate to motion in free space (high stiffness) and contact with hard environments (requiring high compliance). This enables fast and sensitive interaction with the environment. Since the robot is a redundant arm with seven DoF, it can still be freely moved in the null-space while having a Cartesian compliance imposed (Ott *et al.*, 2004). In order to control also the null-space configuration of the arm, additional null-space stiffness is defined.

### 5.2 Position control

The main challenge regarding the position control is related to the fact that the robot exhibits significant elasticity due to the lightweight design. Therefore, a flexible joint model has to be assumed and the controller structure has to provide effective vibration damping and positioning accuracy in presence of elasticity. This is achieved by using a state feedback controller, with the states given by motor position and velocity as well as joint torque (after the gearbox) and its derivative  $\{q, \dot{q}, \tau, \dot{\tau}\}$  (Le-Tien *et al.*, 2007). This topic was addressed for the DLR LWRs as well. However, due to the special design of MIRO robots, the joints 2-3, 4-5, and 6-7 exhibit a strong coupling not only of the motion, but also of the elasticity and damping matrix. Therefore, the joint controllers cannot be designed separately any more, but the coupling of the two joints has to be considered, leading to a multi input-multi output (MIMO) design (Figure 10).


For the coupled joints, one has to distinguish between the motor positions measured by the encoders and the motor positions written in link coordinates. Accordingly, a coordinate transformation has to be done for transforming link side torques of the joints to motor torques. The two coordinate transformations are given by:

$$\begin{cases} q = Tq_m \\ \tau_m = T^T \tau \end{cases} \quad \text{with } T = \begin{bmatrix} 0.5 & 0.5 \\ -0.5 & 0.5 \end{bmatrix} \in \mathbb{R}^{2 \times 2}$$

Herein,  $q_m \in \mathbb{R}^2$  is the motor position,  $\tau_m \in \mathbb{R}^2$  the motor torque,  $q \in \mathbb{R}^2$  the motor position transformed to link coordinates and  $\tau \in \mathbb{R}^2$  the link torque. The controller design starts from the model written in joint coordinates on which a second coordinate transformation, using modal analysis is done for the coupled joints in order to attain optimal performance.

The feedback terms turn out to have very intuitive physical interpretations: torque feedback reduces the apparent inertia of the motors, as well as the joint friction. Since the torque is measured directly in the joints with 3 kHz rate (in contrast to usual techniques of measuring force/torque at the tip) there is no significant delay. The bandwidth of torque control is around 30 Hz. Motor position feedback is equivalent to a physical spring while velocity feedback produces energy dissipation (viscous friction). These interpretations allow a straight forward passivity and stability analysis based on Lyapunov theory and enable also a consistent generalization to Cartesian impedance control. In order to exemplify the

Figure 9 The controller structure for the DLR MIRO




**Notes:** Whereas  $\phi$  is the link position,  $q$  the motor position,  $w^T^{\text{TCP}}$  the desired Cartesian configuration,  $K_d$  and  $D_d$  the desired Cartesian stiffness and damping matrices,  $f_d$  the desired TCP force vector,  $K_0$  and  $D_0$  the desired null-space stiffness and damping matrices

contribution of motors and links to the total inertia, we evaluated the reflected Cartesian inertia in a typical working configuration. The maximal contribution of the link inertia is 5.8 kg; the maximal contribution of the motor inertia is 17.5 kg. However, as mentioned in the control section, the torque feedback scales down the reflected motor inertia, such that the total maximal reflected inertia for the controlled robot is 10.2 kg. Moreover, in case of fast impacts of the robot, only the reflected link inertia of 5.8 kg is relevant (Haddadin *et al.*, 2007), due to the intrinsic joint elasticity.

Moreover, by appropriate scaling of the feedback gains, the controller structure is used to implement position, torque or impedance control. For example, high torque and torque derivative gains, zero position gain and positive velocity feedback (for compensation of viscous friction) provide a torque controller, while high position and velocity gains are used for position control, together with lower torque feedback gains for vibration damping. On the other side, the desired torque for the controller is commanded according to the expected robot dynamics (e.g. if the robot is not moving, this corresponds to the gravity torques).

## 6. Summary

The DLR MIRO presented in this paper is a new versatile lightweight research robot tailored for surgical applications. The compact and slim design, realized by the design of differential couple joints and highly integrated electronics, allows close interaction of robot and surgeon directly at the operation table, where space is sparse. With its low weight of 10 kg, the accelerated masses are reduced significantly in comparison to an industrial robot, which enhances the safety in the unstructured environment of an operating room both for the user and the patient. Despite its low weight, a maximum payload of 30 N and a workspace similar to those of the human arm has been achieved. By this the DLR MIRO will be capable to accomplish a broad range of possible surgical applications. The integration of joint side position and TSs beside the common motor PSs, enables impedance and position control on joint and Cartesian level. Based on the DLR MIRO, we are currently developing a prototype setup for endoscopic heart surgery consisting of three robot arms equipped with active endoscopic instruments and camera. This setup will be presented on the AUTOMATICA fair 2008 in Munich.

**Figure 10** The structure of the MIMO state feedback controller for the coupled joints

## References

- Albers, J., Schmidt, T., Hassfeld, S., Heid, F. and Vahl, C. (2007), "Sternotomy and craniotomy by an autonomous robot", *Zeitschrift für Herz-, Thorax- und Gefäßchirurgie*, Vol. 21.
- Albu-Schäffer, A., Ott, Ch. and Hirzinger, G. (2007), "A unified passivity based control framework for position, torque and impedance control of flexible joint robots", *International Journal of Robotics Research*, Vol. 26 No. 1, pp. 23-39.
- Benveniste, A. and Berry, G. (2001), "The synchronous approach to reactive and real-time systems", *Proceedings of the IEEE*, Vol. 79, pp. 1270-82.
- Butterfaß, J., Grebenstein, M., Liu, H. and Hirzinger, G. (2001), "DLR-Hand II: next generation of a dextrous robot hand", *Proceedings of the 2001 IEEE International Conference on Robotics and Automation*, Seoul, Korea.
- Craig, J.J. (1986), *Introduction to Robotics: Mechanics and Control*, Addison Wesley Longman Publishing Co., Boston, MA.
- ECSS (2003), *ECSS E-50-12A SpaceWire – Links, Nodes, Routers and Networks*, European Cooperation for Space Standardization, Noordwijk.
- Frumento, S., Micheli, R., Konietzschke, R., Hagn, U., Ortmaier, T. and Hirzinger, G. (2006), "A co-robotic positioning device for carrying surgical end-effectors", *Proceedings of ESDA 2006*, Torino.
- Green, P.S., Hill, J.W., Jensen, J.F. and Shah, A. (1995), "Telepresence surgery", *IEEE Engineering in Medicine and Biology Magazine*, Vol. 14 No. 3.
- Grunwald, G., Schreiber, G., Albu-Schäffer, A. and Hirzinger, G. (2004), "Touch: the intuitive type of human and robot interaction", in Grunwald, G., Prassler, E., Lawitzky, G., Stopp, A., Grunwald, G., Hägele, M., Dillmann, R. and Iossifidis, I. (Eds), *Advances In Human Robot Interaction*, Springer Tracts in Advanced Robotics, Vol. 14, Springer, Berlin.
- Haddadin, S., Albu-Schäffer, A. and Hirzinger, G. (2007), "Safety evaluation of physical human-robot interaction via crash-testing", *Proceedings of the 2007 Robotics: Science and Systems Conference (RSS2007)*, Atlanta, USA.
- Hirzinger, G., Sporer, N., Albu-Schäffer, A., Hähle, M., Krenn, R., Pascucci, A. and Schedl, M. (2002), "DLR's torque-controlled light weight robot III – are we reaching the technological limits now?", *Proceedings of the 2002 IEEE International Conference on Robotics and Automation ICRA*, Washington DC, USA, May, pp. 1710-6.
- IC-Haus (2003), "BiSS interface, protocol description, release B2", IC-Haus, available at: [www.BiSS-ic.de/](http://www.BiSS-ic.de/)
- Jakopec, M., Harris, S.J., Baena, F.R., Gomes, P., Cobb, J. and Davies, B.L. (2002), "Preliminary results of an early clinical experience with the AcrobotTM system for total knee replacement surgery", *Proceedings of Fifth International Conference on Medical Image Computing and Computer-Assisted Intervention (MICCAI)*, Tokyo, Japan, September 2002, pp. 256-63.
- Jörg, S., Nickl, M. and Hirzinger, G. (2006), "Flexible signal-oriented hardware abstraction for rapid prototyping of robotic systems", *Proceedings of the International Conference on Intelligent Robots and Systems*, Beijing, 2006.



- Konietschke, R., Ortmaier, T., Weiss, H., Hirzinger, G. and Engelke, R. (2004), "Manipulability and accuracy measures for a medical robot in minimally invasive surgery", *Proceedings, 9th International Symposium on Advances in Robot Kinematics (ARK), Sestri Levante, Italy*, June 28-July 1, 2004.
- Le-Tien, L., Albu-Schäffer, A. and Hirzinger, G. (2007), "MIMO state feedback controller for a flexible joint robot with strong joint coupling", paper presented at IEEE International Conference on Robotics and Automation, April.
- Ortmaier, T., Weiß, H., Hagn, U., Nickl, M., Albu-Schäffer, A., Ott, C., Jörg, S., Konietschke, R., LeTien, L. and Hirzinger, G. (2006), "A hands-on-robot for accurate placement of pedicle screws", *Proceedings of the ICRA-2006 IEEE International Conference on Robotics and Automation, ICRA 2006, Orlando, FL, USA*, pp. 4179-86.
- Ott, Ch., Albu-Schäffer, A. and Hirzinger, G. (2004), "A passivity based Cartesian impedance controller for flexible joint robots – Part I: torque feedback and gravity compensation", *IEEE International Conference on Robotics and Automation*, pp. 2659-65.
- Phee, L., Xiao, D., Yuen, J., Chan, C.F., Ho, H., Thng, C.H., Cheng, C. and Ng, W.S. (2005), "Ultrasound guided robotic system for transperineal biopsy of the prostate",

- IEEE International Conference on Robotics and Automation 2005*, Vol. 2, pp. 1315-20.
- Quang, N. and Dittrich, J. (1999), *Praxis der feldorientierten Drehstromantriebsregelungen*, Expert, Renningen.
- Sangiovanni-Vincentelli, A., Carloni, L., De Bernardinis, F. and Sgroi, M. (2004), "Benefits and challenges for platform-based design", *Proceedings of the 41st Annual ACM IEEE Conference on Design Automation, San Diego, CA, USA*, pp. 409-14.
- Schurr, M.O., Breitwieser, H., Melzer, A., Kunert, W., Schmitt, M., Voges, U. and Buess, G. (1996), "Experimental telemanipulation in endoscopic surgery", *Surgical Laparoscopy and Endoscopy*, Vol. 6, pp. 167-75.
- Seibold, U., Kübler, B. and Hirzinger, G. (2005), "Prototype of instrument for minimally invasive surgery with 6-axis force sensing capability", *ICRA 2005 IEEE International Conference on Robotics and Automation, CD ICRA 2005, Barcelona, April 18-22*, pp. 498-503.
- Yoshikawa, T. (1990), *Foundations of Robotics: Analysis and Control*, The MIT Press, Cambridge, MA.

### Corresponding author

U. Hagn can be contacted at: [ulrich.hagn@dlr.de](mailto:ulrich.hagn@dlr.de)