

Towards an SI-Traceable Reference Measurement System for Seven Serum Apolipoproteins Using Bottom-Up Quantitative Proteomics: Conceptual Approach Enabled by Cross-Disciplinary/Cross-Sector Collaboration

Christa M. Cobbaert ^{a,*} Harald Althaus,^b Ilijana Begcevic Brkovic,^{c,d} Uta Ceglarek,^{c,d} Stefan Coassin ^e, Vincent Delatour,^f Liesbet Deprez ^g Ioannis Dikaios ^g Julia Ditttrich,^{c,d} Andrew N. Hoofnagle,^h Gerhard M. Kostner ⁱ Florian Kronenberg ^e Zsusanna Kuklenyik,^j Urban Prinzing,^k Hubert W. Vesper,^j Ingrid Zegers,^f and L. Renee Ruhaak ^a, for the IFCC Working Group for Standardization of Apolipoproteins by Mass Spectrometry

ABSTRACT: Current dyslipidemia management in patients with atherosclerotic cardiovascular disease (ASCVD) is based on traditional serum lipids. Yet, there is some indication from basic research that serum apolipoproteins A-I, (a), B, C-I, C-II, C-III, and E may give better pathophysiological insight into the root causes of dyslipidemia. To facilitate the future adoption of clinical serum apolipoprotein (apo) profiling for precision medicine, strategies for accurate testing should be developed in advance.

Recent discoveries in basic science and translational medicine set the stage for the IFCC Working Group on Apolipoproteins by Mass Spectrometry. Main drivers were the convergence of unmet clinical needs in cardiovascular disease (CVD) patients with enabling technology and metrology. First, the residual cardiovascular risk after accounting for established risk factors demonstrates that the current lipid panel is too limited to capture the full complexity of lipid metabolism in patients. Second, there is a need for accurate test results in highly polymorphic and atherogenic apolipoproteins such as apo(a). Third, sufficient robustness of mass spectrometry technology allows reproducible protein

quantification at the molecular level. Fourth, several calibration hierarchies in the revised ISO 17511:2020 guideline facilitate metrological traceability of test results, the highest achievable standard being traceability to SI.

This article outlines the conceptual approach aimed at achieving a novel, multiplexed Reference Measurement System (RMS) for seven apolipoproteins based on isotope dilution mass spectrometry and peptide-based calibration. This RMS should enable standardization of existing and emerging apolipoprotein assays to SI, within allowable limits of measurement uncertainty, through a sustainable network of Reference Laboratories.

Introduction

TRADITIONAL DIAGNOSIS OF DYSLIPIDEMIA IN CARDIOVASCULAR DISEASE

After many years, the evaluation of atherosclerotic cardiovascular disease (ASCVD) risk still centers around the procedurally defined classification based on density

^aDepartment of Clinical Chemistry and Laboratory Medicine, Leiden University Medical Center, Leiden, The Netherlands; ^bSiemens Healthcare Diagnostics Products GmbH, Marburg, Germany; ^cInstitute of Laboratory Medicine, Clinical Chemistry and Molecular Diagnostics, University Hospital Leipzig, Leipzig, Germany; ^dLIFE-Leipzig Research Center for Civilization Diseases, University of Leipzig, Leipzig, Germany; ^eInstitute of Genetic Epidemiology, Department of Genetics and Pharmacology, Medical University of Innsbruck, Innsbruck, Austria; ^fLaboratoire National de Métrologie et d'Essais, Paris, France; ^gEuropean Commission, Joint Research Centre (JRC), Geel, Belgium; ^hDepartment of Laboratory Medicine and Pathology, University of Washington, Seattle, WA; ⁱGottfried Schatz Research Center (for Cell Signaling, Metabolism and Aging), Division of Molecular Biology and Biochemistry, Medical University of Graz, Graz, Austria; ^jDivision of Laboratory Sciences, Centers for Disease Control and Prevention, Atlanta, GA; ^kRoche Diagnostics GmbH, Penzberg, Germany.

*Address correspondence to this author at: Leiden University Medical Centre, Department of Clinical Chemistry and Laboratory Medicine, Postzone E2-P, Albinusdreef 2, 2333 ZA Leiden, The Netherlands. E-mail c.m.cobbaert@lumc.nl.

[†]Disclaimer: The findings and conclusions in this report are those of the author(s) and do not necessarily represent the official position of the Centers for Disease Control and Prevention/the Agency for Toxic Substances and Disease Registry. Use of trade names is for identification only and does not imply endorsement by the Centers for Disease Control and Prevention, the Public Health Service, and the US Department of Health and Human Services.

Received August 4, 2020; accepted September 15, 2020.

DOI: 10.1093/clinchem/hvaa239

of lipoprotein particles [High Density Lipoproteins (HDL), Low Density Lipoproteins (LDL), Very Low Density Lipoproteins (VLDL), chylomicrons (CM)] and their quantification based on cholesterol and triglyceride content (1). The traditional lipid profile [HDL-cholesterol (HDL-C), LDL-cholesterol (LDL-C), total cholesterol, and total triglycerides] allows effective identification of individuals who are at high risk for ASCVD due to familial hypercholesterolemia (2, 3). However, even in the age of intensive lipid lowering therapies, there is still a high proportion of patients who die prematurely or experience a major cardiovascular event while on lipid lowering therapy. In elderly populations with several comorbidities (i.e., obesity, diabetes, or chronic kidney disease), the traditional lipid profile is a weak predictor of ASCVD-related mortality (4). Meanwhile, basic research has greatly expanded our understanding of human lipid metabolism, which has been accelerated by the use of modern gene editing and omics technologies (5, 6). In clinical care, these advances in basic research and new drug developments will reach our patients only if our clinical laboratory measurements evolve accordingly. The traditional cholesterol-centered approach has placed a high bar for any newly emerging cardiovascular biomarker to overcome. This is due to the specificity, accuracy, precision, interlaboratory reproducibility, commutability, feasibility for automation, point of care use, and cost of analysis of the cholesterol assay. As a result, new assay technologies and tests will be adopted more swiftly if they both offer solutions for unmet clinical needs in existing care pathways and if tests results are anchored to higher order reference materials and methods in a sustainable way.

DIAGNOSTIC SIGNIFICANCE OF AN APOLIPOPROTEIN (APO)

PANEL FOR ASCVD: A COMING-OF-AGE STORY

Apolipoproteins (apos) A-I and B provide scaffolding support for lipoproteins and are essential for the biogenesis, routing, and metabolism of HDL and non-HDL particles circulating in blood. Serum apoA-I and B measurements allow more direct quantification of HDL and LDL particle numbers, respectively (7). However, apoA-I and B themselves do not provide information on specific irregularities of HDL and non-HDL functions. Efforts to assess functional characteristics of lipoproteins based on size, density, charge of subfractions, or other functional considerations (e.g., cholesterol efflux capacity) have not achieved an acceptable level of interlaboratory comparability, due to lack of clear analytical definition of subclasses across different analytical platforms (8). A more feasible way to assess functional characteristics of lipoproteins (without subfractionation) is the measurement of other apolipoproteins along with apoA-I and B. Exchangeable apolipoproteins, such as

apolipoproteins (apos) C-I, C-II, and C-III and E bind to apoA-I and B-containing lipoproteins and regulate the concerted action of lipase enzymes, transfer proteins, and receptors (9). From genetic studies, especially Mendelian randomization studies, there is now ample evidence that apos are risk markers in cardiovascular disease. Specifically, evidence has been gathered for apos AI, (a), B, C-I, C-II, CIII, and E (10). ApoC-III and E are now monitored in numerous drug trials and proof of concept intervention studies (11–15). However, the lack of standardized analytical measurements prevents the compilation of data from different laboratories into large scale epidemiologic studies, which will be necessary for these newer apolipoprotein biomarkers to be accepted as risk markers.

Currently there is heightened interest in Lp(a) research and its clinical application, as it is now considered a strong genetically determined risk factor for cardiovascular disease (CVD) (16). The lipoprotein(a) particle is composed of an LDL particle with an additional apolipoprotein associated with it, which is called apolipoprotein(a) (apo(a)) and shows a high structural homology with plasminogen. While Lp(a) is known to be synthesized in the liver, the site and mechanism of catabolism is more controversial. After more than 50 years of research, the physiological function of Lp(a) is still unclear (17). One astonishing characteristic of Lp(a) is the more than 1000-fold range of concentrations between individuals, from almost zero to more than 3000 mg/L (16). In addition, the distribution of Lp(a) is almost Gaussian in Blacks but skewed in Caucasians; roughly 50% of Caucasians have concentrations below 100 mg/L and about 25% have concentrations above 300 mg/L.

Lp(a) concentrations are not significantly influenced by age, sex, fasting state, inflammation (18, 19), or lifestyle factors, such as diet and physical activity. However, the concentrations are under strict genetic control. Family studies have revealed a heritability estimate of Lp(a) concentrations of about 90%. The discovery of the size polymorphism of apo(a) in serum, which is based on a variable number of so-called kringle IV type 2 (K-IV₂) repeats in the *LPA* gene, resulted in the identification of the *LPA* gene as the major gene predicting Lp(a) concentrations. Each gene has a few to more than 40 repeats, with each repeat stretching 5.6 kB. This results in a highly polymorphic and informative copy number variation.

Interestingly, there is a pronounced inverse correlation between the number of K-IV₂ repeats and the serum/plasma concentration of Lp(a). Individuals expressing a low number of K-IV₂ repeats resulting in small apo(a) isoforms (up to 22 K-IV₂ repeats) have on average markedly higher serum Lp(a) concentrations, and thereby a higher CVD risk, compared with

individuals carrying only large apo(a) isoforms (more than 22 K-IV₂ repeats) (16). This K-IV₂ size polymorphism of apo(a) explains about 20%–80% of the variability of serum Lp(a) concentrations, depending on ethnicity. However, recent technological advancements have revealed a large number of genetic variants within the K-IV₂ repeats. Some of them can have a pronounced effect on Lp(a) concentrations resulting (e.g., in low Lp(a) concentrations even in a person carrying a small apo(a) isoform) (20, 21). Many Mendelian randomization studies provided strong support for an association between high Lp(a) concentrations and CVD risk (16).

The variable number of identical K-IV₂ repeats interferes with the accurate measurement of Lp(a) in serum and plasma (16), with lower Lp(a) concentrations being overestimated and higher concentrations being underestimated by K-IV₂ dependent tests, and masked its clinical utility for nearly 2 decades. Commercial IVD tests and technologies should be improved and properly standardized so that they are no longer prone to confounding by the apo(a) size polymorphism in native specimens, calibrators and reference materials.

ADOPTION OF SERUM/PLASMA APOLIPOPROTEINS IN CLINICAL PRACTICE

For primary prevention, the measurement of serum apoB and Lp(a) has been proposed to overcome the limitations of traditional lipid biomarkers and to better identify individuals at risk (22). In addition, proteomics using Multiple Reaction Monitoring (MRM)-MS platforms, which allows for simultaneous quantitation of multiple proteins, was selected by *Nature Methods* as the technology of the year 2012. Applied to epidemiological cohorts, the technique has demonstrated the association of serum apolipoprotein profiles with incident CVD in ongoing prospective community-based surveys such as the Bruneck or the EPIC-Norfolk study (23, 24). This parallels current efforts to obtain more detailed information from individual patient samples that could help us provide more customized diagnoses and treatments to our patients (*precision medicine*). Current technologies (next generation protein diagnostics with, for example, MRM-MS) and a more detailed understanding of the causes of dyslipidemia, allow for precision medicine to be applied, which was not possible a few years ago (10).

Consistent, effective, and sustainable diagnosis and treatment can only be achieved if lab tests measure well-defined measurands and have well-known trueness and uncertainty within allowable limits that match their clinical use. To that end, establishing metrological traceability is imperative (25). Metrological traceability requires the establishment of an advanced Reference Measurement System that accommodates evolution of

our understanding of basic science as well as advances in the science of measurement, which will facilitate personalized patient management. It is also essential to make this Reference System robust, reproducible, and sustainable in a network of calibration labs across the globe.


PAVING THE WAY TO AN SI-TRACEABLE REFERENCE MEASUREMENT SYSTEM FOR MULTIPLE CLINICALLY RELEVANT APOLIPOPROTEINS

Metrological traceability of serum/plasma apolipoproteins: status update. It is well-known that metrological traceability of test results is essential for anchoring test results to internationally recognized, higher order materials, and methods in line with ISO 17511:2020 calibration hierarchies (see Supplemental Document for links to web resources). In addition, the measurement uncertainty (MU) on the reported test results should remain within acceptable limits to ensure that the test is *fit-for-clinical purpose*. To accomplish this, IFCC-WHO standardization efforts were initiated in the 1990s for serum/plasma apoA-I and B (26, 27) and Lp(a) (28). The contemporary metrological traceability chains for these 3 analytes are shown in Fig. 1 (left) and demonstrate traceability of test results to WHO-IFCC secondary reference materials, coded by WHO and IFCC as “SP1-01”, “SP3-07” and “SRM2B”.

For apoA-I, SP1-01 was prepared by lyophilizing a human serum pool to ensure long-term stability. The material was value assigned using a highly standardized radioimmunoassay, calibrated with purified apoA-I, for which the mass value had been determined by amino acid analysis. Using SP1-01, with an apoA-I concentration of 1.50 g/L as a common calibrator, inter-assay variation could be reduced from 9% to 5.4%, showing the success of this strategy (26).

For apoB, SP3-07 was prepared as a shock-frozen liquid preparation, because lyophilization was known to affect apoB structure (29). An apoB value of 1.22 g/L was assigned to SP3-07, using a nephelometric method that was calibrated with freshly isolated low-density lipoprotein for which the apoB mass value was determined by a standardized sodium dodecyl sulfate-Lowry procedure. Using SP3-07 as a common calibrator, the inter-assay variation of commercial tests could be reduced from 20% to 7% (27).

For Lp(a) standardization, SRM2B, a lyophilized human serum pool with preservatives, was selected from 4 manufactured Lp(a) materials as the material providing best harmonization results in 27 Lp(a) tests. Lp(a) isoforms consisted of 3 major apo(a) polymorphs of nearly equal gel intensity, containing 16, 17, and 18 K-IV₂ domains, and 3 minor polymorphs of 14, 20, and 32 K-IV₂ domains, respectively (28). Value assignment of SRM2B was performed through ELISA



measurements that were calibrated against 2 individually freshly isolated Lp(a) preparations from a donor exhibiting a single apo(a) isoform. The absolute mass of the isolates was determined using amino acid analysis, and 2 different K-IV₂ number-independent ELISAs were used for value assignment of SRM2B (30). However, when SRM2B was used to achieve uniformity of calibration, inter-laboratory CVs of up to 31% were still observed during the measurement of 30 fresh frozen serum samples with 22 assays (31). A significant portion of the inter-laboratory variance might lie in the variability of assay reactivity toward lipoprotein(a) particles with different numbers of apo(a) kringle domains, but another portion likely stems from assay nonlinearity and variable end-user adherence to the standard operating procedure (e.g., for predilutions to ensure that measurements are done in the optimal linear range). The need for next generation protein tests that measure apo(a) in a kringle-independent manner is obvious.

An important limitation of the WHO-IFCC materials discussed above is their unknown commutability. Also, some of these materials are no longer available or their general availability to the lab community is very limited. Existing stocks of material are running low, which creates the need for new materials. Further, new

reference materials now need to meet current ISO requirements that were not in place at the time these other materials were created (coded. SP1-01, SP3-07, and SRM2B). For apolipoproteins C-I, C-II, C-III, and E standardization, no secondary WHO-IFCC reference materials exist at this moment.

Evolution in metrology: striving for SI-traceability of serum/plasma apolipoprotein test results. The science of measurement has evolved. To that end, ISO Technical Committee 212 working group 2 developed several international ISO standards, among them ISO 15193 (with requirements for reference measurement procedures), ISO 15194 (with requirements for reference materials), ISO 15195 and 17025 (describing competences of calibration laboratories), and ISO 17511:2020 (with requirements for establishing metrological traceability of values assigned to calibrators, trueness control materials and human samples). All ISO standards contribute to a global infrastructure and guidance for standardization of medical tests. The primary goal is to accomplish metrological traceability of test results across the globe. Metrological traceability is defined as the property of a measurement result whereby the result can be related to a reference through a documented

unbroken chain of calibrations, each contributing to measurement uncertainty. Measurement traceability is important because it assures that medical test results are clinically effective and safe by agreeing with national or international standards within the statement of uncertainty in measurement. Without traceability, a laboratory can claim anything they want in a test or calibration report. To implement the metrological traceability concept, calibration hierarchies are essential. The most complete calibration hierarchy is the one that allows the community to trace test results back to a well-defined system of units, based on the International System of Quantities (SI), as adopted by the General Conference on Weights and Measures (CGPM) [VIM 3, JCGM 200:2012].

The reference materials developed for apoA-I, apoB, and Lp(a), as well as the reference measurement procedures, were *state-of-the-art* in the 1990s. Today, it is clear that for some medical tests, methods such as liquid chromatography coupled to liquid chromatography (LC)-MRM-MS, with its higher selectivity, have become promising alternatives for radioimmunoassays and ELISAs. As the current Reference Measurement Systems for apoA-I, apoB, and Lp(a) only allow traceability to secondary WHO-IFCC reference materials, there is a need to reconsider the existing Reference Measurement Systems (RMS) and to achieve SI-traceability in a sustainable manner. By evolving into Next Generation Protein measurement technology, RMS should improve and allow molecular characterization of the measurands (apos) of interest.

An IFCC working group on quantitating Apolipoproteins by Mass Spectrometry was formed in 2017 to establish such a mass spectrometry-based RMS. The objectives are 2-fold: To develop an RMS for a panel of clinically relevant serum apos A-I, (a), B, C-I, C-II, C-III, and E (including qualitative phenotyping where needed). Measurement results should be traceable to SI as outlined in ISO 17511:2020, and the envisioned calibration hierarchy in Fig. 1 (right). Other traceability chains will be used in cases where traceability to SI cannot be achieved. A second objective will be to evaluate clinical performance and clinical utility of serum apolipoprotein panel(s) for CVD risk stratification and treatment, in comparison to or together with contemporary blood lipids once the RMS is in place. The working group is cross-disciplinary and includes representation from metrology institutes and the Joint Research Centre (JRC, former IRMM, Geel, Belgium), candidate reference laboratories that developed LC-MRM-MS-based tests for absolute quantification of serum apolipoproteins, experts in apolipoprotein biochemistry and genetics, and IVD-representatives. The first step in the development of a reference measurement system was the establishment of a common accuracy

base, with pre-agreed starting points. An overview of the common accuracy base is presented in Table 1, and the proposed strategy to achieve apolipoprotein standardization is outlined in Fig. 2.

It was agreed that an affordable and robust reference measurement system should be developed in line with technological advances and the latest ISO standards on metrological traceability and standardization. LC-MRM-MS, using isotope dilution, is currently a widely accepted methodology for standardization of protein tests in serum/plasma (Fig. 3). The technology is attractive, as it allows quantification of proteins in an antibody-independent manner, minimizes by design batch-to-batch variations (e.g., proteolytic enzymes used for digestion, LC-columns used for separation), is independent of specific suppliers, and can be reproduced by other laboratories using that technology. Moreover, isotope dilution mass spectrometry (IDMS) allows specific (molecular) characterization of the measurand(s), which currently cannot be achieved using immunoassay-based techniques. Third, the technique allows multiplexing. Particularly because the technology easily enables multiplexed quantification of apolipoproteins, IDMS was the first method of choice for establishing an SI-traceable RMS for multiple apolipoproteins.

A major challenge of quantitative proteomics using LC-MRM-MS is that proteins are enzymatically digested into their peptides, which has been termed the bottom-up proteomics strategy (Fig. 3). As the measurands are altered during the measurement process, demonstrating completeness of digestion is an essential prerequisite for accurate quantitation. From a metrological viewpoint, this methodology may be advantageous too: if the digestion of intact apolipoproteins into proteotypic peptides is sufficiently optimized to enable equimolar conversion with consistent reproducibility, a peptide-based calibration strategy is possible to accomplish SI-traceability. Main advantages of peptide-based primary reference materials as compared to recombinant protein-based reference materials are the fact that peptide-based reference materials can be obtained more readily, and their purity assessment is much easier as compared to that of recombinant apolipoproteins. Peptide-based calibration will be achievable if peptide-based primary reference materials can be developed for each apolipoprotein (32). It was therefore decided, as a starting point, that primary reference materials should ideally be synthetic peptides of high purity, allowing SI-traceability for the setup of a sustainable and global RMS for apolipoproteins. See Fig. 1.

IDMS-based protein quantification through bottom-up proteomics relies on several assumptions that should be met to take full advantage of this powerful technology and correctly implement the metrological traceability concept. First, the apolipoproteins in

Table 1. Predefined starting points for the establishment of an SI-traceable IDMS-based Reference Measurement System for multiple serum apolipoproteins^a.

Principle n	Starting points	Rational choices agreed by the IFCC Working Group members for establishing a Mass-Spectrometry (MS)-based Reference Measurement System (RMS). All definitions are explained in VIM3, JCGM 200:2012.
1	General	The RMS for serum apo standardization by MS has to be feasible and sustainable and should be set up preferentially in at least 3 calibration laboratories.
2		Metrological traceability to the highest calibration hierarchy in ISO 17511:2020 should be the aim (i.e., apolipoprotein test results should preferentially be traceable to SI). The apolipoprotein measurands have to be defined and quantitated unequivocally, notwithstanding the heterogenous nature of proteins.
3		Beyond the standardization of the analytical RMP/RMS, standardization of the Total Testing Process is needed, including the preanalytical conditions, the type of matrix (serum), the intended measurands, the units (molar), and the reference values and decision limits.
4		Terminology should be compliant with relevant ISO documents, such as ISO 17511; 15193; 15194; 15195; 17025, VIM3, and the IFCC C-NPU terminology and silver book. ^b
5	Reference Materials	The goal is to prepare primary reference materials for each individual apolipoprotein. These primary reference materials will be prepared and stored individually, to ensure stability. However, for use in the RMP, these materials will be mixed to form reference standards in serum for constructing calibration curves.
6		Peptide standards are the primary reference materials of choice, however, if that is not feasible, recombinant proteins will be considered.
7		Commutable and human serum-based reference materials are developed as secondary reference materials.
8	Reference Method	The goal is to establish one harmonized bottom-up proteomics procedure for 7 apolipoproteins by mass spectrometry across the calibration labs involved.
9		We aim to develop a multiplexed primary Reference Measurement Procedure (RMP) for detecting and/or quantifying serum apos (a), A-I, B, C-I, C-II, C-III, E, and the apoE phenotype. The analytical selectivity should be guaranteed through use of appropriate quantifying peptides.
10		The RMP should enable metrological traceability of serum apo test results to SI, within allowable measurement uncertainty (i.e., enough uncertainty budget should be left for the manufacturers and end-users of routine apo tests). As a rule of thumb and in case that biological variation data are available, less than 50% of the allowable uncertainty budget derived from biological variation should be consumed by the candidate RMP.

^aThe starting points in this table were defined after thorough preparational discussions with the working group members and corporate members, taking into consideration the pre-agreed objectives.


^bWeb-links to the respective regulatory documents are provided in the [Supplemental Information](#).


serum-based secondary reference materials should be present in their intact and native form to ensure accurate quantification. Second, the quantifying peptides should be specific for the endogenous apolipoproteins and selected carefully to reflect the intended measurand(s) accurately. Third, equimolar digestion from apolipoprotein into quantifying peptide should be achieved using standardized sample preparations and independent of individual patients'

matrices. These aspects should all be thoroughly investigated during the development of the RMS.

STEPWISE APPROACH TO ESTABLISHING AN MS-BASED REFERENCE SYSTEM

Defining the measurands (quantities intended to be measured). The first step in the development of a measurement system is the accurate definition of the





measurand. Because proteins exist in multiple proteoforms, which each have a different mass, it was decided to quantify each of the proteins in molar units. Specifically for apo(a), which has a size polymorphism affecting its quantification, the peptides should be from a unique KIV₂-independent domain. For apoE, 2 clinically relevant amino acid polymorphisms are known, resulting in the genetic variants ϵ 2, ϵ 3, and ϵ 4, represented by phenotypes E2, E3, and E4. Therefore, apoE phenotype identification was also deemed important.

Peptides will be selected for each measurand. However, the selection of the representative peptides is not a trivial task, as was outlined previously (33). Importantly, 2 of the 3 major assumptions for bottom-up protein quantification should be met: the proteotypic peptide should reflect the intended measurand, and digestion from protein to peptide should be equimolar. The first criterion can be met theoretically, while the second should be evaluated empirically. To ensure peptides will be selected that accurately reflect the intended measurand, the following criteria were agreed upon: the selected peptides should be unique for the target protein (proteotypic peptides) and at least 2 proteotypic peptides should be quantified per protein to ensure analytical specificity. Moreover, to avoid misrepresentation, the peptides should ideally not contain cysteine, methionine, or tryptophan residues, not be affected by (common or rare) mutations, and not carry any post-translational modifications. Exceptions to these criteria should only be allowed if no other peptides are available. For apoE phenotyping, the genetically variable peptides CLAVYQAGAR (E2), LAVYQAGAR (E3 and E4),

LGADMEDVR (E4), and LGADMEDVCGR (E2 and E3) should be measured (34).

Development of primary and secondary reference materials, including purity and stability testing, evaluation of commutability and value-assignment, according to relevant ISO guidelines. A major requirement for establishing, implementing, and maintaining a reference system relying on an MS-based reference measurement procedure (RMP) is to have primary reference materials (RMs) of well-characterized purity [17511:2020] (Fig. 2). Primary calibrators consist of purified peptides (or proteins) that are used to calibrate the reference measurement procedure. Different approaches are available for determination of the concentration of calibration solutions: mass balance, amino acid analysis, quantitative nuclear magnetic resonance, and elemental analysis (32). While technically challenging, rigorous purity assessment of primary calibrators is an essential requisite to establish metrological traceability of results to SI units (32) and to obtain consistent results over prolonged periods of time during which different batches of the primary calibrators will be used.

Provided that the RMP spends less than 50% of the total allowable uncertainty derived from biological variation data, the RMP can then be used to assign target values to 2 types of secondary RMs (serum-based matrix materials): secondary calibrators used to calibrate IVD-tests and trueness verifiers used to verify accuracy in external quality assessment (EQA) programs or as part of the post-market vigilance of IVD tests. Both secondary calibrators and trueness verifiers should have

appropriate stability, homogeneity, and commutability properties, in compliance with ISO 15194.

Commutability relates to the closeness of agreement between results for the candidate RMs and results for clinical samples (CSs) when measured with ≥ 2 measurement procedures. Assessment of commutability requires a dedicated study in which the RMs and a sufficient number of CSs (and matrices) are measured with the RMP and all relevant IVD tests. The study design and the criteria applied for assessing the commutability should be consistent with the required analytical performance specifications (APS) of the IVD tests (35).

The availability of well-characterized and commutable secondary calibrators is no guarantee for successful standardization of commercial IVD tests. Variations in calibration procedure, including the number and level of calibration points or the type of diluent, may still cause relevant differences in measurement results. A harmonized calibration protocol might be needed to further improve inter-method comparability. The calibrated IVD tests should also have an adequate selectivity for measurand. Selectivity differences between an IVD test and the RMP will persistently lead to biased results on individual CSs, even after calibration. These sample-specific effects can become visible in a split-sample comparison of the RMP and IVD test(s) on a large group of fresh, unprocessed CSs. If the sample-specific effects are at a clinically undesirable level, a modification of the medical test improving its selectivity is required.

Development of a common IDMS-based candidate reference method in a network of calibration laboratories. To facilitate protein standardization, a full RMS should be in place, consisting of both primary and secondary RMs and a RMP (Fig. 2). We aim to develop an IDMS-based candidate RMP according to the highest ISO 17511:2020 calibration hierarchy and C62A quality requirements. Moreover, the RMP should be compliant with ISO 15193 for reference measurement procedures and ISO 15195 for reference laboratories.

To ensure a sustainable RMP, the procedure is developed in a network of 3 independent candidate reference laboratories: 2 in Europe (Leiden University Medical Centre (LUMC) in Leiden, The Netherlands, and University Hospital Leipzig (UKL) in Leipzig, Germany) and 1 in the United States (Centers for Disease Control and Prevention (CDC), Atlanta, GA, USA). The first step was agreement on the starting points for a common MS-based RMP, taking into account the Terms of Reference of the IFCC WG. During method development, common target peptides will be selected, and common transitions will be

developed. The use of stable isotope labelled peptides as internal standards is essential, not only for reliable and accurate qualification and quantification, but also to account for variations in ionization efficiency. Initial method comparisons will be performed using mixtures of synthetic peptides. These mixtures may also be used to monitor instrument performance in the respective labs according to system suitability testing procedures. When sample preparation procedures are harmonized, the enzymatic digestion will be optimized to reach equimolar apoprotein-to-peptide conversion. Given that the WG aims for peptide-based primary calibration, transfer of concentrations from quantifying peptide to intact protein through the candidate RMP assumes identical behavior between the peptide-calibrators and the endogenous peptides (representing the proteins) produced through enzymatic digestion. This is the most crucial step of the candidate RMP development (36) that has to be proven with dedicated experiments for each of the apolipoproteins. Behavior of peptides as calibrators, protein digestion kinetics, and recovery of the surrogate proteotypic peptides will be examined to provide evidence that the sample preparation procedure guarantees complete protein digestion.

Analytical performance validation of a common candidate RMP. Similar to any medical test, the candidate RMP should be *fit-for-purpose*, considering its higher order role in the calibration hierarchy. Based on our objectives and starting points for a common accuracy base, an envisioned traceability chain is drafted (Fig. 1). Due to error propagation, each step in the chain introduces additional MU, hence the error budget for the candidate RMP should, as a rule of thumb, be maximally 25%–50% of the total allowable error budget per measurand. We hypothesize that a limited MU contribution of the candidate RMP is achievable, as such uncertainties were achieved in previous evaluations of MS-based apolipoprotein tests already (37–39), and demonstrate long-term stable performance of such a method (39).

Once a common, multiplexed LC-MRM-MS-based candidate reference method is developed, a detailed Standard Operating Procedure will be written, according to ISO 15193, 15195, and 17025 requirements. To evaluate the analytical performance of the common candidate RMP, each laboratory will assess analytical sensitivity, linearity, limit of detection and quantification, repeatability, and uncertainty. Ring trials are periodically organized to monitor accuracy of test results produced by candidate reference laboratories running the candidate RMP. All 3 reference labs should fulfil the predefined APS before the candidate RMP can be submitted to the IFCC Scientific Division and the protein

review team of the Joint Commission on Traceability in Laboratory Medicine (JCTLM).

An approach to improve the accuracy and reliability of routine IVD-tests by applying the Reference Measurement System for calibration and monitoring: future perspective. Adequate implementation of the RMS and the Metrological Traceability concept demands a 3-step process. In the first step, the measurands are defined and the reference measurement system, consisting of reference methods and materials, is established. In the second step, the reference measurement system is used to calibrate and verify the analytical performance of assays operated by IVD manufacturers and laboratories providing in-house developed tests. In the third step, the accuracy and analytical performance of testing in patient care and research settings is monitored on its fitness with clinical needs. This 3-step process, which is called standardization, has been successfully applied to improve global reliability and accuracy of cholesterol, HbA1c, and other clinical analytes, and can be applied to apolipoproteins. It needs to be noted that in any measurement process, the instrumentation, reagents, and laboratory operations can change over time, which can introduce a change in analytical performance. Therefore, this standardization process needs to be continuously applied.

This implies that the RMS needs to be maintained and operational, which can best be achieved through a network of reference laboratories that ensures consistency of analytical performance of reference measurements over time. This network performs regular interlaboratory comparison studies where network members need to demonstrate appropriate analytical performance and metrological traceability in line with relevant ISO standards. Examples of successfully operated reference laboratory networks are the IFCC Network for Standardization of HbA1c, the network of the National Glycohemoglobin Standardization Program, and the CDC Cholesterol Reference Method Laboratory Network. The reference laboratories also assign reference values to commutable serum materials for the purpose of (re)calibration or trueness verification.

Recalibrated IVD manufacturers need to ensure that the level of accuracy and reliability established at the manufacturer in the second step of the standardization process is properly transferred to the tests deployed by end-users in patient care, public health, and research. This can be assessed by monitoring the analytical performance of end-users. Programs such as the CDC Lipids Standardization Program (CDC LSP), Accuracy-based Monitoring Program (CDC AMP), or Accuracy-based External Quality Assessment (EQA) programs successfully monitor the analytical performance of individual

laboratories. While CDC LSP and CDC AMP assess the analytical performance using multiple time points, EQA programs typically assess laboratories using a single time point. The combination of longitudinal monitoring programs and cross-sectional EQA programs provides comprehensive information about measurement accuracy and reliability. Once the entire reference system for apolipoproteins is established (Fig. 2) a formal monitoring program will be introduced. Lp(a) is already part of several accuracy-based EQA programs, and it is being included in the CDC LSP program.

Conclusion

In parallel with technological innovations and our enhanced understanding of the role of blood apolipoproteins in the pathophysiology of ASCVD, the science of measurement of these potential risk markers also is evolving. Therefore, existing WHO-IFCC based Reference Measurement Systems need to be improved and redefined. Hence, the IFCC working group on Apolipoproteins by Mass Spectrometry has established a scientific and technical basis for SI-standardization of 7 relevant serum apolipoproteins. Currently, a complete Apolipoprotein Reference Measurement System is under development, consisting of a harmonized multiplex Reference Measurement Procedure, uniplex primary Reference Materials, and secondary serum-based Reference Materials. The analytical selectivity of the apolipoprotein measurements is guaranteed with quantitative LC-MRM-MS, based on bottom-up proteomics and peptide calibrators. The metrological traceability of contemporary and emerging apolipoprotein tests will be accomplished according to the latest standardization insights as laid down in ISO 17025:2018 and 17511:2020. A harmonized, multiplex Reference Measurement Procedure will be available by the end of 2020, whereas the development of the primary and secondary RMs has been postponed due to the COVID pandemic and is still ongoing. The development of RMs for apo(a) has been prioritized, followed by RMs for apos A-I and B, and finally for the apos C and E. Primary RMs are expected to be available by the end of 2021, whereas secondary RMs are estimated to be available by the end of 2022. The IFCC WG on Apolipoprotein standardization by mass spectrometry strives to have the SI-traceable Reference Measurement System for apo(a) standardization of medical tests in place in 2021, whereas the complete Reference Measurement System for apo A-I, B, C-I, C-II, C-III, and E test standardization will be ready for use by the end of 2022/early 2023 (Fig. 2). Meanwhile, the network of calibration labs running the Reference Measurement System will further improve its analytical

performance, periodically run ring trials, and establish Apolipoprotein Standardization Programs. The convergence of RMS development and implementation is essential for global apolipoprotein standardization in the near future. As the analytical performance and clinical performance of medical tests are interdependent and interrelated key features that have to be considered thoroughly during biomarker-to-test development, concomitant apolipoprotein standardization according to the latest metrology insights is a natural “evolutionary” step in the science of measurement and a prerequisite for improving patient management in this era of precision cardiology.

Supplemental Material

[Supplemental material](#) is available at *Clinical Chemistry* online.

Nonstandard Abbreviations: Apo, apolipoprotein; Apo(a), apolipoprotein(a); Apos, apolipoproteins; APS, analytical performance specifications; ASCVD, atherosclerotic cardiovascular disease; CDC, Centers for Disease Control and Prevention, Atlanta, Georgia, USA; CDC LSP, CDC lipid standardization program; CDC AMP, CDC accuracy-based monitoring program; CGPM, General Conference on Weights and Measures; CLSI, Clinical and Laboratory Standards Institute; CM, chylomicrons; CS, clinical sample; CVD, cardiovascular disease; DGKL, German Society for Clinical Chemistry and Laboratory Medicine; ELISA, enzyme-linked immunosorbent assay; EQA, External Quality Assessment; HbA1c, Hemoglobin A1c; HDL, High Density Lipoprotein; HDL-C, High Density Lipoprotein cholesterol; IDMS, isotope dilution mass spectrometry; IFCC, International Federation of Clinical Chemistry and Laboratory Medicine; ISO, International Organization for Standardization; IVD, In Vitro Diagnostic; JCGM, Joint Committee for Guides in Metrology; JCTLM, Joint Commission on Traceability in Laboratory Medicine; JRC, Joint Research Centre, Geel, Belgium; K-IV₂, kringle IV type 2; LC, liquid chromatography; LDL, Low Density Lipoprotein; LDL-C, Low Density Lipoprotein Cholesterol; LNE, Laboratoire National de Métrologie et d’Essais; LPA, lipoprotein(a)

gene; Lp(a), lipoprotein(a) particle; LUMC, Leiden University Medical Centre; MRM, multiple reaction monitoring; MS, mass spectrometry; MU, measurement uncertainty; Non-HDL, Non-High Density Lipoprotein; SI, international system of quantities (*Système Internationale*); RM(s), Reference Material(s); RMP, Reference Measurement Procedure; RMS, Reference Measurement System; SRM, Standard Reference Material; UKL, University Hospital Leipzig; VLDL, Very Low Density Lipoprotein; VLDL-C, Very Low Density Lipoprotein Cholesterol; WG, Working Group; WHO, World Health Organization.

Author Contributions: All authors confirmed they have contributed to the intellectual content of this paper and have met the following 4 requirements: (a) significant contributions to the conception and design, acquisition of data, or analysis and interpretation of data; (b) drafting or revising the article for intellectual content; (c) final approval of the published article; and (d) agreement to be accountable for all aspects of the article thus ensuring that questions related to the accuracy or integrity of any part of the article are appropriately investigated and resolved.

All authors have accepted responsibility for the entire content of this manuscript and approved its submission. A.N. Hoofnagle, financial support; Z. Kuklennyk, provision of study material or patients; L.R. Ruhaak, administrative support.

Authors’ Disclosures or Potential Conflicts of Interest: Upon manuscript submission, all authors completed the author disclosure form. Disclosures and/or potential conflicts of interest:

Employment or Leadership: A.N. Hoofnagle, *Clinical Chemistry*, AACC; U. Prinzing, Roche Diagnostics Penzberg Germany.

Consultant or Advisory Role: G. Kostner, IFCC working group for apo-Lp standardization; H.W. Vesper, IFCC Committee on Bone Metabolism.

Stock Ownership: None declared.

Honoraria: None declared.

Research Funding: University Hospital Leipzig (UKL), Germany received financial support from the German Society for Clinical Chemistry and Laboratory Medicine (DGKL). Andy Hoofnagle’s institution received funding from NIDDK/NIH.

Expert Testimony: A.N. Hoofnagle, Kilpatrick Townsend & Stockton LLP.

Patents: None declared.

Role of Sponsor: The funding organizations played no role in the design of study, choice of enrolled patients, review and interpretation of data, preparation of manuscript, or final approval of manuscript.

References

- Langlois MR, Nordestgaard BG, Langsted A, Chapman MJ, Aakre KM, Baum H, et al. Quantifying atherogenic lipoproteins for lipid-lowering strategies: consensus-based recommendations from EAS and EFLM. *Clin Chem Lab Med* 2020;58:496–517.
- Watts GF, Gidding S, Wierzbicki AS, Toth PP, Alonso R, Brown WV, et al. Integrated guidance on the care of familial hypercholesterolaemia from the international FH foundation. *Eur J Prev Cardiol* 2015;22:849–54.
- Wilemon KA, Patel J, Aguilar-Salinas C, Ahmed CD, Alkhnifsawi M, Almahmeed W, et al. Reducing the clinical and public health burden of familial hypercholesterolemia: a global call to action. *JAMA Cardiol* 2020;5:217–29.
- Duncan MS, Vasan RS, Xanthakis V. Trajectories of blood lipid concentrations over the adult life course and risk of cardiovascular disease and all-cause mortality: observations from the Framingham study over 35 years. *J Am Heart Assoc* 2019;8:e011433.
- Savoji H, Mohammadi MH, Rafatian N, Toroghi MK, Wang EY, Zhao Y, et al. Cardiovascular disease models: a game changing paradigm in drug discovery and screening. *Biomaterials* 2019;198:3–26.
- Barba I, Andres M, Garcia-Dorado D. Metabolomics and heart diseases: from basic to clinical approach. *Curr Med Chem* 2019;26:46–59.
- Mora S, Buring JE, Ridker PM. Discordance of low-density lipoprotein (LDL) cholesterol with alternative LDL-related measures and future coronary events. *Circulation* 2014;129:553–61.
- Mora S. Advanced lipoprotein testing and subfractionation are not (yet) ready for routine clinical use. *Circulation* 2009;119:2396–404.
- Zannis VI, Fotakis P, Koukos G, Kardassis D, Ehnholm C, Jauhainen M, Chroni A. HDL biogenesis, remodeling, and catabolism. *Handb Exp Pharmacol* 2015;224:53–111.
- Renee Ruhaak L, van der Laarse A, Cobbaert CM. Apolipoprotein profiling as a personalized approach to the diagnosis and treatment of dyslipidaemia. *Ann Clin Biochem* 2019;56:338–56.
- Su X, Peng D. The exchangeable apolipoproteins in lipid metabolism and obesity. *Clin Chim Acta* 2020;503:128–35.
- Marais AD. Apolipoprotein E in lipoprotein metabolism, health and cardiovascular disease. *Pathology* 2019;51:165–76.
- Yamamoto R, Sacks FM, Hu FB, Rosner B, Furtado JD, Aroner SA, et al. High density lipoprotein with apolipoprotein C-III is associated with carotid intima-media

- thickness among generally healthy individuals. *Atherosclerosis* 2018;269:92-9.
14. Mendivil CO, Rimm EB, Furtado J, Sacks FM. Apolipoprotein E in VLDL and LDL with apolipoprotein C-III is associated with a lower risk of coronary heart disease. *J Am Heart Assoc* 2013;2:e000130.
 15. Mendivil CO, Rimm EB, Furtado J, Chiuve SE, Sacks FM. Low-density lipoproteins containing apolipoprotein C-III and the risk of coronary heart disease. *Circulation* 2011;124:2065-72.
 16. Kronenberg F, Utermann G. Lipoprotein(a): resurrected by genetics. *J Intern Med* 2013;273:6-30.
 17. Kostner KM, Kostner GM, Wierzbicki AS. Is Lp(a) ready for prime time use in the clinic? A pro-and-cons debate. *Atherosclerosis* 2018;274:16-22.
 18. Kronenberg F. Lipoprotein(a) in various conditions: to keep a sense of proportions. *Atherosclerosis* 2014;234:249-51.
 19. Langsted A, Kamstrup PR, Nordestgaard BG. Lipoprotein(a): fasting and nonfasting levels, inflammation, and cardiovascular risk. *Atherosclerosis* 2014;234:95-101.
 20. Coassin S, Erhart G, Weissensteiner H, Eca Guimaraes de Araujo M, Lamina C, Schonherr S, et al. A novel but frequent variant in LPA KIV-2 is associated with a pronounced Lp(a) and cardiovascular risk reduction. *Eur Heart J* 2017;38:1823-31.
 21. Coassin S, Schonherr S, Weissensteiner H, Erhart G, Forer L, Lusso JL, et al. A comprehensive map of single-base polymorphisms in the hypervariable LPA kringle IV type 2 copy number variation region. *J Lipid Res* 2019;60:186-99.
 22. Stone NJ, Grundy SM. The 2018 AHA/ACC/multi-society cholesterol guidelines: looking at past, present and future. *Prog Cardiovasc Dis* 2019;62:375-83.
 23. Baig F, Mayr M. What are the prospects of apolipoprotein profiling for cardiovascular disease? *Expert Rev Mol Diagn* 2017;17:805-7.
 24. Pechlaner R, Tsimikas S, Yin X, Willeit P, Baig F, Santer P, et al. Very-low-density lipoprotein-associated apolipoproteins predict cardiovascular events and are lowered by inhibition of APOC-III. *J Am Coll Cardiol* 2017;69:789-800.
 25. Cobbaert C, Smit N, Gillery P. Metrological traceability and harmonization of medical tests: a quantum leap forward is needed to keep pace with globalization and stringent IVD-regulations in the 21st century! *Clin Chem Lab Med* 2018;56:1598-602.
 26. Marcovina SM, Albers JJ, Henderson LO, Hannon WH. International Federation of Clinical Chemistry standardization project for measurements of apolipoproteins A-I and B. III. Comparability of apolipoprotein A-I values by use of international reference material. *Clin Chem* 1993;39:773-81.
 27. Marcovina SM, Albers JJ, Kennedy H, Mei JV, Henderson LO, Hannon WH. International Federation of Clinical Chemistry standardization project for measurements of apolipoproteins A-I and B. IV. Comparability of apolipoprotein B values by use of international reference material. *Clin Chem* 1994;40:586-92.
 28. Tate JR, Berg K, Couderc R, Dati F, Kostner GM, Marcovina SM, et al. International Federation of Clinical Chemistry and Laboratory Medicine (IFCC) standardization project for the measurement of lipoprotein(a). Phase 2: selection and properties of a proposed secondary reference material for lipoprotein(a). *Clin Chem Lab Med* 1999;37:949-58.
 29. Marcovina SM, Albers JJ, Dati F, Ledue TB, Ritchie RF. International Federation of Clinical Chemistry standardization project for measurements of apolipoproteins A-I and B. Clin Chem 1991;37:1676-82.
 30. Dati F, Tate JR, Marcovina SM, Steinmetz A. International Federation of Clinical Chemistry and Laboratory Medicine. First WHO/IFCC international reference reagent for lipoprotein(a) for immunoassay-Lp(a) SRM 2B. *Clin Chem Lab Med* 2004;42:670-6.
 31. Marcovina SM, Albers JJ, Scanu AM, Kennedy H, Giaculli F, Berg K, et al. Use of a reference material proposed by the International Federation of Clinical Chemistry and laboratory medicine to evaluate analytical methods for the determination of plasma lipoprotein(a). *Clin Chem* 2000;46:1956-67.
 32. Josephs RD, Martos G, Li M, Wu LQ, Melanson JE, Quaglia M, et al. Establishment of measurement traceability for peptide and protein quantification through rigorous purity assessment-a review. *Metrologia* 2019;56:044006.
 33. Lange V, Picotti P, Doman B, Aebersold R. Selected reaction monitoring for quantitative proteomics: a tutorial. *Mol Syst Biol* 2008;4:222.
 34. Ruhaak LR, Smit NPM, Suchiman HED, Pieterse MM, Romijn F, Beekman M, Cobbaert CM. MS-based proteomics: a metrological sound and robust alternative for apolipoprotein E phenotyping in a multiplexed test. *Clin Chem Lab Med* 2019;57:e102-e104.
 35. Nilsson G, Budd JR, Greenberg N, Delatour V, Rej R, Panteghini M, et al. IFCC working group recommendations for assessing commutability; part 2: using the difference in bias between a reference material and clinical samples. *Clin Chem* 2018;64:455-64.
 36. van den Broek I, Smit NP, Romijn FP, van der Laarse A, Deelder AM, van der Burgt YE, Cobbaert CM. Evaluation of interspecimen trypsin digestion efficiency prior to multiple reaction monitoring-based absolute protein quantification with native protein calibrators. *J Proteome Res* 2013;12:5760-74.
 37. van den Broek I, Romijn FP, Nouta J, van der Laarse A, Drijfhout JW, Smit NP, et al. Automated multiplex LC-MS/MS assay for quantifying serum apolipoproteins A-I, B, C-I, C-II, C-III, and E with qualitative apolipoprotein E phenotyping. *Clin Chem* 2016;62:188-97.
 38. Ceglarek U, Dittrich J, Becker S, Baumann F, Kortz L, Thiery J. Quantification of seven apolipoproteins in human plasma by proteotypic peptides using fast LC-MS/MS. *Proteomics Clin Appl* 2013;7:794-801.
 39. Toth CA, Kuklennyik Z, Jones JJ, Parks BA, Gardner MS, Schieltz DM, et al. On-column trypsin digestion coupled with LC-MS/MS for quantification of apolipoproteins. *J Proteomics* 2017;150:258-67.
 40. Ruhaak LR, Smit NPM, Romijn F, Pieterse MM, van der Laarse A, van der Burgt YEM, Cobbaert CM. Robust and accurate 2-year performance of a quantitative mass spectrometry-based apolipoprotein test in a clinical chemistry laboratory. *Clin Chem* 2018;64:747-9.