



Published in final edited form as:

*Curr Opin Urol.* 2009 March ; 19(2): 192–195.

## Treatment protocols to reduce renal injury during shock wave lithotripsy

James A. McAteer<sup>a</sup>, Andrew P. Evan<sup>a</sup>, James C. Williams Jr<sup>a</sup>, and James E. Lingeman<sup>b</sup>

<sup>a</sup>Department of Anatomy and Cell Biology, Indiana University School of Medicine

<sup>b</sup>Methodist Hospital Institute for Kidney Stone Disease Indianapolis, Indiana, USA

### Abstract

**Purpose of review**—Growing concern over the acute and long-term adverse effects associated with shock wave lithotripsy calls for treatment strategies to reduce renal injury and improve the efficiency of stone breakage in shock wave lithotripsy.

**Recent findings**—Experimental studies in the pig model show that lithotripter settings for power and shock wave rate and the sequence of shock wave delivery can be used to reduce trauma to the kidney. Step-wise power ramping as is often used to acclimate the patient to shock waves causes less tissue trauma when the initial dose is followed by a brief (3–4 min) pause in shock wave delivery. Slowing the firing rate of the lithotripter to 60 shock waves/min or slower is also effective in reducing renal injury and has the added benefit of improving stone breakage outcomes. Neither strategy to reduce renal injury – not power ramping with ‘pause-protection’ nor delivering shock waves at reduced shock wave rate – have been tested in clinical trials.

**Summary**—Technique in lithotripsy is critically important, and it is encouraging that simple, practical steps can be taken to improve the safety and efficacy of shock wave lithotripsy.

### Keywords

adverse effects; kidney trauma; lithotripsy; shock waves

### Introduction

It is well established that unwanted renal and extrarenal side effects can occur as a consequence of shock wave lithotripsy (SWL), and the topic of SWL injury has been the subject of numerous in-depth reviews (Reviewed in [1–5]). Most of what is known about the structural and functional characteristics of renal injury in SWL and the factors that influence the severity of tissue damage come from research with experimental animals [2,5,6,7,8]. This work has shown that injury can be affected by a variety of risk factors, such as age, the size of the kidney, the number of renal units and the presence of renal disease. It has also been shown that renal damage is dependent on the number of shock waves and the shock wave firing rate and power setting of the lithotripter. For example, measurements of lesion size in the acute (4 h posttreatment) juvenile pig SWL injury model have shown that treatment with a conservative dose of shock waves using the Dornier HM3 lithotripter (Dornier MedTech America, Kennesaw, Georgia, USA; 1000 shock waves, 24 kV, 120 shock waves/min) produced a hemorrhagic lesion measuring approximately 0.3% of the renal

Correspondence to Professor James A. McAteer, PhD, Department of Anatomy and Cell Biology, Indiana University School of Medicine, 635 Barnhill Drive, MS 5055, Indianapolis, IN 46202 5120, USA Tel: +1 317 274 7935; fax: +1 317 278 2040; jmcateer@iupui.edu.

parenchyma (functional renal volume, FRV) [6]. Doubling the dose (2000 shock waves) increased the lesion 20-fold (~6% FRV), and raising the dose to 8000 shock waves further doubled the lesion (~14% FRV) [7]. In the same animal model, but with a dose of 2000 shock waves, lesion size was observed to be dependent on the power setting of the lithotripter, increasing from approximately 0.3% FRV at 12 kV to approximately 2.25% FRV at 18 kV and approximately 6% FRV at 24 kV [8]. Thus, the severity of acute renal injury in SWL is dose dependent and is affected by the parameters of shock wave delivery.

The long-term consequences of SWL injury are not nearly as well understood, but there is solid evidence to show that acute SWL injury initiates an inflammatory response with progression to scar formation and permanent loss of functional renal mass [9]. Dose-dependent parenchymal fibrosis has been observed in dogs [10] and rabbits [11], and there is report of complete atrophy of renal papillae in pigs 3 months following treatment [1]. Clinical reports of long-term effects in SWL have not been extensive, but are concerning just the same. Studies indicate a potential link between SWL and new-onset hypertension with the suggestion that increased age may be a significant risk factor (reviewed in [1,3]). In addition, multiple lithotripsies have been implicated as a risk factor in the exacerbation of stone disease [12]. That is, patients who had received multiple SWL treatment sessions showed a greater tendency over time for transformation from calcium oxalate stone disease to a calcium phosphate or brushite phenotype. There has also been the suggestion that SWL may lead to onset of diabetes mellitus [13]. In this retrospective 19-year follow-up study of patients who underwent SWL for renal stones, lithotripsy increased the risk of developing diabetes mellitus compared with non-SWL controls. In addition, the occurrence of diabetes mellitus in these patients was related to the total number of shock waves they received and the power setting of the lithotripter.

Thus, both the severity of acute renal injury and the potential for long-term adverse effects appear to be dependent on the dose of shock waves applied. This is ample reason to look for ways to reduce the shock wave dose and to identify strategies that will minimize injury that occurs during SWL.

### **Pretreatment with shock waves protects against subsequent shock wave lithotripsy injury**

The first suggestion that shock waves could be used to trigger a protective response by the kidney came in a preliminary study using the pig model in which it was observed that treatment of the lower pole (2000 shock waves, 24 kV, Dornier HM3) produced a lesion measuring approximately 7% of the renal parenchyma, but immediate retargeting and delivery of the same dose of shock waves to the upper pole resulted in a lesion measuring only 0.1% FRV. Subsequently, four groups of juvenile pigs (6–7 weeks of age) were studied to test the idea that an initial dose of shock waves at a low power setting (12 kV) might prove sufficient to induce protection against injury during continued treatment [14]. An initial dose of either 2000 or 500 shock waves at 12 kV was delivered to the lower pole followed by retargeting and treatment of the upper pole with 2000 shock waves at 24 kV. In both cases, injury to the second pole was significantly reduced compared with that when no pretreatment was performed. Similarly, when a dose of either 500 or 100 shock waves at 12 kV was delivered to the lower pole followed by a dose of 2000 shock waves (24 kV) to the same pole, the injury was significantly less than that when no pretreatment was used. Thus, pretreatment of the kidney with shock waves at a low power setting protected the kidney from injury caused by subsequent shock waves. The effect occurred when both doses of shock waves were delivered to the same pole, and the threshold for this protective effect was 100 shock waves or less. The physiologic mechanism for this response was not assessed in the study, but work has begun in this area [15\*].

## Clinical implications

These findings suggest that initiating SWL with the delivery of a ‘preemptive’ dose of shock waves – as few as 100 shock waves at low power setting – could protect against injury caused by subsequent shock waves. The treatment protocol at many centers involves some variation on this theme, that is, using a regimen of power ramping or step-wise SWL. Such protocols have been in place for many years and likely came about as a means to help patients adapt to SWL performed under minimal anesthesia [16]. The precise sequence of shock wave delivery (number of shock waves at various power settings) in such a step-wise protocol may vary considerably depending on the preference and experience of the urologist. This includes whether or not there is any pause in shock wave delivery as the power is ramped up, and many groups make this simple adjustment ‘on-the-fly’. From the standpoint of initiating a protective effect, the timing of shock wave administration maybe key. That is, in the pig experiments described above, the initial dose of shock waves was always followed by a 3–4 min pause (dictated by the time it took to retarget the opposite pole of the kidney). In experiments in which both the initial and subsequent shock wave doses were delivered to the same pole, this 3–4 min pause was retained as part of the treatment protocol. As described next, this brief pause in shock wave delivery appears to be the critical factor in the SWL ‘protection protocol’. As such, there maybe no reason to expect that step-wise power ramping without such a pause – a scenario that to our knowledge has yet to be addressed in clinical trials – would be protective.

### **‘ Pause - protection ’ : allowing the kidney to respond to initial shock wave exposure may be necessary to afford protection from injury**

Subsequent studies in the pig model addressed the timing of shock wave delivery when a pretreatment dose was applied [17\*]. Experiments were performed to determine if the pause in treatment that was part of the original experimental protocol (see above) played a role in the renal response to shock waves. To test this, pigs were treated with an initial dose of 100 shock waves at 24 kV followed 3 min later by 2000 shock waves also at 24 kV. Injury was significantly reduced compared with no pretreatment ( $0.51 \pm 0.14\%$  versus  $3.93 \pm 1.29\%$  FRV,  $P = 0.0135$ ) [17\*]. This showed that the protection afforded by shock wave pretreatment was not because the pretreatment dose was at a lower power setting but was, instead, a consequence of the delay between the initial dose and the main dose.

It would seem to be a very simple and otherwise noncomplicating step to incorporate a brief pause in treatment following delivery of the first 100 or so shock waves in virtually any SWL treatment protocol. If the results from the renal injury studies in pigs discussed above are any indication of the potential for improved safety in SWL for patients, adding 3–4 min to the treatment protocol could bear substantial benefit.

Step-wise power ramping protocols, in which shock waves are delivered first at a low power setting then the amplitude is gradually increased, appear to give improved stone breakage as well. Artificial stones *in vitro* [18] or implanted in pig kidneys [19] broke better with gradual ramping, and there is a recent report of improved clinical outcome with such a protocol [20]. It is highly unlikely that addition of ‘pause-protection’ to such a protocol would reduce the efficiency of stone breakage. Thus, step-wise power ramping with ‘pause-protection’ could be both protective and a means to enhance stone comminution.

### **Slowing shock wave rate reduces renal injury in the pig model**

With constraints on facilities at busy stone centers, it is likely that most patients are treated at a firing rate of 120 shock waves/min. Recent studies with pigs indicate, however, that

choice of shock wave rate in the clinical range (30–120 shock waves/min) can affect lesion size and slowing the firing rate to 30–60 SW/min delivers a significant reduction in renal injury [21\*\*]. The first suggestion that lowering the firing rate of the lithotripter below 120 shock waves/min could reduce injury came as an unexpected finding during studies to characterize the performance of the low-pressure wide-focalzone electromagnetic XX-ES lithotripter [22]. In those tests, pigs were treated at the same settings used to treat patients with this machine (27 shock waves/min, ~17 MPa at 9.3 kV, 1500 shock waves) [23], and for comparison pigs were treated using the Dornier HM3 to deliver the same number of shock waves at roughly similar settings (30 shock waves/min, ~32 MPa at 18 kV). The lesion in the XX-ES animals was too small to effectively quantify, but as this lithotripter did not maintain consistent output at 120 shock waves/min, a direct comparison of the effect of shock wave rate on renal injury could not be made. However, lesion size in the HM3 group was also very low (<0.1% FRV) compared with previous studies using 120 shock waves/min in which the lesion measured approximately 5% FRV [5,6]. A follow-up study with the HM3 subsequently demonstrated a significant reduction in lesion size when pigs were treated (2000 shock waves, 24 kV) at 30 shock waves/min compared with 120 shock waves/min ( $0.08 \pm 0.02\%$  versus  $4.6 \pm 1.7\%$  FRV,  $P < 0.005$ ) [21\*\*]. Thus, reducing the shock wave rate had a significant protective effect.

## Clinical implications

Several prospective clinical trials [24–26] supported by an independent metaanalysis[27\*] have now reported a significant improvement in stone breakage outcomes when patients are treated at 60 shock waves/min compared with 120 shock waves/min. This strongly suggests a stone breakage advantage in reducing the firing rate of the lithotripter. The potential for reduced injury at slow shock wave rate is added reason to use slower rate. Many urologists would likely find it difficult to treat their patients at the extremely slow rate of 30 shock waves/min shown to be protective. Therefore, it is encouraging that recent studies in the pig model show that injury is also reduced at 60 shock waves/min compared with 120 shock waves/min (0.42 versus 3.93% FRV,  $P = 0.034$ ). Thus, there are solid laboratory data to show that injury is reduced when slower shock wave rate is used, and this finding is complemented by clinical studies reporting improved outcomes at reduced shock wave rate. On the basis of these data, it seems reasonable to suggest that slowing the firing rate of the lithotripter improves both the safety and the efficacy of SWL.

## Improved acoustic coupling can improve the efficiency of treatment

Acoustic coupling with modern dry-head lithotripters is not as efficient as when a water bath is used. Defects (air pockets) at the coupling interface arise when coupling is first established and worsen if the patient moves or is repositioned. In-vitro studies using a test tank, with an acoustic window as a surrogate patient, have demonstrated that the quality of coupling is highly variable and that coupling defects can interfere with the transmission of shock waves to the target, thereby requiring more shock waves to break stones [28]. Laboratory tests have shown that simple measures – such as minimizing handling of the gel, using an excess volume of gel and employing the inflation feature of the treatment head to smooth out the gel can significantly improve coupling [29\*\*].

## Clinical Implications

The problem of coupling is three-fold: inefficient coupling diminishes the effectiveness of treatment, the high variability of coupling likely contributes to variability in clinical outcomes and ineffective coupling leads to exposure of the patient to more shock waves than are needed, thereby increasing the potential for adverse effects – and is made more difficult by the fact that there is currently no method to monitor the coupling interface during SWL.

## Conclusion

The safety and effectiveness of SWL can be improved by attention to technique. Animal studies clearly demonstrate that the sequence of shock wave administration can have a dramatic effect on the severity of renal injury. Step-wise power ramping can be a useful strategy to reduce injury but appears only to be effective when the initial dose of shock waves is followed by a brief pause in treatment ('pause-protection'). Slowing the firing rate of the lithotripter to 60 shock waves/min or slower is equally effective in reducing renal injury and offers the advantage of a significant improvement in stone breakage outcomes documented by multiple prospective clinical trials. The poor quality and variability of acoustic coupling is an additional challenge that needs to be addressed. This aspect of SWL is largely taken for granted but is a factor that is critical to successful outcomes.

## Acknowledgments

This work was supported by NIH grant DK 43881 and is a research and education initiative of the International Kidney Stone Institute.

## References

References and recommended reading

Papers of particular interest, published within the annual period of review, have been highlighted as:

- of special interest
- of outstanding interest

Additional references related to this topic can also be found in the Current World Literature section in this issue (p. 634).

1. Lingeman, JE.; Matlaga, B.; Evan, AP. Surgical management of urinary lithiasis. In: Walsh, PC.; Retik, AB.; Vaughan, ED.; Wein, J., editors. *Campbell's urology*. W.B. Saunders Company; Philadelphia: 2006. p. 1431-1507. Chapter 44
2. Evan, AP.; McAteer, JA. Q-Effects of shock wave lithotripsy. In: Coe, FL.; Favus, MJ.; Pak, CYC.; Parks, JH.; Preminger, GM., editors. *Kidney stones: medical and surgical management*. Lippincott-Raven; Philadelphia: 1006. p. 549-570.
3. McAteer JA, Evan AP. The acute and long-term adverse effects of shock wave lithotripsy. *Seminars Nephrol*. 2008; 28:200–213.
4. McAteer, JA.; Evan, AP.; Willis, LR., et al. Shock wave injury to the kidney in SWL: review and perspective. In: Evan, AP.; Lingeman, JE.; Williams, JC., editors. *Renal stone disease: Proceedings of the 1st International Urolithiasis Research Symposium; American Institute of Physics Conference Proceedings*; Melville, New York. 2007. p. 287-301.
5. Evan, AP.; Willis, LR. Extracorporeal shock wave lithotripsy: complications. In: Smith, AD.; Badlani, GH.; Bagley, DH.; Clayman, RV.; Docimo, SG.; Jordan, GH.; Kavoussi, LR.; Lee, BR.; Lingeman, JE.; Preminger, GM.; Segura, JW., editors. *Smith's textbook on endourology*. BC Decker, Inc.; Hamilton, Ontario, Canada: 2007. p. 353-365.
6. Connors BA, Evan AP, Blomgren PM, et al. Reducing shock number dramatically decreases lesion size in a juvenile kidney model. *J Endourol*. 2006; 20:607–611. [PubMed: 16999608]
7. Willis LR, Evan AP, Connors BA, et al. Shock-wave lithotripsy: dose-related effects on renal structure, hemodynamics and tubular function. *J Endourol*. 2005; 19:90–101. [PubMed: 15735392]
8. Connors BA, Evan AP, Willis LR, et al. The effect of discharge voltage on renal injury and impairment caused by lithotripsy in the pig. *J Am Soc Nephrol*. 2000; 11:310–318. [PubMed: 10665938]

9. Evan AP, Willis LR, Lingeman JE, McAteer JA. Renal trauma and the risk of long-term complications in shock wave lithotripsy. *Nephron*. 1998; 78:1–8. [PubMed: 9453396]
10. Newman R, Hackett R, Senior D, et al. Pathological effects of ESWL on canine renal tissue. *Urology*. 1987; 29:194–200. [PubMed: 3811098]
11. Morris JA, Husmann DA, Wilson WT, Preminger GM. Temporal effects of shock wave lithotripsy. *J Urol*. 1991; 145:881–883. [PubMed: 1672386]
12. Parks JH, Worcester EM, Coe FL, et al. Clinical implications of abundant calcium phosphate in routinely analyzed kidney stones. *Kidney Int*. 2004; 66:777–785. [PubMed: 15253733]
13. Krambeck AE, Gettman MT, Rohlinger AL, et al. Diabetes mellitus and hypertension associated with shock wave lithotripsy of renal and proximal ureteral stones at 19 years of followup. *J Urol*. 2006; 175:1742–1747. [PubMed: 16600747]
14. Willis LR, Evan AP, Connors BA, et al. Prevention of lithotripsy-induced renal injury by pretreating kidneys with low-energy shock waves. *J Am Soc Nephrol*. 2006; 17:663–673. [PubMed: 16452495]
- 15•. Handa RK, Bailey MR, Paun M, et al. Pretreatment with low-energy shock waves induces vasoconstriction during standard SWL: a treatment protocol known to reduce lithotripsy-induced renal injury. *BJU Int*. 2009; 103:1270–1274. [PubMed: 19154458] This article provides a suggestion that vasoconstriction plays a role in the SWL ‘protection protocol’.
16. Mobley TB, Myers DA, Grine WB, et al. Low energy lithotripsy with the lithostar: treatment results with 19,962 renal and ureteral calculi. *J Urol*. 1993; 149:1419–1424. [PubMed: 8501779]
- 17•. Connors BA, Evan AP, Blomgren PM, et al. Effect of initial shock wave voltage on shock wave lithotripsy-induced lesion size during step-wise voltage ramping. *BJU Int*. (in press). A demonstration that a 3–4 min pause following delivery of an initial dose of shock waves is critical to the ‘protection protocol’.
18. Zhou Y, Cocks FH, Preminger GM, Zhong P. The effect of treatment strategy on stone comminution efficiency in shock wave lithotripsy. *J Urol*. 2004; 172:349–354. [PubMed: 15201809]
19. Maloney ME, Marguet CG, Zhou Y, et al. Progressive increase of lithotripter output produces better in-vivo stone comminution. *J Endourol*. 2006; 20:603–606. [PubMed: 16999607]
20. Demirci D, Mustafa S, Yalcin E, et al. Comparison of conventional and step-wise shockwave lithotripsy in management of urinary calculi. *J Endourol*. 2007; 21:1407–1410. [PubMed: 18044996]
- 21••. Evan AP, McAteer JA, Connors BA, et al. Renal injury in SWL is significantly reduced by slowing the rate of shock wave delivery. *BJU Int*. 2007; 100:624–627. [PubMed: 17550415] This report of research with the pig model shows that slowing the shock wave rate to 30 shock waves/min virtually eliminates the acute renal injury seen with a dose of 2000 shock waves. In a follow-up study recently submitted for publication, it has been observed that treatment at 60 shock waves/min is also protective.
22. Evan AP, McAteer JA, Connors BA, et al. Independent assessment of a wide-focus, low-pressure electromagnetic lithotripter: absence of renal bioeffects in the pig. *BJU Int*. 2007; 101:382–388. [PubMed: 17922871]
23. Eisenmenger W, Du XX, Tang C, et al. The first clinical results of ‘wide focus and low pressure’ ESWL. *Ultrasound Med Biol*. 2002; 28:769–774. [PubMed: 12113789]
24. Pace KT, Ghiculete D, Harju M, Honey RJD’A. Shock wave lithotripsy at 60 or 120 shocks per minute: a randomized, double-blind trial. *J Urol*. 2005; 174:595–599. [PubMed: 16006908]
25. Madbouly K, El-Tiraifi AM, Seida M, et al. Slow versus fast shock wave lithotripsy rate for urolithiasis: a prospective randomized study. *J Urol*. 2005; 173:127–130. [PubMed: 15592053]
26. Yilmaz E, Batislam E, Basar M, et al. Optimal frequency in extracorporeal shock wave lithotripsy: prospective randomized study. *Urology*. 2005; 66:1160–1164. [PubMed: 16360432]
- 27•. Semins MJ, Trock BJ, Matlaga BR. The effect of shock wave rate on the outcome of shock wave lithotripsy: a metaanalysis. *J Urol*. 2008; 179:194–197. [PubMed: 18001796] This critical metaanalysis supports the findings of several prospective clinical trials (see [24–26]) reporting improved stone breakage outcomes when the firing rate of the lithotripter is slowed to 60 shock waves/min.

28. Pishchalnikov YA, Neucks JS, VonDerHaar RJ, et al. Air pockets trapped during routine coupling in dry-head lithotripsy can significantly reduce the delivery of shock wave energy. *J Urol.* 2006; 176:2706–2710. [PubMed: 17085200]
- 29••. Neucks JS, Pishchalnikov YA, Zancanaro AJ, et al. Improved acoustic coupling for shock wave lithotripsy. *Urol Res.* 2008; 36:61–66. [PubMed: 18172634] In-vitro survey of protocols commonly used to couple the lithotripter to the patient, showing that typical procedures of applying lithotripsy gel are ineffective and highly variable and that simple techniques can be used to significantly improve the quality of coupling.